

Instructions



YBT250F Software Upgrade for NetTek[®] YBT250 Field Transmitter & Interference Tester

075-0721-04

This document applies to firmware version 1.4XX
and above.

www.tektronix.com



075072104

Copyright © Tektronix. All rights reserved. Licensed software products are owned by Tektronix or its subsidiaries or suppliers, and are protected by national copyright laws and international treaty provisions.

Tektronix products are covered by U.S. and foreign patents, issued and pending. Information in this publication supercedes that in all previously published material. Specifications and price change privileges reserved.

TEKTRONIX and TEK are registered trademarks of Tektronix, Inc.

Contacting Tektronix

Tektronix, Inc.
14200 SW Karl Braun Drive
P.O. Box 500
Beaverton, OR 97077
USA

For product information, sales, service, and technical support:

- In North America, call 1-800-833-9200.
- Worldwide, visit www.tektronix.com to find contacts in your area.

YBT250 Software Upgrade Kit

This document explains how to upgrade a YBT250 to Version 1.4XX or above software.

Products

YBT250 all

Minimum Tool and Equipment List

Required tools and equipment	Part number
Windows-based PC	--- ---
NetTek Analyzer Platform with YBT250 module installed	--- ---
RS-232 Serial Cable (NetTek Analyzer Platform standard accessory)	012-1379-XX

Kit Parts List

Quantity	Part number	Description
1 ea	075-0721-XX	KIT INSTRUCTION: YBT250F UPGRADE
1 ea	020-2464-XX	MANUAL, TECH: USER YBT250, W/ CD-ROM
1 ea		UPGRADE LABEL SHEET (NEW SOFT KEY, ADDITIONAL OPTIONS AND APPLICATIONS)

Time Required

The software upgrade process requires about 30 minutes to complete.

Installation Instructions

Refer to the YBT250 user manual or Online help for option information.

On-screen installation instructions will guide you through upgrading or reinstalling the YBT250 software:

1. Insert the NetTek YBT250 and RF Scout Installation SW CD-ROM (Tektronix part number, 063-3476-XX) into your PC and follow the on-screen instructions.
2. If no instructions appear after you insert the CD-ROM, use Windows Explorer to display the contents of the CD-ROM and double-click the **setup.exe** file to begin the installation. Follow the on-screen instructions.
3. Remove the NEW SOFT KEY label (that contains the soft key number/letter sequence) from the label sheet and place it onto the SOFT KEY area of the existing Soft Key/Options and Applications label on the back of the module. See Figure 1.




Figure 1: YBT250 new label positions.

4. Remove the NEW (ADDITIONAL) OPTIONS label from the label sheet and place it anywhere in the indicated area below or to the right of the existing Soft Key/Options and Applications label (see Figure 1). Do not place the NEW (ADDITIONAL) OPTIONS label anywhere on the existing label; you do not want to cover any options or applications that are listed on the main Soft Key/Options and Applications label. Do not let the label cover any screw heads.

5. Discard the label sheet.

■ End of document ■

