

NO: SAMM 605

Page: 1 of 7

LABORATORY LOCATION:  
(PERMANENT LABORATORY)TEKTRONIX MALAYSIA CALIBRATION  
LABORATORYTEKTRONIX INSTRUMENTS MALAYSIA SDN. BHD.  
UNIT 1-15-6 SUNTECH @ PENANG CYBERCITY  
LINTANG MAYANG PASIR 3  
11950 BAYAN BARU  
PENANG, MALAYSIA

FIELD OF CALIBRATION: ELECTRICAL

This laboratory has demonstrated its technical competence to operate in accordance with MS ISO/IEC 17025:2017 (ISO/IEC 17025:2017).

This laboratory's fulfillment of the requirements of ISO/IEC 17025 means the laboratory meets both the technical competence requirements and management system requirements that are necessary for it to consistently deliver technically valid test results and calibrations. The management system requirements in ISO/IEC 17025 are written in language relevant to laboratory operations and operate generally in accordance with the principles of ISO 9001 (see Joint ISO-ILAC-IAF Communiqué dated April 2017).

\* The expanded uncertainties are based on an estimated confidence probability of approximately 95% and have a coverage factor of  $k=2$  unless stated otherwise.

SCOPE OF CALIBRATION: ELECTRICAL

PERMANENT LABORATORY AND SITE CATEGORY I

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty( $\pm$ )*	Remarks
DC Voltage Source	0 mV to 100 mV 100 mV to 1 V 1 V to 10 V 10 V to 100 V 100 V to 1000 V	5.5 $\mu$ V/V + 0.78 $\mu$ V 4.7 $\mu$ V/V + 0.78 $\mu$ V 4.7 $\mu$ V/V + 1.6 $\mu$ V 6.2 $\mu$ V/V + 24 $\mu$ V 6.2 $\mu$ V/V + 78 $\mu$ V + $9.3 \times 10^{-6}$ * $V_{in}^2$ $\mu$ V/V	Measurement using Agilent 3458A
	0 mV to 200 mV 200 mV to 2 V 2 V to 20 V 20 V to 200 V 200 V to 1000 V	4.5 $\mu$ V/V + 0.1 $\mu$ V 3 $\mu$ V/V + 0.4 $\mu$ V 3 $\mu$ V/V + 4 $\mu$ V 4.5 $\mu$ V/V + 40 $\mu$ V 4.5 $\mu$ V/V + 1000 $\mu$ V	Measurement using Fluke 8508A
DC Current Source	0 nA to 100 nA 0.1 $\mu$ A to 1 $\mu$ A 1 $\mu$ A to 10 $\mu$ A 10 $\mu$ A to 100 $\mu$ A 100 $\mu$ A to 1 mA 1 mA to 10 mA 10 mA to 100 mA 100 mA to 1 A	27 $\mu$ A/A + 0.031 nA 20 $\mu$ A/A + 0.000031 $\mu$ A 20 $\mu$ A/A + 0.000078 $\mu$ A 20 $\mu$ A/A + 0.00062 $\mu$ A 20 $\mu$ A/A + 0.0039 $\mu$ A 20 $\mu$ A/A + 0.039 $\mu$ A 31 $\mu$ A/A + 0.39 $\mu$ A 90 $\mu$ A/A + 7.8 $\mu$ A	Measurement using Agilent 3458A

NO: SAMM 605

Page: 2 of 7

## SCOPE OF CALIBRATION: ELECTRICAL

## PERMANENT LABORATORY AND SITE CATEGORY I

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty( $\pm$ )*	Remarks
DC Current Source (continue)	0 $\mu$ A to 200 $\mu$ A 200 $\mu$ A to 2mA 2 mA to 20 mA 20 mA to 200 mA 200 mA to 2 A 2 A to 20 A	12 $\mu$ A/A + 0.00042 $\mu$ A 12 $\mu$ A/A + 0.00042 $\mu$ A 13 $\mu$ A/A + 0.042 $\mu$ A 36 $\mu$ A/A + 0.042 $\mu$ A 170 $\mu$ A/A + 0.17 $\mu$ A 380 $\mu$ A/A + 420 $\mu$ A	Measurement using Fluke 8508A
Resistance Source	0 $\Omega$ to 10 $\Omega$ 10 $\Omega$ to 100 $\Omega$ 100 $\Omega$ to 1 k $\Omega$ 1 k $\Omega$ to 10 k $\Omega$ 10 k $\Omega$ to 100 k $\Omega$ 100 k $\Omega$ to 1 M $\Omega$ 1 M $\Omega$ to 10 M $\Omega$ 10 M $\Omega$ to 100 M $\Omega$	14 $\mu\Omega/\Omega$ + 39 $\mu\Omega$ 12 $\mu\Omega/\Omega$ + 390 $\mu\Omega$ 10 $\mu\Omega/\Omega$ + 390 $\mu\Omega$ 10 $\mu\Omega/\Omega$ + 3.9 m $\Omega$ 10 $\mu\Omega/\Omega$ + 39 m $\Omega$ 14 $\mu\Omega/\Omega$ + 1.6 $\Omega$ 41 $\mu\Omega/\Omega$ + 78 $\Omega$ 390 $\mu\Omega/\Omega$ + 780 $\Omega$	Measurement using Agilent 3458A
	0 $\Omega$ to 2 $\Omega$ 2 $\Omega$ to 20 $\Omega$ 20 $\Omega$ to 200 $\Omega$ 200 $\Omega$ to 2 k $\Omega$ 2 k $\Omega$ to 20 k $\Omega$ 20 k $\Omega$ to 200 k $\Omega$ 200 k $\Omega$ to 2 M $\Omega$ 2 M $\Omega$ to 20 M $\Omega$ 20 M $\Omega$ to 200 M $\Omega$	15 $\mu\Omega/\Omega$ + 4 $\mu\Omega$ 9 $\mu\Omega/\Omega$ + 14 $\mu\Omega$ 7.5 $\mu\Omega/\Omega$ + 50 $\mu\Omega$ 7.5 $\mu\Omega/\Omega$ + 500 $\mu\Omega$ 7.5 $\mu\Omega/\Omega$ + 5 m $\Omega$ 7.5 $\mu\Omega/\Omega$ + 50 m $\Omega$ 8.5 $\mu\Omega/\Omega$ + 1 $\Omega$ 15 $\mu\Omega/\Omega$ + 100 $\mu\Omega$ 60 $\mu\Omega/\Omega$ + 10 k $\Omega$	Measurement using Fluke 8508A
AC Voltage Source	0 V to 1000 V (See Matrix A)	See Matrix A	Measurement using Agilent 3458A
	0 V to 1000 V (See Matrix B)	See Matrix B	Measurement using Fluke 8508A
AC Current Source	0 $\mu$ A to 1 A (See Matrix C)	See Matrix C	Measurement using Agilent 3458A
	0 $\mu$ A to 1 A (See Matrix D)	See Matrix D	Measurement using Fluke 8508A

Scan this QR Code or visit [www.ism.gov.my/cab-directories](http://www.ism.gov.my/cab-directories) for the current scope of accreditation

NO: SAMM 605

Page: 3 of 7

**SCOPE OF CALIBRATION: ELECTRICAL****PERMANENT LABORATORY AND SITE CATEGORY I****Matrix A – AC Voltage Source**

Range	Frequency								
	Hz			kHz			MHz		
	10 to 20	20 to 40	40 to 100	0.1 to 20	20 to 50	50 to 100	0.1 to 0.25	0.25 to 0.5	0.5 to 1
0 mV to 10 mV	3.1 + 0.025	1.2 + 0.02	0.47 + 0.02	0.16 + 0.02	1.2 + 0.02	5.5 + 0.028	31 + 0.055	-	-
10 mV to 100 mV	3.1 + 0.016	1.2 + 0.016	0.47 + 0.0078	0.16 + 0.0078	1.2 + 0.031	4.7 + 0.062	16 + 0.39	-	-
100 mV to 1 V	-	1.2 + 0.16	0.47 + 0.078	0.16 + 0.078	1.2 + 0.31	4.7 + 0.62	16 + 3.9	24 + 4.7	39 + 16
1 V to 10 V	-	1.2 + 1.6	0.47 + 0.78	0.16 + 0.78	1.2 + 3.1	4.7 + 6.2	16 + 39	24 + 47	39 + 160
10 V to 100 V	-	1.2 + 1.6	0.47 + 7.8	0.24 + 7.8	1.2 + 31	4.7 + 62	16 + 390	-	-
100 V to 1000 V	-	-	0.63 + 160	0.47 + 160	1.2 + 310	4.7 + 1600	-	-	-

The expanded uncertainties given in this table are expressed in mV/V + mV

**Matrix B – AC Voltage Source**

Range	Frequency							
	Hz		kHz				MHz	
	10 to 40	40 to 100	0.1 to 2	2 to 10	10 to 30	30 to 100	100 to 300	0.3 to 1
0 to 200 mV	0.13 + 0.004	0.11 + 0.004	0.11 + 0.002	0.11 + 0.004	0.31 + 0.008	0.71 + 0.02	-	-
200 mV to 2 V	0.11 + 0.02	0.085 + 0.02	0.065 + 0.02	0.085 + 0.02	0.21 + 0.04	0.51 + 0.2	-	-
2 V to 20 V	0.11 + 0.2	0.085 + 0.2	0.065 + 0.2	0.085 + 0.2	0.21 + 0.4	0.51 + 2	3 + 20	10 + 200
20 V to 200 V	0.11 + 2	0.085 + 2	0.065 + 2	0.085 + 2	0.21 + 4	0.51 + 20	-	-

Range	Frequency	
	kHz	
	0.04 to 10	10 to 30
200 V to 300 V	0.095 + 40	
300 to 1000 V	$[0.095 + (0.0000004 \cdot (V_{in} - 300)^2)] + 40$	$[0.21 + (0.0000024 \cdot (V_{in} - 300)^2)] + 80$

The expanded uncertainties given in this table are expressed in mV/V + mV

NO: SAMM 605

Page: 4 of 7

**SCOPE OF CALIBRATION: ELECTRICAL****PERMANENT LABORATORY AND SITE CATEGORY I****Matrix C – AC Current Source**

Range	Frequency				
	Hz			kHz	
	10 to 20	20 to 45	45 to 100	0.1 to 5	5 to 20
0 $\mu$ A to 100 $\mu$ A	3.1 + 0.024	1.2 + 0.024	0.47 + 0.024	0.47 + 0.024	-
100 $\mu$ A to 1 mA	3.1 + 0.16	1.2 + 0.16	0.47 + 0.16	0.24 + 0.16	0.47 + 0.16
1 mA to 10 mA	3.1 + 1.6	1.2 + 1.6	0.47 + 1.6	0.24 + 1.6	0.47 + 1.6
10 mA to 100 mA	3.1 + 16	1.2 + 16	0.47 + 16	0.24 + 16	0.47 + 16
100 mA to 1 A	3.1 + 160	1.3 + 160	0.63 + 160	0.78 + 160	2.4 + 160

**Matrix D – AC Current Source**

Range	Frequency
	kHz
	0.01 to 10
0 $\mu$ A to 200 $\mu$ A	0.48 + 0.02
200 $\mu$ A to 2 mA	0.28 + 0.2
2 mA to 20 mA	0.28 + 2
20 mA to 200 mA	0.25 + 20

Range	Frequency	
	kHz	
	0.01 to 2	2 to 10
200 mA to 2 A	0.6 + 0.2	0.71 + 0.2
2 A to 20 A	0.8 + 2	2.5 + 2

The expanded uncertainties given in this table are expressed in mA/A +  $\mu$ A

NO: SAMM 605

Page: 5 of 7

## SCOPE OF CALIBRATION: ELECTRICAL

## PERMANENT LABORATORY AND SITE CATEGORY I

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty( $\pm$ )*	Remarks
DC Voltage Measurement	0 mV to 220 mV 220 mV to 2.2 V 2.2 V to 11 V 11 V to 22 V 22 V to 220 V 220 V to 1100 V	7.5 $\mu$ V/V + 0.4 $\mu$ V 5 $\mu$ V/V + 0.7 $\mu$ V 3.5 $\mu$ V/V + 2.5 $\mu$ V 3.5 $\mu$ V/V + 4 $\mu$ V 5 $\mu$ V/V + 40 $\mu$ V 6.5 $\mu$ V/V + 400 $\mu$ V	Measurement using Fluke 5720A
DC Current Measurement	0 $\mu$ A to 220 $\mu$ A 220 $\mu$ A to 2.2 mA 2.2 mA to 22 mA 22 mA to 100 mA 100 mA to 220 mA  220 mA to 1 A 1 A to 2.2 A	40 $\mu$ A/A + 6 nA 35 $\mu$ A/A + 7 nA 35 $\mu$ A/A + 40 nA 45 $\mu$ A/A + 0.7 $\mu$ A (45 $\mu$ A/A + 200 $\mu$ A/A * I <sup>2</sup> ) + 0.7 $\mu$ A 80 $\mu$ A/A + 12 $\mu$ A (80 $\mu$ A/A + 10 $\mu$ A/A * I <sup>2</sup> ) + 12 $\mu$ A	Measurement using Fluke 5720A
	2.2 A to 11 A	360 $\mu$ A/A + 480 $\mu$ A	Measurement using Fluke 5725A
Resistance Measurement (Specific Value)	1 $\Omega$ 1.9 $\Omega$ 10 $\Omega$ 19 $\Omega$ 100 $\Omega$ 190 $\Omega$ 1 k $\Omega$ 1.9 k $\Omega$ 10 k $\Omega$ 19 k $\Omega$ 100 k $\Omega$ 190 k $\Omega$ 1 M $\Omega$ 1.9 M $\Omega$ 10 M $\Omega$ 19 M $\Omega$ 100 M $\Omega$	95 $\mu$ $\Omega$ / $\Omega$ 95 $\mu$ $\Omega$ / $\Omega$ 23 $\mu$ $\Omega$ / $\Omega$ 23 $\mu$ $\Omega$ / $\Omega$ 10 $\mu$ $\Omega$ / $\Omega$ 10 $\mu$ $\Omega$ / $\Omega$ 8.5 $\mu$ $\Omega$ / $\Omega$ 8.5 $\mu$ $\Omega$ / $\Omega$ 8.5 $\mu$ $\Omega$ / $\Omega$ 8.5 $\mu$ $\Omega$ / $\Omega$ 11 $\mu$ $\Omega$ / $\Omega$ 11 $\mu$ $\Omega$ / $\Omega$ 20 $\mu$ $\Omega$ / $\Omega$ 21 $\mu$ $\Omega$ / $\Omega$ 40 $\mu$ $\Omega$ / $\Omega$ 47 $\mu$ $\Omega$ / $\Omega$ 100 $\mu$ $\Omega$ / $\Omega$	Measurement using Fluke 5720A
AC Voltage Measurement	2.2 mV to 220 V (See Matrix E)	See Matrix E	Measurement using Fluke 5720A
	220 V to 1100 V (See Matrix E)	See Matrix E	Measurement using Fluke 5725A

NO: SAMM 605

Page: 6 of 7

**SCOPE OF CALIBRATION: ELECTRICAL****PERMANENT LABORATORY AND SITE CATEGORY I**

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty( $\pm$ )*	Remarks
AC Current Measurement	0 $\mu$ A to 2.2 A (see Matrix F)	see Matrix F	Measurement using Fluke 5720A
	2.2 A to 11 A (see Matrix F)	see Matrix F	Measurement using Fluke 5725A

**Matrix E – AC Voltage Measurement**

Range	Frequency							
	Hz		kHz			MHz		
	10 to 20	20 to 40	0.04 to 20	20 to 50	50 to 100	0.1 to 0.3	0.3 to 0.5	0.5 to 1
2.2 mV to 22 mV	0.24 + 0.004	0.09 + 0.004	0.08 + 0.004	0.2 + 0.004	0.5 + 0.005	1.1 + 0.01	-	-
22mV to 220 mV	0.24 + 0.012	0.09 + 0.007	0.08 + 0.007	0.2 + 0.007	0.46 + 0.017	0.9 + 0.02	-	-
220 mV to 2.2 V	0.24 + 0.04	0.09 + 0.015	0.045 + 0.008	0.075 + 0.01	0.11 + 0.03	0.42 + 0.08	1 + 0.2	1.7 + 0.3
2.2 V to 22 V	0.24 + 0.4	0.09 + 0.15	0.045 + 0.05	0.075 + 0.1	0.1 + 0.2	0.28 + 0.6	1 + 2	1.5 + 3.2
22 V to 220 V	0.24 + 4	0.09 + 1.5	0.052 + 0.6	0.08 + 1	0.15 + 2.5	-	-	-

Range	Frequency				
	kHz				
	0.04 to 1	1 to 20	20 to 30	30 to 50	50 to 100
220 V to 750 V	-	-	-	0.6 +11	2.3 + 45
750 V to 1100 V	0.09 + 4	0.17 + 6	0.6 + 11	-	-

The expanded uncertainties given in this table are expressed in mV/V + mV

NO: SAMM 605

Page: 7 of 7

**SCOPE OF CALIBRATION: ELECTRICAL****PERMANENT LABORATORY AND SITE CATEGORY I****Matrix F – AC Current Measurement**

Range	Frequency				
	Hz		kHz		
	10 to 20	20 to 40	0.04 to 1	1 to 5	5 to 10
0 $\mu$ A to 220 $\mu$ A	0.25 + 0.016	0.16 + 0.01	0.12 + 0.008	0.28 + 0.012	1.1 + 0.065
220 $\mu$ A to 2.2 mA	0.25 + 0.04	0.16 + 0.035	0.12 + 0.035	0.2 + 0.11	1.1 + 0.65
2.2 mA to 22 mA	0.25 + 0.4	0.16 + 0.35	0.12 + 0.35	0.2 + 0.55	1.1 + 5
22 mA to 220 mA	0.25 + 4	0.16 + 3.5	0.12 + 2.5	0.2 + 3.5	1.1 + 10

Range	Frequency		
	kHz		
	0.02 to 1	1 to 5	5 to 10
220 mA to 2.2 A	0.26 + 35	0.45 + 80	7 + 160

Range	Frequency		
	kHz		
	0.04 to 1	1 to 5	5 to 10
2.2 A to 11 A	0.46 + 170	0.95 + 380	3.6 + 750

The expanded uncertainties given in this table are expressed in mA/A +  $\mu$ A**Signatories:**

1. Ching Gin Kong
2. Tan Chin Seong