



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

TEKTRONIX INDIA PVT LTD, SALARPURIA PREMIA, SURVEY NO 16,
KADUBEESANAHALLI, VARTHUR HOBLI, BENGALURU, BENGALURU URBAN,
KARNATAKA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-2130

Page No

1 of 50

Validity

22/05/2020 to 21/05/2022

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrum	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
Permanent Facility					
1	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Measure)	AC Active Power @50 Hz 60 V to 240 V/ 100mA to 20 A / 0.2 PF	direct method	2.4 W to 960 W	0.6 % to 0.24 %
2	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Measure)	AC Active Power @50 Hz 60 V to 240 V/ 100mA to 20 A / 0.5 PF	direct method	6 W to 2400 W	0.6 % to 0.24 %
3	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Measure)	AC Active Power @50 Hz 60 V to 240 V/ 100mA to 20 A / 0.8PF	direct method	9.6 W to 3800 W	0.6 % to 0.24 %
4	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Measure)	AC Active Power @50 Hz 60 V to 240 V/ 100mA to 20 A / UPF	Power Analyser PA 4400	6 W to 4800 W	2.05 % to 1.25 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

TEKTRONIX INDIA PVT LTD, SALARPURIA PREMIA, SURVEY NO 16,
KADUBEESANAHALLI, VARTHUR HOBLI, BENGALURU, BENGALURU URBAN,
KARNATAKA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-2130

Page No

2 of 50

Validity

22/05/2020 to 21/05/2022

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrum	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
5	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Measure)	AC current, Fluke 8508A	direct method	1 A to 20 A(10 Hz- 5 kHz)	0.02 % to 0.3 %
6	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Measure)	AC current, Fluke 8508A	direct method	1 mA(300 Hz- 10 kHz) to 10 mA(10 Hz-10 kHz)	0.05 % to 0.023 %
7	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Measure)	AC current, Fluke 8508A	direct method	10 mA(300 Hz- 10 kHz) to 1 A(10 Hz- 10 kHz)	0.023 % to 0.056 %
8	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Measure)	AC current, Fluke 8508A	direct method	100 uA to 1 mA (300 Hz- 1kHz)	0.05%
9	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Measure)	AC Voltage, Fluke 5790A	direct method	1 mV(10 kHz-100 kHz) to 1 V(10 kHz-100 kHz)	0.42 % to 0.016 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

TEKTRONIX INDIA PVT LTD, SALARPURIA PREMIA, SURVEY NO 16, KADUBEESANAHALLI, VARTHUR HOBLI, BENGALURU, BENGALURU URBAN, KARNATAKA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-2130

Page No

3 of 50

Validity

22/05/2020 to 21/05/2022

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrum	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
10	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Measure)	AC Voltage, Fluke 5790A	direct method	1 V to 100V (10 Hz-40 Hz)	0.038 % to 0.01 %
11	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Measure)	AC Voltage, Fluke 5790A	direct method	10 mV to 1 V(10 Hz-40 Hz)	0.038%
12	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Measure)	AC Voltage, Fluke 5790A	Direct Method	10 mV to 1V (40 Hz-20 kHz)	0.023 % to 0.008 %
13	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Measure)	AC Voltage, Fluke 5790A	direct method	10 mV to 20 V(100 kHz- 1MHz)	0.19 % to 0.14 %
14	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Measure)	AC Voltage, Fluke 5790A	Direct method	10 V(20 kHz-100 kHz) to 500 V(10 kHz-100 kHz)	0.009 % to 0.03 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

TEKTRONIX INDIA PVT LTD, SALARPURIA PREMIA, SURVEY NO 16, KADUBEESANAHALLI, VARTHUR HOBLI, BENGALURU, BENGALURU URBAN, KARNATAKA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-2130

Page No

4 of 50

Validity

22/05/2020 to 21/05/2022

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured / Instrum	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
15	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Measure)	AC Voltage, Fluke 5790A	direct method	1mV to 10mV (10 Hz-40 Hz)	0.44 % to 0.038 %
16	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Measure)	AC Voltage, Fluke 5790A	Direct Method	1V to 1000 V(40 Hz-20 kHz)	0.008 % to 0.015 %
17	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Measure)	AC Voltage, Fluke 5790A	Direct method	1V to 10 V (10 kHz-100 kHz)	0.016 % to 0.009%
18	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Measure)	AC Voltage, Fluke 5790A	direct method	1 mV to 10 mV(100 kHz- 1MHz)	0.44 % to 0.19 %
19	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Measure)	AC Voltage, Fluke 8508A 5790A	direct method	1mV to 10 mV(40 Hz- 20 KHz)	0.17 % to 0.023 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

TEKTRONIX INDIA PVT LTD, SALARPURIA PREMIA, SURVEY NO 16, KADUBEESANAHALLI, VARTHUR HOBLI, BENGALURU, BENGALURU URBAN, KARNATAKA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-2130

Page No

5 of 50

Validity

22/05/2020 to 21/05/2022

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured / Instrum	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
20	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	AC current, Fluke 5522A	direct method,	10 A to 20 A(40Hz-1kHz)	0.059 % to 0.20 %
21	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	AC Current, Fluke 5522A with 5500A coil	Direct method	20A A to 1000 (@50Hz) A	0.35 % to 0.77 %
22	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	AC Current, Fluke 5720A	direct method	100uA to 1mA (40 Hz-1kHz)	0.109 % to 0.024 %
23	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	AC Current, Fluke 5730A	Direct method	100 mA to 1 A(40Hz-1kHz)	0.03 % to 0.04 %
24	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	AC Current, Fluke 5730A	direct method	1mA to 100 mA (40Hz - 1 kHz)	0.024 % to 0.026 %
25	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	AC Current, Fluke 5730A	Direct method	200uA to 100mA (10Hz-40Hz)	0.11 % to 0.045 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

TEKTRONIX INDIA PVT LTD, SALARPURIA PREMIA, SURVEY NO 16, KADUBEESANAHALLI, VARTHUR HOBLI, BENGALURU, BENGALURU URBAN, KARNATAKA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-2130

Page No

6 of 50

Validity

22/05/2020 to 21/05/2022

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured / Instrum	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
26	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	AC Current, Fluke 5730A & 5725A	direct method	20mA to 10 A(5kHz-10kHz)	0.03 % to 0.42 %
27	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	AC Current, Fluke 5730A & 5725A	direct method	20 mA to 10 A(1kHz-10 kHz)	0.056 % to 0.11 %
28	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	AC Current, Fluke 5730A, 5725A	Direct method, using AC shunts	1A to 10 A (40Hz-1kHz)	0.05 % to 0.059 %
29	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	AC Voltage, Fluke 5730A	Direct method	1 mV to 1 V(10 Hz-40 Hz)	0.49 % to 0.052 %
30	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	AC Voltage, Fluke 5730A	direct method	1mV to 2 mV (40 Hz - 1 kHz)	0.6 % to 0.25 %
31	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	AC Voltage, Fluke 5730A	Direct method	20 mV to 20V(10 Hz-40 Hz)	0.05 % to 0.03 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

TEKTRONIX INDIA PVT LTD, SALARPURIA PREMIA, SURVEY NO 16, KADUBEESANAHALLI, VARTHUR HOBLI, BENGALURU, BENGALURU URBAN, KARNATAKA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-2130

Page No

7 of 50

Validity

22/05/2020 to 21/05/2022

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured / Instrum	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
32	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	AC Voltage, Fluke 5730A	direct method	20 V to 1000 V (40 Hz - 1 kHz)	0.0065 % to 0.0090 %
33	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	AC Voltage, Fluke 5730A	direct method	20 V to 200V (10 Hz-40 Hz)	0.03 % to 0.025 %
34	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	AC Voltage, Fluke 5730A	direct method	200 mV to 2V (40 Hz - 1 kHz)	0.015 % to 0.006 %
35	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	AC Voltage, Fluke 5730A	direct method	200mV to 20V (100 kHz-1 MHz)	0.46 % to 0.25 %
36	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	AC Voltage, Fluke 5730A	direct method	20mV to 200 mV (100 kHz-1 MHz)	0.46 % to 0.27 %
37	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	AC Voltage, Fluke 5730A	Direct method	2mV to 20 mV(40 Hz - 1 kHz)	0.25 % to 0.03 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

TEKTRONIX INDIA PVT LTD, SALARPURIA PREMIA, SURVEY NO 16,
KADUBEESANAHALLI, VARTHUR HOBLI, BENGALURU, BENGALURU URBAN,
KARNATAKA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-2130

Page No

8 of 50

Validity

22/05/2020 to 21/05/2022

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured / Instrum	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
38	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	AC Voltage, Fluke 5730A	Direct method	2mV to 20mV(10 Hz-40 Hz)	0.26 % to 0.05 %
39	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	AC Voltage, Fluke 5730A	direct method	2mV to 2 V (1 kHz - 100 kHz)	0.45 % to 0.015 %
40	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	AC Voltage, Fluke 5730A	direct method	2mV (100 kHz - 1 MHz) to 20 mV (100 kHz - 1MHz)	1.75 % to 0.46 %
41	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	AC Voltage, Fluke 5730A	direct method	2V to 20V (40 Hz - 1 kHz)	0.006 % to 0.0065 %
42	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	AC Voltage, Fluke 5730A, 5725A	direct method	100 V to 500V (1 kHz - 100 kHz)	0.02 % to 0.27 %
43	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	AC Voltage,Fluke 5730A	direct method	20 mV to 200mV (40 Hz - 1 kHz)	0.03 % to 0.015 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

TEKTRONIX INDIA PVT LTD, SALARPURIA PREMIA, SURVEY NO 16, KADUBEESANAHALLI, VARTHUR HOBLI, BENGALURU, BENGALURU URBAN, KARNATAKA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-2130

Page No

9 of 50

Validity

22/05/2020 to 21/05/2022

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured / Instrum	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
44	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	AC Voltage,Fluke 5730A	direct method	2V to 100V (1 kHz - 100 kHz)	0.015 % to 0.02 %
45	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	Capacitance, Capacitor Standard GR 1417	Direct Method	100mF to 1F (@ 1kHz)	0.28 % to 0.66 %
46	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	Capacitance, Fluke 5522A, Capacitor Standard GR 1417	Direct method	10 nF(@ 1kHz) to 1uF (@ 100Hz)	0.41%
47	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	Capacitance, Fluke 5522A, Capacitor Standard GR 1417	Direct method	1uF (@ 100kHz) to 100 mF(@ 1kHz)	0.41 % to 0.28 %
48	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	Capacitance, Fluke 5522A, Capacitor Standard GR 1417	Direct Method	220 pF(@ 1KHz) to 10nF (@ 1KHz)	5.8 % to .41 %
49	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	Inductance, Standard Inductor GENRAD 1482 series	Direct method	100 μH to 10 H	0.3 % to 0.12 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

TEKTRONIX INDIA PVT LTD, SALARPURIA PREMIA, SURVEY NO 16, KADUBEESANAHALLI, VARTHUR HOBLI, BENGALURU, BENGALURU URBAN, KARNATAKA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-2130

Page No

10 of 50

Validity

22/05/2020 to 21/05/2022

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrum	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
50	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	Capacitance @ 1kHz, RLC Digi bridge Quadtech 1693	Direct Method	10 nF to 100 nF	0.41 % to 0.042 %
51	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	Capacitance @ 1KHz, RLC Digi bridge Quadtech 1693	Direct method	10 pF to 100 pF	0.25 % to 0.21 %
52	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	Capacitance @ 1kHz, RLC Digi bridge Quadtech 1693	Direct method	100 nF to 1000 nF	0.042 % to 0.033 %
53	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	Capacitance @ 1kHz, RLC Digi bridge Quadtech 1693	Direct method	100 pF to 10 nF	0.21 % to 0.07 %
54	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	Capacitance @ 1KHz, RLC Digi bridge Quadtech 1693	direct method	1000 nF to 100 uF	0.033 % to 0.32 %
55	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	DC current, Agilent 3458	direct method	100 nA to 1 µA	0.02 % to 0.0015 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

TEKTRONIX INDIA PVT LTD, SALARPURIA PREMIA, SURVEY NO 16,
KADUBEESANAHALLI, VARTHUR HOBLI, BENGALURU, BENGALURU URBAN,
KARNATAKA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-2130

Page No

11 of 50

Validity

22/05/2020 to 21/05/2022

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrum	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
56	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	DC current, Fluke 8508A, Fluke A40B Shunts, Vaiseshika 9410.	Direct method	100 μ A to 1 A	0.0010 % to 0.0012 %
57	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	DC current, Fluke 8508A, Fluke A40B Shunts, Vaiseshika 9410.	direct method	20 A to 100 A	0.32 % to 0.33 %
58	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	DC current, Fluke 8508A, Fluke A40B Shunts, Vaiseshika 9410. Fluke 742	Direct method	1 μ A to 100 μ A	0.0015 % to 0.0012 %
59	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	DC current, Fluke 8508A, Fluke A40B Shunts, Vaiseshika 9410. Fluke 742	direct method	1 A to 20 A	0.0012 % to 0.0047 %
60	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	DC Voltage, Fluke 8508A	Direct Method	10 mV to 100 mV	0.0052 % to 0.0005 %
61	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	DC Voltage, Fluke 8508A	direct method	100 μ V to 1 mV	0.037 % to 0.0082 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

TEKTRONIX INDIA PVT LTD, SALARPURIA PREMIA, SURVEY NO 16,
KADUBEESANAHALLI, VARTHUR HOBLI, BENGALURU, BENGALURU URBAN,
KARNATAKA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-2130

Page No

12 of 50

Validity

22/05/2020 to 21/05/2022

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrum	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
62	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	DC Voltage, Fluke 8508A	Direct Method	100 mV to 1000 V	0.0005 % to 0.00053 %
63	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	DC Voltage, Fluke 8508A, 80K-40, Fluke 752, Null detector, DC Ref Std, Fluke 7000N	Direct Method	1 mV to 10 mV	0.004 % to 0.00052 %
64	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	Inductance, RLC Digi bridge Quadtech 1693	direct method	1 H to 10 H	0.041 % to 0.22 %
65	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	Inductance, RLC Digi bridge Quadtech 1693	direct method	10 mH to 1 H	0.23 % to 0.041 %
66	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	Inductance, RLC Digi bridge Quadtech 1693	direct method	100 µH to 10 mH	0.22 % to 0.23 %
67	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	Resistance, Fluke 8508A,	direct method	1 Ohm to 10 Ohm	0.0044 % to 0.0017 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

TEKTRONIX INDIA PVT LTD, SALARPURIA PREMIA, SURVEY NO 16,
KADUBEESANAHALLI, VARTHUR HOBLI, BENGALURU, BENGALURU URBAN,
KARNATAKA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-2130

Page No

13 of 50

Validity

22/05/2020 to 21/05/2022

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured / Instrum	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
68	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	Resistance, Fluke 8508A,	direct method	10 M Ohm to 100 M Ohm	0.004 % to 0.0198 %
69	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	Resistance, Fluke 8508A,	direct method	10 Ohm to 100 Ohm	0.0017 % to 0.0015 %
70	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	Resistance, Fluke 8508A,	direct method	100 K Ohm to 10 M Ohm	0.0033 % to 0.004 %
71	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	Resistance, Fluke 8508A,	direct method	100 M Ohm to 1 G Ohm	.0198 % to .17 %
72	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	Resistance, Fluke 8508A,	direct method	100 Ohm to 100 K Ohm	15 ppm to 33 ppm
73	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	DC Current, Fluke 5730A, Fluke 5720A	Direct Method	100 µA to 1 mA	117 ppm to 50 ppm



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

TEKTRONIX INDIA PVT LTD, SALARPURIA PREMIA, SURVEY NO 16,
KADUBEESANAHALLI, VARTHUR HOBLI, BENGALURU, BENGALURU URBAN,
KARNATAKA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-2130

Page No

14 of 50

Validity

22/05/2020 to 21/05/2022

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrum	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
74	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	DC Current, Fluke 5730A	Direct method	1 mA to 1 A	15 ppm
75	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	DC Current, Fluke 5730A	direct method	10 μ A to 100 μ A	74 ppm to 117 ppm
76	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	DC Current, Fluke 5730A, Fluke 5522A	direct method	10 A to 20 A	472 ppm to 1400 ppm
77	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	DC Current, Fluke 5730A, Fluke 5522A with 5500A Coil	direct method	20 A to 1000 A	0.35 % to 0.77 %
78	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	DC Current, Fluke 5730A, Fluke 5725A	direct method	1 A to 10 A	110 ppm to 470 ppm
79	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	DC Voltage, Fluke 5730A	Direct method	1 mV to 100 mV	5 ppm



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

TEKTRONIX INDIA PVT LTD, SALARPURIA PREMIA, SURVEY NO 16,
KADUBEESANAHALLI, VARTHUR HOBLI, BENGALURU, BENGALURU URBAN,
KARNATAKA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-2130

Page No

15 of 50

Validity

22/05/2020 to 21/05/2022

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured / Instrum	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
80	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	DC Voltage, Fluke 5730A, 5720A, Fluke 742 with divider, 8508A Null detector	Direct method	100 mV to 100 V	5 ppm to 5.3 ppm
81	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	DC Voltage, Fluke 5730A, 5720A, Fluke 742 with divider, 8508A Null detector	Direct method	100 V to 1000 V	2 ppm
82	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	Resistance, 742A-10	DIRECT METHOD	10 Ohm	2 ppm
83	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	Resistance, 8508A-7000K	Direct method	1 G Ohm	200 ppm
84	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	Resistance, Fluke 5730A and 5522A	direct method	1 Ohm to 10 Ohm	0.011 % to 0.0099 %
85	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	Resistance, Fluke 5730A and 5522A	DIRECT METHOD	10 Ohm to 100 Ohm	0.009 % to 0.0128 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

TEKTRONIX INDIA PVT LTD, SALARPURIA PREMIA, SURVEY NO 16,
KADUBEESANAHALLI, VARTHUR HOBLI, BENGALURU, BENGALURU URBAN,
KARNATAKA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-2130

Page No

16 of 50

Validity

22/05/2020 to 21/05/2022

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrum	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
86	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	Resistance, Fluke 5730A and 5522A	DIRECT METHOD	100 M Ohm to 1 G Ohm	136 ppm to 322 ppm
87	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	Resistance, Fluke 742A - 1 K	DIRECTMETHOD	1 k Ohm	9 ppm
88	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	Resistance, Fluke 742A - 10M	Direct method	10 M Ohm	13 ppm
89	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	Resistance, Fluke 742A -1M	Direct method	1 M Ohm	2 ppm
90	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	Resistance, Fluke 742A, 8508A-7000K	DIRECT METHOD	1 Ohm	2 ppm
91	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	Resistance, Fluke 742A, 8508A-7000K	direct method	10 K Ohm	8 ppm



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

TEKTRONIX INDIA PVT LTD, SALARPURIA PREMIA, SURVEY NO 16,
KADUBEESANAHALLI, VARTHUR HOBLI, BENGALURU, BENGALURU URBAN,
KARNATAKA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-2130

Page No

17 of 50

Validity

22/05/2020 to 21/05/2022

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured / Instrum	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
92	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	Resistance, Fluke 742A-100	DIRECT METHOD	100 Ohm	2 ppm
93	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	Resistance, Fluke 742A-100K	Direct method	100 k Ohm	2 ppm
94	ELECTRO-TECHNICAL-ELECTRICAL EQUIPMENT (Measure)	AC Active Power at 50 Hz 60 V to 240 V/100 mA to 20 A/ 0.8 PF	Power Analyser PA 4400 by direct method	4.8 W to 3840 W	0.60 % to 0.24 %
95	ELECTRO-TECHNICAL-OTHERS (Measure)	Phase Angle (V-V)	Phase Meter , Clark Hess 6000 by Direct method	- 180 ° to +180 °	0.034°
96	ELECTRO-TECHNICAL-RF/MICROWAVE (1 GHZ AND ABOVE) (Measure)	Amplitude Modulation With CF:1GHz (rate 1kHz)	Measuring Receiver HP 8902A by direct method	Modulation Depth: 10 % to 90 %	3.5 % to 3.6 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

TEKTRONIX INDIA PVT LTD, SALARPURIA PREMIA, SURVEY NO 16, KADUBEESANAHALLI, VARTHUR HOBLI, BENGALURU, BENGALURU URBAN, KARNATAKA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-2130

Page No

18 of 50

Validity

22/05/2020 to 21/05/2022

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured / Instrum	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
97	ELECTRO-TECHNICAL-RF/MICROWAVE (1 GHZ AND ABOVE) (Measure)	Frequency Modulation at CF: 1GHz	Measuring Receiver, HP 8902A by direct method	FM Deviation 10 kHz to 200 kHz	6.0 % to 6.1 %
98	ELECTRO-TECHNICAL-RF/MICROWAVE (1 GHZ AND ABOVE) (Measure)	RF Attenuation at 100 kHz to 43 GHz	E8257D, MG3694 C Signal Generators and Power meters with Sensors 8487D & NRPZ-55 by Comparison Method	1 dB to 50 dB	0.25 dB to 0.45 dB
99	ELECTRO-TECHNICAL-RF/MICROWAVE (1 GHZ AND ABOVE) (Measure)	RF Frequency	Frequency Counter 53132 A & R&S FW 43 Spectrum Analyser by Direct Method	1 Hz to 20 GHz	0.058 ppm to 0.0029 ppm
100	ELECTRO-TECHNICAL-RF/MICROWAVE (1 GHZ AND ABOVE) (Measure)	RF Frequency	Frequency Counter 53151 A & R&S FW 43 Spectrum Analyser by Direct Method	20 GHz to 43.5 GHz	0.0029 ppm to 0.0058 ppm
101	ELECTRO-TECHNICAL-RF/MICROWAVE (1 GHZ AND ABOVE) (Measure)	RF Power @ 100 kHz to 40 GHz	Keysight Power meter with 8487D Power sensor and Power meter with NRP- 50 T Power sensor by Direct/ Comparison method	+18 dBm to -70 dBm	5.5 % to 8.7 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

TEKTRONIX INDIA PVT LTD, SALARPURIA PREMIA, SURVEY NO 16,
KADUBEESANAHALLI, VARTHUR HOBLI, BENGALURU, BENGALURU URBAN,
KARNATAKA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-2130

Page No

19 of 50

Validity

22/05/2020 to 21/05/2022

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured / Instrum	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
102	ELECTRO-TECHNICAL-RF/MICROWAVE (1 GHZ AND ABOVE) (Measure)	RF Power at > 40 GHz to 50 GHz	Keysight Power meter with 8487D Power sensor and Power meter with NRZ-55 Power sensor by direct/ Comparison method	+15 dBm to -60 dBm	5.6 % to 9 %
103	ELECTRO-TECHNICAL-RF/MICROWAVE (1 GHZ AND ABOVE) (Measure)	VSWR @ 10 MHz to 50 GHz	Maury microwave loads 2611A Series, RF bridge 560-97N-1, Type-N Calibration Kit Agilent 85056A Mechanical Calibration kit (DC-50GHz) by Comparison method	1.05 to 2.0	0.32 LU to 0.62 LU
104	ELECTRO-TECHNICAL-RF/MICROWAVE (1 GHZ AND ABOVE) (Source)	Oscilloscope Amplitude (DC Signal)	Fluke 9500 Scope calibrator with 9530 Head by Direct Method	1 mV to 100 mV	0.21 % to 0.06 %
105	ELECTRO-TECHNICAL-RF/MICROWAVE (1 GHZ AND ABOVE) (Source)	Oscilloscope Amplitude (DC Signal)	Fluke 9500 Scope calibrator with 9530 Head by Direct Method	100 mV to 190 V	0.06 % to 0.03 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

TEKTRONIX INDIA PVT LTD, SALARPURIA PREMIA, SURVEY NO 16,
KADUBEESANAHALLI, VARTHUR HOBLI, BENGALURU, BENGALURU URBAN,
KARNATAKA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-2130

Page No

20 of 50

Validity

22/05/2020 to 21/05/2022

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured / Instrum	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
106	ELECTRO-TECHNICAL-RF/MICROWAVE (1 GHZ AND ABOVE) (Source)	Oscilloscope Bandwidth with Reference @50 kHz and 10 MHz	Anritsu MG3694C Signal Generator with R&S Power meter NRP 50 T and E 8257 D Signal Generator by direct method	>1 GHz to 18 GHz	0.26 dB to 0.35 dB
107	ELECTRO-TECHNICAL-RF/MICROWAVE (1 GHZ AND ABOVE) (Source)	Oscilloscope Bandwidth with Reference @50 kHz and 10 MHz	Fluke 9500 Scope calibrator with 9530 Head by Direct Method	10 MHz to 1 GHz	0.22 dB to 0.26 dB
108	ELECTRO-TECHNICAL-RF/MICROWAVE (1 GHZ AND ABOVE) (Source)	Oscilloscope Scope Bandwidth with Reference @50 kHz and 10 MHz	Anritsu MG3694C Signal Generator with R&S Power meter NRP 50 T and E 8257 D Signal Generator by direct method	>18 GHz to 26 GHz	0.35 dB to 0.45 dB
109	ELECTRO-TECHNICAL-RF/MICROWAVE (1 GHZ AND ABOVE) (Source)	Oscilloscope Time Base	Fluke 9500 Scope calibrator with 9530 Head by Direct Method	1 nS to 5 S	1 ppm to 3.5 ppm



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

TEKTRONIX INDIA PVT LTD, SALARPURIA PREMIA, SURVEY NO 16,
KADUBEESANAHALLI, VARTHUR HOBLI, BENGALURU, BENGALURU URBAN,
KARNATAKA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-2130

Page No

21 of 50

Validity

22/05/2020 to 21/05/2022

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured / Instrum	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
110	ELECTRO-TECHNICAL-RF/MICROWAVE (1 GHZ AND ABOVE) (Source)	RF Attenuation at 100 kHz to 18 GHz	HP 8496B and HP8494 B Step Attenuator by direct method	1 dB to 50 dB	0.25 dB to 0.45 dB
111	ELECTRO-TECHNICAL-RF/MICROWAVE (1 GHZ AND ABOVE) (Source)	RF Frequency	Keysight Signal Generator E8257D by Direct Method	1 Hz to 20 GHz	0.058 ppm to 0.0029 ppm
112	ELECTRO-TECHNICAL-RF/MICROWAVE (1 GHZ AND ABOVE) (Source)	RF Frequency	Keysight Signal Generator E8257D	20 GHz to 43.5 GHz	0.0029 ppm to 0.0058 ppm
113	ELECTRO-TECHNICAL-RF/MICROWAVE (1 GHZ AND ABOVE) (Source)	RF Power @ >40 GHz to 50 GHz	Keysight E8257D and Anritsu MG 3694 C RF Signal Generator with Keysight Power meter with 8487D Power sensor and Power meter with NRZ-55 Power sensor by direct method	+15 dBm to -60 dBm	7.33 % to 9 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

TEKTRONIX INDIA PVT LTD, SALARPURIA PREMIA, SURVEY NO 16, KADUBEESANAHALLI, VARTHUR HOBLI, BENGALURU, BENGALURU URBAN, KARNATAKA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-2130

Page No

22 of 50

Validity

22/05/2020 to 21/05/2022

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured / Instrum	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
114	ELECTRO-TECHNICAL-RF/MICROWAVE (1 GHZ AND ABOVE) (Source)	RF Power @ 100 kHz to 40 GHz	Keysight E8257D and Anritsu MG 3694 C RF Signal Generator with Keysight Power meter with 8487D Power sensor and Power meter with NRZ-55 Power sensor by direct method	+18 dBm to -70 dBm	5.5 % to 8.81 %
115	ELECTRO-TECHNICAL-RF/MICROWAVE (1 GHZ AND ABOVE) (Source)	VSWR @10MHz to 18 GHz	VSWR Sources by direct method	Fixed values 1.05, 1.10 to 1.50, 2.0	0.095 LU to 0.40 LU
116	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure)	Temperature Simulation, Fluke 8508A,/Zero bath	direct method	E type (-) 250°C to 1000°C	0.03°C
117	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure)	Temperature Simulation, Fluke 8508A,/Zero bath	direct method	J type (-) 200°C to 1200°C	0.014°C



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

TEKTRONIX INDIA PVT LTD, SALARPURIA PREMIA, SURVEY NO 16,
KADUBEESANAHALLI, VARTHUR HOBLI, BENGALURU, BENGALURU URBAN,
KARNATAKA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-2130

Page No

23 of 50

Validity

22/05/2020 to 21/05/2022

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrum	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
118	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure)	Temperature Simulation, Fluke 8508A,/Zero bath	Direct Method	K type (-) 200°C to 1372°C	0.018°C
119	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure)	Temperature Simulation, Fluke 8508A,/Zero bath	direct method	N type (-) 200°C to 1300°C	0.026°C
120	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure)	Temperature Simulation, Fluke 8508A,/Zero bath	direct method	R type 0 °C to 1750°C	0.23°C
121	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure)	Temperature Simulation, Fluke 8508A,/Zero bath	direct method	RTD (-) 200 °C to 800 °C	0.014°C
122	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure)	Temperature Simulation, Fluke 8508A,/Zero bath	direct method	S type 0°C to 1750°C	0.13°C
123	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure)	Temperature Simulation, Fluke 8508A,/Zero bath	Direct method	T type (-) 250°C to 400°C	0.015°C



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

TEKTRONIX INDIA PVT LTD, SALARPURIA PREMIA, SURVEY NO 16,
KADUBEESANAHALLI, VARTHUR HOBOLI, BENGALURU, BENGALURU URBAN,
KARNATAKA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-2130

Page No

24 of 50

Validity

22/05/2020 to 21/05/2022

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrum	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
124	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Source)	Temperature Simulation, Fluke 5522A	Direct Method	J type (-) 200 °C to 1200 °C	0.32°C
125	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Source)	Temperature Simulation, Fluke 5522A	Direct Method	K type (-)200 °C to 1372 °C	0.47°C
126	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Source)	Temperature Simulation, Fluke 5522A	direct method	N type (-) 200 °C to 1300 °C	0.47°C
127	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Source)	Temperature Simulation, Fluke 5522A	direct method	R type 0°C to 1750°C	0.67°C
128	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Source)	Temperature Simulation, Fluke 5522A	Direct Method	RTD (-) 200 °C to 800 °C	0.37°C
129	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Source)	Temperature Simulation, Fluke 5522A	direct method	S type 0 °C to 1750 °C	0.56°C



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

TEKTRONIX INDIA PVT LTD, SALARPURIA PREMIA, SURVEY NO 16,
KADUBEESANAHALLI, VARTHUR HOBLI, BENGALURU, BENGALURU URBAN,
KARNATAKA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-2130

Page No

25 of 50

Validity

22/05/2020 to 21/05/2022

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrum	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
130	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Source)	Temperature Simulation, Fluke 5522A	direct method	T Type (-)250 °C to 400 °C	0.74°C
131	ELECTRO-TECHNICAL-TIME & FREQUENCY (Measure)	Time Period, Counter Tektronix MCA 3040	direct method	1 nS to 5 S	1 ppm to 0.8 ppm
132	THERMAL-TEMPERATURE	Oven, Bath, Dry Well Furnace (Single Point) / S-Type Thermocouple & 8 1/2 DMM	Comparison method	400 °C to 1000 ° C	1.78° C
133	THERMAL-TEMPERATURE	Oven, Bath, Dry Well Furnace (Single Point) / SPRT & 8 1/2 digit DMM.	Comparison method.	-40 °C to 400 °C	0.31°C
134	THERMAL-TEMPERATURE	Temperature - RTD, Thermocouple, (with or without Indicator) Temperature Gauge with dry well, SPRT & 8 1/2 DMM, S-Type Thermocouple & 8 1/2 DMM	Comparison Method	-40 °C to 400 °C	0.30°C



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

TEKTRONIX INDIA PVT LTD, SALARPURIA PREMIA, SURVEY NO 16,
KADUBEESANAHALLI, VARTHUR HOBLI, BENGALURU, BENGALURU URBAN,
KARNATAKA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-2130

Page No

26 of 50

Validity

22/05/2020 to 21/05/2022

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured / Instrum	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
135	THERMAL-TEMPERATURE	Temperature - Thermocouples (with or without Indicator) Temperature Gauge with S-Type Thermocouple & 8 1/2 DMM	Comparison method	400 °C to 1000 ° C	1.76° C



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

TEKTRONIX INDIA PVT LTD, SALARPURIA PREMIA, SURVEY NO 16,
KADUBEESANAHALLI, VARTHUR HOBLI, BENGALURU, BENGALURU URBAN,
KARNATAKA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-2130

Page No

27 of 50

Validity

22/05/2020 to 21/05/2022

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrum	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
Site Facility					
1	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Measure)	AC current, Fluke 8508A	direct method	1 A to 20 A(10 Hz- 5 kHz)	0.02 % to 0.3 %
2	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Measure)	AC current, Fluke 8508A	direct method	1 mA(300 Hz- 10 kHz) to 10 mA(10 Hz-10 kHz)	0.05 % to 0.023 %
3	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Measure)	AC current, Fluke 8508A	direct method	10 mA(300 Hz- 10 kHz) to 1 A(10 Hz- 10 kHz)	0.023 % to 0.056 %
4	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Measure)	AC current, Fluke 8508A	direct method	100 uA to 1 mA (300 Hz- 1kHz)	0.05%



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

TEKTRONIX INDIA PVT LTD, SALARPURIA PREMIA, SURVEY NO 16,
KADUBEESANAHALLI, VARTHUR HOBLI, BENGALURU, BENGALURU URBAN,
KARNATAKA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-2130

Page No

28 of 50

Validity

22/05/2020 to 21/05/2022

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured / Instrum	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
5	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Measure)	AC Voltage, Fluke 5790A	direct method	1 mV(10 kHz-100 kHz) to 1 V(10 kHz-100 kHz)	0.42 % to 0.016 %
6	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Measure)	AC Voltage, Fluke 5790A	direct method	1 V to 100V (10 Hz-40 Hz)	0.038 % to 0.01 %
7	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Measure)	AC Voltage, Fluke 5790A	direct method	10 mV to 1 V(10 Hz-40 Hz)	0.038%
8	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Measure)	AC Voltage, Fluke 5790A	Direct Method	10 mV to 1V (40 Hz-20 kHz)	0.023 % to 0.008 %
9	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Measure)	AC Voltage, Fluke 5790A	direct method	10 mV to 20 V(100 kHz- 1MHz)	0.19 % to 0.14 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

TEKTRONIX INDIA PVT LTD, SALARPURIA PREMIA, SURVEY NO 16, KADUBEESANAHALLI, VARTHUR HOBLI, BENGALURU, BENGALURU URBAN, KARNATAKA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-2130

Page No

29 of 50

Validity

22/05/2020 to 21/05/2022

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured / Instrum	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
10	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Measure)	AC Voltage, Fluke 5790A	Direct method	10 V(20 kHz-100 kHz) to 500 V(10 kHz-100 kHz)	0.009 % to 0.03 %
11	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Measure)	AC Voltage, Fluke 5790A	direct method	1mV to 10mV (10 Hz-40 Hz)	0.44 % to 0.038 %
12	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Measure)	AC Voltage, Fluke 5790A	Direct Method	1V to 1000 V(40 Hz-20 kHz)	0.008 % to 0.015 %
13	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Measure)	AC Voltage, Fluke 5790A	Direct method	1V to 10 V (10 kHz-100 kHz)	0.016 % to 0.009%
14	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Measure)	AC Voltage, Fluke 5790A	direct method	1 mV to 10 mV(100 kHz- 1MHz)	0.44 % to 0.19 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

TEKTRONIX INDIA PVT LTD, SALARPURIA PREMIA, SURVEY NO 16,
KADUBEESANAHALLI, VARTHUR HOBLI, BENGALURU, BENGALURU URBAN,
KARNATAKA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-2130

Page No

30 of 50

Validity

22/05/2020 to 21/05/2022

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured / Instrum	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
15	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Measure)	AC Voltage, Fluke 8508A 5790A	direct method	1mV to 10 mV(40 Hz- 20 KHz)	0.17 % to 0.023 %
16	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	AC current, Fluke 5522A	direct method,	10 A to 20 A(40Hz-1kHz)	0.059 % to 0.20 %
17	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	AC Current, Fluke 5522A with 5500A coil	Direct method	20A A to 1000 (@50Hz) A	0.35 % to 0.77 %
18	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	AC Current, Fluke 5720A	direct method	100uA to 1mA (40 Hz-1kHz)	0.109 % to 0.024 %
19	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	AC Current, Fluke 5730A	Direct method	100 mA to 1 A(40Hz-1kHz)	0.03 % to 0.04 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

TEKTRONIX INDIA PVT LTD, SALARPURIA PREMIA, SURVEY NO 16, KADUBEESANAHALLI, VARTHUR HOBLI, BENGALURU, BENGALURU URBAN, KARNATAKA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-2130

Page No

31 of 50

Validity

22/05/2020 to 21/05/2022

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured / Instrum	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
20	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	AC Current, Fluke 5730A	direct method	1mA to 100 mA (40Hz - 1 kHz)	0.024 % to 0.026 %
21	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	AC Current, Fluke 5730A	Direct method	200uA to 100mA (10Hz-40Hz)	0.11 % to 0.045 %
22	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	AC Current, Fluke 5730A & 5725A	direct method	20mA to 10 A(5kHz-10kHz)	0.03 % to 0.42 %
23	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	AC Current, Fluke 5730A & 5725A	direct method	20 mA to 10 A(1kHz-10 kHz)	0.056 % to 0.11 %
24	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	AC Current, Fluke 5730A, 5725A	Direct method, using AC shunts	1A to 10 A (40Hz-1kHz)	0.05 % to 0.059 %
25	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	AC Voltage, Fluke 5730A	Direct method	1 mV to 1 V(10 Hz-40 Hz)	0.49 % to 0.052 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

TEKTRONIX INDIA PVT LTD, SALARPURIA PREMIA, SURVEY NO 16, KADUBEESANAHALLI, VARTHUR HOBLI, BENGALURU, BENGALURU URBAN, KARNATAKA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-2130

Page No

32 of 50

Validity

22/05/2020 to 21/05/2022

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured / Instrum	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
26	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	AC Voltage, Fluke 5730A	direct method	1mV to 2 mV (40 Hz - 1 kHz)	0.6 % to 0.25 %
27	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	AC Voltage, Fluke 5730A	Direct method	20 mV to 20V(10 Hz-40 Hz)	0.05 % to 0.03 %
28	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	AC Voltage, Fluke 5730A	direct method	20 V to 1000 V (40 Hz - 1 kHz)	0.0065 % to 0.0090 %
29	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	AC Voltage, Fluke 5730A	direct method	20 V to 200V (10 Hz-40 Hz)	0.03 % to 0.025 %
30	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	AC Voltage, Fluke 5730A	direct method	200 mV to 2V (40 Hz - 1 kHz)	0.015 % to 0.006 %
31	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	AC Voltage, Fluke 5730A	direct method	200mV to 20V (100 kHz-1 MHz)	0.46 % to 0.25 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

TEKTRONIX INDIA PVT LTD, SALARPURIA PREMIA, SURVEY NO 16,
KADUBEESANAHALLI, VARTHUR HOBBLI, BENGALURU, BENGALURU URBAN,
KARNATAKA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-2130

Page No

33 of 50

Validity

22/05/2020 to 21/05/2022

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured / Instrum	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
32	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	AC Voltage, Fluke 5730A	direct method	20mV to 200 mV (100 kHz-1 MHz)	0.46 % to 0.27 %
33	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	AC Voltage, Fluke 5730A	Direct method	2mV to 20 mV(40 Hz - 1 kHz)	0.25 % to 0.03 %
34	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	AC Voltage, Fluke 5730A	Direct method	2mV to 20mV(10 Hz- 40 Hz)	0.26 % to 0.05 %
35	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	AC Voltage, Fluke 5730A	direct method	2mV to 2 V (1 kHz - 100 kHz)	0.45 % to 0.015 %
36	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	AC Voltage, Fluke 5730A	direct method	2mV (100 kHz - 1 MHz) to 20 mV (100 kHz - 1MHz)	1.75 % to 0.46 %
37	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	AC Voltage, Fluke 5730A	direct method	2V to 20V (40 Hz - 1 kHz)	0.006 % to 0.0065 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

TEKTRONIX INDIA PVT LTD, SALARPURIA PREMIA, SURVEY NO 16, KADUBEESANAHALLI, VARTHUR HOBLI, BENGALURU, BENGALURU URBAN, KARNATAKA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-2130

Page No

34 of 50

Validity

22/05/2020 to 21/05/2022

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured / Instrum	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
38	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	AC Voltage, Fluke 5730A, 5725A	direct method	100 V to 500V (1 kHz - 100 kHz)	0.02 % to 0.27 %
39	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	AC Voltage,Fluke 5730A	direct method	20 mV to 200mV (40 Hz - 1 kHz)	0.03 % to 0.015 %
40	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	AC Voltage,Fluke 5730A	direct method	2V to 100V (1 kHz - 100 kHz)	0.015 % to 0.02 %
41	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	Capacitance, Capacitor Standard GR 1417	Direct Method	100mF to 1F (@ 1kHz)	0.28 % to 0.66 %
42	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	Capacitance, Fluke 5522A, Capacitor Standard GR 1417	Direct method	10 nF(@ 1kHz) to 1uF (@ 100Hz)	0.41%
43	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	Capacitance, Fluke 5522A, Capacitor Standard GR 1417	Direct method	1uF (@ 100kHz) to 100 mF(@ 1kHz)	0.41 % to 0.28 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

TEKTRONIX INDIA PVT LTD, SALARPURIA PREMIA, SURVEY NO 16,
KADUBEESANAHALLI, VARTHUR HOBOLI, BENGALURU, BENGALURU URBAN,
KARNATAKA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-2130

Page No

35 of 50

Validity

22/05/2020 to 21/05/2022

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrum	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
44	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	Capacitance, Fluke 5522A, Capacitor Standard GR 1417	Direct Method	220 pF(@ 1KHz) to 10nF (@ 1KHz)	5.8 % to .41 %
45	ELECTRO-TECHNICAL-ALTERNATING CURRENT (< 1 GHZ) (Source)	Inductance, Standard Inductor GENRAD 1482 series	Direct method	100 μH to 10 H	0.3 % to 0.12 %
46	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	Capacitance @ 1kHz, RLC Digi bridge Quadtech 1693	Direct Method	10 nF to 100 nF	0.41 % to 0.042 %
47	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	Capacitance @ 1KHz, RLC Digi bridge Quadtech 1693	Direct method	10 pF to 100 pF	0.25 % to 0.21 %
48	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	Capacitance @ 1kHz, RLC Digi bridge Quadtech 1693	Direct method	100 nF to 1000 nF	0.042 % to 0.033 %
49	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	Capacitance @ 1kHz, RLC Digi bridge Quadtech 1693	Direct method	100 pF to 10 nF	0.21 % to 0.07 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

TEKTRONIX INDIA PVT LTD, SALARPURIA PREMIA, SURVEY NO 16, KADUBEESANAHALLI, VARTHUR HOBLI, BENGALURU, BENGALURU URBAN, KARNATAKA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-2130

Page No

36 of 50

Validity

22/05/2020 to 21/05/2022

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured / Instrum	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
50	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	Capacitance @ 1KHz, RLC Digi bridge Quadtech 1693	direct method	1000 nF to 100 uF	0.033 % to 0.32 %
51	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	DC current, Agilent 3458	direct method	100 nA to 1 µA	0.02 % to 0.0015 %
52	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	DC current, Fluke 8508A, Fluke A40B Shunts, Vaiseshika 9410.	Direct method	100 µA to 1 A	0.0010 % to 0.0012 %
53	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	DC current, Fluke 8508A, Fluke A40B Shunts, Vaiseshika 9410.	direct method	20 A to 100 A	0.32 % to 0.33 %
54	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	DC current, Fluke 8508A, Fluke A40B Shunts, Vaiseshika 9410. Fluke 742	Direct method	1 µA to 100 µA	0.0015 % to 0.0012 %
55	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	DC current, Fluke 8508A, Fluke A40B Shunts, Vaiseshika 9410. Fluke 742	direct method	1 A to 20 A	0.0012 % to 0.0047 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

TEKTRONIX INDIA PVT LTD, SALARPURIA PREMIA, SURVEY NO 16,
KADUBEESANAHALLI, VARTHUR HOBLI, BENGALURU, BENGALURU URBAN,
KARNATAKA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-2130

Page No

37 of 50

Validity

22/05/2020 to 21/05/2022

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrum	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
56	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	DC Voltage, Fluke 8508A	Direct Method	10 mV to 100 mV	0.0052 % to 0.0005 %
57	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	DC Voltage, Fluke 8508A	direct method	100 µV to 1 mV	0.037 % to 0.0082 %
58	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	DC Voltage, Fluke 8508A	Direct Method	100 mV to 1000 V	0.0005 % to 0.00053 %
59	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	DC Voltage, Fluke 8508A, 80K-40, Fluke 752, Null detector, DC Ref Std, Fluke 7000N	Direct Method	1 mV to 10 mV	0.004 % to 0.00052 %
60	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	Inductance, RLC Digi bridge Quadtech 1693	direct method	1 H to 10 H	0.041 % to 0.22 %
61	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	Inductance, RLC Digi bridge Quadtech 1693	direct method	10 mH to 1 H	0.23 % to 0.041 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

TEKTRONIX INDIA PVT LTD, SALARPURIA PREMIA, SURVEY NO 16,
KADUBEESANAHALLI, VARTHUR HOBLI, BENGALURU, BENGALURU URBAN,
KARNATAKA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-2130

Page No

38 of 50

Validity

22/05/2020 to 21/05/2022

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured / Instrum	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
62	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	Inductance, RLC Digi bridge Quadtech 1693	direct method	100 µH to 10 mH	0.22 % to 0.23 %
63	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	Resistance, Fluke 8508A,	direct method	1 Ohm to 10 Ohm	0.0044 % to 0.0017 %
64	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	Resistance, Fluke 8508A,	direct method	10 M Ohm to 100 M Ohm	0.004 % to 0.0198 %
65	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	Resistance, Fluke 8508A,	direct method	10 Ohm to 100 Ohm	0.0017 % to 0.0015 %
66	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	Resistance, Fluke 8508A,	direct method	100 K Ohm to 10 M Ohm	0.0033 % to 0.004 %
67	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	Resistance, Fluke 8508A,	direct method	100 M Ohm to 1 G Ohm	.0198 % to .17 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

TEKTRONIX INDIA PVT LTD, SALARPURIA PREMIA, SURVEY NO 16,
KADUBEESANAHALLI, VARTHUR HOBLI, BENGALURU, BENGALURU URBAN,
KARNATAKA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-2130

Page No

39 of 50

Validity

22/05/2020 to 21/05/2022

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured / Instrum	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
68	ELECTRO-TECHNICAL-DIRECT CURRENT (Measure)	Resistance, Fluke 8508A,	direct method	100 Ohm to 100 K Ohm	15 ppm to 33 ppm
69	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	DC Current, Fluke 5730A, Fluke 5720A	Direct Method	100 µA to 1 mA	117 ppm to 50 ppm
70	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	DC Current, Fluke 5730A	Direct method	1 mA to 1 A	15 ppm
71	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	DC Current, Fluke 5730A	direct method	10 µA to 100 uA	74 ppm to 117 ppm
72	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	DC Current, Fluke 5730A, Fluke 5522A	direct method	10 A to 20 A	472 ppm to 1400 ppm
73	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	DC Current, Fluke 5730A, Fluke 5522A with 5500A Coil	direct method	20 A to 1000 A	0.35 % to 0.77 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

TEKTRONIX INDIA PVT LTD, SALARPURIA PREMIA, SURVEY NO 16,
KADUBEESANAHALLI, VARTHUR HOBLI, BENGALURU, BENGALURU URBAN,
KARNATAKA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-2130

Page No

40 of 50

Validity

22/05/2020 to 21/05/2022

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured / Instrum	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
74	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	DC Current, Fluke 5730A, Fluke 5725A	direct method	1 A to 10 A	110 ppm to 470 ppm
75	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	DC Voltage, Fluke 5730A	Direct method	1 mV to 100 mV	5 ppm
76	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	DC Voltage, Fluke 5730A, 5720A, Fluke 742 with divider, 8508A Null detector	Direct method	100 mV to 100 V	5 ppm to 5.3 ppm
77	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	DC Voltage, Fluke 5730A, 5720A, Fluke 742 with divider, 8508A Null detector	Direct method	100 V to 1000 V	2 ppm
78	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	Resistance, 742A-10	DIRECT METHOD	10 Ohm	2 ppm
79	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	Resistance, Fluke 5730A and 5522A	direct method	1 Ohm to 10 Ohm	0.011 % to 0.0099 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

TEKTRONIX INDIA PVT LTD, SALARPURIA PREMIA, SURVEY NO 16, KADUBEESANAHALLI, VARTHUR HOBLI, BENGALURU, BENGALURU URBAN, KARNATAKA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-2130

Page No

41 of 50

Validity

22/05/2020 to 21/05/2022

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured / Instrum	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
80	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	Resistance, Fluke 5730A and 5522A	DIRECT METHOD	10 Ohm to 100 Ohm	0.009 % to 0.0128 %
81	ELECTRO-TECHNICAL-DIRECT CURRENT (Source)	Resistance, Fluke 5730A and 5522A	DIRECT METHOD	100 M Ohm to 1 G Ohm	136 ppm to 322 ppm
82	ELECTRO-TECHNICAL-OTHERS (Measure)	Phase Angle (V-V)	Phase Meter , Clark Hess 6000 by Direct method	- 180 ° to +180 °	0.034°
83	ELECTRO-TECHNICAL-RF/MICROWAVE (1 GHZ AND ABOVE) (Measure)	Amplitude Modulation With CF:1GHz (rate 1kHz)	Measuring Receiver HP 8902A by direct method	Modulation Depth: 10 % to 90 %	3.5 % to 3.6 %
84	ELECTRO-TECHNICAL-RF/MICROWAVE (1 GHZ AND ABOVE) (Measure)	Frequency Modulation at CF: 1GHz	Measuring Receiver, HP 8902A by direct method	FM Deviation 10 kHz to 200 kHz	6.0 % to 6.1 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

TEKTRONIX INDIA PVT LTD, SALARPURIA PREMIA, SURVEY NO 16,
KADUBEESANAHALLI, VARTHUR HOBLI, BENGALURU, BENGALURU URBAN,
KARNATAKA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-2130

Page No

42 of 50

Validity

22/05/2020 to 21/05/2022

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured / Instrum	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
85	ELECTRO-TECHNICAL-RF/MICROWAVE (1 GHZ AND ABOVE) (Measure)	RF Attenuation at 100 kHz to 43 GHz	E8257D, MG3694 C Signal Generators and Power meters with Sensors 8487D & NRPZ-55 by Comparison Method	1 dB to 50 dB	0.25 dB to 0.45 dB
86	ELECTRO-TECHNICAL-RF/MICROWAVE (1 GHZ AND ABOVE) (Measure)	RF Frequency	Frequency Counter 53132 A & R&S FW 43 Spectrum Analyser by Direct Method	1 Hz to 20 GHz	0.058 ppm to 0.0029 ppm
87	ELECTRO-TECHNICAL-RF/MICROWAVE (1 GHZ AND ABOVE) (Measure)	RF Frequency	Frequency Counter 53151 A & R&S FW 43 Spectrum Analyser by Direct Method	20 GHz to 43.5 GHz	0.0029 ppm to 0.0058 ppm
88	ELECTRO-TECHNICAL-RF/MICROWAVE (1 GHZ AND ABOVE) (Measure)	RF Power @ 100 kHz to 40 GHz	Keysight Power meter with 8487D Power sensor and Power meter with NRP- 50 T Power sensor by Direct/ Comparison method	+18 dBm to -70 dBm	5.5 % to 8.7 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

TEKTRONIX INDIA PVT LTD, SALARPURIA PREMIA, SURVEY NO 16, KADUBEESANAHALLI, VARTHUR HOBOLI, BENGALURU, BENGALURU URBAN, KARNATAKA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-2130

Page No

43 of 50

Validity

22/05/2020 to 21/05/2022

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrum	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
89	ELECTRO-TECHNICAL-RF/MICROWAVE (1 GHZ AND ABOVE) (Measure)	RF Power at > 40 GHz to 50 GHz	Keysight Power meter with 8487D Power sensor and Power meter with NRZ-55 Power sensor by direct/ Comparison method	+15 dBm to -60 dBm	5.6 % to 9 %
90	ELECTRO-TECHNICAL-RF/MICROWAVE (1 GHZ AND ABOVE) (Measure)	VSWR @ 10 MHz to 50 GHz	Maury microwave loads 2611A Series, RF bridge 560-97N-1, Type-N Calibration Kit Agilent 85056A Mechanical Calibration kit (DC-50GHz) by Comparison method	1.05 to 2.0	0.32 LU to 0.62 LU
91	ELECTRO-TECHNICAL-RF/MICROWAVE (1 GHZ AND ABOVE) (Source)	Oscilloscope Amplitude (DC Signal)	Fluke 9500 Scope calibrator with 9530 Head by Direct Method	1 mV to 100 mV	0.21 % to 0.06 %
92	ELECTRO-TECHNICAL-RF/MICROWAVE (1 GHZ AND ABOVE) (Source)	Oscilloscope Amplitude (DC Signal)	Fluke 9500 Scope calibrator with 9530 Head by Direct Method	100 mV to 190 V	0.06 % to 0.03 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

TEKTRONIX INDIA PVT LTD, SALARPURIA PREMIA, SURVEY NO 16,
KADUBEESANAHALLI, VARTHUR HOBLI, BENGALURU, BENGALURU URBAN,
KARNATAKA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-2130

Page No

44 of 50

Validity

22/05/2020 to 21/05/2022

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured / Instrum	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
93	ELECTRO-TECHNICAL-RF/MICROWAVE (1 GHZ AND ABOVE) (Source)	Oscilloscope Bandwidth with Reference @50 kHz and 10 MHz	Anritsu MG3694C Signal Generator with R&S Power meter NRP 50 T and E 8257 D Signal Generator by direct method	>1 GHz to 18 GHz	0.26 dB to 0.35 dB
94	ELECTRO-TECHNICAL-RF/MICROWAVE (1 GHZ AND ABOVE) (Source)	Oscilloscope Bandwidth with Reference @50 kHz and 10 MHz	Fluke 9500 Scope calibrator with 9530 Head by Direct Method	10 MHz to 1 GHz	0.22 dB to 0.26 dB
95	ELECTRO-TECHNICAL-RF/MICROWAVE (1 GHZ AND ABOVE) (Source)	Oscilloscope Scope Bandwidth with Reference @50 kHz and 10 MHz	Anritsu MG3694C Signal Generator with R&S Power meter NRP 50 T and E 8257 D Signal Generator by direct method	>18 GHz to 26 GHz	0.35 dB to 0.45 dB
96	ELECTRO-TECHNICAL-RF/MICROWAVE (1 GHZ AND ABOVE) (Source)	Oscilloscope Time Base	Fluke 9500 Scope calibrator with 9530 Head by Direct Method	1 nS to 5 S	1 ppm to 3.5 ppm



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

TEKTRONIX INDIA PVT LTD, SALARPURIA PREMIA, SURVEY NO 16,
KADUBEESANAHALLI, VARTHUR HOBLI, BENGALURU, BENGALURU URBAN,
KARNATAKA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-2130

Page No

45 of 50

Validity

22/05/2020 to 21/05/2022

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured / Instrum	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
97	ELECTRO-TECHNICAL-RF/MICROWAVE (1 GHZ AND ABOVE) (Source)	RF Attenuation at 100 kHz to 18 GHz	HP 8496B and HP8494 B Step Attenuator by direct method	1 dB to 50 dB	0.25 dB to 0.45 dB
98	ELECTRO-TECHNICAL-RF/MICROWAVE (1 GHZ AND ABOVE) (Source)	RF Frequency	Keysight Signal Generator E8257D by Direct Method	1 Hz to 20 GHz	0.058 ppm to 0.0029 ppm
99	ELECTRO-TECHNICAL-RF/MICROWAVE (1 GHZ AND ABOVE) (Source)	RF Frequency	Keysight Signal Generator E8257D	20 GHz to 43.5 GHz	0.0029 ppm to 0.0058 ppm
100	ELECTRO-TECHNICAL-RF/MICROWAVE (1 GHZ AND ABOVE) (Source)	RF Power @ >40 GHz to 50 GHz	Keysight E8257D and Anritsu MG 3694 C RF Signal Generator with Keysight Power meter with 8487D Power sensor and Power meter with NRZ-55 Power sensor by direct method	+15 dBm to -60 dBm	7.33 % to 9 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

TEKTRONIX INDIA PVT LTD, SALARPURIA PREMIA, SURVEY NO 16,
KADUBEESANAHALLI, VARTHUR HOBLI, BENGALURU, BENGALURU URBAN,
KARNATAKA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-2130

Page No

46 of 50

Validity

22/05/2020 to 21/05/2022

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrum	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
101	ELECTRO-TECHNICAL-RF/MICROWAVE (1 GHZ AND ABOVE) (Source)	RF Power @ 100 kHz to 40 GHz	Keysight E8257D and Anritsu MG 3694 C RF Signal Generator with Keysight Power meter with 8487D Power sensor and Power meter with NRZ-55 Power sensor by direct method	+18 dBm to -70 dBm	5.5 % to 8.81 %
102	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure)	Temperature Simulation, Fluke 8508A,/Zero bath	direct method	E type (-) 250°C to 1000°C	0.03°C
103	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure)	Temperature Simulation, Fluke 8508A,/Zero bath	direct method	J type (-) 200°C to 1200°C	0.014°C
104	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure)	Temperature Simulation, Fluke 8508A,/Zero bath	Direct Method	K type (-) 200°C to 1372°C	0.018°C
105	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure)	Temperature Simulation, Fluke 8508A,/Zero bath	direct method	N type (-) 200°C to 1300°C	0.026°C



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

TEKTRONIX INDIA PVT LTD, SALARPURIA PREMIA, SURVEY NO 16,
KADUBEESANAHALLI, VARTHUR HOBLI, BENGALURU, BENGALURU URBAN,
KARNATAKA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-2130

Page No

47 of 50

Validity

22/05/2020 to 21/05/2022

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured / Instrum	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
106	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure)	Temperature Simulation, Fluke 8508A,/Zero bath	direct method	R type 0 °C to 1750°C	0.23°C
107	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure)	Temperature Simulation, Fluke 8508A,/Zero bath	direct method	RTD (-) 200 °C to 800 °C	0.014°C
108	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure)	Temperature Simulation, Fluke 8508A,/Zero bath	direct method	S type 0°C to 1750°C	0.13°C
109	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Measure)	Temperature Simulation, Fluke 8508A,/Zero bath	Direct method	T type (-) 250°C to 400°C	0.015°C
110	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Source)	Temperature Simulation, Fluke 5522A	Direct Method	J type (-) 200 °C to 1200 °C	0.32°C
111	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Source)	Temperature Simulation, Fluke 5522A	Direct Method	K type (-)200 °C to 1372 °C	0.47°C



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

TEKTRONIX INDIA PVT LTD, SALARPURIA PREMIA, SURVEY NO 16,
KADUBEESANAHALLI, VARTHUR HOBLI, BENGALURU, BENGALURU URBAN,
KARNATAKA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-2130

Page No

48 of 50

Validity

22/05/2020 to 21/05/2022

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrum	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
112	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Source)	Temperature Simulation, Fluke 5522A	direct method	N type (-) 200 °C to 1300 °C	0.47°C
113	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Source)	Temperature Simulation, Fluke 5522A	direct method	R type 0°C to 1750°C	0.67°C
114	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Source)	Temperature Simulation, Fluke 5522A	Direct Method	RTD (-) 200 °C to 800 °C	0.37°C
115	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Source)	Temperature Simulation, Fluke 5522A	direct method	S type 0 °C to 1750 °C	0.56°C
116	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION (Source)	Temperature Simulation, Fluke 5522A	direct method	T Type (-)250 °C to 400 °C	0.74°C
117	ELECTRO-TECHNICAL-TIME & FREQUENCY (Measure)	Time Period, Counter Tektronix MCA 3040	direct method	1 nS to 5 S	1 ppm to 0.8 ppm



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

TEKTRONIX INDIA PVT LTD, SALARPURIA PREMIA, SURVEY NO 16, KADUBEESANAHALLI, VARTHUR HOBLI, BENGALURU, BENGALURU URBAN, KARNATAKA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-2130

Page No

49 of 50

Validity

22/05/2020 to 21/05/2022

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrum	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
118	THERMAL-SPECIFIC HEAT & HUMIDITY	Specific Heat & Humidity, °C &%RH, Thermo Hygrometer, Environmental chambers, Temperature/ Humidity Indicator, Temperature Chamber	Comparison Method	0 °C to 50 °C at 50%RH	0.94°C
119	THERMAL-SPECIFIC HEAT & HUMIDITY	Specific Heat & Humidity, °C &%RH, Thermo Hygrometer, Environmental chambers, Temperature/ Humidity Indicator, Temperature Chamber	Comparison method	20 %RH to 95 %RH at 25°C	1.45%RH
120	THERMAL-TEMPERATURE	Oven, Bath (single point) / Using S type Thermocouple with Data logger.	Comparison Method	400 °C to 1000 ° C	1.74°C
121	THERMAL-TEMPERATURE	Oven, Bath, Dry Well Furnace (Single Point) / RTD with data logger.	Comparison method	-40 °C to 400 ° C	1.52°C
122	THERMAL-TEMPERATURE	Temperature Chamber, Oven, Furnace (multipoint) / Using RTD with Logger 2638A .	Comparision method	-70 ° C to 180 ° C	0.83° C



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

TEKTRONIX INDIA PVT LTD, SALARPURIA PREMIA, SURVEY NO 16,
KADUBEESANAHALLI, VARTHUR HOBOLI, BENGALURU, BENGALURU URBAN,
KARNATAKA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-2130

Page No

50 of 50

Validity

22/05/2020 to 21/05/2022

Last Amended on

-

* CMCs represent expanded uncertainties expressed at approximately the 95% level of confidence, using a coverage factor of $k = 2$.

