



**National Accreditation Board for
Testing and Calibration Laboratories**

(A Constituent Board of Quality Council of India)



CERTIFICATE OF ACCREDITATION

TEKTRONIX (INDIA) PVT. LTD.

has been assessed and accredited in accordance with the standard

ISO/IEC 17025:2005

"General Requirements for the Competence of Testing & Calibration Laboratories"

for its facilities at

Salarpuria Premia, Survey No. 16, Kadubeesana Halli, Varthur Hobli,
Sarjapur Outer Ring Road, Bangalore , Karnataka

in the field of

CALIBRATION

Certificate Number CC-2130

Issue Date 22/05/2018

Valid Until 21/05/2020

This certificate remains valid for the Scope of Accreditation as specified in the annexure subject to continued satisfactory compliance to the above standard & the relevant requirements of NABL.

(To see the scope of accreditation of this laboratory, you may also visit NABL website www.nabl-india.org)

Signed for and on behalf of NABL



89076970200020000482

Anil Relia

Anil Relia
Chief Executive Officer



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Tektronix (India) Pvt. Ltd., Salarpuria Premia, Survey No. 16, Kadubeesana Halli, Varthur Hobli, Sarjapur Outer Ring Road, Bangalore , Karnataka

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2130 **Page** 1 of 13

Validity 22.05.2018 to 21.05.2020 **Last Amended on** 22.06.2018

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability ($\pm$)	Remarks
<u>ELECTRO TECHNICAL CALIBRATION</u>				
I.	SOURCE			
1.	DC Voltage [§]	1mV to 100mV 100mV to 1V 1V to 100V 100V to 1000V	0.0474 % to 0.0014.3 % 0.0014.3 % to 0.0005 % 0.0005 % to 0.0007 % 0.0007 % to 0.00083 %	Using Calibrator Fluke 5730A, 5720A by Direct Method
	DC Voltage [*]	10mV to 100mV 100mV to 1V 1V to 100V 100V to 1000V	0.0485 % to 0.0014 % 0.0014 % to 0.0007 % 0.0007 % to 0.0008 % 0.0008 % to 0.00085 %	Using Calibrator Fluke 5720A by Direct Method
2.	AC Voltage [#]	10 Hz to 40 Hz 1mV to 2mV 2mV to 20mV 20mV to 20V 20V to 200V	0.78 % to 0.26 % 0.26 % to 0.05 % 0.05 % to 0.03 % 0.03 % to 0.025 %	Using Calibrator Fluke 5730A, 5720A, 5522A by Direct Method
		40 Hz to 1kHz 1mV to 2mV 2mV to 20mV 20mV to 200mV 200mV to 2V 2V to 20V 20V to 1000V	0.6 % to 0.25 % 0.25 % to 0.03 % 0.03 % to 0.015 % 0.015 % to 0.006 % 0.006 % to 0.0065 % 0.0065 % to 0.0090%	
		1 kHz to 100 kHz 2mV to 2V 2V to 100V	0.45 % to 0.015 % 0.015 % to 0.02 %	

Rjeshwar Kumar
Convener

Avijit Das
Program Manager



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Tektronix (India) Pvt. Ltd., Salarpuria Premia, Survey No. 16, Kadubeesana Halli, Varthur Hobli, Sarjapur Outer Ring Road, Bangalore , Karnataka

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2130

Page

2 of 13

Validity 22.05.2018 to 21.05.2020

Last Amended on 22.06.2018

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability ($\pm$)	Remarks
		100 kHz to 1 MHz 2mV to 20mV 20mV to 200mV 200mV to 20 V	1.75 % to 0.46 % 0.46 % to 0.27 % 0.36 % to 0.23 %	
3.	DC Current [#]	10 μ A to 100 μ A 100 μ A to 1mA 1mA to 100mA 100mA to 1A 1A to 3A 3A to 10A 10A to 20A >20A to 1000A	0.074 % to 0.0117 % 0.0117 % to .0050 % 0.0052 % to .0065 % 0.0065 % to 0.0110 % 0.0110 % to 0.0603 % 0.0603 % to 0.0472 % 0.0472 % to 0.14 % 0.14 % to 0.58 %	Using Calibrator Fluke 5720A by Direct Method Using Calibrator Fluke 5522A with 5500A Current Coil by Direct Method
4.	AC Current ^{\$}	10Hz to 40Hz 200 μ A to 100mA	0.109 % to 0.026 %	Using Calibrator Fluke 5720A by Direct Method
	AC Current [*]	10Hz to 40Hz 100 μ A to 100mA	0.11 % to 0.045 %	Using Calibrator Fluke 5720A by Direct Method
	AC Current [#]	40Hz to 1kHz 100 μ A to 1mA 1mA to 100mA 100mA to 1A 1A to 10A 10A to 20A 1kHz to 5kHz 20mA to 10A 5kHz to 10kHz 20mA to 10A	0.109 % to 0.024 % 0.024 % to 0.019 % 0.03 % to 0.04 % 0.05 % to 0.059 % 0.59 to 0.20 % 0.026 % to 0.11 % 0.03 % to 0.42 %	Using Calibrator Fluke 5730A by Direct Method Calibrator Fluke 5730A


Rjeshwar Kumar
Convener


Avijit Das
Program Manager



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory

Tektronix (India) Pvt. Ltd., Salarpuria Premia, Survey No. 16, Kadubeesana Halli, Varthur Hobli, Sarjapur Outer Ring Road, Bangalore, Karnataka

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number

CC-2130

Page

3 of 13

Validity

22.05.2018 to 21.05.2020

Last Amended on 22.06.2018

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability ($\pm$)	Remarks
		50Hz >20A to 1000A	0.35 % to 0.77 %	Using Calibrator Fluke 5522A with 5500A Current Coil by Direct Method
5.	Resistance ^s	1 Ω to 10M Ω 10M Ω to 100M Ω 100M Ω to 1G Ω	0.011 % to 0.0009 % 0.0009 % to 0.0128 % 0.0128 % to .0322 %	Using Calibrator Fluke 5730A, 5522A by Direct Method
		1 Ω 10 Ω 100 Ω 1K Ω 10K Ω 100K Ω 1M Ω 10M Ω 1G Ω	0.00068 % 0.0005 % 0.0009 % 0.0009 % 0.0008 % 0.0009 % 0.0012 % 0.0013 % 0.031 %	Fixed Resistors, 742A/ 8508A-7000K
	Resistance [*]	1 Ω to 10M Ω	0.01 % to .004 %	Using Calibrator Fluke 5720A
		10M Ω to 100M Ω 100M Ω to 1G Ω	0.004 % to 0.014 % 0.014 % to 0.85 %	Calibrator Fluke 5522A /8508A by Direct Method
6.	Capacitance [#]	1kHz 220PF to 10nF 10nF to 1 μ F 1 μ F to 100Mf	5.8 % to 0.41 % 0.41 % 0.41 % to 0.28 %	Using Calibrator Fluke 5522A
	Capacitance ^s	100mF to 1F	0.28 % to 0.66 %	Using Capacitor Standard GR 1417 by Direct Method

Rjeshwar Kumar
Convener

Avijit Das
Program Manager



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Tektronix (India) Pvt. Ltd., Salarpuria Premia, Survey No. 16, Kadubeesana Halli, Varthur Hobli, Sarjapur Outer Ring Road, Bangalore , Karnataka

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2130

Page

4 of 13

Validity 22.05.2018 to 21.05.2020

Last Amended on 22.06.2018

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability ($\pm$)	Remarks
7.	Inductance [#]	1kHz 100 μ H to 10H	0.3 % to 0.12 %	Using Standard Inductance Genrad 1482 Series by Direct Method
8.	DC Power [#]	1V to 600V 1mA to 1000A 1mW to 10W 10W to 1kW 1kW to 12kW	0.038% to 0.43% 0.03 % to 0.45 % 0.45% to 0.41%	Using Calibrator Fluke 5522A with 5500 Coil
9.	AC Power [#]	50 Hz to 1 kHz @ UPF 120 V to 240 V 0.01 A to 20 A 1.2 W to 4.8 kW	0.095 % to 0.12 %	Using Calibrator Fluke 5522A with 5500 Coil
		50Hz to 1kHz @ 0.8PF 120V to 240V 0.1A to 20A 9.6W to 3.8kW	0.099 % to 0.24 %	
		50Hz to 1kHz @ 0.5PF 120V to 240V 0.1A to 20A 6W to 2.4kW	0.38 % to 0.39 %	
		50Hz to 1kHz @ 0.2PF 120V to 240V 0.1A to 20A 2.4W to 960W	0.56 % to 0.58 %	



Rjeshwar Kumar
Convener



Avijit Das
Program Manager



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Tektronix (India) Pvt. Ltd., Salarpuria Premia, Survey No. 16, Kadubeesana Halli, Varthur Hobli, Sarjapur Outer Ring Road, Bangalore , Karnataka

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2130

Page 5 of 13

Validity 22.05.2018 to 21.05.2020

Last Amended on 22.06.2018

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability ($\pm$)	Remarks
10.	Temperature Simulation [#] K-Type Thermocouple J-Type Thermocouple T-Type Thermocouple R-Type Thermocouple S-Type Thermocouple N-Type Thermocouple RTD-Type	(-) 200°C to 1372°C (-) 200°C to 1200°C (-) 250°C to 400°C (-) 0°C to 1750°C (-) 0°C to 1750°C (-) 200°C to 1300°C (-) 200°C to 800°C	0.47 °C 0.32 °C 0.74 °C 0.67 °C 0.56 °C 0.47 °C 0.37 °C	Using Calibrator Fluke 5522A by Direct Method
11.	Oscilloscope [#] Amplitude (Dc Signal) Bandwidth Time Marker	1mV to 100mV 100mV to 190V 50kHz & 10MHz Ref 50kHz to 8GHz 8GHz to 18GHz 18GHz to 26 GHz 1ns to 5s	0.21 % to 0.06 % 0.06 % to 0.03 % 0.22 dB to 0.26 dB 0.26 dB to 0.35 dB 0.35 dB to 0.45 dB 0.0001 % to 0.00003 %	Using Oscilloscope Calibrator Fluke 9500B/9530 Heads, Bandwidth Signal Generator E8257D, MA2438A ML2444D Time Marker
12.	Frequency ^{\$} Frequency [*]	1Hz to 10MHz 10MHz to 20GHz 20 GHz to 40 GHz	0.0000058 % to 0.00000058 % 0.00000058 % to 0.00000029% 0.00000029% to 0.00000046%	Using Function Generator AFG 3252, Signal Generator E8257D By Direct Method
		1Hz to 10MHz 10MHz to 20GHz 20GHz to 40 GHz	0.00012 % to 0.00006 % 0.00006 % to 0.00035 % 0.00035 % to 0.00046 %	Using Function Generator AFG3252, Signal Generator MG3694C By Direct Method


Rjeshwar Kumar
Convener


Avijit Das
Program Manager



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Tektronix (India) Pvt. Ltd., Salarpuria Premia, Survey No. 16, Kadubeesana Halli, Varthur Hobli, Sarjapur Outer Ring Road, Bangalore , Karnataka

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2130

Page 6 of 13

Validity 22.05.2018 to 21.05.2020

Last Amended on 22.06.2018

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability ($\pm$)	Remarks
13.	Rf Power ^s	10MHz to 18GHz Impedance 50 Ω -50dBm to 12dBm 18GHz to 40 GHz Impedance 50 Ω -50 dBm to 12 dBm	5.5 % to 6.0 % 6.0 % to 8.7 %	Using Signal Generator E8257D Power meter with Sensor ML2438A, MA2444D by Direct Method
14.	Rf Attenuation [#]	10MHz to 18GHz 1dB to 50dB	0.25 dB to 0.45 dB	Using Signal Generator E8257D, Power Meter MA 2438A with Sensor ML 2444d, Step Attenuator 8494B, 8496B.
15.	VSWR ^s	10MHz to 18GHz Fixed Values 1.05(0.023), 1.10, 1.20, 1.50, 2.0(0.33)	0.095 LU to 0.40 LU	Using Maury Microwave 2611A Series by Direct Method
II.	MEASURE			
1.	Dc Voltage ^s	1mV to 10mV 10mV to 1V 1V to 1000V	0.0082% to 0.0052% 0.0052% to 0.0005% 0.0005% to 0.00053 %	Using DMM Fluke 8508A by Direct Method
	Dc Voltage [*]	10mV to 10V 1V to 1000V	0.04 % to 0.0019 % 0.001 % to 0.0035 %	Using DMM Fluke 8508A by Direct Method
2.	AC Voltage ^s	10 Hz to 40 Hz 1 mV to 10 mV 10 mV to 1 V 1 V to 100 V	0.16 % to 0.038 % 0.038 % to 0.01 % 0.01 % to 0.01 %	Using AC Measurement Standard Fluke 5790A by Direct Method

Rjeshwar Kumar
Convener

Avijit Das
Program Manager



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Tektronix (India) Pvt. Ltd., Salarpuria Premia, Survey No. 16, Kadubeesana Halli, Varthur Hobli, Sarjapur Outer Ring Road, Bangalore, Karnataka

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2130

Page 7 of 13

Validity 22.05.2018 to 21.05.2020

Last Amended on 22.06.2018

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability ($\pm$)	Remarks
		40 Hz to 10 kHz 1 mV to 10 mV 10 mV to 1 V 1 V to 1000 V	0.17 % to 0.023 % 0.023 % to 0.008 % 0.008 % to 0.015 %	
		10 kHz to 100 kHz 1 mV to 1 V 1 V to 10 V 10 V to 100 V	0.42 % to 0.016 % 0.016 % to 0.009 % 0.009 % to 0.03 %	
		100 kHz to 1 MHz 1 mV to 10 mV 10 mV to 10 V	0.44 % to 0.19 % 0.19 % to 0.14 %	
	Ac Voltage*	1 kHz 100 mV to 1 V 1 V to 10 V 10 V to 100 V 100 V to 700 V 700 V to 1000 V	0.21 % to 0.44 % 0.04 % to 0.011 % 0.011 % to 0.03 % 0.03 % to 0.05 % 2.03 % to 0.084 %	Using DMM Agilent 3458A /Fluke 8508A by Direct Method
3.	DC Current [#]	100 μ A to 1 mA 1 mA to 100 mA 100 mA to 1 A 1 A to 20 A >20 A to 100 A	0.045 % to 0.003 % 0.003 % to 0.005 % 0.005 % to 0.0027 % 0.0027 % to 0.005 % 0.32 % to 0.33 %	Using DMM Fluke 8508A/Fluke A40 Shunts by Direct Method Using Vaiseshika 9410


Rjeshwar Kumar
Convener


Avijit Das
Program Manager



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Tektronix (India) Pvt. Ltd., Salarpuria Premia, Survey No. 16, Kadubeesana Halli, Varthur Hobli, Sarjapur Outer Ring Road, Bangalore , Karnataka

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2130

Page 8 of 13

Validity 22.05.2018 to 21.05.2020

Last Amended on 22.06.2018

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability ($\pm$)	Remarks
4.	AC Current ^s	300Hz to 1kHz 100 μ A to 1mA	0.105 % to 0.056 %	Using DMM Fluke 8508A/Fluke A40 Shunts by Direct Method
		10Hz to 5kHz 1mA to 10mA 10mA to 1A 1A to 20A	0.036 % 0.036 % to 0.026 % 0.026 % to 0.080 %	
	Ac Current [*]	1kHz 100 μ A to 1mA 1mA to 100mA 100mA to 1A 11A to 20A	0.105 % to 0.06 % 0.06 % to 0.06 % 0.06 % to 0.15 % 0.15 % to 0.25 %	Using DMM Fluke 8508A by Direct Method
5.	Resistance [#]	1 Ω to 10 Ω 10 Ω to 100 Ω 100 Ω to 100k Ω 100k Ω to 10M Ω 10M Ω to 100M Ω 100M Ω to 1G Ω	0.0044 % to 0.0017 % 0.0017 % to 0.0015 % 0.0015 % to 0.0033 % 0.0033 % to 0.0040 % 0.0040 % to 0.0198 % 0.0198 % to 0.17 %	Using DMM Fluke 8508A by Direct Method
6.	Capacitance [#]	1kHz 10pF to 100pF 100pF to 10nF 10nF to 100nF 100nF to 1000nF 1000nF to 100 μ F	0.22 % to 0.21 % 0.21 % to 0.03 % 0.035 % to 0.042 % 0.042 % to 0.033 % 0.033 % to 0.32 %	Using RLC Digi Bridge Quadtech 1693 by Direct Method
7.	Inductance [#]	100 μ H to 10mH 10mH to 1H 1H to 10H	0.22 % to 0.23 % 0.23 % to 0.041 % 0.041 % to 0.22 %	Using RLC Digi Bridge Quadtech 1693 by Direct Method


Rjeshwar Kumar
Convener


Avijit Das
Program Manager



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Tektronix (India) Pvt. Ltd., Salarpuria Premia, Survey No. 16, Kadubeesana Halli, Varthur Hobli, Sarjapur Outer Ring Road, Bangalore , Karnataka

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2130

Page 9 of 13

Validity 22.05.2018 to 21.05.2020

Last Amended on 22.06.2018

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability ($\pm$)	Remarks
8.	Phase Angle ^s	± 180 Deg	0.03 Deg	Using phase meter clark Hess/6000 by Direct Method
9.	Period ^s	1ns to 5s	0.00008 %	Using Counter Tektronix MCA3040 by Direct Method
10.	Temperature Simulation [#] K-Type Thermocouple J-Type Thermocouple E-Type Thermocouple R-Type Thermocouple S-Type Thermocouple T-Type Thermocouple N-Type Thermocouple RTD-Type	(-)200 °C to 1372 °C (-)200 °C to 1200 °C (-)250 °C to 1000 °C 0 °C to 1750 °C 0 °C to 1750 °C (-)250 °C to 400 °C (-)200 °C to 1300 °C (-)200 °C to 800 °C	0.013 °C to 0.018 °C 0.012 °C to 0.014 °C 0.03 °C to 0.01 °C 0.23 °C to 0.1 °C 0.03 °C to 0.13 °C 0.015 °C 0.02 °C to 0.026 °C 0.001 °C to 0.014 °C	Using DMM Fluke 8508A/Zero Bath by Direct Method
11.	AC Power ^s	@ UPF, 50Hz 60V to 25 V, 0.1A to 20A 6W to 12W 12W to 4.8W @0.2pF, 50Hz 60V to 250V, 0.1A to 20A 1.2W to 2.4W 2.4W to 240W 240W to 960W @0.5 pF, 50 Hz 60V to 240V, 0.1A to 20A 6W to 60W 60W to 2.45W	2.05 % to 1.25 % 1.25 % to 0.15 % 1.19 % to 0.64 % 0.64 % to 0.154 % 0.154 % to 0.176 % 0.95 % to 0.37 % 0.37 % to 0.16 % 2.6 % to 0.367 %	Using Power Analyzer PA 4000 by Direct Method


Rjeshwar Kumar
Convener


Avijit Das
Program Manager



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Tektronix (India) Pvt. Ltd., Salarpuria Premia, Survey No. 16, Kadubeesana Halli, Varthur Hobli, Sarjapur Outer Ring Road, Bangalore , Karnataka

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2130

Page 10 of 13

Validity 22.05.2018 to 21.05.2020

Last Amended on 22.06.2018

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability ($\pm$)	Remarks
		@0.8 pF, 50Hz 60V to 240V, 0.1A to 20A 9.6W to 896W 96W to 960W 960W to 3.8kW	0.37 % to 0.15 % 0.15 % to 0.151 %	
12.	Frequency ^s	1Hz to 10MHz 10MHz to 20GHz 20GHz to 40GHz	0.0000058 % to 0.00000058 % 0.00000058 % to 0.00000029 % 0.00000029 % to 0.00000046 %	Using Rubidium Fluke 910R/ Counter Tek MCA 3040 counter/ Analyzer
	Frequency [*]	1Hz to 10MHz 10MHz to 20GHz 20GHz to 40 GHz	0.00015 % to 0.000065 % 0.000065 % to 0.000035 % 0.000035 % to 0.000049 %	Using Microwave Counter/Analyzer Tektronix MCA3040
13.	Rf Power ^s	10MHz to 18GHz Impedance 50 Ohm (-)50dBm to 12dBm 18MHz to 40GHz Impedance 50 Ohm (-)50dBm to 12dBm	5.5 % to 6.0 % 6.0 % to 8.7 %	Using Signal Generator E8257D, Power Meter with Sensor ML 2438A and MA2444D by Direct method
	Rf Power [*]	10MHz to 18GHz Impedance 50 Ohm (-)50dBm to 12dBm 18MHz to 40GHz Impedance 50 Ohm -50dBm to 12dBm	6.0 % to 7.0 % 7.0 % to 9.0 %	Using Signal Generator MG3694C, Power Meter with Sensor ML 2438A and MA2444D by Direct Method



Rjeshwar Kumar
Convener



Avijit Das
Program Manager



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory

Tektronix (India) Pvt. Ltd., Salarpuria Premia, Survey No. 16, Kadubeesana Halli, Varthur Hobli, Sarjapur Outer Ring Road, Bangalore, Karnataka

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2130

Page

11 of 13

Validity

22.05.2018 to 21.05.2020

Last Amended on 22.06.2018

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability ($\pm$)	Remarks
14.	RF Attenuation ^s	10MHz to 18GHz 1dB to 50dB 10MHz to 40GHz 1dB to 50dB	0.25 dB to 0.45 dB	Using Signal Generator E8257D Power Meter with Sensor Step Attenuator 8494B, 8496B by Direct Method
	RF Attenuation [*]	10MHz to 18GHz 1dB to 50dB	0.25 dB to 0.50 dB	Using Signal Generator MG3694C, Power Meter with Sensor Step Attenuator 8494B, 8496B by Direct Method
15.	VSWR ^s	10MHz to 18GHz Fixed Values 1.05(0.023), 1.10, 1.20, 1.50, 2.0(0.40)	0.085 LU to 0.38 LU	Using VSWR Bridge Anristu 560-97N-1, Maury Microwave Load 2611 A series by Direct Method
16.	Modulation [#] AM/FM	CF (Rate @ 1kHz) 1GHz (10 % to 90 %)	3.5 % to 3.6 %	Using Measuring receiver HP 8902A by Direct Method
	Frequency Modulation	CF (1GHz) 10kHz to $\leq$ 200kHz	6.0 % to 6.1 %	

Rjeshwar Kumar
Convener

Avijit Das
Program Manager



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Tektronix (India) Pvt. Ltd., Salarpuria Premia, Survey No. 16, Kadubeesana Halli, Varthur Hobli, Sarjapur Outer Ring Road, Bangalore , Karnataka

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2130

Page 12 of 13

Validity 22.05.2018 to 21.05.2020

Last Amended on 22.06.2018

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability ($\pm$)	Remarks
<u>THERMAL CALIBRATION</u>				
I.	TEMPERATURE			
1.	Temperature RTD, Thermocouple, (With Or Without Indicator) Temperature Gauge ^{\$}	0°C to 400°C 400°C to 700°	0.29 °C 1.68 °C	Using SPRT & 8 ½ DMM by Comparison Method Using S-Type Thermocouple & 8 ½ DMM by Comparison Method
2	Oven, Bath ,Dry well , Furnace ^{\$} (Single Point)	(-)40°C to 400°C 400°C to 1000°C	0.25 °C 1.74 °C	Using SPRT & 8 ½ DMM by Direct Method Using S-Type Thermocouple & 8 ½ DMM by Comparison Method
3.	Temperature Chamber, Oven, Bath* (Multipoint)	(-)70 °C to 180°C	1.52 °C	Using RTD With LOGGER 2638A by Direct Method


Rjeshwar Kumar
Convener


Avijit Das
Program Manager



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Tektronix (India) Pvt. Ltd., Salarpuria Premia, Survey No. 16, Kadubeesana Halli, Varthur Hobli, Sarjapur Outer Ring Road, Bangalore , Karnataka

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2130

Page 13 of 13

Validity 22.05.2018 to 21.05.2020

Last Amended on 22.06.2018

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability ($\pm$)	Remarks
4.	Specific Heat-Humidity , Humidity Chamber* (Centre of the Chamber)	20 % RH to 95%RH @25°C	1.55 % RH	Using Hygropalm Rotronic, HP22-A by direct Method

* Measurement Capability is expressed as an uncertainty ($\pm$) at a confidence probability of 95%

\$ Only in Permanent Laboratory

* Only for Site Calibration

The laboratory is also capable for site calibration however, the uncertainty at site depends on the prevailing actual environmental conditions and master equipment used.

Rjeshwar Kumar
Convener

Avijit Das
Program Manager