

Deutscher Kalibrierdienst (DKD)
Accreditation Body
represented in

Deutscher AkkreditierungsRat



Accreditation

The Accreditation Body of **Deutscher Kalibrierdienst** hereby accredits

TEKTRONIX GmbH
Heinrich-Pesch-Straße 9-11
50739 Köln

according to DIN EN ISO/IEC 17025:2005 for calibrations in the field / fields:

electrical DC and LF quantities, time and frequency

Part of the certificate is: Annex 05 (2 pages), 2009-12-15

DAR registration number: DKD-K-14401
DKD accredited since: 1995-07-10

Braunschweig, 2009-12-15

Head of Accreditation Body
by proxy

Dr. Martin Czaske



Deutscher Kalibrierdienst

Accreditation Body

DKD

Annex 05

of 2009-12-15 to the accreditation certificate of the calibration laboratory

Registration number:

DKD-K-14401

Page 1 of 2

at

TEKTRONIX GmbH
Heinrich-Pesch-Straße 9-11
50739 Köln
Germany

Phone: +49221 9477-264
Fax: +49221 9477-420
E-mail: ralf.riedel@tek.com

measured quantities:
DC voltage,
DC current,
DC resistance,
frequency,
rise time
oscilloscope

Head: Dipl.-Ing. Ralf Riedel
Deputy: Dipl.-Ing. Wolfgang Werner

Accredited since: 1995-07-10

Permanent Laboratory

measured quantity / calibration item	range	measurement conditions / procedure	best measurement capability ¹⁾	remarks
DC voltage voltage sources	10 mV to 100 mV >100 mV to 1 V >1 V to 10 V >10 V to 100 V >100 V to 1100 V		$5 \cdot 10^{-8} \cdot U + 3 \mu\text{V}$ $4 \cdot 10^{-8} \cdot U + 9 \mu\text{V}$ $5 \cdot 10^{-8} \cdot U + 15 \mu\text{V}$ $5 \cdot 10^{-8} \cdot U + 0,15 \text{ mV}$ $7 \cdot 10^{-8} \cdot U + 1,5 \text{ mV}$	$U = \text{measured value}$
measuring instruments	10 mV to 220 mV >220 mV to 2,2 V >2,2 V to 22 V >22 V to 220 V >220 V to 1100 V		$4 \cdot 10^{-8} \cdot U + 2 \mu\text{V}$ $5 \cdot 10^{-8} \cdot U + 3 \mu\text{V}$ $5 \cdot 10^{-8} \cdot U + 10 \mu\text{V}$ $5 \cdot 10^{-8} \cdot U + 0,15 \text{ mV}$ $6 \cdot 10^{-8} \cdot U + 1 \text{ mV}$	
DC current current sources	1 µA to 100 µA >100 µA to 1 mA >1 mA to 10 mA >10 mA to 100 mA >100 mA to 1 A >1 A to 10 A		$50 \cdot 10^{-6} \cdot I + 9 \text{ nA}$ $50 \cdot 10^{-6} \cdot I + 80 \text{ nA}$ $60 \cdot 10^{-6} \cdot I + 0,5 \mu\text{A}$ $70 \cdot 10^{-6} \cdot I + 2 \mu\text{A}$ $0,24 \cdot 10^{-3} \cdot I + 10 \mu\text{A}$ $0,25 \cdot 10^{-3} \cdot I + 0,3 \text{ mA}$	$I = \text{measured value}$
measuring instruments	1 µA to 220 µA >220 µA to 2,2 mA >2,2 mA to 22 mA >22 mA to 220 mA >220 mA to 2,2 A >2,2 A to 11 A		$45 \cdot 10^{-6} \cdot I + 7 \text{ nA}$ $43 \cdot 10^{-6} \cdot I + 70 \text{ nA}$ $50 \cdot 10^{-6} \cdot I + 0,4 \mu\text{A}$ $65 \cdot 10^{-6} \cdot I + 0,5 \mu\text{A}$ $0,2 \cdot 10^{-3} \cdot I + 4 \mu\text{A}$ $0,21 \cdot 10^{-3} \cdot I + 0,14 \text{ mA}$	
DC resistance resistance	1 Ω to <10 Ω 10 Ω to <100 Ω 100 Ω to <1 MΩ 1 MΩ to <10 MΩ 10 MΩ to 100 MΩ		$20 \cdot 10^{-8} \cdot R$ $15 \cdot 10^{-8} \cdot R$ $25 \cdot 10^{-8} \cdot R$ $40 \cdot 10^{-8} \cdot R$ $0,1 \cdot 10^{-3} \cdot R$	$R = \text{measured value}$
measuring instruments	1 Ω 1,9 Ω 10 Ω; 19 Ω 100 Ω to 190 kΩ 1 MΩ; 1,9 MΩ 10 MΩ; 19 MΩ 100 MΩ		$90 \cdot 10^{-6} \cdot R$ $60 \cdot 10^{-6} \cdot R$ $20 \cdot 10^{-6} \cdot R$ $15 \cdot 10^{-6} \cdot R$ $75 \cdot 10^{-6} \cdot R$ $75 \cdot 10^{-6} \cdot R$ $0,10 \cdot 10^{-3} \cdot R$	decade values and factor 1,9

¹⁾ The best measurement capabilities are stated according to DKD-3 (EA-4/02). These are expanded uncertainties of measurement with a coverage probability of 95% and have a coverage factor of $k = 2$ unless stated otherwise. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.

measured quantity / calibration item	range	measurement conditions / procedure	best measurement capability ¹⁾	remarks
frequency	1 MHz to 10 MHz 0,1 Hz to 2 GHz		$1 \cdot 10^{-11} \cdot f$ $1 \cdot 10^{-8} \cdot f + u_{tr}$	$f =$ measured value in steps of 1 MHz
rise time		periodical signal pulse amplitudes: 250 mV to 1 V 10 mV to 1 V		$t_r =$ measured value
generation measurement	14 ps to 10 ms 18 ps to 10 ms		$3 \cdot 10^{-2} \cdot t_r + 4 \text{ ps}$ $3 \cdot 10^{-2} \cdot t_r + 6 \text{ ps}$	
oscilloscope deflection vertical horizontal	50 mV to 50 V 0,5 ns to 5 s	square wave voltage at 1 kHz	$3 \cdot 10^{-3} \cdot U$ $4 \cdot 10^{-6} \cdot t$	
rise time	18 ps to 10 ms		$3 \cdot 10^{-2} \cdot t_r + 6 \text{ ps}$	
bandwidth	$\leq 1 \text{ GHz}$ $> 1 \text{ GHz}$ to 2 GHz		$5 \cdot 10^{-2} \cdot b$ $6 \cdot 10^{-2} \cdot b$	calculated: $0,34 = t_r \cdot b$ $b =$ bandwidth $t_r =$ rise time
oscilloscope with oscilloscope -calibrator DC voltage	0,0 V >0 mV to <0,1 V 0,1 V to 1 V >1 V to 5,56 V 5,6 V to 222,4 V	into 50 Ω, 1 MΩ into 1 MΩ	$15 \mu\text{V}$ $0,05 \cdot 10^{-2} \cdot U + 26 \mu\text{V}$ $0,022 \cdot 10^{-2} \cdot U + 65 \mu\text{V}$ $0,026 \cdot 10^{-2} \cdot U + 50 \mu\text{V}$ $0,03 \cdot 10^{-2} \cdot U$	Fluke 9500
sine-Frequency response reference: 50 kHz to 10 MHz		into 50 Ω		Fluke 9500/9530 (1 Hz to 3,2 GHz)
1 Hz to 100 MHz 100 MHz to 550 MHz 550 MHz to 1,1 GHz 1,1 GHz to 2,5 GHz 2,5 GHz to 3,2 GHz	4,4 mV to 5,6 V 4,4 mV to 3,4 V 4,4 mV to 3,4 V 4,4 mV to 2,2 V 4,4 mV to 2,2 V		0,22 dB 0,29 dB 0,37 dB 0,48 dB 0,48 dB	
AC voltage (sine)		into 50 Ω		
1 Hz to 550 MHz 550 MHz to 2,5 GHz 2,5 GHz to 3,2 GHz	4,4 mV to 5,6 V 4,4 mV to 3,4 V 4,4 mV to 2,2 V		$3,3 \cdot 10^{-2} \cdot U$ $6,3 \cdot 10^{-2} \cdot U$ $1,1 \cdot 10^{-1} \cdot U$	Fluke 9500/9530 (1 Hz to 3,2 GHz)
resistance measurements	50 Ω 1 MΩ		$0,13 \cdot 10^{-2} \cdot R$ $0,12 \cdot 10^{-2} \cdot R$	Fluke 9500/9530
DC voltage measurements	0 V to 5 V		$1,1 \cdot 10^{-3} \cdot U + 3 \text{ mV}$	Keithley 2000
frequency / period measurements	12 kHz to 3,2 GHz		$0,27 \cdot 10^{-6}$	Fluke 9500 opt. 100

¹⁾ The best measurement capabilities are stated according to DKD-3 (EA-4/02). These are expanded uncertainties of measurement with a coverage probability of 95% and have a coverage factor of $k=2$ unless stated otherwise. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.