

# Instructions



## **AWG50BUP and AWG70BUP Upgrades for AWG5000B and AWG7000B Series Instruments**

**075-0971-00**

[www.tektronix.com](http://www.tektronix.com)



075097100

Copyright © Tektronix. All rights reserved. Licensed software products are owned by Tektronix or its subsidiaries or suppliers, and are protected by national copyright laws and international treaty provisions.

Tektronix products are covered by U.S. and foreign patents, issued and pending. Information in this publication supercedes that in all previously published material. Specifications and price change privileges reserved.

TEKTRONIX and TEK are registered trademarks of Tektronix, Inc.

### **Contacting Tektronix**

Tektronix, Inc.  
14200 SW Karl Braun Drive or P.O. Box 500  
Beaverton, OR 97077 USA

For product information, sales, service, and technical support:

- In North America, call 1-800-833-9200
- Worldwide, visit [www.tektronix.com](http://www.tektronix.com) to find contacts in your area.

# Kit Description

This kit provides instructions to install an option on an AWG5000B or AWG7000B series instrument. Once installed, the upgrade is equivalent to a factory-installed option.

For a list of available options, see the Tektronix Web site ([www.tektronix.com](http://www.tektronix.com)), or your Tektronix representative.

## Product

**AWG5000B and AWG7000B series** B010100 and above

## Kit Parts List

Option	Quantity	Part number	Description
All	1 each	NS	LABEL, OPTION INFORMATION
All	1 each	075-0971-00	MANUAL, KIT INSTRUCTIONS

NS – Not Saleable

# Installation Instructions

To install your new option(s), you need the key number that is provided with this kit. This unique key number is valid only for the specific model number and serial number of your instrument; it will not work on any other instrument. This single key number enables all valid options for your instrument, those that were factory installed and any others that you may have already purchased. In other words, this new key number supercedes any previous key numbers that you may already have.



**CAUTION.** *To avoid losing your options, do not misplace the paper label that is included with this kit. This label contains the option installation key number and should be installed on the rear panel of the instrument for future reference.*

*As a backup, you may want to copy the key number and save it with the other important papers for your instrument. If you ever need to reenable the option, you will need to reenter this key number.*

## Install Option

Follow these steps to enable your new option(s):

1. Power on the instrument.
2. From the menu bar, select the **System** menu, and then select **Option Installation...**
3. The Option Installation dialog box is displayed.




Figure 1: Option Installation dialog box (1)

4. Click **Continue** to display the second dialog box.



**Figure 2: Option Installation dialog box (2)**

5. Follow the on-screen instructions to enter the alphanumeric key number exactly as it is printed on the label.
6. Attach the label to the rear panel over the existing label.

## Verify Option Upgrade

Follow these steps to verify that an option upgrade was successful:

1. Restart the instrument application software.
2. Select the **Help** menu, and then select **About Tektronix AWG...**
3. Verify that the option that you installed is in the Options list (see Figure 3).

The example in Figure 3 shows that option 01 is installed:




Figure 3: About Tektronix AWG

❖ End of document ❖