

DPO7AFP Auxiliary Front Panel

Instructions



DPO7AFP Auxiliary Front Panel

The DPO7AFP is a plug-in panel that provides physical knobs and buttons to control the DPO7000SX series oscilloscope. To connect the DPO7AFP:

1. Close the TekScope application.

NOTE. *The TekScope application must be closed before connecting the DPO7AFP.*

2. Plug the DPO7AFP into any USB port on the DPO7000SX. Wait until the Windows OS recognizes and loads the required drivers.
3. Start the TekScope application. The panel LEDs will light when TekScope connects to the panel.
4. To verify that the DPO7AFP is working, push any channel button and observe that the TekScope application enables or disables the selected channel.

NOTE. *If you disconnect the DPO7AFP while TekScope is running, you must close TekScope, reconnect the DPO7AFP, and restart TekScope to use the DPO7AFP.*

Standard accessories

Accessory	Tektronix part number
Cable, USB 2.0, USB-A Male to USB-B Male; 2 m (6 feet)	174-6053-xx
Instructions manual (this document)	071-3401-xx

Warranty information

This product is warranted for one year. For warranty information, go to www.tektronix.com/warranty.

Environmental considerations

Characteristics	Description
Restriction of hazardous substances	This product is classified as an industrial monitoring and control instrument accessory, and is not required to comply with the substance restrictions of the recast RoHS Directive 2011/65/EU until July 22, 2017.



Equipment Recycling. This product complies with the European Union's requirements according to Directive 2012/19/EU on waste electrical and electronic equipment (WEEE). For more information about recycling options, check the Support/Service section of the Tektronix Web site (www.tektronix.com).

Compliance information

Refer to the DPO7000SX user documentation for information on instrument and accessory compliance.

Contacting Tektronix

Web site: www.tektronix.com

Phone: 1-800-833-9200

Address: Tektronix, Inc.
Department or name (if known)
14200 SW Karl Braun Drive
P.O. Box 500
Beaverton, OR 97077 USA

Email: techsupport@tektronix.com



071-3401-00

Tektronix

