

**RSA6100B Series**  
**Real-Time Signal Analyzers**  
**Printable Help**



077-1023-02

**Tektronix**



**RSA6100B Series**  
**Real-Time Signal Analyzers**  
**Printable Help**

Copyright © Tektronix. All rights reserved. Licensed software products are owned by Tektronix or its subsidiaries or suppliers, and are protected by national copyright laws and international treaty provisions.

Tektronix products are covered by U.S. and foreign patents, issued and pending. Information in this publication supersedes that in all previously published material. Specifications and price change privileges reserved.

TEKTRONIX and TEK are registered trademarks of Tektronix, Inc.

Planar Crown is a registered trademark of Aeroflex Inc.

Instrument help part number: 076-0370-02

### **Contacting Tektronix**

Tektronix, Inc.  
14150 SW Karl Braun Drive  
P.O. Box 500  
Beaverton, OR 97077  
USA

For product information, sales, service, and technical support:

- In North America, call 1-800-833-9200.
- Worldwide, visit [www.tektronix.com](http://www.tektronix.com) to find contacts in your area.

# Table of Contents

## Welcome

|               |   |
|---------------|---|
| Welcome ..... | 1 |
|---------------|---|

## About the Tektronix Signal Analyzer

|                            |   |
|----------------------------|---|
| Product software .....     | 3 |
| Accessories                |   |
| Standard accessories ..... | 3 |
| Options                    |   |
| Options .....              | 5 |
| Documentation and support  |   |
| Documentation .....        | 5 |
| Video tutorials .....      | 6 |

## Orientation

|                                      |    |
|--------------------------------------|----|
| Front panel connectors .....         | 9  |
| Front-Panel Controls .....           | 9  |
| Touch Screen .....                   | 13 |
| Touch-Screen Actions .....           | 13 |
| Elements of the Display .....        | 15 |
| Rear-Panel Connectors .....          | 19 |
| Setting Up Network Connections ..... | 19 |

## Operating Your Instrument

|                                  |    |
|----------------------------------|----|
| Restoring Default Settings ..... | 21 |
| Running Alignments .....         | 21 |
| Presets .....                    | 22 |
| Setting Options .....            | 29 |
| Operating System Restore .....   | 31 |

## Using the Measurement Displays

|                          |    |
|--------------------------|----|
| Selecting Displays ..... | 33 |
|--------------------------|----|

## Taking Measurements

|                              |    |
|------------------------------|----|
| Measurements                 |    |
| Available Measurements ..... | 35 |

## General Signal Viewing

|                |    |
|----------------|----|
| Overview ..... | 45 |
|----------------|----|

|  |     |
|--|-----|
| DPX  |     |
| DPX Primer .....   | 45  |
| DPX Display Overview .....                               | 67  |
| DPX Display .....  | 67  |
| DPX Settings .....                                       | 76  |
| Time Overview  |     |
| Time Overview Display .....                              | 91  |
| Time Overview Settings .....                             | 93  |
| Navigator View .....                                     | 95  |
| Spectrum   |     |
| Spectrum Display .....                                   | 98  |
| Spectrum Settings.....                                   | 100 |
| Spectrogram  |     |
| Spectrogram Display .....                                | 101 |
| Spectrogram Settings.....                                | 104 |
| Amplitude Vs Time  |     |
| Amplitude Vs Time Display .....                          | 108 |
| Amplitude Vs Time Settings .....                         | 109 |
| Frequency Vs Time  |     |
| Frequency Vs Time Display .....                          | 110 |
| Frequency Vs Time Settings.....                          | 111 |
| Phase Vs Time  |     |
| Phase Vs Time Display.....                               | 112 |
| Phase Vs Time Settings .....                             | 113 |
| RF I & Q Vs Time   |     |
| RF I & Q vs Time Display .....                           | 114 |
| RF I & Q vs Time Settings.....                           | 115 |
| Common Controls for General Signal Viewing Displays      |     |
| General Signal Viewing Shared Measurement Settings ..... | 116 |

## Analog Modulation

|                   |     |
|-------------------|-----|
| Overview .....    | 129 |
| AM                |     |
| AM Display .....  | 129 |
| AM Settings ..... | 130 |
| FM                |     |
| FM Display .....  | 135 |
| FM Settings.....  | 137 |
| PM                |     |
| PM Display .....  | 143 |
| PM Settings.....  | 145 |

## RF Measurements

|   |     |
|---|-----|
| Overview .....  | 153 |
| Channel Power and ACPR  |     |
| Channel Power and ACPR (Adjacent Channel Power Ratio) Display ..... | 153 |
| Channel Power and ACPR Settings .....                               | 156 |
| MCPR  |     |
| MCPR (Multiple Carrier Power Ratio) Display .....                   | 159 |
| MCPR Settings .....   | 163 |
| Occupied BW & x dB BW   |     |
| Occupied BW & x dB BW Display .....                                 | 170 |
| Occupied BW & x dB BW Settings .....                                | 173 |
| Spurious  |     |
| Spurious Display .....  | 174 |
| Spurious Display Settings .....                                     | 178 |
| CCDF  |     |
| CCDF Display .....  | 185 |
| CCDF Settings .....   | 186 |
| Phase Noise   |     |
| Phase Noise Display .....   | 187 |
| Phase Noise Settings .....  | 192 |
| Settling Time Measurements  |     |
| Settling Time Measurement Overview .....                            | 194 |
| Settling Time Displays  |     |
| Settling Time Displays .....  | 198 |
| Settling Time Settings .....  | 206 |
| Common Controls for Settling Time Displays                          |     |
| Settling Time Displays Shared Measurement Settings .....            | 206 |
| SEM (Spectrum Emission Mask)  |     |
| SEM Display .....   | 214 |
| Spectrum Emission Mask Settings .....                               | 218 |
| Common Controls for RF Measurements Displays                        |     |
| RF Measurements Shared Measurement Settings .....                   | 225 |

## WLAN Measurements

|                                      |     |
|--------------------------------------|-----|
| WLAN Overview .....                  | 233 |
| WLAN Chan Response                   |     |
| WLAN Channel Response Display .....  | 235 |
| WLAN Channel Response Settings ..... | 237 |
| WLAN Constellation                   |     |
| WLAN Constellation Display .....     | 238 |
| WLAN Constellation Settings .....    | 239 |
| WLAN EVM                             |     |

|   |     |
|---|-----|
| WLAN EVM Display .....                          | 240 |
| WLAN EVM Settings.....                          | 241 |
| WLAN Mag Error                                  |     |
| WLAN Magnitude Error Display .....              | 242 |
| WLAN Magnitude Error Settings.....              | 243 |
| WLAN Phase Error                                |     |
| WLAN Phase Error Display .....                  | 244 |
| WLAN Phase Error Settings.....                  | 245 |
| WLAN Power vs Time                              |     |
| WLAN Power vs Time Display .....                | 246 |
| WLAN Power vs Time Settings.....                | 248 |
| WLAN Spectral Flatness                          |     |
| WLAN Spectral Flatness Display.....             | 249 |
| WLAN Spectral Flatness Settings .....           | 250 |
| WLAN Summary                                    |     |
| WLAN Summary Display .....                      | 251 |
| WLAN Summary Settings.....                      | 255 |
| WLAN Symb Table                                 |     |
| WLAN Symbol Table Display .....                 | 256 |
| WLAN Symbol Table Settings .....                | 258 |
| Common Controls for WLAN Analysis Displays      |     |
| WLAN Analysis Shared Measurement Settings ..... | 258 |

## OFDM Analysis

|                                      |     |
|--------------------------------------|-----|
| Overview .....                       | 269 |
| OFDM Chan Response                   |     |
| OFDM Channel Response Display .....  | 269 |
| OFDM Channel Response Settings.....  | 271 |
| OFDM Constellation                   |     |
| OFDM Constellation Display .....     | 272 |
| OFDM Constellation Settings .....    | 273 |
| OFDM EVM                             |     |
| OFDM EVM Display .....               | 273 |
| OFDM EVM Settings .....              | 274 |
| OFDM Flatness                        |     |
| OFDM Spectral Flatness Display ..... | 275 |
| OFDM Spectral Flatness Settings..... | 276 |
| OFDM Mag Error                       |     |
| OFDM Magnitude Error Display .....   | 277 |
| OFDM Magnitude Error Settings .....  | 278 |
| OFDM Phase Error                     |     |
| OFDM Phase Error Display .....       | 279 |

|   |     |
|---|-----|
| OFDM Phase Error Settings .....                 | 280 |
| OFDM Power                                      |     |
| OFDM Power Display .....                        | 281 |
| OFDM Power Settings .....                       | 282 |
| OFDM Summary                                    |     |
| OFDM Summary Display .....                      | 283 |
| OFDM Summary Settings .....                     | 284 |
| OFDM Symb Table                                 |     |
| OFDM Symbol Table Display .....                 | 285 |
| OFDM Symbol Table Settings .....                | 286 |
| Common Controls for OFDM Analysis Displays      |     |
| OFDM Analysis Shared Measurement Settings ..... | 286 |

## Noise Figure and Gain Measurements

|  |     |
|--|-----|
| Overview .....                                     | 295 |
| Noise Table Display                                |     |
| Noise Table Display .....                          | 296 |
| Noise Table Settings .....                         | 297 |
| Gain Display                                       |     |
| Gain Display .....                                 | 298 |
| Gain Settings .....                                | 299 |
| Noise Figure                                       |     |
| Noise Figure Display .....                         | 300 |
| Noise Figure Settings .....                        | 301 |
| Noise Temperature                                  |     |
| Noise Temperature Display .....                    | 302 |
| Noise Temperature Settings .....                   | 303 |
| Y Factor   |     |
| Y Factor Display .....                             | 304 |
| Y Factor Settings .....                            | 304 |
| Uncertainty Calculator                             |     |
| Uncertainty Calculator Display .....               | 305 |
| Uncertainty Calculator Settings .....              | 307 |
| Common Controls for Noise Figure and Gain Displays |     |
| Noise Figure and Gain Measurement Settings .....   | 309 |

## Pulsed RF

|                            |     |
|----------------------------|-----|
| Overview .....             | 319 |
| Pulse Table Display        |     |
| Pulse Table Display .....  | 319 |
| Pulse Table Settings ..... | 320 |
| Pulse Trace Display        |     |

|   |     |
|---|-----|
| Pulse Trace Display.....                    | 321 |
| Pulse Trace Settings .....                  | 323 |
| Pulse Statistics                            |     |
| Pulse Statistics Display.....               | 323 |
| Pulse Statistics Settings .....             | 325 |
| Common Controls for Pulsed RF Displays      |     |
| Pulsed RF Shared Measurement Settings ..... | 326 |

## **APCO P25 Analysis**

|   |     |
|---|-----|
| APCO P25 Analysis.....                        | 337 |
| Overview .....                                | 337 |
| P25 Standards Presets .....                   | 338 |
| P25 Displays .....                            | 339 |
| P25 Measurements.....                         | 340 |
| P25 Test Patterns.....                        | 348 |
| P25 Constellation                             |     |
| P25 Constellation Display .....               | 351 |
| P25 Constellation Settings.....               | 353 |
| P25 Eye Diagram                               |     |
| P25 Eye Diagram Display .....                 | 354 |
| P25 Eye Diagram Settings.....                 | 355 |
| P25 Power vs Time                             |     |
| P25 Power vs Time Display .....               | 356 |
| P25 Power vs Time Settings.....               | 357 |
| P25 Summary                                   |     |
| P25 Summary Display .....                     | 358 |
| P25 Summary Settings .....                    | 361 |
| P25 Symbol Table                              |     |
| P25 Symbol Table Display.....                 | 361 |
| P25 Symbol Table Settings .....               | 363 |
| P25 Frequency Deviation Vs Time               |     |
| P25 Frequency Dev vs Time Display.....        | 364 |
| P25 Frequency Dev Vs Time Settings.....       | 366 |
| Common Controls for P25 Analysis Displays     |     |
| P25 Analysis Shared Measurement Settings..... | 367 |

## **LTE Analysis**

|  |     |
|--|-----|
| LTE Analysis.....                      | 379 |
| Overview .....                         | 379 |
| LTE Standards preset test setups ..... | 380 |
| LTE displays.....                      | 385 |
| LTE measurements .....                 | 385 |

|   |     |
|---|-----|
| LTE Status Messages .....                 | 392 |
| LTE ACLR                                  |     |
| LTE ACLR display .....                    | 392 |
| LTE ACLR Settings .....                   | 394 |
| LTE Channel Spectrum                      |     |
| LTE Channel Spectrum display .....        | 395 |
| LTE Channel Spectrum Settings .....       | 397 |
| LTE Constellation                         |     |
| LTE Constellation display .....           | 397 |
| LTE Constellation Settings .....          | 399 |
| LTE Power vs Time                         |     |
| LTE Power vs Time display .....           | 399 |
| LTE Power vs Time Settings .....          | 401 |
| Common Controls for LTE Analysis Displays |     |
| LTE Analysis Measurement Settings .....   | 401 |

## Bluetooth Analysis

|  |     |
|--|-----|
| Bluetooth Analysis .....                                   | 419 |
| Overview .....   | 419 |
| Bluetooth Standards presets .....                          | 420 |
| Bluetooth displays .....                                   | 423 |
| Bluetooth measurements and test setups .....               | 423 |
| Bluetooth Status Messages .....                            | 428 |
| Bluetooth (BT) Constellation                               |     |
| Bluetooth Constellation display .....                      | 429 |
| BT Constellation Settings .....                            | 431 |
| Bluetooth (BT) Eye Diagram                                 |     |
| Bluetooth Eye Diagram display .....                        | 431 |
| BT Eye Diagram Settings .....                              | 433 |
| Bluetooth (BT) CF Offset and Drift                         |     |
| Bluetooth Carrier Frequency Offset and Drift display ..... | 434 |
| BT CF Offset and Drift Settings .....                      | 436 |
| Bluetooth (BT) Summary                                     |     |
| Bluetooth Summary display .....                            | 436 |
| BT Summary Settings .....                                  | 441 |
| Bluetooth (BT) Symbol Table                                |     |
| Bluetooth Symbol Table display .....                       | 441 |
| BT Symbol Table Settings .....                             | 444 |
| Bluetooth (BT) Frequency Deviation Vs Time                 |     |
| Bluetooth Frequency Dev vs Time display .....              | 445 |
| BT Frequency Dev Vs Time Settings .....                    | 449 |
| Bluetooth (BT) 20dB Bandwidth                              |     |

|   |     |
|---|-----|
| Bluetooth 20dB Bandwidth display .....          | 449 |
| BT 20dB BW settings .....                       | 451 |
| Bluetooth (BT) InBand Emission                  |     |
| Bluetooth InBand Emission display .....         | 452 |
| Common Controls for Bluetooth Analysis Displays |     |
| Bluetooth Analysis Measurement Settings.....    | 454 |

## Audio Analysis

|   |     |
|---|-----|
| Overview .....                              | 469 |
| Audio Spectrum                              |     |
| Audio Spectrum Display .....                | 469 |
| Audio Spectrum Settings.....                | 470 |
| Audio Summary                               |     |
| Audio Summary Display.....                  | 471 |
| Audio Summary Settings .....                | 472 |
| Common Controls for Audio Analysis Displays |     |
| Audio Analysis Measurement Settings.....    | 473 |

## GP Digital Modulation

|   |     |
|---|-----|
| Overview .....                            | 483 |
| Constellation                             |     |
| Constellation Display.....                | 484 |
| Constellation Settings .....              | 485 |
| Demod I & Q vs Time                       |     |
| Demod I & Q vs Time Display.....          | 486 |
| Demod I & Q vs Time Settings .....        | 488 |
| EVM vs Time                               |     |
| EVM vs Time Display.....                  | 488 |
| EVM vs Time Settings .....                | 489 |
| Eye Diagram                               |     |
| Eye Diagram Display.....                  | 490 |
| Eye Diagram Settings .....                | 492 |
| Frequency Deviation vs Time               |     |
| Frequency Deviation vs Time Display ..... | 492 |
| Frequency Deviation vs Time Settings..... | 494 |
| Magnitude Error vs Time                   |     |
| Magnitude Error vs Time Display .....     | 494 |
| Magnitude Error vs Time Settings .....    | 496 |
| Phase Error vs Time                       |     |
| Phase Error vs Time Display .....         | 496 |
| Phase Error vs. Time Settings .....       | 498 |
| Signal Quality                            |     |

|  |     |
|--|-----|
| Signal Quality Display .....                                   | 498 |
| Signal Quality Settings .....                                  | 503 |
| Symbol Table   |     |
| Symbol Table Display .....                                     | 504 |
| Symbol Table Settings .....                                    | 505 |
| Trellis Diagram  |     |
| Trellis Diagram Display .....                                  | 505 |
| Trellis Diagram Settings .....                                 | 507 |
| Common Controls for GP Digital Modulation Displays             |     |
| GP Digital Modulation Shared Measurement Settings .....        | 508 |
| Standard Settings Button .....                                 | 509 |
| Symbol Maps  |     |
| Symbol Maps .....  | 526 |
| User Filters   |     |
| Overview: User Defined Measurement and Reference Filters ..... | 532 |
| User Filter File Format .....                                  | 533 |

## Marker Measurements

|   |     |
|---|-----|
| Using Markers   |     |
| Using Markers .....   | 537 |
| Controlling Markers with the Touchscreen Actions Menu .....               | 538 |
| Measuring Signal Density, Frequency and Power on a DPX Bitmap Trace ..... | 538 |
| Measuring Frequency and Power in the Spectrum Display .....               | 540 |
| Common Marker Actions   |     |
| Marker Action Controls .....  | 541 |
| Peak .....  | 541 |
| Next Peak .....   | 541 |
| Marker to Center Frequency .....  | 541 |
| Define Markers Control Panel  |     |
| Enabling Markers and Setting Marker Properties .....                      | 541 |
| Markers Toolbar   |     |
| Using the Markers Toolbar .....   | 543 |
| Noise Markers in the Spectrum Display                                     |     |
| Measuring Noise Using Delta Markers in the Spectrum Display .....         | 544 |

## Mask Testing

|                              |     |
|------------------------------|-----|
| The Mask Test Tool .....     | 547 |
| Mask Test Settings .....     | 547 |
| Define Tab (Mask Test) ..... | 547 |
| Actions Tab .....            | 553 |

## Analyzing Data

|                                   |     |
|-----------------------------------|-----|
| Analysis Settings                 |     |
| Analysis Settings.....            | 555 |
| Analysis Time Tab .....           | 555 |
| Spectrum Time Tab .....           | 557 |
| Frequency Tab.....                | 557 |
| Units Tab.....                    | 561 |
| Analyzing Data Using Replay       |     |
| Replay Overview .....             | 561 |
| Replay Menu .....                 | 564 |
| Acq Data.....                     | 564 |
| DPX Spectra.....                  | 564 |
| Replay All Selected Records ..... | 565 |
| Replay Current Record .....       | 565 |
| Replay from Selected.....         | 565 |
| Pause .....                       | 565 |
| Stop.....                         | 565 |
| Select All .....                  | 565 |
| Select Records from History.....  | 566 |
| Replay Toolbar .....              | 566 |

## Amplitude Corrections

|  |     |
|--|-----|
| Amplitude Settings .....               | 569 |
| Internal Settings Tab .....            | 569 |
| External Gain/Loss Correction Tab..... | 572 |
| External Gain Value .....              | 573 |
| Apply External Corrections To.....     | 573 |
| External Loss Tables .....             | 574 |
| External Probe Correction Tab .....    | 576 |

## Controlling the Acquisition of Data

|   |     |
|---|-----|
| Acquisition Controls in the Run Menu              |     |
| Continuous Versus Single Sequence.....            | 577 |
| Run .....   | 577 |
| Resume.....                                       | 577 |
| Abort .....                                       | 577 |
| Acquisition Controls in the Acquire Control Panel |     |
| The Acquire Control Panel .....                   | 578 |
| Sampling Parameters Tab .....                     | 579 |
| Advanced Tab (Acquire).....                       | 580 |
| FastSave .....                                    | 581 |
| FastSave Tab .....                                | 584 |

|  |     |
|--|-----|
| Acquire.....                                 | 584 |
| Using Triggers to Capture Just What You Want |     |
| Triggering                                   |     |
| Triggering.....                              | 586 |
| Frequency Mask Trigger .....                 | 589 |
| Mask Editor (Frequency Mask Trigger).....    | 590 |
| Trigger Settings .....                       | 593 |
| Event Tab .....                              | 594 |
| Time Qualified Tab .....                     | 603 |
| Advanced Tab (Triggering) .....              | 604 |
| Actions Tab (Triggering).....                | 605 |

## Managing Data, Settings, and Pictures

|  |     |
|--|-----|
| Saving and Recalling Data, Settings, and Pictures..... | 607 |
| Data, Settings, and Picture File Formats .....         | 609 |
| Printing Screen Shots .....                            | 613 |

## Reference

|  |     |
|--|-----|
| Online Help .....                                  | 615 |
| About the Tektronix RTSA .....                     | 615 |
| Connecting Signals                                 |     |
| Configure In/Out Settings.....                     | 616 |
| Connecting an RF Signal .....                      | 616 |
| Connecting a Signal Using a TekConnect Probe ..... | 618 |
| Connecting External Trigger Signals .....          | 619 |
| Digital I and Q Outputs .....                      | 619 |
| IF Output.....                                     | 620 |
| Other Outputs .....                                | 621 |
| Mapping Measurements                               |     |
| Mapping Measurements.....                          | 622 |
| Menus  |     |
| Menu Overview .....                                | 622 |
| File Menu  |     |
| File Menu .....                                    | 623 |
| View Menu  |     |
| View Menu .....                                    | 627 |
| Run Menu   |     |
| Run Menu.....                                      | 629 |
| Replay   |     |
| Replay Menu .....                                  | 631 |
| Markers Menu                                       |     |
| Markers Menu.....                                  | 631 |

|  |     |
|--|-----|
| Setup Menu   |     |
| Setup Menu .....   | 631 |
| Presets Menu   |     |
| Presets Menu .....                                       | 631 |
| Tools Menu   |     |
| Tools Menu .....   | 632 |
| Window Menu  |     |
| Window Menu .....  | 634 |
| Help Menu  |     |
| Help Menu .....  | 634 |
| Troubleshooting  |     |
| Error and Information Messages .....                     | 635 |
| Displaying the Windows Event Viewer .....                | 643 |
| Dealing with Sluggish Instrument Operation .....         | 645 |
| On/Standby Switch  |     |
| On/Standby Switch .....                                  | 645 |
| Upgrading the Product Software                           |     |
| How to Find Out if Software Upgrades are Available ..... | 646 |
| Changing Settings  |     |
| Settings .....   | 646 |
| Remote Login   |     |
| Remote Login .....                                       | 648 |

## Glossary

## Index

## Welcome

This help provides in-depth information on how to use the RSA6100B Series Real-Time Spectrum Analyzers. For a shorter introduction to the Signal Analyzer, refer to the *RSA6100B Series Real Time Signal Analyzer Quick Start User Manual*. To see tutorial examples of how to use your analyzer to take measurements in different application areas, refer to the *RSA6100B Series Real Time Signal Analyzer Application Examples Reference*.



## Product software

The instrument includes the following software:

- **System software:** The RSA6100B Series product software runs on a specially configured version of Windows 7. As with standard Windows 7 installations, you can install other compatible applications, but the installation and use of non-Tektronix software is not supported by Tektronix. If you need to reinstall the operating system, follow the operating system restore procedure provided in [Operating System Restore \(see page 31\)](#). Do not substitute any version of Windows that is not specifically provided by Tektronix for use with your instrument.
- **Product software:** The product software is the instrument application. It provides the user interface (UI) and all other instrument control functions. You can minimize or even exit/restart the instrument application as your needs dictate.

Occasionally, new versions of software for your instrument may become available at our Web site. Visit [www.tektronix.com/software](http://www.tektronix.com/software) for information.

### Software and hardware upgrades

Tektronix may offer software or hardware upgrade kits for this instrument. Contact your local Tektronix distributor or sales office for more information.

### Software upgrades

Tektronix may offer upgrade kits for this software. Contact your local Tektronix distributor or sales office for more information.

## Standard accessories

The standard accessories for the instruments are shown below. For the latest information on available accessories, see the [Tektronix Web site](#).

### Quick Start User Manual

- English - Option L0, Tektronix part number 071-3359-XX
- Japanese - Option L5, Tektronix part number 071-3361-XX
- Russian, Option L10, Tektronix part number 071-3364-XX
- Simplified Chinese - Option L7, Tektronix part number 071-3362-XX

## Application Examples Reference

- English – Tektronix part number 071-2834-XX
- Simplified Chinese - Option L7, Tektronix part number 071-2835-XX
- Japanese - Option L5, Tektronix part number 071-2836-XX
- Russian, Option L10, Tektronix part number 071-2837-XX

## Product Documentation CD-ROM

The Product Documentation CD-ROM contains a collection documentation available for your product, in PDF format. See [Documentation \(see page 5\)](#) for a list of the documents included on the CD-ROM.

---

**NOTE.** To check for updates to the instrument documentation, browse to [www.tektronix.com/manuals](http://www.tektronix.com/manuals) and search by your instrument's model number.

---

## Important documents folder

- Certificate of Calibration documenting NIST traceability, 2540-1 compliance, and ISO9001 registration

## Power cords

- North America - Option A0, Tektronix part number 161-0104-00
- Universal Euro - Option A1, Tektronix part number 161-0104-06
- United Kingdom - Option A2, Tektronix part number 161-0104-07
- Australia - Option A3, Tektronix part number 161-0104-05
- 240V North America - Option A4, Tektronix part number 161-0104-08
- Switzerland - Option A5, Tektronix part number 161-0167-00
- Japan - Option A6, Tektronix part number 161-A005-00
- China - Option A10, Tektronix part number 161-0306-00
- India - Option A11, Tektronix part number 161-0324-00
- No power cord or AC adapter - Option A99

## Optical wheel mouse

## RF input connectors

- Planar Crown® RF input connector - Type N (RSA6106B, RSA6114B), Tektronix part number 131-4329-00
- Planar Crown® RF input connector - 3.5 mm female (RSA6120B), Tektronix part number 131-9062-00

# Options

To view a listing of the software options installed on your instrument, select **Help > About Your Tektronix Real-Time Analyzer**. There is a label on the rear-panel of the instrument that lists installed hardware options.

Options can be added to your instrument. For the latest information on available option upgrades, see [Tektronix Web site](#).

# Documentation

In addition to the instrument Help, the following documents are available. Many documents are provided on the documentation CD provided with the instrument. For the most up to date documentation, visit the Tektronix Web site [www.tektronix.com/downloads](http://www.tektronix.com/downloads).

- **Quick Start User Manual** (071-3359-XX - English). This manual has information about installing and operating your instrument. This manual is also available in Simplified Chinese (071-3362-XX), Japanese (071-3361-XX), and Russian (071-3364-XX). These manuals are available in both print and a printable PDF file.
- **Application Examples Reference** (071-2834-XX). This manual provides examples of how to solve problems using an RSA5100A and RSA6100B Series Signal Analyzer. The Application Examples Reference is also available in Simplified Chinese (071-2835-XX), Japanese (071-2836-XX), and Russian (071-2837-XX). These are available in both print and a printable PDF file.
- **Programmer Manual** (077-0523-XX). This manual provides information to use commands for remotely controlling your instrument. This is available as a printable PDF file.
- **Service Manual** (077-0648-XX). This manual includes procedures to service the instrument to the module level. This is available as a printable PDF file.
- **Specifications and Performance Verification Technical Reference Manual** (077-0647-XX). This manual includes both the specifications and the performance verification procedures. This is available as a printable PDF file.
- **Declassification and Security Instructions** (077-0521-XX) This document helps customers with data security concerns to sanitize or remove memory devices from the instrument. This is available as a printable PDF file.

The most recent versions of the product documentation, in PDF format, can be downloaded from [www.tektronix.com/manuals](http://www.tektronix.com/manuals). You can find the manuals by searching on the product name.

### Other documentation

Your instrument includes primary and supplemental information on CD-ROM:

- Documents CD (Tektronix part number 063-4546-XX)

## Video tutorials

You can browse the Tektronix YouTube channel ([www.youtube.com/user/tektronix](http://www.youtube.com/user/tektronix)) to find video tutorials about various topics related to your product. You can also subscribe to the Tektronix YouTube channel to keep up with new postings.

### Searching for topics

For example, you can watch a video tutorial about using the WLAN Presets. To find a video on this topic, do the following. The following image shows you what the Tektronix YouTube Channel looks like.

1. Click on the search icon located just above the video you see when the page first loads.

---

**NOTE.** *This icon allows you to search the Tektronix YouTube channel specifically. The search icon located at the top of the page allows you to search all of YouTube .*

---

2. Type in the keyword “WLAN” in the search field.
3. Click the search icon to start the search.
4. Videos related to the topic will appear. Click a video to view it.

**Tektronix**

Subscribe1,518

Home


Videos

Discussion

About

WLAN

Search



7:59 / 7:59

### How to use an oscilloscope






119,684 views · 2 years ago

Tony and Ian from Tektronix walk through what is an oscilloscope, how to use an oscilloscope, and other oscilloscope basics. You'll see how you might use an oscilloscope to look at an RC circuit, something many engineers encounter early in their studies. From there, learn about the most important oscilloscope controls and how to use them.

What is an Oscilloscope?  
At its heart, an oscilloscope is a device that sho...  
[Read more](#)


### Tektronix Tutorials

Tektronix explains the basics about every day test and measurement equipment that play an important role for engineers.






## Front panel connectors




| Item | Connector | Description  |
|------|-----------|--|
| 1    | Trig Out  | Trigger output connector. 50 $\Omega$ , BNC, High > 2.0 V, Low < 0.4 V, (output current 1 mA). |
| 2    | Trig In   | External Trigger input connector, -2.5 V to +2.5 V (user settable).                            |
| 3    | USB 2.0   | USB 2.0 connector.   |
| 4    | USB 2.0   | USB 2.0 connector.   |
| 5    | RF Input  | RF input connector 50 $\Omega$ .   |

## Front-Panel Controls




0519-044

| Reference | Item            | Function   | Menu Equivalent              |
|-----------|-----------------|--|------------------------------|
| 1         | Media           | DVD-RW or removable hard disk drive.   |                              |
| 2         | Displays        | Opens the Displays dialog box enabling you to select which displays to open.   | Setup > Displays             |
| 3         | Settings        | Opens/closes the Settings control panel for the selected display.  | Setup > Settings             |
| 4         | Trigger         | Opens/closes the Trigger control panel.  | Setup > Trigger              |
| 5         | Acquire         | Opens/closes the Acquire control panel.  | Setup > Acquire              |
| 6         | Analysis        | Opens/closes the Analysis control panel.   | Setup > Analysis             |
| 7         | Freq            | Press to adjust the measurement frequency.   | Setup > Analysis > Frequency |
| 8         | Span (Spectrum) | Press to adjust the span or press and hold to display the Freq & Span control panel for the General Signal Viewing displays. |                              |
| 9         | Amplitude       | Opens/closes the Amplitude control panel.  | Setup > Amplitude            |
| 10        | BW (Spectrum)   | Press to adjust the bandwidth or press and hold to display the BW control panel for the General Signal Viewing displays.     |                              |




| Reference | Item                      | Function   | Menu Equivalent        |
|-----------|---------------------------|--|------------------------|
| 11        | Run/Stop                  | Starts and stops acquisitions.   | Run > Start Run > Stop |
| 12        | Peak (Markers section)    | Moves the active marker to the maximum peak of the trace in the selected display. If markers are turned off, the marker reference (MR) will appear at the maximum peak.  | Markers > Peak         |
| 13        | Select (Markers section)  | Selects the next marker. If markers are turned off, the MR marker (marker reference) will appear.  |                        |
| 14        | Define (Markers section)  | Opens the Markers control panel. If markers are turned off, the MR marker (reference) will appear.   |                        |
| 15        | Control knob              | Changes values in numeric and list controls. Pressing the knob (clicking it) is the same as pressing the Enter key on a keyboard.  |                        |
| 16        | Arrow keys                | Move the Markers. The Up arrow moves the selected marker to the next highest peak. The down arrow moves the selected marker to the next lower peak value. The right and left arrows move the selected marker to the next peak. |                        |
| 17        | Increment/decrement keys  | Increments or decrements the selected value  |                        |
| 18        | Delete, (Markers section) | Deletes the selected marker  |                        |
| 19        | Add, (Markers section)    | Add a marker to the selected trace   |                        |
| 20        | Replay                    | Replays the current acquisition record   |                        |
| 21        | Single                    | Sets the Run mode to Single Sequence   |                        |



05-10-046

| Reference | Item   | Function   | Menu Equivalent |
|-----------|--------|--|-----------------|
| 22        | Keypad | Enters values in numeric controls.   |                 |
| 23        | Enter  | Completes data entry in controls.<br>Same as pressing the Enter key on an external keyboard. |                 |



| Reference | Item             | Function   | Menu Equivalent       |
|-----------|------------------|--|-----------------------|
| 24        | Recall           | Opens the Recall dialog box.                                   | File > Recall         |
| 25        | Save             | Opens the Save As dialog box.                                  | File > Save As        |
| 26        | Touch Screen Off | Turns the touch screen on and off. It is off when lighted.     |                       |
| 27        | Help             | Displays the help.   | Help > Online Manual  |
| 28        | Applic           | Sets the instrument to the selected Application Preset values. | Presets > Application |
| 29        | DPX              | Sets the instrument to the selected DPX Preset values.         | Presets > DPX         |
| 30        | User             | Sets the instrument to the selected User Preset values.        | Presets > User        |
| 31        | Preset           | Returns the instrument to the default or preset values.        | Preset                |

## Touch Screen

You can use touch to control the instrument in addition to the front-panel controls, mouse, or extended keyboard. Generally, touch can be used anywhere that click is mentioned in this help.

To disable the touch screen, push the front-panel **Touch Screen Off** button. When the touch screen is off, the button is lighted. You can still access the on-screen controls with a mouse or keyboard.

You can adjust the touch screen operation to your personal preferences. To adjust the touch screen settings, from Windows, select **Start > Control Panel > Touch Screen Calibrator**.


---

**NOTE.** *If the instrument is powered on in Windows Safe Mode, the touch screen is inoperative. You will need to use a mouse or keyboard to restore normal operation.*






---

## Touch-Screen Actions

You can use the touch screen to change marker settings and how waveforms are displayed by using the Touch-screen Actions menu.




To use the Touch-screen Actions menu, touch the display in a graph area and hold for one second, then remove your finger. You can also use a mouse to display the Touch-screen Action menu by clicking the right mouse button.

| Icon  | Menu            | Description   |
|---|-----------------|---|
|  | Select          | Selects markers and adjusts their position.   |
|  | Span Zoom       | Zooms the graph area about the selected point. Touch the graph display at a point of interest and drag to increase or decrease the span about the point of interest. Span Zoom adjusts the span control and can affect the acquisition bandwidth.   |
|  | CF Pan          | Adjusts the Center Frequency according to horizontal movement.  |
|  | Zoom            | Adjusts horizontal and vertical scale of the graph. The first direction with enough movement becomes the primary scale of adjustment. Adjustment in the secondary direction does not occur until a threshold of 30 pixels of movement is crossed.<br><br>Dragging to the left or down zooms out and displays a smaller waveform (increases the scale value). Dragging to the right or up zooms in and displays a larger waveform (decreases the scale value). |
|  | Pan             | Adjusts horizontal and vertical position of the waveform. The first direction with enough movement becomes the primary direction of movement. Movement in the secondary direction does not occur until a threshold of 30 pixels of movement is crossed.   |
| -   | Reset Scale     | Returns the horizontal and vertical scale and position settings to their default values.  |
| -   | Marker to peak  | Moves the selected marker to the highest peak. If no marker is turned on, this control automatically adds a marker.   |
| -   | Next Peak       | Moves the selected marker to the next peak. Choices are Next left, Next right, Next lower (absolute), and Next higher (absolute).   |
| -   | Add marker      | Defines a new marker located at the horizontal center of the graph.   |
| -   | Delete marker   | Removes the last added marker.  |
| -   | All markers off | Removes all markers.  |
|   | Trigger On This | Use to visually define trigger parameters in the DPX display (present only in the DPX Spectrum display).  |

### Touch-Screen Menu for Spurious Display


The Touch-screen actions menu in the Spurious display has some minor changes compared to the standard version used in other displays.




| Icon | Menu               | Description   |
|------|--------------------|---|
| -    | Single-range       | Changes the current multi-range display to a single range display. The displayed range is the range in which you display the touchscreen-actions menu. Selecting Single-range from the menu is equivalent to selecting Single on the Settings > Parameters tab. |
| -    | Multi-range        | Changes the current single-range display to a multi-range display. Selecting Multi-range from the menu is equivalent to selecting Multi on the Settings > Parameters tab.   |
| -    | Marker -> Sel Spur | Moves the selected marker to the selected spur.   |

## Elements of the Display

The main areas of the application window are shown in the following figure.




Specific elements of the display are shown in the following figure.



| Ref number | Setting              | Description   |
|------------|----------------------|---|
| 1          | Recall               | Displays the Open window in order to recall setup files, acquisition data files, or trace files.  |
| 2          | Save                 | Opens the Save As dialog in order to save setup files, pictures (screen captures), acquisition data files, or export measurement settings or acquisition data.  |
| 3          | Undo / Redo          | Undoes or redoes the previous edit to a display or measurement settings, a preset, or a measurement change.   |
| 4          | Displays             | Opens the Select Displays dialog box so that you can select measurement displays.   |
| 5          | Markers              | Opens or closes the Marker toolbar at the bottom of the window.   |
| 6          | Settings             | Opens the Settings control panel for the selected display. Each display has its own control panel.  |
| 7          | Trigger              | Opens the Trigger control panel so that you can define the trigger settings.  |
| 8          | Acquire              | Opens the Acquire control panel so that you can define the acquisition settings.  |
| 9          | Analysis             | Opens the Analysis control panel so that you can define the analysis settings such as frequency, analysis time, and units.  |
| 10         | Amplitude            | Opens the Amplitude control panel so that you can define the Reference Level, configure internal attenuation, and enable/disable the (optional) Preamplifier.   |
| 11         | Frequency            | Displays the frequency at which measurements are made. For spectrum displays, this is called "Center Frequency". To change the value, click the text and use the front panel knob to dial in a frequency. You can also enter a frequency with the front panel keypad or use the front panel up and down buttons.  |
| 12         | Reference Level      | Displays the reference level. To change the value, click the text and enter a number from the keypad or use the front panel up and down buttons.  |
| 13         | Preset               | Recalls the <a href="#">Main (see page 23)</a> preset.  |
| 14         | Replay               | Runs a new measurement cycle on the last acquisition data record using any new settings.  |
| 15         | Run / Stop           | Starts and stops data acquisitions. When the instrument is acquiring data, the button label has green lettering. When stopped, the label has black lettering. You can specify the run conditions in the Run menu. For example, if you select Single Sequence in the Run menu, when you click the Run button, the instrument will run a single measurement cycle and stop. If you select Continuous, the instrument will run continuously until you stop the acquisitions. |
| 16         | Check mark indicator | The check mark indicator in the upper, left-hand corner of the display indicates the display for which the acquisition hardware is optimized.<br><br><b>NOTE.</b> When Best for multiple windows is selected in the Amplitude control panel's RF & IF Optimization control, none of the measurement displays shows a checkmark, as there is not a single optimized measurement.   |
| 17         | Show                 | Shows / hides the selected trace.   |

## Rear-Panel Connectors



| Item | Description   |
|------|---|
| 1    | AC Input, main power connector  |
| 2    | GPIB  |
| 3    | IF Output (Option 55)   |
| 4,5  | Digital I and Q Outputs (Option 05)   |
| 6    | +28 V DC Output, switched   |
| 7    | Microphone in; Headphone, audio output; and Line In connectors                |
| 8    | External Trigger 2 Input  |
| 9    | COM 2, serial port for connecting peripherals                                 |
| 10   | VGA external monitor output (resolution not limited to VGA)                   |
| 11   | PS2 keyboard input  |
| 12   | USB 2.0 ports for mouse and other peripherals (printers, external hard disks) |
| 13   | Ref Out, reference frequency output   |
| 14   | Ref In, reference frequency input   |
| 15   | LAN, Ethernet network connector   |

## Setting Up Network Connections

Because the instrument is based on Windows, you configure network connections for the instrument the same way you would for any PC based on Windows. See **Help and Support** in the Windows **Start** menu to access the Windows Help System for information on setting up network connections.



## Restoring Default Settings

To restore the software to its factory default settings:

1. Select **Presets > Preset Options**.
2. In the Presets tab of the Options control panel, click to view the **Preset type** drop down menu and select **Main**.
3. Click to view the **Presets** drop down menu and select **Original**.
4. Click the red X icon in the top right corner of the Options control panel to close the panel.
5. Select **Presets > Main** from the menu bar to return the software to its original factory default settings.

---

**NOTE.** You can also press the Preset button on the front-panel or click the **Preset** button on the right-hand side of the display menu bar to load the Main preset.

---

---

**NOTE.** The Original Main preset resets all settings and clears all acquisition data (previously recalled waveform files). Settings that have not been saved will be lost.

---

## Running Alignments

Alignments are adjustment procedures. Alignments are run by the instrument using internal reference signals and measurements and do not require any external equipment or connections.

These are the settings for Alignments:

- Automatically align as needed
- Run alignments only when the **Align Now** button is pressed

If **Automatically align as needed** is selected, alignments run whenever the signal analyzer detects a sufficient change in ambient conditions to warrant an alignment.

If **Run alignments only when "Align Now" button is pressed** is selected, the signal analyzer never runs an alignment unless you manually initiate an alignment using the Align Now button.

---

**NOTE.** There are a few critical adjustments that must run occasionally even if *Automatically align* is not enabled.

---

### Alignment Status

When the signal analyzer needs to run an alignment, it displays a message on screen. If no message is displayed, you can assume that the signal analyzer is properly aligned.

---

**NOTE.** *If you must use the instrument before it has completed its 20-minute warm-up period, you should perform an alignment to ensure accurate measurements.*

---

### Initiating an Alignment

1. Select **Tools > Alignments**.
2. Select the **Align Now** button.

The signal analyzer will run an alignment procedure. Status messages are displayed while the alignment procedure is running. If the instrument fails the alignment procedure, an error message will be displayed. If the instrument fails an alignment, run Diagnostics (Tools > Diagnostics) to see if you can determine why the alignment failed.

---

**NOTE.** *While an alignment is running, both the IF and IQ outputs are disabled.*

---

**Alignments during warm-up.** During the 20-minute warm-up period, the signal analyzer will use the alignment data generated during the previous use of the instrument as it warms to operating temperature (if Auto mode is selected). During the specified period for warm-up, the instrument performance is not warranted.

**Alignments during normal operation.** Once the signal analyzer reaches operating temperature  $\pm 3$  degrees C (as detected inside the instrument), an alignment will be run. If an alignment becomes necessary during a measurement cycle (if Auto mode is selected), the measurement is aborted and an alignment procedure is run. Once an alignment procedure is completed, the measurement cycle restarts.

Instruments with an HD option have certain components that require alignment for  $\pm 1$  degree C change. In these cases, once the signal analyzer reaches operating temperature  $\pm 1$  degree C (as detected inside the instrument), an alignment will be run.

---

**NOTE.** *The first time the instrument runs after a software upgrade (or reinstall), the instrument will perform a full alignment after the 20-minute warm-up period. This alignment cannot be aborted and it occurs even if alignments are set to run only when manually initiated.*

---

### Alignments Are Not Calibrations

Alignments are adjustment procedures run by the instrument using internal reference signals and measurements. Calibrations can only be performed at a Tektronix service center and require the use of traceable test equipment (signal sources and measuring equipment) to verify the performance of the instrument.

## Presets

### Menu Bar: Presets

The analyzer includes a set of configurations or presets that are tailored to specific applications. These configurations, referred to as Presets, open selected displays and load settings that are optimized to address specific application requirements.

### Available Presets

Select **Presets > Preset Options** from the menu bar to access the available types of Presets:

- Main
- DPX
- Standards
- Application
- User

You can set if a preset is immediately executed when selected, or if a list of presets is displayed from which you can select the Preset to recall. Available presets are described in the following table.

| Presets                             | Description   |
|-------------------------------------|---|
| <b>Main</b>                         |   |
| Current                             | This Preset sets the instrument to display a Spectrum display with settings matched to show a Spectrum display with settings appropriate for typical spectrum analysis tasks. This preset was updated from the original factory preset with version 3.2 of the instrument software. |
| Original                            | This Preset is the original factory preset used with software versions 1.0 through 3.2. This version of the factory preset is included to allow users to maintain compatibility with existing remote control software.  |
| <b>DPX</b>                          |   |
| Open the DPX display                | The Open the DPX display opens the DPX display without closing existing displays.   |
| Swept                               | The DPX Swept Preset displays the DPX Spectrum display with the span set to maximum and the center frequency set to 1/2 the span.   |
| Real Time                           | The DPX Real Time Preset displays the DPX Spectrum display with the center frequency set to 1.5 GHz and the span set to the maximum available real-time bandwidth.  |
| Zero Span                           | The DPX Zero Span Preset displays the DPX Zero Span display with the position set to 0 s and the sweep set to 1 ms.   |
| <b>Standards</b>                    |   |
| <a href="#">WLAN (see page 233)</a> | This preset sets the instrument to display the WLAN Summary, WLAN Constellation, and SEM displays. After you select the standards and bandwidth, the software configures these displays to apply the parameters appropriate for typical WLAN analysis tasks.                        |
| <a href="#">P25 (see page 379)</a>  | This preset sets the instrument to display the MCPR, Time Overview, P25 Summary, and P25 Constellation displays. After you select the standard and modulation type, the software configures these displays to apply the parameters appropriate for typical P25 analysis tasks.      |

| Presets   | Description   |
|---|---|
| <a href="#">Bluetooth (see page 379)</a>              | This preset set the instrument to display a particular set of displays depending on which combination of standard (Low Energy or Basic Rate) and test setup (one of up to eight) you select. After you select the standard and test setup, the software configures the displays to apply the parameters appropriate for typical Bluetooth analysis tasks. |
| <b>Application</b>                                    |   |
| <a href="#">Time-Frequency Analysis (see page 28)</a> | The Time-Frequency preset configures the instrument with settings suited to analyzing signal behavior over time.  |
| <a href="#">Spectrum Analysis (see page 27)</a>       | The Spectrum Analysis application preset provide you with the settings commonly used for general purpose spectrum analysis.   |
| <a href="#">Modulation Analysis (see page 26)</a>     | The Modulation Analysis setup application preset provides you with the most common displays used during modulation analysis. Only present when Option 21 is installed.  |
| <a href="#">Pulse Analysis (see page 26)</a>          | The Pulse Analysis application preset provides you with the most common displays used during pulse analysis, and makes changes to the default parameters to settings better optimized for pulsed signal analysis. Only present when Option 20 is installed.   |
| Spur Search Multi Zone 9k-1GHz                        | The Spur Search application preset configures the instrument to show the Spurious display with the frequency range set to 9 kHz to 1 GHz.   |
| <a href="#">Phase Noise (see page 26)</a>             | The Phase Noise application preset opens the Phase Noise display, and makes changes to the default parameters to settings better optimized for phase noise analysis. Only present when Option 11 is installed.  |
| <a href="#">Noise Figure (see page 28)</a>            | The Noise Figure application preset opens the Gain, Noise Figure, and Noise Table displays, and makes changes to the default parameters to settings better optimized for noise figure analysis. Only present when Option 14 is installed.   |
| <b>User</b>   |   |
| User Preset 1   | This Preset is provided as an example for you to create your own Presets. This preset displays the Spectrum, Spectrogram, Frequency vs Time, and Time Overview displays.  |
| User Preset 2   | This Preset is provided as an example for you to create your own Presets. This preset displays the Spurious display configured to test for Spurious signals across four ranges.   |

## Preset Options

Select the **Presets > Preset Options** menu to open the Options control panel. This panel allows you to select the following. Once you have chosen these settings, you can access any preset or list of presets from **Presets** on the menu bar.

- **Preset type:** Select the Preset type.
- **Presets:** Select which preset you want to display for that particular preset type.
- **Preset action:** Recalling Presets results in either of two actions. One action is to immediately execute a Preset. The second action displays a list of Presets from which you select the Preset you want to recall. You can select from **Recall selected preset** or **Show list**.

## Configuring a User Preset

After you have selected a preset:

- Set the measurement frequency using the front-panel knob or keypad.
- Adjust the span to show the necessary detail.

## Recalling a Preset

To recall a preset, select **Presets** and then the desired preset type.

---

**NOTE.** You can set which presets to recall from the **Presets > Preset Options (see page 24)** control panel.

---

To recall a named (User) preset from the front panel, press the button on the front panel matching the preset type you want to recall. For example, to recall a DPX preset type, press the DPX button.

---

**NOTE.** You can also click the **Preset** button on the right-hand side of the menu bar to load the Main preset.

---

---

**NOTE.** The only Presets recalled by the front-panel Preset button, the Preset icon in the icon bar, and the \*RST remote command, are the Main Presets. Application, DPX, Standards, and User Presets can only be recalled using selections in the Presets menu.

---

## Creating User Presets

You can add your own presets to the list that appears in the User Presets dialog box. Configure the analyzer as needed for your application and create a Setup file in **C:\RSA6100B Files\User Presets**. The name you give the file will be shown in the User Presets list on the Presets tab of the Options control panel. For instructions on how to save a Setup file, see [Saving Data \(see page 607\)](#).

## Standards Presets

The Standards presets allow you to recall preconfigured displays for the standards that you select. You can select from the following standards groups.

- **WLAN:** The IEEE wireless LAN (WLAN) standards specify the wireless interface between two wireless clients or a wireless client and a base station. The WLAN presets allow you to access displays preconfigured for the WLAN standards and bandwidths you select.

---

**NOTE.** More information is available about WLAN standards [here \(see page 233\)](#). You can also watch a video tutorial about WLAN Presets at [www.youtube.com/user/tektronix](http://www.youtube.com/user/tektronix). Click [here \(see page 6\)](#) for information about searching the Tektronix YouTube channel for videos.

---

- **P25:** The Project 25 (P25) TIA-102 standards specify the design of interoperable digital two-way wireless communications products. The P25 presets allow you to access displays preconfigured for the P25 standards you select. You can choose either the Phase 1 (FDMA) standard or the Phase 2 (TDMA) standard.

---

**NOTE.** More information is available about P25 standards [here \(see page 379\)](#).

---

## Modulation Analysis

The Modulation Analysis application preset opens the following displays:

- **DPX display:** Shows you a continuous spectrum monitoring of the specified carrier frequency.
- **Signal Quality:** Shows a summary of modulation quality measurements (EVM, rho, Magnitude Error, Phase Error, and others).
- **Constellation:** Shows the I and Q information of the signal analyzed in an I vs. Q format.
- **Symbol Table:** Shows the demodulated symbols of the signal.

To use the Modulation Analysis preset (assuming the Preset action is set to Show list in the Presets tab of the Options control panel):

1. Select **Presets > Application**. Select **Modulation Analysis** and then click **OK**.
2. Set the measurement frequency using the front-panel knob or keypad. Your signal should appear in the DPX display.
3. Set the reference level so that the peak of your signal is about 10 dB below the top of the DPX display.
4. Set the modulation parameters for your signal. This includes the Modulation Type, Symbol Rate, Measurement Filter, Reference Filter and Filter Parameter. All of these settings are accessed by pressing the Settings button.

For most modulated signals, the Modulation Analysis application preset should present a stable display of modulation quality. Additional displays can be added by using the Displays button, and other settings can be modified to better align with your signal requirements.

## Phase Noise

The Phase Noise application preset opens the Phase Noise display.

## Pulse Analysis

The Pulse Analysis application preset opens the following displays:

- **DPX:** The DPX display is opened with the maximum available span.
- **Time Overview:** Shows amplitude vs. time over the analysis period.
- **Pulse Trace:** Shows the trace of the selected pulse and a readout of the selected measurement from the pulse table.
- **Pulse Measurement Table:** This shows a full report for the user-selected pulse measurements.

You can make a selected pulse and measurement appear in the Pulse Trace display by highlighting it in the Pulse Measurement Table. Key pulse-related parameters that are set by the Pulse Analysis application preset are:

- **Measurement Filter:** No Filter.
- **Measurement Bandwidth:** This is set to the maximum real-time bandwidth of the instrument (40 MHz in a base instrument or 110 MHz with instruments with Option 110). Note: The label on the “Measurement Bandwidth” setting is just “Bandwidth”. Like the main instrument Preset command and the other application presets, the Pulse Analysis application preset also sets most other instrument controls to default values.
- **Analysis Period:** This is set to 2 ms to ensure a good probability of catching several pulses for typical signals.

To use the Pulse Analysis preset (assuming the Preset action is set to Show list in the Presets tab of the Options control panel):

1. Select **Presets > Application**. Select **Pulse Analysis** and then click **OK**.
2. Set the Center Frequency control to the carrier frequency of your pulsed signal.
3. Set the Reference Level to place the peak of the pulse signal approximately 0-10 dB down from the top of the Time Overview display.

You may need to trigger on the signal to get a more stable display. This is set up in the Trigger control panel. (“Trig” button). Using the Power trigger type with the RF Input source works well for many pulsed signals.

4. Set the Analysis Period to cover the number of pulses in your signal that you want to analyze. To do this, click in the data entry field of the Time Overview window and set the analysis length as needed.

## Spectrum Analysis

The Spectrum Analysis application preset opens a Spectrum display and sets several parameters. The Spectrum Analysis preset sets the analyzer as follows.

- **Spectrum Analysis :** Sets the frequency range to maximum for the analyzer, and sets the RF/IF optimization to **Minimize Sweep Time**.

To use the Spectrum Analysis preset (assuming the Preset action is set to Show list in the Presets tab of the Options control panel):

1. Select **Presets > Application**. Select **Spectrum Analysis** and then click **OK**.
2. Set the measurement frequency using the front-panel knob or keypad.
3. Adjust the span to show the necessary detail.

## Time-Frequency Analysis

The Time-Frequency Analysis application preset opens the following displays:

- Time Overview: Shows a time-domain view of the analysis time 'window'.
- Spectrogram: Shows a three-dimensional view of the signal where the X-axis represents frequency, the Y-axis represents time, and color represents amplitude.
- Frequency vs. Time: This display's graph plots changes in frequency over time and allows you to make marker measurements of settling times, frequency hops, and other frequency transients.
- Spectrum: Shows a spectrum view of the signal. The only trace showing in the Spectrum graph after selecting the Time-Frequency Analysis preset is the Spectrogram trace. This is the trace from the Spectrogram display that is selected by the active marker. Stop acquisitions with the Run button because its easier to work with stable results. In the Spectrogram display, move a marker up or down to see the spectrum trace at various points in time.

The analysis period is set to 5 ms.

To use the Time-Frequency Analysis preset (assuming that Time-Frequency Analysis is the selected preset on the list of Application Presets and Preset action is set to Recall selected preset):

1. Select **Presets > Application**. Select **Time-Frequency Analysis** and then click **OK**.
2. When the preset's displays and settings have all been recalled and acquisitions are running, adjust the center frequency and span to capture the signal of interest.
3. Set the Reference Level to place the peak of the signal approximately 0-10 dB down from the top of the Spectrum graph.
4. If the signal is transient in nature, you might need to set a trigger to capture it. For more information on triggering in the time and frequency domain, see [Triggering \(see page 586\)](#).

When the signal has been captured, the spectrogram shows an overview of frequency and amplitude changes over time. To see frequency transients in greater detail, use the Frequency vs. Time display.

The Time-Frequency Analysis preset sets the analysis period to 5 ms. The Spectrum Span is 40 MHz. The RBW automatically selected for this Span is 300 kHz. For a 300 kHz RBW, the amount of data needed for a single spectrum transform is 7.46  $\mu$ s. A 5 ms Analysis Length yields 671 individual spectrum transforms, each one forming one trace for the Spectrogram to display as horizontal colored lines. This preset scales the Spectrogram time axis (vertical axis) to -2, which means that the Spectrogram has done two levels of time compression, resulting in one visible line for each four transforms. This results in 167 lines in the Spectrogram for each acquisition, each covering 29.84  $\mu$ s.

## Noise Figure

The Noise Figure and Gain Measurements application preset opens the following displays:


- **Gain:** This display shows gain versus frequency of the signal. The gain measurement is the ratio of output power to input power in an amplifier or circuit element.
- **Noise Figure:** This display shows the noise factor in dB. This measurement can help you assess the low level sensitivity of the DUT. Lower noise figure is found in better performing DUTs. Noise factor is defined as the ratio of the input signal to noise ratio to the output signal to noise ratio (Input SNR/Output SNR).
- **Noise Table:** This table lists selected measurements in a spreadsheet format, showing the numeric value at step frequencies for Gain, Noise Figure, Y Factor, Noise Temperature, PHot, and PCold.

You can read more about Noise Figure and Gain measurements [here \(see page 295\)](#).

## Setting Options

### Menu Bar: Tools > Options

There are several settings you can change that are not related to measurement functions. The Option settings control panel is used to change these settings.



| Settings tab    | Description   |
|-----------------|---|
| Presets         | Use this tab to configure Presets. You can specify the action to take when a preset is recalled and which preset to recall when the Preset button is selected.  |
| Analysis Time   | Use this tab to specify the method used to automatically set the analysis and spectrum offsets when the <a href="#">Time Zero Reference (see page 555)</a> is set to Trigger.   |
| Save and Export | Use this tab to specify whether or not save files are named automatically and what information is saved in acquisition data files.  |
| GPIB            | Use this tab to set the primary GPIB address for the instrument.  |
| Security        | Selecting the <b>Hide Sensitive readouts</b> check box causes the instrument to replace measurement readouts with a string of asterisks.  |
| Prefs           | Use this tab to select different color schemes for the measurement graphs and specify how markers to automatically jump to the <a href="#">next peak (see page 542)</a> when you drag them. When this setting is deselected, you can drag a marker to any point on the trace. |

### Presets

The Presets tab in the Options control panel allows you to specify actions taken when you press the Preset button. You can read more about this tab [here \(see page 24\)](#).

## Analysis Time

The Analysis Time tab in the Options control panel is used to specify the method used to automatically set the analysis and spectrum offsets when the [Time Zero Reference \(see page 555\)](#) is set to Trigger. The available settings are:

- Include trigger point – Selects an algorithm that uses the measurements to determine how far in advance of the trigger to set the analysis offset. The analyzer tries to ensure that data about the trigger point is included in the analyses.
- Start at trigger point (legacy) – The method used by the instrument in prior versions, which sets the Analysis Offset to zero when possible. The analyzer tries to ensure that data following the trigger point is included in the analyses. Use this method if your measurements or procedures depend on past behavior of the Auto Analysis Offset function.

## Save and Export

The Save and Export tab allows you to specify whether or not files are saved with an automatically generated name, and how much data is saved in an acquisition data file.

**All files.** The Automatically increment filename/number function can automatically name saved files by appending a number to a base file name. Use this tab to enable/disable automatic naming of files. For example, if **Automatically Increment Filename Number** is disabled, when you select Save from the File menu, you will have to enter a name for the file.

**Acquisition data files.** This setting specifies whether saved data files include the entire acquisition record or only the data for the analysis length (a subset of the acquisition record). You can choose from the following:

- IQ records: Includes IQ records
- DPX spectra: Includes DPX spectra
- Both: Includes both IQ records and DPX spectra

You can also select to include an entire IQ record or just the analysis length of it.

**TIQ acquisition data files.** Specifies which data records to save. You can choose from the following:

- Current acquisition: Saves the current acquisition.
- Current frame: If Fast Frame is enabled, saves only the current frame. The current frame is the one most recently analyzed.
- Selected frames: If Fast Frame is enabled, saves the specified frames.
- All in history: Saves all acquisition records in the history.
- Save TIQ file now: Invokes the Save As dialog box with the Save as type drop-down list set to TIQ.

## Security

The Security tab enables you to hide sensitive readouts in displays with readouts, such as the OFDM Summary display.

## Prefs

The Prefs tab enables you to set properties that apply to all displays.

**Color scheme.** The Color scheme setting provides three color schemes for the measurement graphs. The color scheme setting does not change the overall instrument application or Windows color scheme.

- Thunderstorm – This scheme displays graphs in shades of blue. This provides a less vibrant color scheme than the default setting.
- Blizzard – This scheme displays graphs with a white background to save ink when printing.
- Classic – The default setting. This scheme displays the graph area with a black background.

**Markers snap to peaks when dragged.** When selected, this setting causes makers to automatically jump to the [next peak \(see page 542\)](#) when you drag them. When this setting is deselected, you can drag a marker to any point on the trace.

## Operating System Restore

The instrument contains an operating system restore file on a separate partition of the hard drive.

The preferred method to restore the instrument operating system is to use the hard disk restore file.



---

**CAUTION.** *Using the restore process reformats the hard drive and reinstalls the operating system. All saved data is lost. If possible, save important files to external media before performing a system restore.*

---

1. Restart the instrument. During the boot-up process you will see the following message at the top of the screen: Starting Acronis Loader... press F5 for Acronis Startup Recovery Manager.

---

**NOTE.** *To successfully complete the system restore, you must use the Windows version of the Acronis software. Using a generic Macintosh keyboard starts the DOS version of the Acronis software. Do not use a Macintosh keyboard.*

---

2. Repeatedly press the F5 key until the Acronis True Image Tool opens. There is a 5-second time period from when the message appears until the instrument proceeds with the normal instrument startup. If the instrument does not open the Acronis application, power off the instrument, then power on the instrument and try again.
3. Click Restore.
4. In the Confirmation dialog box, click Yes to restore the instrument operating system, or No to exit the restore process. The restore process takes approximately 30 minutes; the actual time depends on the instrument configuration.




## Selecting Displays

**Menu Bar:** Setup > Displays

**Application Toolbar:** 

Use the **Select Displays** dialog to choose the displays that appear on the screen.



To select displays:

1. Press the **Displays** button or select **Setup > Displays**.
2. Select one of the choices under **Measurements**. The measurement chosen determines the choices available in **Available displays**.
3. Double-click the desired display in the **Available displays** box or select the desired display and click **Add**.
4. Click **OK**.

### Interactions Between Displays

Different displays can require different settings, for example acquisition bandwidth, analysis length, or resolution bandwidth, to achieve optimum results. The application automatically adjusts some settings

to optimize them for the selected display. The check mark indicator in the upper, left-hand corner of the display indicates the display for which the application is optimized. Depending on application settings, some displays might stop displaying results if they are not the selected display.

## Available Measurements

The automatic measurements available include RF power measurements, OFDM analysis, WLAN analysis, APCO P25 analysis, Bluetooth analysis, audio analysis, analog modulation measurements, digital modulation measurements, noise figure and gain measurements, and pulsed RF measurements.

### Power measurements

| Measurement                         | Description  |
|-------------------------------------|--|
| Channel Power                       | The total RF power in the selected channel (located in the Chan Pwr and ACPR display).   |
| Adjacent Channel Power Ratio (ACPR) | Measure of the signal power leaking from the main channel into adjacent channels.  |
| Multi-Carrier Power Ratio (MCPR)    | The ratio of the signal power in the reference channel or group of channels to the power in adjacent channels.   |
| Peak/Avg Ratio                      | Ratio of the peak power in the transmitted signal to the average power in the transmitted signal (located in the CCDF display).  |
| CCDF                                | The <i>Complementary Cumulative Distribution Function</i> (CCDF). CCDF shows how much time a signal spends at or above a given power level relative to the average power of a measured signal. |

**OFDM analysis**

| <b>Measurement</b> | <b>Description</b>  |
|--------------------|---|
| Channel Response   | Plots the channel response (magnitude or phase) versus the subcarrier or frequency. Here, the channel refers to all sources of signal frequency response impairment up to the analyzer input, including the transmitter itself, as well as any transmission medium through which the signal travels between the transmitter and the analyzer.   |
| Constellation      | Shows the WLAN signal modulation amplitude and phase in I (horizontal) versus Q (vertical) form. For multicarrier WLAN OFDM signals, the points show all data symbol subcarriers' modulation. For single-carrier 802.11b, each point corresponds to a single modulated chip.  |
| EVM                | The normalized RMS value of the error vector between the measured signal and the ideal reference signal over the analysis length. The EVM is generally measured on symbol or chip instants and is reported in units of percent and dB. EVM is usually measured after best-fit estimates of the frequency error and a fixed phase offset have been removed. These estimates are made over the analysis length. Displays RMS and Peak values with location of Peak value. |
| Flatness           | Ratio of the peak power in the transmitted signal to the average power in the transmitted signal  |
| Mag Error          | The RMS magnitude difference between the measured signal and the reference signal magnitude. Displays RMS and Peak values with location of Peak value.  |
| Phase Error        | The RMS phase difference between the measured signal and the ideal reference signal. Displays RMS and Peak values with location of Peak value.  |
| Power              | shows the data symbols' individual subcarrier Power values versus symbol interval (time) and subcarrier (frequency).  |

## WLAN measurements

| Measurement      | Description   |
|------------------|---|
| Channel Response | Plots the channel response (magnitude or phase) versus the subcarrier or frequency. Here, the channel refers to all sources of signal frequency response impairment up to the analyzer input, including the transmitter itself, as well as any transmission medium through which the signal travels between the transmitter and the analyzer.   |
| Constellation    | Shows the WLAN signal modulation amplitude and phase in I (horizontal) versus Q (vertical) form. For multicarrier WLAN OFDM signals, the points show all data symbol subcarriers' modulation. For single-carrier 802.11b, each point corresponds to a single modulated chip.  |
| EVM              | The normalized RMS value of the error vector between the measured signal and the ideal reference signal over the analysis length. The EVM is generally measured on symbol or chip instants and is reported in units of percent and dB. EVM is usually measured after best-fit estimates of the frequency error and a fixed phase offset have been removed. These estimates are made over the analysis length. Displays RMS and Peak values with location of Peak value. |
| Flatness         | Ratio of the peak power in the transmitted signal to the average power in the transmitted signal  |
| Mag Error        | The RMS magnitude difference between the measured signal and the reference signal magnitude. Displays RMS and Peak values with location of Peak value.  |
| Phase Error      | The RMS phase difference between the measured signal and the ideal reference signal. Displays RMS and Peak values with location of Peak value.  |
| Power vs Time    | The signal power amplitude versus time. For 802.11b signals, the packet Power-On and Power-Down ramp times are also measured.   |
| Summary          | Shows several measurements of WLAN signal quality.  |
| Symbol Table     | Shows decoded data values for each data symbol in the analyzed signal packet. For OFDM (non-802.11b) signals, results are presented with subcarrier (frequency) indices in the horizontal dimension and symbol (time) intervals in the vertical dimension. For 802.11b signals, the Preamble, Header, and Data (PSDU) symbol values are presented sequentially, with symbol indices in the left column.   |

## Audio measurements

| Measurement    | Description   |
|----------------|---|
| Audio Spectrum | Shows audio modulation characteristics. You can choose to show just the spectrum of the audio signal or show the audio spectrum of the signal and the results of distortion measurements. The Audio Spectrum display can show a table listing the frequency of a Harmonic Distortion (HD) and Non-Harmonic Distortion (NHD) and its level. The Spectrum graph indicates these harmonics and non-harmonics with special markers. |

## Digital Modulation measurements

### Measurements for all modulation types except nFSK, C4FM, OQPSK and SOQPSK

| Measurement      | Description   |
|------------------|---|
| EVM              | The normalized RMS value of the error vector between the measured signal and the ideal reference signal over the analysis length. The EVM is generally measured on symbol or chip instants and is reported in units of percent and dB. EVM is usually measured after best-fit estimates of the frequency error and a fixed phase offset have been removed. These estimates are made over the analysis length. Displays RMS and Peak values with location of Peak value. |
| Phase Error      | The RMS phase difference between the measured signal and the ideal reference signal. Displays RMS and Peak values with location of Peak value.  |
| Mag Error        | The RMS magnitude difference between the measured signal and the reference signal magnitude. Displays RMS and Peak values with location of Peak value.  |
| MER (RMS)        | The MER is defined as the ratio of I/Q signal power to I/Q noise power; the result is indicated in dB.  |
| IQ Origin Offset | The magnitude of the DC offset of the signal measured at the symbol times. It indicates the magnitude of the carrier feed-through signal.   |
| Frequency Error  | The frequency difference between the measured carrier frequency of the signal and the user-selected center frequency of the instrument.   |
| Gain Imbalance   | The gain difference between the I and Q channels in the signal generation path. Constellations with gain imbalance show a pattern with a width that is different from height.   |
| Quadrature Error | The orthogonal error between the I and Q channels. The error shows the phase difference between I and Q channels away from the ideal 90 degrees expected from the perfect I/Q modulation. Not valid for BPSK modulation type.   |
| Rho              | The normalized correlated power of the measured signal and the ideal reference signal. Like EVM, Rho is a measure of modulation quality. The value of Rho is less than 1 in all practical cases and is equal to 1 for a perfect signal measured in a perfect receiver.  |

**Measurements for QPSK and SOQPSK modulation types**

| Measurement      | Description   |
|------------------|---|
| EVM              | The normalized RMS value of the error vector between the measured signal and the ideal reference signal over the analysis length. The EVM is generally measured on symbol or chip instants and is reported in units of percent and dB. EVM is usually measured after best-fit estimates of the frequency error and a fixed phase offset have been removed. These estimates are made over the analysis length. Displays RMS and Peak values with location of Peak value.                       |
| Offset EVM       | Offset EVM is like EVM except for a difference in the time alignment of the I and Q samples. For EVM, I and Q samples are collected at the same time, for every symbol decision point (twice the symbol rate for offset modulations). For Offset EVM, the I and Q symbol decision points are time-aligned before collecting the I and Q samples. In this case, one I and one Q sample is collected for each symbol (half as many samples as the same number of symbols for (non-offset) EVM). |
| Phase Error      | The RMS phase difference between the measured signal and the ideal reference signal. Displays RMS and Peak values with location of Peak value.  |
| Mag Error        | The RMS magnitude difference between the measured signal and the reference signal magnitude. Displays RMS and Peak values with location of Peak value.  |
| MER (RMS)        | The MER is defined as the ratio of I/Q signal power to I/Q noise power; the result is indicated in dB.  |
| IQ Origin Offset | The magnitude of the DC offset of the signal measured at the symbol times. It indicates the magnitude of the carrier feed-through signal.   |
| Frequency Error  | The frequency difference between the measured carrier frequency of the signal and the user-selected center frequency of the instrument.   |
| Gain Imbalance   | The gain difference between the I and Q channels in the signal generation path. Constellations with gain imbalance show a pattern with a width that is different from height.   |
| Quadrature Error | The orthogonal error between the I and Q channels. The error shows the phase difference between I and Q channels away from the ideal 90 degrees expected from the perfect I/Q modulation. Not valid for BPSK modulation type.   |
| Rho              | The normalized correlated power of the measured signal and the ideal reference signal. Like EVM, Rho is a measure of modulation quality. The value of Rho is less than 1 in all practical cases and is equal to 1 for a perfect signal measured in a perfect receiver.  |

**Measurements for nFSK modulation types**

| Measurement       | Description   |
|-------------------|---|
| Peak FSK err      | Peak value of the frequency deviation error at the symbol point.  |
| RMS FSK Err       | RMS value of the frequency deviation error at the symbol point.   |
| Peak Mag Err      | The Peak magnitude difference between the measured signal and the reference signal magnitude.   |
| RMS Mag Err       | The RMS magnitude difference between the measured signal and the reference signal magnitude.  |
| Freq Error        | The frequency difference between the measured carrier frequency of the signal and the user-selected center frequency of the instrument.   |
| Freq Deviation    | Frequency distance from the center frequency at the symbol point.   |
| Symbol Rate Error | This compares the user-entered symbol rate to the instrument calculated symbol rate of the analyzed signal.   |
| Symbol Rate       | When in Auto-symbol rate, the instrument calculates the symbol rate of the signal and the instrument calculates the error between the user entered value and the instrument calculated value. |

**Measurements for C4FM modulation type**

| Measurement             | Description  |
|-------------------------|--|
| RMS Error Magnitude     | RMS value of the frequency deviation error at the symbol point.                  |
| Carrier Frequency Error | Frequency difference between averaged signal frequency and the center frequency. |
| Deviation               | Frequency distance from the center frequency at the symbol point.                |
| Length                  | Number of symbols in the analysis area.  |

**Analog Modulation measurements****Measurements for AM modulation**

| Measurement | Description  |
|-------------|--|
| +AM         | Positive peak AM value.  |
| -AM         | Negative peak AM value.  |
| Total AM    | Total AM value, which is equal to the peak-peak AM value divided by 2. |

**Measurements for FM modulation**

| Measurement | Description                                    |
|-------------|--|
| +Pk         | Positive peak frequency deviation.             |
| -Pk         | Negative peak frequency deviation.             |
| RMS         | RMS value of the frequency deviation.          |
| Pk-Pk/2     | Peak-to-peak frequency deviation divided by 2. |
| Pk-Pk       | Peak-to-peak frequency deviation.              |

**Measurements for PM modulation**

| Measurement | Description                       |
|-------------|-----------------------------------|
| +Pk         | Positive peak phase deviation.    |
| -Pk         | Negative peak phase deviation.    |
| RMS         | RMS value of the phase deviation. |
| Pk-Pk       | Peak-to-peak phase deviation.     |

**Noise Figure and Gain measurements**

| Measurement       | Description   |
|-------------------|---|
| Noise Temperature | This measures thermal noise in the system.  |
| Y Factor          | This measurement method relies on a measurement of Noise Powers. These measurements during the calibration portion of the Y Factor technique are close to the noise floor limit; therefore, a low noise, low level measuring receiver or a spectrum analyzer with a low noise sensitive preamplifier is required.                         |
| Gain              | The measures the ratio of output power to input power in an amplifier or circuit element.   |
| Noise Figure      | This measures how much noise is added by an amplifier or other system component. Noise Figure is Noise Factor expressed in dB. Noise factor is defined as the ratio of the input signal to noise ratio to the output signal to noise ratio (Input SNR/Output SNR) for a reference load at an equivalent source noise temperature of 290K. |
| Uncertainty       | Shows the uncertainty in noise and gain measurements and also provides values showing the contribution of various elements to the overall uncertainty.  |

**APCO P25 measurements**

| Measurement   | Description   |
|---|---|
| RF output power   | Measure of RF output power when the transmitter is connected to the standard load during defined duty cycle.  |
| Operating frequency accuracy  | Measure of the ability of the transmitter to operate on its assigned frequency.   |
| Unwanted emissions (ACPR)   | Ratio of the total power of a transmitter under prescribed conditions and modulation to that of the output power that falls within a prescribed bandwidth centered on the nominal frequency of adjacent channels.       |
| Frequency deviation   | Measurement of the amount of frequency deviation that results for a Low Deviation and High Deviation test pattern.  |
| Modulation fidelity   | Measures the degree of closeness to which the modulation follows the ideal theoretical modulation determined by the rms difference between the actual deviation and the expected deviation for the transmitted symbols. |
| Symbol rate accuracy  | Measures the ability of the transmitter to operate at the assigned symbol rate (4.8 kHz for Phase 1, 6 kHz for Phase 2).  |
| Transmitter power and encoder attack time                           | Measures the time required for a transmitter to prepare and transmit information on the radio channel after changing state from standby to transmit (applies to conventional mode).                                     |
| Transmitter power and encoder attack time with busy/idle operations | Measures the time required for a transmitter to prepare and transmit information on the radio channel after the receiving channel changes state from busy to idle.  |

| Measurement                                     | Description  |
|---|--|
| Transmitter throughput delay                    | Measures the time it requires for audio changes in the microphone to be encoded and transmitted over the air.  |
| Transient frequency behavior                    | Measures the difference of the actual transmitter frequency and assigned transmitter frequency as a function of time when the RF output power is switched on or off.   |
| HCPM transmitter logical channel time alignment | Measures the ratio of total transmitter power under prescribed conditions and modulation to the peak power that falls in a prescribed bandwidth centered on the nominal frequency of the adjacent channel during the transmitter power ramping interval. |

## Bluetooth measurements

| Measurement                        | Description   |
|------------------------------------|---|
| Modulation characteristics         | Verifies that the modulation characteristics of the transmitted signal are correct. This measurement can only be done if the payload has the bit pattern 10101010 or 11110000.  |
| Carrier frequency offset and drift | Verifies that the carrier frequency offset and carrier drift of the transmitted signal is within the specified limits for the Basic Rate and Low Energy standards. This test can be done only if the payload contains 10101010 bit pattern.   |
| Output power                       | Verifies the maximum peak and average power emitted from the EUT. The standard recommends this test be done for a PRBS payload pattern.   |
| In-band emission / ACPR            | Verifies that the in-band spectral emission is within limits. The standard document recommends that this measurement be done with Hopping off, finding the integrated power in 1 MHz band (with RBW 100 kHz) in 80 channels starting from 2401 MHz to 2481 MHz. The integrated power values calculated in the adjacent channels are compared against recommended limits (except the three channels around transmitted frequency). This measurement is referred to as ACPR in the Basic Rate standards document. |
| 20 dB bandwidth                    | Verifies if the emissions inside the operating frequency range are within limits. This measurement is done with Hopping off. The difference between frequency points at which the power level drops to 20 dB below the peak power of the emission is found as 20 dB bandwidth.  |
| Frequency range                    | These measurements verify if the emissions inside the operating frequency range are within the limits.  |
| Power density                      | This measurement verifies the maximum RF output power density.  |
| Out-of-band spurious emission      | This measurement can be done for FCC or ETSI masks using the Spurious display.  |
| Relative power                     | Verifies the relative power in the GFSK and PSK part of the Enhanced Data Rate signal. This measurement is supported only when an Enhanced Data Rate signal is detected.  |

## LTE measurements

| Measurement                           | Description  |
|---------------------------------------|--|
| Cell ID detection                     | The Cell ID is detected from the input LTE signal.<br>For TDD and FDD.   |
| Adjacent Channel Leakage Ratio (ACLR) | The Adjacent Channel integrated power is calculated and shown. The relative power compared to the reference signal is also computed. The computed power is compared against limits suggested by the selected standard and pass/fail is reported.<br>For TDD and FDD. |

| Measurement                      | Description   |
|----------------------------------|---|
| Channel Power                    | The channel power is calculated in the channel bandwidth.<br>For TDD and FDD.   |
| Occupied Bandwidth               | The Occupied bandwidth is calculated as bandwidth containing 99% of the total integrated power in the selected span around the selected center frequency.<br>For TDD and FDD. |
| Operating Band Unwanted Emission | The power in the offset regions is calculated and presented and compared against limits set in the offset and limits table and pass/fail is reported.<br>For TDD and FDD.     |
| T <sub>OFF</sub>                 | The power in off-slot region is computed and compared against selected limits.<br>For TDD only.   |

### Pulse measurements

| Measurement               | Description   |
|---------------------------|---|
| Average ON Power          | The average power transmitted during pulse on.  |
| Peak Power                | Maximum power during pulse on.  |
| Average Transmitted Power | The average power transmitted, including both the time the pulse is on and the time it is off, and all transition times.  |
| Pulse Width               | The time from the rising edge to the falling edge at the –3 dB / –6 dB level (50%) of the user selected 100% level. Level is user selectable for Volts or Watts.  |
| Rise Time                 | The time required for a signal to rise from 10% to 90% (or 20% to 80%) of the user selected 100% level.   |
| Fall Time                 | The time required for a signal to fall from 90% to 10% (or 80% to 20%) of the user selected 100% level.   |
| Repetition Interval       | The time from a pulse rising edge to the next pulse rising edge.  |
| Repetition Rate           | The inverse of repetition interval.   |
| Duty Factor (%)           | The ratio of the width to the pulse period, expressed as a percentage.  |
| Duty Factor (Ratio)       | The ratio of the pulse width to the pulse period.   |
| Ripple                    | Ripple is the peak-to-peak ripple on the pulse top. It does not include any preshoot, overshoot, or undershoot. By default, the first 25% and the last 25% of the pulse top is excluded from this measurement to eliminate distortions caused by these portions of the pulse.<br><br>If the Amplitude units selected in the Amplitude panel (affects all amplitude measurements for the analyzer) are linear, the Ripple results will be in %Volts. For log units, the Ripple results will be in %Watts. The default for the general Units control is dBm, so the Ripple results default is %Watts.<br><br>See also <a href="#">Ripple (see page 663)</a> . |
| Ripple dB                 | The Ripple measurement expressed in dB.   |
| Droop                     | Droop is the power difference between the beginning and the end of the pulse On time. A straight-line best fit is used to represent the top of the pulse. The result is a percentage referenced to the Average ON Power.  |
| Droop dB                  | The Droop measurement expressed in dB.  |

| Measurement                         | Description   |
|-------------------------------------|---|
| Overshoot                           | The amount by which the signal exceeds the 100% level on the pulse rising edge. Units are %Watts or %Volts.   |
| Overshoot dB                        | The Overshoot measurement expressed in dB.  |
| Pulse-Pulse Phase Difference        | The phase difference between the selected pulse and the first pulse in the analysis window. The instantaneous phase is measured at a user-adjustable time following the rising edge of each pulse.  |
| Pulse-Pulse Freq Difference         | The difference between the frequency of the current pulse and frequency of the previous pulse. The instantaneous frequency is measured at a user-adjustable time following the rising edge of each pulse.   |
| RMS Freq Error                      | The RMS Frequency Error measurement is the RMS average of the Freq Error vs. Time trace, computed over the Measurement Time.  |
| Max Freq Error                      | The maximum frequency error is the difference between the measured carrier frequency of the signal and the user-selected center frequency of the analyzer.  |
| RMS Phase Error                     | The RMS Phase Error measurement is the RMS average of the Phase vs Time trace, computed over the Measurement Time.  |
| Max Phase Error                     | The phase is measured at each point during the pulse's ON time. The phase error for each point is the difference between the measured phase value and the calculated ideal phase value. After the phase error is calculated for all points in the acquisition record, the largest error in the positive direction and the largest in the negative direction are determined. Whichever of these two values has the greater absolute value is designated the Max Phase Error.   |
| Freq Deviation                      | The Frequency Deviation measurement is the difference between the maximum and minimum measured values of the signal frequency during the Measurement Time.  |
| Delta Frequency (Non-chirped pulse) | <p>The Delta Frequency measurement is the difference from the measurement frequency to each pulse frequency. Pulse frequency is calculated across the time defined by the Frequency Domain Linearity setting in the Define tab.</p> <p>The measurement is available for modulation types CW (Constant Phase), CW (Changing phase), and Other (manual) setting in the Freq Estimation tab.</p> <p>The measurement is not specified for chirp or other signals and no answer is returned when frequency estimation is set to Chirp.</p> <p>If frequency estimation is set to Other, then Frequency Offset must be set to 0 Hz and the Range can be set to <math>\pm 40\%</math> of the acquisition bandwidth.</p> <p>A least-square fit of slope of phase vs. time over the measurement period is used for the measurement of the individual pulse frequency. Frequency difference is calculated as the difference between the reference frequency and the calculated frequency of the pulse.</p> |
| Phase Deviation                     | The Phase Deviation is the difference between the maximum and minimum Phase values measured during the ON time of a pulse.  |
| Impulse Response Amplitude          | The difference in dB between the levels of the main lobe and highest side lobe.   |
| Impulse Response Time               | The difference in time between the main lobe and highest side lobe.   |
| Time                                | This is the time in seconds relative to the time reference point in the first acquisition record in the data set.   |

## Overview

The displays in General Signal Viewing (Displays > Measurements > General Signal Viewing) are:

- Amplitude vs Time
- DPX Spectrum
- Frequency vs Time
- Phase vs Time
- RF I & Q vs Time
- Spectrogram
- Spectrum
- Time Overview

These displays provide extensive time-correlated multi-domain views that connect problems in time, frequency, phase and amplitude for enabling you to more quickly understand cause and effect when troubleshooting.

## DPX Primer

With the DPX display (which displays only DPX waveforms saved by RSA5000/RSA6000/SPECMON Real-Time Signal Analyzers), you can detect and accurately measure transients as brief as 3.7  $\mu$ s. The instrument computes up to 292,000 spectrums per second of the digitized input signal. Then it displays all these spectrums as a color-graded bitmap that reveals low-amplitude signals beneath stronger signals sharing the same frequency at different times.

The strong signal in the DPX spectrum graph, shown in Figure 1, is a repeating pulse at a fixed frequency. There is also a lower-power CW signal that steps very quickly through the same span. During the pulse's on time, the power of the two signals is additive, resulting in nearly undetectable differences in the pulse envelope shape. But during the time the pulse is off, the sweeping signal is detected and shown in its true form. Both signals are visible in the bitmap because at least one full cycle of their activities occurs within a single DPX display update.




Figure 1

Compare the display of a traditional swept spectrum analyzer (Figure 2) and that of a real-time signal analyzer with a DPX display (Figure 3). The signal captured is a typical WLAN interchange between a nearby PC and a more-distant network access point (AP). The laptop signal is nearly 30 dB stronger than the AP's signal because it is closer to the measuring antenna.




Figure 2




Figure 3

The traditional swept spectrum analyzer display, Figure 2, uses line traces that can show only one level for each frequency point, representing the largest, the smallest or the average power. After many sweeps, the

Max Hold trace shows a rough envelope of the stronger laptop signal. +Peak detection was selected for the other trace in an attempt to capture the weaker but more frequent AP signal, but the bursts are very brief, so the likelihood of seeing one in any particular sweep is small. It will also take a long time to statistically capture the entire spectrum of a bursted signal due to the architecture of the swept spectrum analysis.

The DPX display, Figure 3, reveals much more insight on the same signal. Since it is a bitmap image instead of a line trace, you can distinguish many different signals occurring within each update period and/or different version of the same signal varying over time. The heavy band running straight across the lower third of the graph is the noise background when neither the laptop nor the AP is transmitting. The red lump of energy in the middle is the ON shape of the AP signal. Finally, the more delicate spectrum above the others is the laptop transmissions. In the color scheme used for this demonstration (“Temperature”), the hot red color indicates a signal that is much more frequent than signals shown in cooler colors. The laptop signal, in yellow, green and blue, has higher amplitude but doesn't occur nearly as often as the AP transmissions because the laptop was downloading a file when this screen capture was taken.

## How DPX Works

This section explains how DPX displays are created. The input RF signal is conditioned and down-converted as usual for a signal analyzer, then digitized. The digitized data is sent through an FPGA that computes very fast spectral transforms, and the resulting frequency-domain waveforms are rasterized to create the bitmaps.

The DPX bitmap that you see on screen is composed of pixels representing x, y, and z values for frequency, amplitude, and Density. A multi-stage process, shown in Figures 4a - 4d, creates this bitmap, starting with analog-to-digital conversion of the input signal.

### Simplified Flow of Multi-stage Processing from RF Input Through to Spectrum Processing:




Figure 4a. RF signals are downconverted and sampled into a continuous data stream.




Figure 4b. Samples are segmented into data records for FFT processing based on the selected resolution bandwidth.




Figure 4c. Data records are processed in the DPX transform engine




Figure 4d. Overlapping the FFTs shortens the minimum event duration required for 100% probability of intercept.

**Collecting spectral data.** Sampling and digitization is continuous. The digitized data stream is chopped into data records whose length is based on the desired resolution bandwidth (RBW). An additional requirement is placed on FFT length by the desired number of points in a trace. Table 1 shows this relationship and the FFT length is reported in the display if desired. Then the DPX transform engine performs a discrete Fourier transform on each record, continually producing spectral waveforms.

**Table 1: Minimum FFT length versus trace length – independent of span and RBW**

| Trace length (points) | Minimum FFT length |
|-----------------------|--------------------|
| 801                   | 1,024              |
| 2,401                 | 4,096              |
| 4,001                 | 8,192              |
| 10,401                | 16,384             |

As long as spectral transforms are performed faster than the acquisition data records arrive, the transforms can overlap each other in time, so no events are missed in between. Minimum event length for guaranteed capture depends on the length of the data records being transformed. An event must last through two consecutive data records in order for its amplitude to be accurately measured. Shorter events are detected and visible on screen, but may be attenuated. The DPX Spectrum RBW setting determines the data record length; narrow RBW filters have a longer time constant than wide RBW filters. This longer time constant requires longer FFTs, reducing the transform rate. Additional detail on minimum signal duration is provided in [Guaranteed Capture of Fast Events \(see page 54\)](#).

The spectral waveforms are plotted onto a grid of counting cells called the “bitmap database”. The number held by each database cell is the z-axis count. For simplicity, the small example grid used here in Figure 6 is 11x10, so our spectral waveforms will each contain 11 points. A waveform contains one (y) amplitude value for each (x) frequency. As waveforms are plotted to the grid, the cells increment their values each time they receive a waveform point.




Figure 5. Example 3-D Bitmap Database after 1 (left) and 9 (right) updates. Note that each column contains the same total number of “hits”.

The grid on the left shows what the database cells might contain after a single spectrum is plotted into it. Blank cells contain the value zero, meaning that no points from a spectrum have fallen into them yet.

The grid on the right shows values that our simplified database might contain after an additional eight spectral transforms have been performed and their results stored in the cells. One of the nine spectrums happened to be computed as a time during which the signal was absent, as you can see by the string of “1” occurrence counts at the noise floor.

**Frame updates.** The maximum rate for performing the variable-length frequency transforms that produce those waveforms can be greater than 292,000 per second. Measurement settings that slow this transform rate include narrowing the RBW and increasing the number of points for the line traces available in the DPX display along with the bitmap. Even at their slowest, spectral transforms are performed orders of magnitude faster than a physical display can respond, and also too fast for humans to see, so there's no need to update the screen or measurements at this rate. Instead, the grid collects thousands of waveforms into “frames”, each covering about 50 milliseconds (ms). A 50 ms frame contains the counts from up to 14,600 waveforms. After each frame's waveforms have been mapped into the grid, the cell occurrence counts are converted to colors and written to the DPX bitmap, resulting in a bitmap update rate of around 20 per second.

Frame length sets the time resolution for DPX measurements. If the bitmap shows that a -10 dBm signal at 72.3 MHz was present for 10% of one frame's duration (5 ms out of 50 ms), it isn't possible to determine just from the DPX display whether the actual signal contained a single 5 ms pulse, one hundred 50 microsecond ( $\mu$ s) pulses, or something in between. For this information, you need to examine the spectral details of the signal or use another display with finer time resolution, such as Frequency vs. Time or Amplitude vs. Time.

**Converting occurrence counts to color.** About 20 times per second, the grid values are transferred to the next process step, in which the z-axis values are mapped to pixel colors in the visible bitmap, turning data into information (Figure 6). In this example, warmer colors (red, orange, yellow) indicate more occurrences. The color palette is user-selectable, but for now we will assume the default “temperature” palette.

| Number of Occurrences | Color      |
|-----------------------|------------|
| 0                     | black      |
| 1                     | blue       |
| 2                     | light blue |
| 3                     | cyan       |
| 4                     | green blue |
| 5                     | green      |
| 6                     | yellow     |
| 7                     | orange     |
| 8                     | red orange |
| 9                     | red        |

**Figure 6. Example Color-mapping algorithm**

The result of coloring the database cells, Figure 7, according to the number of times they were written into by the nine spectrums, one per pixel on the screen, creates the DPX displays.




Figure 7. Color-coded low-resolution example (left) and a real DPX display (right).

In addition to the choice of palette, there are z-axis scaling adjustments for Maximum, Minimum, and Curve. Maximum sets the occurrence value that will be mapped to the highest color in the palette. Minimum sets the occurrence value for the lowest color. In the “temperature” palette, the highest color is deep red and the lowest is dark blue. Occurrence values less than the selected Minimum are represented with black pixels, while pixels that exceed the selected Maximum are red in hue but somewhat transparent. Values between Maximum and Minimum are represented by the other colors of the palette.

Adjusting the Minimum above the black default allows you to concentrate most of your color resolution over a small range of medium or higher occurrence rates to visually discriminate between different signals that have nearly equal probability values.

To see why adjustable color scaling is useful compare Figures 8 and 9. On the Scale tab, the Max control is set to 100% in Figure 8. The range of colors now covers the full z-axis range of densities from 0 to 100%. The signals used to create this bitmap are fairly diffused in both frequency and amplitude, so most pixels have low occurrence counts or density values and the upper half of the color palette is unused.




Figure 8. DPX spectrum bitmap with default color scale settings.

When the Auto Color button is selected, the Maximum control's value is set to the highest pixel value in the current bitmap, shown in Figure 9. Now none of the available colors remain unused. The entire palette is mapped to the occurrence values present at the time the button is selected, providing better

visual resolution for low densities. Selecting the Autoscale button in the DPX display scales all three axes based on current results.




Figure 9. The Auto Color function optimizes the color scale settings.

### Color Mapping Curves

The mapping between z-axis values and color does not have to be linear. The Curve control lets you choose the shape of the mapping equation. A Curve setting of 1 selects the straight-line relationship. Higher Curve numbers pull the curve upwards and to the left, concentrating color resolution on lower densities. Settings less than 1 invert the curve, moving the focus of the color range towards higher density values. Figure 10 shows the mapping curves.




Figure 10. Representative color mapping curves for the “Temperature” palette.

Using the same signal shown in Figures 8 and 9, the impact of the Curve control can be observed. With the Curve control set to 1 in the Scale tab, shown in Figure 11, the mapping between color and density is linear, so the colors spread evenly across the full density range. The color distribution is visible in the colored palette illustration to the left of the Curve control in the Settings panel.




Figure 11. Over a narrow Signal Density range, the color curve is set to 1.

When the Curve control is set to 0.5, as shown in Figure 12, the best color resolution is in the upper half of the density range, and only the dark blues are assigned to densities below 50%. Note the difference in the palette illustration.




Figure 12. Adjusting to values less than 1, increases the contrast for viewing events in the top half of the selected density range.

In Figure 13, the Curve control is increased to 3. The majority of colors shifts to the lower half of the density scale, but various shades of orange and red are still available for densities above 50%.




Figure 13. For color curve settings greater than 1, better contrast is provided for events near the low end of the density range.

## Swept DPX

DPX Spectrum is not limited in span by its real-time bandwidth. Like the regular Spectrum display, DPX Spectrum steps through multiple real-time frequency segments, building a wide-span display with line traces and the bitmap. See Figure 14.




Figure 14. Off-air ambient signals over a 1 GHz span in the swept DPX display.

The analyzer “dwells” in each frequency segment for one or more DPX frames, each containing the results of up to 14,600 spectral transforms. Dwell time is adjustable, so you can monitor each segment of the sweep for up to 100 seconds before moving to the next step. While dwelling in a segment, the probability of intercept for signals within that frequency band is the same as in normal, real-time spans: 100% capture of events as short as 10.3  $\mu$ sec.

A full pixel bitmap is created for every segment and compressed horizontally to the number of columns needed for displaying the frequency segment. Compression is done by averaging pixel densities of the points being combined together. The final swept bitmap contains a representation of the same pixel bitmap resolution, just like the non-swept bitmaps. Line traces are also created in full for each segment, and then horizontally compressed to the user-selected number of trace points for the full span.

A complex algorithm for determining the number and width of each frequency segment has been implemented. The variables in the equation include user-adjustable control settings like Span, RBW, and number of trace points, RF and IF optimization, and Acquisition BW. Installed hardware options also can affect the span segmentation. The number of segments ranges from 10 to 50 for each 1 GHz in a sweep.

A helpful piece of information for operators is the actual Acquisition Bandwidth used for capturing each segment. “Acq BW” is shown in the Acquire control panel on the Sampling Parameters tab. Acq BW is typically set automatically by the instrument, based on the needs of all the open displays, but can also be set manually. In either case, the displayed bandwidth is used for every frequency segment in the swept DPX display, though in practice, the displayed portion of the segment is somewhat narrower than the actual Acquisition BW, for performance reasons.

The entire instrument frequency range of many GHz can be covered in a DPX sweep. The Dwell Time control sets the amount of time DPX spends in each segment. This control, circled in Figure 15, can be set between 50 ms and 100 seconds.




Figure 15. During swept DPX operation, the Dwell time control adjusts the observation time of each frequency segment used to construct the composite DPX display.

## Guaranteed Capture of Fast Events

The main reason that swept-tuned and step-tuned spectrum analyzers can't provide 100% Probability of Intercept, POI, for a signal that isn't continuously present is that they spend only a short period of time tuned to each segment of their frequency span during each sweep. If something happens in any part of the span other than where it is tuned at that instant, that event will not be detected or displayed. There is also a period of time between sweeps, retrace time, during which the analyzer is not paying attention to the input signal. FFT-based analyzers, including vector signal analyzers, also miss signals during the time between acquisitions. Their POI depends on a combination of factors including span, number of FFT points, acquisition time, memory read/write time, and signal processing speed. Vector analyzers process information sequentially, so when read/write from data and processing is occurring, data is not being acquired.

RSAs, on the other hand, capture data across all frequencies within their real-time span during every acquisition. With Tektronix' exclusive Frequency Mask trigger and DPX Density trigger, POI increases to 100%, insuring capture of any spectral event matching the trigger definition. When operating in free run as a simple signal analyzer, the RSA has a POI similar to other FFT-based analyzers, with gaps between each acquisition. Processing is done concurrent with the acquisitions.

## Guaranteed Capture in DPX Real-Time Spans

The DPX display captures any signal that is at least 3.7 microseconds long (with Option 09) and within the real-time bandwidth. This performance is possible because the instrument computes up to 292,000 spectrum transforms per second. The faster the spectrum updates, the shorter the time between acquisitions and the greater the probability that any signal will be detected.

The following table shows the specified minimum signal duration (MSD) for 100% probability of intercept under various combinations of Span and RBW in DPX for a representative signal analyzer. As you can see, MSD is affected by multiple factors.

### Minimum signal duration specifications

| Minimum Event Duration 100% POI (μs) |           |            |              |           |         |
|--------------------------------------|-----------|------------|--------------|-----------|---------|
| Span (MHz)                           | RBW (kHz) | FFT Length | Spectrum/sec | Base unit | Opt. 09 |
| 110 MHz                              | 10000     | 1024       | 292,969      | 17.3      | 3.7     |
|                                      | 1000      | 1024       | 292,969      | 19.5      | 5.8     |
|                                      | 300       | 2048       | 146,484      | 28.5      | 14.8    |
|                                      | 100       | 4096       | 73,242       | 37.6      | 37.6    |
|                                      | 30        | 16384      | 18,311       | 134.6     | 134.6   |
|                                      | 20        | 32768      | 18,311       | 229.2     | 229.2   |
| 40 MHz                               | 5000      | 1024       | 292,969      | 17.5      | 3.9     |
|                                      | 1000      | 1024       | 292,969      | 19.4      | 5.8     |
|                                      | 300       | 1024       | 146,484      | 25        | 11.4    |
|                                      | 100       | 2048       | 73,242       | 37.6      | 30.8    |
|                                      | 30        | 4096       | 36,621       | 93.6      | 93.6    |
|                                      | 20        | 8192       | 18,311       | 147.3     | 147.3   |
|                                      | 10        | 16384      | 18,311       | 194.5     | 294.5   |
|                                      | 20        | 4096       | 73,242       | 133.9     | 133.9   |
|                                      | 10        | 8192       | 36,621       | 267.8     | 267.8   |

To demonstrate the POI in action, a challenging bi-stable signal is used. A CW sinusoid sits at 2.4453 GHz most of the time, but every 1.28 seconds, its frequency changes for about 100  $\mu$ s before returning to normal. The duty factor of this transient is less than 0.01%.

Figure 16 shows a swept analyzer set up for a 5-second sweep of its MaxHold trace. It shows that there is something occurring around the signal. This sweep rate was empirically determined to be the optimum rate for reliable capture of this signal in the shortest time. Faster sweep times can reduce the probability of intercept and result in fewer intersections of the sweep with the signal transient.




Figure 16. Swept spectrum display of the infrequent transient.

The DPX display shown in Figure 17 shows the exact same event, also captured over a 5 second period. A lot more information can be discovered about the transient. It is obvious at first glance that the signal is hopping by about 3 MHz, with 1.2 MHz of frequency overshoot on transitions




Figure 17. The DPX spectrum display after 5 seconds. The MaxHold trace is cyan.

### Guaranteed Capture in DPX Swept Spans

Probability of intercept (POI) for signals within a single segment, while DPX is dwelling in that segment, is the same as for non-swept DPX operation. 100% POI for events as brief as 3.7 microseconds long (with Option 09). But just as in traditional swept analyzers, during the time the acquisition is tuned to any one segment, the analyzer is not monitoring signals in any of the other segments, so probability of capture in segments other than the current one is zero. Because of the wide real-time bandwidth, the number of segments needed to cover the span is much less than for swept analyzers, so the overall probability of intercept is significantly better for DPX sweeps.

Another factor affecting POI is number of trace points. The bitmap is always 801 points wide, but the line traces allow user selection for number of points. 801 is the default and the other choices are 2401, 4001, and 10401. Frequency transforms for traces containing more than 801 points take longer, and this lower waveform update rate increases the minimum signal duration proportionally. This caution applies for swept and non-swept operation. The trace length control is on the Prefs tab in the DPX control panel.

### DPX Density Measurements

“Density” is a measure of the amount of time during a defined measurement period during which signals are present within a particular area of the DPX Spectrum bitmap. A clean CW tone gives a 100% reading, while a pulse that is on for one microsecond out of every millisecond reads 0.1%. This section describes how density is computed from hit counts.

If we plot 41 more waveforms into the example grid we used previously in Figure 6 (in addition to the nine we already plotted), each column ends with a total of 50 hits (Figure 18). The density for any one cell in a column is its own count value divided by 50, expressed in percent as shown in Figure 19. The math is very simple: a cell with 24 counts has a 48% density. In practice, instead of batches of 50 waveforms, we collect a frame of thousands of waveforms before each update to the density bitmap.

|    |    |    |    |    |    |    |    |    |    |    |
|----|----|----|----|----|----|----|----|----|----|----|
|    |    |    |    |    |    |    |    |    |    |    |
|    |    |    |    |    | 4  |    |    |    |    |    |
|    |    |    |    |    | 6  |    |    |    |    |    |
|    |    |    |    |    | 13 |    |    |    |    |    |
|    |    |    |    |    | 18 |    |    |    |    |    |
|    |    |    |    | 1  | 3  | 1  |    |    |    |    |
|    |    | 1  | 2  | 4  |    | 4  |    |    |    |    |
| 1  | 2  | 1  | 43 | 41 | 2  | 38 | 6  | 1  | 2  | 2  |
| 47 | 46 | 47 | 5  | 4  | 3  | 7  | 43 | 47 | 46 | 47 |
| 2  | 2  | 1  |    |    | 1  |    | 1  | 2  | 2  | 1  |

Figure 18. Grid showing cell counts after 50 waveforms. For each column, the sum of z-axis values is 50.

|     |     |     |     |     |     |     |     |     |     |     |
|-----|-----|-----|-----|-----|-----|-----|-----|-----|-----|-----|
| 0%  | 0%  | 0%  | 0%  | 0%  | 0%  | 0%  | 0%  | 0%  | 0%  | 0%  |
| 0%  | 0%  | 0%  | 0%  | 0%  | 8%  | 0%  | 0%  | 0%  | 0%  | 0%  |
| 0%  | 0%  | 0%  | 0%  | 0%  | 12% | 0%  | 0%  | 0%  | 0%  | 0%  |
| 0%  | 0%  | 0%  | 0%  | 0%  | 26% | 0%  | 0%  | 0%  | 0%  | 0%  |
| 0%  | 0%  | 0%  | 0%  | 0%  | 36% | 0%  | 0%  | 0%  | 0%  | 0%  |
| 0%  | 0%  | 0%  | 0%  | 2%  | 6%  | 2%  | 0%  | 0%  | 0%  | 0%  |
| 0%  | 0%  | 2%  | 4%  | 8%  | 0%  | 8%  | 0%  | 0%  | 0%  | 0%  |
| 2%  | 4%  | 2%  | 86% | 82% | 4%  | 76% | 12% | 2%  | 4%  | 4%  |
| 94% | 92% | 94% | 10% | 8%  | 6%  | 14% | 86% | 94% | 92% | 94% |
| 4%  | 4%  | 2%  | 0%  | 0%  | 2%  | 0%  | 2%  | 4%  | 4%  | 2%  |

Figure 19. Grid after converting occurrence counts to percent density values. The sums of the cell density measurements within each column are all 100%.

### Measuring Density with Markers

Hit counts are cleared after every frame update, as long as Persistence is not turned on. The density value for any pixel is simply the percent of time it was occupied during the most recent 50 ms frame. Markers can be used to see the Density value for one or more individual points on the screen, enabling measurements of the signal density at any interesting point in the DPX display.

In Figure 20, Wireless LAN signals are analyzed in the presence of a Bluetooth radio signal in the 2.4 GHz ISM band.




Figure 20. DPX display of WLAN and Bluetooth signals, with a marker on the highest signal.

The “Marker to Peak” function was used to find the peak signal recorded in the display. The marker readout in the upper left corner of Figure 20 shows the Density, Amplitude, and Frequency for the pixel you selected with the marker. By adding additional markers, you can measure the signal density differences between multiple signals of interest.

### Marker Peak Search in the DPX Bitmap

Markers on the DPX bitmap can search for peaks, similar to marker peak searching on spectrum line traces. For a human, it is pretty easy to discern “signals” in the bitmap picture. Your brain intuitively identifies strings of contiguous bright pixels. This isn't so easy for a computer. The first thing the RSA must do for any peak search is analyze pixel density values to identify apparent signals. Then it can sift through these density peaks for the amplitude peaks you want to find.

Z-axis density values for the pixels in each column of the bitmap are internally converted into histograms to find density peaks indicating the presence of signals. Table 2 shows the five middle columns from the example grid we used to illustrate density measurements in a previous section (Figure 19). The density values for each pixel in the middle, highlighted column are plotted on the y axis in the bar chart in Figure 21. The bar chart x axis is bitmap row number, numbering from the top of the table.

Table 2: Bitmap section showing density values.

|     |     |            |     |     |
|-----|-----|------------|-----|-----|
| 0%  | 0%  | <b>0%</b>  | 0%  | 0%  |
| 0%  | 0%  | <b>8%</b>  | 0%  | 0%  |
| 0%  | 0%  | <b>12%</b> | 0%  | 0%  |
| 0%  | 0%  | <b>26%</b> | 0%  | 0%  |
| 0%  | 0%  | <b>36%</b> | 0%  | 0%  |
| 0%  | 2%  | <b>6%</b>  | 2%  | 0%  |
| 4%  | 8%  | <b>0%</b>  | 8%  | 0%  |
| 86% | 82% | <b>4%</b>  | 76% | 12% |

**Table 2: Bitmap section showing density values. (cont.)**

|     |    |           |     |     |
|-----|----|-----------|-----|-----|
| 10% | 8% | <b>6%</b> | 14% | 86% |
| 0%  | 0% | <b>2%</b> | 0%  | 2%  |




Figure 21. Bar chart of the density values in the bolded column of Table 2.

Assume that Density Threshold is set to 5% and Density Excursion to 5% also. Starting with  $x=1$  in the bar chart, test each bar against the threshold. The threshold criteria is met at  $x=2$ . Keep testing until you find a bar that is shorter than the previous bar by at least the Excursion setting. In this case it is  $x=6$ . This tells us that a “signal” covers rows 2 through 5. Its density peak is at row 5.

Now you can look for another peak. Continue looking at bars to the right and you will find a density value at row 9 that meets the threshold criteria, but since there are no bars to the right of it that meet the excursion criteria, we can't declare row 9 a signal because it fails to meet the excursion criteria. If row 1 had 1% density, then row 9 would be a density peak.

Once density peaks are found for all columns in the bitmap, we can start looking for the amplitude peaks. When the **Peak** button is selected, the analyzer checks the histograms of every column in the bitmap and finds the density peak with the highest amplitude. The amplitude search has its own versions of Threshold and Excursion settings, but in dBm and dB units. When **Next Peak Down** command is given, the search will scan inside the current column for the next density peak. **Next Peak Right** examines each column to the right of the current marker location to locate density peaks that also meet the amplitude peak criteria.

To demonstrate the value of marker peak search in the DPX bitmap, we will use the time-multiplexed signals showing multiple amplitude levels from an example earlier in this manual. The **Peak** button and its menu equivalent place the active marker on the peak signal in Figure 22. The peak signal is the density peak of highest amplitude in the bitmap.




Figure 22. The marker was positioned by selecting the Peak button. Density, frequency, and amplitude measurements at the marker are displayed in the upper left corner of the graph.

The Marker Toolbar, at the bottom of Figure 22, allows easy navigation of peak signals (**Peak Left**, **Peak Right**, **Next Peak Up**, or **Next Peak Down**). Selecting the arrow keys enables the marker to search for amplitude/density peaks at other frequencies, while the **Next Peak Up** and **Next Peak Down** arrows enable the marker to search for other high-density points at the same frequency.

In the Define Peaks tab of the Define Markers control panel, Figure 24, you can adjust the density threshold and excursion controls to modify search behavior. The amplitude threshold and excursion controls also apply to DPX marker searches. Smoothing keeps the marker from finding multiple peaks within the same apparent signal by averaging an adjustable number of pixel densities together, but it does not affect the single-pixel measurement readout displayed by the marker.




Figure 24. Amplitude and Signal Density controls can be adjusted to define Peak search behavior.

### Density Measurements over an Adjustable Area (“The Box”)

The density for a single pixel is its ratio of actual hits vs. possible hits over a defined time period, and markers display these density values. For measuring density over an area larger than one pixel, the DPX display includes a measurement box you can resize and drag around in the DPX display with your mouse or finger.

If you could make the box so narrow that it contained only points within a single column of pixels, the density of this area would be the sum of the included pixels' density values. For example, if the box was three pixels tall and the density values for these pixels were 4, 2, and 7% respectively, the overall density for the three-pixel area would be 13%. Imagine a box one pixel wide and as tall as the graph. Assume that the input signal's amplitude was such that all hits fell at or near the vertical center of the screen. Since 100% of the waveforms written to the bitmap passed through the box, the density for the box is 100%.

$$\text{Density of an area} = \frac{(\text{Sum of densities of all pixels})}{(\text{Number of columns})}$$

When you widen the box to cover a broader range of frequencies, software computes the density sum for the included pixels in each column inside the box. The aggregate density value for this box is the average density, calculated by adding the column density sums then dividing by the number of columns. For a 100% result, there must not be any hits above the top edge of the box or below its bottom edge. In other words, every waveform drawn across the graph entered the box through its left side and exited the box through its right side, with no excursions out the top or bottom. Figure 24 demonstrates this principle on a CW signal. As you can see on the left-hand side, no amplitudes exist above or below the box; the density of the signal is 100%. On the right hand side, there are signals below the box, therefore the density is less than 100%.




Figure 24. Density of signals defined within an area. Left: Correct measurement of a CW signal. All columns in the box include the signal. Right: Incorrect measurement area. The measurement is accurate, but probably not what you expected. Some columns in the box contain no hits, so they contribute zeros to the calculation of average density.

The density measurement box' vertical size and location are always set in dB and dBm, no matter what units you have selected for measurements. (Amplitude control panel > Units tab.) The box is not draggable when the selected units are linear (such as Amps, Volts, Watts...), though you can still adjust its size and location using the Frequency and Amplitude controls in both the DPX Settings > Density and Trigger > Event tabs. Since the vertical scale is non-linear, a box of constant amplitude changes visual height as it changes vertical position, a disconcerting effect if you are trying to drag it.




Figure 25. DPX Density control panel is used to define the area of interest for DPX density measurements.

A readout will appear somewhere in the graph. If the box is off-screen, the readout will be accompanied by an arrow pointing towards the invisible box. Grab this readout with your mouse or finger and drag the density readout to the area you want to measure.

To adjust the box size, a mouse is the easiest way to drag the sides and corners of the rectangle. For precise settings, use the knob, arrow keys, or keyboard to adjust frequency and amplitude values for the rectangle. These controls are located in the right half of the Density tab in the control panel.

## Persistence

Previous sections of this topic have assumed that persistence was not applied to the DPX bitmap. Without persistence, hit counts in the grid are cleared after each frame update. Now we will describe how persistence modifies this behavior, starting with infinite persistence because it is simpler than variable persistence.

Hit counts are not cleared between frames if infinite persistence is enabled. When the instrument is set up for continuous acquisitions, hits keep collecting until you stop acquisitions or click the Clear button above the DPX display. Software keeps track of the total number of waveforms computed during the entire collection period. Density equals the total number of hits to a cell divided by the total number of waveforms.

Variable persistence is trickier. A single-occurrence signal shown in the bitmap does not disappear suddenly upon the next frame update, nor does it linger forever. It fades gradually away. The user sets a time constant for the Dot Persistence control which determines how long it takes for signals to fade. Fading is accomplished by reducing the hit count in every cell, after each frame update, by a factor based on the persistence time constant. The longer the time constant, the less the hit counts are reduced.




Figure 26. Example of fast transient discovery with and without variable persistence turned on. In the display on the left, with variable persistence of 10 seconds, the occasional sub-second transient that spikes up above the normal signals is held in the display rather than disappearing as soon as the signal goes away. The display on the right, with persistence turned off, requires watching the display continually to see the brief signal.

Not only are single-occurrence signals allowed to remain in the display for awhile by variable persistence, additional hits keep piling on. The result is that cell values are no longer pure hit counts; they include counts due to new hits from waveforms plus proportionally reduced counts from prior frames. As part of translating hit counts into density values, a new software algorithm uses a finite-series equation to discriminate between the effects of persistence and the arrival of new hits. The inflationary effects of persistence on cell counts are removed, so density readings represent the true ratio of actual hits to possible hits over the persistence interval.

The density computation for variable persistence is a very good estimate of true signal density, with errors of less than 0.01%. For exact density measurements, use either no persistence or infinite persistence.

Another subtlety of persistence is its smoothing effect on the density measurement of intermittent signals. Consider a pulse that is on for 10 ms and off for 90 ms of each 100 ms cycle. We'll make the simplifying assumption that the pulse ON time always falls entirely within a single DPX frame update (50 ms). If persistence is not applied, the density measurement is computed on each individual frame. The results will be 20% for each frame containing the ON time and 0% for the other frames. If infinite persistence

is enabled, however, the density measurement will settle to 10% after the second frame, and remain at this value for as long as the pulsing continues. With persistence, the density is effectively computed over many frames.

### Persistence Effects on Density

Persistence does not alter colors in a density-based bitmap. Its effect is to extend the amount of time over which densities are calculated, leaving signal events visible for the persistence duration.

Before the introduction of density measurements and extra-long hit counters, persistence caused colors to “bloom”, becoming more and more intense over time as the hit counts increased. Longer persistence intervals caused increased blooming, turning crisp signals into fat red stripes. When hit counts are converted to density values, the display is not subject to this effect. As long as the input signals maintain reasonably stable repetition rates and duty ratios, their density values will also remain stable despite ever-increasing hit counts in the underlying grid cells.

If you are accustomed to the original hit-count-based persistence displays, it may seem counter-intuitive that repeating signals in a density-based bitmap will not get brighter and redder over time with infinite persistence. A quick review of the density algorithm explains why: the hit count is divided by the total number of waveforms over the persistence interval. For example, if a signal occupies a pixel 50% of the time over a period of 15 minutes, the density reading will be 50% throughout the entire 15 minutes, though the underlying hit count is steadily increasing.

### Z-Axis Resolution

Another factor that can cause color bloom is overflow of the hit counters. If a pixel could only count up to 1000 hits, its density and color values would clip at 100% after just 1000 hits, even if waveform points continue to arrive in the same pixel location. With waveform points being written to the bitmap at rates approaching 300k/sec, counts add up really fast for highly-repetitive signals. Deeper counters permit higher hit counts, so overflow happens much later, as shown in Table 4.

**Table 3: Comparison of DPX z-axis resolution and its effect on saturation.**

| Hit Count  | 36-bit custom float (equivalent to 33-bit integer) |
|--|--|
| Maximum Hit Count  | $2^{33}$   |
| Minimum Time until Overflow (for pixels with 100% density) | 8.1 hours  |

Clipping due to overflow of the counters in one or more cells will not occur until hours have passed, or even days.

One more benefit to having deeper hit counters is better visual resolution of density. RSAs with the highest-performance DPX hardware installed use floating-point numbers to count hits, allowing us to count billions of waveforms while retaining one-hit resolution, providing better than 99 dB of dynamic range for density measurements. Density measurements in  $\mu\%$ ,  $n\%$ , and even  $f\%$  ranges are quite possible for extremely rare signals captured with infinite persistence.

With straight-line mapping between density and color (Curve setting of 1), resolution is fixed by the number of colors in the palette. For non-linear mappings (Curve settings higher or lower than 1), most

of the colors are concentrated at either the low or high end of the density scale, so you can visually discriminate finer differences between density values in that range.

## Persistence Adjustments

Dot Persistence can be enabled for the “Bitmap” trace using the Settings control panel. The Persistence can be displayed as Infinite or Variable. For Variable Persistence, you can select the time constant for fading in seconds as shown in Figure 27.




Figure 27. The trace settings control panel allows user control of persistence parameters.

Figure 28 demonstrates the observed behavior of variable persistence when a CW signal, represented in the first frame, is turned off. Even if the event was instantaneous and was confined within a single frame, you will observe the color changing to indicate lower and lower density values, until the signal finally disappears entirely.




Figure 28. With variable persistence, a brief CW signal captured by DPX remains in the display for an adjustable period of time before fading away.

## DPX Density Trigger

The standard DPX display shows you a clear picture of transients and other hard-to-find signals and goes well beyond helping you discover these difficult to find signals by actually triggering on their appearance to capture them into acquisition memory for in-depth analysis. If you can see it in the DPX bitmap, you can trigger on it.

Other trigger methods can detect signals that exceed an amplitude threshold, or even a sophisticated amplitude-vs-frequency mask, but they can't find a signal at a particular frequency if another signal of higher amplitude is sometimes present at that same frequency. The Runt trigger addresses some of these signal-under-signal cases, but not all. As shown in Figure 29, the DPX Density trigger can discriminate signals within a precise amplitude-frequency range without the operator having to know any characteristics of the target signal besides where it might show up in the DPX Spectrum graph.




Figure 29. Example of Density Trigger. Left: A free-run DPX display showing pulses with varying frequency. Occasionally, a short pulse in the middle appears for a split instant, but it is hard to capture it with just a Run/Stop button. Right: The triggered DPX displays shows the low-amplitude pulse that was not apparent in the untriggered display. The analyzer was set to trigger whenever the average density in the user-drawn box measured 50% or higher.

The DPX Density trigger uses the same screen-based measurement box as the DPX Density measurement. While the target signal is absent, the density measurement characterizes the “normal” signals within the box. When the target signal finally appears, the density value increases. The trigger system monitors the density measurement and activates a trigger whenever the density value exceeds the adjustable density threshold. The only thinking you have to do is to set this threshold to a level somewhere between the normal density readings and the density due to the trouble-making signal. However, the instrument software can compute the threshold value automatically.

### Trigger On This™

The Trigger On This™ function allows you to point and click to set up the DPX Density trigger. By right-clicking on a spot within the DPX display, or pressing and holding your finger on the touchscreen display for about a second, a menu selection will appear. Selecting **Trigger On This** causes a DPX Density box to appear and automatically adjusts the threshold. The DPX display will now only update whenever the automatic threshold is exceeded. Subsequently, if needed for your signal, open the Trigger control panel to adjust the density threshold or the size of the measurement box until the event is reliably captured.

### Automatic Threshold Adjustment by Trigger On This™

The trigger density threshold automatically set by Trigger On This is 80% of the measured value. If the signal was present at the moment you selected Trigger On This, the threshold will be 20% less than the signal density, so the next time the signal is present long enough (or present enough times) to exceed the threshold density, it will cause a trigger. If the signal happened to be missing when you selected Trigger On This, the threshold value will be even lower. If you clicked in a part of the display with no signal activity at all, the threshold will be set to zero. Any signal that shows up here will fire the trigger, as shown in Figure 32.




Figure 30. The analyzer triggered when the density in the DPX measurement box exceeded the threshold set by Trigger On This. You can see in the Spectrogram and Frequency-vs-Time displays that the signal event which caused the trigger was a quick frequency hop. The Time Overview shows that the signal amplitude never changed, so a power level trigger would not have worked.

### DPX Density Trigger Timing

The time resolution for DPX density measurements is the frame length, around 50 ms. A basic implementation of the DPX Density trigger concept is also frame-based, so a trigger event that occurs anywhere within a frame will not be recognized until the end of the frame. Therefore, the worst case trigger uncertainty is 50 ms.

DPX Density trigger doesn't always have to wait until the end of a frame before firing. For the common configuration of triggering when the measured density is higher than the threshold, the density measurement in the trigger can be computed many times within each frame and it can fire the trigger as soon as the threshold is exceeded.

Consider the case where the threshold is zero. As soon as a single waveform causes a hit within the measurement box, we know that the density is greater than zero. It takes a little longer to test for a 5 or 10% density, and even more time for thresholds at or near 100%.

The DPX Density trigger can also be set to fire when the measured density is below the threshold value. This is useful when you suspect that your signal is missing some of the time. For a signal that is supposed to be CW, you can set the trigger controls to acquire when the density measurement of the signal peak drops below 100%. When using the “lower than” form of the DPX Density trigger, the time resolution is

one frame because of the following logic: We can't be sure the actual density is less than, say, 15% until at least 85% of the full test time has elapsed. In order to keep things simple and fast in the trigger module, the RSA just waits until the end of each 50 ms frame to do the "lower than" comparisons.

### Persistence and DPX Density Trigger

The smoothing effect of persistence on density measurements can help in determining a good threshold value. With persistence turned off, an infrequent signal's density reading jumps between higher and lower values as it turns on and off, and it can be hard to read these flashing numbers. By turning persistence on, you instruct the instrument to average the density over a longer time period. This density result is somewhere between the ON and OFF density values - the very definition of a good trigger threshold.

Unlike the DPX Density measurement, the DPX Density trigger is not affected in any way by persistence. Density calculations in the trigger system are made with hit count data received from each individual DPX frame, before any persistence is applied. Even when the density measurement reading in the display is averaged over many frames due to persistence, the trigger is computing density for each frame and comparing these quick snapshots against the threshold setting.

## DPX Display Overview

The DPX display enables you to see how traces change over time and thus displays signal events that cannot be seen on a swept spectrum analyzer. A DPX Spectrum indicates how traces change in two ways. First, it uses color shading to show how consistent the shape of a trace is. Second, it uses persistence to hold signals on the screen so you can see them longer.

## DPX Display

The DPX display works by using a two-dimensional array to represent points on the display. Each time a trace writes to a point on the display, a counter in the array is incremented. A color is assigned to each point in the display based on the value of its counter. Thus, as acquisitions occur over time, a colored waveform, the Bitmap, develops on the display that shows how frequently each display point has been written to.

An important feature of the DPX display is dot persistence. Dot persistence sets how long a point on the display will be visible. You can set the Dot Persistence to be Variable or Infinite. In variable persistence mode, you specify a decay period that limits how long a point will be displayed. In infinite persistence mode, once a point in the display has been written to, it will remain visible indefinitely.

The DPX display can plot the trace in the following views:

- **Spectrum** – This view plots power on the vertical axis versus frequency on the horizontal axis. This display is similar to a standard Spectrum display.
- **Zero Span** – This view plots power on the vertical axis versus time on the horizontal axis. This display shows how the power level at the center frequency changes with time.


- **Frequency** - This view plots frequency on the vertical axis versus time on the horizontal axis. This displays how frequency changes over time, where the center frequency is displayed at the center of the vertical axis.
- **Phase** - This view plots phase on the vertical axis versus time on the horizontal axis. This displays how phase changes over time, where the zero degree phase position is displayed at the center of the vertical axis.
- **DPXogram** - This view is a spectrogram version (time plotted along the vertical axis versus frequency on the horizontal axis) of the DPX Spectrum trace. DPX spectrogram creates the spectrogram in real time, and does not require an acquisition to be transferred into memory and analyzed. Because of this real time processing, there are no gaps in the spectral lines, even for monitoring periods that can last for several days.

The DPXogram has the following limitations:


- DPXogram cannot sweep a range greater than the maximum real-time bandwidth. When the instrument is sweeping, the DPXogram display shows the *Disabled - data is from swept acquisition* message.
- No overlap, however, DPXogram can display multiple Spectrums/line.
- The DPXogram monitors in real time and can be used as a monitor while triggers are occurring. It cannot be set to display only triggered lines.
- **Split** - This view consists of two DPX views. A DPXogram view appears on the top half of the display and a DPX Spectrum view appears on the bottom half of the display.

To display a DPX view:

1. Select **Freq** and use the front panel knob or number keys to set the measurement frequency.
2. Select the **Displays** button or **Setup > Displays**. This displays the **Select Displays** dialog box.
3. From the **Measurements** box, select **General Signal Viewing**.
4. Select **DPX** from the **Available displays** box.
5. Click the **Add** button. This will add the DPX icon to the Selected Displays box (and remove it from the Available displays box).
6. Click the **OK** button. This displays the DPX Spectrum view.
7. Select the desired view from the drop-down list on the left side of the graph.




DPX Zero Span view



DPX Split View


## Elements of the DPX Display



| Item | Display element         | Description   |
|------|-------------------------|---|
| 1    | Vert Position           | Sets the top of graph value. This is only a visual control for panning the graph. The Reference Level is adjusted in the Toolbar and the Ampl control panel. By default, Vert Position = Ref Level.   |
| 2    | dB/div                  | Sets the vertical scale value. The maximum value is 20.00 dB/division.  |
| 3    | RBW                     | Sets the resolution bandwidth. Note that when the RBW is set to Auto, its value is italicized.  |
| 4    | DPX view                | Selects the DPX view. Choices are Spectrum, Zero Span, Frequency, Phase, DPXogram, and Split.   |
| 5    | Autoscale               | Adjusts the Vertical and Horizontal scaling to display the entire trace on screen.  |
| 6    | Pos/CF                  | Spectrum: Center Frequency - Adjusts the analyzer center frequency. For Zero Span, Frequency, or Phase the Position is in seconds.  |
| 7    | Span/Scale, Sweep/Scale | Spectrum display: Span - Adjusts frequency range of the measurement. Scale - If Horizontal scale has been manually adjusted in Settings > Scale, then this control adjusts the visual graph scaling without affecting the Span. Zero Span, Frequency, Phase displays: Sweep - adjusts the trace duration in seconds. Scale - adjusts the visual graph scaling without affecting the Sweep time. |
| 8    | Clear                   | Erases the bitmap and traces in the graph and restarts multi-trace functions (Avg, Hold).   |
| 9    | Function                | Readout of the Detection and Function selections for the selected trace.  |

| Item | Display element | Description  |
|------|-----------------|--|
| 10   | Show            | Controls whether the selected Trace is visible or not. When trace is Off, the box is not checked.  |
| 11   | Trace           | Selects a trace. Touching here pops up a context menu listing the available traces, whether they are enabled or not. If user selects a trace that is not currently enabled, it will be made enabled. |

### Additional Elements of the DPXogram Split Display



| Item | Display element | Description   |
|------|-----------------|---|
| 1    | Spectrums/line  | Appears only when the display is stopped. Readout of the number of spectrum lines represented by each line of the DPXogram display. This value changes when the Time/div or Time resolution settings are changed.   |
| 2    | Color scale     | Legend at the right side of the DPX Spectrum display. This element illustrates the relationship between the colors in the DPXogram plot and the amplitude axis of the DPX Spectrum plot. This scale changes with Color (DPXogram) palette selection and Max and Min settings on the Ampl Scale tab. |
| 3    | DPXogram trace  | The selected line in the DPXogram graph can be shown in the DPX Spectrum graph of the Split view. The most recent DPXogram line, usually at the bottom of the graph, is selected by default. If any markers are on, the selected marker determines the selected line.                               |

### Time Resolution of DPXogram Display

Due to the large amount of data produced by the DPX hardware during acquisitions, a compressed version of the plot is shown while running. This plot is limited to 500 lines, with each line having 267 points. However, a much longer record, with higher frequency resolution is being collected. As soon as the instrument is stopped, this underlying data is shown, replacing the temporary version. There are 50 lines in each vertical division of the 2-D DPXogram plot, so the time resolution of the graph is Time/div

divided by 50. However, you can set the instrument to collect multiple spectra per line, allowing you to zoom in later on this high-time-resolution data.

$$\text{Time Resolution} = \frac{10 \times \text{Time/Division}}{500}$$

When the DPXogram display is stopped, the analyzer can display the full resolution of the captured data. The Time Resolution readout applies only when the DPXogram is running.

**Effects of Changing Time Resolution.** The Time Resolution control affects acquisition parameters for the DPX hardware. This means that if you change the Time Resolution value while the instrument is stopped, the new value applies to the NEXT acquisition, and might not represent the results currently shown in the display.

Time resolution can be changed either directly, by manually adjusting the Time Resolution control, or automatically, by changing the Time/div control. Auto is the default, yielding one spectrum per line in the display. When the Time Resolution is decreased below its auto value, multiple spectra are collected to create each line in the DPXogram graph. Once you stop the instrument, you can decrease the Time/div value or use Zoom to see increased time resolution.

If the time resolution is set to a very small number while the Time/div is set to a large value, you might notice that there is a limit to the number of spectra that can be collected. This limit depends on the number of trace points selected. For 801-point spectra, 60,000 underlying spectra can be collected. The number of 2401-point spectra collected is 20,000, and for 4001-point spectra the number is 12,000. When the limit is reached, the oldest spectra are discarded as newer spectra are captured.

### Touchscreen Actions on Markers in the Graph Area

| Action  | Description  |
|---|--|
| Mouse click within 1/2 div. of a marker                 | Selects the marker and updates the marker display to show the selected marker's values.  |
| Touch marker to select and then use knob, or arrow keys | Adjust the setting associated with the Marker.   |
| Touch and drag a marker                                 | Changes marker position to the "drop point". You can use Tools > Options > Prefs to change whether markers jump from one peak to the next while dragging or move smoothly along the trace. |

### Available Traces for Display – Standard Instrument


Five traces can be shown in the DPX display— one bitmap and four line traces. The default traces are Bitmap and +Peak detection. The other three traces are –Peak detection, average detection, and math.

| Trace            | Description   |
|------------------|---|
| Bitmap           | Displays the density of acquired data. The number of data points acquired at each pixel (representing a particular amplitude level at a specific frequency at a point in time) is indicated by color. |
| +Peak detected   | Line trace. Displays the maximum values acquired in each update. Normal and Hold functions are available for this trace.  |
| -Peak detected   | Line trace. Displays the minimum values acquired in each update. Normal and Hold functions are available for this trace.  |
| Average detected | Line trace. Displays the average of all the values acquired in each update. Normal and Hold functions are available for this trace.   |
| Math Trace       | Line trace. Displays the difference between two traces. The two traces used are set in the Traces tab of the Settings panel.  |

### Determining Which Trace Types Are Displayed

You can see the status of all the traces by selecting the Trace drop-down list. Traces that are not displayed are preceded by "Enable". In the following figure, you can tell that the Bitmap and -Peak Trace traces are displayed but the +Peak, Average, and Math Traces are not displayed.

Selecting **Enable -Peak Trace** from the Trace list displays the -Peak detected values trace.




You can see whether a trace is enabled by looking at its **Show** check box. The "selected trace" is selected in the Trace list. The Show check box is checked when the selected trace is enabled. To the right of the show box are readouts for detection of the selected trace (+Pk, Avg (VRMS), ) and its function (Hold, Normal,). You can enable/disable the selected trace by checking or unchecking **Show**.



### Selecting Traces for Display – Standard Instrument

To select a trace for display:

1. Use the Settings control panel:
  - Select **Setup > Settings** or click the **Settings** button.
  - Select the **Traces** tab.
  - Select the trace from the drop-down menu.
  - Select the **Show** check box.
2. Select a trace from the Trace drop-down list.



### Specifying How +Peak, –Peak, and Average Traces Are Displayed

You control how the +Peak, –Peak and Average (Avg (VRMS)) traces are displayed from the Traces tab of the Settings control panel. From the Traces tab, you can also specify whether these traces display results from single updates or results collected over multiple updates.

To change how the traces are displayed:

1. Select **Setup > Settings** or click the **Settings** button.
2. Select the **Traces** tab.
3. Select the trace type from the drop-down list.
4. If you select +Peak Trace or -Peak Trace, use the **Function** drop-down list to select either **Normal** or **Hold**.
  - a. Select **Normal** to set the trace to display the maximum/minimum values acquired in each individual update.
  - b. Select **Hold** to set the trace to display the maximum/minimum values acquired over time. The trace values are updated only if they exceed the existing values.
5. If you select Avg (VRMS) or Avg (of logs), use the **Function** drop-down list to select either **Normal**, **Average (VRMS)**, or **Avg (of logs)**.
  - a. Select **Normal** to set the trace to display the average values acquired in each update.
  - b. Select **Average (VRMS)** or **Avg (of logs)** to set the trace to display an average of the average values. Use the Count box to enter the number of times the trace is averaged.

### Available Traces for Display

Five traces can be shown in the DPX Spectrum, Zero Span, Frequency, and Phase displays— one bitmap and four line traces. The default traces are Bitmap and Trace 1. The other three traces are 2, 3, and Math. Line traces 1, 2, and 3 have user-selectable Detection and Function settings. The final line trace is Math, allowing you to subtract one line trace from another.

For the DPXogram display, only one trace is available – the DPXogram trace.

For the Split display, which consists of a DPXogram display on the top half and a DPX Spectrum display on the bottom half, you can display the Bitmap trace, Trace 1, Trace 2, Trace 3, Math trace, and Ogram Line (the selected line from the DPXogram display) on the bottom half of the display.

### Selecting Traces for Display

This is done almost the same as with the standard instrument, except that the choices available for Trace are different. Instead of +Pk, -Pk, and Avg traces, you select Trace 1, 2, and 3 and independently set the Detection method for each of these traces to +Pk, -Pk, Avg (VRMS), or Avg (of logs).

**Reference.** [Changing the DPX Spectrum Display Settings \(see page 470\)](#)

## DPX Settings

**Menu Bar:** Setup > Settings


**Front Panel / Application Toolbar:** Settings

The measurement settings for the DPX display are shown in the following table.

| Settings tab   | Description   |
|--|---|
| <a href="#">Freq &amp; Span (see page 118)</a>           | Sets frequency and span parameters for the DPX display. This tab appears for the Spectrum and DPXogram displays.  |
| <a href="#">Params (see page 77)</a>                     | Sets sweep time and scroll settings. This tab appears only for the DPX Zero Span, DPX Frequency and DPX Phase displays.   |
| <a href="#">Freq &amp; BW (see page 78)</a>              | The Freq & BW tab specifies frequency and bandwidth parameters for the DPX Zero Span, DPX Frequency, and DPX Phase views.   |
| <a href="#">BW (see page 122)</a>                        | Sets Resolution Bandwidth.  |
| <a href="#">Traces (Bitmap) Tab (see page 78)</a>        | Allows you to configure the Bitmap Trace.   |
| <a href="#">Traces Tab (see page 81)</a>                 | Allows you to select the number and types of traces to display and their functions.   |
| <a href="#">Traces (Math) Tab (see page 122)</a>         | Allows you to configure the Math Trace.   |
| <a href="#">Horiz &amp; Vert Scale Tab (see page 84)</a> | Sets the vertical and horizontal scale parameters for all the DPX views.  |
| <a href="#">Bitmap Scale Tab (see page 85)</a>           | Sets the DPX Bitmap display parameters.   |
| <a href="#">Amplitude Scale Tab (see page 86)</a>        | The Amplitude Scale tab allows you to change the vertical scale and offset, enable the 3-D Waterfall display, and set the color scheme used for the DPXogram trace. |
| <a href="#">Time &amp; Freq Scale Tab (see page 87)</a>  | The Time and Freq Scale tab allows you to change the vertical and horizontal scale settings, number of points in the trace, and Time resolution.                    |
| <a href="#">Prefs (see page 125)</a>                     | Specifies whether certain display elements are visible.   |
| <a href="#">Density (see page 88)</a>                    | Specifies location and size for the DPX Density measurement box. (Spectrum display only.)   |
| <a href="#">Audio Demod (see page 89)</a>                | Enables and sets parameters for audio demodulation function.  |

## Params Tab - DPX Zero Span, DPX Frequency and DPX Phase Views

The Params tab sets the sweep time for the DPX Zero Span, Frequency and Phase views. Use the Params tab to set the scroll mode settings for the DPX Zero Span, Frequency and Phase views. In Scroll mode, points of the trace are plotted as they occur, as opposed to normal mode where the trace is plotted after an acquisition is completed.



### Params tab settings Zero Span, Frequency and Phase views

| Setting                        | Description  |
|--------------------------------|--|
| Sweep time                     | Sets the time period for the measurement. By default, the Horizontal Scale is equal to Sweep time and the sweep covers 10 divisions. When the graph is zoomed in, the sweep time extends beyond the left and/or right edges of the graph. Range: 100 ns – 2000 s. Default: 1 ms. |
| Trace motion for Sweep ≥ 1 sec | Specifies how the trace is displayed when the sweep time is equal to or greater than 1 second.   |
| Normal                         | Select Normal to scroll the position at which data points are added to the trace. In this mode, a caret (^) moves below the graph to indicate the latest position.   |
| Roll                           | Select Roll to scroll the trace as points are added at the right side of the graph. When Roll is selected, the trace moves to the left as points are added to the trace at the right side of the graph.  |
| None                           | Select None to display the trace without motion.   |

## Freq & BW Tab - DPX Zero Span, DPX Frequency and DPX Phase Views

The Freq & BW tab specifies frequency parameters for some of the DPX display views.

### Freq & BW tab settings for the bitmap trace

| Setting                     | Description   |
|-----------------------------|---|
| Center Frequency            | Sets the frequency at the center of the measurement bandwidth.  |
| Step Size                   | Sets the increment size when changing the Frequency using the knob or mouse wheel. Arrow keys have an increment 10 times this setting.  |
| Auto                        | When Auto is enabled, the step size is adjusted automatically based on Spectrum's span setting.   |
| Measurement BW, no filter   | This setting allows you to override the automatic bandwidth calculation and directly enter a bandwidth value without the time-domain filter. If you enter a value for the measurement bandwidth, be aware that the actual bandwidth of data provided to the measurement will be at least as wide as the value you request and may be as much as two times wider than requested. This override of the selected measurement bandwidth is done so that the instrument uses sufficient bandwidth relative to the chosen symbol rate to ensure good signal quality measurements. |
| RBW (Time-domain BW) filter | RBW (Time-domain BW) is a filter used to process the input signal before the system analyzes the signal. The filter value determines the acquisition bandwidth that the view requires. Range: 1 Hz to 60 MHz.   |
| Actual BW                   | Shows the actual bandwidth being used for the display.  |

## Traces Tab - Bitmap

The Traces Tab allows you to set the display characteristics of displayed traces in the DPX display. The Traces tab for the DPX display has two versions: one for the DPX Bitmap trace (described in this topic) and one [for line traces \(see page 81\)](#).

Traces tab – Bitmap trace

Traces tab – DPXogram trace

### Traces tab settings for the bitmap trace

| Setting                              | Description   |
|--------------------------------------|---|
| Trace                                | Select Bitmap or DPXogram to set the parameters of the DPX Bitmap or DPXogram trace.  |
| Show                                 | Specifies whether or not the selected trace is displayed.   |
| Detection (DPXogram trace only)      | Sets the <a href="#">Detection method (see page 119)</a> used for the trace. Available detection methods are +Peak, -Peak, and Avg (VRMS).  |
| Intensity (standard instrument only) | Use Intensity to control the visibility of events. An increased intensity level allows a single, short event to be seen. This also makes such an event subject to the persistence controls. This allows you to see the effect of the Persistence controls on infrequent events. Range is 1-100%. Resolution: 1. |
| Dot Persistence                      | Allows a dot to remain visible if it is not updated with new data. Choices for this setting are Variable and Infinite (not available for the DPXogram trace).   |
| Variable                             | The Variable dot persistence setting controls how long a point in the display is visible before fading (not available for the DPXogram trace).<br>Range: 100 ms – 60 s. Resolution 0.1.   |
| Infinite                             | The Infinite dot persistence setting prevents a point in the display from fading (not available for the DPXogram trace).  |
| Freeze                               | Halts updates to the selected trace.  |
| Save Trace As                        | Saves the selected trace to a file for later recall and analysis.   |
| Show Recalled Trace                  | Displays a saved trace instead of a live trace (not available for the DPXogram trace).  |

### Dot Persistence

Dot Persistence is the characteristic of the DPX display that determines how long a pixel in the display remains visible.


To set the Persistence:

1. Select **Setup > Settings**.
2. Select the **Traces** tab.
3. Select **Dot Persistence**.
4. Select either **Infinite** or **Variable**.
5. If you select **Variable**, enter a value in the text box. The **Variable** persistence value can be set from 0.05 to 100 seconds.

## Saving Traces

To save a trace for later analysis:

1. Select the **Save Trace As** button. This displays the Save As dialog box.




2. Navigate to the location where you want to save the file.
3. Type a name for the saved trace and click **Save**.

## Show Recalled Trace

You can recall a previously saved trace for comparison to a live trace. First, specify a trace for recall and second, enable **Show Recalled Trace**.

To select a trace for recall:


1. Select the **...** button to display the Open dialog box.



2. In the Open dialog, navigate to the location of the saved trace.
3. Select the desired trace file.
4. Select **OK** to complete your selection.
5. Select the **Show recalled trace** check box.
6. Verify that the trace's **Show** check box is selected (either on this tab or next to the drop-down list located at the top-left corner of the graph).

## Traces Tab

The Traces Tab allows you to set the display characteristics of displayed traces in the DPX display. The Traces tab for the DPX display has two versions. One for non-Bitmap traces (described in this topic) and one for [the DPX Bitmap trace \(see page 78\)](#).



Traces tab

| Setting              | Description  |
|----------------------|--|
| Trace drop-down list | Selects which trace to configure. The available traces are Bitmap, Trace 1, Trace 2, Trace 3, and Math.  |
| Show                 | Specifies whether or not the trace shown in the Trace setting is displayed.  |
| Freeze               | Halts updates to the selected trace.   |
| Detection            | Sets the Detection method used for the trace. Available detection methods are +Peak, -Peak, and Avg (VRMS). Not all detection methods are available in all displays. |
| Function             | Selects the trace processing method. Available settings are: Normal, Average, and Hold.  |
| Save Trace As        | Saves the selected trace to a file for later recall and analysis.  |
| Show Recalled Trace  | Displays a saved trace instead of a live trace.  |
| Count                | Enables user adjustable number of averages. This setting is only present when Function = Average.  |

## Detection

Trace Detection occurs when the trace is being decimated by the measurement. For example, if the maximum number of trace points is 100,000, and the selected analysis region is 200,000 samples, the measurement must decimate the 200,000 resulting trace points by 2 to prevent exceeding the 100,000 trace point limit. Since only one value can be selected for each trace point, an algorithm must be used to select (detect) the appropriate value to use.

The available detection methods are:

- **+Peak** – Each point on the trace is the result of detecting the positive peak value present in the set of IQ samples available to that trace point.
- **-Peak** – Each point on the trace is the result of detecting the negative peak value present in the set of IQ samples available to that trace point.
- **Avg (VRMS)** [Average  $V_{RMS}$ ] – Each point on the trace is the result of determining the RMS Voltage value for all of the IQ samples available to the trace point. When displayed in either linear (Volts, Watts) or Log (dB, dBm), the correct RMS value results. When the averaging function is applied to a trace, the averaging is performed on the linear (Voltage) values, resulting in the correct average for RMS values.

---

**NOTE.** *The Detection setting does not affect the trace until the spectrum length is longer than the Auto setting.*

---

## DPX Trace Processing

The +Peak, -Peak, and Average traces can be processed to display in different ways. The Function setting controls trace processing.


- **Hold** - Displays the value in the trace record for each display point. Each new trace display point is compared to the previous maximum value and the greater value is retained for display and subsequent comparisons. Available for traces using +Peak or -Peak detection.
- **Normal** - Displays the trace record for each display point without additional processing. Available for all detection selections.
- **Average** - Default setting for the Average. Multiple traces are averaged together to generate the displayed trace. There is one vertical value for each underlying frequency data point. Once the specified number of traces have been acquired and averaged to generate the displayed trace, each new trace takes the place of the oldest trace in the calculation. The **Count** setting specifies how many traces are averaged. Available for traces using Average detection.

Trace averaging uses the exponential method. If Count = 10, the newest trace's contribution to the averaged trace is 10%. When Count is not checked, the algorithm assumes the maximum number of traces contributing to the average is  $2^{32}$ .

## Saving Traces

To save a trace for later analysis:

1. Select the **Save Trace As** button. This displays the Save As dialog box.




2. Type a name for the saved trace and click **Save**.

## Recalling Traces

You can recall a previously saved trace for comparison to a live trace. First, specify a trace for recall and second, enable Show Recalled Trace.

To select a trace for recall:


1. Click the ... button to display the Open dialog box.



2. Navigate to the desired file and click **Open**.
3. Check the **Show Recalled Trace** check box.
4. Verify that the trace's **Show** check box is selected (either on this tab or next to the drop-down list located at the top-left corner of the graph).

## Horiz & Vert Scale Tab


The Horiz & Vert Scale tab allows you to change the vertical scale settings used for the Bitmap trace. Changing the scale settings changes how the trace appears on the display but does not change control settings such as Measurement Frequency.



| Setting     | Description   |
|-------------|---|
| Vertical    | Controls the vertical position and scale of the trace display.  |
| Scale       | Changes the vertical scale.   |
| Offset      | Adjusts the Reference Level away from the top of the trace display.   |
| Reset Scale | Sets Scale to its default value and Offset to zero. Disabled when Units (Setup > Analysis > Units) is set to Watts or Volts.                  |
| Autoscale   | Resets the scale of the vertical axis to contain the complete trace. Disabled when Units (Setup > Analysis > Units) is set to Watts or Volts. |
| Horizontal  | Controls the horizontal position and scale of the trace display.  |
| Scale       | Changes the horizontal scale.   |
| Position    | Adjusts the horizontal position of the signal. This does not change the center frequency.   |
| Autoscale   | Resets the scale of the horizontal axis to contain the complete trace.  |

## Bitmap Scale Tab

The Bitmap Scale tab allows you to set the color scheme used for the Bitmap trace. Changing the DPX bitmap Color, Max and Min scale settings changes how the trace appears on the display but does not change control settings such as Measurement Frequency.



| Setting                     | Description  |
|-----------------------------|--|
| DPX Bitmap (Signal Density) | Controls the appearance and scale of the DPX Bitmap trace.   |
| Color                       | Allows you to select the color palette used for the DPX Bitmap trace.  |
| Max                         | Sets the hit density represented by the top of the color scale. "Clipping" occurs for densities greater than this value.<br>Range: 1p% - 100%; Default: 100%.  |
| Min                         | Sets the hit density represented by the bottom of the color range. Range: 0 - 80%; Default: 0.   |
| Curve                       | Adjusts how colors are mapped to the signal density. The mapping can be linear (Curve = 1), or it can be set to concentrate the resolution on the lower level of the range (Curve > 1) or the mapping can be set to show the best resolution on the upper range of density or hit count (Curve < 1). |
| Auto Color                  | Adjusts the Max and Min settings to display the broadest range of colors.  |

## Amplitude Scale Tab

The Amplitude Scale tab allows you to change the vertical and horizontal scale settings, enable the 3-D Waterfall display, and set the color scheme used for the DPXogram trace.

| Setting          | Description   |
|------------------|---|
| Height           | Height controls apply only to the 3-D Waterfall display.  |
| Scale            | Changes the vertical scale for trace Amplitude in the graph (not the vertical scale for Time).  |
| Position         | Specifies the level displayed at the bottom edge of the graph. (Bottom front edge in the 3-D view).   |
| Autoscale        | Adjusts the vertical position bottom for linear units like Amps and Volts. Adjust the vertical position top for log units like dBm. dBm is the default. |
| 3-D Waterfall    | Displays the DPXogram in a 3-D format.  |
| Northeast        | Shifts the perspective of the 3-D graph so that the oldest traces move back and to the right.   |
| Northwest        | Shifts the perspective of the 3-D graph so that the oldest traces move back and to the left.  |
| Reset Scale      | Resets the Height and Color settings to their default values.   |
| Color (DPXogram) |   |
| Color            | Displays a drop-down list that allows you to set the color scheme used for the DPXogram trace.  |
| Max              | Sets the power level represented by the top of the color scale.   |
| Min              | Sets the power level represented by the bottom of the color scale.  |

## Time & Freq Scale Tab


The Time and Freq Scale tab allows you to change the vertical and horizontal scale settings, set the time resolution and number of trace points of the DPXogram display.

| Setting                | Description  |
|------------------------|--|
| Vertical (time)        |  |
| Time/div               | For most Spectrogram applications. Primary time scale control is Time/div. Time scale can be zoomed in or out when acquisitions are stopped.   |
| Position               | The position of the DPXogram record at the bottom of the display. Position cannot be changed while acquisitions are active, and is reset to zero when acquisitions are started again.  |
| Time at position       | Displays the time of the DPXogram line shown at the bottom of the graph. This time is relative to the Time Zero Reference of the current acquisition. If Position is set to a negative value, the Time at position readout will be blanked.                  |
| Reset Scale            | Sets the Time/div and Position settings to their default values.   |
| Trace Points           | Sets the number of trace points computed for each DPXogram line. These are the points used for marker measurements and for results export.   |
| Time resolution        | Specifies the length of time represented by each line in the graph.  |
| Auto                   | Sets the time represented by each line in the graph to be adjusted by the analyzer (checked) or manually (when unchecked). When Auto is enabled, Time Resolution change based on Time/div.   |
| Capacity               | Readout of the total length of time that can be captured. This readout is provided so that you can see how changing the Trace Points and Time resolution affects the amount of data that can be captured. Capacity is represented in the format dd:hh:mm:ss. |
| Horizontal (frequency) |  |
| Scale                  | Sets the displayed frequency range of the graph. This control affects only visual scaling, and does not change the acquisition or analysis parameters.   |
| Position               | Sets the frequency displayed at the center of the graph. Changing this value does not change the Frequency setting.  |
| Autoscale              | Sets the frequency scale to the Spectrogram Span value.  |

## Density Tab

The Density tab specifies the parameters of the measurement box used for measuring average signal density of an area in the bitmap. The measurement box is also used by DPX Density triggering.

To measure the average signal density over a rectangular portion of the DPX bitmap, you can adjust the size and location of the measurement area using these controls, or by dragging the measurement box in the graph. You move the box by dragging the readout. You adjust the size of the box by dragging the corners or edges.




| Setting          | Description  |
|------------------|--|
| Show Measurement | Shows or hides the measurement box in the graph when Triggering is not set to DPX Density. If Triggering is set to DPX Density, the measurement box is always visible. |
| Frequency        | Specifies the frequency at the center of the measurement box. The +/- value specifies the width of the measurement box.  |
| Amplitude        | Specifies the amplitude of the center of the measurement box. The +/- value specifies the height of the measurement box.   |

## Audio Demod Tab

Audio demodulation can help you identify unknown radio signals. You control the audio demodulation function with the Audio Demod tab. You access the Audio Demod tab from the Settings control panel of the DPX display.

**NOTE.** *Audio Demodulation is available only in real-time acquisition mode (not swept acquisition), and the Trigger mode must be set to Free Run (not Triggered).*



| Setting       | Description  |
|---------------|--|
| On/Off        | Enables/disables audio demodulation.   |
| Audio Gain    | Adjusts the volume of the demodulated audio.   |
| AM            | Selects Amplitude Modulation as the demodulation method.   |
| FM            | Selects Frequency Modulation as the demodulation method.   |
| Tune with     | Specifies how the frequency to be demodulated is specified. You can select markers or the frequency control.   |
| Receiver Freq | Readout of the frequency to be demodulated.  |
| Receiver BW   | Adjusts the equivalent receiver bandwidth for the audio demodulation. The range of values is 1 kHz to 500 kHz. |

### AM / FM

Note that these buttons select the demodulation method; they do not specify a frequency band.

### Tune with

The choices for this setting are: one of the markers (MR, M1, M2, M3, M4) or the Frequency control (either the front-panel knob or the Freq control in the application).

To use a marker to specify the frequency to be demodulated:

1. Select **Markers > Define Markers** to display the Define Markers control panel.
2. Select **Add** to turn on the next marker.

---

**NOTE.** The first marker defined will always be **MR**. The MR marker is the reference for delta marker readouts.

---


3. Select the trace to which the marker should be attached from the drop-down list.
4. Click **Add** to add additional markers. Assign them to the appropriate trace.
5. Click the close button to remove the Define Markers control panel.

**Select markers.** Select one of the markers in the Tune with drop-down list and then set the location of the marker to the frequency you want to demodulate. If the marker you select is not enabled, the analyzer will tune to the frequency the marker was last set to.


### Receiver BW

The Receiver BW setting will not change the acquisition bandwidth. Thus, if the acquisition bandwidth is too narrow, audio demodulation will be disabled. The following three cases illustrate the relationship between the Acquisition BW, Receiver Frequency, and Receiver BW.


**Case 1.** In this case, the upper limit of the Receiver BW exceeds the upper limit of the acquisition bandwidth. When this occurs, audio demodulation is disabled and an error message is displayed.



**Case 2.** In this case, the upper and lower limits of the Receiver BW exceed the upper and lower limits of the acquisition bandwidth. When this occurs, audio demodulation will be disabled and an error message is displayed.



**Case 3.** In this case, the upper and lower limits of the Receiver BW fall within the upper and lower limits of the acquisition bandwidth. This is the desired relationship between the three settings, and audio demodulation is enabled.



## Time Overview Display


The Time Overview display shows the entire acquisition record and shows you how the spectrum time and analysis time fit within the acquisition record. This enables you to see how you can adjust the spectrum time and analysis time to measure portions of the data.

You can specify the maximum number of trace points in the Time Overview display. You can set the maximum number of trace points to 1K, 10K, 100K, 1M points or to Never decimate. If the Acquisition Length includes more than 10,000 sample points (and Max trace points is not set to Never decimate), the trace is decimated (using the +Peak method, similar to +Peak detection in a Spectrum display) to 10,000 points. This decimated trace is what is used for marker measurements.

The Time Overview window displays the Spectrum Length and Analysis Length. The Spectrum Length is the period of time within the acquisition record over which the spectrum is calculated. The Analysis Length is the period of time within the acquisition record over which all other measurements (such as Amplitude vs. Time) are made. The Spectrum Length and Analysis Length can be locked together so that the data used to produce the Spectrum display is also used for measurement displays; however, they

do not have to be tied together. They are by default specified separately and used to analyze different parts of the acquisition record.

### Elements of the Time Overview Display



| Item | Element   | Description   |
|------|---|---|
| 1    | Time  | Select the type of time analysis to be performed. You can select Analysis, Spectrum, or Linked.   |
| 2    | Offset  | Sets the offset of the selected analysis time control.  |
| 3    | Length  | Sets the length of the selected analysis time control.  |
| 4    | Position and Scale  | Adjusts the vertical scale and position.  |
| 5    | Scale indicators  | Shows the vertical scale.   |
| 6    | Autoscale button  | Resets the horizontal scale to display the entire acquisition record and the vertical scale to show all trace points.   |
| 7    | Horizontal offset   | Adjusts the horizontal offset.  |
| 8    | Results Timeline  | This fuchsia line indicates the portion of the record actually used for calculating the selected result. For example: if a pulse measurement is selected, it shows the period of the specific pulse. For a constellation display, it shows the points included in the demodulation. |
| 9    | Amplitude vs. Time graph  | The trace represents the entire acquisition record (at full horizontal scale). The graph indicates the Analysis Length or Spectrum Length on the graph with a darker background.  |
| 10   | Horizontal Scale  | Adjusts the span of the graph. By decreasing the scale, the graph essentially becomes a window that you can move over the acquisition record by adjusting the offset.   |
| 11   | Spectrum Length and Offset Indicator (red line, top of graph)                 | This red line indicates the Spectrum Length and Offset. The longer the time, the longer the bar. Adjusting the offset shifts the bar left or right.   |
| 12   | Analysis Length and Offset Indicator (blue line, top of graph below red line) | The blue line indicates the Analysis Length and Offset. The longer the time, the longer the bar. Adjusting the offset shifts the bar left or right.   |

**NOTE.** The area with black background (not gray) in the Amplitude vs. Time Graph highlights the control selected in the Analysis Time Control drop-down list.

[Changing the Time Overview Display Settings \(see page 93\)](#)

## Time Overview Settings

**Menu Bar:** Setup > Settings

**Front Panel :** Settings

**Application Toolbar:** 

The Setup settings for Time Overview are shown in the following table.

| Settings tab                         | Description  |
|--------------------------------------|--|
| <a href="#">Scale (see page 124)</a> | Adjusts the vertical and horizontal scale and offset of the display. |
| <a href="#">Trace (see page 96)</a>  | Allows you to select the types of trace to display and its function. |
| <a href="#">Prefs (see page 125)</a> | Specifies whether or not certain display elements are shown.         |


## Navigator View

The Time Overview Navigator View places the Time Overview display across the top of the application screen. This allows you to adjust the area of interest in the Navigator View and see the results simultaneously in the other displays. For example, in the following image, adjusting the mask in the Navigator View moves the trace and markers in all of the other displays.



### Show Navigator View

To show the Navigator View, select **View > Navigator View**. To remove the Navigator View, deselect Navigator View in the View menu.



## Trace Tab

The Trace Tab allows you to set the display characteristics of displayed traces.

The screenshot shows the 'Trace' tab in a software interface. It contains several controls: a 'Show' checkbox (checked), a 'Freeze' checkbox (unchecked), and a 'Save Trace As...' button. Below these are two rows of settings. The first row has 'Detection:' with a dropdown menu set to '+Peak', and 'Count' with a checked checkbox and a text box containing '10'. The second row has 'Function:' with a dropdown menu set to 'Avg (VRMS)', and another text box with a three-dot menu icon.

| Setting             | Description   |
|---------------------|---|
| Show                | Shows / hides the selected trace.   |
| Detection           | Sets the Detection method used for the trace. Not available for saved traces. Available detection methods are +Peak, -Peak, +/-Peak, Avg (VRMS), and Sample. Not all detection methods are available in all displays. |
| Function            | Selects the trace processing method. Available settings are: Normal, Average, Max Hold, and Min Hold.   |
| (Number of Traces)  | Sets the number of traces averaged to generate the displayed trace. (Present only when Function is set to Average.)   |
| Freeze              | Halts updates to the selected trace.  |
| Save Trace As       | Saves the selected trace to a file for later recall and analysis.   |
| Show Recalled trace | Displays a saved trace instead of a live trace.   |

### Detection

Trace Detection occurs when the trace is being decimated by the measurement. For example, if the maximum number of trace points is 100,000, and the selected analysis region is 200,000 samples, the measurement must decimate the 200,000 resulting trace points by 2 to prevent exceeding the 100,000 trace point limit. Since only one value can be selected for each trace point, an algorithm must be used to select (detect) the appropriate value to use.

The IQ samples in a data acquisition can be detected in a variety of ways. The number of IQ samples available to each trace point varies with both analysis length and trace length. For example, with Spectrum Length set to 'Auto' in the Analysis menu, the instrument analyzes just enough samples to produce one IQ sample pair per trace point. In this case, the detection method chosen has very little effect, as the +Peak, -Peak, Avg (VRMS) and Sample values are all equal. Changing the Spectrum Length causes the available detection methods to differ in value because they have a larger set of samples for the various detection methods to process.

The available detection methods (depending on the display) are:

- **+Peak** – Each point on the trace is the result of detecting the positive peak value present in the set of IQ samples available to that trace point.
- **-Peak** – Each point on the trace is the result of detecting the negative peak value present in the set of IQ samples available to that trace point.
- **+/-Peak** – Selects the highest and lowest values of all the samples contained in two consecutive acquisition intervals.
- **Avg (VRMS)** [Average  $V_{RMS}$ ] – Each point on the trace is the result of determining the RMS Voltage value for all of the IQ samples available to the trace point. When displayed in either linear (Volts, Watts) or Log (dB, dBm), the correct RMS value results. When the averaging function is applied to a trace, the averaging is performed on the linear (Voltage) values, resulting in the correct average for RMS values.
- **Sample** – The result is calculated based on the first sample available in the set of IQ samples for each trace point.


## Trace Processing

Traces can be processed to display in different ways. The Function setting controls trace processing.

- **Normal** - Each new trace is displayed and then replaced by the next trace. Each data point contains a single vertical value.
- **Average** - Multiple traces are averaged together to generate the displayed trace. There is one vertical value for each underlying frequency data point. Once the specified number of traces have been acquired and averaged to generate the displayed trace, each new trace takes the place of the oldest trace in the calculation. The **Number of Traces** setting specifies how many traces averaged.
- **Max Hold** - Displays the maximum value in the trace record for each display point. Each new trace display point is compared to the previous maximum value and the greater value is retained for display and subsequent comparisons.
- **Min Hold** - Displays the minimum value in the trace record for each display point. Each new trace display point is compared to the previous minimum value and the lesser value is retained for display and subsequent comparisons.

## Saving Traces

To save a trace for later analysis:


1. Select the Save icon . This displays the Save As dialog box.
2. Navigate to the desired folder or use the default.
3. Type a name for the saved trace and click **Save**.

## Recalling Traces

You can recall a previously saved trace for analysis or comparison to a live trace.

To select a trace for recall:

1. Select the trace into which the recalled trace will be loaded, from the Trace drop-down list.
2. Check the **Show** check box.
3. Click the ... button to display the Open dialog box.




4. Navigate to the desired file and click **Open**.
5. Check the **Show Recalled Trace** check box.
6. Verify that the trace's **Show** check box is selected (either on this tab or next to the drop-down list located at the top-left corner of the graph).

## Spectrum Display

To display a spectrum:

1. Click the **Displays** button or select **Setup > Displays**.
2. From the Measurements box, select **General Signal Viewing**.
3. Double-click the **Spectrum** icon in the **Available displays** box. This adds the Spectrum icon to the **Selected displays** box (and removes it from the Available displays box). Alternatively, you can click the Spectrum icon and then click the Add button to select Spectrum for display.
4. Click the **OK** button.

## Elements of the Spectrum Display



| Item | Display element | Description  |
|------|-----------------|--|
| 1    | dB/div          | Sets the vertical scale value. The maximum value is 20.00 dB/division.   |
| 2    | RBW             | Sets the resolution bandwidth. Note that when the RBW is set to Auto, its value is italicized.   |
| 3    | VBW             | Enables the VBW (Video Bandwidth) filter. See Setup > Settings > <a href="#">BW Tab</a> (see page 122).  |
| 4    | Autoscale       | Adjusts the Vertical and Horizontal scaling to display the entire trace on screen.   |
| 5    | Position        | Default function is CF - center frequency (equivalent to the Freq setting). If Horizontal scaling has been manually adjusted in Settings > Scale, then Offset will replace CF as the setting at the bottom-left corner of the screen.                                |
| 6    | Span / Scale    | Default function is Span - frequency difference between the left edge of the display and the right edge. If Horizontal scaling has been manually adjusted in Settings > Scale, then Scale will replace Span as the setting at the bottom-right corner of the screen. |
| 7    | Clear           | Restarts multi-trace functions (Avg, Hold).  |
| 8    | Function        | Readout of the Detection and Function selections for the selected trace.   |
| 9    | Show            | Controls whether the selected Trace is visible or not. When trace is Off, the box is not checked.  |
| 10   | Trace           | Selects a trace. Touching here pops up a context menu listing the available traces, whether they are enabled or not. If user selects a trace that is not currently enabled, it will be made enabled.   |

### Touchscreen Actions on Markers in the Graph Area

| Action  | Description   |
|---|---|
| Mouse click within 1/2 div. of a marker                 | Selects the marker and updates the marker display to show the selected marker's values. |
| Touch marker to select and then use knob, or arrow keys | Adjust the setting associated with the Marker.  |
| Touch and drag a marker                                 | Changes marker position to the "drop point".  |

[Changing the Spectrum Display Settings \(see page 100\)](#)

## Spectrum Settings

**Menu Bar:** Setup > Settings

**Front Panel:** Settings

**Application Toolbar:** 

The settings for the Spectrum display are shown in the following table.

| Settings tab                                   | Description   |
|--|---|
| <a href="#">Freq &amp; Span (see page 118)</a> | Sets frequency and span parameters for the Spectrum Analysis display. |
| <a href="#">BW (see page 122)</a>              | Sets Resolution Bandwidth and windowing parameters.                   |
| <a href="#">Traces (see page 119)</a>          | Sets Trace display parameters.  |
| <a href="#">Traces (Math) (see page 122)</a>   | Sets the traces used to create the Math trace.                        |
| <a href="#">Scale Tab (see page 101)</a>       | Sets vertical and horizontal scale and position parameters.           |
| <a href="#">Prefs Tab (see page 125)</a>       | Specifies whether or not certain display elements are shown.          |

## Scale Tab

The Scale tab allows you to change the vertical and horizontal scale settings. Changing the scale settings changes how the trace appears on the display but does not change control settings such as Measurement Frequency. In effect, these controls operate like pan and zoom controls.

The screenshot shows the 'Scale' tab in a software interface. It has two main sections: 'Vertical' and 'Horizontal'. The 'Vertical' section includes a 'Scale' input set to '100.00 dB', a 'Position' input set to '0.00 dBm', and an 'Autoscale' button. The 'Horizontal' section includes a 'Scale' input set to '50.0 MHz', a 'Position' input set to '1.9000 GHz', a 'Log' checkbox (which is unchecked), and an 'Autoscale' button. A 'Reset Scale' button is located between the two sections. At the top of the window, there are tabs for 'Freq & Span', 'BW', 'Traces', 'Scale' (which is selected), and 'Prefs'.

| Setting           | Description   |
|-------------------|---|
| <b>Vertical</b>   |   |
| Scale             | Changes the vertical scale.   |
| Position          | Vertical Position adjusts the top of graph amplitude value. This control allows you to move ("pan") the traces up and down in the graph without changing the Reference Level. |
| Autoscale         | Resets the Offset so that the trace appears below the top of the graph.   |
| <b>Horizontal</b> |   |
| Scale             | Allows you to change the range of frequencies shown in the graph without changing the span or measurement frequency.  |
| Position          | Allows you pan the graph.   |
| Autoscale         | Resets Scale to the Span setting.   |
| Reset Scale       | Resets all settings to their default values.  |
| Log scale         | Resets the display to show the frequency axis in a logarithmic scale.   |

## Spectrogram Display

The Spectrogram is a display with the vertical axis (time) composed of successive spectral displays, each having the amplitude represented by color or intensity. The horizontal axis represents frequency. The most recently acquired spectrum results are added to the bottom of the spectrogram. The addition of a new spectrum can occur at the fastest rate that new spectra can be plotted, or, if you choose, new spectra can be added at a timed rate. The spectrogram view is well-suited to displaying long-term trends of spectral data. The maximum number of lines that can be displayed in a spectrogram is 125,000.

The spectrogram can also be displayed in a 3-D waterfall format. In the 3-D waterfall format, the spectrogram displays the time axis along a simulated Z-axis.

---



**NOTE.** Spectrogram data is shared with the Spectrum display. There is no linkage between DPX and Spectrogram.

---

To display a Spectrogram:

1. Select the **Displays** button or select **Setup > Displays**. This displays the **Select Displays** dialog box.
2. From the **Measurements** box, select **General Signal Viewing**.
3. Double-click the **Spectrogram** icon in the **Available Displays** box. This adds the Spectrogram icon to the Selected displays box.
4. Click the **OK** button. This displays the spectrogram view.
5. To display a 3-D version of the spectrogram, select the 3-D checkbox.

### Elements of the Spectrogram Display



| Item | Display element               | Description   |
|------|-------------------------------|---|
| 1    | Check mark indicator          | The check mark indicator in the upper, left-hand corner of the display shows when the Spectrogram display is the optimized display.<br><br><b>NOTE.</b> When <i>Best for multiple windows</i> is selected in the <i>Amplitude control</i> panel's <i>RF &amp; IF Optimization control</i> , none of the measurement displays shows a checkmark, as there is not a single optimized measurement. |
| 2    | Time/div                      | Sets the length of time represented by each vertical division. Divisions are indicated by tick marks along the left edge of the graph.  |
| 3    | RBW                           | Sets the resolution bandwidth. Note that when the RBW is set to Auto, its value is italicized.  |
| 4    | VBW                           | Enables the VBW (Video Bandwidth) filter. See Setup > Settings > <a href="#">BW Tab</a> (see page 122).   |
| 5    | 3-D checkbox                  | Enables and disables the 3-D view.  |
| 6    | Selected records indicators   | Shows the positions of the start and stop records selected on the Select data records tab. Drag the indicators to select which records will be played by Replay All. Note that these are not visible while acquisitions are running; the instrument must be stopped for the indicators to be visible.   |
| 7    | T                             | Trigger indicator. This icon indicates the trigger point within the current acquisition.  |
| 8    | Pos                           | Position indicates the bottom line visible in graph. Changing this setting scrolls the window up and down through the displayed acquisition records.  |
| 9    | Autoscale                     | Resets Vertical and Horizontal scale and Pos to default values.   |
| 10   | CF                            | Sets the Center Frequency.  |
| 11   | Span                          | Sets the span of the spectrogram display.   |
| 12   | Current data record indicator | A blue line indicates the current data record. When the analysis length is short, the blue line appears as a thin line much like the selected indicator line. When the analysis length is relatively long, the blue line appears more like a blue bar.  |
| 13   | Position scroll bar           | Changes the position of the trace in the window. Changing the position scroll bar is the same as adjusting the Pos setting.   |
| 14   | Selected indicator            | This inverse-colored line indicates the Spectrogram line that will appear in the Spectrum display when the Spectrogram trace is enabled. This line is attached to the selected marker.  |
| 15   | Clear                         | Clears the spectrogram display; however, data records in acquisition history remain in memory and are available for replay. To clear memory, select <b>File &gt; Acquisition Data Info &gt; Delete All Data</b> .   |
| 16   | Marker indicators             | These icons indicate the position of markers in the spectrogram. You can move markers by dragging the desired marker indicator.   |
| 17   | Detection setting             | Displays the selected Detection method (see <a href="#">Settings &gt; Trace</a> (see page 105) tab).  |

| Item | Display element           | Description   |
|------|---------------------------|---|
| 18   | Marker readout            | Marker readout for the selected marker. In the Spectrogram display, the marker readout includes a date and timestamp. The time is displayed in a 24-hour format. The timestamp readout can be shown or hidden independently of the other marker readouts (see <a href="#">Settings &gt; Prefs (see page 125)</a> ). |
| 19   | Time Scale status readout | Three readouts can appear here depending on settings: Time/update, Spectrums/line, and Overlap. See <a href="#">Time Scale Status Readout (see page 104)</a> .  |

### Time Scale Status Readout

Three types of readouts can appear in the display depending on settings:

- Time/update – Displays minutes:seconds when Spectrum Monitor is selected in the **Settings > Time & Freq Scale** tab in the **Vertical (Time)** section.
- Spectrums/line – Displays an integer number when vertical scale is Normal and each line contains the results from one or more frequency transforms (whether zoomed out or not).
- Overlap – Displays the overlap percentage when vertical scale is Normal and each line's transform shares some points with the transforms of lines before and after it (zoomed in). Overlap can only be done when the Analysis Length > 2x RBW frame length.

The Spectrogram can show results from one or multiple acquisitions and it can show one or multiple lines for each acquisition.

[Changing the Spectrogram Display Settings \(see page 104\)](#)

## Spectrogram Settings

**Menu Bar:** Setup > Settings

**Front Panel:** Settings

**Application Toolbar:** 

The Setup settings for the Spectrogram display are visible when Spectrogram is the selected display.

| Settings tab   | Description  |
|--|--|
| <a href="#">Freq &amp; Span (see page 118)</a>       | Sets frequency and span parameters for the Spectrogram display.  |
| <a href="#">BW (see page 122)</a>                    | Sets Resolution Bandwidth and windowing parameters.  |
| <a href="#">Trace (see page 105)</a>                 | Sets Trace display parameters.   |
| <a href="#">Amplitude Scale (see page 106)</a>       | Selects between 2-D and 3-D, sets height scale, position and orientation for 3-D display. Sets color parameters for the spectrogram trace. |
| <a href="#">Time &amp; Freq Scale (see page 106)</a> | Sets the vertical and horizontal scale parameters for the spectrogram trace. The Spectrum Monitor controls are also on this tab.           |
| <a href="#">Prefs (see page 125)</a>                 | Specifies whether or not certain display elements are shown.   |

## Trace Tab

The Spectrogram Trace tab controls let you specify the Detection method, save traces, and recall saved traces for display. The Spectrogram Trace tab does not have all the controls that other Trace tabs contain, however, it does have a unique display element, the Selected Line readout. For details on Detection, Freeze, and saving and recalling traces, see the [Traces Tab \(see page 119\)](#).

The screenshot shows the 'Trace' tab selected in a software interface. The tab bar at the top includes 'Freq & Span', 'BW', 'Trace', 'Amplitude Scale', 'Time & Freq Scale', and 'Prefs'. The 'Trace' tab content area has the following elements:

- Selected Line:** A text field displaying '6/30/2006'.
- Detection:** A dropdown menu currently showing '+Peak'.
- Freeze:** A checkbox that is currently unchecked.
- Save Trace As...:** A button to the right of the Freeze checkbox.
- Show recalled trace:** A checkbox that is currently unchecked.
- Input field:** A text input field at the bottom right, followed by a button with three dots (ellipsis).

### Selected Line

Displays the time at the Selected Line. If Markers are enabled, the selected line is positioned by the selected marker. If no markers are enabled, the selected line is the first line in the current analysis period.

## Amplitude Scale Tab

The Amplitude Scale tab allows you to change the vertical and horizontal scale settings, enable the 3-D Waterfall display, and set the color scheme used for the spectrogram trace.

| Setting           | Description   |
|-------------------|---|
| Height (3-D only) |   |
| Scale             | Changes the vertical scale for trace Amplitude in the graph (not the vertical scale for Time).                  |
| Position          | Specifies the level displayed at the bottom edge of the graph. (Bottom front edge in the 3-D view).             |
| Autoscale         | Adjusts the vertical position and scale of the trace lines to bring them into the visible portion of the graph. |
| 3-D Waterfall     | Displays the spectrogram in a 3-D format.   |
| Northeast         | Shifts the perspective of the 3-D graph so that the oldest traces move back and to the right.                   |
| Northwest         | Shifts the perspective of the 3-D graph so that the oldest traces move back and to the left.                    |
| Reset Scale       | Resets the Height and Color settings to their default values.   |
| Color (Power)     |   |
| Color             | Displays a drop-down list that allows you to set the color scheme used for the spectrogram trace.               |
| Max               | Sets the maximum power level represented by the top of the color scale.   |
| Min               | Sets the minimum power level represented by the bottom of the color scale.                                      |

## Time & Freq Scale Tab

The Time and Freq Scale tab allows you to change the vertical and horizontal scale settings, enable the 3-D Waterfall display, and set the color scheme used for the spectrogram trace.

The screenshot shows the 'Time & Freq Scale' tab with the following settings:

- Vertical (Time):**
  - Mode: ☒ Normal
  - Time/div: 72.0 us
  - Overlap: 94 %
  - Visible elapsed time: 396.7 us
  - Position: 12.3 div
  - Time at position: 1.261 ms
  - Time/update: 0 m : 1 s
  - Buttons: Reset Scale, Autoscale
- Horizontal (Frequency):**
  - Scale: 10.00 MHz
  - Position: 2.44530 GHz
  - Button: Autoscale

| Setting                       | Description  |
|-------------------------------|--|
| <b>Vertical (Time)</b>        |  |
| Normal                        | For most Spectrogram applications. Primary time scale control is Time/div. Time scale can be zoomed in or out.   |
| Spectrum Monitor              | For long-term signal monitoring applications. In spectrum monitor mode, each line in the spectrogram represents the period of time specified by the Time/update parameter. Time scale can be zoomed out, but not zoomed in (no overlap).   |
| Time/update                   | Sets the time, in minutes and seconds, represented by each line of the spectrogram. Only available in Spectrum Monitor.  |
| Reset Scale                   | Resets the Time/div and Pos settings to their default values.  |
| Autoscale                     | Scales the vertical (time) axis to compress all existing trace lines into the visible area of the graph. Resets the Position value to zero, placing the most recent spectrogram line at the bottom of the spectrogram display. Only Position is affected by Autoscale when Spectrum Monitor is selected. |
| Time/div                      | Sets the time displayed per division.  |
| Visible elapsed time          | Displays the length of time visible in the display. This does not represent the total time available to view.  |
| Position                      | Adjusts vertical position of the trace within the graph area. Setting represents the offset, in divisions, between the bottom of the graph and the bottom (most recent) line in the results trace.   |
| Time at position              | Displays the time of the spectrogram line shown at the bottom of the graph. This time is relative to the Time Zero Reference of the current acquisition.   |
| <b>Horizontal (Frequency)</b> |  |
| Scale                         | Sets the frequency range of the graph without changing the Span value.   |
| Position                      | Sets the frequency displayed at the center of the graph. Changing this value does not change the Freq setting.   |
| Autoscale                     | Sets the frequency scale to the Spectrogram Span value.  |

## Spectrum Monitor

Spectrum Monitor performs long term monitoring. The monitor mode compresses time into each line of the spectrogram, which enables you to monitor long periods of time (from 1 second per line up to 600 minutes per line). With extended memory (Option 53) installed in the instrument, you can capture up to 125,000 lines. With standard memory installed in the instrument, the maximum number of lines you can capture is 31,250.

During each line's collection period, spectrum transforms are computed for each acquisition taken by the instrument. As each transform completes, it is incorporated into the current spectrogram line. How each

line of the spectrogram is created in spectrum monitor mode depends on the detection setting (Settings > Trace). For example, if Detection is set to +Peak, each spectrogram line is effectively a peak hold display of all the spectral data captured since the prior line.

## Amplitude Vs Time Display


The Amplitude vs. Time display plots the signal amplitude against time. The amplitude appears on the vertical axis while time is plotted along the horizontal axis.

Note that the trace(s) in the Amplitude vs. Time display can be set to a maximum of 100,000 points (however, the actual number of trace points can extend up to 1,000,000 points if Max trace points is set to Never Decimate). If the Analysis Length includes more than the selected Max trace points value, the trace is decimated (using the method specified with the Detection control) to be equal to or less than the Max trace points setting (except when Max trace points is set to Never Decimate). This decimated (or undecimated) trace is what is used for marker measurements and for results export. You can set the Max trace points on the **Settings > Prefs** tab.

To show Amplitude vs. Time display:

1. Press the **Displays** button or select **Setup > Displays**.
2. In the **Select Displays** dialog, select **General Signal Viewing** in the **Measurements** box.
3. In the **Available displays** box, double-click the **Amplitude vs. Time** icon or select the icon and click **Add**. The Amplitude vs. Time icon will appear in the **Selected displays** box and will no longer appear under Available displays.
4. Click **OK**.

### Elements of the Display



| Item | Display element           | Description   |
|------|---------------------------|---|
| 1    | Vertical scale adjustment | Adjusts the vertical scaling.   |
| 2    | Span                      | Adjust the bandwidth of the data to be analyzed. (Not the period of time shown in the display.) |
| 3    | Autoscale                 | Adjusts the vertical and horizontal settings to provide the best display.                       |
| 4    | Offset                    | Adjust the horizontal offset.   |
| 5    | Max and Min readouts      | Displays the maximum and minimum amplitudes, as well as when those values occur.                |
| 6    | Scale                     | Sets the time spanned by the graph.   |
| 7    | Clear button              | Restarts multi-trace functions (Avg, Hold).   |
| 8    | Trace function            | Displays the current trace function setting (Settings > Trace tab > Function).                  |

Reference. [Changing Amplitude vs Time Display Settings \(see page 109\)](#)

## Amplitude Vs Time Settings

**Menu Bar:** Setup > Settings

**Front Panel:** Settings

**Application Toolbar:** 

The settings for the Amplitude vs. Time display are shown in the following table.

| Settings tab                                 | Description  |
|--|--|
| <a href="#">Freq &amp; BW (see page 110)</a> | Sets the Bandwidth Method used for setting the measurement bandwidth.  |
| <a href="#">Traces (see page 119)</a>        | Allows you to select the type of trace to display and their functions. |
| <a href="#">Traces (Math) (see page 122)</a> | Sets the traces used to create the Math trace.                         |
| <a href="#">Scale (see page 124)</a>         | Sets the vertical and horizontal scale parameters.                     |
| <a href="#">Prefs (see page 125)</a>         | Specifies whether certain display elements are visible.                |

## Freq & BW Tab

The Freq & BW (Bandwidth) tab allows you to specify the bandwidth parameters used for setting measurement bandwidth. This determines what Acq BW the measurement will request.

| Settings                     | Description   |
|------------------------------|---|
| Time-domain Bandwidth filter | Time-domain BW is a filter used to process the trigger input signal before the trigger system analyzes the signal. The frequency edge trigger point must lie within the range of time domain bandwidth. This makes the range of the frequency edge trigger = Center Frequency $\pm$ (0.5 $\times$ Time Domain Bandwidth)  |
| Measurement BW, no filter    | This control limits the bandwidth of measurements. You use the measurement bandwidth setting to improve the signal-to-noise ratio of the measurement, resulting in lower measurement uncertainty. However, if the measurement bandwidth is set too low, resulting in fewer samples per second, it can reduce the number of points within the measurement length below the 256 stable samples required, thus causing the analysis to fail. |

## Frequency Vs Time Display

The Frequency vs. Time Display shows how the signal frequency varies with time.


Note that the trace(s) in the Frequency vs. Time display can be set to a maximum of 100,000 points (however, the actual number of trace points can extend up to 1,000,000 points if Max trace points is set to Never Decimate). If the Analysis Length includes more than the selected Max trace points value, the trace is decimated (using the method specified with the Detection control) to be equal to or less than the Max trace points setting (except when Max trace points is set to Never Decimate). This decimated (or undecimated) trace is what is used for marker measurements and for results export. You can set the Max trace points on the **Settings > Prefs** tab.

To display the Frequency vs. Time Display:

1. Select the **Displays** button or **Setup > Displays**.
2. In the **Select Displays** dialog, select **General Signal Viewing** in the **Measurements** box.

3. In the **Available displays** box, double-click the **Frequency vs. Time** icon or select the icon and click **Add**. The Frequency vs. Time icon will appear in the **Selected displays** box and will no longer appear under Available displays.
4. Click **OK** to display the Freq vs. Time display.

### Elements of the Display



| Item | Display element                        | Description  |
|------|--|--|
| 1    | Top of graph adjustment                | Use the knob to adjust the frequency range displayed on the vertical axis.                 |
| 2    | Offset adjustment                      | Adjusts the frequency shown at the center of the display.                                  |
| 3    | Autoscale button                       | Adjusts the offset and range for both vertical and horizontal to provide the best display. |
| 4    | Maximum and Minimum frequency readouts | Displays the maximum and minimum values, as well as when those values occur.               |
| 5    | Horizontal Scale                       | Sets the time spanned by the graph.  |
| 6    | Clear button                           | Restarts Average trace.  |
| 7    | Trace function                         | Displays the current trace function setting (Settings > Trace > Function).                 |

[Changing Frequency vs Time Display Settings \(see page 111\)](#)

## Frequency Vs Time Settings

**Menu Bar:** Setup > Settings

**Front Panel: Settings****Application Toolbar:** 

The Setup settings for Frequency vs. Time are shown in the following table.

| Settings tab                                 | Description   |
|--|---|
| <a href="#">Freq &amp; BW (see page 116)</a> | Sets the frequency and bandwidth parameters.                  |
| <a href="#">Trace (see page 119)</a>         | Sets the trace display parameters.                            |
| <a href="#">Scale (see page 124)</a>         | Sets the Vertical and Horizontal scale and offset parameters. |
| <a href="#">Prefs (see page 125)</a>         | Specifies whether certain display elements are visible.       |

## Phase Vs Time Display


The Phase vs. Time display plots the signal phase against time. The phase appears on the vertical axis while time is plotted along the horizontal axis.

Note that the trace(s) in the Phase vs. Time display can be set to a maximum of 100,000 points (however, the actual number of trace points can extend up to 1,000,000 points if Max trace points is set to Never Decimate). If the Analysis Length includes more than the selected Max trace points value, the trace is decimated (using the method specified with the Detection control) to be equal to or less than the Max trace points setting (except when Max trace points is set to Never Decimate). This decimated (or undecimated) trace is what is used for marker measurements and for results export. You can set the Max trace points on the **Settings > Prefs** tab.

To display Phase vs. Time:

1. Press the **Displays** button or select **Setup > Displays**.
2. In the **Select Displays** dialog, select **General Signal Viewing** in the **Measurements** box.
3. In the **Available displays** box, double-click the **Phase vs. Time** icon or select the icon and click **Add**. The Phase vs. Time icon will appear in the **Selected displays** box and will no longer appear under Available displays.
4. Click **OK** to show the Phase vs. Time display.

## Elements of the Display



| Item | Display element            | Description   |
|------|----------------------------|---|
| 1    | Top of graph adjustment    | Adjusts the vertical scale. Use the knob to adjust the value of the top of the graph.                                     |
| 2    | Vertical offset adjustment | Adjusts the phase error shown at the vertical center of the display.  |
| 3    | Autoscale button           | Adjusts the vertical and horizontal settings so that the entire trace fits in the view.                                   |
| 4    | Horizontal Offset          | Adjusts the horizontal position of the trace.   |
| 5    | Max and Min readouts       | Displays the maximum and minimum value of the phase error within the analysis times and the times at which they occurred. |
| 6    | Horizontal Scale           | Sets the time spanned by the graph.   |
| 7    | Clear button               | Restarts multi-trace functions (Avg, Hold).   |
| 8    | Trace function             | Shows the trace function as set on the Settings > Trace tab.  |

[Changing the Phase vs Time Display Settings \(see page 113\)](#)

## Phase Vs Time Settings

**Menu Bar:** Setup > Settings

**Front Panel:** Settings

**Application Toolbar:** 

The settings for the Phase vs. Time display are shown in the following table.

| Settings tab                                 | Description   |
|--|---|
| <a href="#">Freq &amp; BW (see page 116)</a> | Sets the frequency and bandwidth parameters.                  |
| <a href="#">Trace (see page 119)</a>         | Sets the trace display parameters.                            |
| <a href="#">Scale (see page 124)</a>         | Sets the Vertical and Horizontal scale and offset parameters. |
| <a href="#">Prefs (see page 125)</a>         | Specifies whether certain display elements are visible.       |

## RF I & Q vs Time Display


This is a plot of the baseband In-Phase (I) and Quadrature (Q) components of a modulated carrier. The plot is in the time domain, with I and/or Q values the Y-axis.

Note that the trace(s) in the RF I & Q vs. Time display can be set to a maximum of 100,000 points (however, the actual number of trace points can extend up to 1,000,000 points if Max trace points is set to Never Decimate). If the Analysis Length includes more than the selected Max trace points value, the trace is decimated (using the method specified with the Detection control) to be equal to or less than the Max trace points setting (except when Max trace points is set to Never Decimate). This decimated (or undecimated) trace is what is used for marker measurements and for results export. You can set the Max trace points on the **Settings > Prefs** tab.

To display an RF I & Q vs. Time display:

1. Select the **Displays** button or select **Setup > Displays**. This shows the **Select Displays** dialog box.
2. From the **Measurements** box, select **General Signal Viewing**.
3. Double-click the **RF I&Q vs. Time** icon in the **Available Displays** box. This adds the RF I & Q vs. Time icon to the **Selected displays** box.
4. Click the **OK** button.

## Elements of the Display



| Item | Display element                    | Description   |
|------|------------------------------------|---|
| 1    | Top of Graph adjustment            | Use the knob to adjust the vertical scaling.  |
| 2    | Vertical offset adjustment         | Adjusts the level shown at the center of the display.   |
| 3    | Autoscale button                   | Adjusts the offset and scale for both vertical and horizontal to provide the best display.  |
| 4    | Maximum and Minimum level readouts | Displays the maximum and minimum values, within the Analysis Time, as well as the times at which they occurred.                             |
| 5    | Scale                              | Sets the time spanned by the graph.   |
| 6    | Clear button                       | Restarts multi-trace functions (Avg, Hold).   |
| 7    | Trace function                     | Displays the current trace function setting (Settings > Trace > Function). If the traces are averaged, the number of averages is displayed. |
| 8    | Trace Control                      | Selects which trace is displayed (using the drop-down list) and which trace is active (click on the trace name to display a menu).          |

[Changing the RF I & Q vs Time Display Settings \(see page 115\)](#)

## RF I & Q vs Time Settings

**Menu Bar:** Setup > Settings

**Front Panel:** Settings

**Application Toolbar:** 

The Setup settings for RF I&Q vs. Time are shown in the following table.

| Settings tab                                 | Description   |
|--|---|
| <a href="#">Freq &amp; BW (see page 116)</a> | Sets the frequency and bandwidth parameters.                  |
| <a href="#">Trace (see page 119)</a>         | Sets the trace display parameters.                            |
| <a href="#">Scale (see page 124)</a>         | Sets the Vertical and Horizontal scale and offset parameters. |
| <a href="#">Prefs (see page 125)</a>         | Specifies whether certain display elements are visible.       |

## General Signal Viewing Shared Measurement Settings

The control panel tabs in this section are identical or very similar for each of the displays in the General Signal Viewing folder (Setup > Displays). Some tabs are shared by all the displays, some tabs are shared by only a couple of displays.

For some tabs, the control values are shared across all the General Signal Viewing displays. For other control values, each display has unique values for the controls. Details are provided for the specific tabs.

### Common controls for general signal viewing displays

| Settings tab                                   | Description   |
|--|---|
| <a href="#">Freq &amp; Span (see page 118)</a> | Sets the frequency and span parameters.                               |
| <a href="#">Trace (see page 119)</a>           | Sets the trace display parameters.                                    |
| <a href="#">Traces – Math (see page 122)</a>   | Sets the traces used to create the Math trace.                        |
| <a href="#">BW (see page 122)</a>              | Sets the Bandwidth Method used for setting the measurement bandwidth. |
| <a href="#">Scale (see page 124)</a>           | Sets the Vertical and Horizontal scale and offset parameters.         |
| <a href="#">Prefs (see page 125)</a>           | Specifies whether certain display elements are visible.               |

## Freq & BW Tab — Freq vs Time, Phase vs Time, RF I & Q vs Time Display

The Freq & BW tab provides access to settings that control frequency settings for the Freq vs Time, Phase vs Time, RF I & Q vs Time display.

The screenshot shows the 'Freq & BW' tab selected in a software interface. The tab bar at the top includes 'Freq & BW', 'Traces', 'Scale', and 'Prefs'. The main area contains the following controls:

- Measurement Freq:** A text box containing '1.000 GHz'.
- Measurement BW, no filter:** A text box containing '500 MHz'.
- Link to Span:** A checkbox that is currently unchecked.
- Set to max BW:** A button located below the bandwidth text box.

| Setting          | Description   |
|------------------|---|
| Measurement Freq | The frequency at the which measurements are made. This value is the same as the setting for Frequency in the Application bar.   |
| Measurement BW   | This control limits the bandwidth of measurements. You use the measurement bandwidth setting to improve the signal-to-noise ratio of the measurement, resulting in lower measurement uncertainty. However, if the measurement bandwidth is set too low, resulting in fewer samples per second, it can reduce the number of points within the measurement length below the 256 stable samples required, thus causing the analysis to fail. |
| Link to Span     | When enabled, the measurement bandwidth of the RF I & Q display is determined by the span of the analyzer. When unchecked, the measurement bandwidth is specified by the user, and no additional filter is applied.   |
| Set to max BW    | Sets the measurement bandwidth to the maximum acquisition bandwidth of the instrument.  |

## Freq & Span Tab

The Freq & Span tab provides access to settings that control frequency settings for the trace display. The control values set in this tab are shared by all the General Signal Viewing displays.

The screenshot shows the 'Freq & Span' tab selected. It contains the following controls:

- Center:** 1.9000 GHz
- Span:** 85.0 MHz
- Start:** 1.8575 GHz
- Stop:** 1.9425 GHz
- Step Size:** 5.0000 MHz
- Dwell time:** 50.0 ms (swept only)
- Auto checkboxes:** Checked for Step Size and Dwell time.
- Max Span button:** Located next to the Span field.

| Setting                       | Description  |
|-------------------------------|--|
| Center                        | The frequency at the center of the selected Span.  |
| Start                         | The lowest frequency in the span.  |
| Stop                          | The highest frequency in the span.   |
| Step Size                     | Sets the increment/decrement size for Center, Start and Stop values.   |
| Span                          | The difference between the start and stop frequencies. This is the measurement bandwidth for the general signal viewing displays.  |
| Dwell time (DPX display only) | The amount of time the DPX sweep remains in a frequency segment collecting data and updating the bitmap and traces before moving on to the next higher frequency segment. The minimum Dwell Time setting is 50 ms, the normal update rate for DPX computations. Maximum Dwell Time per frequency segment is 100 s. The number of frequency segments and their start/stop frequencies are internally determined based on span, acquisition bandwidth, RBW and other parameters. |
| Max Span                      | Sets the Span to the maximum value.  |

### Center, Start, Stop, and Span Frequencies Are Correlated

Changing the values for Center frequency, Start frequency, Stop frequency or Span will change the values for the other settings, depending on which setting you change. For example, if you change the Center frequency, the Start and Stop frequencies will be adjusted automatically to maintain the same Span.

Note however that if the Start and Stop frequencies are changed so that they are closer than the minimum span setting, the Start and Stop frequencies will be adjusted to maintain the minimum Span setting.

| Setting Changed Manually | Settings Changed Automatically As a Result | Setting Not Automatically Changed |
|--------------------------|--|-----------------------------------|
| Start                    | Center, Span                               | Stop                              |
| Stop                     | Center, Span                               | Start                             |
| Center                   | Start, Stop                                | Span                              |
| Span                     | Start, Stop                                | Center                            |

## Traces Tab

The Traces Tab allows you to set the display characteristics of displayed traces.

The screenshot shows the 'Traces' tab in a software interface. It contains several controls for 'Trace 1': a 'Show' checkbox (checked), a 'Freeze' checkbox (unchecked), a 'Save Trace As...' button, a 'Detection' dropdown menu set to '+Peak', and a 'Function' dropdown menu set to 'Normal'. There is also an unchecked 'Show recalled trace' checkbox and a small button with three dots.

| Setting             | Description   |
|---------------------|---|
| Trace               | Selects a trace. (This setting is not present for every display.)   |
| Show                | Shows / hides the selected trace.   |
| Detection           | Sets the Detection method used for the trace. Not available for saved traces. Available detection methods are +Peak, -Peak, +/-Peak, Avg (VRMS), Sample, CISPR Avg, CISPR Pk, and CISPR QPk. Not all detection methods are available in all displays. |
| Function            | Selects the trace processing method. Available settings are: Normal, Average, Max Hold, and Min Hold.   |
| Count               | Sets the number of traces averaged to generate the displayed trace. (Present only when Function is set to Average, Min Hold, or Max Hold.)  |
| Freeze              | Halts updates to the selected trace.  |
| Save Trace As       | Saves the selected trace to a file for later recall and analysis.   |
| Show Recalled trace | Displays a saved trace instead of a live trace.   |

### Trace

Available traces for Spectrum are: Trace 1, Trace 2, Trace 3, Math, and Spectrogram. Other displays support fewer traces. Traces 1-3 are based on the input signal and enable you to display the input signal using different processing. For example, you could display Trace 1 with Function set to Normal, Trace 2 with Function set to Max Hold and Trace 3 with Function set to Min Hold.

The Math trace is the result of subtracting one trace from another.

The Spectrogram trace applies only to the Spectrum display and is available only if the Spectrogram display is shown. The Spectrogram trace shows the trace selected in the Spectrogram as a spectrum trace.

### Detection

Trace Detection is used to reduce the results of a measurement to the desired number of trace points. For example, if the maximum number of trace points is 100,000, and a measurement over the selected analysis length yields 200,000 points, the measurement must decimate these 200,000 trace points by 2 to prevent exceeding the 100,000 trace point limit. Since only one value can be represented for each trace point, an algorithm must be used to select (detect) the appropriate value to use.

The results array from an analysis can be detected (or “decimated”) in a variety of ways. The number of results points produced for each trace point varies with both analysis length and trace length. For example, the frequency transform used for the Spectrum display produces just one output value for each desired trace point. In this case, the detection method chosen has no effect, as no decimation is required. Increasing the Analysis Length (or for the Spectrum display, the Spectrum Length), causes the available detection method's output traces to differ from each other because they have a larger set of samples for the various detection methods to process.

The available detection methods (depending on the display) are:

- **+Peak** – The highest value is selected from the results to be compressed into a trace point.
- **-Peak** – The lowest value is selected from the results to be compressed into a trace point.
- **+/-Peak** – Both the highest and lowest values are selected from the results to be compressed into a trace point.
- **Avg (VRMS)** [Average  $V_{RMS}$ ] – Each point on the trace is the result of determining the RMS Voltage value for all of the results values it includes. When displayed in either linear (Volts, Watts) or Log (dB, dBm), the correct RMS value results.
- **Avg (of logs)** – The detector is used to emulate legacy spectrum analyzer results and for the specification of displayed average noise level. In older swept analyzers, a voltage envelope detector is used in the process of measuring signal level, and the result is then converted to Watts and then to dBm. Averaging is then applied to the resultant traces.

For CW signals, this method results in an accurate power measurement. However, with random noise and digitally modulated carriers, errors result from this 'average of logs' method. For random noise, the average of logs methods results in power levels -2.51 dB lower than that measured with a power meter, or with a signal analyzer that measures the rms value of a signal, and performs averaging on the calculated power in Watts and not dBm or other log-power units.

This detector should be used when following a measurement procedure that specifies it, or when checking the Displayed Averaged Noise Level (DANL) of the instrument. The 'average of logs' detection and trace function is used for DANL specification to provide similar results to other spectrum/signal analyzers for comparison purposes. Use of the Average of Logs method of measurement is not recommended for digitally modulated carriers, as power measurement errors will occur.

---

**NOTE.** *The Detection setting does not affect the trace until the spectrum length is longer than the Auto setting.*

---

- **Sample** – The first value is selected from the set of results to be compressed into a trace point.
- **CISPR Peak** – The trace value is calculated by the methods described for peak detectors in the CISPR documents.

## Trace Processing

Traces can be processed to display in different ways. The Function setting controls trace processing.

- **Normal** - Each new trace is displayed and then replaced by the next trace. Each data point contains a single vertical value.
- **Average** - Multiple traces are averaged together to generate the displayed trace, which will contain just one vertical value for each underlying frequency data point. Once the specified number of traces have been acquired and averaged to generate the displayed trace, additional traces contribute to the running average, except in Single Sequence run mode. In the case of Single Sequence, the instrument stops running after the specified number of traces have been averaged together. The **Number of Traces** setting specifies how many traces are averaged. The averaging is performed on the linear (Voltage) values, resulting in the correct RMS average).
- **Max Hold** - Displays the maximum value in the trace record for each display point. Each new trace's display point is compared to the previous maximum value and the greater value is retained for display and subsequent comparisons.
- **Min Hold** - Displays the minimum value in the trace record for each display point. Each new trace's display point is compared to the previous minimum value and the lesser value is retained for display and subsequent comparisons.

### Saving Traces

To save a trace for later analysis:


1. Select the **Save Trace As** button. This displays the Save As dialog box.
2. Navigate to the desired folder or use the default.
3. Type a name for the saved trace and click **Save**.

### Recalling Traces

You can recall a previously saved trace for analysis or comparison to a live trace.

To select a trace for recall:

1. Select the trace into which the recalled trace will be loaded, from the Trace drop-down list.
2. Check the **Show** check box.
3. Click the ... button to display the Open dialog box.




4. Navigate to the desired file and click **Open**.

5. Check the **Show Recalled Trace** check box.
6. Verify that the trace's **Show** check box is selected (either on this tab or next to the drop-down list located at the top-left corner of the graph).

## Traces Tab - Math Trace

This tab is not a distinct tab, it is just how the Traces tab appears when Math is selected in the Traces drop-down list.




Trace 4 is a mathematically-derived trace defined as Trace A minus Trace B. You can select Trace 1, 2, or 3 to serve as either Trace A or Trace B.

| Setting           | Description   |
|-------------------|---|
| Trace             | When set to Trace 4 (Math), this tab is displayed.                |
| Show              | Shows / hides the selected trace.                                 |
| Freeze            | Halts updates to the selected trace.                              |
| Save Trace As     | Saves the selected trace to a file for later recall and analysis. |
| Trace minus Trace | Selects which traces serve as Trace A and Trace B.                |

## BW Tab

The BW (bandwidth) tab allows you to change Resolution Bandwidth and Video Bandwidth settings, and set the windowing method used by the transform process by selecting a filter shape (not present for all displays).



| Setting        | Description  |
|----------------|--|
| RBW            | Sets the Resolution Bandwidth value to be used in the spectrum analysis view. The value is italicized when Auto is selected.   |
| Auto           | When Auto is checked, the RBW is calculated as a percentage of the Span. Kaiser is selected as the windowing method. When Auto is unchecked, the RBW is set by the user. Selecting any Window other than Kaiser changes the RBW setting to manual. |
| Span/RBW ratio | If Auto is checked, this value is used to calculate the RBW. If Auto is unchecked, this setting is not selectable.   |
| Filter Shape   | Specifies the windowing method used for the transform (when Auto is unchecked). (Spectrum and Spectrogram displays only.)  |
| VBW            | Adjusts the VBW (Video Bandwidth) value. (Spectrum and Spectrogram displays only.)   |

### Filter Shape Settings

In the analyzer, computationally efficient discrete Fourier transform algorithms such as FFT (Fast Fourier Transform) or CZT (Chirp-Z Transform) are generally employed to transform time-domain signals into frequency-domain spectra. There is an assumption inherent in the mathematics of Discrete Fourier Transforms that the data to be processed is a single period of a periodically repeating signal. The discontinuities between successive frames will generally occur when the periodic extension is made to the signal. These artificial discontinuities generate spurious responses not present in the original signal, which can make it impossible to detect small signals in the presence of nearby large ones. This phenomenon is called spectral leakage.

Applying a filter, such as Kaiser, to the signal to be transformed is an effective method to combat the spectral leakage problem. Generally the filter has a bell shape. Multiplying the transform frame by the filter function eliminates or reduces the discontinuities at the ends of the frame, however, at the expense of increased RBW.

### Filter Shape Characteristics

The choice of filter shape depends on its frequency response characteristics such as side-lobe level, equivalent noise bandwidth and maximum amplitude error. Use the following guidelines to select the best filter shape.

| Filter Shape       | Characteristics   |
|--------------------|---|
| Kaiser (RBW)       | Best side-lobe level, shape factor closest to the traditional Gaussian RBW.   |
| -6dB RBW (MIL)     | These filters are specified for bandwidth at their -6 dB point, as required by military EMI regulations.  |
| CISPR              | These filters comply with the requirements specified in the P-CISPR 16 -1-1 document for EMI measurements.  |
| Blackman-Harris 4B | Good side-lobe level.   |
| Uniform (None)     | Best frequency resolution, poor side-lobe level and amplitude accuracy.   |
| Flat-Top           | Best amplitude accuracy, best representation of brief events captured near the beginning or end of the time-domain data frame, poor frequency resolution. |
| Hanning            | Good frequency resolution, high side-lobe roll-off rate.  |

## VBW

The VBW setting enables/disables the Video Bandwidth filter. VBW is used in traditional swept analyzers to reduce the effect of noise on the displayed signal. The VBW algorithm in the analyzer emulates the VBW filters of traditional swept analyzers.

The maximum VBW value is the current RBW setting. The minimum VBW value is 1/10,000 of the RBW setting. VBW is disabled when the Filter shape is set to CISPR.

---

**NOTE.** If you are following a procedure that says to "set VBW to three times the RBW value or greater", it means that the test should be conducted with no VBW effects. In the analyzer, this condition is met by disabling the VBW function.

---

## Scale Tab

The Scale tab allows you to change the vertical and horizontal scale settings. Changing the scale settings changes how the trace appears on the display but does not change control settings such as Measurement Frequency. In effect, these controls operate like pan and zoom controls.

The Scale tab values are unique to each display. Also, note that each display uses horizontal and vertical units that are appropriate for the display. For example, for the Spectrum display uses power (dBm) units and frequency (Hz) units; the Amplitude vs. Time display uses power (dBm) and time (seconds) units; and the Phase vs. Time display uses phase (degrees) and time (seconds) units.


| Setting     | Description   |
|-------------|---|
| Vertical    | Controls the vertical position and scale of the trace display.          |
| Scale       | Changes the vertical scale.   |
| Offset      | Vertical Offset adjusts the reference level away from top of the graph. |
| Autoscale   | Resets the scale of the vertical axis to contain the complete trace.    |
| Horizontal  | Controls the span of the trace display and position of the trace.       |
| Zoom Start  | Sets the starting frequency for ...                                     |
| Zoom Stop   | Sets the stop frequency for ...   |
| Log         | Resets the scale of the horizontal axis to contain the complete trace.  |
| Reset Scale | Resets all settings to their default values.                            |

## Prefs Tab


The Prefs tab enables you to change parameters of the measurement display. The parameters available on the Prefs tab vary depending on the selected display, but include such items as enabling/disabling Marker Readout, switching the Graticule display on/off, and Marker Noise mode.

Each of the General Signal Viewing displays maintains its own separate values for the controls that appear on the Prefs tab. Some parameters appear with most displays while others appear with only one display.

For example, in the following image, the Show Marker readout in graph check box appears in the Prefs tab for every display. However, the Show Power Trigger level check box only appears on the Amplitude vs Time Prefs tab.



The following image shows the Prefs tab for the Time Overview display.



The following table explains the controls that can appear on the Prefs tab.

| Setting   | Description   |
|---|---|
| Show:   | Selects the horizontal settings that appear below the graph area. You can choose Start, Stop or Center, Span.   |
| Trace points                                      | Sets the number of trace points used for marker measurements and for results export.  |
| Max trace points<br>(Time Overview Display only)  | Sets the maximum number of trace points used for marker measurements and for results export.  |
| Show trace legend                                 | Enables display of a legend in the measurements area that shows the Detection method and Function setting for displayed traces. The color of the legend text matches the color of the associated trace. |
| Show graticule                                    | Select to display or hide the graticule.  |
| Navigator View<br>(Time Overview Display only)    | Places the Time Overview display across the top of the application window, above all other active displays.   |
| Show Parameter Readouts                           | For the DPX display, enables/disables the display of DPX parameters. The parameters readout shows 100% Probability of Intercept, Transforms/s, and FFT Points.  |
| Show Marker readout in graph<br>(selected marker) | Shows or hides the readout for the selected marker in the graph area.   |
| Show timestamp in graph<br>(selected line)        | For spectrogram displays, this readout shows or hides the timestamp associated with the selected line or marker position.   |
| Show Power Trigger Level                          | Displays or hides a green line in the graph that indicates the level at which the power trigger is set. The line is not displayed if Trigger is set to Free Run.  |
| Marker Noise mode                                 | Select to enable or disable the Marker Noise mode. Use this mode to measure noise on the trace. See <a href="#">Using Noise Markers in the Spectrum Display (see page 544)</a> .                        |

### Show Parameter Readouts (DPX Spectrum Only)

The elements of the Show Parameter Readouts are:

- 100% POI – This is the minimum event duration required to ensure 100% Probability of Intercept or event capture. This duration depends on interaction between span, RBW, and FFT length. You can use the POI time value to determine if you are at risk of either missing a narrow transient or of its captured amplitude being less than the actual signal.

### Minimum signal duration vs. RBW

|         |           |            |               | Minimum event duration for 100% POI (μs) |         |
|---------|-----------|------------|---------------|--|---------|
| Span    | RBW (kHz) | FFT length | Spectrums/sec | Base unit                                | Opt. 09 |
| 110 MHz | 10000     | 1024       | 292,969       | 17.3                                     | 3.7     |
|         | 1000      | 1024       | 292,969       | 19.5                                     | 5.8     |
|         | 300       | 2048       | 146,484       | 28.5                                     | 14.8    |
|         | 100       | 4096       | 73,242        | 37.6                                     | 37.6    |
|         | 30        | 16384      | 18,311        | 134.6                                    | 134.6   |
|         | 20        | 32768      | 18,311        | 229.2                                    | 229.2   |

**Minimum signal duration vs. RBW (cont.)**

| Span   | RBW (kHz) | FFT length | Spectrums/sec | Minimum event duration for 100% POI (μs) |         |
|--------|-----------|------------|---------------|--|---------|
|        |           |            |               | Base unit                                | Opt. 09 |
| 40 MHz | 5000      | 1024       | 292,969       | 17.5                                     | 3.9     |
|        | 1000      | 1024       | 292,969       | 19.4                                     | 5.8     |
|        | 300       | 1024       | 146,484       | 25                                       | 11.4    |
|        | 100       | 2048       | 73,242        | 37.6                                     | 30.8    |
|        | 30        | 4096       | 36,621        | 93.6                                     | 93.6    |
|        | 20        | 8192       | 18,311        | 147.3                                    | 147.3   |
|        | 10        | 16384      | 18,311        | 194.5                                    | 294.5   |
|        | 20        | 4096       | 73,242        | 133.9                                    | 133.9   |
|        | 10        | 8192       | 36,621        | 267.8                                    | 267.8   |

- **Transforms/s** – This is the number of transforms/second being performed by the DPX engine. It is a function of the selected resolution bandwidth and number of trace points. The desired RBW determines the number of points required in each transform, which alters the transform rate. Additionally, the selected trace length may place requirements on the number of points required in the transform independent of the selected resolution bandwidth.
- **FFT Points** – The approximate number of points in the FFT of the input signal. FFT points provides an indication of frequency resolution, but it is not an exact measurement. Note that higher numbers for FFT points mean that the instrument is taking a longer string of input samples for each FFT, which decreases the time resolution.

**Minimum FFT length versus trace length - independent of span and RBW**

| Trace length (points) | Minimum FFT length |
|-----------------------|--------------------|
| 801                   | 1,024              |
| 2,401                 | 4,096              |
| 4,001                 | 8,192              |
| 10,401                | 16,384             |



## Overview

The displays in Analog Modulation (Displays > Measurements > Analog Modulation) are:

- AM
- FM
- PM

The Analog Modulation displays provide measurements and time-domain trace displays.


## AM Display

The Amplitude Modulation Display is a graph of Modulation Factor vs Time. The AM display includes three numeric readouts.

To show the AM display:

1. Select the **Displays** button or select **Setup > Displays**. This shows the **Select Displays** dialog box.
2. From the **Measurements** box, select **Analog Modulation**.
3. Double-click the **AM** icon in the **Available Displays** box. This adds the AM icon to the **Selected displays** box.
4. Click the **OK** button. This shows the AM display.

### Elements of the Display



| Item | Element                 | Description  |
|------|-------------------------|--|
| 1    | Check mark indicator    | The check mark indicator in the upper, left-hand corner of the display shows when the AM display is the optimized display.<br><br><b>NOTE.</b> When Best for multiple windows is selected in the Amplitude control panel's RF & IF Optimization control, none of the measurement displays shows a checkmark, as there is not a single optimized measurement. |
| 2    | Top of Graph            | Sets the %AM indicated at the top of the graph by increasing or decreasing the vertical scale. Changing the top value affects the bottom of graph value because the graph scales about vertical center. Also, note that the top of graph setting interacts with the internal vertical scale setting (which is not user settable).                            |
| 3    | Position                | Specifies the %AM shown at the center of the graph display.  |
| 4    | Bottom Readout          | Displays the value of the modulation factor shown at the bottom of graph.  |
| 5    | Measurement readouts    | Displays numeric values for the +AM (positive modulation factor), -AM (negative modulation factor), and Total AM.  |
| 6    | Position                | Specifies the horizontal position of the trace on the graph display.   |
| 7    | Scale                   | Adjusts the horizontal range of the graph. By decreasing the scale (time across the entire graph), the graph essentially becomes a window that you can move over the trace by adjusting the position.  |
| 8    | Clear                   | Erases the trace from the graph.   |
| 9    | Trace Detection readout | Displays the Settings > Trace > Detection setting.   |


[Changing the AM Settings \(see page 130\)](#)

## AM Settings

**Menu Bar:** Setup > Settings

**Application Toolbar:** 


The AM Settings control panel provides access to settings that control parameters of the AM Display.



| Setting                                   | Description   |
|---|---|
| <a href="#">Parameters (see page 131)</a> | Sets the Carrier Amplitude Detection method. You can choose either Average or Median.   |
| <a href="#">Trace (see page 132)</a>      | Sets Trace display parameters.  |
| <a href="#">Scale (see page 134)</a>      | Sets vertical and horizontal scale and position parameters.   |
| <a href="#">Prefs (see page 135)</a>      | Specifies whether or not certain display elements are shown. Specifies the maximum number of points shown in the display graph. |

## Parameters Tab

The Parameters tab enables you to specify two parameters that control the carrier amplitude detection.



| Setting        | Description  |
|----------------|--|
| Average        | Selects the Average method for computing the average baseline for measurements.    |
| Median         | Selects the Median method for computing the average baseline for measurements.     |
| Measurement BW | Specifies the bandwidth about the center frequency at which measurements are made. |

## AM Modulation

An amplitude modulated carrier can be described mathematically by:

$$X(t) = [A + a(t)] \cos \omega_0 t$$

A, in the above equation, represents the carrier amplitude, a(t) represents the time-varying modulation and  $\omega_0$  represents the carrier frequency.

The signal modulation envelope is given by:

$$E_M(t) = A + a(t)$$

There are several ways to express the AM modulation depth, expressed as a percentage.

*Peak method*

$$AMModulationDepth = \frac{Max[A + a(t)]}{A}$$

*Trough Method*

$$AMModulationDepth = \frac{Min[A + a(t)]}{A}$$

*Max-Min Method*

$$AMModulationDepth = \frac{Max[A + a(t)] - Min[A + a(t)]}{2A}$$

In each case, the value of the carrier amplitude,  $A$ , needs to be estimated from the input signal.

The instrument represents signals in a sampled form. The sampled envelope can be expressed as a function of sample index  $K$  and sampling period  $T$  as:

$$E_M[KT] = A + a[KT]$$

The instrument allows the choice of two methods for estimating the carrier amplitude:

#### Average Method

$$A = Average \{E_M[KT]\}$$

#### Median Method

$$A = Median \{E_M[KT]\}$$

In both cases, the instrument computes the average and the median over the analysis period. It should be noted that the two methods give the same result for sinusoidal modulation where the sampling frequency is much higher than the modulation frequency.

## Trace Tab

The Trace Tab allows you to set the display characteristics of a trace.



| Setting             | Description  |
|---------------------|--|
| Show                | Shows/hides the trace. If the instrument continues to run, the measurement results below the graph display continue to update even if the trace is hidden. |
| Detection           | Sets the Detection method used for the trace. Available detection methods are +Peak, -Peak, and Avg (VRMS). Not available for saved traces.                |
| Function            | Selects the trace processing method. The only available setting is Normal.   |
| Freeze              | Halts updates to the trace.  |
| Save Trace As       | Saves the trace to a file for later recall and analysis.   |
| Show recalled trace | Displays a saved trace instead of a live trace.  |

## Detection

Detection refers to the method of processing the data acquisition points when creating a trace. The IQ samples in a data acquisition can be detected in a variety of ways. The number of IQ samples available to each trace point varies with both analysis length and trace length. For example, with Spectrum Length set to 'Auto' in the Analysis menu, the instrument analyzes just enough samples to produce one IQ sample pair per trace point. In this case, the detection method chosen has very little effect, as the +Peak, -Peak, and Avg (VRMS) are all equal. Changing the Spectrum Length causes the available detection methods to differ in value because they have a larger set of samples for the various detection methods to process.

The available detection methods are:

- **+Peak** – Each point on the trace is the result of detecting the positive peak value present in the set of IQ samples available to that trace point.
- **-Peak** – Each point on the trace is the result of detecting the negative peak value present in the set of IQ samples available to that trace point.
- **Avg (VRMS)** [Average  $V_{RMS}$ ] – Each point on the trace is the result of determining the RMS Voltage value for all of the IQ samples available to the trace point. When displayed in either linear (Volts, Watts) or Log (dB, dBm), the correct RMS value results. When the averaging function is applied to a trace, the averaging is performed on the linear (Voltage) values, resulting in the correct average for RMS values.

## Trace Processing

Traces can be processed to display in different ways. The Detection setting controls trace decimation, when needed. When the trace points each cover more than one sample data point, the vertical results values for multiple data points are combined into each trace point. Each trace point ends up with a single vertical value.

- **Average** - Each trace point is computed by averaging together the multiple results points it represents.
- **+Peak** - Each trace point represents the highest vertical value among the results it includes.
- **-Peak** - Each trace point represents the lowest vertical value among the results it includes.

## Saving Traces

To save a trace for later analysis:

1. Select the **Save Trace As** button. This displays the Save As dialog box.
2. Navigate to the desired folder or use the default.
3. Type a name for the saved trace and click **Save**.

## Recalling Traces

You can recall a previously saved trace for analysis or comparison to a live trace.

To select a trace for recall:

1. Click the ... button to display the Open dialog box.
2. Navigate to the desired file and click **Open**.
3. Check the **Show Recalled Trace** check box.
4. Verify that the **Show** check box is selected.

## Scale Tab


The Scale tab allows you to change the vertical and horizontal scale settings. Changing the scale settings changes how the trace appears on the display but does not change control settings such as Measurement Frequency. In effect, these controls operate like pan and zoom controls.

The screenshot shows the 'Scale' tab in a software interface. It has four tabs at the top: 'Parameters', 'Trace', 'Scale' (selected), and 'Prefs'. The 'Scale' tab is divided into two main sections: 'Vertical' and 'Horizontal'. Each section contains a 'Scale' input field, a 'Position' input field, and an 'Autoscale' button. In the 'Vertical' section, the Scale is set to '143 %' and the Position is set to '-1.00 %'. In the 'Horizontal' section, the Scale is set to '4.600 ms' and the Position is set to '400.000 us'.

| Setting    | Description   |
|------------|---|
| Vertical   |   |
| Scale      | Changes the range shown between the top and bottom of the graph.                    |
| Position   | Adjusts the level shown at the center of the graph.                                 |
| Autoscale  | Resets the Position so that the entire trace is in the graph.                       |
| Horizontal |   |
| Scale      | Changes the range shown between the left and right sides of the graph.              |
| Position   | Adjusts the position of the acquisition record shown at the left edge of the graph. |
| Autoscale  | Resets the Scale and Position settings to provide the optimum display.              |

## Prefs Tab

The Prefs tab enables you to change appearance characteristics of the AM display.



| Setting  | Description  |
|--|--|
| Show graticule                                 | Shows or hides the graticule.  |
| Show Marker readout in graph (selected marker) | Shows or hides the readout for the selected marker in the graph area.  |
| Max trace points                               | The trace in the AM display can be set to a maximum of 100,000 points (however, the actual number of trace points can extend up to 500,000 points if Max trace points is set to Never Decimate). If the Analysis Length includes more than the selected Max trace points value, the trace is decimated (using the method specified with the Detection control) to be equal to or less than the Max trace points setting (except when Max trace points is set to Never Decimate). This decimated (or undecimated) trace is what is used for marker measurements and for results export. |


## FM Display

The Frequency Modulation Display shows Frequency Deviation vs. Time. The vertical axis units are Hertz and the horizontal axis units are seconds. When taking measurements, only the first burst in the Analysis period is analyzed. No trace points are shown for data outside the first detected burst, nor are measurements made on data outside the first detected burst.

To show the FM display:

1. Select the **Displays** button or select **Setup > Displays**. This shows the **Select Displays** dialog box.
2. From the **Measurements** box, select **Analog Modulation**.
3. Double-click the **FM** icon in the **Available Displays** box. This adds the FM icon to the **Selected displays** box.
4. Click the **OK** button. This shows the FM display.

## Elements of the Display



| Item | Element                 | Description  |
|------|-------------------------|--|
| 1    | Check mark indicator    | The check mark indicator in the upper, left-hand corner of the display shows when the FM display is the optimized display.<br><br><b>NOTE.</b> When Best for multiple windows is selected in the Amplitude control panel's RF & IF Optimization control, none of the measurement displays shows a checkmark, as there is not a single optimized measurement.   |
| 2    | Top of Graph control    | Sets the frequency indicated at the top of the graph. Since the Position value at the vertical center of this graph remains constant as the Top of Graph value is adjusted, the Vertical Scale increases as the Top of Graph value increases, which also affects the bottom of graph readout. Vertical Scale can also be controlled from the Settings control panel's Scale tab.   |
| 3    | Position                | Specifies the frequency shown at the center of the graph display. Changing this value moves the trace up and down in the graph, which affects the Top of Graph and Bottom of Graph values as well.   |
| 4    | Bottom of Graph Readout | Displays the frequency value at the bottom of the graph.   |
| 5    | Measurement readouts    | Displays numeric values for the +Pk (positive frequency deviation), -Pk (negative frequency deviation), RMS (RMS value of the deviation), Pk-Pk (peak-to-peak frequency deviation), and Pk-Pk/2 (peak-to-peak frequency deviation divided by two).   |
| 6    | Position                | Specifies the horizontal position of the trace on the graph display.   |
| 7    | Scale                   | Adjusts the time range of the graph. By decreasing the scale (full-scale time over 10 divisions), the graph essentially becomes a window that you can move over the acquisition record by adjusting the horizontal position.   |
| 8    | Clear                   | Clears the trace and numeric measurement results.  |
| 9    | Trace Detection readout | Displays the Settings > Trace > Detection setting.   |
| 10   | Freq Error              | This readout can show Freq Error or Freq Offset. When it displays Freq Error, it shows the difference between the instrument Frequency setting and the measured value of the signal's carrier frequency. When it displays Freq Offset, it shows the frequency offset specified on the Settings > Parameters tab. If Freq Error is displayed, it also indicates that the Carrier frequency detection setting is Auto. If Freq Offset is displayed, it indicates that the Carrier frequency detection setting is manual. |

[Changing the FM Settings \(see page 137\)](#)

## FM Settings

**Menu Bar:** Setup > Settings

**Application Toolbar:** 

The FM Settings control panel provides access to settings that control parameters of the FM Display.

Parameters Trace Scale Prefs

Burst detect threshold: -100 dBc

Frequency offset: -50.54 Hz ☒ Auto

Measurement BW: 100.0 kHz

| Setting                                   | Description   |
|---|---|
| <a href="#">Parameters (see page 473)</a> | Sets the burst detection threshold, measurement bandwidth, and carrier frequency detection method.                              |
| <a href="#">Trace (see page 140)</a>      | Sets Trace display parameters.  |
| <a href="#">Scale (see page 142)</a>      | Sets vertical and horizontal scale and position parameters.   |
| <a href="#">Prefs (see page 143)</a>      | Specifies whether or not certain display elements are shown. Specifies the maximum number of points shown in the display graph. |

## Parameters Tab

The Parameters tab enables you to specify parameters for carrier detection.

The screenshot shows the 'Parameters' tab selected in a software interface. It contains the following controls:

- Burst detect threshold:** A text input field containing '-100 dBc'.
- Measurement BW:** A text input field containing '100.0 kHz'.
- Frequency offset:** A text input field containing '-50.54 Hz'.
- Auto:** A checked checkbox next to the text 'Auto'.
- Load from marker:** A button located below the Frequency offset field.

| Setting                | Description  |
|------------------------|--|
| Burst detect threshold | This parameter specifies the power level used to determine whether a burst is present. A valid burst is required to take measurements. The burst detected first is used for the analysis. The units for burst detect threshold are dBc, relative to the maximum acquisition sample data points level.  |
| Measurement BW         | Specifies the bandwidth about the center frequency at which measurements are made.   |
| Frequency offset       | Displays the carrier signal's offset from the instrument's measurement frequency. Auto is used to select the method for determining the carrier frequency. When Auto is selected, the instrument determines the carrier frequency by analyzing the signal. When Auto is deselected, you can set the carrier frequency offset using Frequency Offset. |
| Load from Marker       | Pressing this button sets the frequency offset to the frequency offset of the selected marker. (Pressing this button automatically deselects Auto.)  |

### Frequency Offset

In Auto (Auto is selected), the instrument scans the measurement bandwidth about the measurement frequency and looks for the highest-powered signal. This is defined as the carrier frequency. In Manual (Auto is deselected), the carrier frequency is specified by adding/subtracting the specified Frequency offset from the measurement frequency. Range:  $-(\text{Measurement BW} \times 1.1)/2$  to  $+(\text{Measurement BW} \times 1.1)/2$ . If the **Load from Marker** button is pressed, the frequency offset is determined from the phase difference between the two markers on screen compared to the time between the two markers. This function is useful for removing frequency slope from the phase modulated signal.

### Load from Marker

When the Auto check box is selected, the instrument scans the measurement bandwidth about the center frequency and looks for the highest-powered signal. This is defined as the carrier frequency. When the Auto check box is deselected, the carrier frequency is specified by adding/subtracting the specified Frequency offset from the center frequency. Pressing the Load from Marker button loads the marker frequency into the Frequency offset box.

## Trace Tab

The Trace Tab allows you to set the display characteristics of displayed traces.



| Setting             | Description  |
|---------------------|--|
| Show                | Shows / hides the trace. If the instrument continues to run, the measurement results below the graph display continue to update even if the trace is hidden. |
| Detection           | Sets the Detection method used for the trace. Available detection methods are +Peak, -Peak, and Avg (VRMS). Not available for saved traces.                  |
| Function            | Selects the trace processing method. The only available setting is Normal.   |
| Freeze              | Halts updates to the trace.  |
| Save Trace As       | Saves the trace to a file for later recall and analysis.   |
| Show Recalled trace | Displays a saved trace instead of a live trace.  |

### Detection

Detection refers to the method of processing the data acquisition points when creating a trace. The IQ samples in a data acquisition can be detected in a variety of ways. The number of IQ samples available to each trace point varies with both analysis length and trace length. For example, with Spectrum Length set to 'Auto' in the Analysis menu, the instrument analyzes just enough samples to produce one IQ sample pair per trace point. In this case, the detection method chosen has very little effect, as the +Peak, -Peak, and Avg (VRMS) are all equal. Changing the Spectrum Length causes the available detection methods to differ in value because they have a larger set of samples for the various detection methods to process.

The available detection methods are:

- **+Peak** – Each point on the trace is the result of detecting the positive peak value present in the set of IQ samples available to that trace point.
- **-Peak** – Each point on the trace is the result of detecting the negative peak value present in the set of IQ samples available to that trace point.
- **Avg (VRMS)** [Average  $V_{RMS}$ ] – Each point on the trace is the result of determining the RMS Voltage value for all of the IQ samples available to the trace point. When displayed in either linear (Volts, Watts) or Log (dB, dBm), the correct RMS value results. When the averaging function is applied to a trace, the averaging is performed on the linear (Voltage) values, resulting in the correct average for RMS values.

## Trace Processing

Traces can be processed to display in different ways. The Function setting controls trace processing.

- **Normal** - Each new trace is displayed and then replaced by the next trace. Each data point contains a single vertical value.
- **Average** - Each trace point is computed by averaging together the multiple results points it represents.
- **+Peak** - Each trace point represents the highest vertical value among the results it includes.
- **-Peak** - Each trace point represents the lowest vertical value among the results it includes.

## Saving Traces

To save a trace for later analysis:

1. Select the **Save Trace As** button. This displays the Save As dialog box.
2. Navigate to the desired folder or use the default.
3. Type a name for the saved trace and click **Save**.

## Recalling Traces

You can recall a previously saved trace for analysis or comparison to a live trace.

To select a trace for recall:

1. Click the ... button to display the Open dialog box.
2. Navigate to the desired file and click **Open**.
3. Check the **Show Recalled Trace** check box.
4. Verify that the **Show** check box is selected.

## Scale Tab

The Scale tab allows you to change the vertical and horizontal scale settings. Changing the scale settings changes how the trace appears on the display but does not change control settings such as Measurement Frequency. In effect, these controls operate like pan and zoom controls.

The screenshot shows the 'Scale' tab in a software interface. It has four tabs at the top: 'Parameters', 'Trace', 'Scale' (which is selected), and 'Prefs'. The 'Scale' tab is divided into two main sections: 'Vertical' and 'Horizontal'. Each section contains a 'Scale' input field, a 'Position' input field, and an 'Autoscale' button. In the 'Vertical' section, the Scale is set to '10.0 kHz' and the Position is '0.000 Hz'. In the 'Horizontal' section, the Scale is set to '35.280 us' and the Position is '92.160 us'.

| Setting    | Description   |
|------------|---|
| Vertical   |   |
| Scale      | Changes the range shown between the top and bottom of the graph.                    |
| Position   | Adjusts the frequency shown at the center of the graph.                             |
| Autoscale  | Resets the Position so that the entire trace fits within the graph.                 |
| Horizontal |   |
| Scale      | Changes the range shown between the left and right sides of the graph.              |
| Position   | Adjusts the position of the acquisition record shown at the left edge of the graph. |
| Autoscale  | Resets the Scale and Position settings to provide the optimum display.              |

## Prefs Tab

The Prefs tab enables you to change appearance characteristics of the FM display.


The screenshot shows the 'Prefs' tab selected in a software interface. The tab bar at the top includes 'Parameters', 'Trace', 'Scale', and 'Prefs'. The 'Prefs' tab content area contains three settings: a checked checkbox for 'Show graticule', a checked checkbox for 'Show Marker readout in graph (selected marker)', and a dropdown menu for 'Max trace points' currently set to '100K'.

| Setting  | Description  |
|--|--|
| Show graticule                                 | Shows or hides the graticule.  |
| Show Marker readout in graph (selected marker) | Shows or hides the readout for the selected marker in the graph area.  |
| Max trace points                               | The trace in the FM display can be set to a maximum of 100,000 points (however, the actual number of trace points can extend up to 500,000 points if Max trace points is set to Never Decimate). If the Analysis Length includes more than the selected Max trace points value, the trace is decimated (using the method specified with the Detection control) to be equal to or less than the Max trace points setting (except when Max trace points is set to Never Decimate). This decimated (or undecimated) trace is what is used for marker measurements and for results export. |

## PM Display

The Phase Modulation Display shows Phase vs. Time. The vertical axis units are degrees and the horizontal axis units are seconds. When taking measurements, only the first burst in the Analysis period is analyzed. No trace points are shown for data outside the first detected burst, nor are measurements made on data outside the first detected burst.

## Elements of the Display



| Item | Element                 | Description  |
|------|-------------------------|--|
| 1    | Check mark indicator    | The check mark indicator in the upper, left-hand corner of the display shows when the PM display is the optimized display.<br><br><b>NOTE.</b> When Best for multiple windows is selected in the Amplitude control panel's RF & IF Optimization control, none of the measurement displays shows a checkmark, as there is not a single optimized measurement.   |
| 2    | Top of Graph control    | Sets the phase value indicated at the top of the graph. Since the Position value at the vertical center of this graph remains constant as the Top of Graph value is adjusted, the Vertical Scale increases as the Top of Graph value increases, which also affects the bottom of graph readout. Vertical Scale can also be controlled from the Settings control panel's Scale tab.   |
| 3    | Position                | Specifies the phase shown at the center of the graph display. Changing this value moves the trace up and down in the graph, which affects the Top of Graph and Bottom of Graph values as well.   |
| 4    | Bottom of Graph Readout | Displays the phase value at the bottom of the graph.   |
| 5    | Measurement readouts    | Displays numeric values for the +Pk (positive phase deviation), -Pk (negative phase deviation), Pk-Pk (peak-to-peak phase deviation), and RMS (RMS value of the phase deviation).  |
| 6    | Position                | Specifies the horizontal position of the trace on the graph display.   |
| 7    | Scale                   | Adjusts the time range of the graph. By decreasing the scale (full-scale time over 10 divisions), the graph essentially becomes a window that you can move over the acquisition record by adjusting the horizontal position.   |
| 8    | Clear                   | Clears the measurement results.  |
| 9    | Trace Detection readout | Displays the Trace Detection setting (see Settings > Trace > Detection).   |
| 10   | Freq Error              | This readout can show Freq Error or Freq Offset. When it displays Freq Error, it shows the difference between the instrument Frequency setting and the measured value of the signal's carrier frequency. When it displays Freq Offset, it shows the frequency offset specified on the Settings > Parameters tab. If Freq Error is displayed, it also indicates that the Carrier frequency detection setting is Auto. If Freq Offset is displayed, it indicates that the Carrier frequency detection setting is manual. |

[Changing the PM Settings \(see page 145\)](#)

## PM Settings

**Menu Bar:** Setup > Settings

**Application Toolbar:** 

The PM Settings control panel provides access to settings that control parameters of the PM Display.

Parameters

Trace

Scale

Prefs

Burst detect threshold: -100 dBc

Measurement BW: 1.000 MHz

Frequency offset: -383.2 Hz ☐ Auto

Phase offset: 128 ° ☐ Auto

Load Δ from marker

Load from marker

| Setting                                   | Description  |
|---|--|
| <a href="#">Parameters (see page 146)</a> | Sets the Carrier Frequency Detection method. You can choose either Automatic or Manual. Sets burst threshold and phase detection method. |
| <a href="#">Trace (see page 147)</a>      | Sets Trace display parameters.   |
| <a href="#">Scale (see page 150)</a>      | Sets vertical and horizontal scale and position parameters.  |
| <a href="#">Prefs (see page 151)</a>      | Specifies whether or not certain display elements are shown. Specifies the maximum number of points shown in the display graph.          |

## Parameters Tab

The Parameters tab enables you to specify parameters that control the carrier frequency and phase detection.

Parameters

Trace

Scale

Prefs

Burst detect threshold: -100 dBc

Measurement BW: 1.000 MHz

Frequency offset: -383.2 Hz ☐ Auto

Phase offset: 128 ° ☐ Auto

Load Δ from marker

Load from marker

| Setting                   | Description  |
|---------------------------|--|
| Burst detect threshold    | Specifies the power level used to determine whether a burst is present. A valid burst is required to take measurements. The burst detected first is used for the analysis. The units for burst detect threshold are dBc, relative to the maximum acquisition sample data points level.   |
| Measurement BW            | Specifies the bandwidth about the center frequency at which measurements are made.   |
| Frequency offset          | Displays the carrier signal's frequency offset from the instrument's Measurement Frequency. Auto is used to select the method for determining the carrier frequency. When Auto is selected, the instrument determines the carrier frequency by analyzing the signal. When Auto is deselected, you can set the carrier frequency offset using Frequency Offset. |
| Load $\Delta$ from Marker | Pressing this button disables the Auto Frequency Offset function and sets the Frequency offset to the frequency offset calculated from the difference between MR (the marker reference) and the selected marker (or M1 if MR is currently the selected marker).  |
| Phase offset              | Displays the phase offset. When Auto is selected, the instrument determines the phase offset. When Auto is deselected, the user sets the phase offset with this control.   |
| Load from marker          | Pressing this button sets the phase offset to the phase offset of the selected marker. (Pressing this button automatically deselects Auto.)  |

## Frequency Offset

In Auto (Auto is selected), the instrument scans the measurement bandwidth about the Measurement Frequency and looks for the highest-powered signal. This is defined as the carrier frequency. In Manual (Auto is deselected), the carrier frequency is specified by adding/subtracting the specified Frequency offset from the Measurement Frequency. Range:  $-(\text{Measurement BW} \times 1.1)/2$  to  $+(\text{Measurement BW} \times 1.1)/2$ . If the **Load  $\Delta$  from marker** button is pressed, the frequency offset is determined from the phase difference between the two markers on screen compared to the time between the two markers. This function is useful for removing frequency slope from the phase modulated signal.

## Phase Offset

In Auto (Auto is selected), the instrument sets the phase offset to fit the waveform to the screen centering  $0^\circ$  phase on the vertical axis. The actual phase offset is indicated on the Phase Offset readout. In Manual (Auto is deselected), the instrument sets the phase offset using the value specified in Phase offset. For example, when Phase offset is set to  $10^\circ$ , the waveform shifts upward by  $10^\circ$  on screen. Selecting **Load from marker** sets the phase offset to the value of the selected marker. Range:  $-180$  to  $+180^\circ$ .

## Trace Tab

The Traces Tab allows you to set the display characteristics of displayed traces.

| Setting             | Description  |
|---------------------|--|
| Show                | Shows / hides the trace. If the instrument continues to run, the measurement results below the graph display continue to update even if the trace is hidden. |
| Detection           | Sets the Detection method used for the trace. Available detection methods are +Peak, -Peak, and Avg (VRMS). Not available for saved traces.                  |
| Function            | Selects the trace processing method. The only available setting is Normal.   |
| Freeze              | Halts updates to the trace.  |
| Save Trace As       | Saves the trace to a file for later recall and analysis.   |
| Show Recalled trace | Displays a saved trace instead of a live trace.  |

## Detection

Detection refers to the method of processing the data acquisition points when creating a trace. The IQ samples in a data acquisition can be detected in a variety of ways. The number of IQ samples available to each trace point varies with both analysis length and trace length. For example, with Spectrum Length set to 'Auto' in the Analysis menu, the instrument analyzes just enough samples to produce one IQ sample pair per trace point. In this case, the detection method chosen has very little effect, as the +Peak, -Peak, and Avg (VRMS) are all equal. Changing the Spectrum Length causes the available detection methods to differ in value because they have a larger set of samples for the various detection methods to process.

The available detection methods are:

- **+Peak** – Each point on the trace is the result of detecting the positive peak value present in the set of IQ samples available to that trace point.
- **-Peak** – Each point on the trace is the result of detecting the negative peak value present in the set of IQ samples available to that trace point.
- **Avg (VRMS)** [Average  $V_{RMS}$ ] – Each point on the trace is the result of determining the RMS Voltage value for all of the IQ samples available to the trace point. When displayed in either linear (Volts, Watts) or Log (dB, dBm), the correct RMS value results. When the averaging function is applied to a trace, the averaging is performed on the linear (Voltage) values, resulting in the correct average for RMS values.

## Trace Processing

Traces can be processed to display in different ways. The Function setting controls trace processing.

- **Normal** - Each new trace is displayed and then replaced by the next trace. Each data point contains a single vertical value.
- **Average** - Each trace point is computed by averaging together the multiple results points it represents.
- **+Peak** - Each trace point is represents the highest vertical value among the results it includes.
- **-Peak** - Each trace point is represents the lowest vertical value among the results it includes.

### Saving Traces

To save a trace for later analysis:

1. Select the **Save Trace As** button. This displays the Save As dialog box.
2. Navigate to the desired folder or use the default.
3. Type a name for the saved trace and click **Save**.

### Recalling Traces

You can recall a previously saved trace for analysis or comparison to a live trace.

To select a trace for recall:

1. Click the ... button to display the Open dialog box.
2. Navigate to the desired file and click **Open**.
3. Check the **Show Recalled Trace** check box.
4. Verify that the **Show** check box is selected.

## Scale Tab

The Scale tab allows you to change the vertical and horizontal scale settings. Changing the scale settings changes how the trace appears on the display but does not change control settings such as Measurement Frequency. In effect, these controls operate like pan and zoom controls.

The screenshot shows the 'Scale' tab in a software interface. It has four tabs at the top: 'Parameters', 'Trace', 'Scale' (which is selected), and 'Prefs'. The 'Scale' tab is divided into two main sections: 'Vertical' and 'Horizontal'. Each section contains a 'Scale' input field, a 'Position' input field, and an 'Autoscale' button. In the 'Vertical' section, the Scale is set to '22.8 °' and the Position is set to '-24.7 m°'. In the 'Horizontal' section, the Scale is set to '1.840 ms' and the Position is set to '160.000 us'.

| Setting    | Description   |
|------------|---|
| Vertical   |   |
| Scale      | Changes the range shown between the top and bottom of the graph.                  |
| Position   | Adjusts the phase shown at the center of the graph.                               |
| Autoscale  | Resets the Position so that the entire trace in the graph.                        |
| Horizontal |   |
| Scale      | Changes the range shown between the left and right sides of the graph.            |
| Position   | Adjusts the phase shown at the left edge of the graph.                            |
| Autoscale  | Resets the Scale and Position settings to show the entire trace within the graph. |

## Prefs Tab

The Prefs tab enables you to change appearance characteristics of the PM display.

The screenshot shows the 'Prefs' tab selected in a software interface. The tab bar at the top includes 'Parameters', 'Trace', 'Scale', and 'Prefs'. The main area of the 'Prefs' tab contains three settings: a checked checkbox for 'Show graticule', a checked checkbox for 'Show Marker readout in graph (selected marker)', and a 'Max trace points' dropdown menu currently set to '100K'.

| Setting  | Description  |
|--|--|
| Show graticule                                 | Shows or hides the graticule.  |
| Show Marker readout in graph (selected marker) | Shows or hides the readout for the selected marker in the graph area.  |
| Max trace points                               | The trace in the PM display can be set to a maximum of 100,000 points (however, the actual number of trace points can extend up to 500,000 points if Max trace points is set to Never Decimate). If the Analysis Length includes more than the selected Max trace points value, the trace is decimated (using the method specified with the Detection control) to be equal to or less than the Max trace points setting (except when Max trace points is set to Never Decimate). This decimated (or undecimated) trace is what is used for marker measurements and for results export. |



## Overview

The displays in RF Measurements (Displays > Measurements > RF Measurements) are:

- CCDF
- Channel Power and ACPR
- Frequency and Phase Settling Time (Option 12)
- MCPR
- Occupied Bandwidth
- Phase Noise (Option 11)
- SEM (Spectrum Emission Mask)
- Spurious


The RF Measurements power measurements and signal statistics help you characterize components and systems.

### Power Measurements

| Measurement                  | Description  |
|------------------------------|--|
| Channel Power                | The total RF power in the selected channel (located in the ACPR display).  |
| Adjacent Channel Power Ratio | Measure of the signal power leaking from the main channel into adjacent channels.  |
| Multi-Carrier Power Ratio    | The ratio of the signal power in the reference channel or group of channels to the power in adjacent channels.   |
| Peak/Avg Ratio               | Ratio of the peak power in the transmitted signal to the average power in the transmitted signal (located in the CCDF display).  |
| CCDF                         | The <i>Complementary Cumulative Distribution Function</i> (CCDF). CCDF shows how much time a signal spends at or above a given power level relative to the average power of a measured signal. |

## Channel Power and ACPR (Adjacent Channel Power Ratio) Display

Use the **Channel Power and ACPR** measurement to measure channel power by itself, or adjacent channel leakage ratio with one main channel.




## Measuring Adjacent Channel Power Ratio

1. Select the **Displays** button.
2. Select **RF Measurements** from the **Measurements** box.
3. Double-click **Chan Power and ACPR** in the Available displays box. Click **OK** to complete your selection.
4. Press the front-panel **Freq** button and use the front panel keypad or knob to adjust the frequency to that of your main channel.
5. Press the **Settings** button. This displays the control panel for Chan Power and ACPR (the tab displayed will be the tab displayed the last time the Settings panel was opened).

6. To set the number of adjacent channels, select the **Channels** tab. Enter the number of channels in the **Number of adjacent pairs** value box. If zero is entered for the number of adjacent pairs, the resultant measurement will be channel power only.

**NOTE.** As you add adjacent channels, the span of the display is adjusted so that all the channels can be seen. Use the horizontal scale and offset to zoom the display in on any portion of the trace.

7. To set the spacing between channel centers, enter the required value in the **Channel Spacing** value box.

8. To set the channel bandwidth, enter the required value in the **Channel Bandwidth** value box.
9. After you have configured the channel settings, click the close button () in the Settings panel or press the **Settings** button again to remove the settings panel.
10. Press the **Run** button to take the measurements.

## Viewing Results

Measurement results are displayed in a table below the graph and within the graph itself (which can be enabled/disabled in the Settings > Prefs tab). To see all measurements, you might need to scroll the table. The size of the results table can be changed by dragging the horizontal divider bar between the graph and table areas.

|                                    |    |           |           |           |
|------------------------------------|----|-----------|-----------|-----------|
| Channel Power (Main):<br>-1.94 dBm | Ch | Offset    | Lower     | Upper     |
|                                    | A1 | 2.000 MHz | -65.69 dB | -64.82 dB |

| Heading                  | Description  |
|--------------------------|--|
| Channel                  | Identifies the displayed channels. A1 means the first adjacent channel. A2 means second adjacent channel. Adjacent channels are numbered according to their offset from the Main channel. The closest channel is numbered 1. The next closest channel is numbered 2; and so forth. |
| Lower                    | The power measured for the lower adjacent channel. Adjacent channel power measurements are displayed in dB relative to the Main channel.   |
| Upper                    | The power measured for the upper adjacent channel. Adjacent channels power measurements are displayed in dB relative to the Main channel.  |
| Avg Channel Power (Main) | The power measured for the Main channel. The detection type used for measuring channel power and adjacent channel powers is Average. This means that the average of the linear (pre-log) samples is used to determine the power. The Main channel power is displayed in dBm.       |

[Setting Channel Power and ACPR Settings Parameters \(see page 156\)](#)

## RF Channel Power Measurement

The RF channel power gives an indication of the total average (and other measures) RF power in a given channel.

For some communications systems, there is an “out-of-service” total power measurement defined in the specifications that calls for a specified constant modulation. In this case, the output power should be relatively constant. For many measurements, this may not be the case, and the Power Measurement results will vary as the signal varies.

Average power is the square root of the sum of the squares of the voltage samples over the measurement time.

The defined “channel” width for the Power Measurement defines the bandwidth and shape of the filter used to remove any RF power on frequencies outside of the channel bandwidth.

## Channel Power

The total RF power in the selected frequency band. The detection type used for measuring channel power is Average. This means that the average of the linear (pre-log) samples is used to determine the channel power. To measure channel power, use the ACLR measurement, and set the number of adjacent channels to zero. This results in only the channel power being measured.

## Average Channel Power

The total RF power in the selected channel (located in the ACPR display).

## Adjacent Channel Leakage Power Ratio

Adjacent Channel Leakage power Ratio (ACLR) is the ratio of the mean power centered on the assigned channel frequency to the mean power centered on an adjacent channel frequency. In the 3GPP specification, both the main channel and adjacent channels are required to be filtered with RRC (Root Raised Cosine) filters.

## Adjacent Channel Power

Measure of the signal power leaking into nearby channels.

## Channel Power and ACPR Settings

**Menu Bar:** Setup > Settings

**Front Panel:** Settings

**Application Toolbar:** 

The settings for the Channel Power and ACPR display are shown in the following table.

Freq & RBW | Measurement Params | Channels | Scale | Prefs

Meas Freq: 1.50000 GHz      RBW: 30.000 kHz      ☒ Auto

Step: 2.0000 MHz      ☒ Auto      ☐ VBW:

| Settings tab                                      | Description  |
|---|--|
| <a href="#">Freq &amp; RBW (see page 165)</a>     | Specify the frequency and resolution bandwidth used for the ACPR measurement.  |
| <a href="#">Measurement Params (see page 166)</a> | Specify several parameters that control the measurement, such as channel filter, chip rate, averaging, and correcting for noise floor. |
| <a href="#">Channels (ACPR) (see page 157)</a>    | This tab specifies the BW and offset parameters of the Channels for the selected ACPR measurement.                                     |
| <a href="#">Scale (see page 230)</a>              | Specifies the vertical and horizontal scale settings.  |
| <a href="#">Prefs (see page 231)</a>              | Specifies whether certain display elements are visible.  |

**Restore defaults.** Sets parameters for a 1-channel WCDMA measurement.

## Channels Tab for ACPR

The Channels tab is where you specify parameters for the channels measured in the ACPR display.

Freq & RBW | Measurement Params | Channels | Scale | Prefs


Number of adjacent pairs: 1

Channel Bandwidth: 1.000 MHz

Channel Spacing: 2.000 MHz

| Setting                  | Description  |
|--------------------------|--|
| Number of adjacent pairs | Specifies the number of adjacent channel pairs. Range: 1 - 50; Resolution 1.       |
| Channel Bandwidth        | Specifies the frequency width of each channel (all channels share the same value). |
| Channel Spacing          | Specifies the difference in frequency between the centers of each channel.         |

The following figure illustrates the settings controlled from the Channels tab.



### Changing the Number of Adjacent Pairs

1. Select **Number of adjacent pairs** number entry box.
2. Enter a value for the number of adjacent pairs using the knob or the keypad.

---

**NOTE.** As you change the value for the number of adjacent pairs, the analyzer will update the spectrum display to identify the adjacent channels.

---

3. Select the Close box when you have finished making changes.

### Changing the Channel Bandwidth

1. Select the **Channel Bandwidth** number entry box.
2. Enter a value for the number of adjacent pairs using the knob or the keypad.

---

**NOTE.** As you change the value for the number of adjacent pairs, the analyzer will update the display to indicate the channel bandwidth.

---

3. Select the Close box when you have finished making changes.

### Channel Spacing

1. Select **Channel Spacing** number entry box.
2. Enter a value for the difference in center frequency between channels using the knob or the keypad.

---

**NOTE.** As you change the value for the number of adjacent pairs, the analyzer will update the display to indicate the channel bandwidth.

---


3. Select **Close** when you have finished making changes.

## MCPR (Multiple Carrier Power Ratio) Display

Use the MCPR measurement to measure adjacent channel power ratio for multiple main channels or when adjacent channel pairs have different offsets and/or bandwidths. The MCPR display show the Reference Power and the ratio of each adjacent channel to the Reference Power. You can select whether the Reference Power is the total of all active channels or a single channel.

### Measuring Multiple Carrier Power Ratio

1. Press the front-panel **Displays** button.
2. From the **Select Displays** window, select **RF Measurements** or **P25 Analysis** from the Measurements box.
3. Double-click the **MCPR** icon in the **Available displays** box. Click **OK** to complete your selection.
4. Press the front-panel **Freq** button and use the front panel keypad or knob to adjust the frequency to that of the main channel. Select **Meas. Freq** to set the measurement frequency. Use the front panel knob to adjust the frequency.
5. Press the **Settings** button. This displays the control panel for MCPR (the tab displayed will be the tab displayed the last time the Settings panel was opened).



6. Click on the **Channels** tab. To set the number of Main channels, enter the number of Main channels in the **Number** text entry box under **Main Channels**.


---

**NOTE.** As you add channels, the span of the display will be adjusted so that all the channels can be seen.

---

7. To set the bandwidth of all main channels, enter a value in the **BW** value box.
8. To set the spacing between the main channels, enter a value in the **Spacing** value box.
9. To prevent a main channel power level from being included in the Power Reference (Total) calculation, select the channel to be excluded in the **Inactive** box.

10. To add adjacent channels, click the **Add** button under **Adj Chans**.




11. To edit Frequency Offset or Bandwidth, click within the value box and use the front-panel knob or your keyboard to set the value. Click **Close** to save your changes.


---


**NOTE.** *Frequency Offset is the difference between the center frequency of the selected channel and the Measurement Frequency. All channels (Main or Adjacent) are specified by their offset from the Measurement Frequency.*

---

12. To enable and define power limits, check the **Enable Power Limits** box and enter the desired Upper and Lower limits. You can read more about power limits [here \(see page 162\)](#).




13. After you have configured the Main and Adjacent channels, click the close button (  ) in the Settings panel or the **Settings** button to remove the settings panel.

14. Click the Run button (  ) to take the measurements.

## Viewing Results

MCPR measurement results are displayed both in the spectrum graph and in a table below the spectrum display. Individual Main channel power measurements appear in the graph under their channel identifiers. Individual adjacent channel power ratios also appear in the graph under the adjacent channel identifier.



The following table details the entries in the results table.


| Heading         | Description  |
|-----------------|--|
| Ch              | Identifies the adjacent channels. Adjacent channels are numbered according to their offset from the Main channel. The closest channel is numbered 1. The next closet channel is numbered 2; and so forth. The prefix "L" means lower; "U" means upper. |
| Offset          | The offset is the difference in frequency between the Measurement Frequency and the center frequency of the channel.   |
| Lower           | Displays the ratio of the lower adjacent channel power to the Power Reference.   |
| Upper           | Displays the ratio of the upper adjacent channel power to the Power Reference.   |
| L. Limit        | Displays the lower power limit.  |
| U. Limit        | Displays the upper power limit.  |
| Power Reference | Depending on the selected Power Ref, this readout displays either the power of the selected main channel or the total power in the active main channels.   |

Depending on the number of adjacent channels, to see all the measurements, you might need to drag the divider bar between the main display and the results table at the bottom of the window to change the height of the results table.

[Setting MCPR Measurement Parameters \(see page 166\)](#)

### Setting Power Limits

When power limits are enabled (**Settings > Channels > Add**), the comparison of the actual results of an adjacent channel is done against the limits. If there is a failure, the Lower and Upper power cells in the MCPR table will be red and the failure segments of the trace display will be red. The words **Fail** in red will also appear in the top left corner of the display (even if one adjacent band fails). See the following image for an example.



## Multiple Carrier Power Ratio


The ratio of the signal power in an adjacent channel to the power in one or more main channels.

## MCPR Settings

**Menu Bar: Setup > Settings**

**Front Panel / Application Toolbar: Settings**

The Setup control panel tabs for MCPR (Multiple Carrier Power Ratio) are:



| Settings tab                                      | Description  |
|---|--|
| <a href="#">Freq &amp; RBW (see page 165)</a>     | Specify the frequency and resolution bandwidth used for the MCPR measurements. |
| <a href="#">Measurement Params (see page 166)</a> | Specifies parameters controlling how the MCPR measurement is made.             |
| <a href="#">Channels (see page 168)</a>           | Specifies the parameters of the channels to be measured.                       |
| <a href="#">Scale (see page 230)</a>              | Specifies the vertical and horizontal scale and offset values.                 |
| <a href="#">Prefs (see page 231)</a>              | Specifies whether or not certain display elements are shown.                   |

**Restore defaults.** Sets parameters for a 4-carrier WCDMA measurement.

## Freq & RBW Tab for ACPR and MCPR Displays

The Freq & RBW tab specifies frequency parameters for the Channel Power & ACPR measurements and MCPR measurement.

The screenshot shows the 'Freq & RBW' tab selected. It contains the following controls:

- Meas Freq:** A text box containing '1.00000 GHz'.
- RBW:** A text box containing '30.000 kHz' and a checked 'Auto' checkbox.
- Step:** A text box containing '2.0000 MHz' and a checked 'Auto' checkbox.
- VBW:** An unchecked checkbox.

| Setting   | Description  |
|-----------|--|
| Meas Freq | Specifies the center/measurement frequency.  |
| Step      | The Step control sets the increment/decrement size for the knob adjustment of the center frequency. Arrow key adjustment of frequency is also affected (it is 10x the knob adjustment). If Auto is enabled, the signal analyzer will adjust the Step size as required. |
| RBW       | Select Auto or Manual. Adjusts the RBW for the entire measurement. This setting is Independent of the Spectrum view's RBW setting.   |
| VBW       | Adjusts the VBW (Video Bandwidth) value. VBW Maximum = current RBW value. VBW Minimum = 1/10,000 RBW setting.  |

**NOTE.** While the center frequency and step settings on this control tab are shared between ACPR, MCPR and Occupied Bandwidth, the RBW setting is not shared. The RBW setting is unique for each measurement.

### VBW

The VBW setting enables/disables the Video Bandwidth filter. VBW is used in traditional swept analyzers to reduce the effect of noise on the displayed signal. The VBW algorithm in the analyzer emulates the VBW filters of traditional swept analyzers.

The maximum VBW value is the current RBW setting. The minimum VBW value is 1/10,000 of the RBW setting.

**NOTE.** If you are following a procedure that says to “set VBW to three times the RBW value or greater”, it means that the test should be conducted with no VBW effects. In the RSA6100B Series instruments, this condition is met by disabling the VBW function.

## Measurement Params for ACPR and MCPR Displays

The Measurement Params tab is where you set parameters that control the ACPR and MCPR measurements.

|  |                         |                   |          |                 |                    |
|--|-------------------------|-------------------|----------|-----------------|--------------------|
| Freq & RBW                                       | Measurement Params      | Channels          | Scale    | Prefs           |                    |
| Requires triggered synchronous signal            |                         |                   |          |                 |                    |
| Average:   | Time Domain             | Number:           | 10       | Channel Filter: | Root-raised Cosine |
| <input type="checkbox"/> Correct for Noise Floor |                         |                   |          |                 |                    |
| Power Reference:                                 | Total (active channels) | Filter Parameter: | 0.22     |                 |                    |
|  |                         | Symbol/Chip Rate: | 3.84 MHz |                 |                    |

|  |                         |                   |          |                 |                    |
|--|-------------------------|-------------------|----------|-----------------|--------------------|
| Freq & RBW                                       | Measurement Params      | Channels          | Scale    | Prefs           |                    |
| Average:   | Frequency Domain        | Number:           | 10       | Channel Filter: | Root-raised Cosine |
| <input type="checkbox"/> Correct for Noise Floor |                         |                   |          |                 |                    |
| Power Reference:                                 | Total (active channels) | Filter Parameter: | 0.22     |                 |                    |
|  |                         | Symbol/Chip Rate: | 3.84 MHz |                 |                    |

| Parameter               | Description   |
|-------------------------|---|
| Average                 | Enables/disables measurement averaging. Averaging can be enabled in either the Frequency Domain or Time Domain.   |
| Frequency-domain        | This setting takes the average linear value of the traces (so that rms values are preserved). The number of averages is user-defined. Frequency domain averaging is available in spans larger (or smaller) than the maximum real time bandwidth. This is the mode to use unless you need to extract maximum dynamic range from an ACPR measurement.   |
| Time-domain             | This setting takes the average linear value of the traces. It is useful if you need to extract maximum dynamic range from an ACPR measurement. The number of traces is user defined. But, the signals must be triggered and repeating. That is, the signal needs to be exactly the same for each acquisition. When this condition is met, each waveform contains the same signal, but the random noise changes from acquisition to acquisition and the average value of the random noise is lowered, while the signal value remains constant. Time domain averaging is not available in spans wider than the maximum real-time bandwidth. |
| Number                  | Specifies the number of acquisitions to be averaged when Averaging is set to Frequency Domain. Specifies the number of successive capture records that are averaged when Averaging is set to Time Domain.   |
| Correct for Noise Floor | Enables/disables the Noise Floor Correction function.   |
| Power Reference         | Specifies power against which the results are compared. Choices available are each of the existing main channels and "Total (active channels)".<br>Only present when MCPR is the selected display.  |
| Channel Filter          | Specify the channel filter used. Select between None (default) and Root-raised Cosine.  |
| Filter Parameter        | Value entry box for the Root-raised Cosine filter Adjacent Channel Filter.<br>Present only when the Channel Filter is set to Root-raised Cosine.  |
| Symbol/Chip Rate        | Value entry box for the Root-raised Cosine chip rate.<br>Present only when the Channel Filter is set to Root-raised Cosine.   |

### Correct for Noise Floor

When this setting is enabled, the instrument applies a correction to the ACPR or MCPR measurement to reduce the effect of instrument noise on the results. It generates this correction by taking a preliminary acquisition to measure the instrument noise floor. Once this is done, the measurement proceeds, applying the correction to each result. When any relevant settings (reference level, attenuator, frequency, or span) are changed, the instrument performs a new noise measurement and correction.

The noise correction signal is created by switching off the input to the RTSA and performing acquisitions of the instrument internal noise. A minimum of 100 acquisitions are averaged to create the noise reference signal. It is possible to increase the number of acquisitions for creation of the noise reference signal. When frequency domain averaging is enabled and the number of averages exceeds 100, the number of frequency domain averages becomes the total number of acquisitions to determine the instrument internal noise. Noise is measured for each channel defined by the measurement. The noise reference from each channel is subtracted from the incoming signal power for each channel to create the corrected result. All calculations are performed in Watts, and then converted to the desired units.

The amount of noise correction is limited to 12 dB to avoid the possibility of a negative power measurement. This is a rare condition that could occur if the subtraction of the reference power from the channel power results in a negative value (or ‘infinite’ dBm).


The measured values differ from the displayed trace in two ways. First, any filtering applied to the channels is not displayed on the trace. Second, the single value of measured noise for a channel is subtracted from each trace point in the channel, rather than offsetting the entire channel by a single amount. This produces a smooth trace with no discontinuities at the channel edges.

## Channels Tab for MCPR

**Path:** Setup > Settings > Channels


**Application Toolbar / Front Panel: Settings**

The Channels tab is where you specify the parameters for the channels measured in the MCPR display.



| Setting                  | Description   |
|--------------------------|---|
| <b>Main Channels</b>     |   |
| Number                   | Sets the number of Main channels.   |
| BW                       | Sets the bandwidth of the Main channels.  |
| Spacing                  | Sets the frequency difference between the centers of the Main channels.   |
| Inactive                 | Selected channels are ignored by the measurement routine.   |
| <b>Adj Chans</b>         |   |
| Add                      | Opens the Edit Channel Parameters dialog, which you use to specify the parameters of a new pair of adjacent channels. You can add up to 50 adjacent channel pairs.  |
| Edit                     | Opens the Edit Channel Parameters dialog box. (You must first select a channel to enable the Edit button.)  |
| Delete                   | Deletes the selected main channel or adjacent channel pair.   |
| Channel Definition table | A table of the parameters for every channel. Offset is the frequency difference between the Center Frequency and center of the selected channel. Nothing in the table is editable directly. You can set the upper and lower limits by clicking on the Add button. |

The following figure illustrates the settings controlled from the Channels tab.



### Editing Channel Parameters

To edit the offset or bandwidth of a Main channel:

1. Select the value in the **BW** or **Spacing** number entry boxes.
2. Adjust the value as required. All the Main channels share the same BW and Spacing.
3. Select the appropriate check box to make a channel Inactive, as appropriate.
4. Click the Close button when you have finished making changes.

To edit the offset or bandwidth of an Adjacent channel:

1. Select the channel to be edited in the channel definition table.
2. Select **Edit**. This displays the **Edit Channel Parameters** dialog box.
3. Adjust the values for the **Frequency Offset**, **Bandwidth**, and upper and lower **Limits**, as required.
4. Click the Close button when you have finished making changes.

### Adding Channels

To add a main channel:

1. Select the **Number** entry box.
2. Use the knob to enter a value for the desired number of Main channels.
3. Click the Close button when you have finished making changes.

To add an adjacent channel pair:

1. Select **Add**. This displays the **Edit Channel Parameters** dialog box.
2. Use the knob to enter values for the **Frequency Offset**, **Bandwidth**, and upper and lower **Limits**, as required. The Frequency Offset is measured from the Meas Frequency of the Main channels.
3. Click the Close button when you have finished making changes.

### Deleting Channels

To delete a main channel:

1. Select the **Number** entry box.
2. Use the knob to reduce the number of channels or enter a new value using the front-panel keypad.
3. Click the Close button when you have finished making changes.

To delete an adjacent channel:

1. Select the channel to be deleted in the channel definition table.
2. Select **Delete**. This removes the selected channel from the channel table.
3. Click the Close button when you have finished making changes.


## Occupied BW & x dB BW Display

The Occupied BW display shows the bandwidth within which 99% (a user-defined value, the default is 99%) of the power transmitted within the measurement bandwidth falls.

To show the Occupied BW display:

1. Press the **Displays** button or select **Setup > Displays**.
2. In the **Select Displays** dialog, select **RF Measurements** in the **Measurements** box.
3. In the **Available displays** box, double-click the **Occupied Bandwidth** icon or select the icon and click **Add**. The Occupied Bandwidth icon will appear in the **Selected displays** box and will no longer appear under Available displays.
4. Click **OK** to display the Occupied Bandwidth.

## Elements of the Display



| Item | Display element        | Description   |
|------|------------------------|---|
| 1    | Check mark indicator   | The check mark indicator denotes the display for which the acquisition hardware is optimized. This indicator appears only when the display is the selected display.<br><br><b>NOTE.</b> When Best for multiple windows is selected in the Amplitude control panel's RF & IF Optimization control, none of the measurement displays shows a checkmark, as there is not a single optimized measurement. |
| 2    | Averaging readout      | Appears when averaging is enabled (Settings > Parameters tab > Average results). Displays "Avg: n of m" while averaging the results and "Avg: m" once the requested number of results have been averaged.   |
| 3    | Autoscale              | Adjusts the vertical and horizontal settings so that the entire trace fits in the graph.  |
| 4    | Trace function readout | Displays either Normal or MaxHold to indicate trace function.   |
| 5    | Clear                  | Resets count for Average and MaxHold functions. Enabled only when Averaging or MaxHold is enabled. Pressing Clear will clear the trace and, if acquisition is running, restart the averaging or hold process.   |
| 6    | Graph                  | Displays the input signal. Shaded areas indicate the measurement bandwidth (Settings > Parameters tab > Measurement BW).  |
| 7    | Blue lines             | The blue lines indicate where the selected results (selected in the Main results area) are being measured on the trace.   |
| 8    | Grid divider           | Determines the portion of the display allocated to the graph and detailed results area. You can move the grid divider all the way to the top or bottom and any position in between.   |
| 9    | Detailed results area  | Displays additional measurements results.   |
| 10   | Position and dB/div    | Position sets the top of graph value. The dB/div setting is the vertical scale value.   |
| 11   | Main results area      | Displays results for Occupied Bandwidth (OBW) and x dB bandwidth. Use the two radio buttons to select which of the two results are illustrated in the graph with the blue lines and arrows. There are also controls for adjusting OBW % Power, x dB BW, and RBW.  |
| 12   | VBW                    | Enables the VBW (Video Bandwidth) filter. Displays current VBW filter setting. See Setup > Settings > <a href="#">Freq &amp; RBW tab (see page 165)</a> .   |

### Detailed Results Readouts

| Measurement       | Description   |
|-------------------|---|
| Freq Error        | The frequency difference between the measured carrier frequency of the signal and the user-selected center frequency of the analyzer. |
| OBW Level left    | The signal level at the left edge of the occupied bandwidth.  |
| OBW Level right   | The signal level at the right edge of the occupied bandwidth.   |
| OBW Ref Power     | The total power measured within the measurement bandwidth.  |
| x dB BW Ref Power | The peak power measured within the measurement bandwidth.   |

[Changing the Occupied Bandwidth Settings \(see page 173\)](#)

## Occupied Bandwidth

Bandwidth within which some defined percentage of the power transmitted on a single channel lies.

## Occupied BW & x dB BW Settings

**Menu Bar:** Setup > Settings

**Front Panel / Application Toolbar:** Settings

The control panel tabs for the Occupied Bandwidth display are shown in the following table.

| Settings tab                                  | Description  |
|---|--|
| <a href="#">Freq &amp; RBW (see page 226)</a> | Allows you to specify the Center Frequency, Step size and RBW.   |
| <a href="#">Parameters (see page 174)</a>     | Allows you to specify the Occupied BW % Power, x dB level, Measurement BW, enable averaging and the Max Hold function. |
| <a href="#">Scale (see page 230)</a>          | Allows you to set the vertical and horizontal scale parameters.  |
| <a href="#">Prefs (see page 231)</a>          | The Prefs tab enables you to set characteristics of the measurement display.   |

## Parameters Tab

The Parameters tab enables you to specify parameters that control the Occupied BW measurement.

The screenshot shows the 'Parameters' tab selected. The settings are as follows:

- Occupied BW % Power: 99.0 %
- x dB level: -26.0 dB
- Measurement BW: 10.00 MHz
- Max Hold spectral data: ☒
- Average results: ☒
- Count: 10

| Setting                | Description  |
|------------------------|--|
| Occupied BW % Power    | Specifies the proportion of power within the occupied bandwidth (referenced against the total power in the measurement bandwidth). |
| x dB level             | x dB level defines the x dB BW level search threshold.   |
| Measurement BW         | Specifies the frequency range used by the measurement.   |
| Max Hold spectral data | Enables the Max Hold function.   |
| Average results        | Enables/disables results averaging. Note that this is not an averaging of the trace, but an averaging of the results.              |
| Count                  | Specifies the number of results averaged to calculate the Occupied BW. Range: 2 to 10,000.   |

### x dB Level

The x dB level determines the x dB BW. The instrument analyzes the spectrum trace to locate the frequencies at which the level is x dB down from the peak level calculated over the measurement bandwidth. The frequency difference between the upper and lower crossing thresholds is the x dB BW.

Range: -80.0 to -1.0 dB; Resolution: 0.1%; Inc/dec small: 0.1%, large: 1%; Default: -26 dB

### Max Hold Spectral Data

Max Hold displays the maximum value in the acquisition record for each display point. Each new trace display point is compared to the previous maximum value and the greater value is retained for display and subsequent comparisons.

## Spurious Display

To show the Spurious display:

1. Click the **Displays** button or select **Setup > Displays**.
2. From the Measurements box, select **RF Measurements**.


- Double-click the **Spurious** icon in the **Available displays** box. This adds the Spurious icon to the **Selected displays** box (and removes it from the Available displays box). Alternatively, you can click the Spurious icon and then click the Add button to select Spectrum for display.
- Click the **OK** button.








### Elements of the Spurious Display



| Item | Display element         | Description  |
|------|-------------------------|--|
| 1    | dB/ div                 | Sets the vertical scale value. The maximum value is 20.00 dB/division.   |
| 2    | Vert Position           | Sets the top of graph value. This is only a visual control for panning the graph. The Reference Level is adjusted in the Toolbar and the Ampl control panel. By default, Vert Position = Ref Level.  |
| 3    | Check mark indicator    | Indicates the display for which the acquisition hardware is optimized.<br><br><b>NOTE.</b> When Best for multiple windows is selected in the Amplitude control panel's RF & IF Optimization control, none of the measurement displays shows a checkmark, as there is not a single optimized measurement. |
| 4    | Pass / Fail readout     | Indicates whether one or more spurs have exceeded a limit specified on the <b>Settings &gt; Limits</b> tab.  |
| 5    | Ref:                    | If enabled, displays the power reference level.  |
| 6    | Trace function          | Indicates the trace processing method. If Average or Max Hold is selected, this readout displays the number of traces being processed (averaged or compared for the Max Hold value).   |
| 7    | Clear                   | Restarts multi-trace functions (Average, Max Hold).  |
| 8    | Stop                    | Adjusts the graph stop frequency. This setting affects only visual scaling, not measurement parameters.  |
| 9    | Green bar               | Indicates the range that is selected on the Settings > Ranges and Settings > Limits tabs.  |
| 10   | Spur table              | Displays spur measurements. See the following table for details.   |
| 11   | Start                   | Adjusts the graph start frequency. This setting affects only visual scaling, not measurement parameters.   |
| 12   | Autoscale               | Adjusts the Vertical and Horizontal scaling to display the entire trace on screen. Selects Multi-range display mode.   |
| 13   | Bottom-of-graph readout | Indicates the amplitude at the bottom of the graph. This value changes with the dB/div and Vertical Position settings.   |

### Elements of the Spurious Graph Display



| Item | Display element   | Description   |
|------|---|---|
| 1    |  | The green line or shading indicates an Abs (absolute) limit. The absolute limit is enabled when the mask is set to Abs, Abs & Rel, or Abs OR Rel. Note that absolute and relative amplitude masks can overlap.      |
| 2    |  | The gray-shaded area indicates a portion of the spectrum where no measurements are being taken.   |
| 3    |  | The cyan line or shading indicates a Relative limit. The relative limit is enabled when the mask is set to Rel, Abs & Rel, or Abs OR Rel. Note that absolute and relative amplitude masks can overlap.              |
| 4    | abs rel   | Reminder that green-shaded areas highlight absolute limits and cyan-shaded areas highlight relative limits.   |
| 5    |  | This indicates the selected spur, when it is in violation of the limits. In Run mode, this is the peak spur violation. The selected marker is highlighted in the Spur table below the graph with a blue background. |
| 6    |  | A spur marker. Indicates a spur that does not exceed the mask settings. See the <a href="#">Settings &gt; Ranges and Limits tab (see page 181)</a> .  |
| 7    |  | A violation marker. Indicates a spur that exceeds the mask settings. See the <a href="#">Settings &gt; Ranges and Limits tab (see page 181)</a> .   |
| --   |  | Indicates the selected spur when it is not in violation of the limits. The selected marker is highlighted in the Spur table below the graph with a blue background.   |
| 8    | 1, 2, 3...  | A Spur number. The number indicates the row in the spur table that corresponds to the spur. The instrument can display up to 999 spurs.   |
| 9    | A, B, C, D...   | Identifies the enabled ranges.  |
| 10   | Green bar   | Indicates the range selected on the <a href="#">Settings &gt; Ranges and Limits (see page 181)</a> tab.   |
| 11   | Ref   | Indicates the location of the power reference. See <a href="#">Settings &gt; Reference (see page 179)</a> .   |

### Elements of the Spur Table

| Spur | Range | Freq Abs     | Freq Rel | Ampl Abs   | Ampl Rel  | Limit Abs  | Limit Rel |
|------|-------|--------------|----------|------------|-----------|------------|-----------|
| 1    | B     | 1.751145 GHz | – Hz     | -5.91 dBm  | 4.09 dB   | -15.00 dBm | -15.00 dB |
| 2    | E     | 2.053875 GHz | – Hz     | -8.99 dBm  | 1.01 dB   | -15.00 dBm | -10.00 dB |
| 3    | D     | 1.903950 GHz | – Hz     | -10.48 dBm | -0.48 dB  | -50.00 dBm | -57.00 dB |
| 4    | A     | 1.748826 GHz | – Hz     | -39.57 dBm | -29.57 dB | -57.00 dBm | -30.00 dB |
| 5    | D     | 1.899062 GHz | – Hz     | -40.31 dBm | -30.31 dB | -50.00 dBm | -57.00 dB |

| Column          | Description   |
|-----------------|---|
| Spur            | A number that identifies a spur in the graph area. The instrument can display a maximum of 999 spurs.   |
| Range           | The letter representing the frequency range where the spur is located.  |
| Freq Abs        | The absolute frequency at which the spur occurs.  |
| Freq Rel        | The relative frequency at which the spur occurs. The relative frequency is the difference between the absolute frequency (absolute) and the carrier frequency.  |
| Ampl Abs        | The absolute amplitude of the spur.   |
| Ampl Rel        | The relative amplitude at which the spur occurs. The relative amplitude is the difference between the absolute amplitude and the carrier amplitude. See the Settings > Limits tab.  |
| Lim Abs         | The value of the absolute amplitude limit at the spur frequency. This value can vary even with small spur frequency changes if the start and stop limit values are different. See the <a href="#">Settings &gt; Ranges and Limits tab (see page 181)</a> .          |
| Lim Rel         | The relative amplitude limit at which the spur occurs. This value can vary even with small spur frequency changes if the start and stop relative amplitude limit values are different. See the <a href="#">Settings &gt; Ranges and Limits tab (see page 181)</a> . |
| Blue background | The cell in the Spur column with a blue background identifies the selected spur.  |
| Red background  | Cells in the Spur column with a red background identify violations.<br>Cells in the results area with a red background identify the measurement that exceeded a limit.  |

### Rearranging the Columns in the Spur Table

You can rearrange the order of the columns in the Spur Table. To move a column, click on the column heading and drag it to the desired position.

### Sorting the Rows in the Spur Table

You can sort the rows in the Spur table by clicking on the column heading. For example, if you click on the Freq Abs heading, the results in the Spur table will be sorted by frequency. If you click on Range, the rows will be sorted by range. Clicking a second time on the same heading reverses the order.

[Changing the Spurious Display Settings \(see page 178\)](#)

## Spurious Display Settings

**Menu Bar:** Setup > Settings

**Front Panel:** Settings

**Application Toolbar:** 


The control panel tabs for the Spurious display are shown in the following table.

| Settings tab                              | Description   |
|---|---|
| <a href="#">Parameters (see page 179)</a> | Specifies whether the graph displays one range or multiple ranges. Specifies whether all spurs are shown or only spurs over specified limits.                                     |
| <a href="#">Reference (see page 179)</a>  | Specifies the Power Reference level.  |
| <a href="#">Ranges (see page 181)</a>     | Specifies start and stop frequencies of ranges and which ranges to take measurements in. Sets parameters that define a spur. Review, save and load the Range table from this tab. |
| <a href="#">Limits (see page 181)</a>     | Specifies Pass/Fail limit parameters.   |
| <a href="#">Trace (see page 227)</a>      | Specifies the trace Function.   |
| <a href="#">Scale (see page 230)</a>      | Specifies the vertical and horizontal scale settings.   |
| <a href="#">Prefs (see page 231)</a>      | Specifies the appearance features of the graph area.  |

**Restore defaults.** Sets the Spurious parameters to their default values.

## Parameters Tab

The Parameters tab enables you to specify settings that control the Spurious measurement.



| Setting                 | Description   |
|-------------------------|---|
| <b>Frequency Ranges</b> |   |
| Multi                   | Enables the instrument to display all enabled ranges in the graph.  |
| Single                  | Limits the graph to the display of only one range. The displayed range is set to the range that contains the selected spur.               |
| <b>List Spurs</b>       |   |
| All spurs               | Displays any spur that exceeds the Threshold and Excursion values set on the Ranges tab.  |
| Over Limit              | Displays only spurs that exceed both the Threshold and Excursion values set on the Ranges tab and the limits specified on the Limits tab. |

## Reference Tab

The Reference tab enables you to specify the Power Reference parameter. The Power reference parameter is used to calculate relative values in the Spur table and to calculate relative amplitude limits.

Parameters Reference Ranges and Limits Trace Scale Prefs

Power reference: Manual level Carrier Manual level No reference

Reference power level: -10.00 dBm

| Setting         | Description  |
|-----------------|--|
| Power Reference | Specifies the power reference level used for relative values and limits.   |
| No reference    | Power level readings are calculated only for absolute values, no relative values are calculated.                         |
| Manual level    | The reference level for relative measurements is specified by the Reference power level setting.                         |
| Carrier         | The reference level for relative measurements is based on the power level calculated at the specified carrier frequency. |

### Setting the Power Reference Level to No Reference

Parameters Reference Ranges and Limits Trace Scale Prefs

Power reference: No reference

To set the power reference to No Reference:

- Select **No reference** from the **Power reference** drop-down list.

### Setting the Power Reference Level to Manual Level

Parameters Reference Ranges and Limits Trace Scale Prefs

Power reference: Manual level

Reference power level: -10.00 dBm

To set the power reference to manual level:

1. Select **Manual level** from the **Power reference** drop-down list. This displays the **Reference power level** entry box.
2. Specify the desired power level in the **Reference power level** entry box.

### Setting the Power Reference Level to Carrier

| Parameters       |              | Reference | Ranges and Limits | Trace                                    | Scale | Prefs |
|------------------|--------------|-----------|-------------------|--|-------|-------|
| Power reference: | Carrier      |           |                   |  |       |       |
| Frequency:       | 1.466000 GHz |           | Res BW:           | 85 kHz                                   |       |       |
| Channel width:   | 8.500 MHz    |           |                   | <input checked="" type="checkbox"/> Auto |       |       |
| Threshold:       | -10.00 dBm   |           | Integration BW:   | 2 MHz                                    |       |       |
|                  |              |           |                   | Detection: +Peak                         |       |       |

To set the power reference to Carrier:

1. Select **Carrier** from the **Power reference** drop-down list.
2. To set the Carrier frequency, specify the necessary values in the **Frequency** and **Channel width** entry boxes .
3. Specify the **Threshold** level.
4. Specify the **Integration BW**. The Integration BW is the bandwidth over which the carrier is measured.
5. If necessary, specify the Resolution Bandwidth by changing the value in the **Res BW** entry box. Auto is the recommended setting.
6. Set the **Detection** method. Select **+Peak**, **Avg (VRMS)**, or **Avg (of logs)**.

---

**NOTE.** If the power level measured over the Integration BW about the carrier frequency is below the threshold level, the instrument concludes that no carrier is present and no relative measurements will be taken. Absolute measurements will still be taken.

---

## Ranges and Limits Tab

Use the Ranges and Limits tab to specify the parameters that control the Spurious measurement. From the Ranges and Limits tab, you specify the start and stop frequencies for ranges, whether or not a range is enabled, the parameters that specify what constitutes a spur, measurement filter shape and bandwidth, and the detection method used for processing signals, as well as the pass/fail parameters for Limit Testing.

| Parameters Reference Ranges and Limits Trace Scale Prefs |    |            |           |              |          |      |          |          |        |  |
|--|----|------------|-----------|--------------|----------|------|----------|----------|--------|--|
|  | On | Start (Hz) | Stop (Hz) | Filter Shape | BW (Hz)  | Auto | Detector | VBW (Hz) | VBW On |  |
| ▶ A  | X  | 1.492500G  | 1.507500G | RBW          | 100.000k | X    | +Peak    | 100k     |        |  |
| B  | X  | 1.512500G  | 1.537500G | RBW          | 200.000k | X    | +Peak    | 200k     |        |  |
| C  |    | 0.000000   | 0.000000  | RBW          | 1.000M   | X    | +Peak    | 1M       |        |  |
| D  |    | 0.000000   | 0.000000  | RBW          | 1.000M   | X    | +Peak    | 1M       |        |  |
| E  |    | 0.000000   | 0.000000  | RBW          | 1.000M   | X    | +Peak    | 1M       |        |  |

**Setting****Description**

|              |  |
|--------------|--|
| Expand       | Displays the Ranges and Limits Table in a new, resizeable window.  |
| Reset layout | You can reorder columns in the Ranges and Limits Table by dragging the columns to a new position. Clicking Reset Layout returns the column order to the factory default order. |
| Load         | Click to load a saved Ranges and Limits table from a file.   |
| Save         | Click to save the current Ranges and Limits table to a file.   |

To specify the ranges and limits for the Spurious measurement:

- Edit the values in the Ranges and Limits table.

You can edit the Ranges and Limits table in two ways:

- Edit values directly from the Ranges and Limits tab. You will need to scroll the table to access all settings.
- Click the **Expand** button. This displays the Ranges and Limits Table in a new window that can be sized to the fullscreen.

| Ranges and Limits Table |    |            |           |              |          |      |          |          |        |  |
|-------------------------|----|------------|-----------|--------------|----------|------|----------|----------|--------|--|
|                         | On | Start (Hz) | Stop (Hz) | Filter Shape | BW (Hz)  | Auto | Detector | VBW (Hz) | VBW On |  |
| ▶ A                     | X  | 1.492500G  | 1.507500G | RBW          | 100.000k | X    | +Peak    | 100k     |        |  |
| B                       | X  | 1.512500G  | 1.537500G | RBW          | 200.000k | X    | +Peak    | 200k     |        |  |
| C                       |    | 0.000000   | 0.000000  | RBW          | 1.000M   | X    | +Peak    | 1M       |        |  |
| D                       |    | 0.000000   | 0.000000  | RBW          | 1.000M   | X    | +Peak    | 1M       |        |  |
| E                       |    | 0.000000   | 0.000000  | RBW          | 1.000M   | X    | +Peak    | 1M       |        |  |
| F                       |    | 0.000000   | 0.000000  | RBW          | 1.000M   | X    | +Peak    | 1M       |        |  |
| G                       |    | 0.000000   | 0.000000  | RBW          | 1.000M   | X    | +Peak    | 1M       |        |  |
| H                       |    | 0.000000   | 0.000000  | RBW          | 1.000M   | X    | +Peak    | 1M       |        |  |
| I                       |    | 0.000000   | 0.000000  | RBW          | 1.000M   | X    | +Peak    | 1M       |        |  |
| J                       |    | 0.000000   | 0.000000  | RBW          | 1.000M   | X    | +Peak    | 1M       |        |  |
| K                       |    | 0.000000   | 0.000000  | RBW          | 1.000M   | X    | +Peak    | 1M       |        |  |
| L                       |    | 0.000000   | 0.000000  | RBW          | 1.000M   | X    | +Peak    | 1M       |        |  |
| M                       |    | 0.000000   | 0.000000  | RBW          | 1.000M   | X    | +Peak    | 1M       |        |  |
| N                       |    | 0.000000   | 0.000000  | RBW          | 1.000M   | X    | +Peak    | 1M       |        |  |
| O                       |    | 0.000000   | 0.000000  | RBW          | 1.000M   | X    | +Peak    | 1M       |        |  |
| P                       |    | 0.000000   | 0.000000  | RBW          | 1.000M   | X    | +Peak    | 1M       |        |  |
| Q                       |    | 0.000000   | 0.000000  | RBW          | 1.000M   | X    | +Peak    | 1M       |        |  |
| R                       |    | 0.000000   | 0.000000  | RBW          | 1.000M   | X    | +Peak    | 1M       |        |  |
| S                       |    | 0.000000   | 0.000000  | RBW          | 1.000M   | X    | +Peak    | 1M       |        |  |
| T                       |    | 0.000000   | 0.000000  | RBW          | 1.000M   | X    | +Peak    | 1M       |        |  |

The following table describes the parameters that are set in the Ranges and Limits Table.

### Ranges and Limits Table Settings

| Setting                  | Description  |
|--------------------------|--|
| On                       | Specifies whether or not measurements are taken in the specified range.  |
| Start (Hz)               | Readout of the start frequency for the selected range.   |
| Stop (Hz)                | Readout of the stop frequency for the selected range.  |
| Filter shape             | Specifies the filter shape used for the Spurious measurement.  |
| BW (Hz)                  | Specifies the bandwidth used for the selected filter shape.  |
| Auto                     | Sets the BW automatically. If CISPR is selected for Filter shape, this control is disabled.  |
| Detector                 | Sets the processing method used for compressing excess intermediate data into the desired number of trace points.  |
| VBW (Hz)                 | Adjusts the VBW (Video Bandwidth) value. VBW Maximum: RBW current value; VBW Minimum: 1/10,000 RBW setting. Disabled when Filter shape is set to CISPR.  |
| VBW On                   | Specifies whether the VBW filter is applied.   |
| Thrshld (dBm)            | Threshold specifies the level that must be exceeded for a signal peak to be recognized as a spur. A signal peak must also exceed the Excursion setting to be considered a spur.  |
| Excrsn (dB)              | Excursion specifies the peak to peak magnitude that must be exceeded for a signal peak to be recognized as a spur. A signal transition must also exceed the Threshold setting to be considered a spur.                           |
| Mask                     | Selects the type of limits used for Pass/Fail testing. Spurs that exceed the mask settings are considered violations. The available choices are shown below.   |
| Abs                      | Spurs that exceed the Absolute limits settings are identified as violations.   |
| Rel                      | Spurs that exceed the Relative limits settings are identified as violations.   |
| Abs & Rel                | Spurs that exceed both the Absolute and Relative limits settings are identified as violations.   |
| Abs OR Rel               | Spurs that exceed either the Absolute or Relative limits settings are identified as violations.  |
| Off                      | Turns off limit testing for the selecting range. Measurements are still taken in the range, but violations are not flagged. Signal peaks that exceed the Threshold and Excursions values will still be identified as spurs.      |
| (Abs Start and Abs Stop) | Absolute amplitude limits are not based on the measured carrier amplitude or manual reference. Start and stop values can be different.   |
| Abs Start (dBm)          | Abs Start specifies the limit at the start frequency of the range.   |
| Abs Stop (dBm)           | Abs Stop specifies the limit at the stop frequency of the range.   |
| Abs Same                 | Sets the limit at the start and stop frequencies to the value set for the start frequency.   |
| (Rel Start and Rel Stop) | Relative amplitude limits are calculated from the Power Reference. If the Power Reference is set to the Carrier level, the relative amplitude limits will change with the carrier level. Start and stop values can be different. |
| Rel Start (dB)           | Specifies the limit at the start frequency of the range.   |
| Rel Stop (dB)            | Specifies the limit at the stop frequency of the range.  |
| Rel Same                 | Sets the limit at the stop frequency to the value set for the start frequency.   |
| Save As                  | Click to save the current Ranges and Limits table to a file.   |
| Load                     | Click to load a saved Ranges and Limits table from a file.   |
| Done                     | Save changes and close the Ranges and Limits Table window.   |

## Changing the Range Start and Stop Frequencies

To change the range start and stop frequencies, edit the Start and Stop frequencies in the Ranges and Limits Table:

1. Click the **Expand** button on the **Settings > Range and Limits** tab. This displays the Ranges and Limits Table.
2. Click the **On** box for a range to take measurements in the range.
3. Click on the Start or Stop frequency setting to change it. Type in a number for the frequency and a letter as a multiplier. You can use k, m, or g to set the frequency multiplier.
4. Click **Done** to save your changes.

## Specifying Spur Requirements

A spur is a signal peak that exceeds both the Threshold and Excursion settings in the Ranges and Limits table. The Threshold and Excursion settings are absolute values; they are not calculated relative to a reference. The Threshold and Excursion settings are specific to the selected range. If you want to use different settings for spurs in different ranges, you have to set the values separately for each range. The Excursion control is used to avoid interpreting a single spur as multiple narrower spurs by requiring the amplitude to drop by the Excursion amount between spurs. Raising the Threshold value means that fewer, larger signals will be identified as spurs.

To specify the spur requirements for a range:

1. Click the **Expand** button on the **Range and Limits** tab. This displays the Ranges and Limits Table.
2. Select the **Range** (A–T) for which you want to specify the spur requirements.
3. Set the **Thrshld** value.
4. Set the **Excrsn** value.

## Setting Limits

Use the Limits settings in the Ranges and Limits Table to specify the pass/fail parameters for the Spurious measurement. When the Mask setting is set to any value except off, the instrument identifies any signal peak that exceeds the specified limits as a violation and displays Fail on the screen. If no signal peak exceeds the limits, the instrument displays Pass on the screen.

## Performing Pass/Fail Limit Testing

To set limits:

1. Click the **Expand** button on the **Range and Limits** tab. This displays the Ranges and Limits Table.
2. Adjust the Start and Stop frequencies as required for each range you want to test.
3. For ranges that you wish to test, verify that the **On** box is checked.

4. For each range you wish to test, select the desired **Mask** type from the drop-down list.
5. Set the limits as desired in the **Abs Start**, **Abs Stop**, **Rel Start**, and **Rel Stop** boxes. The values you can edit depend on the mask type you select.
6. Verify that the **Thrshld** and **Excrsn** values are set as required. These are the values that define a spur.
7. Click **Done** to save your changes and close the Range and Limits Table window.
8. Click **Run** to begin testing.


## CCDF Display

The CCDF (Complementary Cumulative Distribution Function) is a statistical characterization that plots power level on the x-axis and probability on the y-axis of a graph. Each point on the CCDF curve shows what percentage of time a signal spends at or above a given power level. The power level is expressed in dB relative to the average signal power level. The CCDF Display also shows the Average power level and the Peak/Average ratio.

To show the CCDF display:

1. Press the **Displays** button or select **Setup > Displays**. This shows the **Select Displays** dialog box.
2. From the **Measurements** box, select **RF Measurements**.
3. Double-click the **CCDF** icon in the **Available Displays** box. This adds the CCDF icon to the **Selected displays** box.
4. Click the **OK** button.

### Elements of the CCDF Display



| Item | Display element    | Description   |
|------|--------------------|---|
| 1    | 100%               | The top of the graph represents the 100% probability that the signal will be at or above the average signal level.                                    |
| 2    | Power level select | Adjust to display the value of the CCDF curve at a specific power level. The selected power level is indicated by a small triangle on the CCDF curve. |
| 3    | CCDF value         | Readout of the value of the CCDF curve at the selected power level.   |
| 4    | Avg                | Readout of the Average power level and the CCDF value at the Average power level.   |
| 5    | Readout Table      | Readout of the CCDF curve at six points.  |
| 6    | Pk / Avg           | Readout of the Peak to Average power ratio.   |
| 7    | Measurement time   | Readout of the length of time used for the measurement. Displays --- s when the Measurement time is set to continuous.                                |

[Changing the CCDF Display Settings \(see page 186\)](#)

## CCDF Settings

**Menu Bar:** Setup > Settings

**Front Panel:** Settings

**Application Toolbar:** 


The control panel tabs for the CCDF Display are shown in the following table.

| Settings tab                              | Description   |
|---|---|
| <a href="#">Parameters (see page 187)</a> | Specifies the time to be measured.                                      |
| <a href="#">Traces (see page 227)</a>     | Select the trace to be measured and select reference trace for display. |
| <a href="#">Prefs (see page 231)</a>      | Specifies whether on not certain display elements are shown.            |

**Restore defaults.** Sets the CCDF parameters to their default values.

## Parameters Tab

The Parameters tab enables you to specify several parameters that control signal acquisition for the CCDF Display.



| Setting    | Description  |
|------------|--|
| Single     | When Single is selected, the CCDF measurement is based on the Analysis Time parameters set on Analysis control panel (Setup > Analysis > Analysis Time), shared by all displays. In Single, CCDF is time-correlated with the other open displays.                          |
| Total Time | When Total Time is selected, you can adjust its value. The value set here does not affect the Analysis Length value set in the Analysis control panel. If the Total Time value is longer than the Analysis Length, CCDF uses multiple acquisitions to collect signal data. |
| Continuous | When Continuous is selected, the CCDF measurement combines each new record (Actual Analysis Length) into the existing result. It does not erase and start over until user presses STOP or CLEAR.   |
| Span       | Specifies the range of frequencies used for analysis.  |

### Single

To have statistically meaningful results, CCDF works best when calculated on at least 1 ms of data. Depending on the Analysis Time setting, the CCDF display may present error messages:

- If the Actual Analysis Time is greater than 1 ms, the CCDF Display will calculate the trace without any problems.
- If the Actual Analysis time is less than 1 ms, the CCDF Display will measure the data as best it can, but if it receives fewer than 20 samples, the CCDF display will clear the trace and report "Not enough samples".


## Phase Noise Display

Phase Noise measures and graphs the power in a 1 Hz bandwidth referenced to the carrier power. This is repeated for many frequency points across a user-adjustable span.

To show the Phase Noise display:

1. Click the **Displays** button or select **Setup > Displays**.
2. From the Measurements box, select **RF Measurements**.
3. Double-click the **Phase Noise** icon in the **Available displays** box. This adds the Phase Noise icon to the **Selected displays** box (and removes it from the Available displays box). Alternatively, you can click the Phase Noise icon and then click the **Add** button to select Phase Noise for display, or drag the Phase Noise icon to the Selected Displays box.
4. Click the **OK** button.

### Elements of the Phase Noise Display



| Item | Display element                        | Description  |
|------|--|--|
| 1    | Check mark indicator                   | Indicates the display for which the acquisition hardware is optimized.<br><br><b>NOTE.</b> When Best for multiple windows is selected in the Amplitude control panel's RF & IF Optimization control, none of the measurement displays shows a checkmark, as there is not a single optimized measurement. |
| 2    | Vert Position (dBc/Hz)                 | Sets the top of graph value. This is only a visual control for panning the graph.  |
| 3    | Trace selector                         | Selects a trace.   |
| 4    | Show check box                         | Shows/hides the selected trace.  |
| 5    | Detection method / Average count       | Displays the Detector method used for the trace. If Average detection is enabled, the number of averaged traces appears here.  |
| 6    | Measurement Marker readout             | Readout of the value of the phase noise at the selected offset. The Phase Noise measurement marker does not affect the results reported in the readouts or table below the graph, nor is it associated with markers in any other display.  |
| 7    | Clear                                  | Restarts multi-trace function (Average).   |
| 8    | Horizontal stop                        | Readout of the Right Horizontal frequency value. This setting affects only visual scale, not measurement bandwidth. (See <a href="#">Settings &gt; Scale (see page 142)</a> )  |
| 9    | Measurement marker                     | Indicates the position of the measurement point.   |
| 10   | Integration frequency range indicators | The gray-shaded areas denote the frequencies outside the integration frequency range. (See <a href="#">Settings &gt; Frequency (see page 557)</a> )  |
| 11   | Phase Noise Table                      | Displays the phase noise values at different offsets.  |
| 12   | Horizontal start                       | Readout of the Left Horizontal frequency value. Affects only visual scaling, not measurement bandwidth. (See <a href="#">Settings &gt; Scale (see page 142)</a> )  |
| 13   | Measurement results readouts           | Display of the different measured values.  |

### Elements of the Measurement Results Readout / Phase Noise Table

|                  |           |         |               |               |
|------------------|-----------|---------|---------------|---------------|
| Carrier power:   | -6.93 dBm | Offset  | Trace 1       | Trace 2       |
| Freq Error:      | -392.8 Hz | 10 Hz   | -63.93 dBc/Hz | -63.46 dBc/Hz |
| RMS Phase Noise: | 2.554 °   | 100 Hz  | -86.75 dBc/Hz | -84.91 dBc/Hz |
| Jitter:          | 2.902 ps  | 1 kHz   | -88.56 dBc/Hz | -83.80 dBc/Hz |
| Residual FM:     | 20.47 kHz | 10 kHz  | -73.18 dBc/Hz | -69.89 dBc/Hz |
|                  |           | 100 kHz | -95.98 dBc/Hz | -94.04 dBc/Hz |

| Readout         | Description  |
|-----------------|--|
| Carrier power   | The power level at the carrier frequency.  |
| Freq Error      | The measured carrier frequency error.  |
| RMS Phase Noise | The RMS value of the phase noise, measured over the Integration BW.  |
| Jitter          | The jitter in seconds, measured over the Integration BW. Includes the effects of random and coherent jitter.   |
| Residual FM     | The residual frequency modulation, measured over the Integration BW.   |
| Trace 1         | The phase noise on trace 1 at different frequency offset values. Note that if Trace 1 is set to +/- Peak detection, the value displayed in the table is the +Peak value. |
| Trace 2         | The phase noise on trace 2 at different frequency offset values. Note that if Trace 2 is set to +/- Peak detection, the value displayed in the table is the +Peak value. |

### [Changing the Phase Noise Settings \(see page 192\)](#)

### How Phase Noise Is Measured

If carrier tracking is on, the instrument measures phase noise relative to the highest-amplitude signal within about 20 MHz of the user-entered frequency. The difference between center frequency and the frequency of the measured carrier is reported as frequency error. If carrier tracking is not on, the instrument measures phase noise relative to the amplitude at the user-entered frequency.

The unit will tune to a number of offsets from the carrier frequency, and calculate the spectrum for the frequency range about each offset using a resolution bandwidth appropriate for the offset frequency. This is done to optimize speed while maintaining acceptable frequency resolution. A table of the values used for measurement frequency offset, resolution bandwidths, hardware IF filter bandwidths and RF attenuations is shown below.

| Measurement frequency offset range | Measured span | RBW use for spectrum calculation | Hardware IF filter | RF Attenuation                            |
|------------------------------------|---------------|----------------------------------|--------------------|---|
| 10 Hz to 100 Hz                    | 312.5 kHz     | 1 Hz                             | 40 MHz             | Auto from ref level                       |
| 100 Hz to 1 kHz                    | 312.5 kHz     | 2 Hz                             | 40 MHz             | Auto from ref level                       |
| 1 kHz to 10 kHz                    | 312.5 kHz     | 20 Hz                            | 40 MHz             | Auto from ref level                       |
| 10 kHz to 100 kHz                  | 3 MHz         | 200 Hz                           | 40 MHz             | Auto from ref level                       |
| 100 kHz to 1 MHz                   | 3 MHz         | 1 kHz                            | 40 MHz             | Auto from ref level                       |
| 1.0 MHz to 1.9 MHz                 | 300 kHz       | 1 kHz                            | 1 MHz              | Auto from ref level (-5 dB) <sup>1</sup>  |
| 1.9 MHz to 20.0 MHz                | 300 kHz       | 1 kHz                            | 1 MHz              | Auto from ref level (-20 dB) <sup>1</sup> |
| 20 MHz to 100 MHz                  | 20 MHz        | 10 kHz                           | 40 MHz             | Auto from ref level (-5 dB) <sup>1</sup>  |
| 100 MHz to 1000 MHz                | 20 MHz        | 100 kHz                          | 40 MHz             | Auto from ref level (-5 dB) <sup>1</sup>  |

<sup>1</sup> Attenuator adjustment when 'optimize dynamic range' is selected


After acquisition and calculation, the spectrum is corrected to display power in a 1 Hz bandwidth referenced to the carrier power (dBc/Hz). The correction algorithm does not discriminate between noise signals and coherent signals in making the dBc/Hz correction. Since all measured signals are corrected as

if they were noise, coherent signal measurement results can contain errors in level. This is a key difference between an SA-based measurement and a dedicated phase noise tester.

The largest difference between operation of a swept spectrum analyzer and a real-time signal analyzer when measuring phase noise is a result of the wide acquisition bandwidths found in the real-time signal analyzer.

The real-time signal analyzer can select between 1 MHz, 40 MHz and 110 MHz IF filters when performing acquisitions.

By contrast, a traditional spectrum analyzer uses an IF bandwidth that can range from a few kHz wide up to 10 MHz wide, depending upon the selected resolution bandwidth. When the IF bandwidth is small it is able to reject out-of-band carriers so they don't overload the final IF and digitizers in the instrument. This allows the signal analyzer to begin to reduce attenuation as the phase noise measurement tunes farther away from the carrier, improving dynamic range. The same thing is done in the real-time signal analyzer, but the 1 MHz final IF section in the real-time signal analyzer can't reject signal within about 1 MHz whereas a conventional spectrum analyzer can reject signals within a few kHz of the desired carrier. This is the reason for the shape of the phase noise plot seen in the following illustration.



In the previous illustration, the effect of choosing 'optimize dynamic range' is seen. In the offsets between 10 Hz and 1 MHz, the 40 MHz acquisition bandwidth is used, and the RF attenuator is set based on the reference level of the instrument to prevent overload of the RF/IF signal path.

At an offset of >1 MHz, a 1 MHz IF bandwidth is applied, 5 dB of RF attenuation is removed from the signal path and the noise 'floor' of the measurement is reduced. At 2 MHz offset from carrier, 20 dB of attenuation is removed with a resultant increase in dynamic range. Finally, at 20 MHz offsets, the 40 MHz filter is used again with 20 dB attenuation removed. In order to be able to use the 40 MHz BW filter at a

20 MHz offset, the measurement frequency is moved so that the carrier is 30 MHz from the measurement point, and well down the filter rejection skirt.

## Phase Noise Settings

**Menu Bar:** Setup > Settings

**Front Panel:** Settings

**Application Toolbar:** 

The control panel tabs for the Phase Noise display are shown in the following table.

| Settings tab                              | Description  |
|---|--|
| <a href="#">Frequency (see page 557)</a>  | Specify Measurement BW and Integration BW. Enable carrier frequency tracking and specify the Carrier threshold level.                                  |
| <a href="#">Parameters (see page 194)</a> | Specifies the Optimization method.   |
| <a href="#">Traces (see page 227)</a>     | Specifies trace parameters such as detection method and whether smoothing is enabled. Traces can also be saved and loaded for reference from this tab. |
| <a href="#">Scale (see page 230)</a>      | Specify vertical and horizontal scale settings.  |
| <a href="#">Prefs (see page 231)</a>      | Specify appearance features of the graph area.   |

**Restore defaults.** Sets the Phase Noise parameters to their default values.

## Frequency Tab

Use the Frequency tab to specify parameters used for taking the Phase Noise measurement.


The screenshot shows the 'Frequency' tab selected in a software interface. The 'Parameters' sub-tab is active. The interface includes the following controls:

- Measurement BW:** A text box containing '100.0 Hz'.
- Start Offset:** A text box containing '100.0 MHz'.
- Stop Offset:** A text box containing '100.0 MHz'.
- Integration BW:** A text box containing '100.0 Hz'.
- Track carrier frequency:** A checked checkbox.
- Carrier threshold:** A text box containing '-20.00 dBm'.

| Setting                 | Description  |
|-------------------------|--|
| Measurement BW          | This sets the frequency range for the phase noise measurement range. The instrument's Frequency control is used to set the "carrier frequency of interest". The widest available start/stop range is from 10 Hz greater than the carrier frequency ("Stop Offset") to 1 GHz above the carrier ("Stop Offset"). Very small start offsets will take significant time to measure as very narrow resolution bandwidths are used in to achieve small-offset measurements. Large-offset stop frequencies cause small effects on measurement time, but will increase the displayed scale of the phase noise plot. |
| Start Offset            | Sets the starting point for the Phase Noise plot. Offset is from the carrier frequency. The minimum start offset is 10 Hz. The maximum start offset is 100 MHz. The setting is limited to powers of ten (1, 10, 100...).   |
| Stop Offset             | Sets the stopping point for the Phase Noise plot. Offset is from the carrier frequency. The minimum start offset is 100 Hz. The maximum stop offset is 1 GHz. The Stop offset value must be at least 10X the Start offset value. The setting is limited to powers of ten (1, 10, 100...).  |
| Integration BW          | Integration bandwidth sets the frequency range over which integrated results for phase noise (labeled "RMS Phase Noise") and jitter are calculated. Integration start and stop cannot exceed the measurement bandwidth. If you set the integration bandwidth greater than the measurement, the error message 'Setup error: Integration BW exceeds Measurement BW' will appear.   |
| Start Offset            | Specifies the Integration BW's start offset from the carrier frequency.  |
| Stop Offset             | Specifies the Integration BW's stop offset from the carrier frequency.   |
| Track carrier frequency | The instrument finds the carrier signal and measures its frequency at least once each time the Phase Noise measurement is taken and reports the error between the Freq setting and the measured frequency of the carrier. When Track carrier frequency is enabled, the measurement adjusts its tuning to counteract the carrier error or drift. When carrier tracking is disabled, the measurement is made using the Freq setting specified by the user.   |
| Carrier threshold       | The minimum power level required for the signal to be recognized as a carrier. If no signal is found above this level within 20 MHz of the Freq setting, the Phase Noise measurement will fail.  |

## Parameters Tab

The Parameters tab enables you to specify acquisition settings that control the Phase Noise measurement.



**Average.** Specifies whether averaging is used or not. This setting affects integrated numeric results and the trace display. The range for this setting is 2 to 100. The default setting is disabled.

## Settling Time Measurement Overview

The Settling Time measurement (Option 12) is used to measure frequency and phase settling time of frequency-agile oscillators and subsystems. Automated measurements can reduce user-to-user measurement variations, improve repeatability and measurement confidence, and save time in gathering results.

### Measurement Definitions

**Settling time.** The time measured from a reference point to when the signal of interest settles to within user-defined tolerance.

**Settled frequency.** The final reference frequency, determined automatically or manually. The tolerance band is centered about this settled frequency.

**Settled phase.** The final reference phase, determined automatically or manually. The tolerance band is centered about this settled phase.


### Mask Testing

Mask testing is a tool to automatically determine the pass/fail state of the Device-Under-Test. The mask is defined over three or fewer user-defined time zones, with an independent limit band for each time zone. If the mask is violated, the area of violation is highlighted in red on screen, and a Fail message is displayed. The pass/fail status of the mask test can be queried over GPIB for rapid results. To configure Mask testing, see the [Mask tab \(see page 210\)](#).


The instrument's Search function can watch the Settling Time measurement for either pass or fail results, and perform actions such as stopping or saving data when the defined condition occurs. See [Mask Test Limits Settings \(see page 547\)](#) for details on configuring Mask Test.

## Frequency Settling Time Theory of Operation

The following paragraphs describe how a frequency and phase settling measurement is made.



Frequency settling time flow diagram



Phase settling time flow diagram

Acquisition hardware collects I and Q samples at the measurement frequency. Sample Rate is based on the measurement bandwidth setting.


The instantaneous phase is computed as:

$$Ph_{(n)} = atan\left(\frac{Q_{(n)}}{I_{(n)}}\right)$$

The instantaneous frequency is the derivative of the phase:

$$Fr_{(n)} = (Ph_{(n-1)} - Ph_{(n+1)}) \times \left(\frac{F_{sampling}}{(2 \times 2PI)}\right)$$

Data smoothing is applied (if selected by the user). This filtering method uses  $\pm n$  adjacent samples to produce an average value that is used to replace the values of the original  $2n$  samples. An example is shown below.




If averaging is turned on, the smoothed data array is averaged with data arrays of previous acquisitions on a point-by-point basis. Smoothing has the advantage of reducing noise-like variations in the measurement, but carries the risk of masking real measurement data. The amount of smoothing applied must be determined by evaluating the waveform you see for artifacts that result from the process.

Next, Settled Frequency is calculated. Settled frequency is the average frequency determined by looking back in time from the point set by the Measurement Length control over a minimum of 256 samples. This value is used to determine the frequency level about which to center the Tolerance range. The measurement then searches backward through the data until it finds the first point that exceeds the tolerance range. A maximum of 20 million acquisition data samples can be processed by the measurement.

Settling Time is measured as the time from the start reference to the point where the frequency remains within the tolerance band. The start reference can be any of three points:

- The trigger point
- The beginning of the analysis period. This result can only be computed when the instrument can find a previous settled state within the measurement period
- The point at the end of the previous frequency hop where the frequency first leaves the tolerance band. This result can only be computed when the instrument can find a previous settled state within the measurement period.



### Settling time start reference points

All measurements are now complete and the numeric settling time values can be sent to the display. However, no result traces have been produced yet. Detection is used to reduce the measured points (which can number up to 10 million) to a number more easily processed by the display processor. +/- Peak detection is always used, preserving the maximum positive and negative excursions for display. Reducing the number of points in the result reduces the time resolution of the markers. If averaging is on, the measurement results are computed on the average of the undecimated results.

### Phase Settling Time Overview

Phase settling time is determined, as described previously, with the addition of tilt correction.


The instrument estimates tilt (or slope of the phase) using linear least square fitting over (at least) 256 samples of the phase vs. time waveform, looking backwards from the point set by the Measurement Length. The instrument adjusts the phase waveform so that its phase is of zero value at the point set by the Measurement Length.

## Settling Time Displays

The Frequency Settling Time measurement is used to measure the frequency settling time of frequency-agile oscillators and subsystems. The Phase Settling Time measurement is used to measure the phase settling time of phase-agile oscillators and subsystems. Automated measurements can reduce user-to-user measurement variations, improving repeatability and measurement confidence and saving time in gathering results. To measure phase settling, a phase reference is required. This means that the measurement tool must be frequency- referenced to the device under test, or must act as the frequency reference to the device under test. To lock the reference to an external source, go to Setup > Configure In/Out > Frequency Reference and choose **External**. Values from 1 MHz to 25 MHz in 1 MHz steps can be used as an external reference.


To show a Settling Time display:

1. Click the **Displays** button or select **Setup > Displays**.
2. From the Measurements box, select **RF Measurements**.
3. Double-click the **Freq Settling** icon or the **Phase Settling** icon in the **Available displays** box. This adds the selected display to the **Selected displays** box (and removes it from the Available displays box).
4. Click the **OK** button.



0070\_035

Frequency settling time display



Phase settling time display

## Elements of the Display

| Item | Display element                   | Description  |
|------|-----------------------------------|--|
| 1    | Top of graph                      | Sets the frequency/phase that appears at the top of the graph. However, note that the top of graph setting, vertical scale setting (Settings > Scale tab), and Vertical Position settings interact. Adjusting this value changes the frequency/phase at the top of the graph by adjusting the scale setting.<br><b>Phase Settling Time:</b> Sets |
| 2    | Vertical Position                 | Sets the frequency/phase value at the vertical center of the graph.  |
| 3    | Autoscale                         | Adjusts the Vertical and Horizontal scaling to display the entire trace on screen.   |
| 4    | Horizontal Position               | Sets the horizontal position of the trace on the graph.  |
| 5    | Measurement readouts              | Displays measurement results. See <i>Measurement Readout Text Color</i> below.   |
| 6    | Signal transition start indicator | A blue vertical line that indicates the starting point of the Settling Time measurement that is based on the signal transition away from the previous settled state, if a previous settled state is found within the measurement period.   |
| 7    | Trigger point indicator           | An orange line and "T" (which appears below the graph) that indicates the location of the trigger point in time.   |
| 8    | +/- Tolerance indicators          | A pair of blue horizontal lines that show the tolerance range.   |
| 9    | Settled time indicator            | A blue vertical bar that shows the point on the trace at which the measurement has determined the signal frequency or phase to be settled.   |
| 10   | Measurement length indicator      | A magenta vertical line that specifies the end of the allowed measurement period. This is a control, not a result.   |
| 11   | Scale                             | Adjusts the horizontal (time) scale of the graph.  |
| 12   | Clear                             | Restarts multi-trace functions (Avg, Hold) and erases the trace.   |
| 13   | Function                          | Displays the Detection and Function selections for the selected trace. (Detection is always +/-Peak for Settling Time measurements.)   |
| 14   | Show                              | Controls whether the selected Trace is visible or not. When trace is Off, the box is not checked.  |
| 15   | Trace                             | Selects a trace. Touching here displays a menu that shows the available traces and whether they are enabled or not. If you select a trace that is not currently enabled, it will be enabled.   |

## Measurement Readout Text Color


The measurement readouts can appear in either black text or red text. Measurement results that appear in black text indicate:

- the signal is settled AND
- the signal is settled longer than the minimum settled duration AND
- the settling time is valid OR the settling time from trigger is valid

If any of the conditions listed are not met, the measurement results text is red.

### Elements of the Frequency Settling Time Graph


The following illustration shows the various indicators that appear in the Frequency Settling Time graph.



**NOTE.** The blue vertical bar that indicates the end of the previous settled state may not always appear. For frequency settling time, this can happen because the starting frequency is outside the measurement bandwidth.

### Elements of the Phase Settling Time Graph

The following illustration shows the various indicators that appear in the Phase Settling Time graph. For most phase settling time measurements, there is not a previous settled state within the measurement period, making it impossible to measure the settling time from the previous settled value. However, the settling time from trigger is typically available.



**NOTE.** The blue vertical bar indicating the measurement start time may not appear. Zero phase is referenced to the measured hop, as the phase will not typically have a settled value during the previous hop.


### Measuring Settling Time

1. Select the **Displays** button.
2. Select **RF Measurements** from the **Measurements** box.
3. Double-click the **Freq Settling** icon or **Phase Settling** icon in the Available displays box. Click **OK** to complete your selection.
4. Select **General Signal Viewing** from the **Measurements** box.
5. Double-click the **Time Overview** icon in the Available displays box. Click **OK** to complete your selection.
6. Press the front-panel **Freq** button and use the front-panel keypad or knob to adjust the frequency to that of the signal you want to measure.

7. Select the Trig button.
8. In the Trigger control panel, select Triggered. Configure the trigger settings as needed to achieve a triggered signal.

For the next step in measuring settling time, you must set the starting point of the measurement. To do this, you will use the Time Overview display.

9. Select the Time Overview display.
10. At the left edge of the graph, drag across the graph to adjust the analysis offset. Adjust the Analysis Offset to locate the start of the settling time measurement just ahead of the signal transition where you want to measure the settling time. Any portion of the signal in the gray area will be excluded from the settling time measurement.



#### Setting the starting point of the settling time measurement

Even if the signal transition is difficult to see on the Time Overview display, you can set the starting point based on the location of the Trigger indicator below the graph. Or open the Frequency vs. Time graph to see the frequency transition.

11. Click in the Settling Time display to select it.
12. Select the **Settings** button to display the **Settling Time Settings** control panel.
13. On the **Define** tab, adjust the [Meas BW \(see page 207\)](#) to improve the signal-to-noise ratio as appropriate. Make sure the bandwidth covers the settled frequency, particularly when the instrument Measurement Frequency is not locked to the spectrum Center Frequency (Analysis control panel > Frequency tab).
14. Set the [Target reference \(see page 208\)](#) to Auto or Meas freq as appropriate. If you select Meas freq, also set Offset to fine-tune the measurement frequency if necessary.
15. Set the [Tolerance \(see page 208\)](#) as required for your measurement.


If your signal must be settled for a period of time before it is considered settled, you can specify a minimum settled duration for the settling time measurement.

16. Select the Time Params tab. Adjust the [Min settled duration \(see page 209\)](#) as required.


To get the measurement made in the correct location on the signal, you must adjust the measurement length. You do this by dragging the magenta line in the Settling Time graph.

17. Drag the measurement length indicator (magenta line) to the correct location on the signal. Choose a point at which the signal appears to be well settled. The measurement will then find the exact time at which the signal actually settled to within the specified tolerance.

Note the difference in the settling time in the two illustrations that follow. In both examples, the Target reference is set to Auto. The only difference is the measurement length, as shown by the magenta line.



Frequency settling time display before setting the measurement length



Frequency settling time display after setting the measurement length

# Settling Time Settings

**Menu Bar:** Setup > Settings

**Front Panel:** Settings

**Application Toolbar:** 

The control panel tabs for the Frequency and Phase Settling Time displays are shown in the following table.

| Settings tab                               | Description  |
|--|--|
| <a href="#">Define (see page 207)</a>      | Sets the measurement parameters that characterize the settling time measurement.   |
| <a href="#">Time Params (see page 209)</a> | Sets measurement end-time and minimum settled duration parameters.   |
| <a href="#">Mask (see page 210)</a>        | Enable or disables mask testing and sets the parameters that specify the three zones used for mask testing.  |
| <a href="#">Trace (see page 210)</a>       | Specifies trace display characteristics and which traces are displayed.  |
| <a href="#">Scale (see page 230)</a>       | Sets vertical and horizontal scale and position parameters.  |
| <a href="#">Prefs (see page 231)</a>       | Specifies whether or not to show certain display elements, the maximum number of points in the exported trace, and the displayed precision of the settling time measurement. |

# Settling Time Displays Shared Measurement Settings

The control panel tabs in this section are shared by the Frequency and Phase Settling Time displays. These control panel tabs are not shared with other RF Measurement displays. With few exceptions, each display shares values for each of the controls on these tabs. For example, Vertical Scale values are independent for each of the displays (degrees for Phase Settling Time and Hertz for Frequency Settling Time).

### Common controls for settling time measurement displays

| Settings tab                               | Description  |
|--|--|
| <a href="#">Define (see page 207)</a>      | Specifies the parameters that characterize the settling time measurement.                        |
| <a href="#">Time Params (see page 209)</a> | Specifies the Measurement length and minimum settled duration for the settling time measurement. |
| <a href="#">Mask (see page 210)</a>        | Specifies the parameters used for Mask testing.  |
| <a href="#">Trace (see page 210)</a>       | Specify which traces to show and how they are computed.  |
| <a href="#">Scale (see page 213)</a>       | Specifies the vertical and horizontal scale settings.  |
| <a href="#">Prefs (see page 214)</a>       | Specifies display parameters of the Settling Time displays.                                      |

## Define Tab for Settling Time Displays

The Define tab for the Frequency Settling Time measurement controls several parameters that affect how the measurement is made.

The screenshot shows the 'Define' tab selected in a software interface. The tab bar includes 'Define', 'Time Params', 'Mask', 'Trace', 'Scale', and 'Prefs'. The main area contains the following controls:

- Meas Freq:** A text box containing '2.44700 GHz'.
- Target reference:** A dropdown menu with 'Auto' selected.
- Offset:** A text box containing '0.0000 Hz'.
- Meas BW:** A text box containing '40.00 MHz'.
- actual:** A text box containing '40.00 MHz'.
- Tolerance (+/-):** A text box containing '500.0 kHz'.

### Meas Freq

Select the frequency of the signal to be measured. Measurement Frequency can be selected as the target frequency (see *Target Reference*).

**NOTE.** If Lock Center Frequency of Spectrum Display to Measurement Frequency is unchecked, then Measurement Frequency is independent of Center Frequency) Range: Entire input frequency of the instrument

### Meas BW

This control limits the bandwidth of the Settling Time measurement. You use the measurement bandwidth setting to improve the signal-to-noise ratio of the measurement, resulting in lower measurement uncertainty. However, if the measurement bandwidth is set too low, resulting in fewer samples per second, it can reduce the number of points within the measurement length below the 256 stable samples required, thus causing the analysis to fail. The measurement bandwidth is 100 Hz to 110 MHz (Option 110) or 100 Hz to 40 MHz (Standard). To see the effect of measurement bandwidth on measurement uncertainty, see the *Specifications and Performance Verification* manual. This manual is supplied on the Documents CD or can be downloaded from [www.tek.com/manuals](http://www.tek.com/manuals).

## Target Reference

The target reference is a calculated or user-specified value of final settled frequency within the measurement period. When set to **Auto**, the instrument determines this value by averaging at least the last 256 samples at the end of the measurement period. When set to **Meas freq**, the Target Reference is set to the Measurement Frequency of the instrument. You can enter a manual offset from the target frequency to fine-tune the Target Reference when **Meas freq** is selected.

## Tolerance(+/-)

The Tolerance is the frequency or phase range within which the signal must remain to be considered settled. This is set as a  $\pm$  value by the user. In the frequency settling time measurement, the tolerance band is centered about the target value. In the phase settling time measurement, the tolerance band is centered about the settled phase value.

The target frequency depends on the Target reference value setting:

- If Target Reference is set to **AUTO**, then the target value is the settled value (the frequency at which the signal is considered settled).
- If Target Reference is set to **MeasFreq**, then the Target value = **Meas Frequency** + **Offset**.

## Offset

Offset is used for fine-tuning the Target Reference when **Meas freq** is selected. The value for Offset is entered manually.

## Time Params Tab for Settling Time Displays

The Time Params tab is used to set the Measurement length and minimum settled duration parameters.

The screenshot shows a software interface with a tabbed menu at the top containing 'Define', 'Time Params', 'Mask', 'Trace', 'Scale', and 'Prefs'. The 'Time Params' tab is selected. Below the tabs, there are two input fields. The first is labeled 'Measurement length:' with a value of '1.000 ms' and a description '(from Analysis Offset to End of Measurement Marker)'. Below this field, the text 'actual: 938.4 us' is displayed. The second input field is labeled 'Min settled duration:' with a value of '0.0000 s' and a description '(from Settled Point to End of Measurement Marker)'.

### Measurement Length

Measurement length locates the end of the measurement period for settling time. This value is set independently of the analysis period for other measurements in the analyzer. This allows other measurements to display multiple frequency hops during an analysis, while focusing the settling time analysis on a single hop. You can set the measurement length either by entering a value in the entry box, or by dragging the magenta bar to the desired location in the graph. Traces and the display to the right of the measurement time are shown in grey to indicate that this section of the waveform is not being used in the measurement. If your signal does not settle to within your tolerance band, make sure that the measurement time ends during the settled portion of the waveform. The beginning of the settling time measurement is set with the analysis offset control, the same as other measurements in the instrument. Use analysis offset to exclude earlier hops from the measurement. See [Analysis Time \(see page 555\)](#) for setting the analysis offset control.

If there is more than one phase or frequency hop in the signal to be analyzed, use the Measurement Length control to instruct the analyzer which hop to measure. First, adjust the main Analysis Offset control to a point in time slightly earlier than the transition you wish to measure. Do this in the Time Overview display or the Analysis Settings control panel. Then adjust the Measurement Length so the end of the Settling Time measurement falls within the apparent settled period after the signal transition. The reason it is important for the user to set this control is that the Settling Time measurements start at this point and look backwards in time (towards the signal transition) to find the latest trace point that is not inside the tolerance band.

---

**NOTE.** *If your signal does not settle to within your tolerance bands, make sure that the measurement time ends during the settled portion of the waveform.*

---

### Min Settled Duration

The minimum settled duration is the amount of time the signal must remain within the tolerance band to be considered settled. Some user specifications require that the signal be settled for a minimum period. If this is not required in your application, set this value to the default of 0 seconds.

## Mask Tab for Settling Time Displays

The Mask tab is used to specify the parameters used for Mask testing.

| Mask time reference  |          | Time     | Limit (+/-) |
|--|----------|----------|-------------|
| <input type="checkbox"/> Enable mask test<br><input checked="" type="radio"/> Trigger<br><input type="radio"/> Signal transition | Start 1: | 0.0000 s | 500.0 kHz   |
|  | Start 2: | 100.0 us | 500.0 kHz   |
|  | Start 3: | 1.000 ms | 500.0 kHz   |
|  | Stop:    | 13.18 ms |             |

### Enable Mask Test

Enables or disables mask testing.

### Start 1, 2, 3

Sets the starting point for the time zone of interest. If you don't want to use all three zones, set the Start Time for the unneeded zones to a time after the end of the Measurement Length.

### Stop

Sets the end point for the final time zone of interest.

### Limit (+/-)

Sets the mask violation limits for each time zone.

### Mask Time Reference

When mask testing is enabled, the Mask time reference selection specifies the starting point of the mask. You can choose to start the mask test at the trigger point or at the point on the trace where the signal began its transition from its previous settled frequency or phase level to the level you intend to measure.

This capability is useful because some applications do not have a trigger signal available. For those situations where a trigger signal is available, the specification for the device-under-test may require that measurements be referenced to that signal.


## Trace Tab for Settling Time Displays

The Traces Tab allows you to set the display characteristics of displayed traces.

|  |             |      |   |       |       |
|--|-------------|------|---|-------|-------|
| Define   | Time Params | Mask | Trace   | Scale | Prefs |
| <input type="checkbox"/> Smooth: (points) 2<br><input type="checkbox"/> Average: (count) 2 |             |      | Trace: Trace 1 <input checked="" type="checkbox"/> Show <input type="checkbox"/> Freeze<br><input type="checkbox"/> Show recalled trace <input type="text"/> ...<br><input type="button" value="Save Trace As..."/> |       |       |

## Smooth

Smooth is a low-pass filter function that uses  $n$  points in the trace to determine the smoothed value. For example, in the case of  $n=3$ , 3 consecutive points are averaged to determine the result of the smoothed output. This is very useful for single-shot signals that contain significant noise where multiple trace averages cannot be used. However, smoothing does have limits, and the user should visually determine whether smoothing is removing noise from the measurement, or if it is affecting the signal. The following illustration shows the effect smoothing has on the displayed trace.



The range for the Smooth function is 2–1000. The setting resolution is 1. The knob increments this value by 1, and the arrow keys increment it in steps of 1, 2, 5, 10. When smoothing is enabled, it applies to both traces. Smooth is turned off by default.

## Average

This sets the number of acquisitions to be averaged together to produce the result. Averaging reduces random variations but maintains the correct waveshape for repeating signal aberrations. To use averaging, the signal must be repetitive. For example, the tuning direction must always be same for each measurement

to be averaged. Signals that vary in tuning direction will tend to average positive-going effects with negative-going effects and the settling time characteristics will cancel out.

The range for the Average function is 2–10000. The resolution is 1. The knob increments this value by 1, and the arrow keys increment it in steps of 1, 2, 5, 10. When averaging is enabled, it applies to both traces. Average is turned off by default.

## Trace

Selects a trace for display. Choices are Trace 1 and 2. Either trace can be a recalled trace.

## Saving Traces

To save a trace for later analysis:


1. Select the **Save Trace As** button. This displays the Save As dialog box.
2. Navigate to the desired folder or use the default.
3. Type a name for the saved trace and click **Save**.

## Recalling Traces

You can recall a previously saved trace for analysis or comparison to a live trace.

To select a trace for recall:

1. Select the trace into which the recalled trace will be loaded, from the Trace drop-down list.
2. Check the **Show** check box.
3. Click the ... button to display the Open dialog box.



4. Navigate to the desired file and click **Open**.
5. Check the **Show Recalled Trace** check box.
6. Verify that the trace's **Show** check box is selected (either on this tab or next to the drop-down list located at the top-left corner of the graph).


## Scale Tab for Settling Time Displays

The Scale tab allows you to change the vertical and horizontal scale settings. Changing the scale settings changes how the trace appears on the display but does not change control settings such as Measurement Frequency. In effect, these controls operate like pan and zoom controls.

| Setting       | Description  |
|---------------|--|
| Vertical      |  |
| Scale         | Changes the vertical scale or span of the graph.   |
| Position      | Sets the frequency (or phase) at the vertical center of the graph.   |
| Relative Freq | Sets the vertical (position) display readouts to show frequency values relative to the center frequency or as absolute values. |
| Autoscale     | Resets the Scale and Position to center the trace in the graph.  |
| Horizontal    |  |
| Scale         | Sets the range of time shown in the graph.   |
| Position      | Allows you to pan the graph left and right.  |
| Autoscale     | Resets Scale and Position to the automatically determined settings.  |

# Prefs Tab for Settling Time Displays

The Prefs tab allows you to change display parameters of the Settling Time displays.



The following table explains the controls that can appear on the Prefs tab.


| Setting  | Description   |
|--|---|
| Max trace points                               | Sets the number of trace points used for marker measurements and for results export.  |
| Show graticule                                 | Select to display or hide the graticule.  |
| Show Marker readout in graph (selected marker) | Shows or hides the readout for the selected marker in the graph area.   |
| Settling Time decimal places                   | Specifies the precision of the settling time measurement readout that appears below the graph. This setting does not affect the precision of internal calculations. |

## SEM Display

The SEM (Spectrum Emission Mask) display measures out-of-band emissions immediately outside the channel bandwidth that result from the modulation process and non-linearity in the transmitter. The Spectrum Emission Mask display allows you to define up to six pair of masks, or offsets, that specify the limits for out-of-band emissions and performs a pass/fail tests of the signal against the defined offsets and limits.

To show the Spectrum Emission Mask display:

1. Select **Displays** or select **Setup > Displays**.
2. In the **Select Displays** dialog, select **RF Measurements** or **WLAN Analysis** in the Measurements box.
3. In the Available displays box, double-click the **SEM** icon or select the icon and click **Add**. The SEM icon will appear in the Selected displays box and will no longer appear under Available displays.
4. Click **OK** to show the Spectrum Emission Mask display.



## Elements of the Display

| Item | Display element         | Description  |
|------|-------------------------|--|
| 1    | Check mark indicator    | Indicates the display for which the acquisition hardware is optimized.<br><br><b>NOTE.</b> When <i>Best for multiple windows</i> is selected in the <i>Amplitude control</i> panel's <i>RF &amp; IF Optimization control</i> , none of the measurement displays shows a checkmark, as there is not a single optimized measurement.                           |
| 2    | Top-of-graph setting    | Sets the level that appears at the top of the graph.   |
| 3    | dB/div                  | Sets the vertical scale value. The maximum value is 20.00 dB/division.   |
| 4    | Bottom-of-graph readout | Indicates the amplitude at the bottom of the graph. This value changes with the dB/div and Vertical Position settings.   |
| 5    | Autoscale               | Properly offsets the peak of the display from the top of the screen.<br>If the screen has been enlarged through a manual vertical (dB/div) setting entry, the view is not vertically rescaled when Autoscale is used.  |
| 6    | CF / Pos                | Default function is CF - center frequency (equivalent to the Measurement Freq setting). If Horizontal scaling has been manually adjusted in Setup > Settings > Scale, then Pos replaces CF as the setting at the bottom-left corner of the screen. Adjusting Pos shifts the trace left or right in the graph, but does not change the measurement frequency. |
| 7    | Span / Scale            | Default function is Span - frequency difference between the left edge of the display and the right edge. If Horizontal scaling has been manually adjusted in Settings > Scale, then Scale will replace Span as the setting at the bottom-right corner of the screen.   |
| 8    | Clear                   | Erases the current results from the display.   |
| 9    | Function readout        | Readout of the Function selection for the Reference channel (Setup > Processing > Function).<br>Spurious   |
| 10   | Detection readout       | Readout of the Detection selection for the Reference channel (Setup > Processing > Detection).   |
| 11   | Pass / Fail readout     | Readout indicating whether any part of the trace has exceeded the any of the limits defined in the Offsets & Limits table (Setup > Settings > Offsets & Limits Table).   |

## Pass / Fail Readout

This readout displays either **Pass** or **Fail XX%**. When the signal exceeds the mask in any offset, it is considered a mask violation, and results in **Fail XX%** being displayed. If there are no mask violations, **Pass** is displayed.

**How the fail percentage is determined.** Each offset has two sides, one lower and one upper. For example, Offset A has two sides: AL (lower side) and AU (upper side). In the SEM display, there is a maximum of six offsets (A-F) and each offset can have a maximum of two sides (if Both is selected for Side in the Offsets & Limits table). This results in a total of six offsets multiplied by two sides per offset, which results in 12 offsets. Mask violations can occur in any of these offsets. The failure percentage is calculated as follows:

Fail percentage = Number of offsets with mask violations x 100 / Total number of enabled offsets

For example: Assume there are three enabled offsets: A, B, C, and all these offsets have both lower and upper sides. Thus, the total number of enabled offsets is equal to six. Suppose there are mask violations in the offsets AL and CU, then the number of offsets having mask violations is equal to 2. Therefore, Fail percentage =  $2 * 100 / 6 = 33\%$ .

## Results Display

The Results Display shows the measurement results for the SEM display.

|                  |     |        |             |            |                |                 |               |                 |               |                 |
|------------------|-----|--------|-------------|------------|----------------|-----------------|---------------|-----------------|---------------|-----------------|
| Ref: Total Power | Row | Offset | Start Freq  | Stop Freq  | Peak Abs (dBm) | Margin Abs (dB) | Peak Rel (dB) | Margin Rel (dB) | @Freq         | Integ Abs (dBm) |
| -16.93 dBm       | 1   | EL     | -12.500 MHz | -8.000 MHz | -73.83         | -62.33          | -91.81        | --              | -8.010000 MHz | -68.44          |
|                  | 2   | DL     | -8.000 MHz  | -4.000 MHz | -73.17         | -61.67          | -91.16        | --              | -7.460000 MHz | -67.61          |
|                  | 3   | CL     | -4.000 MHz  | -3.515 MHz | -87.63         | -63.13          | -105.62       | --              | -3.965444 MHz | -77.11          |
|                  | 4   | BL     | -3.515 MHz  | -2.715 MHz | -87.77         | -75.17          | -105.76       | --              | -2.722000 MHz | -75.14          |
|                  | 5   | AL     | -2.715 MHz  | -2.515 MHz | -70.18         | -57.68          | -88.16        | --              | -2.515000 MHz | -70.74          |
|                  | 6   | AU     | 2.515 MHz   | 2.715 MHz  | -70.73         | -58.23          | -87.67        | --              | 2.515000 MHz  | -70.54          |
|                  | 7   | BU     | 2.715 MHz   | 3.515 MHz  | -87.66         | -67.24          | -104.59       | --              | 3.243000 MHz  | -74.72          |
|                  | 8   | CU     | 3.515 MHz   | 4.000 MHz  | -88.01         | -63.51          | -106          | --              | 3.976356 MHz  | -77.04          |

### Readout

### Description

|            |  |
|------------|--|
| Ref:       | A two-part readout to the left of the table that displays information about the Reference Channel. The first part of the readout displays the Measurement Type selected on the Parameters tab. The second part of the readout displays the measurement result for the Reference channel. |
| Row        | A sequence number for ordering the offsets. Click to organize by row number, resort rows from high to low or low to high.  |
| Offset     | The Offset name. A combination of the offset (A-F) and the location (U for upper and L for Lower).   |
| Start Freq | The start frequency of the offset (relative to the Measurement Freq).  |
| Stop Freq  | The stop frequency of the offset (relative to the Measurement Freq).   |
| Peak Abs   | Absolute power peak power at worst case margin of the offset   |
| Margin Abs | The worst case difference between the trace and the absolute limit line. Note that an offset can have different start and stop levels, thus, the margin can vary if the offset start and stop levels are different.  |
| Peak Rel   | Relative peak power of worst case margin of the offset. Relative with respect to the reference power level.  |
| Margin Rel | The worst-case difference between the trace and the relative limit line. Because an offset can have different start and stop levels, the margin can vary if the offset start and stop levels are different.  |
| @Freq      | The frequency at which the worst margin occurs.  |
| Integ Abs  | The absolute integrated power of the offset.   |
| Integ Rel  | The relative integrated power of the offset (with respect to the reference power).   |

**Reordering results in the results table.** You can sort the results in the Results table by clicking the column heading containing the values you want to sort on. You can sort results for the following columns: Row, Offset, Start Frequency, Stop Frequency, Peak Abs, and Peak Rel.

**Rearranging columns in the results table.** You can rearrange the order of the columns and their size. To rearrange columns, select the column heading you want to move and drag it to the new position; the new location is indicated by a dark blue bar that appears between column headings.

You can resize the columns by selecting the divider between columns and dragging it to a new position.

## Spectrum Emission Mask Settings

### Front Panel: Settings

#### Application Toolbar:

The settings for the Spectrum Emission Mask display are visible when SEM is the selected display. To display the SEM Settings, select **Setup > Settings**.

| Settings tab  | Description  |
|---|--|
| <a href="#">Parameters Tab (see page 219)</a>             | Specifies several characteristics that control how the measurement is made.                                |
| <a href="#">Processing Tab (see page 220)</a>             | Specifies settings for detection on the Reference channel and the offsets. Specifies the function setting. |
| <a href="#">Ref Channel Tab (see page 221)</a>            | Specifies how the measurements on the reference channel are performed.                                     |
| <a href="#">Offsets &amp; Limits Table (see page 221)</a> | Specifies characteristics of offsets and mask limits.  |
| <a href="#">Scale Tab (see page 224)</a>                  | Specifies the vertical and horizontal scale settings.  |
| <a href="#">Prefs Tab (see page 225)</a>                  | Specifies the appearance features of the graph area and the maximum trace points.                          |


## Parameters Tab - SEM

The Parameters tab specifies several characteristics that control how the Spectrum Emission Mask measurements are made.

| Setting                 | Description   |
|-------------------------|---|
| Meas Freq               | Specify the frequency of the signal to be measured.   |
| Step                    | Sets the increment size when changing the Frequency using the knob or mouse wheel.  |
| Auto                    | When Auto is enabled, the step size is adjusted automatically based on the span setting.  |
| Real-Time               | When Real-Time mode is enabled, the entire SEM span is measured using a real-time/contiguous acquisition. Not all described parameters are available in Real-Time mode.   |
| Measurement Type        | Specifies the type of measurement for the reference channel; used as a reference for the offsets.   |
| Total Power             | Sets the reference to the integrated power of the reference channel within the reference's integration bandwidth.   |
| PSD                     | Sets the reference to the mean of the power spectral density (dBm/Hz) of the reference channel. The trace and the absolute and relative limits line are also in dBm/Hz  |
| Peak                    | Sets the reference to the Peak power of the reference channel.  |
| Offset definition       | Defines the relative position of the start/stop frequency of an offset. Choose the offset definition based on the standard.   |
| Ref center to OS center | Specifies that the start/stop frequencies are defined from the center frequency of the reference channel to the center of the filter BW.  |
| Ref center to OS edge   | Specifies that the start/stop frequencies are defined from the center frequency of the reference channel to the center of the (inner) edge of the filter BW of the offset.  |
| Ref Edge to OS center   | Specifies that the start/stop frequencies are defined from the edge of the reference channel to the center of the filter BW of the offset.  |
| Ref Edge to OS edge     | Specifies that the start/stop frequencies are defined from the edge of the reference channel to the (inner) edge of the filter BW of the offset.  |
| Filter Shape            | Specifies the shape of the filter determined by the window that is applied to the data record, in the spectrum analysis, to reduce spectral leakage. 3GPP specifies a Gaussian window shape be applied to the reference channel measurements. |
| Gaussian                | This filter shape provides optimal localization in the frequency domain.  |
| Rectangular             | This filter shape provides the best frequency, worst magnitude resolution. This is essentially the same as no window.   |

## Processing Tab - SEM

The Processing tab controls the Detection settings for the Reference Channel and Offsets, as well as selecting the Function.



### Settings tab

### Description

#### Detection (in each sweep)

##### Ref Channel

Specifies the Detection method used for the Reference Channel. Detection is used to produce the desired measurement result (peak or average) and to reduce the results of a measurement to the desired number of trace points.

##### Avg (VRMS)

For each sweep, each point of the trace is the result of determining the RMS voltage value for the last 'n counts' of the collected traces for the same point. When 'n count' has not been reached, partial averaging results are displayed.

##### +Peak

Selects the +Peak detection method. With this method, the highest value is selected from the results to be compressed into a trace point.

##### Offsets

Specifies the detection method used for the offsets.

##### Avg (VRMS)

Selects the Average Vrms detection method. With this method, each point on the trace in each offset is the result of determining the RMS Voltage value for all of the results values it includes.

##### +Peak

Selects the +Peak detection method. With this method, the highest value in each offset is selected from the results to be compressed into a trace point.

#### Function (across multiple sweeps)

##### Function

Selects the trace processing method. Available settings are: Normal, Avg (VRMS), and Hold.

##### Normal

When a new trace has been computed, it replaces the previous trace.

##### Max Hold

With each sweep, each trace point in the new trace is compared to the point's value in the old trace and the greater value is retained for display and subsequent comparisons.

##### Avg (VRMS)

For each sweep, each point on the trace is the result of determining the RMS Voltage value for all of the collected traces' values for the same point.

## Ref Channel Tab

The Ref Channel tab controls how the measurement of the Reference channel is performed.

| Settings tab    | Description   |
|-----------------|---|
| Power Reference | The value used to calculate relative measurements.  |
| Auto            | When Auto is unchecked, you can enter a value for the reference power, and the measured reference power is not used or displayed  |
| RBW             | Sets the RBW for the Reference Channel.   |
| VBW             | Enables/disables the Video Bandwidth filter. VBW is used in traditional swept analyzers to reduce the effect of noise on the displayed signal. The VBW algorithm in the analyzer emulates the VBW filters of traditional swept analyzers. When the checkbox next to VBW is not checked, the VBW filter is not applied. SEM standards specify what value of VBW should be applied, if any. |
| Channel width   | Specifies the width of the reference channel.   |
| Integration BW  | Specifies the integration bandwidth used to compute the total power in the reference channel.   |
| Channel Filter  | Specifies the measurement filter used in the Reference Channel. Choices are None and Root-raised Cosine.  |
| Filter param    | Specifies the alpha value (or roll-off factor) of the Root-raised cosine filter. Present only when Channel Filter is set to Root-raised Cosine.   |
| Chip rate       | Specifies the chip rate used for the Root-raised Cosine filter. Present only when Channel Filter is set to Root-raised Cosine.  |

## Offsets & Limits Table Tab - SEM

The Offsets & Limits Table tab is used to specify parameters that define Offsets and masks.

|     | On | Start (Hz)  | Stop (Hz)   | Side | RBW (Hz) | n (RBWs) | Filter BW (n x RBW) | VBW (Hz) | VBW On | Mask |
|-----|----|-------------|-------------|------|----------|----------|---------------------|----------|--------|------|
| ▶ A | ✓  | 9.000000M   | 11.000000M  | Both | 100.000k | 1        | 100.000k            | 30.000k  | ✓      | Rel  |
| B   | ✓  | 11.000000M  | 20.000000M  | Both | 100.000k | 1        | 100.000k            | 30.000k  | ✓      | Rel  |
| C   | ✓  | 20.000000M  | 30.000000M  | Both | 100.000k | 1        | 100.000k            | 30.000k  | ✓      | Rel  |
| D   | ✓  | 30.000000M  | 50.000000M  | Both | 100.000k | 1        | 100.000k            | 30.000k  | ✓      | Rel  |
| E   |    | 50.000000M  | 100.000000M | Both | 100.000k | 1        | 100.000k            | 30.000k  | ✓      | Rel  |
| F   |    | 100.000000M | 150.000000M | Both | 100.000k | 1        | 100.000k            | 30.000k  | ✓      | Rel  |

Expanded display of Offsets &amp; Limits Table

| Setting              | Description   |
|----------------------|---|
| <b>Buttons</b>       |   |
| Expand               | Displays the Offsets & Limits Table in a new, resizable window.   |
| Reset Layout         | You can reorder columns in the Offsets & Limits Table by dragging the columns to a new position. Clicking Reset Layout returns the column order to the factory default order.   |
| Load                 | Click to load a saved Offsets & Limits table from a file.   |
| Save As              | Click to save the current Offsets & Limits table to a file.   |
| Done                 | When the table is expanded, click Done when you have finished editing the table to save your changes and close the expanded table display.  |
| <b>Table columns</b> |   |
| On                   | Specifies whether or not measurements are taken in the specified offset.  |
| Start (Hz)           | Start Frequency of the selected offset.   |
| Stop (Hz)            | Stop Frequency of the selected offset.  |
| Side                 | Specifies whether the specified range appears on both side of the carrier frequency or just one side (left or right).   |
| RBW (Hz)             | Specifies the RBW for the selected range.   |
| n (RBWS)             | An integer value that specifies how many times to multiply the RBW to set the Filter bandwidth.   |
| Filter BW (n x RBW)  | Displays the Filter BW. Filter BW is the equivalent BW of each point in the offset. When $n > 1$ , an integration technique is used to achieve the Filter BW using narrower RBWs. For most SEM measurements, this value is set to 1. The 3GPP standard suggests $n$ greater than 1. |
| VBW (Hz)             | Adjusts the VBW (Video Bandwidth) value. VBW Maximum: RBW current value; VBW Minimum: 1/10,000 RBW setting.   |
| VBW On               | Specifies whether the VBW filter is applied.  |
| Mask                 | Select the type of limits used for Pass/Fail testing. Signal excursions that exceed the mask settings are considered violations. The available choices are shown below.   |
| Abs Start (dBm)      | Specifies the mask level at the inner edge of the selected offset.  |
| Abs Stop (dBm)       | Specifies the mask level at the outer edge of the selected offset.  |
| Abs Same             | Specifies whether the value for the inner and outer edges of the offset must be the same value. When enabled, the Abs Stop cell is grayed-out and set to the value of the Abs Start cell.   |
| Rel Start (dBm)      | Specifies the mask level for the Relative mask at the inner edge of the selected offset.  |
| Rel Stop (dBm)       | Specifies the mask level for the Relative mask at the outer edge of the selected offset.  |
| Rel Same             | Specifies whether the inner and outer edges of the offset must be the same value. When enabled, the Rel Stop cell is grayed-out and its value is set to the value of the Rel Start cell.  |

## Scale Tab - SEM

Use the Scale tab to change the vertical and horizontal scale settings. Changing the scale settings changes how the trace appears on the display. In effect, these controls operate like pan and zoom controls.

| Setting     | Description   |
|-------------|---|
| Vertical    | Controls the vertical position and scale of the trace display.                |
| Scale       | Changes the vertical scale units.   |
| Position    | Adjusts the reference level away from top of the graph.                       |
| Autoscale   | Resets the scale of the vertical axis to contain the complete trace.          |
| Horizontal  | Controls the span of the trace display and position of the trace.             |
| Scale       | Specifies the frequency range displayed in the graph                          |
| Position    | Specifies the frequency shown at the center of the graph.                     |
| Autoscale   | Resets the scale of the horizontal axis to optimize the display of the trace. |
| Reset Scale | Resets all settings to their default values.                                  |

## Prefs Tab - SEM

The Prefs tab enables you to change parameters of the measurement display. The parameters available on the Prefs tab vary depending on the selected display, but include such items as enabling/disabling Marker Readout, switching the Graticule display on/off, and Marker Noise mode. Some parameters appear with most displays while others appear with only one display.

Parameters Processing Ref Channel Offsets & Limits Table Scale Prefs

Max Trace points: 801 (per range)

☒ Show graticule

☒ Show Marker readout in graph (selected marker)

Show limits: Line only

| Setting                      | Description  |
|------------------------------|--|
| Show graticule               | Controls the vertical position and scale of the trace display.   |
| Max Trace points (per range) | When the spectrum analysis produces more than the selected maximum number of points, the method specified in Detection control is used to decimate the result. This setting applies to both the Reference channel and offsets. |
| Show Marker readout in graph | Adjusts the reference level away from top of the graph.  |
| Show limits                  | Controls the appearance of mask limits in the graph.   |
| Shaded                       | Shows limits using a shaded area. Green shading is used to identify Absolute limits. Blue shading is used to indicate Relative limits.   |
| Line only                    | Shows limits using only a line. A green line is used to identify Absolute limits. A blue line is used to indicate Relative limits.   |
| None                         | No lines or shading are used to indicate limits in the graph. Violations of the mask are still identified by red shading.  |

## RF Measurements Shared Measurement Settings

The control panel tabs in this section are shared by multiple displays in the RF Measurement folder (Setup > Displays). With few exceptions, each display maintains separate values for each of the controls on these tabs. For example, Scale values are independent for each of the displays.

### Common controls for RF measurement displays

| Settings tab                                  | Description  |
|---|--|
| <a href="#">Freq &amp; RBW (see page 226)</a> | Allows you to specify the Center Frequency, Step size and RBW.   |
| <a href="#">Traces (see page 227)</a>         | Specifies trace parameters such as detection method and whether smoothing is enabled. Traces can also be saved and loaded for reference from this tab. |
| <a href="#">Scale (see page 230)</a>          | Specify vertical and horizontal scale settings.  |
| <a href="#">Prefs (see page 231)</a>          | Specify appearance features of the graph area.   |

## Freq & RBW Tab

The Freq & RBW tab specifies frequency parameters for some of the RF Measurements displays.

The screenshot shows the 'Freq & RBW' tab selected. It contains the following controls:

- Meas Freq:** 1.00000 GHz
- Step:** 2.0000 MHz
- RBW:** 30.000 kHz
- VBW:** (empty field)
- Auto checkboxes:** Checked for Step and RBW.

| Setting   | Description   |
|-----------|---|
| Meas Freq | Specifies the measurement frequency.  |
| Step      | The Step control sets the increment/decrement size for the knob adjustment of the center frequency. Arrow key adjustment of frequency is also affected (it is 10x the knob adjustment). If Auto is enabled, the analyzer will adjust the Step size as required. |
| RBW       | Select Auto or Manual. Adjusts the RBW for the entire measurement. This setting is Independent of the Spectrum view's RBW setting.  |
| VBW       | Adjusts the VBW (Video Bandwidth) value.  |

**NOTE.** While the measurement frequency and step settings on this control tab are shared between ACPR, MCPR and Occupied Bandwidth, the RBW setting is not shared. The RBW setting is unique for each measurement.

### VBW

The VBW setting enables/disables the Video Bandwidth filter. VBW is used in traditional swept analyzers to reduce the effect of noise on the displayed signal. The VBW algorithm in the analyzer emulates the VBW filters of traditional swept analyzers.

The maximum VBW value is the current RBW setting. The minimum VBW value is 1/10,000 of the RBW setting.

**NOTE.** If you are following a procedure that says to "set VBW to three times the RBW value or greater", it means that the test should be conducted with no VBW effects. In the analyzer, this condition is met by disabling the VBW function.

## Traces Tab

The Traces tab enables you to select traces for display. You can choose to display live traces and/or recalled traces. The [Trace tab \(see page 210\)](#) for Settling Time displays is described in the Settling Time controls section.

**NOTE.** Not all RF Measurements support multiple traces and/or save and recall of traces.

| Setting                      | Description  |
|------------------------------|--|
| Trace                        | Selects a trace for display. Choices are Trace 1, 2, and Gaussian (CCDF display only). Trace 1 and 2 can be recalled traces.   |
| Show                         | Shows / hides the selected trace.  |
| Freeze                       | Halts updates to the selected trace.   |
| Save Trace As                | Saves the selected trace to a file for later recall and analysis.  |
| Show recalled trace          | Used to select a saved trace for display.  |
| Detection (Phase Noise only) | Sets the Detector used for the trace. Only +/- Peak and Avg (VRMS) are available for the Phase Noise display. This setting is not available for saved traces. +/-Peak detection results in a trace with two Y values for each X location, with shading between.                                  |
| Function                     | Selects the trace processing method (Spurious display only). Possible settings are: Normal, Average (VRMS), Avg (of logs), and Max Hold.   |
| Smoothing (Phase Noise only) | When enabled, diminishes point-to-point trace variations. Smoothing +/- Peak traces treats + and - lines independently. The smoothing value adjusts how many points are in each filter set. Range: 1 - 20; Default: 5.   |
| Smooth: (points)             | Available for Phase and Frequency Settling Time displays only.   |
| Average: (count)             | Smooth is a low-pass filter function that uses n points in the trace to determine the smoothed value.<br>Average sets the number of acquisitions to be averaged together to produce the result.<br>See <a href="#">Trace Tab for Settling Time Displays (see page 210)</a> for complete details. |

## Detection

Trace Detection occurs when the trace is being decimated by the measurement. For example, if the maximum number of trace points is 100,000, and the selected analysis region is 200,000 samples, the measurement must decimate the 200,000 resulting trace points by 2 to prevent exceeding the 100,000 trace point limit. Since only one value can be selected for each trace point, an algorithm must be used to select (detect) the appropriate value to use.

The IQ samples in a data acquisition can be detected in a variety of ways. The number of IQ samples available to each trace point varies with both analysis length and trace length. For example, with Spectrum Length set to 'Auto' in the Analysis menu, the instrument analyzes just enough samples to produce one IQ sample pair per trace point. In this case, the detection method chosen has very little effect, as the +Peak, -Peak, Avg (VRMS) and Sample values are all equal. Changing the Spectrum Length causes the available detection methods to differ in value because they have a larger set of samples for the various detection methods to process.

The available detection methods (depending on the display) are:

- **+/-Peak** – Selects the highest and lowest values of all the samples contained in an acquisition interval.
- **Avg (VRMS)** [Average  $V_{RMS}$ ] – Each point on the trace is the result of determining the RMS Voltage value for all of the IQ samples available to the trace point. When displayed in either linear (Volts, Watts) or Log (dB, dBm), the correct RMS value results. When the averaging function is applied to a trace, the averaging is performed on the linear (Voltage) values, resulting in the correct average for RMS values.

## Trace Processing

The trace can be processed to display in different ways. The Function setting controls trace processing.

- **Max Hold** - Displays the maximum value in the trace record for each display point. Each new trace display point is compared to the previous maximum value and the greater value is retained for display and subsequent comparisons.
- **Normal** - Displays the trace record for each display point without additional processing.
- **Avg (VRMS)** [Average  $V_{RMS}$ ] – Each point on the trace is the result of determining the RMS Voltage value for all of the IQ samples available to the trace point. When displayed in either linear (Volts, Watts) or Log (dB, dBm), the correct RMS value results. When the averaging function is applied to a trace, the averaging is performed on the linear (Voltage) values, resulting in the correct average for RMS values.
- **Avg (of logs)** – This is a trace function used to emulate legacy spectrum analyzer results and for the specification of displayed average noise level. In older swept analyzers, a voltage envelope detector is used in the process of measuring signal level, and the result is then converted to Watts and then to dBm. Averaging is then applied to the resultant traces.

For CW signals, this method results in an accurate power measurement. However, with random noise and digitally modulated carriers, errors result from this 'average of logs' method. For random noise, the average of logs methods results in power levels -2.51 dB lower than that measured with a power meter, or with a spectrum analyzer that measures the rms value of a signal, and performs averaging on the calculated power in Watts and not dBm or other log-power units.

This detector should be used when following a measurement procedure that specifies it, or when checking the Displayed Averaged Noise Level (DANL) of the instrument. The 'average of logs' detection and trace function is used for DANL specification to provide similar results to other spectrum/signal analyzers for comparison purposes. Use of the Average of Logs method of measurement is not recommended for digitally modulated carriers, as power measurement errors will occur.

---

**NOTE.** *The Detection setting does not affect the trace until the spectrum length is longer than the Auto setting.*

---

## Saving Traces

To save a trace for later analysis:


1. Select the **Save Trace As** button. This displays the Save As dialog box.
2. Navigate to the desired folder or use the default.
3. Type a name for the saved trace and click **Save**.

## Recalling Traces

You can recall a previously saved trace for analysis or comparison to a live trace.

To select a trace for recall:


1. Select the trace into which the recalled trace will be loaded, from the Trace drop-down list.
2. Check the **Show** check box.
3. Click the ... button to display the Open dialog box.



1. Navigate to the desired file and click **Open**.
2. Check the **Show Recalled Trace** check box.
3. Verify that the trace's **Show** check box is selected (either on this tab or next to the drop-down list located at the top-left corner of the graph).

## Gaussian Trace

The Gaussian trace is a reference trace that you can compare your trace against. Because the Gaussian trace is for reference, there are no settings for the trace other than whether or not it is shown.



## Scale Tab

The Scale tab allows you to change the vertical and horizontal scale settings. Changing the scale settings changes how the trace appears on the display but does not change control settings such as Center Frequency.



| Setting             | Description  |
|---------------------|--|
| Vertical            | Controls the vertical position and scale of the trace display.   |
| Scale               | Changes the vertical scale units. This is only accessible when the vertical units are set to dBm.  |
| Offset              | Adjusts the Reference Level away from the top of the trace display.  |
| Autoscale           | Resets the scale of the vertical axis to contain the complete trace.   |
| Horizontal          | Controls the span of the trace display and position of the trace.  |
| Scale               | Allows you to, in effect, change the span.   |
| Offset              | Allows you to pan a zoomed trace without changing the Measurement Frequency. Offset is only enabled when the span, as specified by Freq / div, is less than the acquisition bandwidth. |
| Autoscale           | Resets the scale of the horizontal axis to contain the complete trace.   |
| Log (Spurious only) | Resets the display to show the frequency axis in a logarithmic scale.  |

## Prefs Tab

The Prefs tab enables you to change parameters of the measurement display. The parameters available on the Prefs tab vary depending on the selected display, but include such items as enabling/disabling Marker Readout, switching the Graticule display on/off, and Marker Noise mode. Some parameters appear with most displays while others appear with only one display.

The screenshot shows the 'Prefs' tab in a software interface. It contains several settings:

- Show:** A dropdown menu set to 'Start, Stop'.
- Trace points:** A dropdown menu set to '801'.
- ☒ **Show graticule**
- ☒ **Show Marker readout in graph (selected marker)**
- ☒ **Show trace legend**
- ☐ **Marker Noise mode (Requires dBm Ampl units and Average trace detect)**

| Setting  | Description   |
|--|---|
| Show graticule   | Shows or hides the graticule.   |
| Show Marker readout in graph (selected marker)   | Shows or hides the readout for the selected marker in the graph area.   |
| Max trace points   | In general, time-domain traces can include up to 100K points and frequency-domain traces can include up to 10K points. When the analysis produces a results array longer than the selected trace length, the trace is decimated (using the method specified with the Detection control) to be equal to or less than the Max trace points setting (except when Max trace points is set to Never Decimate). This decimated (or undecimated) trace is what is used for marker measurements and for results export. |
| Settling Time decimal places (Frequency and Phase Settling Time displays only)         | Specifies the number of decimal places shown in the results readout below the graph. This setting does not affect the precision of internal calculations.   |
| Show trace legend (CCDF (Complementary Cumulative Distribution Function) display only) | Displays or hides the trace legend.   |



## WLAN Overview

The WLAN Analysis option allows you to evaluate WLAN signals, performing standards based transmitter measurements in the time, frequency, and modulation domains. WLAN measurements can be made on signals defined by 802.11a/b/g/j/n/p (standard IEEE 802.11-2012) and 802.11ac (draft IEEE P802.11ac-D4.0 or later). User controls allow you to modify signal parameters for analysis of signals. The analysis results give multiple views of WLAN signal characteristics to allow the diagnosis of signal imperfections and impairments quickly and easily. Display controls allow you to selectively display the analysis results to help locate trouble-spots in the signal.

### WLAN Standards

The following options support the given standards:

- **Option SV23:** Supports IEEE 802.11a, g, j, and p OFDM signals and 802.11b DSSS/CCK signals
- **Option SV24:** Supports IEEE 802.11n signals with bandwidths of 20 MHz and 40 MHz
- **Option SV25:** Supports IEEE 802.11ac signals with bandwidths of 20, 40, 80, and 160 MHz

### WLAN Standards Presets

The **Standards Presets** button located on the WLAN Settings Control Panel allows you to recall preconfigured displays for the standards and bandwidths that you select. You can also access these presets from **Presets** on the display menu bar. You can read more information about the [Standards Presets \(see page 22\)](#).

The following table shows the center frequency and displays that are automatically loaded for each of the listed standards and bandwidths. SEM masks are also loaded and are explained [here \(see page 339\)](#).

**Table 4: WLAN presets standards, bandwidths, center frequencies, and displays**

| Standard | Bandwidth (MHz) | Center frequency (GHz) | Displays loaded                                      |
|----------|-----------------|------------------------|--|
| 802.11a  | 20              | 5.18                   | SEM, Time Overview, WLAN Constellation, WLAN Summary |
| 802.11g  | 20              | 2.412                  |  |
| 802.11j  | 10              | 5.18                   |  |
|          | 20              |                        |  |
| 802.11p  | 5               | 5.8                    |  |
|          | 10              |                        |  |
|          | 20              |                        |  |
| 802.11n  | 20              | 2.412                  |  |
|          | 40              |                        |  |
| 802.11ac | 20              | 5.18                   |  |
|          | 40              | 5.19                   |  |
|          | 80              | 5.21                   |  |
|          | 160             | 5.25                   |  |
| 802.11b  |                 | 2.412                  |  |

**SEM mask parameters.** The SEM wireless standard masks that are applied to the WLAN signal depend on the standard you select when you configure the preset. Once you select a standard and bandwidth, the application will automatically load the mask that best fits the signal for which the standard applies. All mask parameters are derived from IEEE standards and loaded for you. This provides you the assurance that you are evaluating the signal with the most appropriate mask.

## WLAN Displays

The displays in WLAN Analysis (Setup > Displays > Measurements) are:

- [SEM \(see page 214\)](#)
- [WLAN Channel Response \(see page 235\)](#)
- [WLAN Constellation \(see page 238\)](#)
- [WLAN EVM \(see page 240\)](#)
- [WLAN Magnitude Error \(see page 242\)](#)
- [WLAN Phase Error \(see page 244\)](#)
- [WLAN Power versus Time \(see page 246\)](#)
- [WLAN Spectral Flatness \(see page 249\)](#)

- [WLAN Summary \(see page 251\)](#)
- [WLAN Symbol Table \(see page 256\)](#)

## WLAN Channel Response Display

The WLAN Channel Response display plots the channel response (magnitude and phase) versus the subcarrier or frequency. Here, the channel refers to all sources of signal frequency response impairment up to the analyzer input, including the transmitter itself, as well as any transmission medium through which the signal travels between the transmitter and the analyzer.

---


**NOTE.** *WLAN Channel Response is only available for OFDM (non-802.11b) signals.*

---

To show the WLAN Channel Response display:

1. If you are analyzing a stored data file, recall the data file now. Otherwise, go to the next step.
2. Press the **Displays** button or select **Setup > Displays**.
3. In the **Select Displays** dialog, select **WLAN Analysis** in the **Measurements** box.
4. In the Available displays box, double-click the **WLAN Chan Response** icon or select the icon and click **Add**. The WLAN Chan Response icon will appear in the **Selected displays** box and will no longer appear under Available displays.
5. Click **OK** to show the WLAN Chan Response display.
6. Set the Frequency appropriate for the signal.
7. Select **Setup > Settings** to display the control panel.
8. Select the Modulation Params tab. Set the Standard, Guard Interval, Subcarrier Spacing, and Bandwidth controls as appropriate for the input signal.

## Elements of the Display



0519-048

| Item     | Display element                    | Description   |
|----------|------------------------------------|---|
| <b>1</b> | <b>Magnitude graph (top graph)</b> |   |
| 1a       | Top of graph                       | Sets the level that appears at the top of the magnitude graph. This is only a visual control for panning the graph. The Reference Level is adjusted in the Toolbar and the Ampl control panel. By default, Vert Position = Ref Level.   |
| 1b       | Position (dB)                      | Sets the vertical position value. The maximum value is 20.00 dB. The readout indicates the subcarrier (for non-b standards only) or frequency shown at the bottom of the display.   |
| 1c       | Bottom of graph readout            | Indicates the magnitude at the bottom of the top graph. This value changes with the dB and vertical Position settings.  |
| <b>2</b> | <b>Phase graph (bottom graph)</b>  |   |
| 2a       | Top of graph                       | Sets the phase value indicated at the top of the graph. Since the Position value at the vertical center of this graph remains constant as the Top of Graph value is adjusted, the Vertical Scale increases as the Top of Graph value increases, which also affects the bottom of graph readout. |
| 2b       | Position (°)                       | Specifies the phase shown at the center of the graph display. Changing this value moves the trace up and down in the graph, which affects the Top of Graph and Bottom of Graph values as well.  |
| 2c       | Bottom of bottom graph readout     | Indicates the phase at the bottom of the bottom graph. This value changes with the Position setting.  |
| 3        | Autoscale                          | Adjusts the Vertical and Horizontal scaling to optimize the trace display on screen.  |
| 4        | Start (Position)                   | Shifts the trace left or right in the graph. The readout indicates the subcarrier (for non-b standards only) or frequency shown at the left edge of the display.  |
| 5        | Stop (Scale)                       | Specifies the number of subcarriers (for non-b standards only) shown in the graph.  |

## WLAN Channel Response Settings

The WLAN Channel Response Settings control panel provides access to settings that control parameters of the Channel Response Display.

| Settings tab                                     | Description  |
|--|--|
| <a href="#">Modulation Params (see page 259)</a> | Specifies the input signal standard and additional user-settable signal parameters.  |
| <a href="#">Analysis Params (see page 260)</a>   | Specifies parameters used by the instrument to analyze the input signal.   |
| <a href="#">Data Range Tab (see page 262)</a>    | Specifies which symbols and subcarriers of the signal to display.  |
| <a href="#">Analysis Time Tab (see page 263)</a> | Specifies parameters that define the portion of the acquisition record that is used for analysis. Also allows you to specify the time units (Symbols or Seconds) for WLAN Analysis displays. |

| Settings tab                              | Description  |
|---|--|
| <a href="#">Traces Tab (see page 264)</a> | Enables you to select from magnitude or phase trace, save a trace, and recall an trace |
| <a href="#">Scale Tab (see page 265)</a>  | Specifies the Zoom scale, and vertical and horizontal positions of the display.        |
| <a href="#">Prefs Tab (see page 268)</a>  | Specifies the units of the display and whether elements of the graphs are displayed.   |


## WLAN Constellation Display

The WLAN Constellation Display shows the WLAN signal modulation amplitude and phase in I (horizontal) versus Q (vertical) form. For multicarrier WLAN OFDM signals, the points show all data symbol subcarriers' modulation. For single-carrier 802.11b, each point corresponds to a single modulated chip.

To show the WLAN Constellation display:

1. If you are analyzing a stored data file, recall the data file now. Otherwise, go to the next step.
2. Press the **Displays** button or select **Setup > Displays**.
3. In the **Select Displays** dialog, select **WLAN Constellation** in the **Measurements** box.
4. In the Available displays box, double-click the **WLAN Constellation** icon or select the icon and click **Add**. The WLAN Constellation icon will appear in the **Selected displays** box and will no longer appear under Available displays.
5. Click **OK** to show the WLAN Constellation display.
6. Set the **Frequency** appropriate for the signal.
7. Select **Setup > Settings** to display the control panel.
8. Select the **Modulation Params** tab. Set the Standard, Guard Interval, Channel Bandwidth, and Subcarrier Spacing controls as appropriate for the input signal.

## Elements of the Display



| Item | Display element | Description  |
|------|-----------------|--|
| 1    | Plot            | Constellation graph.   |
| 2    | Marker readout  | If markers are enabled, the marker readout shows the selected Marker, Symbol, Subcarrier, Type, Value, Magnitude, and Phase for the symbol at the marker location. Located to the left of the constellation plot or below it, depending on the size of the window. |
| 3    | Marker controls | Define and position markers.   |

## WLAN Constellation Settings

### Front Panel: Settings

#### Application Toolbar:

The WLAN Constellation Settings control panel provides access to settings that control parameters of the Constellation Display.

| Settings tab                                     | Description   |
|--|---|
| <a href="#">Modulation Params (see page 259)</a> | Specifies the input signal standard and additional user-settable signal parameters.   |
| <a href="#">Analysis Params (see page 260)</a>   | Specifies parameters used by the instrument to analyze the input signal.  |
| <a href="#">Data Range (see page 262)</a>        | Specifies which symbols and subcarriers of the signal to display.   |
| <a href="#">Analysis Time Tab (see page 263)</a> | Specifies parameters that define the portion of the acquisition record that is used for analysis. Also allows you to specify the Units (Symbols or Seconds) for WLAN Analysis displays. |
| <a href="#">Trace Tab (see page 264)</a>         | Enables you to freeze the display or hide the measurement or average trace.   |
| <a href="#">Scale Tab (see page 265)</a>         | Specifies the Zoom scale, and vertical and horizontal positions of the display.   |
| <a href="#">Prefs Tab (see page 268)</a>         | Specifies the units of the display and whether elements of the graphs are displayed.  |

## WLAN EVM Display

The WLAN EVM display shows the data symbols' individual subcarrier Error Vector Magnitude values versus symbol interval (time) and subcarrier (frequency).

---


**NOTE.** For 802.11b analysis, the subcarrier graph is not displayed.

---

To show the WLAN EVM display:

1. If you are analyzing a stored data file, recall the data file now. Otherwise, go to the next step.
2. Press the **Displays** button or select **Setup > Displays**.
3. In the **Select Displays** dialog, select **WLAN EVM** in the **Measurements** box.
4. In the Available displays box, double-click the **WLAN EVM** icon or select the icon and click **Add**. The WLAN EVM icon will appear in the **Selected displays** box and will no longer appear under Available displays.
5. Click **OK** to show the WLAN EVM display.
6. Set the **Frequency** appropriate for the signal.
7. Select **Setup > Settings** to display the control panel.
8. Select the **Modulation Params** tab. Set the Standard as appropriate. Set the Standard, Guard Interval, Channel Bandwidth, and Subcarrier Spacing controls as appropriate for the input signal.

## Elements of the Display



| Item | Display element  | Description   |
|------|------------------|---|
| 1    | Top of graph     | Sets the EVM value that appears at the top of the graph. This is only a visual control for panning the graph.   |
| 2    | Bottom of graph  | Sets the EVM value that appears at the bottom of the graph. This is only a visual control for panning the graph.  |
| 3    | Autoscale        | Adjusts the vertical and horizontal settings to provide the best display.   |
| 4    | Start (Position) | Specifies the horizontal position of the trace on the graph display.  |
| 5    | Stop (Scale)     | Adjusts the horizontal range of the graph. By decreasing the scale, the graph essentially becomes a window that you can move over the analysis results by adjusting the position. |

## WLAN EVM Settings

### Front Panel: Settings

#### Application Toolbar:

The settings for the WLAN EVM display are shown in the following table.

| Settings tab                                     | Description   |
|--|---|
| <a href="#">Modulation Params (see page 259)</a> | Specifies the type of modulation used for the input signal and other parameters.  |
| <a href="#">Analysis Params (see page 260)</a>   | Specifies parameters used by the instrument to analyze the input signal.  |
| <a href="#">Data Range (see page 262)</a>        | Specifies which symbols and subcarriers of the signal to display.   |
| <a href="#">Analysis Time (see page 263)</a>     | Specifies parameters that define the portion of the acquisition record that is used for analysis. Also allows you to specify the Units (Symbols or Seconds) for WLAN Analysis displays. |
| <a href="#">Trace (see page 264)</a>             | Enables you to freeze the display or hide the measurement or average trace.   |
| <a href="#">Scale (see page 265)</a>             | Specifies the vertical, subcarrier (for non-b standards only) and symbols scale and position settings.  |
| <a href="#">Prefs (see page 268)</a>             | Specifies the units of the display and whether elements of the graphs are displayed.  |

## WLAN Magnitude Error Display

The WLAN Magnitude Error display shows the data symbols' individual subcarrier Magnitude Error values versus symbol interval (time) and subcarrier (frequency).

---


**NOTE.** For 802.11b analysis, the subcarrier graph is not displayed.

---

To show the WLAN Magnitude Error display:

1. If you are analyzing a stored data file, recall the data file now. Otherwise, go to the next step.
2. Press the **Displays** button or select **Setup > Displays**.
3. In the **Select Displays** dialog, select **WLAN Analysis** in the **Measurements** box.
4. In the Available displays box, double-click the **WLAN Mag Error** icon or select the icon and click **Add**. The WLAN Mag Error icon will appear in the **Selected displays** box and will no longer appear under Available displays.
5. Click **OK** to show the WLAN Mag Error display.
6. Set the **Frequency** appropriate for the signal.
7. Select **Setup > Settings** to display the control panel.
8. Select the **Modulation Params** tab. Set the Standard, Guard Interval and Spacing and Bandwidth controls as appropriate for the input signal.

## Elements of the Display



| Item | Display element | Description   |
|------|-----------------|---|
| 1    | Top of graph    | Sets the Magnitude Error value that appears at the top of the graph. This is only a visual control for panning the graph.   |
| 2    | Bottom of graph | Sets the Magnitude Error value that appears at the bottom of the graph. This is only a visual control for panning the graph.  |
| 3    | Autoscale       | Adjusts the vertical and horizontal settings to provide the best display.   |
| 4    | Pos             | Specifies the horizontal position of the trace on the graph display.  |
| 5    | Scale           | Adjusts the horizontal range of the graph. By decreasing the scale, the graph essentially becomes a window that you can move over the analysis results by adjusting the position. |

## WLAN Magnitude Error Settings

### Front Panel: Settings

#### Application Toolbar:

The settings for the WLAN Mag Error display are shown in the following table.

| Settings tab                                     | Description   |
|--|---|
| <a href="#">Modulation Params (see page 259)</a> | Specifies the type of modulation used for the input signal and other parameters.  |
| <a href="#">Analysis Params (see page 260)</a>   | Specifies parameters used by the instrument to analyze the input signal.  |
| <a href="#">Data Range (see page 262)</a>        | Specifies which symbols and subcarriers of the signal to display.   |
| <a href="#">Analysis Time Tab (see page 263)</a> | Specifies parameters that define the portion of the acquisition record that is used for analysis. Also allows you to specify the Units (Symbols or Seconds) for WLAN Analysis displays. |
| <a href="#">Trace (see page 264)</a>             | Enables you to display or hide the measurement or average trace.  |
| <a href="#">Scale Tab (see page 265)</a>         | Specifies the vertical, subcarrier (for non-b standards only), and symbols scale and position settings.   |
| <a href="#">Prefs Tab (see page 268)</a>         | Specifies the units of the display and whether elements of the graphs are displayed.  |

## WLAN Phase Error Display

The WLAN Phase Error display shows the data symbols' individual subcarrier Phase Error values versus symbol interval (time) and subcarrier (frequency).

---


**NOTE.** For 802.11b analysis, the subcarrier graph is not displayed.

---

To show the WLAN Phase Error display:

1. If you are analyzing a stored data file, recall the data file now. Otherwise, go to the next step.
2. Press the **Displays** button or select **Setup > Displays**.
3. In the **Select Displays** dialog, select **WLAN Analysis** in the **Measurements** box.
4. In the Available displays box, double-click the **WLAN Phase Error** icon or select the icon and click **Add**. The WLAN Phase Error icon will appear in the **Selected displays** box and will no longer appear under Available displays.
5. Click **OK** to show the WLAN Phase Error display.
6. Set the **Frequency** appropriate for the signal.
7. Select **Setup > Settings** to display the control panel.
8. Select the **Modulation Params** tab. Set the Standard, Guard Interval and Spacing and Bandwidth controls as appropriate for the input signal.

## Elements of the Display



| Item | Display element | Description   |
|------|-----------------|---|
| 1    | Top of graph    | Sets the Phase Error value that appears at the top of the graph. This is only a visual control for panning the graph.   |
| 2    | Bottom of graph | Sets the Phase Error value that appears at the bottom of the graph. This is only a visual control for panning the graph.  |
| 3    | Autoscale       | Adjusts the vertical and horizontal settings to provide the best display.   |
| 4    | Pos             | Specifies the horizontal position of the trace on the graph display.  |
| 5    | Scale           | Adjusts the horizontal range of the graph. By decreasing the scale, the graph essentially becomes a window that you can move over the analysis results by adjusting the position. |

## WLAN Phase Error Settings

### Front Panel: Settings

#### Application Toolbar:

The settings for the WLAN Phase Error display are shown in the following table.

| Settings tab                                     | Description   |
|--|---|
| <a href="#">Modulation Params (see page 259)</a> | Specifies the type of modulation used for the input signal and other parameters.  |
| <a href="#">Analysis Params (see page 260)</a>   | Specifies parameters used by the instrument to analyze the input signal.  |
| <a href="#">Data Range (see page 262)</a>        | Specifies which symbols and subcarriers of the signal to display.   |
| <a href="#">Analysis Time Tab (see page 263)</a> | Specifies parameters that define the portion of the acquisition record that is used for analysis. Also allows you to specify the Units (Symbols or Seconds) for WLAN Analysis displays. |
| <a href="#">Trace (see page 264)</a>             | Enables you to display or hide the measurement or average trace.  |
| <a href="#">Scale Tab (see page 265)</a>         | Specifies the vertical, subcarrier (for non-b standards only), and symbols scale and position settings.   |
| <a href="#">Prefs Tab (see page 268)</a>         | Specifies the units of the display and whether elements of the graphs are displayed.  |

## WLAN Power vs Time Display

The WLAN Power vs Time display shows the signal power amplitude versus time. For 802.11b signals, the packet Power-On and Power-Down ramp times are also measured.

To show the WLAN Power vs Time display:

1. If you are analyzing a stored data file, recall the data file now. Otherwise, go to the next step.
2. Press the **Displays** button or select **Setup > Displays**.
3. In the **Select Displays** dialog, select **WLAN Analysis** in the **Measurements** box.
4. In the Available displays box, double-click the **WLAN Power vs Time** icon or select the icon and click **Add**. The WLAN Power vs Time icon will appear in the **Selected displays** box and will no longer appear under Available displays.
5. Click **OK** to show the WLAN Power vs Time display.
6. Set the **Frequency** appropriate for the signal.
7. Select **Setup > Settings** to display the control panel.
8. Select the **Modulation Params** tab. Set the Standard, Guard Interval and Spacing and Bandwidth controls as appropriate for the input signal.

## Elements of the Display



| Item | Display element              | Description  |
|------|------------------------------|--|
| 1    | Top of graph, first setting  | Sets the Power level that appears at the top of the graph, in dBm. This is only a visual control for panning the graph.  |
| 2    | Top of graph, second setting | Sets the vertical Scale of the graphs, in dB/div. This is only a visual control for panning the graph.   |
| 3    | View                         | <p>Selects the specific view of the packet burst within the display:</p> <ul style="list-style-type: none"> <li>- Full Burst displays the entire packet, with vertical lines indicating length of the packet</li> <li>- Rising Edge zooms the display into the interval around the packet rising edge, with vertical lines indicating the 10% to 90% Power-On Ramp time</li> <li>- Falling Edge zooms the display into the interval around the packet falling edge, with vertical lines indicating the 90% to 10% Power-Down Ramp time</li> </ul>  |
| 4    | Bottom of graph readout      | Shows the Power level at the bottom of the graph in dBm.   |
| 5    | Autoscale                    | Adjusts the vertical and horizontal settings to provide the best display.  |
| 6    | Bottom of graph, left side   | Sets the starting time of the graph in seconds   |
| 7    | Bottom of graph, right side  | Sets the scale (width) of the graph in seconds   |
| 8    | Table below graph            | <ul style="list-style-type: none"> <li>- Burst Power: Average power of the burst packet, in dBm</li> <li>- Burst Width: Measured time width of the burst packet from Power-On to Power-Down (or end of waveform, if that occurs before Power-Down), in seconds</li> </ul> <p><b>NOTE.</b> Power-On Ramp Time and Power-Down Ramp Time values are only available for 802.11b analysis.</p> <ul style="list-style-type: none"> <li>- Power-On Ramp Time: Time interval for signal level to increase from 10% to 90% of maximum packet power, in seconds</li> <li>- Power-Down Ramp Time: Time interval for signal level to decrease from 90% to 10% of maximum packet power, in seconds. This value is not available if the analysis record does not include the packet power-down portion.</li> </ul> |

## WLAN Power vs Time Settings

### Front Panel: Settings

### Application Toolbar:

The settings for the WLAN Power vs Time display are shown in the following table.

| Settings tab                                     | Description   |
|--|---|
| <a href="#">Modulation Params (see page 259)</a> | Specifies the type of modulation used for the input signal and other parameters.  |
| <a href="#">Analysis Params (see page 260)</a>   | Specifies parameters used by the instrument to analyze the input signal.  |
| <a href="#">Data Range (see page 262)</a>        | Specifies which symbols and subcarriers of the signal to display.   |
| <a href="#">Analysis Time (see page 263)</a>     | Specifies parameters that define the portion of the acquisition record that is used for analysis. Also allows you to specify the Units (Symbols or Seconds) for WLAN Analysis displays. |
| <a href="#">Scale (see page 265)</a>             | Specifies the vertical, subcarrier (for non-b standards only), and symbols scale and position settings.   |
| <a href="#">Prefs (see page 268)</a>             | Specifies the units of the display and whether elements of the graphs are displayed.  |

## WLAN Spectral Flatness Display

The WLAN Spectral Flatness display shows the average power levels of subcarriers across the signal bandwidth and indicates if they remain within the limits defined for a particular standard.

The averaged subcarrier level is shown as a line on the display, while the individual subcarrier levels are shown as points. The Pass or Fail result indicates whether the average line remains between the upper and lower limit mask boundaries.

---


**NOTE.** *WLAN Spectral Flatness is only available for OFDM (non-802.11b) signals.*

---

To show the WLAN Spectral Flatness display:

1. If you are analyzing a stored data file, recall the data file now. Otherwise, go to the next step.
2. Press the **Displays** button or select **Setup > Displays**.
3. In the **Select Displays** dialog, select **WLAN Analysis** in the **Measurements** box.
4. In the Available displays box, double-click the **WLAN Flatness** icon or select the icon and click **Add**. The WLAN Flatness icon will appear in the **Selected displays** box and will no longer appear under Available displays.
5. Click **OK** to show the WLAN Flatness display.
6. Set the **Frequency** appropriate for the signal.
7. Select **Setup > Settings** to display the control panel.
8. Select the **Modulation Params** tab. Set the Standard, Guard Interval, Subcarrier Spacing, and Channel Bandwidth controls as appropriate for the input signal.

## Elements of the Display



| Item | Display element              | Description   |
|------|------------------------------|---|
| 1    | Top of graph, first setting  | Sets the Flatness value that appears at the top of the graph in dB. This is only a visual control for panning the graph.  |
| 2    | Top of graph, second setting | Sets the lower range of the Flatness value that appears at the top of the graph in dB/div. This is only a visual control for panning the graph.                                   |
| 3    | Bottom of graph              | Shows the Flatness value set at the bottom of the graph.  |
| 4    | Autoscale                    | Adjusts the vertical and horizontal settings to provide the best display.   |
| 5    | Posn                         | Specifies the horizontal position of the trace on the graph display.  |
| 6    | Scale                        | Adjusts the horizontal range of the graph. By decreasing the scale, the graph essentially becomes a window that you can move over the analysis results by adjusting the position. |

## WLAN Spectral Flatness Settings

### Front Panel: Settings

### Application Toolbar:

The settings for the WLAN Spectral Flatness display are shown in the following table.

| Settings tab                                     | Description  |
|--|--|
| <a href="#">Modulation Params (see page 259)</a> | Specifies the type of modulation used for the input signal and other parameters.   |
| <a href="#">Analysis Params (see page 260)</a>   | Specifies parameters used by the instrument to analyze the input signal.   |
| <a href="#">Data Range (see page 262)</a>        | Specifies which symbols and subcarriers of the signal to display.  |
| <a href="#">Analysis Time Tab (see page 263)</a> | Specifies parameters that define the portion of the acquisition record that is used for analysis.<br><br><b>NOTE.</b> The Units control on this tab only affects the Analysis Length (AFAIK). Use the Units control in the Prefs tab to affect the displays. |
| <a href="#">Trace (see page 264)</a>             | Enables you to display or hide the measurement or average trace.   |
| <a href="#">Scale Tab (see page 265)</a>         | Specifies the vertical, subcarrier (for non-b standards only), and symbols scale and position settings.  |
| <a href="#">Prefs Tab (see page 268)</a>         | Specifies the units of the display and whether elements of the graphs are displayed.   |

## WLAN Summary Display

The WLAN Summary display shows several measurements of WLAN signal quality. The summary display and its contents will vary by standard selected. Not all of the display contents are provided in this section.

To show the WLAN Summary display:

1. If you are analyzing a stored data file, recall the data file now. Otherwise, go to the next step.
2. Press the **Displays** button or select **Setup > Displays**.
3. In the **Select Displays** dialog, select **WLAN Analysis** in the **Measurements** box.
4. In the Available displays box, double-click the **WLAN Summary** icon or select the icon and click **Add**. The WLAN Summary icon will appear in the **Selected displays** box and will no longer appear under Available displays.
5. Click **OK** to show the WLAN Summary display.
6. Set the **Frequency** appropriate for the signal.
7. Select **Setup > Settings** to display the control panel.
8. Select the **Modulation Params** tab. Set the Standard, Guard Interval and Spacing and Bandwidth controls as appropriate for the input signal.

## WLAN Summary Display for 802.11a/g/j/p/n/ac Signals

|                          |             |                            |           |                                      |     |                                      |  |
|--------------------------|-------------|----------------------------|-----------|--------------------------------------|-----|--------------------------------------|--|
| <b>Standard:</b> 802.11n |             | <b>Bandwidth:</b> 40 MHz   |           | <b>Guard Interval:</b> Auto from SIG |     | <input type="button" value="Clear"/> |  |
| <b>Burst Power:</b>      | -0.01 dBm   | <b>Peak-to-Average:</b>    | 9.80 dB   | <b>L-SIG Data</b>                    |     |                                      | <b>Parity:</b> <span style="color: green;">Pass</span> |
| <b>Burst Index:</b>      | 1           | <b>IQ Origin Offset:</b>   | -82.25 dB | <b>Rate</b>                          | 13  | <b>Reserved</b>                      | 0  |
| <b>Frequency Error:</b>  | -140.27 mHz | <b>Common Pilot Error:</b> | 0.337 %   | <b>Length</b>                        | 312 | <b>Parity</b>                        | 1  |
| <b>Symbol Clk Error:</b> | -0.070 ppm  |                            |           | <b>Tail</b>                          | 0   |                                      |  |

|                   |            |               |             |
|-------------------|------------|---------------|-------------|
| <b>EVM</b>        |            |               |             |
|                   | <b>All</b> | <b>Pilots</b> | <b>Data</b> |
| <b>RMS</b>        | -39.91 dB  | -40.40 dB     | -39.89 dB   |
| <b>Peak</b>       | -27.33 dB  | -31.09 dB     | -27.33 dB   |
| <b>Pk@Sym/Sub</b> | 32 / 58    | 85 / 53       | 32 / 58     |
| <b>Avg RMS</b>    | -39.91 dB  |               |             |
| 1 of 1 Bursts     |            |               |             |
| <b>Max RMS</b>    | -39.91 dB  |               |             |

|                    |     |   |   |
|--------------------|-----|---|---|
| <b>HT-SIG Data</b> |     | <b>CRC:</b> <span style="color: green;">Pass</span> |   |
| <b>MCS</b>         | 0   | <b>CBW</b>  | 1 |
| <b>Length</b>      | 672 | <b>Smoothing</b>                                    | 0 |
| <b>Not Sound</b>   | 0   | <b>Reserved</b>                                     | 1 |
| <b>Aggregation</b> | 0   | <b>STBC</b>   | 0 |
| <b>FEC Coding</b>  | 0   | <b>Short GI</b>                                     | 0 |
| <b>Ness</b>        | 0   | <b>CRC</b>  | 1 |
| <b>Tail</b>        | 0   |   |   |

|                             |                |                              |                  |
|-----------------------------|----------------|------------------------------|------------------|
| <b>Packet Format:</b> HT_MF |                | <b>Data Modulation:</b> BPSK |                  |
|                             |                | <b>Guard Interval:</b> 1/4   |                  |
|                             | <b>Symbols</b> | <b>EVM</b>                   | <b>Avg Power</b> |
| <b>L-STF</b>                | 2              | -44.57 dB                    | 0.00 dBm         |
| <b>L-LTF</b>                | 2              | -40.02 dB                    | -0.01 dBm        |
| <b>L-SIG</b>                | 1              | -40.43 dB                    | -0.26 dBm        |
| <b>HT-SIG</b>               | 2              | -40.34 dB                    | -0.47 dBm        |
| <b>HT-STF</b>               | 1              | -44.08 dB                    | -0.01 dBm        |
| <b>HT-LTF</b>               | 1              | -42.43 dB                    | 0.00 dBm         |
| <b>Data</b>                 | 100            | -39.91 dB                    | 0.01 dBm         |

### Elements of the Display for 802.11a/g/j/p/n/ac Signals

| Measurement   | Description   |
|---|---|
| Standard  | Display of the standard selected on the Setup > Settings > Modulation Params tab.   |
| Bandwidth   | Display of the channel bandwidth selected on Setup > Settings > Modulation Parameters tab.  |
| Burst Power   | The average power of all symbols in the packet, including Preamble and Data segments.   |
| Peak-to-Average                                       | The ratio of the highest instantaneous signal power level to the average signal power.  |
| Burst Index   | The index of the analyzed packet within the analysis record.  |
| IQ Origin Offset                                      | The average magnitude of the DC subcarrier level relative to total signal power. It indicates the level of carrier feedthrough detected at the center (DC) subcarrier.  |
| Frequency Error                                       | The frequency difference between the measured carrier frequency of the signal and the measurement frequency setting.  |
| Common Pilot Error                                    | The RMS magnitude error of the pilots over all data symbols.  |
| Symbol Clk Error                                      | The symbol clock error in parts per million.  |
| EVM   | <p>The RMS and Peak values of the normalized subcarrier Error Vector Magnitude values. The normalized subcarrier EVM values are calculated as the difference between the detected received signal subcarrier constellation points and ideal reference points estimated by the instrument from the received signal. Values are reported in units of percent or dB. Peak values include the symbol and subcarrier location.</p> <p>RMS and Peak values are displayed for groupings of all subcarriers, Pilots only and Data only. Results are calculated over all Data symbols in the packet.</p> <p>Average RMS and Peak RMS values are accumulated over multiple packet analysis cycles. The Clear button on the display resets these values by clearing the result memory.</p> |
| Packet Format   | Displays the packet format: AG, HT_MF, HT_GF, VHT.  |
| Data Modulation                                       | Displays the modulation used in the Data symbols: BPSK, QPSK, 16QAM, 64QAM, 256QAM.   |
| Guard Interval  | Displays the Guard Interval used by the Data symbols: 1/4, 1/8.   |
| Symbols, EVM, Avg Power                               | Displays the type and number of symbols, EVM-RMS and average power of the Preamble and Data portions of the packet.   |
| SIG Data<br>L-SIG Data<br>HT-SIG Data<br>VHT-SIG Data | Displays the decimal values of the received packet SIGNAL, HT-SIGNAL and VHT-SIGNAL symbols' fields. The Pass/Fail result in each heading line indicates whether the calculated Parity or CRC value matches the received Parity or CRC value of the corresponding SIGNAL symbol grouping.   |

## WLAN Summary Display for 802.11b Signals

|                          |           |  |           |
|--------------------------|-----------|--|-----------|
| <b>Standard:</b> 802.11b |           | <b>Packet Format:</b> Auto Detect <input type="button" value="Clear"/> |           |
| <b>Burst Power:</b>      | -1.22 dBm | <b>Peak-to-Average:</b>  | 1.50 dB   |
| <b>Burst Index:</b>      | 1         | <b>IQ Origin Offset:</b>   | -39.12 dB |
| <b>Frequency Error:</b>  | -2.34 Hz  |  |           |

|                  |            |                 |
|------------------|------------|-----------------|
| <b>EVM</b>       |            |                 |
|                  | <b>All</b> | <b>1k Chips</b> |
| <b>RMS</b>       | -39.12 dB  | -39.02 dB       |
| <b>Peak</b>      | -30.29 dB  | -30.32 dB       |
| <b>Pk @ Chip</b> | 2992       | 66              |
| <b>Avg RMS</b>   | -39.12 dB  |                 |
| 1 of 1 Bursts    |            |                 |
| <b>Max RMS</b>   | -39.12 dB  |                 |

|                               |              |                                 |                  |
|-------------------------------|--------------|---------------------------------|------------------|
| <b>Packet Format:</b> DSSS 2M |              | <b>Data Modulation:</b> DSSS 2M |                  |
|                               | <b>Chips</b> | <b>EVM</b>                      | <b>Avg Power</b> |
| <b>Preamble:</b>              | 1584         | -38.96 dB                       | -1.72 dBm        |
| <b>Header:</b>                | 528          | -39.06 dB                       | -1.23 dBm        |
| <b>Data:</b>                  | 4400         | -39.19 dB                       | -1.23 dBm        |

|                    |     |                  |       |
|--------------------|-----|------------------|-------|
| <b>Header Data</b> |     | <b>CRC:</b> Pass |       |
| <b>Signal</b>      | 20  | <b>Service</b>   | 0     |
| <b>Length</b>      | 400 | <b>CRC</b>       | 17735 |

### Elements of the Display for 802.11b Signals

| Measurement           | Description   |
|-----------------------|---|
| Standard              | Display of the standard selected on the Setup > Settings > Modulation Parameters tab.   |
| Burst Power           | The average power of the entire packet.   |
| Peak-to-Average       | The ratio of the highest instantaneous packet signal power level to the average signal power.   |
| Burst Index           | The index of the analyzed packet within the analysis record.  |
| IQ Origin Offset      | The amount of power at the signal carrier frequency relative to the total power of the signal.  |
| Frequency Error       | The frequency difference between the measured carrier frequency of the signal and the measurement frequency setting.  |
| EVM                   | <p>The RMS and Peak values of the normalized Error Vector Magnitudes taken at signal chip intervals. The normalized EVM values are calculated as the difference between the detected received signal constellation points and ideal reference points estimated by the instrument from the received signal. Values are reported in units of percent or dB. Peak values indicate the associated chip locations.</p> <p>RMS and Peak values are displayed for groupings of All chips, and the first 1000 (1k) chips.</p> <p>Average RMS and Peak RMS values are accumulated over multiple packet analysis cycles. The Clear button on the display resets these values by clearing the result memory.</p> |
| Packet Format         | Displays the packet format: DSSS 1M, DSSS 2M, CCK 5.5M or CCK 11M.  |
| Data Modulation       | Displays the modulation used in the PSDU/Data packet segment.   |
| Chips, EVM, Avg Power | Displays the number of chips, EVM-RMS and average power of the Preamble, Header and Data portions of the packet.  |
| Header Data           | Displays the decimal values of the received packet Header's Signal, Service, Length, and CRC fields. The CRC Pass/Fail result indicates if the received CRC value is equal to the CRC value computed from the received Signal, Service and Length values.   |

## WLAN Summary Settings

### Front Panel: Settings

### Application Toolbar:

The settings for the WLAN Summary display are shown in the following table.

| Settings tab                                     | Description   |
|--|---|
| <a href="#">Modulation Params (see page 259)</a> | Specifies the type of modulation used for the input signal and other parameters.  |
| <a href="#">Analysis Params (see page 260)</a>   | Specifies parameters used by the instrument to analyze the input signal.  |
| <a href="#">Data Range (see page 262)</a>        | Specifies which symbols and subcarriers of the signal to display.   |
| <a href="#">Analysis Time (see page 263)</a>     | Specifies parameters that define the portion of the acquisition record that is used for analysis. Also allows you to specify the Units (Symbols or Seconds) for WLAN Analysis displays.   |
| <a href="#">EVM (see page 267)</a>               | Specifies the EVM units (dB or %) and Max Bursts to Avg. The EVM is generally measured on symbol or chip instants and is usually measured after best-fit estimates of the frequency error and a fixed phase offset have been removed. |
| <a href="#">Prefs (see page 268)</a>             | Specifies the units of the display and whether elements of the graphs are displayed.  |


## WLAN Symbol Table Display

The WLAN Symbol Table display shows decoded data values for each data symbol in the analyzed signal packet. For OFDM signals (all standards except 802.11b), results are presented with subcarrier (frequency) indices in the horizontal dimension and symbol (time) intervals in the vertical dimension. For 802.11b signals, the Preamble, Header, and Data (PSDU) symbol values are presented sequentially, with symbol indices in the left column.

To show the WLAN Symbol Table display:

1. If you are analyzing a stored data file, recall the data file now. Otherwise, go to the next step.
2. Press the **Displays** button or select **Setup > Displays**.
3. In the **Select Displays** dialog, select **WLAN Analysis** in the **Measurements** box.
4. In the Available displays box, double-click the **WLAN Symb Table** icon or select the icon and click **Add**. The WLAN Symbol icon will appear in the **Selected displays** box and will no longer appear under Available displays.
5. Click **OK** to show the WLAN Symbol Table display.
6. Set the **Frequency** appropriate for the signal.
7. Select **Setup > Settings** to display the control panel.
8. Select the **Modulation Params** tab. Set the Standard, Guard Interval and Spacing and Bandwidth controls as appropriate for the input signal.

## WLAN Symbol Table for OFDM signals



| Item | Description   |
|------|---|
| 1    | Subcarrier identifiers (OFDM only). Yellow column indicates Pilot subcarrier locations. |
| 2    | Symbol identifiers.   |
| 3    | Subcarrier data values.   |
| 4    | Marker readout when markers are enabled.  |

## WLAN Symbol Table for 802.11b signals

| Symbols | (Preamble, Header, Data) |   |   |   |   |   |   |   |   |   |   |   |   |  |  |  |
|---------|--------------------------|---|---|---|---|---|---|---|---|---|---|---|---|--|--|--|
| 130     | 0                        | 0 | 0 | 1 | 0 | 1 | 1 | 1 | 0 | 0 | 1 | 1 | 1 |  |  |  |
| 143     | 1                        |   |   |   |   |   |   |   |   |   |   |   |   |  |  |  |
| 144     | 0                        | 0 | 1 | 0 | 1 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 |  |  |  |
| 157     | 0                        | 0 | 0 | 0 | 0 | 0 | 0 | 1 | 0 | 0 | 1 | 1 | 0 |  |  |  |
| 170     | 0                        | 0 | 0 | 0 | 0 | 0 | 1 | 1 | 1 | 0 | 0 | 0 | 1 |  |  |  |
| 183     | 0                        | 1 | 0 | 1 | 0 | 0 | 0 | 1 | 0 |   |   |   |   |  |  |  |
| 192     | 0                        | 2 | 0 | 2 | 1 | 3 | 1 | 1 | 1 | 0 | 1 | 1 | 1 |  |  |  |
| 205     | 3                        | 0 | 0 | 0 | 3 | 0 | 2 | 3 | 3 | 3 | 2 | 0 | 1 |  |  |  |
| 218     | 0                        | 1 | 3 | 0 | 0 | 3 | 3 | 1 | 3 | 0 | 0 | 1 | 3 |  |  |  |
| 231     | 0                        | 2 | 1 | 1 | 1 | 2 | 1 | 1 | 2 | 2 | 2 | 2 | 2 |  |  |  |
| 244     | 2                        | 2 | 3 | 0 | 1 | 2 | 3 | 2 | 2 | 1 | 3 | 0 | 3 |  |  |  |
| 257     | 0                        | 0 | 2 | 3 | 1 | 2 | 2 | 0 | 3 | 3 | 3 | 1 | 1 |  |  |  |
| 270     | 3                        | 0 | 1 | 0 | 3 | 3 | 2 | 0 | 0 | 3 | 3 | 1 | 3 |  |  |  |
| 283     | 2                        | 3 | 0 | 3 | 2 | 0 | 0 | 3 | 1 | 1 | 2 | 2 | 3 |  |  |  |
| 296     | 0                        | 0 | 2 | 1 | 2 | 2 | 3 | 2 | 0 | 2 | 2 | 2 | 1 |  |  |  |
| 309     | 3                        | 3 | 3 | 2 | 0 | 1 | 0 | 0 | 1 | 2 | 1 | 3 | 2 |  |  |  |
| 322     | 3                        | 1 | 0 | 1 | 2 | 1 | 2 | 0 | 0 | 0 | 2 | 1 | 0 |  |  |  |
| 335     | 1                        | 0 | 0 | 0 | 0 | 1 | 1 | 1 | 2 | 0 | 1 | 2 | 1 |  |  |  |
| 348     | 2                        | 1 | 2 | 3 | 1 | 1 | 2 | 1 | 3 | 3 | 1 | 1 | 2 |  |  |  |
| 361     | 0                        | 3 | 1 | 3 | 3 | 1 | 2 | 3 | 1 | 1 | 3 | 2 | 1 |  |  |  |
| 374     | 1                        | 2 | 3 | 2 | 3 | 2 | 2 | 0 | 2 | 2 | 1 | 2 | 0 |  |  |  |
| 387     | 3                        | 2 | 2 | 1 | 3 | 3 | 2 | 0 | 0 | 0 | 2 | 3 | 2 |  |  |  |

| Item | Description   |
|------|---|
| 1    | Symbol number index (from beginning of packet or segment) of first Symbol data value on the line.   |
| 2    | Symbol values, with the value in column 2 of each row corresponding to the symbol number index in column 1, and then the symbol value in column 3 corresponding to the next symbol number index, etc., to the end of each row.<br><br>For example: The symbol number index 130 = 0 in column 2. The symbol number index for column 3 is 131, which equals 0. The symbol number index for column 4 is 132, which equals 1. This pattern continues to the end of the row. |

## WLAN Symbol Table Settings

### Front Panel: Settings

#### Application Toolbar:

The settings for the WLAN Symbol Table display are shown in the following table.

| Settings tab                                     | Description   |
|--|---|
| <a href="#">Modulation Params (see page 259)</a> | Specifies the type of modulation used for the input signal and other parameters.  |
| <a href="#">Analysis Params (see page 260)</a>   | Specifies parameters used by the instrument to analyze the input signal.  |
| <a href="#">Data Range (see page 262)</a>        | Specifies which symbols and subcarriers of the signal to display.   |
| <a href="#">Analysis Time (see page 263)</a>     | Specifies parameters that define the portion of the acquisition record that is used for analysis. Also allows you to specify the Units (Symbols or Seconds) for WLAN Analysis displays. |
| <a href="#">Prefs (see page 268)</a>             | Specifies the units of the display and whether elements of the graphs are displayed.  |

## WLAN Analysis Shared Measurement Settings

### Front Panel: Settings

#### Application Toolbar:


The control panel tabs in this section are shared between the displays in the WLAN Analysis folder (Setup > Displays). Some tabs are shared by all the displays, some tabs are shared by only a subset of displays. The settings available on some tabs change depending on the selected display.

### Common controls for WLAN analysis displays

| Settings tab                                     | Description   |
|--|---|
| <a href="#">Modulation Params (see page 259)</a> | Specifies the type of modulation used for the input signal and other parameters.  |
| <a href="#">Analysis Params (see page 260)</a>   | Specifies parameters used by the instrument to analyze the input signal.  |
| <a href="#">Data Range (see page 262)</a>        | Specifies which symbols and subcarriers of the signal to display.   |
| <a href="#">Analysis Time (see page 263)</a>     | Specifies parameters that define the portion of the acquisition record that is used for analysis. Also allows you to specify the Units (Symbols or Seconds) for WLAN Analysis displays. |
| <a href="#">Trace (see page 264)</a>             | Enables you to display or hide the measurement or average trace.  |
| <a href="#">Scale (see page 265)</a>             | Specifies the vertical, subcarrier (for non-b standards only), and symbols scale and position settings.   |
| <a href="#">EVM (see page 267)</a>               | Specifies the EVM units and max burst averages.   |
| <a href="#">Prefs (see page 268)</a>             | Specifies the units of the display and whether elements of the graphs are displayed.  |

## Modulation Params Tab - WLAN

The Modulation Params tab specifies the type of modulation used by the input signal and other parameters that define the signal format.



Modulation Params Analysis Params Data Range Analysis Time Trace Scale Prefs

Standard: 802.11g Guard Interval: Auto from SIG

Channel Bandwidth: 20 MHz Subcarrier Spacing: 312.500 kHz ☒ Auto

FFT Sample Rate: 20.000 MS/s FFT Length: 64

| Settings           | Description  |
|--------------------|--|
| Standard           | Specifies the standard used for the input signal. Choices are 802.11a/b/g/j/n/p/ac.  |
| Guard Interval     | Specifies the guard interval used in the input signal. You can select the following: <ul style="list-style-type: none"> <li>– Auto from SIG uses the Guard Interval value extracted from the signal</li> <li>– 1/8 allows setting the value manually</li> <li>– 1/4 allows setting the value manually</li> </ul> |
| Channel Bandwidth  | Specifies the nominal channel bandwidth. This setting affects the Subcarrier Spacing value when the Subcarrier Spacing Auto box is checked.  |
| Subcarrier Spacing | Specifies the spacing between subcarriers. When Auto is selected, this setting is automatically calculated according to the Channel Bandwidth value. If Auto is unchecked, a custom subcarrier spacing value can be entered for nonstandard signal definition.   |
| FFT Sample Rate    | Indicates the FFT sample rate, based on the bandwidth or subcarrier setting.   |
| FFT Length         | Indicates the fixed FFT length.  |

## Analysis Params Tab - WLAN

The Analysis Params tab contains parameters that control the analysis of the input signal.

The screenshot shows the 'Analysis Params' tab in a software interface. The tab is active, and the following parameters are visible:

- Burst Selection Index:** 1
- Frequency Error:** 0.000 Hz
- Symbol Analysis Offset:** -50 %
- Swap I & Q:** ☐
- Equalizer Training:** Preamble
- Data Modulation:** Auto Detect
- Pilot Tracking:**
  - ☒ Phase
  - ☐ Amplitude
  - ☐ Timing

| Settings               | Description  |
|------------------------|--|
| Burst Selection Index  | Allows specifying the burst you want to measure when multiple bursts are present in an acquisition. Index number of the first burst is 1, second burst is 2, etc.  |
| Frequency Error        | When the Auto box is checked, the analysis determines the Frequency Error and the measured Frequency Error is displayed. When the Auto box is unchecked, the entered value is used by the analysis as a fixed frequency offset. This is useful when the exact frequency offset of the signal is known.   |
| Symbol Analysis Offset | Specifies the analysis offset in the symbol interval. This value is a percentage of the Guard Interval. The allowable range is -100% to 0%. -100% positions the FFT to start at the beginning of the Guard Interval, 0% positions it to start at the end of the Guard Interval. Default value is -50% which usually gives the best measurement results.  |
| Swap I & Q             | Select the checkbox to swap the I and Q components of a signal. This compensates the input signal for spectral inversion.  |
| Equalizer Training     | Specifies the method used to estimate channel frequency response and equalization. This control can be used to diagnose changes in frequency response over the signal packet.  |
| Preamble               | The instrument uses only the Preamble to estimate channel frequency response. This response is then used to equalize the entire signal packet.   |
| Preamble + Data        | The instrument makes an initial channel frequency response estimate from the Preamble. It then estimates the channel response for each data symbol using the decoded data content to derive equalization for each symbol individually. This allows compensation for time-varying channel response over the packet.   |
| Data Modulation        | Allows choice of automatic or manual method of data symbol modulation identification, as follows: <ul style="list-style-type: none"> <li>– Auto Detect estimates the modulation from the data symbol IQ content.</li> <li>– Auto from SIG sets the modulation as indicated by the embedded SIG preamble symbol format data.</li> <li>– Manual allows specifying the modulation type regardless of the signal content.</li> </ul>   |
| Pilot tracking         | Specifies if pilot subcarriers should be used to correct amplitude, phase, and symbol timing variations over the packet. The choices available are Phase, Amplitude, and Timing. The default setting is Phase correction enabled, Amplitude and Timing correction disabled..   |
| Subcarrier derotation  | Allows some displays to show subcarriers with or without Gamma subcarrier phase rotation removed. Gamma phase rotation is applied to 802.11n and 802.11ac subcarriers in defined subranges depending on the selected Channel Bandwidth >40 MHz. Only Constellation and Symbol Table results are affected by this control. <p>When the box is unchecked, the rotation is not removed, which provides a direct view of the physical modulation on the channel.</p> <p>When the box is checked, the rotation is removed, allowing easier decoding of the underlying data content.</p> |

## Data Range Tab - WLAN

The Data Range tab enables you to control how much signal is analyzed, as well as specify the range of results that are displayed for the WLAN Constellation, EVM, Flatness, Magnitude Error, Phase Error, and Symbol Table graphs. The tab contents vary by standard.

Data Range tab for non-b standards.

| Settings               | Description   |
|------------------------|---|
| Symbols                | Specifies which symbols are displayed in the graphs.  |
| All                    | Select All to display all symbols.  |
| Range                  | Select Range to specify a subset of symbols for display.  |
| Start                  | Specifies the first symbol to include in the display.   |
| Stop                   | Specifies the last symbol to include in the display.  |
| Single                 | Select Single to display a single symbol.   |
| Index                  | Use Index to specify the symbol you want to display.  |
| Subcarriers            | Specifies which subcarriers are displayed.  |
| All                    | Select All to display all subcarriers in the signal.  |
| Pilots                 | Select Pilot to display only pilot subcarriers.   |
| Data                   | Select Data to display only data subcarriers.   |
| Single                 | Specifies a specific subcarrier for display.  |
| Index                  | Specifies the specific subcarrier to be displayed.  |
| Range                  | Specifies a range of subcarriers to be displayed.   |
| Start                  | Specifies the start value of the range to be displayed.   |
| Stop                   | Specifies the ending value of the range to be displayed.  |
| Max symbols to analyze | Specifies how many symbols are analyzed. You can use this setting to speed analysis by limiting the number of symbols being analyzed. |

Data Range tab for 802.11b standards.

| Settings                       | Description   |
|--------------------------------|---|
| Chips                          | Specifies which symbols are displayed in the graphs.  |
| All                            | Select All to display all chips.  |
| Preamble Only                  | Select to include only the Preamble in the display.   |
| Header Only                    | Select to include only the Header in the display.   |
| Data Only                      | Select to include only the Data in the display.   |
| Max number of chips to analyze | Specifies how many chips are analyzed. You can use this setting to speed analysis by limiting the number of chips being analyzed. Check this box and then enter the maximum number of chips to analyze. |

## Analysis Time Tab - WLAN

The Analysis Time tab contains parameters that define how the signal is analyzed in the WLAN Analysis displays.

The screenshot shows the 'Analysis Time' tab selected. It contains the following controls:

- Analysis Length:** A text box containing '200,000 Symbol'.
- Auto:** A checked checkbox.
- Units:** A dropdown menu currently showing 'Symbols'.

| Settings        | Description  |
|-----------------|--|
| Analysis Length | Specifies the length of the analysis period to use in measurements. Length is specified in either symbols or seconds, depending on the Units setting.<br>Use this to specify how long a period of time is analyzed.<br><ul style="list-style-type: none"> <li>– Range: minimum value depends on modulation type</li> <li>– Resolution: 1 symbol</li> </ul> |
| Auto            | When enabled, causes the instrument to set the Analysis Length value based on the requirements of the selected display.  |
| Units           | Specifies the units of the Analysis Length in either Symbols or Seconds.   |

## Trace Tab - WLAN

The Trace tab allows you to set the trace display characteristics of the selected display. This tab is not available for all WLAN displays.

The screenshot shows the 'Trace' tab selected in a software interface. It contains a 'Trace' dropdown menu set to 'Average', a 'Show' checkbox which is checked, and a 'Freeze' checkbox which is unchecked.

| Settings     | Description  |
|--------------|--|
| Trace        | Use this drop-down list to select whether or not the Average trace or the measurement trace is displayed. The measurement trace depends on the selected display. |
| Show Symbols | Select or deselect this checkbox to show or hide symbols.  |
| Freeze       | Selecting Freeze halts updates to the symbols.   |

## Traces Tab - WLAN Channel Response


The Traces tab allows you to set the trace display characteristics of the WLAN Channel Response display.

The screenshot shows the 'Traces' tab selected in a software interface. It contains a 'Trace' dropdown menu set to 'Magnitude', a 'Save Trace As...' button, a 'Recall Trace...' button, and a 'Show recalled trace' checkbox which is unchecked.

| Settings            | Description   |
|---------------------|---|
| Trace               | Selects the Magnitude or Phase trace for saving or recalling. |
| Save Trace As . . . | Select to save a trace.                                       |
| Recall Trace . . .  | Select to recall a trace.                                     |


## Scale Tab - WLAN

The Scale tab allows you to change the scale settings that control how the trace appears on the display but does not change control settings such as Measurement Frequency. There are four versions of the Scale tab for WLAN displays. One version is used for the Constellation display, one for the Spectral Flatness display, one for the Power vs Time display, and one for the EVM, Magnitude Error, Phase Error, and Channel Response displays.




The image shows the 'Scale' tab for the WLAN Power vs Time display. The tab bar at the top includes 'Modulation Params', 'Analysis Params', 'Data Range', 'Analysis Time', 'Scale', and 'Prefs'. The 'Scale' tab is active. It contains two main sections: 'Vertical' and 'Horizontal Full Burst'. The 'Vertical' section has a 'Scale' field set to '100.00 dB' and a 'Position' field set to '-11.00 dBm'. Below these fields are 'Reset' and 'Autoscale' buttons. The 'Horizontal Full Burst' section has a 'Scale' field set to '100.000 us' and a 'Position' field set to '0.000 s'. Below these fields are 'Autoscale' and 'Reset' buttons.

Scale tab for the WLAN Power vs Time display



The image shows the 'Scale' tab for the WLAN EVM, Magnitude Error, Phase Error, and Channel Response displays. The tab bar at the top includes 'Modulation Params', 'Analysis Params', 'Data Range', 'Analysis Time', 'Trace', 'Scale', and 'Prefs'. The 'Scale' tab is active. It contains three main sections: 'Vertical', 'Horizontal Symbol', and 'Horizontal Subcarrier'. The 'Vertical' section has a 'Scale' field set to '7.80 °' and a 'Position' field set to '-3.90 °'. Below these fields are 'Reset' and 'Autoscale' buttons. The 'Horizontal Symbol' section has a 'Start' field set to '0 Symbol' and a 'Stop' field set to '99 Symbol'. Below these fields are 'Reset' and 'Autoscale' buttons. The 'Horizontal Subcarrier' section has a 'Start' field set to '-58 Subcarrier' and a 'Stop' field set to '58 Subcarrier'. Below these fields are 'Reset' and 'Autoscale' buttons. There is also an 'Auto' checkbox which is checked.

Scale tab for WLAN EVM, Magnitude Error, Phase Error, and Channel Response displays



The image shows the 'Scale' tab for the WLAN Constellation display. The tab bar at the top includes 'Modulation Params', 'Analysis Params', 'Data Range', 'Analysis Time', 'Trace', 'Scale', and 'Prefs'. The 'Scale' tab is active. It contains a 'Zoom' section with a 'Scale' field set to '3'. Below the 'Scale' field are 'Vertical Position' and 'Horizontal Position' fields, both set to '0'. Below these fields are 'Reset' and 'Autoscale' buttons.

Scale tab for WLAN Constellation display


The screenshot shows the 'Scale' tab of the WLAN Spectral Flatness display. It contains two main sections: 'Vertical' and 'Horizontal'. The 'Vertical' section has a 'Scale' field set to '100.00 dB', a 'Position' field set to '5.00 dB', and an 'Autoscale' button. The 'Horizontal' section has a 'Scale' field set to '52 Subcarrier', a 'Position' field set to '-26 Subcarrier', and an 'Autoscale' button. Above these sections are tabs for 'Modulation Params', 'Analysis Params', 'Data Range', 'Analysis Time', 'Trace', 'Scale' (which is active), and 'Prefs'.

Scale tab for WLAN Spectral Flatness display

| Settings            | Description   |
|---------------------|---|
| Vertical            | Controls the vertical position and scale of the trace display.  |
| Scale               | Changes the vertical scale of the graph.  |
| Position            | Adjusts the reference level away from top of the graph.   |
| Autoscale           | Resets the scale of the vertical axis to contain the complete trace.  |
| Subcarrier          | Controls the scale of the subcarrier graph and position of the trace.   |
| Scale               | Specifies how many subcarriers are displayed horizontally.  |
| Position            | Specifies the subcarrier that appears at the left edge of the subcarrier graph.   |
| Autoscale           | Resets the scale of the horizontal axis to contain the complete trace.  |
| Symbols             | Controls the scale of the Symbols graph.  |
| Scale               | Specifies the number of symbols that appear in the symbols graph.   |
| Position            | Specifies the symbol that appears at the left edge of the Symbols graph.  |
| Autoscale           | Resets the scale and position settings to optimize the display.   |
| Auto                | When Auto is checked, the scale and position values for the Symbols graph are automatically adjusted to maintain the optimal display. |
| Zoom                | Sets the Constellation display size and position.   |
| Scale               | Sets the magnification value for the Constellation display.   |
| Vertical Position   | Sets the vertical location of the Constellation display within the graph. Range is -3.7 to +3.7.                                      |
| Horizontal Position | Sets the horizontal location of the Constellation display within the graph. Range is -3.7 to +3.7.                                    |

## EVM Tab - WLAN

The EVM Tab enables you to choose between dB and percent for the EVM units and to select to set the maximum bursts to average. The EVM Tab is only available for the WLAN Summary display.




The screenshot shows the EVM Tab interface with the following settings:

- EVM Units:** A dropdown menu set to **dB**.
- Max Bursts to Avg:** A text input field containing the value **20**.
- Count:** An unchecked checkbox.

| Settings          | Description   |
|-------------------|---|
| EVM units         | Specifies whether the displayed EVM units are dB or percent.              |
| Max Bursts to Avg | Specifies the maximum number of bursts to average when Count is selected. |
| Count             | Enables the Max Bursts to Avg function.                                   |

## Prefs Tab - WLAN

The Prefs tab enables you to change appearance characteristics of the WLAN Analysis displays. Not all settings on the Prefs tab shown below appear for every WLAN display.



| Settings            | Description  |
|---------------------|--|
| Time units          | Specifies whether the displayed time units are seconds or symbols.   |
| Freq units          | Specifies whether the displayed frequency units are frequency (Hz) or subcarrier channel.  |
| Radix               | Specifies whether symbol values are displayed in binary or hex format (for example, in the Symbol Table or markers readouts).  |
| Show graphs         | Specifies which graphs are displayed. Select from Symbols, Subcarriers or Both, Magnitude, Phase, or Both. Disabled (no masks), Shaded Region, or Line Only. These selections are only available for OFDM signals. |
| Show graticule      | Specifies to show the graticule on the display when checked.   |
| Show marker readout | Specifies to show the marker readout when checked.   |

## Overview

The OFDM Analysis option provides measurements for OFDM signals specified by 802.11a/g/j (Wifi) and 802.16 (2004) (commonly known as Fixed WiMAX) standards. User controls allow you to modify signal parameters for analysis of non-standard signals. The analysis results give multiple views of OFDM signal characteristics to allow diagnosing signal imperfections and impairments quickly and easily. Display controls provide the ability to selectively display the analysis results to help locate trouble-spots in the signal.

The displays in OFDM Measurements (Displays > Measurements > OFDM Measurements) are:

- [OFDM Channel Response \(see page 269\)](#)
- [OFDM Constellation \(see page 272\)](#)
- [OFDM EVM \(see page 273\)](#)
- [OFDM Spectral Flatness \(see page 275\)](#)
- [OFDM Mag Error \(see page 277\)](#)
- [OFDM Phase Error \(see page 279\)](#)
- [OFDM Power \(see page 281\)](#)
- [OFDM Summary \(see page 283\)](#)
- [OFDM Symbol Table \(see page 285\)](#)

## OFDM Channel Response Display



The OFDM Channel Response display plots the channel response (magnitude or phase) versus the subcarrier or frequency. Here, the channel refers to all sources of signal frequency response impairment up to the analyzer input, including the transmitter itself, as well as any transmission medium through which the signal travels between the transmitter and the analyzer.

To show the OFDM Channel Response display:

1. If you are analyzing a stored data file, recall the data file now. Otherwise, go to the next step.
2. Press the **Displays** button or select **Setup > Displays**.
3. In the **Select Displays** dialog, select **OFDM Analysis** in the **Measurements** box.
4. In the Available displays box, double-click the **OFDM Chan Response** icon or select the icon and click **Add**. The OFDM Chan Response icon will appear in the **Selected displays** box and will no longer appear under Available displays.
5. Click **OK** to show the OFDM Chan Response display.
6. Set the Frequency appropriate for the signal.

7. Select **Setup > Settings** to display the control panel.
8. Select the Modulation Params tab. Set the Standard, Guard Interval and Spacing and Bandwidth controls as appropriate for the input signal.

### Elements of the Display



| Item | Display element                     | Description   |
|------|-------------------------------------|---|
| 1    | Top-of-graph (magnitude)            | Sets the level that appears at the top of the magnitude graph. This is only a visual control for panning the graph. The Reference Level is adjusted in the Toolbar and the Ampl control panel. By default, Vert Position = Ref Level.   |
| 2    | dB/div (magnitude)                  | Sets the vertical scale value. The maximum value is 20.00 dB/division.  |
| 3    | Display selector                    | Selects the display type. Channel Response Magnitude or Phase can be displayed as a Magnitude or Phase graph.   |
| 4    | Bottom-of-graph readout (magnitude) | Indicates the amplitude at the bottom of the graph. This value changes with the dB/div and Vertical Position settings.  |
| 5    | Autoscale                           | Adjusts the Vertical and Horizontal scaling to display the optimize the trace display on screen.  |
| 6    | Pos                                 | Shifts the trace left or right in the graph. The readout indicates the subcarrier or frequency shown at the left edge of the display.   |
| 7    | Scale                               | Specifies the number of subcarriers shown in the graph.   |
| 8    | Top-of-graph (phase)                | Sets the phase value indicated at the top of the graph. Since the Position value at the vertical center of this graph remains constant as the Top of Graph value is adjusted, the Vertical Scale increases as the Top of Graph value increases, which also affects the bottom of graph readout. |
| 9    | Position (phase)                    | Specifies the phase shown at the center of the graph display. Changing this value moves the trace up and down in the graph, which affects the Top of Graph and Bottom of Graph values as well.  |
| 10   | Bottom-of-graph (phase)             | Indicates the phase at the bottom of the graph. This value changes with the Position setting.   |

## OFDM Channel Response Settings

The OFDM Channel Response Settings control panel provides access to settings that control parameters of the Channel Response Display.

| Settings tab                                     | Description  |
|--|--|
| <a href="#">Modulation Params (see page 287)</a> | Specifies the input signal standard and additional user-settable signal parameters.  |
| <a href="#">Advanced Params (see page 288)</a>   | Specifies parameters used by the instrument to analyze the input signal.   |
| <a href="#">Data Range Tab (see page 289)</a>    | Specifies which symbols and subcarriers of the signal to display.  |
| <a href="#">Analysis Time Tab (see page 289)</a> | Specifies parameters that define the portion of the acquisition record that is used for analysis. Also allows you to specify the time units (Symbols or Seconds) for OFDM Analysis displays. |
| <a href="#">Prefs Tab (see page 293)</a>         | Specifies the units of the display and whether elements of the graphs are displayed.   |


## OFDM Constellation Display

Display shows the OFDM signal subcarriers' amplitude and phase in IQ constellation form.

To show the OFDM Constellation display:

1. If you are analyzing a stored data file, recall the data file now. Otherwise, go to the next step.
2. Press the **Displays** button or select **Setup > Displays**.
3. In the **Select Displays** dialog, select **OFDM Constellation** in the **Measurements** box.
4. In the Available displays box, double-click the **OFDM Constellation** icon or select the icon and click **Add**. The OFDM Constellation icon will appear in the **Selected displays** box and will no longer appear under Available displays.
5. Click **OK** to show the OFDM Constellation display.
6. Set the **Frequency** appropriate for the signal.
7. Select **Setup > Settings** to display the control panel.
8. Select the **Modulation Params** tab. Set the Standard, Guard Interval and Spacing and Bandwidth controls as appropriate for the input signal.

### Elements of the Display



| Item | Display element | Description  |
|------|-----------------|--|
| 1    | Plot            | Constellation graph.   |
| 2    | Marker Readout  | If markers are enabled, the marker readout shows the selected Marker, Symbol, Subcarrier, Type, Value, Magnitude, and Phase for the symbol at the marker location. Located to the left of the constellation plot or below it, depending on the size of the window. |

## OFDM Constellation Settings

### Front Panel: Settings

#### Application Toolbar:

The OFDM Constellation Settings control panel provides access to settings that control parameters of the Constellation Display.

| Settings tab                                     | Description   |
|--|---|
| <a href="#">Modulation Params (see page 287)</a> | Specifies the input signal standard and additional user-settable signal parameters.   |
| <a href="#">Advanced Params (see page 288)</a>   | Specifies parameters used by the instrument to analyze the input signal.  |
| <a href="#">Data Range (see page 289)</a>        | Specifies which symbols and subcarriers of the signal to display.   |
| <a href="#">Analysis Time Tab (see page 289)</a> | Specifies parameters that define the portion of the acquisition record that is used for analysis. Also allows you to specify the Units (Symbols or Seconds) for OFDM Analysis displays. |
| <a href="#">Trace (see page 291)</a>             | Enables you to freeze the display or hide the measurement or average trace.   |
| <a href="#">Scale Tab (see page 291)</a>         | Specifies the Zoom scale, and vertical and horizontal positions of the display.   |
| <a href="#">Prefs Tab (see page 293)</a>         | Specifies the units of the display and whether elements of the graphs are displayed.  |

## OFDM EVM Display


The OFDM EVM display shows the data symbols' individual subcarrier Error Vector Magnitude values versus symbol interval (time) and subcarrier (frequency).

To show the OFDM EVM display:

1. If you are analyzing a stored data file, recall the data file now. Otherwise, go to the next step.
2. Press the **Displays** button or select **Setup > Displays**.
3. In the **Select Displays** dialog, select **OFDM EVM** in the **Measurements** box.
4. In the Available displays box, double-click the **OFDM EVM** icon or select the icon and click **Add**. The OFDM EVM icon will appear in the **Selected displays** box and will no longer appear under Available displays.

5. Click **OK** to show the OFDM EVM display.
6. Set the **Frequency** appropriate for the signal.
7. Select **Setup > Settings** to display the control panel.
8. Select the **Modulation Params** tab. Set the Standard as appropriate. Set the Standard, Guard Interval and Spacing and Bandwidth controls as appropriate for the input signal.

### Elements of the Display



| Item | Display element | Description   |
|------|-----------------|---|
| 1    | Top of graph    | Sets the EVM value that appears at the top of the graph. This is only a visual control for panning the graph.   |
| 2    | Bottom of graph | Sets the EVM value that appears at the bottom of the graph. This is only a visual control for panning the graph.  |
| 3    | Autoscale       | Adjusts the vertical and horizontal settings to provide the best display.   |
| 4    | Pos             | Specifies the horizontal position of the trace on the graph display.  |
| 5    | Scale           | Adjusts the horizontal range of the graph. By decreasing the scale, the graph essentially becomes a window that you can move over the analysis results by adjusting the position. |

## OFDM EVM Settings

### Front Panel: Settings

#### Application Toolbar:

The settings for the OFDM EVM display are shown in the following table.

| Settings tab                                     | Description   |
|--|---|
| <a href="#">Modulation Params (see page 287)</a> | Specifies the type of modulation used for the input signal and other parameters.  |
| <a href="#">Advanced Params (see page 288)</a>   | Specifies parameters used by the instrument to analyze the input signal.  |
| <a href="#">Data Range (see page 289)</a>        | Specifies which symbols and subcarriers of the signal to display.   |
| <a href="#">Analysis Time (see page 289)</a>     | Specifies parameters that define the portion of the acquisition record that is used for analysis. Also allows you to specify the Units (Symbols or Seconds) for OFDM Analysis displays. |
| <a href="#">Trace (see page 291)</a>             | Enables you to freeze the display or hide the measurement or average trace.   |
| <a href="#">Scale (see page 291)</a>             | Specifies the vertical, subcarrier, and symbols scale and position settings.  |
| <a href="#">Prefs (see page 293)</a>             | Specifies the units of the display and whether elements of the graphs are displayed.  |

## OFDM Spectral Flatness Display


The OFDM Spectral Flatness display shows the average power levels of subcarriers across the signal bandwidth and indicates if they remain within the limits defined for a particular standard.

The averaged subcarrier level is shown as a line on the display, while the individual subcarrier levels are shown as points. The Pass or Fail result indicates whether the average line remains between the upper and lower limit mask boundaries.

To show the OFDM Spectral Flatness display:

1. If you are analyzing a stored data file, recall the data file now. Otherwise, go to the next step.
2. Press the **Displays** button or select **Setup > Displays**.
3. In the **Select Displays** dialog, select **OFDM Analysis** in the **Measurements** box.
4. In the Available displays box, double-click the **OFDM Flatness** icon or select the icon and click **Add**. The OFDM Flatness icon will appear in the **Selected displays** box and will no longer appear under Available displays.
5. Click **OK** to show the OFDM Flatness display.
6. Set the **Frequency** appropriate for the signal.
7. Select **Setup > Settings** to display the control panel.
8. Select the **Modulation Params** tab. Set the Standard, Guard Interval, Subcarrier Spacing and Channel Bandwidth controls as appropriate for the input signal.

## Elements of the Display



| Item | Display element | Description   |
|------|-----------------|---|
| 1    | Top of graph    | Sets the Flatness value that appears at the bottom of the graph in dB. This is only a visual control for panning the graph.   |
| 2    | Bottom of graph | Shows the Flatness value set at the top of the graph.   |
| 3    | Autoscale       | Adjusts the vertical and horizontal settings to provide the best display.   |
| 4    | Pos             | Specifies the horizontal position of the trace on the graph display.  |
| 5    | Scale           | Adjusts the horizontal range of the graph. By decreasing the scale, the graph essentially becomes a window that you can move over the analysis results by adjusting the position. |

## OFDM Spectral Flatness Settings

### Front Panel: Settings

#### Application Toolbar:

The settings for the OFDM Spectral Flatness display are shown in the following table.

| Settings tab                                     | Description   |
|--|---|
| <a href="#">Modulation Params (see page 287)</a> | Specifies the type of modulation used for the input signal and other parameters.  |
| <a href="#">Advanced Params (see page 288)</a>   | Specifies parameters used by the instrument to analyze the input signal.  |
| <a href="#">Data Range (see page 289)</a>        | Specifies which symbols and subcarriers of the signal to display.   |
| <a href="#">Analysis Time Tab (see page 289)</a> | Specifies parameters that define the portion of the acquisition record that is used for analysis. Also allows you to specify the Units (Symbols or Seconds) for WLAN Analysis displays. |
| <a href="#">Trace (see page 291)</a>             | Enables you to display or hide the measurement or average trace.  |
| <a href="#">Scale Tab (see page 291)</a>         | Specifies the vertical, subcarrier, and symbols scale and position settings.  |
| <a href="#">Prefs Tab (see page 293)</a>         | Specifies the units of the display and whether elements of the graphs are displayed.  |


## OFDM Magnitude Error Display

The OFDM Mag Error display shows the data symbols' individual subcarrier Magnitude Error values versus symbol interval (time) and subcarrier (frequency).

To show the OFDM Magnitude Error display:

1. If you are analyzing a stored data file, recall the data file now. Otherwise, go to the next step.
2. Press the **Displays** button or select **Setup > Displays**.
3. In the **Select Displays** dialog, select **OFDM Analysis** in the **Measurements** box.
4. In the Available displays box, double-click the **OFDM Mag Error** icon or select the icon and click **Add**. The OFDM Mag Error icon will appear in the **Selected displays** box and will no longer appear under Available displays.
5. Click **OK** to show the OFDM Mag Error display.
6. Set the **Frequency** appropriate for the signal.
7. Select **Setup > Settings** to display the control panel.
8. Select the **Modulation Params** tab. Set the Standard, Guard Interval and Spacing and Bandwidth controls as appropriate for the input signal.

## Elements of the Display



| Item | Display element | Description   |
|------|-----------------|---|
| 1    | Top of graph    | Sets the Magnitude Error value that appears at the top of the graph. This is only a visual control for panning the graph.   |
| 2    | Bottom of graph | Sets the Magnitude Error value that appears at the bottom of the graph. This is only a visual control for panning the graph.  |
| 3    | Autoscale       | Adjusts the vertical and horizontal settings to provide the best display.   |
| 4    | Pos             | Specifies the horizontal position of the trace on the graph display.  |
| 5    | Scale           | Adjusts the horizontal range of the graph. By decreasing the scale, the graph essentially becomes a window that you can move over the analysis results by adjusting the position. |

## OFDM Magnitude Error Settings

### Front Panel: Settings

#### Application Toolbar:

The settings for the OFDM Mag Error display are shown in the following table.

| Settings tab                                     | Description   |
|--|---|
| <a href="#">Modulation Params (see page 287)</a> | Specifies the type of modulation used for the input signal and other parameters.  |
| <a href="#">Advanced Params (see page 288)</a>   | Specifies parameters used by the instrument to analyze the input signal.  |
| <a href="#">Data Range (see page 289)</a>        | Specifies which symbols and subcarriers of the signal to display.   |
| <a href="#">Analysis Time Tab (see page 289)</a> | Specifies parameters that define the portion of the acquisition record that is used for analysis. Also allows you to specify the Units (Symbols or Seconds) for OFDM Analysis displays. |
| <a href="#">Trace (see page 291)</a>             | Enables you to display or hide the measurement or average trace.  |
| <a href="#">Scale Tab (see page 291)</a>         | Specifies the vertical, subcarrier, and symbols scale and position settings.  |
| <a href="#">Prefs Tab (see page 293)</a>         | Specifies the units of the display and whether elements of the graphs are displayed.  |


## OFDM Phase Error Display

The OFDM Phase Error display shows the data symbols' individual subcarrier Phase Error values versus symbol interval (time) and subcarrier (frequency).

To show the OFDM Phase Error display:

1. If you are analyzing a stored data file, recall the data file now. Otherwise, go to the next step.
2. Press the **Displays** button or select **Setup > Displays**.
3. In the **Select Displays** dialog, select **OFDM Analysis** in the **Measurements** box.
4. In the Available displays box, double-click the **OFDM Phase Error** icon or select the icon and click **Add**. The OFDM Phase Error icon will appear in the **Selected displays** box and will no longer appear under Available displays.
5. Click **OK** to show the OFDM Phase Error display.
6. Set the **Frequency** appropriate for the signal.
7. Select **Setup > Settings** to display the control panel.
8. Select the **Modulation Params** tab. Set the Standard, Guard Interval and Spacing and Bandwidth controls as appropriate for the input signal.

## Elements of the Display



| Item | Display element | Description   |
|------|-----------------|---|
| 1    | Top of graph    | Sets the Phase Error value that appears at the top of the graph. This is only a visual control for panning the graph.   |
| 2    | Bottom of graph | Sets the Phase Error value that appears at the bottom of the graph. This is only a visual control for panning the graph.  |
| 3    | Autoscale       | Adjusts the vertical and horizontal settings to provide the best display.   |
| 4    | Pos             | Specifies the horizontal position of the trace on the graph display.  |
| 5    | Scale           | Adjusts the horizontal range of the graph. By decreasing the scale, the graph essentially becomes a window that you can move over the analysis results by adjusting the position. |

## OFDM Phase Error Settings

### Front Panel: Settings

#### Application Toolbar:

The settings for the OFDM Phase Error display are shown in the following table.

| Settings tab                                     | Description   |
|--|---|
| <a href="#">Modulation Params (see page 287)</a> | Specifies the type of modulation used for the input signal and other parameters.  |
| <a href="#">Advanced Params (see page 288)</a>   | Specifies parameters used by the instrument to analyze the input signal.  |
| <a href="#">Data Range (see page 289)</a>        | Specifies which symbols and subcarriers of the signal to display.   |
| <a href="#">Analysis Time Tab (see page 289)</a> | Specifies parameters that define the portion of the acquisition record that is used for analysis. Also allows you to specify the Units (Symbols or Seconds) for OFDM Analysis displays. |
| <a href="#">Trace (see page 291)</a>             | Enables you to display or hide the measurement or average trace.  |
| <a href="#">Scale Tab (see page 291)</a>         | Specifies the vertical, subcarrier, and symbols scale and position settings.  |
| <a href="#">Prefs Tab (see page 293)</a>         | Specifies the units of the display and whether elements of the graphs are displayed.  |


## OFDM Power Display

The OFDM Power display shows the data symbols' individual subcarrier Power values versus symbol interval (time) and subcarrier (frequency).

To show the OFDM Power display:

1. If you are analyzing a stored data file, recall the data file now. Otherwise, go to the next step.
2. Press the **Displays** button or select **Setup > Displays**.
3. In the **Select Displays** dialog, select **OFDM Analysis** in the **Measurements** box.
4. In the Available displays box, double-click the **OFDM Power** icon or select the icon and click **Add**. The OFDM Power icon will appear in the **Selected displays** box and will no longer appear under Available displays.
5. Click **OK** to show the OFDM Power display.
6. Set the **Frequency** appropriate for the signal.
7. Select **Setup > Settings** to display the control panel.
8. Select the **Modulation Params** tab. Set the Standard, Guard Interval and Spacing and Bandwidth controls as appropriate for the input signal.
9. If you are analyzing a stored data file, press the **Replay** button to take measurements on the recalled acquisition data file.

### Elements of the Display



| Item | Display element         | Description   |
|------|-------------------------|---|
| 1    | Vertical Position       | Sets the top of graph value. This is only a visual control for panning the graph. The Reference Level is adjusted in the Toolbar and the Ampl control panel. By default, Vert Position = Ref Level. |
| 2    | dB/div                  | Sets the vertical scale value. The maximum value is 20.00 dB/division.  |
| 2    | Bottom of graph readout | Indicates the amplitude at the bottom of the graph. This value changes with the dB/div and Vertical Position settings.  |
| 3    | Autoscale               | Adjusts the vertical and horizontal settings to provide the best display.   |
| 4    | Pos                     | Specifies the horizontal position of the trace on the graph display.  |
| 5    | Scale                   | Adjusts the horizontal range of the graph. By decreasing the scale, the graph essentially becomes a window that you can move over the analysis results by adjusting the position.                   |

## OFDM Power Settings

### Front Panel: Settings

#### Application Toolbar:

The settings for the OFDM Power display are shown in the following table.

| Settings tab                                     | Description   |
|--|---|
| <a href="#">Modulation Params (see page 287)</a> | Specifies the type of modulation used for the input signal and other parameters.  |
| <a href="#">Advanced Params (see page 288)</a>   | Specifies parameters used by the instrument to analyze the input signal.  |
| <a href="#">Data Range (see page 289)</a>        | Specifies which symbols and subcarriers of the signal to display.   |
| <a href="#">Analysis Time (see page 289)</a>     | Specifies parameters that define the portion of the acquisition record that is used for analysis. Also allows you to specify the Units (Symbols or Seconds) for OFDM Analysis displays. |
| <a href="#">Trace (see page 291)</a>             | Enables you to display or hide the measurement or average trace.  |
| <a href="#">Scale (see page 291)</a>             | Specifies the vertical, subcarrier, and symbols scale and position settings.  |
| <a href="#">Prefs (see page 293)</a>             | Specifies the units of the display and whether elements of the graphs are displayed.  |


## OFDM Summary Display

The OFDM Summary display shows several measurements of OFDM signal quality.

To show the OFDM Summary display:

1. If you are analyzing a stored data file, recall the data file now. Otherwise, go to the next step.
2. Press the **Displays** button or select **Setup > Displays**.
3. In the **Select Displays** dialog, select **OFDM Analysis** in the **Measurements** box.
4. In the Available displays box, double-click the **OFDM Summary** icon or select the icon and click **Add**. The OFDM Summary icon will appear in the **Selected displays** box and will no longer appear under Available displays.
5. Click **OK** to show the OFDM Summary display.
6. Set the **Frequency** appropriate for the signal.
7. Select **Setup > Settings** to display the control panel.
8. Select the **Modulation Params** tab. Set the Standard, Guard Interval and Spacing and Bandwidth controls as appropriate for the input signal.

### Elements of the Display



|                             |            |                          |             |                              |  |
|-----------------------------|------------|--------------------------|-------------|------------------------------|--|
| Frequency: 5.90000 GHz      |            | RefLev: 0.00 dBm         |             | Preset   Replay   Run        |  |
| <b>802.11a/g/j</b>          |            |                          |             |                              |  |
| Frequency Error: 1.75 Hz    |            | Symbols: 13              |             | Symbol Clk Error: -0.502 ppm |  |
| IQ Origin Offset: -74.31 dB |            | CPE: 1.509 %             |             |                              |  |
| Average Power: -24.31 dBm   |            | Peak-to-Average: 9.53 dB |             |                              |  |
| <b>Subcarriers:</b>         | <b>All</b> | <b>Pilots</b>            | <b>Data</b> |                              |  |
| RMS EVM:                    | -35.17 dB  | -36.43 dB                | -35.08 dB   |                              |  |
|                             | 1.744 %    | 1.509 %                  | 1.763 %     |                              |  |
| Peak EVM:                   | -23.75 dB  | -27.80 dB                | -23.75 dB   |                              |  |
|                             | 6.492 %    | 4.074 %                  | 6.492 %     |                              |  |
| at symbol:                  | 1          | 5                        | 1           |                              |  |
| at subcarrier:              | 26         | 21                       | 26          |                              |  |

| Measurement      | Description  |
|------------------|--|
| Standard         | Display of the standard selected on the Setup > Settings > Modulation Params tab.  |
| Frequency Error  | The frequency difference between the measured carrier frequency of the signal and the measurement frequency setting.   |
| IQ Origin Offset | The average magnitude of the DC subcarrier level relative to total signal power. It indicates the level of carrier feedthrough detected at the center (DC) subcarrier.   |
| Average Power    | The average power of all symbols in the analysis. Calculated over only the data symbols in the packet.   |
| Symbols          | How many symbols were analyzed.  |
| Symbol Clk Error | The symbol clock error in parts per million.   |
| CPE              | CPE, Common Pilot Error, is the RMS magnitude error of the pilots over all analyzed symbols.   |
| Peak-to-Average  | The ratio of the highest instantaneous signal power level to the average signal power. Calculated over only the data symbols in the packet.  |
| EVM              | <p>The RMS and Peak values of the normalized subcarrier Error Vector Magnitude values. The normalized subcarrier EVM values are calculated as the difference between the detected received signal subcarrier constellation points and ideal reference points estimated by the instrument from the received signal. Values are reported in units of percent and dB. Peak values include the symbol and subcarrier location.</p> <p>RMS and Peak values are displayed for groupings of all subcarriers, Pilots only and Data only. Results are calculated over the entire signal packet analyzed, covering the number of Symbols indicated in the Summary display.</p> |

## OFDM Summary Settings

### Front Panel: Settings

#### Application Toolbar:

The settings for the OFDM Summary display are shown in the following table.


| Settings tab                                     | Description   |
|--|---|
| <a href="#">Modulation Params (see page 287)</a> | Specifies the type of modulation used for the input signal and other parameters.  |
| <a href="#">Advanced Params (see page 288)</a>   | Specifies parameters used by the instrument to analyze the input signal.  |
| <a href="#">Data Range (see page 289)</a>        | Specifies which symbols and subcarriers of the signal to display.   |
| <a href="#">Analysis Time (see page 289)</a>     | Specifies parameters that define the portion of the acquisition record that is used for analysis. Also allows you to specify the Units (Symbols or Seconds) for OFDM Analysis displays. |
| <a href="#">Prefs (see page 293)</a>             | Specifies the units of the display and whether elements of the graphs are displayed.  |

## OFDM Symbol Table Display

The OFDM Symbol Table display shows decoded data bits for each subcarrier in each symbol in the analyzed signal packet. Results are presented with subcarrier (frequency) indices in the horizontal dimension and symbol (time) intervals in the vertical dimension.

To show the OFDM Symbol Table display:

1. If you are analyzing a stored data file, recall the data file now. Otherwise, go to the next step.
2. Press the **Displays** button or select **Setup > Displays**.
3. In the **Select Displays** dialog, select **OFDM Analysis** in the **Measurements** box.
4. In the Available displays box, double-click the **OFDM Symb Table** icon or select the icon and click **Add**. The OFDM Symbol icon will appear in the **Selected displays** box and will no longer appear under Available displays.
5. Click **OK** to show the OFDM Symbol Table display.
6. Set the **Frequency** appropriate for the signal.
7. Select **Setup > Settings** to display the control panel.
8. Select the **Modulation Params** tab. Set the Standard, Guard Interval and Spacing and Bandwidth controls as appropriate for the input signal.



| Item | Description             |
|------|-------------------------|
| 1    | Subcarrier identifiers. |
| 2    | Symbol identifiers.     |

(cont.)

| Item | Description                              |
|------|--|
| 3    | Subcarrier data values.                  |
| 4    | Marker readout when markers are enabled. |

## OFDM Symbol Table Settings

**Front Panel: Settings**

**Application Toolbar:** 

The settings for the OFDM Symbol Table display are shown in the following table.

| Settings tab                                     | Description   |
|--|---|
| <a href="#">Modulation Params (see page 287)</a> | Specifies the type of modulation used for the input signal and other parameters.  |
| <a href="#">Advanced Params (see page 288)</a>   | Specifies parameters used by the instrument to analyze the input signal.  |
| <a href="#">Data Range (see page 289)</a>        | Specifies which symbols and subcarriers of the signal to display.   |
| <a href="#">Analysis Time (see page 289)</a>     | Specifies parameters that define the portion of the acquisition record that is used for analysis. Also allows you to specify the Units (Symbols or Seconds) for OFDM Analysis displays. |
| <a href="#">Prefs (see page 293)</a>             | Specifies the units of the display and whether elements of the graphs are displayed.  |

## OFDM Analysis Shared Measurement Settings

**Front Panel: Settings**

**Application Toolbar:** 

The control panel tabs in this section are shared between the displays in the OFDM Analysis folder (Setup > Displays). Some tabs are shared by all the displays, some tabs are shared by only a couple of displays. The settings available on some tabs change depending on the selected display.

## Common controls for OFDM analysis displays

| Settings tab                                     | Description   |
|--|---|
| <a href="#">Modulation Params (see page 287)</a> | Specifies the type of modulation used for the input signal and other parameters.  |
| <a href="#">Advanced Params (see page 288)</a>   | Specifies parameters used by the instrument to analyze the input signal.  |
| <a href="#">Data Range (see page 289)</a>        | Specifies which symbols and subcarriers of the signal to display.   |
| <a href="#">Analysis Time (see page 289)</a>     | Specifies parameters that define the portion of the acquisition record that is used for analysis. Also allows you to specify the Units (Symbols or Seconds) for OFDM Analysis displays. |
| <a href="#">Trace (see page 291)</a>             | Enables you to display or hide the measurement or average trace.  |
| <a href="#">Scale (see page 291)</a>             | Specifies the vertical, subcarrier, and symbols scale and position settings.  |
| <a href="#">Prefs (see page 293)</a>             | Specifies the units of the display and whether elements of the graphs are displayed.  |

## Modulation Params Tab - OFDM

The Modulation Params tab specifies the type of modulation used by the input signal and other parameters that control the demodulation of the input signal.

| Settings              | Description  |
|-----------------------|--|
| Standard              | Specifies the standard used for the input signal. Choices are 802.11a/g/j and 802.16 (2004).   |
| Guard Interval        | Specifies the guard interval used in the input signal. Choices are 1/4, 1/8, 1/16, 1/32, and User. When you select User, you can enter a percentage value to specify the size of the guard interval as a percentage of the useful symbol interval. The Guard Interval range is 0–100%. |
| Spacing and Bandwidth | Specifies the spacing between subcarriers and the nominal channel bandwidth. Only one of these settings can be set at a time, the other setting is automatically calculated.   |
| Sample Rate           | Indicates the FFT sample rate, based on the bandwidth or subcarrier setting.   |

## Advanced Params Tab - OFDM

The Advanced Params tab contains parameters that control the analysis of the input signal.

The screenshot shows the 'Advanced Params' tab with the following settings:

- Constellation determination:** Auto Detect
- Channel Estimation:** Preamble
- Pilot tracking:**
  - ☒ Phase
  - ☐ Amplitude
  - ☐ Timing
- Symbol analysis offset:** -0.5 %
- Swap I & Q:** ☐

### Settings

### Description

|                             |   |
|-----------------------------|---|
| Constellation determination | Specifies the method used to detect the constellation. Auto Detect analyzes the signal to determine the appropriate constellation, Manual allows you to specify the constellation type. Manual choices are: BPSK, QPSK, 16QAM, and 64QAM.   |
| Symbol analysis offset      | Specifies the analysis offset in the symbol interval. This value is a percentage of the Guard Interval. The allowable range is -100% to 0%. -100% positions the FFT to start at the beginning of the Guard Interval, 0% positions it to start at the end of the Guard Interval. Default value is -50% which usually gives the best measurement results. |
| Channel Estimation          | Specifies the method used to estimate channel frequency response and equalization. This control can be used to diagnose changes in frequency response over the signal packet.   |
| Preamble                    | When you select Preamble, the instrument uses only the Preamble to estimate channel frequency response. This response is then used to equalize the entire signal packet.  |
| Preamble + Data             | When you select Preamble + Data, the instrument makes an initial channel frequency response estimate from the Preamble. Then estimates the channel response for each data symbol using the decoded data content to derive equalization for each symbol individually. This allows compensation for time-varying channel response over the packet.        |
| Pilot tracking              | Specifies if pilot subcarriers should be used to correct amplitude, phase, and symbol timing variations over the packet. The choices available are Phase, Amplitude, and Timing. The default setting is Phase.  |
| Swap I & Q                  | Select the checkbox to swap the I and Q components of a signal. This compensates the input signal for spectral inversion.   |

## Data Range Tab - OFDM

The Data Range tab enables you to control how much signal is analyzed, as well as specify the range of results that are displayed in the Constellation, EVM, Magnitude Error, Phase Error, Power and Symbol Table graphs.

The screenshot shows the 'Data Range' tab selected in the software interface. It contains two dropdown menus: 'Symbols' and 'Subcarriers', both currently set to 'All'. Below these, there is a checkbox labeled 'Max symbols to analyze' which is checked, and a text box next to it containing the value '1004'.

| Settings               | Description   |
|------------------------|---|
| <b>Symbols</b>         | Specifies which symbols are displayed in the graphs.  |
| All                    | Select All to display all symbols.  |
| Range                  | Select Range to specify a subset of symbols for display.  |
| Start                  | Specifies the first symbol to include in the display.   |
| Number                 | Specifies how many symbols to display in the graphs.  |
| Single                 | Select Single to display a single symbol.   |
| Index                  | Use Index to specify the symbol you want to display.  |
| <b>Subcarriers</b>     | Specifies which subcarriers are displayed.  |
| All                    | Select All to display all subcarriers in the signal.  |
| Pilots Only            | Select Pilot to display only pilot subcarriers.   |
| Data Only              | Select Data to display only data subcarriers.   |
| Single                 | Specifies a specific subcarrier for display.  |
| Index                  | Specifies the specific subcarrier to be displayed.  |
| Range                  | Specifies a range to subcarriers to be displayed.   |
| Start                  | Specifies the start value of the range to be displayed.   |
| Stop                   | Specifies the ending value of the range to be displayed.  |
| Max symbols to analyze | Specifies how many symbols are analyzed. You can use this setting to speed analysis by limiting the number of symbols being analyzed. |

## Analysis Time Tab - OFDM

The Analysis Time tab contains parameters that define how the signal is analyzed in the OFDM Analysis displays.

|  |                 |            |              |       |       |       |
|--|-----------------|------------|--------------|-------|-------|-------|
| Modulation Params  | Advanced Params | Data Range | AnalysisTime | Trace | Scale | Prefs |
| Analysis Offset: <input type="text" value="-0.668 Sym"/> <input checked="" type="checkbox"/> Auto      Time Zero Reference: <input type="text" value="Trigger"/> |                 |            |              |       |       |       |
| Analysis Length: <input type="text" value="66.000 Sym"/> <input checked="" type="checkbox"/> Auto      Units: <input type="text" value="Symbols"/>               |                 |            |              |       |       |       |
| Actual: 66.000 Sym   |                 |            |              |       |       |       |

| Settings            | Description  |
|---------------------|--|
| Analysis Offset     | Specifies the location of the first time sample to use in measurements.  |
| Auto                | When enabled, causes the instrument to set the Analysis Offset value based on the requirements of the selected display.  |
| Analysis Length     | Specifies the length of the analysis period to use in measurements. Length is specified in either symbols or seconds, depending on the Units setting.                                      |
| Auto                | When enabled, causes the instrument to set the Analysis Length value based on the requirements of the selected display.  |
| Actual              | This is a displayed value, not a setting. It is the Analysis Length (time or symbols) being used by the analyzer, this value may not match the Analysis Length requested (in manual mode). |
| Time Zero Reference | Specifies the zero point for the analysis time.  |
| Units               | Specifies the units of the Analysis Length to either Symbols or Seconds.   |

## Analysis Offset

Use analysis offset to specify where measurements begin. Be aware that you cannot set the Analysis Offset outside the range of time covered by the current acquisition data. (all time values are relative to the Time Zero Reference).

You can set the Analysis Length so that the requested analysis period falls partly or entirely outside the current range of acquisition data settings. When the next acquisition is taken, its Acquisition Length will be increased to cover the new Analysis Length, as long as the Sampling controls are set to Auto. If the Sampling parameters are set to manual, or if the instrument is analyzing saved data, the actual analysis length will be constrained by the available data length, but in most cases, measurements are able to be made anyway. The instrument will display a notification when measurement results are computed from less data than requested. Range: 0 to [(end of acquisition) - Analysis Length]. Resolution: 1 effective sample (or symbol).

## Analysis Length

Use the analysis length to specify how long a period of time is analyzed. As you adjust this value, the actual amount of time for Analysis Length, in Symbol or Seconds units, is shown below the control in the "Actual" readout. This setting is not available when Auto is checked. Range: minimum value depends on modulation type. Resolution: 1 symbol.


## Time Zero Reference

All time values are measured from this point (such as marker position or horizontal position (in Y vs Time displays). Choices are: Acquisition Start or Trigger.

| Parameter         | Description  |
|-------------------|--|
| Acquisition Start | Offset is measured from the point at which acquisition begins. |
| Trigger           | Offset is measured from the trigger point.                     |

## Trace Tab - OFDM

The Trace tab allows you to set the trace display characteristics of the selected display. This tab is not available for all OFDM displays.



| Settings | Description  |
|----------|--|
| Trace    | Use this drop-down list to select whether or not the Average trace or the measurement trace is displayed. The measurement trace depends on the selected display. |
| Show     | Select or deselect this checkbox to show or hide the trace selected in the Trace drop-down list.   |
| Freeze   | Selecting Freeze halts updates to the trace.   |

## Scale Tab - OFDM

The Scale tab allows you to change the scale settings that control how the trace appears on the display but does not change control settings such as Measurement Frequency. There are two versions of the Scale tab for OFDM displays. One version is used only for the Constellation display and the other version is used for the OFDM EVM, OFDM Magnitude Error, OFDM Phase Error, and OFDM Power displays.

Modulation Params | Advanced Params | Data Range | AnalysisTime | Trace | **Scale** | Prefs

**Vertical**  
 Scale: 41.8 %  
 Position: -21.9 %  
 Autoscale

**Subcarrier**  
 Scale: 200 Subc  
 Position: -100 Subc  
 Autoscale

**Symbols**  
 Scale: 12 Sym  
 Position: 0 Sym  
☒ Auto Autoscale

Scale tab for OFDM EVM, OFDM Mag Error, OFDM Phase Error, and OFDM Power displays

Modulation Params | Advanced Params | Data Range | AnalysisTime | Trace | **Scale** | Prefs

**Zoom**  
 Scale: 3  
 Vertical Position: 0  
 Horizontal Position: 0

Scale tab for OFDM Constellation display

| Settings            | Description   |
|---------------------|---|
| Vertical            | Controls the vertical position and scale of the trace display.  |
| Scale               | Changes the vertical scale of the graph.  |
| Position            | Adjusts the reference level away from top of the graph.   |
| Autoscale           | Resets the scale of the vertical axis to contain the complete trace.  |
| Subcarrier          | Controls the scale of the subcarrier graph and position of the trace.   |
| Scale               | Specifies how many subcarriers are displayed horizontally.  |
| Position            | Specifies the subcarrier that appears at the left edge of the subcarrier graph.   |
| Autoscale           | Resets the scale of the horizontal axis to contain the complete trace.  |
| Symbols             | Controls the scale of the Symbols graph.  |
| Scale               | Specifies the number of symbols that appear in the symbols graph.   |
| Position            | Specifies the symbol that appears at the left edge of the Symbols graph.  |
| Autoscale           | Resets the scale and position settings to optimize the display.   |
| Auto                | When Auto is checked, the scale and position values for the Symbols graph are automatically adjusted to maintain the optimal display. |
| Zoom                | Sets the Constellation display size and position.   |
| Scale               | Sets the magnification value for the Constellation display.   |
| Vertical Position   | Sets the vertical location of the Constellation display within the graph. Range is -3.7 to +3.7.                                      |
| Horizontal Position | Sets the horizontal location of the Constellation display within the graph. Range is -3.7 to +3.7.                                    |

## Prefs Tab - OFDM

The Prefs tab enables you to change appearance characteristics of the OFDM Analysis displays. Not all settings on the Prefs tab shown below appear for every OFDM display.

The screenshot shows the 'Prefs' tab in the OFDM Analysis software. The tab is selected, and the settings are as follows:

- Time units: Symbols
- Freq units: Subcarrier
- Radix: Binary
- Show graphs: Both (selected), Subcarriers, Symbols
- Show graticule: ☒
- Show marker readout: ☒

| Settings            | Description   |
|---------------------|---|
| Time units          | Specifies whether the displayed time units are seconds or symbols.  |
| Freq units          | Specifies whether the displayed frequency units are frequency (Hz) or subcarrier channel.                                     |
| Radix               | Specifies whether symbol values are displayed in binary or hex format (for example, in the Symbol Table or markers readouts). |
| Show graphs         | Specifies which graph types are displayed.  |
| Both                | Displays both the Subcarrier and Symbol graphs.   |
| Subcarriers         | Displays only the subcarrier graph.   |
| Symbols             | Displays only the symbol graph.   |
| Show graticule      | Displays or hides the graticule in the graphs.  |
| Show marker readout | Displays or hides the marker readouts in the graphs.  |



## Overview

The noise contribution from circuit elements is usually defined by noise figure, noise factor, or noise temperature.

For example, the function of an amplifier in a particular system is to amplify the signal presented at its input and deliver it to the load. The thermal noise that is present at the input is amplified along with the input signal. The amplifier also contributes additional noise. The load receives a composite signal made up of the sum of the amplified input signal, the amplified thermal noise, and the additional noise contributed by the amplifier.

Noise figure, noise factor, and noise temperature are figures of merit used to quantify the noise added by the amplifier. This noise can be measured directly using available test equipment.

For example, devices with high noise figure can be measured directly, as long as gain is known, with accurately known bandwidths. The input to the DUT should be terminated in a source that is near the reference temperature (290K).

The Noise Figure and Gain Measurement options in this instrument measure noise figure using the Y Factor method, perform mask testing to user-defined limits, and include an uncertainty calculator.

---

**NOTE.** To read a white paper about Noise measurements and when to use particular types of noise measurements, visit [www.tektronix.com/downloads](http://www.tektronix.com/downloads).

---

## Displays

Displays include noise figure, gain, Y-factor, noise temperature, and tabular results. Single-frequency metering and swept-trace results are available, as well as support for industry-standard noise sources. Measurements for amplifiers and other non-frequency converting devices, and fixed local-oscillator up and down converters are also available.

The displays for Noise Figure and Gain (Displays > Measurements > Noise Figure and Gain) are:

- [Noise Temperature \(see page 302\)](#)
- [Gain \(see page 298\)](#)
- [Noise Figure \(see page 300\)](#)
- [Noise Table \(see page 296\)](#)
- [Uncertainty Calculator \(see page 305\)](#)
- [Y Factor \(see page 304\)](#)


These measurements allow you to measure the noise contributions of circuit elements to a signal.

## Noise Figure and Gain Measurements

The analyzer takes the following noise measurements: Noise Figure, Noise Temperature, Y Factor, and Gain. See the [Available Measurements \(see page 41\)](#) section for brief definitions of each measurement.

## Noise Table Display

The Noise Table lists selected measurements in a spreadsheet format, showing the numeric value at step frequencies for Gain, Noise Figure, Y Factor, Noise Temperature, P<sub>Hot</sub>, and P<sub>Cold</sub>.



|           | Gain     | Noise Figure | Noise Factor | Y Factor |
|-----------|----------|--------------|--------------|----------|
| 2.245 GHz | 26.03 dB | 1.33 dB      | 1.36         | 13.2     |
| 2.384 GHz | 26.25 dB | 1.25 dB      | 1.33         | 13.2     |
| 2.524 GHz | 26.56 dB | 1.44 dB      | 1.39         | 13.1     |
| 2.663 GHz | 26.49 dB | 1.56 dB      | 1.43         | 12.9     |
| 2.803 GHz | 24.92 dB | 1.59 dB      | 1.44         | 12.9     |
| 2.942 GHz | 22.21 dB | 1.69 dB      | 1.48         | 12.7     |
| 3.082 GHz | 19.36 dB | 1.78 dB      | 1.51         | 12.5     |
| 3.221 GHz | 16.14 dB | 2.11 dB      | 1.63         | 12.0     |
| 3.361 GHz | 13.50 dB | 3.07 dB      | 2.03         | 10.9     |
| 3.500 GHz | 11.70 dB | 3.89 dB      | 2.45         | 10.0     |

### Displaying the Noise Table Using a Preset

Using the Noise Figure Application Preset will turn on the Noise Figure, Gain, and Noise Table displays, and set the preamplifier (if present) ON, with the acquisition mode to *minimum noise* and the attenuator to 0 dB for the best noise floor.



**CAUTION.** Make sure that the input of the analyzer is not connected to a device that will overload or damage the preamplifier.

1. Select **Presets >Application** from the menu bar.
2. Select **Noise Figure** from the Preset drop down menu.
3. Click **OK**.


## Displaying the Noise Table

1. Press the **Displays** button or select **Setup > Displays**.
2. Select **Noise Figure and Gain** in the Measurements box.
3. Double-click **Noise Table** in the **Available displays** box to add the display to the Selected displays box.
4. Click **OK**.

## Setting the Frequency for a Local Oscillator

If you are using a local oscillator and need to set the local oscillator frequency, do the following:

1. With the Noise Table selected, press the **Settings** button.
2. Select Up or Down Convertor as the Measurement Type in the **Overview** tab.



3. Click on the Frequency icon in the **Overview** tab image (see above). This will open the **Params** tab.
4. Set the frequency under LO Settings.
5. Click the X located in the right top corner of the Settings control panel to close the panel.

## Noise Table Settings

**Menu Bar:** Setup > Settings

**Front Panel:** Settings

**Application Toolbar:** 

The Settings control panel tabs for the Noise Table Display are shown in the following table.


| Settings tab   | Description  |
|--|--|
| <a href="#">Overview (see page 309)</a>                  | Specifies the measurement type (non-frequency converting or frequency converting devices) to be shown in the display. Also allows you to navigate to the related panel tabs. |
| <a href="#">Freq Setup (see page 310)</a>                | Specifies several parameters that control frequency of the test and number of points per measurement.  |
| <a href="#">Noise Source (see page 311)</a>              | Specifies ENR and noise source mode (constant or from a table).  |
| <a href="#">Ext Gain/Loss Corrections (see page 311)</a> | Specifies gain and loss values for cables, connectors, and external preamplifiers.   |
| <a href="#">Params (see page 313)</a>                    | Specifies local oscillator (LO) and RSA settings, reference temperature, resolution bandwidth (RBW), and average counts.   |
| <a href="#">Internal Settings (see page 314)</a>         | Specifies the reference level (dB), set internal attenuator (manual or auto), optimize RF and IF, and set an internal preamp.  |
| <a href="#">Prefs (see page 317)</a>                     | Specifies display preferences such as graticule on/off, number of displayed points and measurement units.  |

**Calibrate.** Calibrates the equipment setup prior to measurement.

## Gain Display

The Gain display shows gain versus frequency of the signal. The gain measurement is the ratio of output power to input power in an amplifier or circuit element.

### Elements of the Gain Display



| Item | Display element             | Description  |
|------|-----------------------------|--|
| 1    | Check mark indicator        | The check mark indicator in the upper, left-hand corner of the display shows when the Gain display is the optimized display.<br><br><b>NOTE.</b> When Best for multiple windows is selected in the Amplitude control panel's RF & IF Optimization control, none of the measurement displays shows a checkmark, as there is not a single optimized measurement. |
| 2    | dB                          | Sets the vertical scale maximum.   |
| 3    | /Div                        | The distance between graticule lines on the display in dB.   |
| 4    | Position                    | Sets the vertical scale center.  |
| 5    | Min                         | Sets the vertical scale minimum.   |
| 6    | Autoscale                   | Set the axes to values that show all trace points.   |
| 7    | Horizontal offset and scale | Located along the bottom edge of the trace display, these controls specify the offset and scale on the horizontal axis.  |
| 8    | Max                         | Shows the maximum level of the trace.  |
| 9    | Min                         | Shows the minimum level of the trace.  |
| 10   | Mean                        | Shows the mean value of the trace.   |

## Gain Settings

**Menu Bar:** Setup > Settings

**Front Panel:** Settings

**Application Toolbar:** 

The Setup settings for the Gain display are shown in the following table.


| Settings tab   | Description   |
|--|---|
| <a href="#">Overview (see page 309)</a>                  | Specifies the measurement type to be shown in the display. Also allows you to navigate to the related panel tabs.             |
| <a href="#">Freq Setup (see page 310)</a>                | Specifies several parameters that control frequency displayed.  |
| <a href="#">Noise Source (see page 311)</a>              | Specifies ENR and noise source mode (constant or from a table).   |
| <a href="#">Ext Gain/Loss Corrections (see page 311)</a> | Specifies gain and loss values for cables, connectors, and external preamplifiers.  |
| <a href="#">Params (see page 313)</a>                    | Specifies local oscillator (LO) and RSA settings, reference temperature, resolution bandwidth (RBW), and average counts.      |
| <a href="#">Internal Settings (see page 314)</a>         | Specifies the reference level (dB), set internal attenuator (manual or auto), optimize RF and IF, and set an internal preamp. |
| <a href="#">Traces (see page 315)</a>                    | Specify the display trace parameters and also allows you to save traces.  |
| <a href="#">Scale (see page 316)</a>                     | Specifies the scale settings that control how the trace appears on the display but does not change control settings.          |
| <a href="#">Prefs (see page 317)</a>                     | Specifies preferences such as graticule on/off, number of displayed points and measurement units.                             |

**Calibrate.** Calibrates the equipment setup prior to measurement.

## Noise Figure Display

The Noise Figure display shows the noise factor in dB. This measurement can help you assess the low level sensitivity of the DUT. Lower noise figure is found in better performing DUTs. Noise factor is defined as the ratio of the input signal to noise ratio to the output signal to noise ratio (Input SNR/Output SNR).

### Elements of the Noise Figure Display



| Item | Display element             | Description  |
|------|-----------------------------|--|
| 1    | Check mark indicator        | The check mark indicator in the upper, left-hand corner of the display shows when the Noise Figure display is the optimized display.<br><br><b>NOTE.</b> When Best for multiple windows is selected in the Amplitude control panel's RF & IF Optimization control, none of the measurement displays shows a checkmark, as there is not a single optimized measurement. |
| 2    | Max                         | Sets the vertical scale maximum.   |
| 3    | /Div                        | The distance between graticule lines on the display in dB.   |
| 4    | Position                    | Sets the vertical scale center.  |
| 5    | Min                         | Sets the vertical scale minimum.   |
| 6    | Autoscale                   | Set the axes to values that show all trace points.   |
| 7    | Horizontal offset and scale | Located along the bottom edge of the trace display, these controls specify the offset and scale on the horizontal axis.  |

## Noise Figure Settings

**Menu Bar:** Setup > Settings

**Front Panel:** Settings

**Application Toolbar:** 

The settings for the Noise Figure Display are shown in the following table.


| Settings tab   | Description   |
|--|---|
| <a href="#">Overview (see page 309)</a>                  | Specifies the measurement type to be shown in the display. Also allows you to navigate to the related panel tabs.             |
| <a href="#">Freq Setup (see page 310)</a>                | Specifies several parameters that control measurement and displayed frequencies.  |
| <a href="#">Noise Source (see page 311)</a>              | Specifies ENR and noise source mode (constant or from a table).   |
| <a href="#">Ext Gain/Loss Corrections (see page 311)</a> | Specifies gain and loss values for cables, connectors, and external preamplifiers.  |
| <a href="#">Params (see page 313)</a>                    | Specifies local oscillator (LO) and RSA settings, reference temperature, resolution bandwidth (RBW), and average counts.      |
| <a href="#">Internal Settings (see page 314)</a>         | Specifies the reference level (dB), set internal attenuator (manual or auto), optimize RF and IF, and set an internal preamp. |
| <a href="#">Traces (see page 315)</a>                    | Specify the display trace parameters and also allows you to save traces.  |
| <a href="#">Scale (see page 316)</a>                     | Specifies the scale settings that control how the trace appears on the display but does not change control settings.          |
| <a href="#">Prefs (see page 317)</a>                     | Specifies preferences such as graticule on/off, number of displayed points and measurement units.                             |

**Calibrate.** Calibrates the equipment setup prior to measurement.

## Noise Temperature Display

The Noise Temperature display shows the temperature of a source impedance that would give the same SNR as the amplifier or system element, if it were noiseless.

### Elements of the Noise Temperature Display



| Item | Display element             | Description   |
|------|-----------------------------|---|
| 1    | Check mark indicator        | The check mark indicator in the upper, left-hand corner of the display shows when the Noise Temperature display is the optimized display.<br><br><b>NOTE.</b> When <i>Best for multiple windows</i> is selected in the <i>Amplitude control</i> panel's <i>RF &amp; IF Optimization control</i> , none of the measurement displays shows a checkmark, as there is not a single optimized measurement. |
| 2    | Max                         | Sets the vertical scale maximum.  |
| 3    | /Div                        | The distance between graticule lines on the display in K.   |
| 4    | Position                    | Sets the vertical scale center.   |
| 5    | Min                         | Sets the vertical scale minimum.  |
| 6    | Autoscale                   | Set the axes to values that show all trace points.  |
| 7    | Horizontal offset and scale | Located along the bottom edge of the trace display, these controls specify the offset and scale on the horizontal axis.   |
| 8    | Vertical scale              | Sets the vertical scale value.  |

## Noise Temperature Settings

**Menu Bar:** Setup > Settings

**Front Panel:** Settings

**Application Toolbar:** 

The settings for the Noise Temperature Display are shown in the following table.

| Settings tab   | Description   |
|--|---|
| <a href="#">Overview (see page 309)</a>                  | Specifies the measurement type to be shown in the display. Also allows you to navigate to the related panel tabs.             |
| <a href="#">Freq Setup (see page 310)</a>                | Specifies several parameters that control measurement and displayed frequencies.  |
| <a href="#">Noise Source (see page 311)</a>              | Specifies ENR and noise source mode (constant or from a table).   |
| <a href="#">Ext Gain/Loss Corrections (see page 311)</a> | Specifies gain and loss values for cables, connectors, and external preamplifiers.  |
| <a href="#">Params (see page 313)</a>                    | Specifies local oscillator (LO) and RSA settings, reference temperature, resolution bandwidth (RBW), and average counts.      |
| <a href="#">Internal Settings (see page 314)</a>         | Specifies the reference level (dB), set internal attenuator (manual or auto), optimize RF and IF, and set an internal preamp. |
| <a href="#">Traces (see page 315)</a>                    | Specify the display trace parameters and also allows you to save traces.  |
| <a href="#">Scale (see page 316)</a>                     | Specifies the scale settings that control how the trace appears on the display but does not change control settings.          |
| <a href="#">Prefs (see page 317)</a>                     | Specifies preferences such as graticule on/off, number of displayed points and measurement units.                             |


**Calibrate.** Calibrates the equipment setup prior to measurement.

## Y Factor Display

The Y-Factor measurement method uses a noise source that can be switched off and on with a specified amount of excess noise. It is based on two power measurements, each performed with the same port impedances and the same measurement bandwidth. This is specified as the Excess Noise Ratio (ENR). ENR is the ratio of noise from the source to the system thermal noise or  $kTB$ , often expressed in dB.

Y-Factor measurements require an initial calibration where the noise source is measured in the on and off state with the DUT bypassed. This allows removal of the RSA noise figure using the two-stage NF correction, and allows for accurate gain measurement.

### Elements of the Y Factor Display



| Item | Display element             | Description   |
|------|-----------------------------|---|
| 1    | Check mark indicator        | The check mark indicator in the upper, left-hand corner of the display shows when the Y Factor display is the optimized display.<br><br><b>NOTE.</b> When <i>Best for multiple windows</i> is selected in the <i>Amplitude control</i> panel's <i>RF &amp; IF Optimization</i> control, none of the measurement displays shows a checkmark, as there is not a single optimized measurement. |
| 2    | Max                         | Sets the vertical scale maximum.  |
| 3    | /Div                        | The distance between graticule lines on the display in dB.  |
| 4    | Position                    | Sets the vertical scale center.   |
| 5    | Min                         | Sets the vertical scale minimum.  |
| 6    | Autoscale                   | Set the axes to values that show all trace points.  |
| 7    | Horizontal offset and scale | Located along the bottom edge of the trace display, these controls specify the offset and scale on the horizontal axis.   |
| 8    | Vertical scale              | Sets the vertical scale value.  |

## Y Factor Settings

**Menu Bar:** Setup > Settings

**Front Panel: Settings****Application Toolbar:** 

The settings for the Y Factor Display are shown in the following table.

| Settings tab   | Description   |
|--|---|
| <a href="#">Overview (see page 309)</a>                  | Specifies the measurement type to be shown in the display. Also allows you to navigate to the related panel tabs.             |
| <a href="#">Freq Setup (see page 310)</a>                | Specifies several parameters that control measurement and displayed frequencies.  |
| <a href="#">Noise Source (see page 311)</a>              | Specifies ENR and noise source mode (constant or from a table).   |
| <a href="#">Ext Gain/Loss Corrections (see page 311)</a> | Specifies gain and loss values for cables, connectors, and external preamplifiers.  |
| <a href="#">Params (see page 313)</a>                    | Specifies local oscillator (LO) and RSA settings, reference temperature, resolution bandwidth (RBW), and average counts.      |
| <a href="#">Internal Settings (see page 314)</a>         | Specifies the reference level (dB), set internal attenuator (manual or auto), optimize RF and IF, and set an internal preamp. |
| <a href="#">Traces (see page 315)</a>                    | Specify the display trace parameters and also allows you to save traces.  |
| <a href="#">Scale (see page 316)</a>                     | Specifies the scale settings that control how the trace appears on the display but does not change control settings.          |
| <a href="#">Prefs (see page 317)</a>                     | Specifies preferences such as graticule on/off, number of displayed points and measurement units.                             |

**Calibrate.** Calibrates the equipment setup prior to measurement.

## Uncertainty Calculator Display

The noise factor of a DUT is computed from two power measurements (noise source and DUT) and two calibration power measurements (external preamp and RSA). These computations rely on specified characteristics on the noise source ENR, input and output match specifications, and other component specifications used in these measurements. The Uncertainty Calculator accepts the key specifications of each device and their specification limit and converts them to standard uncertainty for the noise figure and gain measurement. The output of the Uncertainty Calculator is the uncertainty tolerance for gain and noise figure based on the confidence level (conversion factor) set by the user.

It is important to understand that the computations rely on the nominal values and tolerance limits of components in the noise figure and gain measurement. The Calculator will accept different forms of component specifications and confidence level in the component specifications. The Calculator converts the component specifications into standard uncertainties for the measurement. All of the component standard uncertainties are combined into a standard measurement uncertainty. The standard measurement uncertainty is multiplied by a coverage factor to yield an *expanded coverage factor* for the measurement. For K=2, approximately 95% confidence is frequently used, as is K=3, approximately 99% confidence (maximum). So, if the noise figure is reported as 2 dB  $\pm$ 0.25 dB at 95% confidence, this means that the “true” value of noise figure is thought to lie between 1.75 dB to 2.25 dB with 95% confidence.

## Elements of the Uncertainty Calculator Display

| Computed Uncertainty |          |       |          |                        |        |
|----------------------|----------|-------|----------|------------------------|--------|
| Noise Figure:        | 0.378 dB | Gain: | 0.782 dB | Measurement Confidence | 95 % ▼ |

| Noise Source     |         |          |        |                    |      |
|------------------|---------|----------|--------|--------------------|------|
| Noise ENR:       | 15 dB   |          |        |                    |      |
| Uncertainty ENR: | 0.15 dB | Coverage | 99 % ▼ | Error Distribution |      |
| Output VSWR:     | 1.5     | 99.7 %   |        | U-Shaped           |      |
|                  |         |          |        | Noise Figure       | Gain |
|                  |         |          |        | 0.117 dB           | 0 dB |

| DUT - Device Under Test |           |          |        |                    |          |
|-------------------------|-----------|----------|--------|--------------------|----------|
| NF:                     | 3.000 dB  |          |        |                    |          |
| Gain:                   | 20.000 dB |          |        |                    |          |
| Input VSWR:             | 1.5       | Coverage | 99.7 % | Error Distribution | U-Shaped |
| Output VSWR:            | 1.5       | 99.7 %   |        | U-Shaped           |          |

| Ext Preamp    |           |          |        |                    |          |
|---------------|-----------|----------|--------|--------------------|----------|
| In Use:       | No ▼      |          |        |                    |          |
| Noise Figure: | 3.500 dB  |          |        |                    |          |
| Gain:         | 21.000 dB |          |        |                    |          |
| Input VSWR:   | 1.3       | Coverage | 99.7 % | Error Distribution | U-Shaped |

| RSA             |           |          |        |                    |          |
|-----------------|-----------|----------|--------|--------------------|----------|
| NF:             | 10.000 dB |          |        |                    |          |
| Uncertainty NF: | 0.070 dB  | Coverage | 95 % ▼ | Error Distribution |          |
| Gain:           | 0.020 dB  | 95 % ▼   |        |                    |          |
| Input VSWR:     | 1.5       | 99.7 %   |        | U-Shaped           |          |
| Sub DANL Floor: | 13.000 dB | 99 % ▼   |        |                    |          |
|                 |           |          |        | Noise Figure       | Gain     |
|                 |           |          |        | 0.161 dB           | 0.102 dB |
|                 |           |          |        | 0.01 dB            | 0 dB     |
|                 |           |          |        | 0 dB               | 0.01 dB  |

| Item | Display element        | Description   |
|------|------------------------|---|
| 1    | Computed Uncertainty   | Shows the noise figure and gain standard uncertainties. You can also select the measurement uncertainty (99%, 95%, or 80%).   |
| 2    | Noise Source           | Sets the noise source ENR, uncertainty of the ENR and source match, coverage (measurement confidence), output Voltage Standing Wave Ratio (VSWR). It also shows the error distribution and the Noise Figure and Gain uncertainties, which will vary according to the values entered for the available parameters. These values can be determined from the manufacturer's data sheet for the noise source you are using.   |
| 3    | DUT- Device Under Test | Sets the input Voltage Standing Wave Ratio (VSWR) and output VSWR for your device under test (DUT). It also shows the coverage (measurement confidence) and error distributions for this element of the uncertainty calculation.  |
| 4    | Ext Preamp             | Sets the external preamp to in use (Yes), and the input Voltage Standing Wave Ratio (VSWR) value. When the external preamp is selected on, you can also set the noise figure and gain of the preamp. It also shows the coverage (measurement confidence) and error distributions for the input Voltage Standing Wave Ratio (VSWR).  |
| 5    | RSA                    | Sets the error parameters for the RSA. These include the input noise figure and noise figure uncertainty, gain uncertainty, input VSWR, and the sub-DANL noise floor error terms of the RSA. Input noise figure and VSWR information can be taken from the RSA data sheet or measured by the user for the specific instrument used. Gain uncertainty contains a predetermined value that is the same for all RSAs. Sub-DANL distortion refers to distortion products present in the RSA that exist below the noise floor and cannot be measured by the user. The worst-case value is given for the term by default. You can also select the coverage (measurement confidence) for Noise Figure uncertainty, Gain measurement, and sub DANL floor. The error distribution is also displayed. |

## Uncertainty Calculator Settings

**Menu Bar:** Setup > Settings

**Front Panel:** Settings

**Application Toolbar:** 

The settings for the Uncertainty Calculator display are shown in the following table.

| Settings tab                                       | Description  |
|--|--|
| <a href="#">Uncertainty (see page 308)</a>         | Shows noise figure and gain uncertainty values (dB). |
| <a href="#">Intermediate Values (see page 308)</a> | Shows standard uncertainties in system.              |

# Uncertainty Tab

The Uncertainty tab shows the noise figure and gain uncertainty values (dB). These values change based on the noise measurement result and the parameters set in the Uncertainty Calculator display.

| Uncertainty   | Intermediate Values |
|---------------|---------------------|
| Noise Figure: | 0.338 dB            |
| Gain:         | 0.632 dB            |

| Setting      | Description  |
|--------------|--|
| Noise Figure | Shows the standard uncertainty for the noise figure measurement. This value also reflects the uncertainty tolerance for noise figure based on the confidence level (conversion factor) set by the user in the Display. |
| Gain         | Shows the standard uncertainty for the gain measurement. This value also reflects the uncertainty tolerance for gain based on the confidence level (conversion factor) set by the user in the Display.                 |

# Intermediate Values Tab

The Intermediate Values tab shows the noise figure and gain uncertainty values (dB). These values change based on the parameters set in the Uncertainty Calculator display and the noise measurement results.

| Uncertainty          | Intermediate Values          |                      |
|----------------------|------------------------------|----------------------|
|                      | Uncertainty for Noise Factor | Uncertainty for Gain |
| Mismatch Source-SA:  | 0 dB                         | 0.465 dB             |
| Mismatch Source-DUT: | 0.234 dB                     | 0.465 dB             |
| Mismatch DUT-SA:     | 0.234 dB                     | 0.465 dB             |

| Setting                      | Description  |
|------------------------------|--|
| Uncertainty for Noise Figure | Shows the uncertainty for three components of the noise figure measurement.    |
| Uncertainty for Gain         | Shows the uncertainty for three components of the gain measurement.            |
| Mismatch Source-SA           | Shows the error between the noise source and the spectrum analyzer.            |
| Mismatch Source-DUT          | Shows the error between the noise source and the device under test (DUT).      |
| Mismatch DUT-SA              | Shows the error between the device under test (DUT) and the spectrum analyzer. |


## Noise Figure and Gain Measurement Settings

The control panel tabs in this section are shared by the displays in the Noise Figure and Gain measurement (Setup > Displays). Changing a setting on one tab changes that setting for all the noise and gain measurement displays.


| Settings tab   | Description   |
|--|---|
| <a href="#">Overview (see page 309)</a>                  | Specifies the measurement type to be shown in the display. Also allows you to navigate to the related panel tabs.             |
| <a href="#">Freq Setup (see page 310)</a>                | Specifies several parameters that control measurement and displayed frequencies.  |
| <a href="#">Noise Source (see page 311)</a>              | Specifies ENR and noise source mode (constant or from a table).   |
| <a href="#">Ext Gain/Loss Corrections (see page 311)</a> | Specifies gain and loss values for cables, connectors, and external preamplifiers.  |
| <a href="#">Params (see page 313)</a>                    | Specifies LO and RSA settings, reference temperature, RBW, and average counts.  |
| <a href="#">Internal Settings (see page 314)</a>         | Specifies the reference level (dB), set internal attenuator (manual or auto), optimize RF and IF, and set an internal preamp. |
| <a href="#">Traces (see page 315)</a>                    | Specifies the display trace parameters and also allows you to save traces.  |
| <a href="#">Scale (see page 316)</a>                     | Specifies the scale settings that control how the trace appears on the display but does not change control settings.          |
| <a href="#">Prefs (see page 317)</a>                     | Specifies display view.   |

## Overview Tab

The Overview tab is used to specify the measurement type that appears in the display. You can also use the images in the tab to quickly navigate to tabs that allow you to adjust parameters at key places in the circuit system. You can select either **Amp/Direct**, **Up Converter**, or **Down Converter** as the measurement type. If you are using an up frequency converter, for example, you would want to select **Up Converter**.



The above image shows the Overview tab with Amp/Direct selected.




The above image shows the Overview tab with Up Convertor selected.

| Setting          | Description  |
|------------------|--|
| Measurement Type | Specify whether the measurement should account for a frequency converter (Up Converter or Down Converter) or not (Amp/Direct). |

## Freq Setup Tab

The Freq Setup tab is used to specify the frequency mode (Sweep, Single, or Frequency Table). If you select the Frequency Table, you can edit and load an existing table or save a table.



| Setting          | Description  |
|------------------|--|
| Frequency Mode   | Set to Single Frequency, Sweep Frequency, or Frequency Table. To set the band of interest for Sweep Frequency, set the Start and Stop frequencies and then specify the Steps value for the number of measurement points you want to have. To set the band of interest for Single Frequency, set the Center Frequency value. To use the Frequency Table, you can edit the current table, load an existing table, or save the current table. |
| Center Frequency | Specify the center frequency (GHz).  |
| Span             | Specify the range over which to measure the signal. (Only available with Sweep mode.)  |
| Start frequency  | Specify the minimum frequency to include in the measurement. (Only available with Sweep mode.)   |
| Stop frequency   | Specify the maximum frequency to include in the measurement. (Only available with Sweep mode.)   |
| Steps            | Specify the number of steps in the measurement range. This value determines how many steps (rows) are in the frequency table. (Only available with Sweep mode.)  |
| Table            | You can edit an existing frequency table, load a previously saved table, or save a table.  |

## Noise Source Tab

The Noise Source tab is used to specify the Noise mode (constant or table) and the ENR constant. If you select Table mode, you can edit and load an existing table or save a table.

| Frequency | ENR      |
|-----------|----------|
| 2.90 GHz  | 15.00 dB |
| 2.90 GHz  | 15.00 dB |
| 2.90 GHz  | 15.00 dB |
| 2.90 GHz  | 15.00 dB |
| 2.90 GHz  | 15.00 dB |

| Setting                    | Description  |
|----------------------------|--|
| Mode                       | Set mode to Constant or Table. For Constant, set the ENR value. For Table, enter the noise source model and serial number.<br>To use the Table, you can edit the current table, load an existing table, or save the current table.<br>You can also load a predefined noise source. These values are useful starting points because they contain the frequency points for common noise sources, along with representative excess noise ratios. However, you must enter your own ENR values. |
| ENR Constant               | Specify the excess noise ratio (ENR) value. The ENR is the difference between the two power levels calculated in a Y-Factor measurement.   |
| Noise Source Model         | Specify the model of the noise source. (Only available with Table mode.)   |
| Noise Source Serial Number | Specify serial number of the noise source. (Only available with Table mode.)   |
| Table                      | You can edit an existing Noise Source table, load a previously saved table, or save a table.   |

## Ext Gain/Loss Corrections Tab


The Ext Gain/Loss Corrections tab is used to specify external gain and loss values. You can set loss values for three common loss scenarios: loss at the DUT input, loss at the DUT output, and loss at calibration. You can also edit directly in an external loss table.

| Constant Values                                     | External Loss Tables (Loss > 0, Gain < 0)  | Temperature                                       |
|---|--|---|
| Loss @ DUT Input: <input type="checkbox"/> 0.0 dB   | <input type="checkbox"/> Ext Loss Table 1 <input type="button" value="Edit..."/> | 290 K <input checked="" type="checkbox"/> Use Ref |
| Loss @ DUT Output: <input type="checkbox"/> 0.0 dB  | <input type="checkbox"/> Ext Loss Table 2 <input type="button" value="Edit..."/> | 290 K <input checked="" type="checkbox"/> Use Ref |
| Loss @ Calibration: <input type="checkbox"/> 0.0 dB | <input type="checkbox"/> Ext Loss Table 3 <input type="button" value="Edit..."/> | 290 K <input checked="" type="checkbox"/> Use Ref |

(Loss < 0, Gain > 0)

| Setting              | Description   |
|----------------------|---|
| Constant values      | Specify up to three external gain values (dB). Use these settings to apply a flat gain/loss correction to the signal. Positive values represent a gain and negative values represent a loss. The range is -50 to +30 dB. Resolution is 0.1 dB.  |
| Loss @ DUT Input     | This specifies the loss at the DUT input. For example, this can be the result of an adapter connected between a waveguide adapter and the DUT input.  |
| Loss @ DUT Output    | This specifies the loss at the DUT output. For example, this can be the result of an adapter connected between the DUT output and the analyzer.   |
| Loss @ Calibration   | This specifies the total loss in the calibrated system. For example, if an adapter must be used during calibration that is not present during measurements, the frequency response of the adapter can be removed with this correction table.  |
| External Loss Tables | Each loss component (at DUT input, DUT output, and at calibration) can use an External Loss Table for correction of loss versus frequency. When an External Loss Table is selected, the analyzer adjusts the signal according to the values in the table. An external loss table allows you to compensate the signal level for variations in cable loss, antenna frequency response or preamp frequency response. You can read more about using Loss Tables <a href="#">here (see page 574)</a> . |
| Temperature          | Specify the temperature (K) of the loss elements. The temperatures are used in adjusting the measured results based on the temperature difference of the loss elements compared to the reference temperature. You can also use the reference temperature for these values by selecting the click-box. The reference temperature is entered in the Params tab.   |

The following image shows the External Loss Table setup window.



# Params Tab

The Params tab is used to specify the local oscillator (LO) and RSA settings, reference temperature, resolution bandwidth (RBW), and average count.

The screenshot shows the 'Params' tab selected in a software interface. The interface has a menu bar with 'Overview', 'Freq Setup', 'Noise Source', 'Ext Gain/Loss Corrections', 'Params', 'Internal Settings', 'Traces', 'Scale', and 'Prefs'. The 'Params' tab is active, displaying two main sections: 'RSA Settings' and 'LO Settings'. The 'RSA Settings' section includes 'Settle Time (On): 27 ms', 'Settle Time (Off): 1.001 s', 'Ref Temperature: 290 K', 'RBW: 1.00 MHz', and 'Average Count: 1'. The 'LO Settings' section includes 'Frequency: 1.500 GHz'.

| Setting         | Description  |
|-----------------|--|
| RSA Settings    | The Settle Time sets the time the RSA will pause before measuring after turning the noise source On or Off to assure measurement stability. The required settling time for your measurement setup will vary with the noise source used. The default values have been set for Noisecom 346B series noise sources. Settling time for your noise source may be different and can be determined by judging the stability of your measurement as the settle time is varied. |
| Ref Temperature | Sets the reference temperature (K).  |
| RBW             | Sets the resolution bandwidth (MHz).   |
| Average Count   | Sets the number of measurements to average.  |
| LO Settings     | Sets the frequency of the local oscillator, which is an oscillator that produces the internal signal that is mixed with an incoming signal to produce the IF signal. Set the Side Band to Upper, Lower, or Double.   |


## Internal Settings Tab

The Internal Settings tab is used to specify the reference level (dB) for the signal analyzer's RF front end, the settings for the internal attenuator, enable/disable the optional internal preamp, and specify the RF & IF Optimization. If you use the application preset for Noise Figure and Gain, these settings will be automatically optimized for the best measurements. The application preset will turn the preamplifier on, set the attenuator to 0 dB, and set RF/IF optimization to minimize noise using 25 MHz bandwidth acquisitions. In instruments that have wider acquisition bandwidths, measurement speed can be improved by setting the RF & IF Optimization to *Best for multiple displays*. This allows use of wider acquisition bandwidths, but causes a slight reduction in ultimate instrument noise floor.


| Setting              | Description   |
|----------------------|---|
| Ref Level            | Specify the reference level (dB) for the signal analyzer's RF front end.  |
| Internal Attenuator  | Set to Auto or manual. When Auto mode is enabled, the attenuation setting is displayed but is not adjustable. When Auto is disabled, the attenuator operates in manual mode, allowing you to set the internal attenuation. The internal attenuation setting range is 0 - 75 dB. |
| RF & IF Optimization | Specifies how the gain and bandwidth should be optimized (in the RF and IF stages of the front end). Read about these settings <a href="#">here (see page 570)</a> .  |
| Internal Preamp      | Switch the preamp on or off. To ensure the best accuracy, turn on the Internal Preamp and allow it to run for 20 minutes to thermally stabilize before taking critical measurements.  |

## Traces Tab

The Traces tab is used to specify the display trace parameters and also allows you to save traces. This tab is available for trace displays only.



This is the Traces tab when Function is set to Normal.




This is the Traces tab when Function is set to .

| Setting  | Description  |
|--|--|
| Trace  | Specify up to three traces and whether or not to show the traces of freeze them. |
| Detection  | Detection is set to Average (VRMS).  |
| Function   | Specify the trace function as Normal, Average (VRMS), Max Hold, or Min Hold.     |
| Count<br>(Only available for Noise<br>Temperature display) | Specify the trace count.   |

# Scale Tab

The Scale tab allows you to change the scale settings that control how the trace appears on the display but does not change control settings such as Frequency. In effect, these controls operate like pan and zoom controls. This tab is only available for trace displays.




| Settings   | Description   |
|------------|---|
| Vertical   | Controls the vertical position and scale of the trace display.  |
| Scale      | Changes the vertical scale of the graph.  |
| Position   | Adjusts the reference level away from top of the graph.   |
| Autoscale  | Resets the scale of the vertical axis to contain the complete trace.  |
| Auto       | When Auto is checked, the scale and position values for the are automatically adjusted to maintain the optimal display. |
| Reset      | Resets horizontal and vertical values to default values.  |
| Horizontal | Controls the horizontal position and scale of the trace display.  |
| Scale      | Changes the horizontal scale of the graph.  |
| Position   | Adjusts the horizontal reference level of the graph.  |
| Autoscale  | Resets the scale of the horizontal axis to contain the complete trace.  |


## Prefs Tab

The Prefs tab is used to set display preferences. You can set preferences to show measurement points, graticule, trace legend, scale, units, or to view a linear display of results instead of dBm depending on the selected display.


The following image shows the Prefs tab for all of the trace displays.



The following image shows the Prefs tab for the Noise Table display.



The following image shows the Prefs tab for the Noise Figure display.





## Overview

The displays in Pulsed RF (Displays > Measurements > Pulsed RF) are:

- Pulse Statistics
- Pulse Table
- Pulse Trace

The Pulse measurements provide deep insight into pulse train behavior.

### Pulse Measurements

The analyzer takes the following pulse measurements: Average ON Power, Peak Power, Average Transmitted Power, Pulse Width, Rise Time, Fall Time, Repetition Interval (Sec), Repetition Rate (Hz), Duty Factor (%), Duty Factor (Ratio), Ripple, Droop, Pulse-Pulse Phase Difference, Pulse-Pulse Frequency Difference, RMS Frequency Error, Maximum Frequency Error, RMS Phase Error, Maximum Phase Error, Frequency Deviation, Delta Frequency, Phase Deviation, and Time. See [Available Measurements \(see page 35\)](#) for definitions.

## Pulse Table Display

The Pulse Table lists selected measurements in a spreadsheet format, showing the pulse numbers and all the results for each measurement on all the detected pulses. Pulses are numbered from one.


| Freq Error: 169.1 Hz (Auto)      Freq: 2.700000169 GHz |             |             |             |              |              |              |              |              |     |  |
|--|-------------|-------------|-------------|--------------|--------------|--------------|--------------|--------------|-----|--|
|  | Avg ON      | Peak        | Avg Tx      | Width        | Rise         | Fall         | Rep Int      | Rep Rate     | Ri  |  |
| 1  | -2.3542 dBm | -2.2729 dBm | -18.732 dBm | 5.1147485 us | 70.286255 ns | 69.794922 ns | 220.95078 us | 4.525895 kHz | 0.1 |  |
| 2  | -2.3518 dBm | -2.2345 dBm | -18.187 dBm | 7.2183543 us | 70.688479 ns | 70.083004 ns | 275.67105 us | 3.627512 kHz | 0.1 |  |
| 3  | -2.3510 dBm | -2.2595 dBm | -18.619 dBm | 5.2991536 us | 71.052249 ns | 69.774167 ns | 223.23629 us | 4.479559 kHz | 0.1 |  |
| 4  | -2.3511 dBm | -2.2816 dBm | -18.731 dBm | 5.1165252 us | 70.684202 ns | 70.438233 ns | 220.96363 us | 4.525632 kHz | 0.1 |  |
| 5  | -2.3513 dBm | -2.2713 dBm | -18.192 dBm | 7.2066669 us | 71.347657 ns | 70.579837 ns | 275.49471 us | 3.629834 kHz | 0.1 |  |
| 6  | -2.3531 dBm | -2.2518 dBm | -18.587 dBm | 5.4644311 us | 70.315551 ns | 70.734863 ns | 228.46991 us | 4.376944 kHz | 0.1 |  |
| 7  | -2.3537 dBm | -2.2510 dBm | -18.727 dBm | 5.1185839 us | 71.486205 ns | 70.560304 ns | 220.90306 us | 4.526873 kHz | 0.1 |  |
| 8  | -2.3530 dBm | -2.3067 dBm | -18.157 dBm | 7.2754278 us | 70.033572 ns | 69.621585 ns | 275.90856 us | 3.624389 kHz | 0.1 |  |
| 9  | -2.3547 dBm | -2.2646 dBm | -18.667 dBm | 5.1216639 us | 70.291136 ns | 69.699709 ns | 218.00389 us | 4.587074 kHz | 0.1 |  |
| 10   | -2.3505 dBm | -2.2762 dBm | -18.735 dBm | 5.1066654 us | 69.887697 ns | 70.510254 ns | 220.97585 us | 4.525381 kHz | 0.1 |  |
| 11   | -2.3541 dBm | -2.2895 dBm | -18.198 dBm | 7.1933723 us | 70.523072 ns | 70.412597 ns | 275.21345 us | 3.633543 kHz | 0.1 |  |
| 12   | -2.3528 dBm | -2.2509 dBm | -18.556 dBm | 5.7297557 us | 70.941773 ns | 69.781493 ns | 237.95580 us | 4.202461 kHz | 0.1 |  |
| 13   | -2.3561 dBm | -2.2638 dBm | -18.444 dBm | 6.0395946 us | 71.277462 ns | 70.528564 ns | 244.25826 us | 4.094027 kHz | 0.1 |  |
| 14   | -2.3530 dBm | -2.3083 dBm | -18.730 dBm | 5.1135430 us | 70.236204 ns | 69.981688 ns | 220.97836 us | 4.525330 kHz | 0.1 |  |
| 15   | -2.3500 dBm | -2.3191 dBm | -18.193 dBm | 7.2002217 us | 71.079711 ns | 71.169431 ns | 275.33952 us | 3.631880 kHz | 0.1 |  |
| 16   | -2.3559 dBm | -2.2768 dBm | -18.572 dBm | 5.6117051 us | 69.582519 ns | 70.191653 ns | 233.68913 us | 4.279189 kHz | 0.1 |  |
| 17   | -2.3563 dBm | -2.3134 dBm | -18.668 dBm | 5.1218622 us | 70.316162 ns | 69.885253 ns | 218.00520 us | 4.587046 kHz | 0.1 |  |
| 18   | -2.3558 dBm | -2.2804 dBm | -18.740 dBm | 5.1066395 us | 70.247189 ns | 69.825440 ns | 220.96526 us | 4.525598 kHz | 0.1 |  |
| 19   | -2.3564 dBm | -2.2648 dBm | -18.195 dBm | 7.2042967 us | 71.071774 ns | 70.909422 ns | 275.35643 us | 3.631656 kHz | 0.1 |  |
| 20   | -2.3509 dBm | -2.2886 dBm | -18.573 dBm | 5.6055314 us | 70.825806 ns | 70.736085 ns | 233.68525 us | 4.279260 kHz | 0.1 |  |
| 21   | -2.3566 dBm | -2.2565 dBm | -18.669 dBm | 5.1220964 us | 71.097411 ns | 70.708005 ns | 218.00013 us | 4.587153 kHz | 0.1 |  |
| 22   | -2.3541 dBm | -2.2646 dBm | -18.735 dBm | 5.1116199 us | 71.083370 ns | 69.742434 ns | 220.97304 us | 4.525439 kHz | 0.1 |  |

## Displaying the Pulse Table

1. Press the **Displays** button or select **Setup > Displays**.
2. Select **Pulsed RF** in the Measurements box.
3. **Double-click Pulse Table** in the **Available displays** box to add the display to the Selected displays box.
4. Click **OK**.

## Selecting the Measurements to Show

1. With the Pulse Table selected, press the **Settings** button.
2. In the Pulse Table Settings control panel, select the **Measurements** tab.



3. Select the measurements you wish to take from the list of measurements or click **Select all** to choose all measurements.
4. Click the close box ( ) to remove the Settings control panel.

[Changing the Pulse Table Display Settings \(see page 320\)](#)

## Pulse Table Settings

**Menu Bar:** Setup > Settings

**Front Panel:** Settings

**Application Toolbar:** 


The Settings control panel tabs for the Pulse Table Display are shown in the following table.

| Settings tab                                   | Description  |
|--|--|
| <a href="#">Measurements (see page 321)</a>    | Selects the measurements to be show in the Pulse Table.                                      |
| <a href="#">Params (see page 326)</a>          | Specifies several parameters that control how pulses are counted and defined.                |
| <a href="#">Define (see page 328)</a>          | Specifies parameters that control where measurements are taken on a pulse.                   |
| <a href="#">Levels (see page 333)</a>          | Specifies parameters that control the method and levels used to calculate some pulse values. |
| <a href="#">Freq Estimation (see page 333)</a> | Specifies the reference used for computing frequency errors.                                 |

**Restore defaults.** Sets the Pulse Table parameters to their default values.

## Measurements Tab

The Measurements tab is used to specify the measurements that appear in the Pulse Table.



### Show in Pulse

Checked measurements appear in the Pulse Table.

### Select all

Click **Select all** to choose all measurements for display in the Pulse Table.


### Clear all

Click **Clear all** to remove all measurements from the Pulse Table display.

## Pulse Trace Display

The Pulse Trace display shows one measurement result and a trace graph illustrating that measurement for a selected pulse number. You can also choose to display arrows and lines in the graph that illustrate where on the pulse the measurement is being taken.

## Elements of the Pulse Trace Display



| Item | Display element             | Description  |
|------|-----------------------------|--|
| 1    | Check mark indicator        | The check mark indicator in the upper, left-hand corner of the display shows when the Pulse Trace display is the optimized display.<br><br><b>NOTE.</b> When <i>Best for multiple windows</i> is selected in the <i>Amplitude control</i> panel's <i>RF &amp; IF Optimization</i> control, none of the measurement displays shows a checkmark, as there is not a single optimized measurement. |
| 2    | Result                      | Use this list to select the measurement to show. The measurement chosen here selects the same result to be displayed in the Statistics view and highlighted in the Table view. Selecting a result in any of these views causes the same result to be selected in all of them.  |
| 3    | Pulse                       | Selects which pulse result is shown. Like the Result, this selection is shared with the Statistics and Table views. The maximum number of pulses that can be measured is limited to 10000.   |
| 4    | Pulse threshold indicator   | This green triangle indicates the power threshold used to detect pulses. See Settings > Params > Power threshold to detect pulses.   |
| 5    | Autoscale                   | Set the axes to values that show all trace points.   |
| 6    | Horizontal offset and scale | Located along the bottom edge of the trace display, these controls specify the offset and scale on the horizontal axis.  |
| 7    | Top of graph adjustment     | Sets the power level shown at the top of the graph. If the number is greater than the Ref Level, a white line will appear in the graph to indicate the Ref Level.  |
| 8    | Vertical scale              | Sets the vertical scale value.   |

[Changing the Pulse Trace Display Settings \(see page 323\)](#)

## Pulse Trace Settings

**Menu Bar:** Setup > Settings

**Front Panel:** Settings

**Application Toolbar:** 


The Setup settings for the Pulse Trace display are shown in the following table.

| Settings tab                                   | Description  |
|--|--|
| <a href="#">Params (see page 326)</a>          | Specifies several parameters that control how pulses are counted and defined.                |
| <a href="#">Define (see page 328)</a>          | Specifies parameters that control where measurements are taken on a pulse.                   |
| <a href="#">Levels (see page 333)</a>          | Specifies parameters that control the method and levels used to calculate some pulse values. |
| <a href="#">Freq Estimation (see page 333)</a> | Specifies the reference used for computing frequency error.                                  |
| <a href="#">Scale (see page 335)</a>           | Specifies the vertical and horizontal scale settings.  |
| <a href="#">Prefs (see page 336)</a>           | Specifies whether or not certain display elements are shown.                                 |

## Pulse Statistics Display

The Pulse Statistics view displays a plot of a pulse measurement's values for every pulse in the analysis period. For example, the Pulse Statistics display will show the rise time measurement for each of 30 pulses, with rise time on the Y axis and pulse number on the X axis. Alternatively, the Statistics Display can show an FFT trace for the values of a measurement over all the pulses analyzed. Numeric readouts of Max, Min, and Average in this display summarize results for the selected pulse measurement.

### Elements of the Pulse Statistics Display



| Item | Display element      | Description  |
|------|----------------------|--|
| 1    | Check mark indicator | The check mark indicator in the upper, left-hand corner of the display shows when the Pulse Statistics display is the optimized display.<br><br><b>NOTE.</b> When Best for multiple windows is selected in the Amplitude control panel's RF & IF Optimization control, none of the measurement displays shows a checkmark, as there is not a single optimized measurement. |
| 2    | Result               | Use this list to select which measurement statistics to display. The measurement chosen here selects the same result to be displayed in the Pulse Trace display and the Pulse Table display. Selecting a result in any of these views causes the same result to be selected in all of them.  |
| 3    | Pulse                | Selects which pulse's result is shown. Like the Result, this selection is shared with the Statistics and Table views.  |
| 4    | Statistics summary   | Display of measurement statistics for the selected Result.   |
| 5    | Vertical range       | Located at the top of the left graph edge and at the bottom of the left graph edge, use these controls to adjust the values at the top of the graph and the bottom of the graph.   |
| 6    | Plot                 | Selects from Trend, Time Trend, FFT, and Histogram for the graph.  |
| 7    | Autoscale            | Set the axes to values that show all trace points.   |
| 8    | Horizontal offset    | Adjusting this value moves the graph right or left.  |
| 9    | Trigger indicator    | Indicates the trigger point on the pulse train if the trigger occurred during the time frame of the pulse measurements.  |
| 10   | Pulse indicator      | For Trend and Time Trend plot types, this indicates the pulse selected by the Pulse setting. For the FFT plot type, this indicator marks the trace point at the selected frequency. For Histogram plot types, this indicator marks the selected results bin.   |
| 11   | Scale                | Adjusts the horizontal scale.  |

[Changing the Pulse Statistics Display Settings \(see page 325\)](#)

## Pulse Statistics Settings

**Menu Bar:** Setup > Settings

**Front Panel:** Settings

**Application Toolbar:** 

The settings for the Pulse Statistics Display are shown in the following table.

| Settings tab                                   | Description  |
|--|--|
| <a href="#">Params (see page 326)</a>          | Specifies several parameters that control how pulses are counted and defined.                |
| <a href="#">Define (see page 328)</a>          | Specifies parameters that control where measurements are taken on a pulse.                   |
| <a href="#">Levels (see page 333)</a>          | Specifies parameters that control the method and levels used to calculate some pulse values. |
| <a href="#">Freq Estimation (see page 333)</a> | Specifies the reference used for computing frequency errors.                                 |
| <a href="#">Scale (see page 335)</a>           | Specifies the vertical and horizontal scale settings.  |
| <a href="#">Prefs (see page 336)</a>           | Specifies whether or not certain elements of the display are shown.                          |

## Pulsed RF Shared Measurement Settings

The control panel tabs in this section are shared by the displays in the Pulsed RF folder (Setup > Displays). Changing a setting on one tab changes that setting for all the Pulsed RF displays.

### Common controls for pulsed RF displays

| Settings tab                                   | Description  |
|--|--|
| <a href="#">Params (see page 326)</a>          | Specifies several parameters that control how pulses are counted and defined.                |
| <a href="#">Define (see page 328)</a>          | Specifies parameters that control where measurements are taken on a pulse.                   |
| <a href="#">Levels (see page 333)</a>          | Specifies parameters that control the method and levels used to calculate some pulse values. |
| <a href="#">Freq Estimation (see page 333)</a> | Specifies the reference used for computing frequency errors.                                 |
| <a href="#">Scale (see page 335)</a>           | Specifies the vertical and horizontal scale settings.  |
| <a href="#">Prefs (see page 336)</a>           | Specifies whether or not certain elements of the display are shown.                          |

## Params Tab

The Params tab enables you to adjust several measurement parameters for Pulsed RF displays.

The screenshot shows the 'Params' tab selected in a control panel. The panel has five tabs: 'Params', 'Define', 'Levels', 'Freq Estimation', and 'Prefs'. The 'Params' tab is active. It contains the following settings:

- Measurement Filter:** A dropdown menu set to 'Gaussian'.
- Power threshold to detect pulses:** A text input field containing '-12 dBc'.
- Bandwidth:** A text input field containing '10 MHz'.
- Minimum OFF time between pulses:** A text input field containing '0.000 s'.
- Max number of pulses:** A checkbox labeled 'Max number of pulses:' followed by a text input field containing '100'.

| Setting                          | Description  |
|----------------------------------|--|
| Measurement Filter               | Specify whether a filter is used to limit the bandwidth of the input signal.                                       |
| Bandwidth                        | Sets the bandwidth of the measurement filter.  |
| Power threshold to detect pulses | Specifies the level used for locating pulses in the data.  |
| Minimum OFF time between pulses  | Specifies the time the signal must fall below the power threshold for two pulses to be considered separate pulses. |
| Max number of pulses             | Specifies the number of pulses to measure within the analysis time.  |

### Measurement Filter

Three choices are available for the measurement filter:

- **No Filter - Max BW:** The widest acquisition bandwidth available is used. The **Bandwidth** setting is disabled, but shows the value in use.
- **No Filter:** - This is the default. The Bandwidth control is enabled for you to specify an acquisition bandwidth.
- **Gaussian** - The Bandwidth control is enabled for you to specify a value. The instrument uses an acquisition bandwidth two times wider than the entered value.

### Power Threshold to Detect Pulses

Specifies the minimum power level the trace must exceed to be detected as a pulse. The range for this setting is: -3 to -70 dBc. The setting resolution is 1 dB. The default value is -10 dB.

### Max Number of Pulses

If the Analysis Time contains fewer pulses than specified, the analyzer will measure all the pulse within the analysis time. If there are more pulses in the Analysis Time, than the specified number, the analyzer measures the specified number of pulses and ignores the rest. The range for this setting is: 1 - 10000. The setting resolution is: 1. If this setting is not checked, the analyzer will measure all pulses within the Analysis Time, up to a maximum of 10000 pulses.

To determine the maximum number of pulses that can be analyzed, use the following equation:

$$\text{Max number of pulses that can analyzed} = \text{Pulse Rate} \times \text{capacity}$$

where:

- Pulse rate is the number of pulses per second (frequency).
- Capacity is a length of time which is displayed on the **Acquire > Sampling Parameters** tab.

Note that the maximum number of pulses is affected by several parameters. For example, measurement bandwidth affects the sample rate. The measurement algorithm also can reduce the maximum number of pulses that can be analyzed (by increasing the sample rate) based on the characteristics of the signal.

Additionally, when FastFrame is enabled, determining the maximum number of pulses is even more challenging. In FastFrame mode, the signal analyzer samples the signal around events of interest and

ignores the signal between events of interest. Thus, if the instrument is only looking at pulses and ignoring the signal between pulses, the number of pulses that can be analyzed depends strongly on the characteristics of the pulse itself (for example, fewer wide pulses can be analyzed than narrow pulses, all other things being equal).

## Define Tab

The Define tab enables you to specify parameters that control where measurements are made on a pulse. The settings available depend on the measurement selected.

The screenshot shows the 'Define' tab for the 'Impulse Response' measurement. The interface includes tabs for Params, Define, Levels, Freq Estimation, Scale, and Prefs. The 'Measurement(s):' dropdown is set to 'Impulse Response'. The 'Time method:' dropdown is set to 'Absolute'. The 'Ref (R):' is set to '100 %'. The 'Start (S):' is set to '0.000 s'. The 'Length (L):' is set to '1.00 us'. There is a checkbox for 'Apply ampl corrections' which is checked. The 'Keep-out time (+/-):' is set to '0.000 s'. A diagram of a pulse waveform is shown with markers R, S, and L. A small inset shows a waveform with a peak and a trough.

The following paragraphs describe the settings that appear on the Define tab according to the Measurement(s) selection.

### Freq-Domain Linearity

The screenshot shows the 'Define' tab for the 'Freq-domain Linearity' measurement. The interface includes tabs for Params, Define, Levels, Freq Estimation, Scale, and Prefs. The 'Measurement(s):' dropdown is set to 'Freq-domain Linearity'. The 'Time method:' dropdown is set to 'Absolute'. The 'Ref (R):' is set to '100 %'. The 'Start (S):' is set to '0.000 s'. The 'Length (L):' is set to '1.00 us'. A diagram of a pulse waveform is shown with markers R, S, and L.

| Setting              | Description  |
|----------------------|--|
| Time method          | Specifies how the measurement's duration is determined. The choices are Absolute and Relative.   |
| Absolute time method |  |
| Ref (R)              | Ref specifies the level along the pulse rising edge that is defined as the point from which Start time is counted. Choices: 50%, 60%, 70%, 80%, 90% 100%; Default: 100%  |
| Start (S)            | Start specifies the time the instrument waits after the rising-edge reference before starting to measure the pulse. The Start time is measured from the point along the pulse rising edge specified by the Ref setting. Start Range: $\pm 100$ ms; Resolution: 3 digits; inc/dec small: 1 ns, large: 1,2,5,10...; Default: 0 |
| Length (L)           | Length specifies the period of time that is used for pulse measurements. The measurement time begins at the Start point and continues for the amount of time specified by Length.  |
| Relative time method |  |
| Length               | In the Relative time method, Length specifies the percentage of the top of the pulse that is used for measurements. The instrument automatically determines the pulse top. The measurement length is centered within the pulse ON time.  |

## Impulse Response

Params

Define

Levels

Freq Estimation

Scale

Prefs

Measurement(s):  

Impulse Response

☒ Apply ampl corrections  
Keep-out time (+/-): 0.000 s

Time method: 

Absolute


Ref (R): 100 %

Start (S): 0.000 s

Length (L): 1.00 us


| Setting                | Description  |
|------------------------|--|
| Apply ampl corrections | Enable/disable corrections that remove errors due to the window function and to the time offset of the side lobe.  |
| Keep-out time (+/-)    | The Keep-out time specifies a region that is ignored when the trace is analyzed for side lobes. The setting defines a region to the left and to the right of the center of the main lobe. Lobes that fall within this time region are not eligible to be the "highest side lobe".  |
| Time method            | Specifies how measurement parameters are determined. The choices are Absolute and Relative.  |
| Absolute time method   |  |
| Ref (R)                | Ref specifies the level along the pulse rising edge that is defined as the point from which Start time is counted. Choices: 50%, 60%, 70%, 80%, 90% 100%; Default: 100%  |
| Start (S)              | Start specifies the time the instrument waits after the rising-edge reference before starting to measure the pulse. The Start time is measured from the point along the pulse rising edge specified by the Ref setting. Start Range: $\pm 100$ ms; Resolution: 3 digits; inc/dec small: 1 ns, large: 1,2,5,10...; Default: 0 |
| Length (L)             | Amount of time that should be included in the measurement. The measurement time begins at the Start point and continues for the amount of time specified by Length.  |
| Relative time method   |  |
| Length                 | In the Relative time method, Length specifies the percentage of the top of the pulse that is used for measurements. The instrument automatically determines the pulse top. The measurement length is centered within the pulse ON time.  |

## Freq Diff, Phase Diff



| Setting       | Description   |
|---------------|---|
| Measure point | Specifies the period in time after the 50% rising edge at which frequency and phase difference measurements are made. |

## Ripple

| Params          | Define | Levels  | Freq Estimation | Scale              | Prefs |
|-----------------|--------|---|-----------------|--------------------|-------|
| Measurement(s): |        |   |                 |                    |       |
| Ripple          |        | Time method: Relative   |                 | Ref (R): 100 %     |       |
|                 |        |  |                 | Start (S): 0.000 s |       |
|                 |        |   |                 | Length (L): 50.0 % |       |


### Setting

### Description

Length

Length specifies the percentage of the top of the pulse that is used for measurements. The instrument automatically determines the pulse top. The measurement length is centered within the pulse ON time. Only the Relative Time method is available for Ripple.


## Overshoot

| Params          | Define | Levels   | Freq Estimation | Scale              | Prefs |
|-----------------|--------|--|-----------------|--------------------|-------|
| Measurement(s): |        |  |                 |                    |       |
| Overshoot       |        | Time method: Relative  |                 | Ref (R): 100 %     |       |
|                 |        |  |                 | Start (S): 0.000 s |       |
|                 |        |  |                 | Length (L): 25.0 % |       |

| Setting              | Description  |
|----------------------|--|
| Time method          | Specifies how measurement parameters are determined. The choices are Absolute and Relative.  |
| Absolute time method |  |
| Ref (R)              | Ref specifies the level along the pulse rising edge that is defined as the point from which Start time is counted. Choices: 50%, 60%, 70%, 80%, 90% 100%; Default: 100%  |
| Start (S)            | Start specifies the time the instrument waits after the rising-edge reference before starting to measure the pulse. The Start time is measured from the point along the pulse rising edge specified by the Ref setting. Start Range: $\pm 100$ ms; Resolution: 3 digits; inc/dec small: 1 ns, large: 1,2,5,10...; Default: 0 |
| Length (L)           | Amount of time that should be included in the measurement. The measurement time begins at the Start point and continues for the amount of time specified by Length.  |
| Relative time method |  |
| Ref                  | Ref specifies the level along the pulse rising edge that is defined as the point from which Start time is counted. Choices: 50%, 60%, 70%, 80%, 90% 100%; Default: 100%  |
| Start                | Start specifies the time the instrument waits after the rising-edge reference before starting to measure the pulse. The Start time is measured from the point along the pulse rising edge specified by the Ref setting. Start Range: $\pm 100$ ms; Resolution: 3 digits; inc/dec small: 1 ns, large: 1,2,5,10...; Default: 0 |
| Length               | In the Relative time method, Length specifies the percentage of the top of the pulse that is used for measurements. The instrument automatically determines the pulse top. The measurement length is centered within the pulse ON time.  |

## Levels Tab

Use the Levels tab to set parameters that control the method and levels used to calculate some pulse values.



| Setting          | Description   |
|------------------|---|
| 100% Level       | Specifies the method used to determine the 100% level(s).   |
| 50% Level        | Specifies the method used to determine the 50% level on the pulse.  |
| Rise/Fall Levels | Select whether to use the 10% to 90% or 20% to 80% points (based on voltage level) to define the rise and fall times. |

### 100% Level

Use the 100% Level settings to select the method used to determine the 100% level(s) used for calculating pulse parameters, for example, Rise, Fall, and Width.

The Pulse Average Amplitude defines the pulse top as the average of the values of all the points along the pulse top. This average is used as the 100% level, from which the 10, 20, 50, 80 and 90% levels are calculated. Pulse measurements are referenced against these various levels. For example, Rise is the time between the 10 and 90% (or 20 and 80%) levels on the rising edge of the pulse. When the Pulse Average Amplitude method is selected, the same 100% level is used for both rising and falling edges.

Because some RF pulse types have droop (a height difference between the beginning and ending points of the pulse top), the 100% percent level on the rising edge may not be equal to the 100% level on the falling edge. The Independent method of pulse point location is designed for pulses with different 100% levels at their rising and falling edges. The Independent method calculates the 100% level for the rising edge separately from the 100% level of the falling edge. As a result, the 10, 20, 50, 80 and 90% levels are also different for the rising and falling edges, allowing for more accurate measurements on pulses with droop.

### 50% Level

Select Voltage to use -6 dB as the 50% point. Select Power to use -3 dB as the 50% level.

## Freq Estimation Tab

Use the Freq Estimation tab to specify parameters used for determining frequency offset.

Params Define Levels Freq Estimation Scale Prefs

Modulation type:  
Linear Chirp

Pulse Frequency Reference

Freq Offset: 169.1 Hz ☒ Auto

Chirp Slope: 0.0000 Hz/us ☒ Auto

| Setting                   | Description  |
|---------------------------|--|
| Modulation type           | Specifies which algorithm to use for estimating frequency error.   |
| Pulse Frequency Reference | Specifies the method used to determine the pulse frequency error and if applicable, the chirp slope.   |
| Auto                      | Selecting Auto causes the instrument to calculate the frequency offset and if applicable, the chirp slope.   |
| Freq Offset               | If Auto is not enabled, specify the value for frequency offset here. If Auto is enabled, the offset is set to zero and this readout displays the calculated frequency error.                                     |
| Chirp BW                  | If Auto is not enabled, specify the value for Chirp Slope here. If Auto is enabled, this readout displays the calculated Chirp slope. This setting is used only when the modulation type is set to Linear Chirp. |

## Modulation Type

Frequency estimation is performed by the instrument using selectable methods, depending on signal type. The selections for modulation type are CW (constant phase), CW (changing phase), Linear Chirp and Other. Select the method of frequency method based upon a best match to your signal based on the following descriptions:

- CW (constant phase): The signal is not designed to change in either frequency or phase during the measured pulse train.
- CW (changing phase): The signal does not change the carrier phase within each pulse, although it could change the phase from one pulse to another pulse. The signal is not designed to make frequency changes.
- Linear Chirp: The signal changes frequency in a linear manner during each pulse. The signal has the same carrier phase at the same time offset from the rising edge of the pulse.
- Other: The signal is not one of the listed types. You must manually enter the Frequency Offset value.

The following table maps the appropriate signal type selection with the signal characteristics.

| Signature | Phase offset from one pulse to another |                   |
|-----------|--|-------------------|
|           | Zero                                   | Any (unknown)     |
| CW        | CW Constant Phase                      | CW Changing Phase |
| LFM       | Linear Chirp                           | N/A               |

## Scale Tab


The Scale tab allows you to change the vertical and horizontal scale settings. Changing the scale settings changes how the trace appears on the display but does not change control settings such as Measurement Frequency. In effect, these controls operate like pan and zoom controls.

The screenshot shows the 'Scale' tab in a software interface. It has two main sections: 'Vertical' and 'Horizontal'. The 'Vertical' section has a 'Scale' input set to '100.00 dB' and a 'Position' input set to '2.73 dBm'. The 'Horizontal' section has a 'Scale' input set to '348 ns', a 'Position' input set to '6.99 ms', and a 'Full scale' section with two radio buttons: 'Selected' (which is selected) and 'Max pulse'. There are 'Autoscale' buttons for both sections and a 'Reset Scale' button in the center.

| Setting                               | Description   |
|---------------------------------------|---|
| Vertical                              | Controls the vertical position and scale of the trace display.  |
| Scale                                 | Changes the vertical scale. The units for this setting depend on the statistic selected from the Result drop-down list in the Pulse Statistics display.   |
| Position                              | Adjusts the Reference Level away from the top of the trace display. The units for this setting depend on the statistic selected from the Result drop-down list in the Pulse Statistics display. |
| Autoscale                             | Resets the scale of the vertical axis to contain the complete trace.  |
| Horizontal                            | Controls the span of the trace display and position of the trace.   |
| Scale                                 | Allows you to change the span.  |
| Position                              | Allows you to pan a zoomed trace.   |
| Full Scale (Pulse Trace display only) | Specifies the Horizontal scale default.   |
| Selected                              | Sets the horizontal scale default to be based on the result value for the currently-select pulse.   |
| Max Pulse                             | Sets the horizontal scale default to be based on the largest value for the selected pulse measurement.  |
| Autoscale                             | Resets the scale of the horizontal axis to contain the complete trace.  |
| Plot (Pulse Statistics display only)  | Specifies the FFT, Trend, Time Trend, or Histogram plot.  |
| Reset Scale                           | Restores all settings to their default values.  |

# Prefs Tab

The Prefs tab enables you to change parameters of the measurement display. The parameters available on the Prefs tab vary depending on the selected display.



## Settings on the Pulsed RF Displays Prefs Tab

The following table describes the settings that appear on the Prefs tab of the Pulsed RF displays: Pulse Statistics and Pulse Trace. There is no Prefs tab for the Pulse Table display.

| Setting   | Description   |
|---|---|
| Show graticule  | Displays or hides the graticule in the trace display.   |
| Show Marker readout in graph  | When a marker is enabled, this setting displays or hides the maker readout, but not the maker itself, on Time Trend plots.  |
| Histogram readout (present only when Histogram is the selected plot type) | Controls the parameters Histogram readout and Bins. Histogram readout can be set to either Count or %. Count indicates the number of hits that fell into each bin. % indicates percentage of the total count (for the acquisition) that fell into each bin. |
| Bins  | Specifies how many "bins" or histogram bars the results are distributed into.   |

# APCO P25 Analysis

## Overview

The Association of Public Safety Communication Officials (APCO) P25 Compliance Testing and Analysis option allows you to evaluate radio signals to ensure they meet the standards set for the public safety communications community for interoperable LMR equipment. This complete set of push-button Telecommunication Industry Association TIA-102 standard-based transmitter measurements includes modulation measurements, power measurements, and timing measurements. These measurements are also compared with the limits that best fit the signal for which the standard applies to provide pass/fail results.

The P25 measurements available with this option can be made on signals defined by the Phase 1 (C4FM) and Phase 2 (HCPM, HDQPSK) P25 standards. With this test suite, test engineers can simplify the execution of a number of transmitter tests while still allowing for controls to modify signal parameters for signal analysis. The analysis results give multiple views of P25 signal characteristics to allow the diagnosis of signal imperfections and impairments quickly and easily. Display controls allow you to selectively display the analysis results to help locate trouble-spots in the signal.

## P25 Topics

The following information about the P25 Analysis option is available:

- [Reference Table of Supported P25 Measurements \(see page 337\)](#)
- [P25 Standards Presets \(see page 338\)](#)
- [P25 Displays \(see page 339\)](#)
- [P25 Settings \(see page 401\)](#)
- [P25 Measurements \(see page 340\)](#)
- [P25 Test Patterns \(see page 348\)](#)

## Reference Table of Supported P25 Measurements

| TIA-102 transmitter measurement                               | Tektronix measurement display(s) (value name if different from TIA-102 measurement name) | Phase 1 | Phase 2 |
|---|--|---------|---------|
| RF output power   | P25 Summary<br>P25 Power vs Time   | Yes     | Yes     |
| Operating frequency accuracy                                  | P25 Summary (Operating Freq Accuracy)  | Yes     | Yes     |
| Modulation emission spectrum                                  | SEM (The SEM display can be found in Select Displays > RF Measurements)                  | Yes     | Yes     |
| Unwanted emissions: Non spurious adjacent channel power ratio | MCPR (The MCPR display can be found in Select Displays > P25 Analysis)                   | Yes     | Yes     |

| TIA-102 transmitter measurement                                     | Tektronix measurement display(s) (value name if different from TIA-102 measurement name)   | Phase 1 | Phase 2 |
|---|--|---------|---------|
| Frequency deviation   | P25 Summary (Freq Dev)   | Yes     | HCPM    |
| Modulation fidelity   | P25 Summary<br>P25 Constellation   | Yes     | Yes     |
| Symbol rate accuracy  | P25 Summary  | Yes     | Yes     |
| Transmitter power and encoder attack time                           | P25 Summary (Phase1 Tx Attack Time)<br>P25 Power vs Time (Power Attack Time and Encoder Attack Time)                                 | Yes     | N/A     |
| Transmitter power and encoder attack time with busy/idle operations | P25 Summary (Phase1 Tx Attack Time (Busy/Idle))<br>P25 Power vs Time (Power Attack Time Busy Idle and Encoder Attack Time Busy Idle) | Yes     | N/A     |
| Transmitter throughput delay  | P25 Summary (Phase1 Tx Throughput Delay)   | Yes     | N/A     |
| Transient frequency behavior  | P25 Freq Dev vs. Time  | Yes     | N/A     |
| HCPM transmitter logical channel peak adjacent channel power ratio  | P25 Summary (HCPM Tx Logic Ch Pk ACPR)   | N/A     | HCPM    |
| HCPM Transmitter logical channel off slot power                     | P25 Power vs Time (Off Slot Power)<br>P25 Summary (HCPM Tx Logic Ch Off Slot)  | N/A     | HCPM    |
| HCPM Transmitter logical channel power envelope                     | P25 Power vs Time (Power Info)<br>P25 Summary (HCPM Tx Logic Ch Pwr Env Limits)  | N/A     | HCPM    |
| HCPM Transmitter logical channel time alignment                     | P25 Summary (HCPM Tx Logic Ch Time Alignt)   | N/A     | HCPM    |

## P25 Standards Presets

The P25 standards preset allows you to access displays preconfigured for the P25 standards you select. You can read more about how Presets work [here \(see page 22\)](#).


The following table shows the bandwidth, modulation type, and displays that are automatically loaded for each of the listed standards. MCPR masks are also loaded and are explained [here \(see page 339\)](#).

**Table 5: P25 standards, modulation type, bandwidth, and displays**

| Standard | Modulation type   | Bandwidth (kHz) | Displays loaded with preset                         |
|----------|-------------------|-----------------|---|
| Phase 1  | C4FM              | 12.5            | MCPR, Time Overview, P25 Constellation, P25 Summary |
| Phase 2  | HCPM (inbound)    | 12.5            | MCPR, Time Overview, P25 Constellation, P25 Summary |
|          | HDQPSK (outbound) | 12.5            |   |

## Retain Current Center Frequency

This setting becomes available when P25 is the selected preset in the Standards Preset window. You can access this window by selecting **Presets > Standards** and then selecting **P25** from the **Preset** drop down menu. To activate the **Retain current Center Frequency Setting**, check the box. This setting allows you to retain the previously used center frequency. By default, the box is unchecked and therefore the four P25 preset displays will load with a center frequency of 850 MHz.



The default adjacent channels table for MCPR is different for RF frequencies in the range of 769 to 806 MHz (called 700 MHz band) and for frequency ranges outside it. The option of retaining center frequency in Standards Preset is therefore useful if you want to load the default table for center frequencies in the 700 MHz band.

**MCPR channel and limit parameters.** The MCPR (ACPR) standard channel and limit parameters that are applied to the P25 signal depend on the standard you select when you configure the preset. Once you select a standard and center frequency, the application will automatically load the parameters and default limits recommended for best performance comparison by the Standard document. All channel and limit parameters are derived from the TIA-102 standard and loaded for you. This provides you the assurance that you are evaluating the signal with the most appropriate parameters.

---

**NOTE.** *Changing analysis and display parameters recalculates the measurement results, but does not affect acquisitions.*

---

## P25 Displays

The displays in P25 Analysis (Setup > Displays > Measurements: P25 Analysis) are:

- [MCPR](#)
- [P25 Constellation \(see page 351\)](#)

- [P25 Power vs Time \(see page 356\)](#)
- [P25 Eye Diagram \(see page 354\)](#)
- [P25 Frequency Dev Vs Time \(see page 364\)](#)
- [P25 Symbol Table \(see page 361\)](#)
- [P25 Summary \(see page 358\)](#)
- [Time Overview \(see page 91\)](#)

## P25 Measurements

The following topics contain important information you should know about specific P25 measurements.

### RF Output Power (Phase 1 and Phase 2)

This is a measure of RF output power when the transmitter is connected to the standard load during defined duty cycle. This measurement is presented as a scalar result in the P25 Summary display. Power variation is shown in the P25 Power vs Time display.

Information of note about this measurement:

- For bursty HCPM (Phase 2 Inbound) signals, the RF Output Power is measured only during the on slot regions centered at the middle of the on slot.
- The result shown in the P25 Summary Display is the average RF Output Power of all the bursts selected in the analysis window.
- If only one on slot region is chosen in the analysis window, then the RF Output Power of only the chosen on slot is reported.
- For non-bursty data, the entire duration of analysis window is considered to obtain RF Output Power.
- Enter external attenuation introduced by the user by entering the attenuation in the Amplitude control panel (**Setup > Amplitude > External Gain/Loss Correction** tab. > **External Gain Value** (dB). Select **All acquired data (filter method)**.
- The Limits tab in the P25 Summary Settings control panel allows you to compare the results against limits set for Pass/Fail.
- This measurement result is not shown in the P25 Summary display when Trigger Measurements are selected.

---

**NOTE.** Read about important information related to HCPM bursty data measurements [here \(see page 350\)](#).

---

### Operating Frequency Accuracy (Phase 1 and Phase 2)

The Operating Frequency Accuracy is the ability of the transmitter to operate on its assigned frequency. This measurement is presented as a scalar result in the P25 Summary display.

- For bursty HCPM (Phase 2 Inbound) signals, the operating frequency accuracy is measured only during the on slot regions centered at the middle of the on slot.
- The result shown in the P25 Summary display is the average of all the bursts selected in the analysis window.
- If only one on slot region is chosen in the analysis window, then the operating frequency accuracy of only the chosen on slot is reported.
- For non-bursty data, the entire duration of the analysis window is considered to obtain the operating frequency accuracy.
- The Limits tab in the P25 Summary Settings control panel allows you to compare the results against limits set for Pass/Fail.
- This measurement result is not shown in the P25 Summary display when Trigger Measurements are selected.

### Unwanted Emissions (ACPR) (Phase 1 and Phase 2)

This measurement is the ratio of the total power of a transmitter under prescribed conditions and modulation to that of the output power that falls within a prescribed bandwidth centered on the nominal frequency of adjacent channels. This measurement is shown in the MCPR display.

- For HCPM signals, Adjacent Channel Power Ratio analysis is only done on the first on slot region. When the Modulation Type is set to HCPM from any P25 display or from Standards Presets, there will be a search for bursts before ACPR analysis is done. If a burst is found, then analysis is done only on first burst. If a burst is not found, it will show the results of the entire analysis length. The Main Preset will remove the HCPM modulation Type setting, allowing MCPR to work in its regular mode.
- The MCPR display will only have one Main channel for P25 signals in addition to relevant adjacent channels.
- The default adjacent channels table for Unwanted Emissions (ACPR) as suggested by the standard can be loaded by using the P25 Standards Preset option (for which MCPR is one of the four displays).
- The default adjacent channels table for MCPR is different for RF frequencies in the range of 769 to 806 MHz (called 700 MHz band) and for frequency ranges outside it.
- To load the default adjacent channels table of a 700 MHz band, first set the center frequency. Next, select **Presets > Standards** to view the Standards Preset window. Lastly, check the **Retain current Center Frequency Setting** box.
- By default, the Retain Center Frequency box is unchecked and therefore the four P25 preset displays will load with a center frequency of 850 MHz and load the adjacent channels table for non-700 MHz band.
- By default, only six adjacent channels (on either side) will be shown when 700 MHz band is analyzed in P25 (for visual clarity). However, analysis is done for all the ten channels and results will be available in the table just under the display. You can zoom out to see the remaining bands.

## Frequency Deviation (Phase 1 (C4FM) and Phase 2 (HCPM))

This measurement shows the amount of frequency deviation that results for a Low Deviation and High Deviation test pattern. This measurement is shown in the P25 Summary display.

- This measurement only applies to HCPM and C4FM signals and High Deviation or Low Deviation test patterns. If this measurement is run for any other test patterns or for HDQPSK signals, comparisons for limits (set in the Limits tab of the control panel) will result in N/A being shown, as the measurement will not be valid.
- This measurement is done using a FM demodulator without any filter. As a result, it will not be exactly the same as the result given in the P25 Frequency Deviation vs Time display. That display shows the frequency deviation after complete demodulation and might use relevant shaping filters.
- Measurement results are only available when High Deviation, Low Deviation, or Symbol Rate test pattern is selected.
- The scalar result in the P25 Summary display will be shown as f1 (Positive Peak) and f2 (Negative Peak) when analyzing the High Deviation test pattern (or the Symbol Rate test pattern for C4FM signals).
- The scalar result in the P25 Summary display will be shown as f3 (Positive Peak) and f4 (Negative Peak) when analyzing the Low Deviation test pattern.
- The scalar results in the P25 Summary display do not have a corresponding graphical view. A closely representative display would be the FM display (Setup > Displays > Analog Modulation).
- The Limits tab in the P25 Summary Settings control panel allows you to compare the results against limits set for Pass/Fail.

## Modulation Fidelity (Phase 1 and Phase 2)

This measures the degree of closeness to which the modulation follows the ideal theoretical modulation determined by the rms difference between the actual deviation and the expected deviation for the transmitted symbols. This measurement is shown in the P25 Summary and P25 Constellation displays.

- This measurement is done on the first on slot region in the analysis window for bursty HCPM data. The analysis is done on 160 symbols centered at the burst.
- This measurement is done on all of the data chosen in the analysis window for non-bursty modulation types. A warning message is issued if the data is less than 164 symbols (the standard recommends at least 164 symbols be present for the measurement of non-bursty signals).
- This measurement is done by performing the frequency demodulation after taking the signal through relevant shaping filters and comparing it with expected frequency deviation points.
- The Limits tab in the P25 Summary Settings control panel allows you to compare the results against limits set for Pass/Fail.
- This measurement result is not shown in the P25 Summary display when Trigger measurements are selected.

---

**NOTE.** Read about important information related to HCPM bursty data measurements [here \(see page 350\)](#).

---

## Symbol Rate Accuracy (Phase 1 and Phase 2)

Symbol Rate Accuracy measures the ability of the transmitter to operate at the assigned symbol rate (4.8 kHz for Phase 1, 6 kHz for Phase 2). It is a scalar result shown in the P25 Summary display.

- This measurement result appears in the P25 Summary display and only applies to High Deviation or Low Deviation test patterns. Test patterns are selected in the Test Patterns tab in the Settings control panel.
- To check if the results are consistent with the symbol rate used, the measurement is done based on the number of zero crossings when a High Deviation or Low Deviation test pattern is used.
- The Limits tab in the P25 Summary Settings control panel allows you to compare the results against limits set for Pass/Fail.
- This measurement only applies to High Deviation or Low Deviation test patterns. If this measurement is run for any other test patterns, comparisons for limits will result in N/A being shown, as the measurement will not be valid.

## Common Trigger Related Measurements

The measurements covered in this subsection have some common information. These measurements are:


- **Transmitter Power and Encoder Attack Time (Phase 1):** This measurement is the time required for a transmitter to prepare and transmit information on the radio channel after changing state from standby to transmit (applies to conventional mode). This measurement result can be seen in the P25 Summary display. A graphical representation of these results can be seen in the P25 Power vs Time display.
- **Transmitter Power and Encoder Attack Time with Busy/Idle Operations (Phase 1):** This measures the time required for a transmitter to prepare and transmit information on the radio channel after the receiving channel changes state from busy to idle. Transmission is inhibited until a status symbol indicates an idle channel. This measurement result can be seen in the P25 Summary display. A graphical representation of these results can be seen in the P25 Power vs Time display.
- **Transmitter Throughput Delay (Phase 1):** This measures the time it requires for audio changes in the microphone to be encoded and transmitted over the air. A calibrated receiver with a known receiver throughput delay is used to monitor the transmitted signal. The aggregate delay of the transmitter under test and calibrated receiver is measured and the desired transmitter throughput delay is then the aggregate delay less the delay of the calibrated receiver. A calibrated Receiver throughput delay from the UI is needed. This measurement result can be seen in the P25 Summary display.
- **Transient Frequency Behavior (Phase 1):** This is a measure of difference of the actual transmitter frequency and assigned transmitter frequency as a function of time when the RF output power is switched on or off. This measurement appears in the P25 Freq Dev vs Time display.
- **HCPM Tx Logical Channel Time Alignment (Phase 2 HCPM):** This measures the ratio of total transmitter power under prescribed conditions and modulation to the peak power that falls in a prescribed bandwidth centred on the nominal frequency of the adjacent channel during the transmitter power ramping interval. This applies only to inbound signals. This measurement result appears in the P25 Summary display.

**Common information.** The following information applies to all of the measurements in the previous list.

- Select the relevant Trigger measurement from the Trig Meas tab of the control panel.
- When this measurement is chosen from the Trig Meas tab of the control panel, only this result will be populated in the P25 Summary display and everything else will have no result.
- The Time Zero Reference under the Analysis Time tab in the control panel is forced to Trigger.
- Align the Trigger to the RSA/MDO4000B as described in the measurement methods standard document.
- When a trigger measurement is chosen from the Trig Meas tab of the control panel, the following status message will appear: *P25: Set instrument in Triggered Mode.*
- You must set the trigger mode from Free Run to Triggered in the Trigger control panel.
- This measurement can only be performed when the Source is Trig In (front) or Trig 2 In (rear).
- Use the Acquisition control panel to select enough acquisition length to ensure that sufficient data is available when the acquisition is triggered.
- This measurement operates in Single Acquisition mode as the acquisition is triggered.
- You can select to do the same experiment multiple times and this measurement will produce an average result of the last 10 single acquisitions. Clicking the **Clear** button on the P25 Summary display will clear the results and start a new measurement.

**Transmitter Power and Encoder Attack Time and Transmitter Power and Encoder Attack Time with Busy/Idle Operations (Phase 1).** The following information is specific to these two measurements.

- The Transmitter Power Attack Time analysis is done by measuring the time taken from the trigger point to the point where the transmitter output power will reach 50% of its maximum value.
- For Encoder Attack Time, the initial frame synchronization word is searched in the demodulated output and the time taken from trigger to the start of the synchronization word is reported as the result.
- When a clear power ramp up is not available, the analysis will report the following error, indicating that no power ramp was received as expected: *Input data too short.* When the synchronization word is not found, the analysis will report the following error: *IQ Processing error.*
- The Limits tab in the P25 Summary Settings control panel allows you to compare the results against limits set for Pass/Fail.
- The P25 Power vs Time display shows marking from trigger point to 50% ramp point and also up to the Synchronization word for Transmitter Power and Encoder Attack Time. Frequency Dev vs Time can also be used to check for the synchronization word. The following image shows this measurement.



**Transmitter Throughput Delay (Phase 1).** The following information is specific to this measurement.


- Select Tx Throughput Delay from the Trig Meas tab of the Settings control panel to get this measurement.
- The original noise source is connected as input to this trigger and the acquisition is triggered when the input from the noise source exceeds the trigger level set by the user.
- The actual input to the RSA comes after the noise signal from the source has passed through the DUT and calibrated receiver as shown in the setup for this measurement in the measurement methods standard document.
- The level in the Trigger control panel set by the user is used for analysis to determine that Input to the RSA (after it has gone through the DUT and calibrated receiver) has hit the desired power levels.
- The time difference between the trigger point and the point when the input to the RSA has exceeded the level set by the user for triggering is calculated.
- The calibrated receiver throughput delay entered by the user in the Trig Meas tab in the Settings control panel (this option appears when Throughput Delay is selected) is subtracted from the calculated time difference and reported as Throughput Delay.
- A limit comparison can be performed by selecting the appropriate limits from the Limits tab in the Settings control panel.

**Transient Frequency Behavior (Phase 1).** The following information is specific to this measurement.

- Select Transmitter Frequency Behavior from the Trig Meas tab of the Settings control panel to get this measurement.
- This measurement can be done for On to Off and Off to On behaviors. Select which behavior to measure by selecting the desired behavior from the Transient Frequency drop down list that appears

when the Transient Frequency Behavior measurement is selected in the Trig Meas tab of the Settings control panel.

- Select enough acquisition length (from the Acquisition control panel) to ensure that sufficient data is available when the acquisition is triggered. For a transient frequency behavior, there is a need for  $t_1+t_2$  amount of time after  $t_{on}$  as defined in the measurement methods standard document. Approximately 100 ms of data after trigger should ensure that all cases are taken care of as suggested by the standard.  $t_1$ ,  $t_2$ , and  $t_3$  durations are fetched based on the RF frequency range.
- The identification of  $t_{on}$  is done by looking for a significant frequency deviation after a certain power level has been achieved. A manual override for the  $t_{on}$  is also provided in the Trigger Meas tab of the Settings control panel when a Transient Frequency Behavior measurement is chosen. This allows you to manually override the  $t_{on}$  that is calculated (by releasing the Auto Option) and place it appropriately based on the P25 Freq Dev vs Time display. The same is true for  $t_{off}$  when the measurement is being done for the On to Off behavior.
- $t_1$  and  $t_2$  regions are identified after  $t_{on}$  and  $t_3$  before  $t_{off}$  (for the On to Off behavior).
- The mean and max frequency deviation is reported in the regions identified.
- A Pass or Fail is also reported by comparing the Mean frequency deviation with the recommended values given in the standard. A red band is shown if a particular region's result is less than the performance recommendation. The following image shows this measurement.



**HCPM Tx Logical Channel Time Alignment (Phase 2 HCPM).** The following information is specific to this measurement. This result can be viewed in P25 Summary and P25 Power vs Time displays.

- This measurement is done in two steps: First by calculating  $tOB\_sync$  using HDQPSK data (the ISCH pattern is looked for) and then by using the result to calculate  $tIB\_sync$  using HCPM data (the SACCH pattern is looked for).

The first step is done by choosing Time Alignment ( $tOB\_sync$  measurement) in the Trig Meas tab in the Settings control panel.

When the second step, Time Alignment ( $t_{\text{error\_0}}$  calculation) or ( $t_{\text{error\_1}}$  calculation) is selected in the Trig Meas tab of the control panel, there is an option to override the  $tOB\_sync$  value calculated in the first step.

- When this measurement is chosen, only this result is provided in the P25 Summary display.
- You must provide the ultraframe boundary to the trigger of the analyzer.
- An autocorrelation of the input RF signal with the reference patterns is performed to get the peak that will determine  $tOB\_sync$  (measured) and  $t_{\text{error\_0}}$  or  $t_{\text{error\_1}}$  from the trigger point.

### HCPM Tx Logical Channel Peak ACPR (Phase 2 HCPM)

This measures the ratio of total transmitter power under prescribed conditions and modulation to the peak power that falls in a prescribed bandwidth centred on the nominal frequency of the adjacent channel during the transmitter power ramping interval. This applies only to inbound signals. This measurement result appears in the P25 Summary display.

- This measurement is done by calculating power in the adjacent channels for the entire duration of data chosen by the user, including the power ramp up and ramp down portions. The standard recommends 360 ms of data for this measurement, unlike the other ACPR measurements for which the analysis is done only for the on slot region.
- The higher and lower adjacent channel power is reported in the P25 Summary display under Power Measurements.
- The two results are then subtracted from the calculated RF output power and the minimum of the two results is presented as the Min Pk ACPR in the P25 Summary display.
- This measurement is only done for bursty HCPM data and not for High Deviation or Low Deviation test patterns.
- The Limits tab in the P25 Summary Settings control panel allows you to compare the results against limits set for Pass/Fail.
- This measurement result is not shown in the P25 Summary display when trigger measurements are chosen.

---

**NOTE.** Read about important information related to HCPM bursty data measurements [here \(see page 350\)](#).

---

### HCPM Transmitter Logical Channel Off Slot power (Phase 2 HCPM)

This measures the power of a TDMA transmitter during the off portion of the TDMA pulse. This measurement only applies to inbound signals. This result can be viewed in P25 Summary and P25 Power vs Time displays.

- The measurement is done as suggested by the standard to calculate  $P_{\text{ONREL}}$  and  $P_{\text{OFFREL}}$
- $P_{\text{TX}}$  is an input expected from the user and this can be given from the Analysis Params tab in the Settings control panel.
- The absolute level of the off slot power is then calculated as  $P_{\text{OFF}} = P_{\text{TX}} - (P_{\text{ONREL}} - P_{\text{OFFREL}})$  dBm

- This measurement is only done for bursty HCPM data and not for High Deviation or Low Deviation test patterns.
- The Limits tab in the P25 Summary Settings control panel allows you to compare the results against limits set for Pass/Fail.
- This measurement result is not shown in the P25 Summary display when trigger measurements are chosen.

---

**NOTE.** Read about important information related to HCPM bursty data measurements [here \(see page 350\)](#).

---

### HCPM Transmitter Logical Channel Power Envelope (Phase 2 HCPM)

This is a measure of how well a portable radio controls the transmitter power as it inserts an inbound HCPM TDMA burst into a frame on a voice channel. This measurement applies to inbound signals only.

- All scalar results relevant to this measurement are shown both in P25 Power vs Time display and the P25 Summary display.
- The measurement results and the duration in which the measurements are made are shown graphically in P25 Power vs Time display. The results are grouped under Power Info and Time Info in the table at the bottom of the display.
- This measurement is only done for bursty HCPM data and not for High Deviation or Low Deviation test patterns.
- The Limits tab in the P25 Summary Settings control panel allows you to compare the results against limits set for Pass/Fail.
- This measurement result is not shown in the P25 Summary display when trigger measurements are chosen.

---

**NOTE.** Read about important information related to HCPM bursty data measurements [here \(see page 350\)](#).

---

## P25 Test Patterns

A variety of test patterns are specified in the TIA-102 documents for use in performance testing of transmitters. These test patterns allow the software to compare the measurement result to the standards limit. Test engineers can select from the test patterns described in the following tables.

---

**NOTE.** Although this table gives the test patterns for measurements as recommended by the Standard, other measurement results are also be provided as additional information for a given test pattern. For example, Modulation Fidelity results can also be provided for High Deviation and Low Deviation test patterns.

---

**Table 6: P25 test patterns, Phase 1 C4FM**

| <b>TIA-102 Phase 1 C4FM test pattern</b>                                  | <b>Phase 1 (C4FM) measurement</b>   |
|---|---|
| Standard transmitter  | RF Output Power   |
|   | Operating Frequency Accuracy  |
|   | Modulation Emission Spectrum  |
|   | Unwanted Emissions (Adjacent Channel Power Ratio)   |
|   | Modulation Fidelity   |
|   | Transmitter Power and Encoder Attack Time   |
| Standard Busy   | Transmitter Power and Encoder Attack Time (With Busy/Idle Operations)   |
| Low Deviation   | Operating Frequency Accuracy  |
|   | Transient Frequency Behavior  |
|   | Frequency Deviation   |
| Standard Idle   | Transmitter Power and Encoder Attack Time (With Busy/Idle Operations)   |
| Standard Transmitter Symbol Rate<br>(Same as High Deviation test pattern) | Symbol Rate Accuracy  |
|   | Frequency Deviation   |
| C4FM Modulation Fidelity  | Modulation Fidelity   |
| Standard Tone   | Can be used for measurements such as RF Output Power, Operating Frequency Accuracy, Modulation Fidelity, and ACPR |
| Other<br>(User created test pattern)                                      | Can be used for measurements such as RF Output Power, Operating Frequency Accuracy, Modulation Fidelity, and ACPR |

**Table 7: P25 test patterns, Phase 2 Inbound and Outbound**

| <b>TIA-102 Phase 2 test pattern</b>         | <b>Phase 2 measurement</b>  |
|---|---|
| Standard Transmitter (Inbound and Outbound) | RF Output Power   |
|   | Operating Frequency Accuracy  |
|   | Modulation Emission Spectrum  |
|   | Unwanted Emissions (Adjacent Channel Power Ratio)   |
|   | Modulation Fidelity   |
| Inbound Standard Tone Ch0                   | Can be used for measurements such as RF Output Power, Operating Frequency Accuracy, Modulation Fidelity, and ACPR |
| Inbound Standard Tone Ch1                   | Can be used for measurements such as RF Output Power, Operating Frequency Accuracy, Modulation Fidelity, and ACPR |
| Outbound Standard Tone                      | Operational Frequency Accuracy  |

**Table 7: P25 test patterns, Phase 2 Inbound and Outbound (cont.)**

| <b>TIA-102 Phase 2 test pattern</b>   | <b>Phase 2 measurement</b>  |
|---------------------------------------|---|
| Inbound Symmetrical Time Slot         | RF Output Power   |
|                                       | Modulation Emission Spectrum  |
|                                       | Modulation Fidelity   |
|                                       | HCPM (Peak ACPR)  |
|                                       | HCPM (Off slot power)   |
|                                       | HCPM (Power envelope)   |
|                                       | Unwanted Emissions (Adjacent Channel Power Ratio)   |
| Low Deviation (Inbound)               | Frequency deviation for HCPM  |
|                                       | Symbol Rate Accuracy  |
| Low Deviation (Outbound)              | Symbol Rate Accuracy  |
| High Deviation (Inbound and Outbound) | Symbol Rate Accuracy  |
| Other<br>(User created test pattern)  | Can be used for measurements such as RF Output Power, Operating Frequency Accuracy, Modulation Fidelity, and ACPR |

### Analysis of HCPM Bursty Data

It important to take into account the following information when analyzing HCPM bursty data.

- It is mandatory that at least one complete on slot and off slot region (including the ramp up and ramp down portion) have to be included in the analysis window for HCPM measurement results to be analyzed.
- If HCPM is the chosen standard in Standards Presets, then a minimum length of 94 ms is set for the analysis window. This is to ensure that a full on slot and off slot region is available irrespective of the selected analysis offset.
- Do not use the High Deviation or Low Deviation test patterns for bursty signals. Test patterns are selected in the P25 Settings control panel on the Test Patterns tab. If selected, analysis would look for a bursty pattern. High Deviation and Low Deviation test patterns are not bursty and therefore do not have the above analysis length restriction.

If you try to set the analysis length to less than 94 ms, the following status message will display:  
*P25:HCPM:Minimum Analysis Length should be  $\geq 94$  ms.*

- Most measurements that are required to be done on on slot regions are done centered at the on slot region of HCPM bursty data.

When HCPM bursty data is analyzed, the following status message will appear on all relevant displays:  
*P25:HCPM modulation analysis is done only on on slot regions.*

## P25 Constellation Display


The P25 Constellation Display shows the P25 signal modulation amplitude. For the HCPM and HDQPSK standards, you can view this display as frequency deviation or as I and Q. For the C4FM standard, only the frequency deviation applies.

To show the P25 Constellation display you can select **Presets > Standards > P25** or do the following:


1. Press the **Displays** button or select **Setup > Displays**.
2. In the **Select Displays** dialog, select **P25 Analysis** in the **Measurements** box.
3. In the Available displays box, double-click the **P25 Constellation** icon or select the icon and click **Add**. The P25 Constellation icon will appear in the **Selected displays** box and will no longer appear under Available displays.
4. Click **OK** to show the P25 Constellation display.
5. Select **Setup > Settings** to display the control panel.
6. Select the **Modulation Params** tab. Set the Standard, Modulation Type, and Measurement and Reference filters as appropriate for the input signal.
7. For Phase 2 signals, select the **Trace** tab and set the **Trace Type** to Freq Dev or IQ.
8. Select the **Test Patterns** tab and select the desired test pattern as appropriate for the input signal.

### Elements of the Display

The following image shows the P25 Constellation display set to IQ for a HDQPSK signal.



The following image shows the same signal as the previous image, but with the display set to Frequency Deviation.



| Item | Display element             | Description   |
|------|-----------------------------|---|
| 1    | Marker Readout              | Located to the left of the constellation plot or below it, depending on the size of the window. If markers are enabled, the marker readout will show frequency deviation results or IQ results (time, mag, phase, symbol marker and symbol value) of the point with the selected marker.  |
| 2    | Measurement results readout | <p>These readouts are located to the left or below of the Constellation plot, depending on the window size. The readout shows Modulation Fidelity (%) and Frequency Error (Hz).</p> <p>The second readout can either be Freq Error where the result is followed by Auto or it could be Freq Offset where the result is followed by Manual. This choice between Freq Error and Freq Offset is made from the Analysis Params tab in the Settings control panel.</p> |
| 3    | Plot                        | Shown as either I vs Q or as Frequency Deviation. The trace type is controlled from the <b>Settings &gt; Trace</b> tab.   |

## P25 Constellation Settings

### Front Panel: Settings

#### Application Toolbar:

The P25 Constellation Settings control panel provides access to settings that control parameters of the Constellation Display.

| Settings tab                                     | Description   |
|--|---|
| <a href="#">Modulation Params (see page 368)</a> | Specifies the input signal standard and additional user-settable signal parameters.   |
| <a href="#">Analysis Params (see page 371)</a>   | Specifies parameters used by the instrument to analyze the input signal.  |
| <a href="#">Analysis Time (see page 371)</a>     | Specifies parameters that define the portion of the acquisition record that is used for analysis. Also allows you to specify the Units (Symbols or Seconds) and Time Zero Reference (Trigger or Acquisition Start) for P25 Analysis displays. |
| <a href="#">Test Patterns (see page 373)</a>     | Specifies the type of test pattern being used. Some analysis differs based on the selected test pattern. Available test patterns vary depending on the selected standard and modulation type.   |
| <a href="#">Trace (see page 374)</a>             | Specifies the trace type (Vectors, Points, and Lines), the number of points per symbol when chosen to view as Vectors or Lines, and to control the view as Freq Dev or IQ.  |
| <a href="#">Prefs (see page 377)</a>             | Specifies the radix of the marker readout and whether elements of the graphs are displayed.   |
| <a href="#">Trig Meas (see page 376)</a>         | Enables you to select from several different trigger measurements.  |

## P25 Eye Diagram Display


The P25 Eye Diagram display shows a digitally modulated signal overlapped on itself to reveal variations in the signal.

To show the P25 Eye Diagram display:


1. Press the **Displays** button or select **Setup > Displays**.
2. From the **Measurements** box, select **P25 Analysis**.
3. Double-click the **P25 Eye Diagram** icon in the **Available Displays** box. This adds the P25 Eye Diagram icon to the **Selected displays** box.
4. Click **OK** button. This displays the P25 Eye Diagram view.
5. Select **Setup > Settings** to display the control panel.
6. Select the **Modulation Params** tab. Set the Standard, Modulation Type, and Measurement and Reference filters as appropriate for the input signal.
7. For Phase 2 signals, select the **Trace** tab and set the **Trace Type** to Freq Dev or IQ.
8. Select the **Test Patterns** tab and choose a test pattern appropriate for the input signal.

### Elements of the Display

The following image shows the P25 Eye display with Trace Type set to IQ.



The following image shows the P25 Eye display with Trace Type set to Freq Dev.



| Item | Display element | Description  |
|------|-----------------|--|
| 1    | Top of graph    | The vertical scale is normalized with no units (for IQ) and with Hz (for Freq dev).  |
| 2    | Position        | Specifies the value shown at the center of the graph display.  |
| 3    | Bottom Readout  | Displays the value indicated by the bottom of graph.   |
| 4    | Autoscale       | Adjusts the vertical and horizontal settings to provide the best display.  |
| 5    | Position        | Displays the horizontal position of the trace on the graph display.  |
| 6    | Scale           | Adjusts the span of the graph in symbols.  |
| 7    | Freq Error      | Displays the difference between the maximum and minimum measured values of the signal frequency during the Measurement Time. The readout can either be Freq Error where the result is followed by Auto or it could be Freq Offset where the result is followed by Manual. This choice between Freq Error and Freq Offset is made from the Analysis Params tab in the Settings control panel. |

## P25 Eye Diagram Settings

### Front Panel: Settings

#### Application Toolbar:

The settings for the P25 Eye Diagram display are shown in the following table.

| Settings tab                                     | Description   |
|--|---|
| <a href="#">Modulation Params (see page 368)</a> | Specifies the input signal standard and additional user-settable signal parameters.   |
| <a href="#">Analysis Params (see page 371)</a>   | Specifies parameters used by the instrument to analyze the input signal.  |
| <a href="#">Analysis Time (see page 371)</a>     | Specifies parameters that define the portion of the acquisition record that is used for analysis. Also allows you to specify the Units (Symbols or Seconds) for P25 Analysis displays.        |
| <a href="#">Trig Meas (see page 376)</a>         | Enables you to select from various trigger measurements. The trigger selections vary by signal type.  |
| <a href="#">Test Patterns (see page 373)</a>     | Specifies the type of test pattern being used. Some analysis differs based on the selected test pattern. Available test patterns vary depending on the selected standard and modulation type. |
| <a href="#">Trace (see page 374)</a>             | Specifies the trace type (Vectors, Points, and Lines), the number of points per symbol when chosen to view as Vectors or Lines, and to control the view as Freq Dev or IQ.                    |
| <a href="#">Scale (see page 374)</a>             | Defines the vertical and horizontal axes.   |
| <a href="#">Prefs (see page 377)</a>             | Specifies the radix of the marker readout and whether elements of the graphs are displayed.   |


## P25 Power vs Time Display

The P25 Power vs Time display shows the signal power amplitude versus time.

To show the P25 Power vs Time display:

1. Press the **Displays** button or select **Setup > Displays**.
2. In the **Select Displays** dialog, select **P25 Analysis** in the **Measurements** box.
3. In the Available displays box, double-click the **P25 Power vs Time** icon or select the icon and click **Add**. The P25 Power vs Time icon will appear in the **Selected displays** box and will no longer appear under Available displays.
4. Click **OK** to show the P25 Power vs Time display.
5. Select **Setup > Settings** to display the control panel.
6. Select the **Modulation Params** tab. Set the Standard, Modulation Type, and Measurement and Reference filters as appropriate for the input signal.
7. Select the **Test Patterns** tab and choose a test pattern appropriate for the input signal.

## Elements of the Display



| Item | Display element                                | Description  |
|------|--|--|
| 1    | Top of graph, first setting                    | Sets the Power level that appears at the top of the graph, in dBm. This is only a visual control for panning the graph.  |
| 2    | Top of graph, second setting                   | Sets the vertical Scale of the graphs, in dB/div. This is only a visual control for panning the graph.   |
| 3    | View<br>(Only available for bursty HCPM data.) | Selects the specific view of the packet burst within the display:<br>Full Burst displays the entire packet, with vertical lines indicating Power ramp up, On Slot, Power ramp down, and Off slot regions.<br>Ramp Up zooms the display into the interval around the packet rising edge.<br>Ramp Down zooms the display into the interval around the packet falling edge. |
| 4    | Bottom of graph readout                        | Shows the Power level at the bottom of the graph in dBm.   |
| 5    | Autoscale                                      | Adjusts the vertical and horizontal settings to provide the best display.  |
| 6    | Bottom of graph, left side                     | Sets the starting time of the graph in seconds   |
| 7    | Bottom of graph, right side                    | Sets the scale (width) of the graph in seconds   |
| 8    | Table below graph                              | This table shows the following results.<br>For non-bursty signals: RF output power<br>For bursty signals: Power Info (Offslot Power and Power Envelope results) and Time Info.   |

## P25 Power vs Time Settings

### Front Panel: Settings

Application Toolbar: 

The settings for the P25 Power vs Time display are shown in the following table.

| Settings tab                                     | Description   |
|--|---|
| <a href="#">Modulation Params (see page 368)</a> | Specifies the input signal standard and additional user-settable signal parameters.   |
| <a href="#">Analysis Params (see page 371)</a>   | Specifies parameters used by the instrument to analyze the input signal.  |
| <a href="#">Analysis Time (see page 371)</a>     | Specifies parameters that define the portion of the acquisition record that is used for analysis. Also allows you to specify the Units (Symbols or Seconds) for P25 Analysis displays.        |
| <a href="#">Trig Meas (see page 376)</a>         | Enables you to select from various trigger measurements. The trigger selections vary by signal type.  |
| <a href="#">Test Patterns (see page 373)</a>     | Specifies the type of test pattern being used. Some analysis differs based on the selected test pattern. Available test patterns vary depending on the selected standard and modulation type. |
| <a href="#">Scale (see page 374)</a>             | Defines the vertical and horizontal axes.   |
| <a href="#">Prefs (see page 377)</a>             | Specifies the radix of the marker readout and whether elements of the graphs are displayed.   |

## P25 Summary Display

The P25 Summary display shows a summary of all the scalar measurements done on the acquired test pattern. The summary display and contents will vary according to the selected standard.

Pass/Fail information is also provided in this display for all enabled scalar measurements. You can set limits and choose which measurement to compare for Pass/Fail from the Limits tab in the P25 Summary Settings control panel. The default limits come from the performance recommendation limits given by the Standard document. The default limits can be reloaded by selecting the P25 Standards Preset option or by loading the default limits table.

To show the P25 Summary display you can select **Presets > Standards > P25** or do the following:

1. Press the **Displays** button or select **Setup > Displays**.
2. In the **Select Displays** dialog, select **P25 Analysis** in the **Measurements** box.
3. In the Available displays box, double-click the **P25 Summary** icon or select the icon and click **Add**. The P25 Summary icon will appear in the **Selected displays** box and will no longer appear under Available displays.
4. Click **OK** to show the P25 Summary display.
5. Select **Setup > Settings** to display the control panel.
6. Select the **Modulation Params** tab. Set the Standard, Modulation Type, and Measurement and Reference filters as appropriate for the input signal.
7. Select the **Test Patterns** tab and select a test pattern. The available test patterns in the drop down list depend of the standard and modulation type you have selected.

## P25 Summary Display

The Modulation Measurements, Power Measurements, and Trigger Measurements components of this display vary depending on which modulation type is selected.


The following image shows an example of the display for a Phase 1 C4FM signal.

| Standard: Phase1                         |           | Bandwidth: 12.5 KHz |  | Modulation Type: C4FM |  | Clear |  |
|--|-----------|---------------------|--|-----------------------|--|-------|--|
| <b>Modulation Measurements</b>           |           |                     |  |                       |  |       |  |
| Modulation Fidelity:                     | 0.465 %   | PASS                |  |                       |  |       |  |
| Operating Freq Accuracy:                 | 0 ppm     | PASS                |  |                       |  |       |  |
| Symbol Rate Accuracy:                    | -- ppm    | N/A                 |  |                       |  |       |  |
| <b>Frequency Deviation</b>               |           |                     |  |                       |  |       |  |
| Positive Peak:                           | -- Hz     | N/A                 |  |                       |  |       |  |
| Negative Peak:                           | -- Hz     | N/A                 |  |                       |  |       |  |
| <b>Power Measurements</b>                |           |                     |  |                       |  |       |  |
| RF Output Power:                         | 10.00 dBm | PASS                |  |                       |  |       |  |
| <b>Trigger Measurements</b>              |           |                     |  |                       |  |       |  |
| <b>Phase1 Tx Attack Time</b>             |           |                     |  |                       |  |       |  |
| Avg Tx Pwr Attack                        | ---       | N/A                 |  |                       |  |       |  |
| Avg Tx Encoder Attack                    | ---       | N/A                 |  |                       |  |       |  |
| --- of 10                                |           |                     |  |                       |  |       |  |
| <b>Phase1 Tx Attack Time (Busy/Idle)</b> |           |                     |  |                       |  |       |  |
| Avg Tx Pwr Attack                        | ---       |                     |  |                       |  |       |  |
| Avg Tx Encoder Attack                    | ---       |                     |  |                       |  |       |  |
| --- of 10                                |           |                     |  |                       |  |       |  |
| <b>Phase1 Tx Throughput Delay:</b> ---   |           |                     |  |                       |  |       |  |
| --- of 10                                |           |                     |  |                       |  |       |  |


The following image shows an example of the display for a Phase 1 C4FM High Deviation signal.

| Standard: Phase1                                 |             | Bandwidth: 12.5 KHz |  | Modulation Type: C4FM |  | Clear |  |
|--|-------------|---------------------|--|-----------------------|--|-------|--|
| <b>Modulation Measurements</b>                   |             |                     |  |                       |  |       |  |
| Modulation Fidelity:                             | 0.030 %     | PASS                |  |                       |  |       |  |
| Operating Freq Accuracy:                         | 0 ppm       | PASS                |  |                       |  |       |  |
| Symbol Rate Accuracy:                            | 0.00 ppm    | PASS                |  |                       |  |       |  |
| <b>Frequency Deviation (Symbol Rate Pattern)</b> |             |                     |  |                       |  |       |  |
| f1 Positive Peak:                                | 2768.93 Hz  | PASS                |  |                       |  |       |  |
| f2 Negative Peak:                                | -2865.77 Hz | PASS                |  |                       |  |       |  |
| <b>Power Measurements</b>                        |             |                     |  |                       |  |       |  |
| RF Output Power:                                 | 10.00 dBm   | PASS                |  |                       |  |       |  |
| <b>Trigger Measurements</b>                      |             |                     |  |                       |  |       |  |
| <b>Phase1 Tx Attack Time</b>                     |             |                     |  |                       |  |       |  |
| Avg Tx Pwr Attack                                | ---         |                     |  |                       |  |       |  |
| Avg Tx Encoder Attack                            | ---         |                     |  |                       |  |       |  |
| --- of 10  |             |                     |  |                       |  |       |  |
| <b>Phase1 Tx Attack Time (Busy/Idle)</b>         |             |                     |  |                       |  |       |  |
| Avg Tx Pwr Attack                                | ---         |                     |  |                       |  |       |  |
| Avg Tx Encoder Attack                            | ---         |                     |  |                       |  |       |  |
| --- of 10  |             |                     |  |                       |  |       |  |
| <b>Phase1 Tx Throughput Delay:</b> ---           |             |                     |  |                       |  |       |  |
| --- of 10  |             |                     |  |                       |  |       |  |

The following image shows an example of the display for a Phase 2 HCPM (Inbound) signal.



The following image shows an example of the display for a Phase 2 HDQPSK (Outbound) signal.



For more information about specific measurement results, see the P25 Measurements section [here \(see page 385\)](#).

## Elements of the Display

| Element                 | Description  |
|-------------------------|--|
| Standard                | Display of the standard selected on the Setup > Settings > Modulation Params tab.        |
| Bandwidth               | Display of the channel bandwidth which is set based on the standard and modulation type. |
| Modulation Type         | Display of the modulation type selected on Setup > Settings > Modulation Parameters tab. |
| Clear                   | Click button to reset measurement. Clears all values.                                    |
| Modulation Measurements | Shows the modulation measurements associated with the signal.                            |
| Power Measurements      | Shows the power measurements associated with the signal.                                 |
| Trigger Measurements    | Shows the trigger measurements associated with the signal.                               |

## P25 Summary Settings

### Front Panel: Settings

#### Application Toolbar:

The settings for the P25 Summary display are shown in the following table.

| Settings tab                                     | Description   |
|--|---|
| <a href="#">Modulation Params (see page 368)</a> | Specifies the input signal standard and additional user-settable signal parameters.   |
| <a href="#">Analysis Params (see page 371)</a>   | Specifies parameters used by the instrument to analyze the input signal.  |
| <a href="#">Analysis Time (see page 371)</a>     | Specifies parameters that define the portion of the acquisition record that is used for analysis. Also allows you to specify the Units (Symbols or Seconds) for P25 Analysis displays.        |
| <a href="#">Trig Meas (see page 376)</a>         | Enables you to select from various trigger measurements. The trigger selections vary by signal type.  |
| <a href="#">Test Patterns (see page 373)</a>     | Specifies the type of test pattern being used. Some analysis differs based on the selected test pattern. Available test patterns vary depending on the selected standard and modulation type. |
| <a href="#">Limits (see page 378)</a>            | Load and define P25 measurement limits for Pass/Fail comparison. You can save defined limits as a .csv file and also load previously saved .csv files.  |

## P25 Symbol Table Display

The P25 Symbol Table display shows decoded data values for each data symbol in the analyzed signal packet.

To show the P25 Symbol Table display:

1. Press the **Displays** button or select **Setup > Displays**.
2. In the **Select Displays** dialog, select **P25 Analysis** in the **Measurements** box.

3. In the Available displays box, double-click the **P25 Symbol Table** icon or select the icon and click **Add**. The P25 Symbol icon will appear in the **Selected displays** box and will no longer appear under Available displays.
4. Click **OK** to show the P25 Symbol Table display.
5. Select **Setup > Settings** to display the control panel.
6. Select the **Modulation Params** tab. Set the Standard, Modulation Type, and Measurement and Reference filters as appropriate for the input signal.
7. Select the **Test Patterns** tab and select a test pattern. The available test patterns in the drop down list depend on the standard and modulation type you have selected.

## P25 Symbol Table

The following image shows the P25 Symbol Table for HCPM (bursty) signals. For this signal type, the symbols are arranged as Bursts vs Symbols. The analysis is done only on the on slot regions of the bursty HCPM data and 160 symbols (centered at the middle of the burst) are reported on the symbol table for every on slot region. You can read more about specific measurements [here \(see page 385\)](#).

| P25:HCPM modulation analysis is done only on Onslot regions |    |    |
|---|----|----|
|   | B1 | B2 |
| 0   | 1  | 3  |
| 1   | 3  | 1  |
| 2   | 2  | 1  |
| 3   | 1  | 0  |
| 4   | 2  | 3  |
| 5   | 3  | 3  |
| 6   | 0  | 0  |
| 7   | 3  | 1  |
| 8   | 2  | 3  |
| 9   | 1  | 2  |
| 10  | 0  | 3  |
| 11  | 2  | 2  |
| 12  | 3  | 0  |
| 13  | 1  | 3  |
| 14  | 2  | 2  |
| 15  | 2  | 2  |
| 16  | 0  | 1  |
| 17  | 0  | 3  |
| 18  | 0  | 2  |
| 19  | 3  | 3  |
| 20  | 2  | 3  |
| 21  | 1  | 0  |
| 22  | 3  | 0  |
| 23  | 1  | 0  |

Marker: MR

Symbol: 18

Time: 69.411 ms

Value: 0

The following image shows the P25 Symbol Table for C4FM (non-bursty) signals. For this signal type and for HDQPSK, there is no grouping into bursts and all symbols that are analyzed are shown.

|     |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |
|-----|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|
| 0   | 3 | 2 | 0 | 0 | 0 | 2 | 1 | 1 | 3 | 2 | 2 | 3 | 0 | 2 | 2 | 1 |
| 16  | 1 | 3 | 3 | 1 | 1 | 3 | 0 | 2 | 2 | 0 | 0 | 3 | 0 | 0 | 0 | 3 |
| 32  | 3 | 2 | 0 | 2 | 0 | 0 | 0 | 3 | 1 | 2 | 0 | 3 | 1 | 0 | 3 | 2 |
| 48  | 3 | 2 | 3 | 3 | 3 | 1 | 3 | 1 | 3 | 1 | 0 | 0 | 1 | 1 | 2 | 3 |
| 64  | 2 | 1 | 2 | 3 | 3 | 3 | 3 | 0 | 2 | 1 | 0 | 2 | 3 | 3 | 3 | 1 |
| 80  | 0 | 1 | 3 | 3 | 0 | 1 | 3 | 1 | 0 | 1 | 2 | 3 | 1 | 1 | 2 | 2 |
| 96  | 0 | 2 | 0 | 0 | 3 | 0 | 1 | 3 | 2 | 3 | 0 | 0 | 2 | 2 | 0 | 3 |
| 112 | 3 | 2 | 3 | 3 | 1 | 0 | 3 | 3 | 2 | 0 | 3 | 0 | 2 | 0 | 0 | 2 |
| 128 | 1 | 2 | 2 | 1 | 2 | 2 | 2 | 0 | 3 | 1 | 1 | 2 | 2 | 2 | 1 | 2 |
| 144 | 3 | 1 | 3 | 1 | 3 | 2 | 2 | 0 | 2 | 1 | 1 | 1 | 2 | 0 | 0 | 3 |
| 160 | 3 | 2 | 2 | 0 | 2 | 3 | 1 | 1 | 3 | 0 | 2 | 2 | 0 | 2 | 3 | 0 |
| 176 | 2 | 1 | 1 | 0 | 0 | 0 | 3 | 2 | 1 | 1 | 3 | 1 | 3 | 3 | 3 | 2 |
| 192 | 3 | 1 | 0 | 3 | 1 | 2 | 2 | 2 | 1 | 3 | 3 | 0 | 3 | 0 | 3 | 0 |
| 208 | 2 | 3 | 3 | 2 | 1 | 3 | 0 | 3 | 1 | 2 | 2 | 2 | 0 | 0 | 0 | 1 |
| 224 | 0 | 2 | 1 | 0 | 1 | 0 | 1 | 3 | 2 | 3 | 2 | 2 | 3 | 1 | 2 | 0 |
| 240 | 3 | 2 | 2 | 0 | 2 | 3 | 2 | 1 | 0 | 1 | 3 | 3 | 2 | 1 | 2 | 0 |
| 256 | 1 | 2 | 3 | 2 | 0 | 0 | 2 | 0 | 0 | 0 | 1 | 0 | 1 | 1 | 0 | 3 |
| 272 | 3 | 2 | 0 | 2 | 1 | 3 | 3 | 0 | 0 | 3 | 2 | 0 | 3 | 0 | 3 | 3 |
| 288 | 3 | 2 | 3 | 0 | 2 | 1 | 1 | 1 | 2 | 0 | 1 | 1 | 1 | 2 | 1 | 3 |
| 304 | 1 | 2 | 3 | 0 | 1 | 0 | 1 | 3 | 1 | 2 | 2 | 1 | 1 | 3 | 3 | 2 |
| 320 | 1 | 1 | 3 | 2 | 3 | 2 | 0 | 0 | 1 | 0 | 0 | 0 | 3 | 3 | 2 | 1 |

Marker: MR      Symbol: 152  
Time: 39.578 ms      Value: 2

## Elements of the Display

| Element | Description   |
|---------|---|
| Marker  | Displays the selected marker label.   |
| Time    | Displays the time in ms or in Symbols based on the Units chosen in the Analysis Time tab of the Settings Control panel. |
| Symbol  | The value shown here reflects the symbol you have selected (highlighted) in the display.                                |
| Value   | Displays the value of the selected symbol.  |

## P25 Symbol Table Settings

### Front Panel: Settings

#### Application Toolbar:

The settings for the P25 Symbol Table display are shown in the following table.

| Settings tab                                     | Description   |
|--|---|
| <a href="#">Modulation Params (see page 368)</a> | Specifies the input signal standard and additional user-settable signal parameters.   |
| <a href="#">Analysis Params (see page 371)</a>   | Specifies parameters used by the instrument to analyze the input signal.  |
| <a href="#">Analysis Time (see page 371)</a>     | Specifies parameters that define the portion of the acquisition record that is used for analysis. Also allows you to specify the Units (Symbols or Seconds) for P25 Analysis displays.        |
| <a href="#">Trig Meas (see page 376)</a>         | Enables you to select from various trigger measurements. The trigger selections vary by signal type.  |
| <a href="#">Test Patterns (see page 373)</a>     | Specifies the type of test pattern being used. Some analysis differs based on the selected test pattern. Available test patterns vary depending on the selected standard and modulation type. |
| <a href="#">Prefs (see page 377)</a>             | Specifies the radix of the marker readout.  |

## P25 Frequency Dev vs Time Display


The P25 Frequency Deviation vs. Time Display shows how the signal frequency varies with time.

To display the P25 Frequency Dev vs. Time Display:


1. Select the **Displays** button or **Setup > Displays**.
2. In the **Select Displays** dialog, select **P25 Analysis** in the **Measurements** box.
3. In the **Available displays** box, double-click the **P25 Frequency Dev vs. Time** icon or select the icon and click **Add**. The P25 Frequency Dev vs. Time icon will appear in the **Selected displays** box and will no longer appear under Available displays.
4. Click **OK** to display the Freq Dev vs. Time display.
5. Select the **Modulation Params** tab. Set the Standard, Modulation Type, and Measurement and Reference filters as appropriate for the input signal.
6. Select the **Test Patterns** tab and select a test pattern. The available test patterns in the drop down list depend of the standard and modulation type you have selected.

### Elements of the Display

The following image shows the P25 Freq Dev vs Time display for Phase 1 C4FM (non-bursty) High Deviation signals.



The following image shows the P25 Freq Dev vs Time display for Phase 2 HCPM (bursty) signals.



---

**NOTE.** For bursty HCPM signals, frequency deviation analysis is done only on 160 symbols centered at every on slot region in the chosen analysis window and not on off slot regions. That is why there is no information shown during off slot regions. You can read more specific information about P25 measurements [here \(see page 340\)](#).

---

### Elements of the Display

| Item | Display element         | Description  |
|------|-------------------------|--|
| 1    | Top of graph adjustment | Adjust the frequency range displayed on the vertical axis.                                 |
| 2    | Position (center)       | Adjust the frequency shown at the center of the display.                                   |
| 3    | Autoscale button        | Adjusts the offset and range for both vertical and horizontal to provide the best display. |
| 4    | Position                | Displays the horizontal position of the trace on the graph display.                        |
| 5    | Scale                   | Adjust the horizontal scale (time).  |

[Changing Frequency vs Time Display Settings \(see page 366\)](#)

## P25 Frequency Dev Vs Time Settings

**Menu Bar:** Setup > Settings

**Front Panel:** Settings

**Application Toolbar:** 

The Setup settings for P25 Frequency Dev vs. Time are shown in the following table.

| Settings tab                                     | Description   |
|--|---|
| <a href="#">Modulation Params (see page 368)</a> | Specifies the input signal standard and additional user-settable signal parameters.   |
| <a href="#">Analysis Params (see page 371)</a>   | Specifies parameters used by the instrument to analyze the input signal.  |
| <a href="#">Analysis Time (see page 371)</a>     | Specifies parameters that define the portion of the acquisition record that is used for analysis. Also allows you to specify the Units (Symbols or Seconds) for P25 Analysis displays.        |
| <a href="#">Trig Meas (see page 376)</a>         | Enables you to select from various trigger measurements. The trigger selections vary by signal type.  |
| <a href="#">Test Patterns (see page 373)</a>     | Specifies the type of test pattern being used. Some analysis differs based on the selected test pattern. Available test patterns vary depending on the selected standard and modulation type. |
| <a href="#">Trace (see page 374)</a>             | Specifies the trace type (Vectors, Points, and Lines), the number of points per symbol when chosen to view as Vectors or Lines, and to control the view as Freq Dev or IQ.                    |
| <a href="#">Scale (see page 374)</a>             | Defines the vertical and horizontal axes.   |
| <a href="#">Prefs (see page 377)</a>             | Specifies the radix of the marker readout and whether elements of the graphs are displayed.   |

## P25 Analysis Shared Measurement Settings

### Front Panel: Settings

#### Application Toolbar:

The control panel tabs in this section are shared between the displays in P25 Analysis (Setup > Displays). Some tabs are shared by all the displays, some tabs are shared by only a subset of displays. The settings available on some tabs change depending on the selected display.

### Common controls for P25 analysis displays

| Settings tab                                     | Description  |
|--|--|
| <a href="#">Modulation Params (see page 368)</a> | Specifies the input signal standard and additional user-settable signal parameters.  |
| <a href="#">Analysis Params (see page 371)</a>   | Specifies parameters used by the application to analyze the input signal.  |
| <a href="#">Analysis Time (see page 371)</a>     | Specifies parameters that define the portion of the acquisition record that is used for analysis. Also allows you to specify the Units (Symbols or Seconds) for P25 Analysis displays. |
| <a href="#">Trace (see page 374)</a>             | Specifies the trace type (Vectors, Points, and Lines), the number of points per symbol when chosen to view as Vectors or Lines, and to control the view as Freq Dev or IQ.             |
| <a href="#">Scale (see page 374)</a>             | Defines the vertical and horizontal axes.  |
| <a href="#">Test Patterns (see page 373)</a>     | Enables you to select from eight different test patterns.  |
| <a href="#">Trig Meas (see page 376)</a>         | Enables you to select from various trigger measurements. The trigger selections vary by signal type.   |
| <a href="#">Prefs (see page 377)</a>             | Specifies the radix of the marker readout and whether elements of the graphs are displayed.  |

## Modulation Params Tab - P25

The Modulation Params tab specifies the type of modulation used by the input signal and other parameters that define the signal format.

The screenshot shows the 'Modulation Params' tab selected in a software interface. The tab bar at the top includes 'Modulation Params', 'Analysis Params', 'Analysis Time', 'Trig Meas', 'Test Patterns', and 'Prefs'. The main area contains several controls:

- Standard:** A dropdown menu set to 'Phase2'.
- Measurement Filter:** A dropdown menu set to 'None'.
- Symbol Rate:** A text field displaying '6 kHz'.
- Channel Bandwidth:** A text field displaying '12.5 KHz'.
- Reference Filter:** A dropdown menu set to 'None'.
- Modulation Type:** A dropdown menu set to 'HCPM (Inbound)'.

| Settings           | Description  |
|--------------------|--|
| Standard           | Specifies the standard used for the input signal: Phase 1, Phase 2.  |
| Channel Bandwidth  | This readout shows the nominal channel bandwidth based on the standard.  |
| Modulation Type    | Specifies the modulation type of the input signal. Choices vary depending on the selected standard. Modulation types for Phase 2 are HCPM (Inbound) and HDQPSK (Outbound). Phase 1 has only C4FM as the modulation type selection. |
| Measurement Filter | Specifies the filter used as a measurement.  |
| Reference Filter   | Specifies the filter used as a reference.  |
| Filter Parameter   | Enter a value used for defining the Reference Filter. (Not present for some filter types)  |
| Symbol Rate        | This is a readout that shows the symbol rate for demodulating digitally modulated signals based on the standard. This rate is always 4.8 kHz for Phase 1 signals and 6 kHz for Phase 2 signals.                                    |

### Symbol Rate

Specifies the symbol rate for demodulating digitally modulated signals based on the standard. The symbol rate and the bit rate are related as follows:

$$(\text{Symbol rate}) = (\text{Bit rate}) / (\text{Number of bits per symbol})$$

The bit rate used for Phase 1 (C4FM) is 9600 bps. For Phase 2 (HCPM and HDQPSK) it is 12000 bps. There are two bits per symbol for all above mentioned modulation types. Therefore, the symbol rate is 4800 Hz for Phase 1 and 6000 Hz for Phase 2.

### Measurement and Reference Filters

The available measurement and reference filters depend on the selected modulation type. The following table shows the recommended filters for the specified modulation types.



**CAUTION.** Although there are other filter types listed in the drop down menu, if you select any filter other than that which is recommended, the measurement results may not be accurate.

| Modulation type | Measurement filters | Reference filters                   |
|-----------------|---------------------|-------------------------------------|
| HDQPSK          | HDQPSK-P25          | None                                |
| HPCM            | None                | None                                |
| C4FM            | C4FM-P25            | RaisedCosine (Filter parameter 0.2) |

The measurement filter is applied before the demodulation bit is detected and the ideal reference is calculated.

The reference filter is applied to the internally generated ideal reference signal before the modulation fidelity is calculated.

## How to Select Filters

In a signal transmitter/receiver system, the baseband signal might be filtered for bandwidth limiting or for another kind of necessary shaping that needs to be applied. Normally, a filter in the transmitter ( $F_t$ ) and a filter in the receiver ( $F_r$ ) are applied.

The Measurement Filter setting in the analyzer corresponds to the baseband filter in the receiver ( $F_r$ ): This setting tells the analyzer what filter your receiver uses. When the analyzer is set to the same filter used by the receiver, the analyzer sees the signal as your receiver would. The Measurement Filter setting should be the same as the filter used in the receiver under normal operation (as opposed to testing).

The Reference Filter setting in the analyzer corresponds to the baseband filter in the transmitter-receiver combination ( $F_r * F_t$ ). The baseband filter for the transmitter-receiver combination is often referred to as the *System Filter*. This filter is called the reference filter because it is used to recreate a reference signal that is compared to the received signal. This recreated reference signal is the *ideal signal* with  $F_r * F_t$  applied; differences between this *ideal signal* and the received signal enables the determination of signal quality, such as modulation fidelity measurements.

## Filter Parameter (C4FM only)

The filter parameter specifies the alpha for the Raised Cosine filter when selected as the Reference filter. Some filter types have a fixed parameter value that is specified by industry standard, while other filter types by definition have no filter parameter. For filter types with no filter parameter, there is no filter parameter control present in the control panel. The recommended Reference filter for C4FM is Raised Cosine and the corresponding filter parameter for C4FM is 0.2.

## Analysis Params Tab - P25

The Analysis Params tab contains parameters that control the analysis of the input signal.

Modulation Params Analysis Params Analysis Time Test Patterns Prefs Trig Meas

Frequency Error:  ☒ Auto PTX:

Measurement BW:

| Settings        | Description  |
|-----------------|--|
| Frequency Error | When the Auto box is checked, the analysis determines the Frequency Error and the measured Frequency Error is displayed. When the Auto box is unchecked, the entered value is used by the analysis as a fixed frequency offset. This is useful when the exact frequency offset of the signal is known. |
| Measurement BW  | Specifies the bandwidth about the center frequency at which measurements are made. Select Manual, Auto, or Link to Span.   |
| PTX             | Specifies the RF Output Power as recommended for transmitters. This is only used in the computation of off slot power in HCPM modulated signals.   |

## Analysis Time Tab - P25

The Analysis Time tab contains parameters that define how the signal is analyzed in the P25 Analysis displays.

Modulation Params Analysis Params Analysis Time Test Patterns Prefs Trig Meas

Analysis Offset:  ☒ Auto Time Zero Reference:

Analysis Length:  ☐ Auto Units:

Available: 31.816 ms

| Settings            | Description  |
|---------------------|--|
| Analysis Offset     | Specifies the location of the first time sample to use in measurements.  |
| Auto                | When enabled, causes the instrument to set the Analysis Offset value based on the requirements of the selected display.  |
| Analysis Length     | Specifies the length of the analysis period to use in measurements. Length is specified in either symbols or seconds, depending on the Units setting.                                      |
| Auto                | When enabled, causes the instrument to set the Analysis Length value based on the requirements of the selected display.  |
| Actual              | This is a displayed value, not a setting. It is the Analysis Length (time or symbols) being used by the analyzer, this value may not match the Analysis Length requested (in manual mode). |
| Time Zero Reference | Specifies the zero point for the analysis time.  |
| Units               | Specifies the units of the Analysis Length to either Symbols or Seconds.   |

### Analysis Offset

Use analysis offset to specify where measurements begin. Be aware that you cannot set the Analysis Offset outside the range of time covered by the current acquisition data. (all time values are relative to the Time Zero Reference).

You can set the Analysis Length so that the requested analysis period falls partly or entirely outside the current range of acquisition data settings. When the next acquisition is taken, its Acquisition Length will be increased to cover the new Analysis Length, as long as the Sampling controls are set to Auto. If the Sampling parameters are set to manual, or if the instrument is analyzing saved data, the actual analysis length will be constrained by the available data length, but in most cases, measurements are able to be made anyway. The instrument will display a notification when measurement results are computed from less data than requested. Range: 0 to [(end of acquisition) - Analysis Length)]. Resolution: 1 effective sample (or symbol).

### Analysis Length

Use the analysis length to specify how long a period of time is analyzed. As you adjust this value, the actual amount of time for Analysis Length, in Symbol or Seconds units, is shown below the control in the "Actual" readout. This setting is not available when Auto is checked. Range: minimum value depends on modulation type. Resolution: 1 symbol.


### Time Zero Reference

All time values are measured from this point (such as marker position or horizontal position (in Y vs Time displays). Choices are: Acquisition Start or Trigger. When a trigger measurement is chosen from the Trig Meas tab of the Settings control panel, Time Zero Reference is forced to Trigger.

| Parameter         | Description  |
|-------------------|--|
| Acquisition Start | Time zero starts from the point at which acquisition begins. |
| Trigger           | Time zero starts from the trigger point.                     |

## Test Patterns Tab - P25

Test patterns allow the software to compare the measurement result to the standards limit. The list of available test patterns varies depending on which standard and modulation type is selected.




| Settings      | Description   |
|---------------|---|
| Test Patterns | Use this drop-down list to select an appropriate test pattern. The list varies depending on which standard and modulation type is selected. |

You can read more about test patterns [here \(see page 348\)](#).

## Trace Tab - P25

The Trace tab allows you to set the trace display characteristics of the P25 trace display. The selections vary depending on the selected display.

The following image shows the tab for the Constellation and Eye Diagram displays. For the other trace displays, the Trace Type setting is not available.



| Setting  | Description   |
|--|---|
| Trace<br>(P25 Constellation and P25 Eye Diagram displays only)       | Select the trace to display.<br>P25 Constellation display only: Select the trace that is hidden or displayed based on whether or not Show is selected.  |
| Show<br>(P25 Constellation and P25 Eye Diagram displays only)        | Specifies whether the trace selected by Trace is displayed or hidden.   |
| Freeze<br>(P25 Constellation display only)                           | Halts updates to the trace selected by the Trace setting. Present for the Constellation display only.   |
| Points/Symbol  | Select how many points to use between symbols when connecting the dots. Values: 1, 2, 4, 8.   |
| Content<br>(P25 Constellation and P25 Freq Dev & Time displays only) | Select whether to display the trace as vectors (points connected by lines), points (symbols only without lines), or lines (lines drawn between symbols, but no symbols are displayed). The choices available depend on the display. |
| Trace Type<br>(P25 Constellation and P25 Eye Diagram displays only)  | Select to specify whether the plots in the Constellation and Eye Diagram displays are shown as I vs Q or as Frequency Deviation.  |

## Scale Tab - P25

The Scale tab allows you to change the scale settings that control how the trace appears on the display but does not change control settings such as Measurement Frequency. There are three versions of the Scale tab for P25 displays.

Scale tab for the P25 Eye Diagram display

Scale tab for the P25 Power vs Time display

Scale tab for the P25 Freq Dev vs Time display


| Settings              | Description   |
|-----------------------|---|
| Vertical              | Controls the vertical position and scale of the trace display.  |
| Scale                 | Changes the vertical scale of the graph.  |
| Position              | Adjusts the reference level away from top of the graph.   |
| Autoscale             | Resets the scale of the vertical axis to contain the complete trace.  |
| Horizontal            | Controls the span of the trace display and position of the trace.   |
| Horizontal Full Burst |   |
| Scale                 | Allows you to, in effect, change the span.  |
| Position              | Allows you to pan a zoomed trace without changing the Measurement Frequency.  |
| Autoscale             | Resets the scale of the horizontal axis to contain the complete trace.  |
| Auto                  | When Auto is checked, the scale and position values for the Symbols graph are automatically adjusted to maintain the optimal display. |
| Reset                 | Resets the vertical and horizontal settings.  |

**NOTE.** The Units used for the horizontal scale can be either Seconds or Symbols. To set the units for the horizontal scale, display the Analysis Time tab. On the tab, select the appropriate units from the Units drop-down list.


## Trig MeasTab - P25

The Trig Meas tab enables you to chose a trigger measurement. The available measurements in the drop-down menu depend on the standard and modulation type selected in the Modulation Params tab. The Trig Meas tab is not available for Phase 2 HDQPSK (Outbound) signals. You can read more about P25 trigger related measurements [here \(see page 343\)](#).

The following image shows the tab for Phase 1 (C4FM) signals.




The following image shows the tab for Phase 1 (HCPM Inbound) signals.



| Settings             | Description                          |
|----------------------|--------------------------------------|
| Trigger Measurements | Select the trigger measurement test. |

## Prefs Tab - P25

The Prefs tab enables you to change appearance characteristics of the P25 Analysis displays. Not all settings on the Prefs tab shown below appear for every P25 display. The Summary display does not have a Prefs tab.



Modulation Params Analysis Params Analysis Time Test Patterns Trace Scale Prefs Trig Meas


Radix: Binary

☒ Show graticule ☒ Show Marker readout in graph (selected marker)

| Setting  | Description  |
|--|--|
| Show graticule                                 | Shows or hides the graticule.  |
| Show Marker readout in graph (selected marker) | Shows or hides the readout for the selected marker in the graph area.  |
| Radix  | Specifies how symbols are displayed in the Symbol Table display and in the Marker readout in the display. There are three choices for Radix: Binary (00,01,10,11), Quaternary (0,1,2,3), and Modulation Symbols (+1,+3,-1,-3). |

## Limits Tab - P25

The Limits tab is only available for the P25 Summary display. It enables you to load an existing limits table, save a limits table, or edit limits values.



| Setting | Description  |
|---------|--|
| Load    | Click to load a saved Limits table from a .csv file.   |
| Save    | Click to save the current Limits table to a .csv file. |

### Edit Limits

To directly edit measurement limits in the table, click on the value in the Limit column that you want to change.

The following table describes the parameters that are set in the Limits Table.

#### Limits Table Settings

| Setting          | Description   |
|------------------|---|
| On               | Click on the cell in the On column next to the measurement to specify whether or not measurements are selected for limit comparison to indicate Pass or Fail.. A check mark means the measurement will be taken. An empty box means it will not be taken. |
| Measurement Name | Specifies the name of the measurement related to the limit. (Not editable.)   |
| Limit            | Specifies the value of the limit to the related measurement. When the cell is selected, the value is shown along with the units.  |

# LTE Analysis

## Overview

The Long Term Evolution (LTE) Downlink RF Measurements Analysis option allows you to evaluate RF signals to ensure that they meet 3GPP measurements. These are described in the TS 36.104 Base Station (BS) radio transmission and reception and test specifications TS36.141 Base Station (BS) conformance testing documents version 12.5. This analysis option supports both LTE TDD and LTE FDD frame structures. This analysis option supports the following measurements.

- Channel Power
- Occupied Bandwidth
- Adjacent Channel Leakage Ratio (ACLR)
- Spectral Emission Mask (Operating Band Unwanted Emission)
- Cell ID
- For TDD LTE Transmitter Off Power

These measurements are also compared with the limits provided by the standard to give pass/fail results (except for Channel Power and OBW).

More detailed information about these measurements is available in the [LTE measurements](#) section and the supported measurements [table](#).

You can also select these measurements from four LTE preset test setups. The test setups load pre-configured displays and control setting as suggested by the standard to accelerate the test setup of the analyzer. The following four test setups are available for this analysis option.

- Cell ID
- ACLR
- Channel Power and TDD Toff Power
- SEM

More detailed information about these test setups is available [here](#).

With the LTE downlink RF Measurement Analysis test suite, test engineers can simplify the execution of a number of transmitter tests while still enabled to modify signal parameters for in-depth signal analysis. The analysis results give multiple views of LTE signal characteristics to allow the diagnosis of signal imperfections and impairments quickly and easily. Display controls allow you to selectively display the analysis results to help locate trouble spots in the signal.

## LTE topics in this Help

The following information about the LTE Analysis option is available:

- [Reference table of supported LTE measurements](#)
- [LTE measurements](#)
- [LTE Standards preset test setups](#)
- [LTE displays](#)
- [LTE measurement control settings](#)

## Supported LTE measurements

The following table gives a brief description of the available LTE measurements. More detailed information can be found here. [LTE measurements](#)

| LTE measurement                       | LTE standard(s) | Description   |
|---------------------------------------|-----------------|---|
| Cell ID detection                     | TDD<br>FDD      | The Cell ID is detected from the input LTE signal.  |
| Adjacent Channel Leakage Ratio (ACLR) | TDD<br>FDD      | The Adjacent Channel integrated power is calculated and shown. The relative power compared to the reference signal is also computed. The computed power is compared against limits suggested by the selected standard and pass/fail is reported. The appropriate settings for this measurement are loaded with the ACLR test setup (Presets>Standards>LTE). |
| Channel Power                         | TDD<br>FDD      | The channel power is calculated in the channel bandwidth.   |
| Occupied Bandwidth                    | TDD<br>FDD      | The Occupied bandwidth is calculated as bandwidth containing 99% of the total integrated power in the selected span around the selected center frequency.   |
| Operating Band Unwanted Emission      | TDD<br>FDD      | The power in the offset regions is calculated and presented and compared against limits set in the offset and limits table and pass/fail is reported. The appropriate settings for this measurement are loaded with the SEM test setup (Presets>Standards>LTE).   |
| T <sub>OFF</sub>                      | TDD             | The power in off-slot region is computed and compared against selected limits.  |


## LTE Standards preset test setups

### Presets > Standards


The LTE Standards preset allows you to access displays preconfigured for the test setup you select. The test setups load the displays and control setting options suggested by the LTE standard to perform the measurements. You can read more about how Presets work [here](#).

There are four test setups for LTE:


#### ■ Cell ID




#### ■ ACLR



## ■ Channel Power and TDD Toff Power



## ■ SEM



The following table shows what automatically loads with each test setup.

**Table 8: LTE preset test setups**

| Test setup                       |   | Displays loaded with preset  |
|----------------------------------|---|--|
| Cell ID                          | Frame structure:<br>- Downlink FDD<br>- Downlink TDD<br>(Default is Downlink FDD.)  | Spectrum, Time Overview, LTE Constellation                                   |
|                                  | Channel bandwidth: 1.4, 3, 5, 10, 15, 20 MHz<br>(Default is 1.4 MHz.)   |  |
| ACLR                             | Frame structure:<br>- Downlink FDD<br>- Downlink TDD<br>(Default is Downlink FDD.)  | LTE ACLR   |
|                                  | Channel bandwidth: 1.4, 3, 5, 10, 15, 20 MHz<br>(Default is 1.4 MHz.)   |  |
|                                  | Base station category:<br>- Wide Area Cat A<br>- Wide Area Cat B — Option 1<br>- Home Area (P-rated ≤ 20 dBm)<br>- Local Area (P-rated ≤ 24 dBm)<br>- Medium Area (P-rated ≤ 38 dBm)<br>(Default is Wide Area Cat A.) |  |
|                                  | Adjacent channel type:<br>- UTRA<br>- E-UTRA<br>(Default is UTRA.)  |  |
|                                  | UTRA chip rate.<br>1.28 Mcps<br>3.84 Mcps<br>7.68 Mcps  |  |
|                                  | Chip rate is only displayed under the following conditions:<br>TDD is frame structure<br>Adjacent channel type selected is UTRA<br>Channel bandwidth is >3 MHz<br>(Default chip rate is 3.84 Mpcs.)                   |  |
| Channel Power and TDD Toff Power | Frame structure:<br>- Downlink FDD<br>- Downlink TDD<br>(Default is Downlink FDD.)  | Time Overview (TDD only), LTE Channel Spectrum, LTE Power vs Time (TDD only) |
|                                  | Channel bandwidth: 1.4, 3, 5, 10, 15, 20 MHz<br>(Default is 1.4 MHz.)   |  |

Table 8: LTE preset test setups (cont.)


| Test setup  | Displays loaded with preset |
|---|-----------------------------|
| SEM<br>Frame structure:<br>- Downlink FDD<br>- Downlink TDD<br>(Default is Downlink FDD.)<br>Channel bandwidth: 1.4, 3, 5, 10, 15, 20 MHz<br>(Default is 1.4 MHz.)<br>Base station category:<br>- Wide Area Cat A<br>- Wide Area Cat B — Option 1<br>- Home Area ( $2 \text{ dBm} \leq P \leq 20 \text{ dBm}$ )<br>- Local Area<br>- Medium Area ( $P \leq 31 \text{ dBm}$ )<br>(Default is Wide Area Cat A.) | SEM                         |

### Retain current center frequency and reference level settings

The **Retain current Center Frequency setting** appears when the LTE Standards Preset is chosen. This setting allows you to retain the previously used center frequency. By default, the Center Frequency setting box is unchecked and the LTE preset displays will load with 1.96 GHz for FDD and 1.9 GHz for TDD.

The **Retain current Ref Level setting** appears when the LTE Standards Preset is chosen. This setting allows you to retain the previously used reference level. By default, the Ref Level setting box is checked.


To activate these settings, check the box next to the desired setting. The following image shows the **Retain current Ref Level setting** box checked.



## LTE displays

The displays in LTE Analysis (Setup > Displays > Measurements: LTE Analysis) are:

- [LTE Channel Spectrum](#)
- [LTE ACLR](#)
- [LTE Constellation](#)
- [LTE Power vs Time](#)
- [SEM](#) (This display is not an LTE specific display, but is available when RF Measurements is selected from the Select Displays dialog (Setup > Displays) (see the following image).)



## LTE measurements

LTE Analysis enables RF measurements and detection of Cell ID in the transmitted signal for both TDD and FDD frame structure LTE signals. For TDD signals, the analysis will also do the  $T_{OFF}$  measurement. The following topics contain important information you should know about specific LTE measurements. You can view a table with all of the available measurements [here](#).

### Channel Power and Occupied Bandwidth (TDD and FDD)

The Channel Power measurement is done by calculating the power in the LTE signal based on the selected channel bandwidth option. The Occupied Bandwidth measurement calculates the frequency range over which 99% of the power is contained in the acquired signal. This measurement is done for the selected channel bandwidth option.

More information about this measurement:

- Select the Channel Power and TDD Toff Power test setup from Presets > Standards > LTE to perform the measurement. This will load the displays found in the [LTE preset test setups table](#).
- This test setup allows you to choose the frame structure and the channel bandwidth.
- The Channel Spectrum display will contain the results of Channel Power and Occupied Bandwidth. The result gives the power in the LTE signal (calculated over the channel bandwidth).
- Results are presented as scalar and are located below the display.
- For TDD signals, the channel power and the occupied bandwidth measurement are done on the on-slot region. If the valid on-slot region is not found, the measurement is done in the selected analysis length.

### ACLR (FDD and TDD)

The Adjacent Channel specification and limits for comparison are set based on selected frame structure, adjacent channel and base station types from Standards Presets. It calculates the integrated power in the different adjacent regions and also indicates pass/fail based on comparison with absolute and relative limits in LTE ACLR display.

More information about this measurement:

- Select the ACLR test setup from Presets > Standards > LTE to perform the measurement. This will load the displays found in the [LTE preset test setups table](#).
- This test setup allows you to select the frame structure, channel bandwidth, base station type, and adjacent channel type with the following restrictions:
  - Frame Structure can be TDD or FDD
  - Channel Bandwidth can be 1.4, 3, 5, 10, 15, or 20 MHz
  - Base station type can be Wide Area Cat A, Wide Area Cat B- Option 1, Home Area (P-rated  $\leq 20$  dBm), Local Area (P-rated  $\leq 24$  dBm), Medium Area (P-rated  $\leq 38$  dBm)
  - Adjacent Channel type can E-UTRA or UTRA
  - UTRA chip rate can be 1.28, 3.84, or 7.68 Mcps

Default rate is 3.84 Mcps for all UTRA adjacent channels of FDD and when channel bandwidth is 1.4 or 3 MHz.

However, you can choose other rates when the frame structure is TDD and when the bandwidth is more than 3 MHz. Otherwise, the standard recommends use of 3.84 Mcps.

- Based on the settings you select from Standards Preset, the offset and bandwidth of the adjacent channels are set in the Offset and Limits tab of the control panel. The settings are different for paired and unpaired spectrum.
- The settings also include the absolute power limit for comparison based on the base station type. The standard recommends the settings in dBm/MHz for each base station type. The power limits are appropriately scaled with the integration bandwidth and presented in dBm in the Offset and Limits table of the display. The integrated power in each adjacent band is compared with the absolute and relative limits as recommended by the standard. The absolute limits are dependent on this choice of

base station type because they apply to different power levels (as the P-rated levels in the drop down menu indicate). The relative limit is also set.

- The results are presented in both tabular and display format. In the tabular display, the reference channel power and the offset, bandwidth, integrated power, and relative power of the adjacent channels are presented. An expanded view format is also available.
- The integrated power in the display is shaded blue in each adjacent channel region.
- Each adjacent channel is clearly shown. Interadjacent channel gaps are shown in gray.
- The limit lines are shown in different colors. Based on the Mask option chosen in the Offset & Limits table, failures in the different bands are shown in red on violation. By default for LTE Standards Presets, the Mask option is Abs & Rel, so the failure is shown only when the calculated power violates both relative and absolute power limits.
- The respective rows that violate both absolute and relative limits are also shown in red. Pass/fail information appears in the top left corner of the display.
- This measurement can be performed in Real Time or Non-Real Time mode. In Real Time mode, a single acquisition required for the entire span needed for this measurement is taken and the measurement is done. In the Non-Real Time mode, a separate acquisition for each adjacent channel region is taken and analyzed.

Non-real time mode is also useful when the bandwidth offered by the instrument does not support the span requirement for this measurement. The span requirement comes from the choice of adjacent channel type and channel bandwidth. In such cases, Non-Real Time mode can be used.

- For UTRA adjacent channels, filtering of the adjacent channels with an RRC filter of the same bandwidth is employed, as suggested by the standard. The chip rate is set to 3.84 Mcps for FDD and when the channel bandwidth is 1.4 or 3 MHz. For TDD, when the channel bandwidth is more than 3 MHz, you can choose the UTRA chip rate from the ACLR test setup (Presets > Standards > LTE). The options are 1.28, 3.84, and 7.68 Mcps.
- The standard recommends this measurement be done on the transmitter ON period when the signal being analyzed is TDD. An on slot detection module in the analysis helps you to do the measurement in the transmitter ON period.
  - If you are doing the measurement in Real Time, you can also load the Time Overview display, which will indicate to you the region in which the measurement is done (a purple line at the bottom). You can see the measurement is done only on the valid transmitter ON period. However, when a bursty data is not available, the measurement is done on the available analysis length.

---

**NOTE.** Real-Time mode can be set from the [Parameters tab](#) of the Settings control panel of the LTE ACLR display. Disabling Real-Time mode is the non-Real-Time setting.

---

- When the measurement is done in non-Real Time mode, separate acquisitions are obtained for each adjacent channel band. In such cases, it cannot be ensured that all the adjacent channel acquisitions are done on the transmitter ON period. It is recommended that you use either Real time mode or an external trigger from the source (when used with non-Real-Time mode) to do the measurement. An external trigger from the source can ensure that all the acquisitions are done just

after the rising edge of the burst. A status message in the display indicates when to use Real-Time mode or external trigger for TDD signals. The status message that is displayed in the LTE ACLR display is "Signal must be triggered or measurement in Real-Time". This message is displayed when TDD is chosen from Standard Presets, when Real-Time is not enabled from the Parameters tab of the control panel, and when the instrument is in Free Run mode.

## Operating Band Unwanted Emission (FDD and TDD)

This measurement finds the power in the offset regions recommended by the standard. The offset channel specification and limits for comparison are set based on the selected channel bandwidth, frame structure, and base station types in Standards Preset. It calculates the integrated power in the different offset regions and also indicates pass/fail information based on comparison with absolute and relative limits in the SEM display.

More information about this measurement:

- Select the SEM test setup from Presets > Standards > LTE to perform the measurement. This will load the displays found in the [LTE preset test setups table](#).
- This test setup allows you to select the frame structure, channel bandwidth, and base station type with the following restrictions:
  - Frame Structure can be TDD or FDD
  - Channel Bandwidth can be 1.4, 3, 5, 10, 15, or 20 MHz
  - Base station type can be Wide Area Cat A, Wide Area Cat B - Option 1, Home Area ( $2 \text{ dBm} \leq P \leq 20 \text{ dBm}$ ), Local Area, Medium Area ( $P \leq 31 \text{ dBm}$ )
- Based on the settings chosen by the user from Standards Presets the offset and bandwidth of the offset channels are set in the Offset and Limits tab of the control panel of SEM display.
- The settings also include the absolute power limit for comparison based on the base station type. The standard recommends the settings in dBm/MHz for each base station type. The power limits are appropriately scaled with the integration bandwidth and presented in dBm in the Offset and Limits table of the display. The integrated power in each adjacent band is compared with the absolute and relative limits as suggested by the standard. The absolute limits are dependent on this choice of base station type as they apply to different power levels (as the P-rated levels in the drop down menu indicate). The relative limit is also set.
- The results are presented both in tabular and display form.
- The Pass/ Fail information is also shown in the top left corner of the display.
- For TDD signals, a detection module in the analysis helps you to do the measurement in the transmitter ON period.
  - If you are doing the measurement in Real Time, you can also load the Time Overview display, which will indicate to you the region in which the measurement is done (a purple line at the bottom). You can see the measurement is done only on the valid transmitter ON period. However, when a bursty data is not available, the measurement is done on the available analysis length.

---

**NOTE.** *Real-Time mode can be set from the [Parameters tab](#) of the Settings control panel of the SEM display. Disabling Real-Time mode is the non-Real-Time setting.*

---

- When the measurement is done in non-Real Time mode, separate acquisitions are obtained for each adjacent channel band. In such cases, it cannot be ensured that all the adjacent channel acquisitions are done on the transmitter ON period. It is recommended that you use either Real-Time mode or an external trigger from the source (when used with non-Real Time mode) to do the measurement. An external trigger from the source can ensure that all the acquisitions are done just after the rising edge of the burst. A status message in the display indicates when to use Real Time mode or external trigger for TDD signals. The status message that is displayed in the LTE SEM display is "Signal must be triggered or measurement in Real-Time". This message is displayed when TDD is chosen from Standard Presets, when Real-Time is not enabled from the Parameters tab of the control panel, and when the instrument is in Free Run mode.

- Settings for all tables provided in the base station conformance testing document (3GPP TS 36.141 v12.5 – for Release 12) has been incorporated from Standards Presets.
  - Wide Area and Local area settings from standard presets are provided as per the standard document.
  - Home Area base station settings loaded from standard presets correspond to Reference power levels between 2 to 20 dBm.
  - Medium Area base station settings loaded from standard presets correspond to reference power levels less than 31 dBm.
- The following settings are provided in a \*.csv file that you can load. These files will be available in the installed directory (Example Files\LTE).
  - Wide Area category B Option 2 (for specific bands specified by the standard). There are three csv files based on operating band and channel bandwidth for Wide Area.
    - LTE\_SEM\_1.4MHz\_CatB\_optn2\_bands\_3\_8.csv
    - LTE\_SEM\_3MHz\_CatB\_optn2\_bands\_3\_8.csv
    - LTE\_SEM\_5to20MHz\_CatB\_optn2\_bands\_1\_3\_8\_32\_33\_34.csv
  - Home Area base station for power level less than 2 dBm. There are six csv files based on operating band and channel bandwidth for Home Area.
    - LTE\_SEM\_1\_4MHz\_above3GHz\_Home2P.csv
    - LTE\_SEM\_1\_4MHz\_below3GHz\_Home2P.csv
    - LTE\_SEM\_3MHz\_above3GHz\_Home2P.csv
    - LTE\_SEM\_3MHz\_below3GHz\_Home2P.csv
    - LTE\_SEM\_5to20MHz\_above3GHz\_Home2P.csv
    - LTE\_SEM\_5to20MHz\_below3GHz\_Home2P.csv
  - Medium Area base station for reference power level between 31 to 38 dBm. There are six csv files based on operating band and channel bandwidth for Medium Area.
    - LTE\_SEM\_1\_4MHz\_above3GHz\_Medium31P38P.csv
    - LTE\_SEM\_1\_4MHz\_below3GHz\_Medium31P38.csv
    - LTE\_SEM\_3MHz\_above3GHz\_Medium31P38.csv
    - LTE\_SEM\_3MHz\_below3GHz\_Medium31P38.csv
    - LTE\_SEM\_5to20MHz\_above3GHz\_Medium31P38.csv
    - LTE\_SEM\_5to20MHz\_below3GHz\_Medium31P38.csv
  - Additional requirements given in Section 6.6.3.5.3 in 3GPP TS 36.141 v12.5 standard document. There are 14 csv files based on the requirements in the standard document. The files are named based on the table numbers of Section 6.6.3.5.3 of the standard document (3GPP TS 36.141 V12.5.0 (2014-09)).

## Cell ID (TDD and FDD)

The dominant Cell ID in the transmitted LTE signal is detected and presented in the Constellation display. Along with the Cell ID scalar results, the group and sector ID are also presented. The constellation of the Primary and Second Synchronization signals are presented too.

More information about this measurement:

- Select the Cell ID test setup from Presets > Standards > LTE to perform the measurement. This will load the displays found in the [LTE preset test setups table](#).
- This test setup allows you to select the frame structure and channel bandwidth with the following restrictions:
  - Frame Structure can be TDD or FDD
  - Channel Bandwidth can be 1.4, 3, 5, 10, 15, or 20 MHz
- Analysis of the LTE signals is done based on the settings you choose in the Standards Presets.
- The constellation for PSS and SSS (Primary and Secondary Synchronization Signals) are shown. You can optionally select to view only the PSS or SSS constellation (selection is made in the Trace tab).
- Equalization based on PSS (Primary Synchronization Signal) data can be enabled using the "Enable Equalization" option (available in the Analysis Params tab of the Settings Control panel.) This Equalization is based on PSS data and applied on other parts of the OFDM symbol.

## TOFF (TDD)

The  $T_{\text{OFF}}$  measurement is done only for the TDD signal. It is a measure of the off slot power in a LTE TDD signal.

More information about this measurement:

- Select the Channel Power and TDD TOFF Power test setup from Presets > Standards > LTE to perform the measurement. This will load the displays found in the [LTE preset test setups table](#).
- This test setup allows you to select the frame structure and channel bandwidth with the following restrictions:
  - Frame Structure can be TDD or FDD
  - Channel Bandwidth can be 1.4, 3, 5, 10, 15, or 20 MHz
- Analysis of the LTE signals is done based on the settings you choose in the Standards Presets.
- By default, the  $T_{\text{OFF}}$  measurement is the average of several non-overlapping and adjoining 70  $\mu\text{s}$  windows in the off slot region. This measurement is done only on the 70  $\mu\text{s}$  window in the center of the off slot region, when the **Average over entire offslot region** option is disabled.
- The scalar result is presented in dBm/MHz.
- The result is compared against limits that can be defined using the Limits tab of the control panel. By default, these limits are set based on the Base Station conformance testing document (3GPP TS 36.141 v12.5 – for Release 12). Pass/fail is shown in the top left corner of the LTE Power vs Time display.

## LTE Status Messages

The following status messages may appear. Each message indicates the related condition (Description) shown in the following table.

| Status message  | Description  |
|---|--|
| LTE Analysis: Signal must be triggered or measured in Real Time     | <p>Shown in the display and status bar when the user attempts to make the ACLR measurement and SEM measurement for TDD signals in non-Real-Time mode and when the instrument is in free run.</p> <p>Since the standard recommends the ACLR measurement to be done in transmitter ON periods only, either the measurement has to be done in Real-Time mode (in which a detection module ensures the measurement is done on the transmitter ON period) or has to be externally triggered if in non-Real-Time mode.</p> |
| LTE Analysis: Recovery done on PSS/SSS on the center 62 carriers    | <p>Shown in the status bar to indicate that the constellation of PSS and SSS shown in the LTE constellation is done based on analysis of PSS and SSS signals in the center 62 carriers of the corresponding OFDM symbols.</p> <p>Therefore, you may see the correct constellation and Cell ID even when a different channel bandwidth is chosen from the control panel because the analysis for PSS and SSS is done only on the center carriers.</p>   |
| LTE Analysis: Analysis Failure - Synchronization Sequence not found | Shown in the status bar to indicate that a valid Cell ID (from PSS and SSS) is not detected.   |

## LTE ACLR display

The LTE Adjacent Channel Leakage Ratio (ACLR) display shows the ratio of the mean power centered on the assigned channel frequency to the mean power centered on an adjacent channel frequency. In the 3GPP specification, both the main channel and adjacent channels are required to be filtered with RRC (Root Raised Cosine) filters.

To show the LTE ACLR display you can select **Presets > Standards > LTE**.

---

**NOTE.** Loading the LTE ACLR display from the **Presets > Standards > LTE** menu is recommended. This loads the control settings based on the selected options.

---


You can also load the LTE ACLR display as follows:

1. Press the **Displays** button or select **Setup > Displays**.
2. In the **Select Displays** window, select **LTE Analysis** in the **Measurements** box.
3. In the Available displays box, double-click the **LTE ACLR** icon or select the icon and click **Add**. The LTE ACLR icon will appear in the **Selected displays** box and will no longer appear under Available displays.

4. Click the **OK** button to show the display.
5. Select **Setup > Settings** to display the control panel.
6. Select **Replay/Run** to take measurements on the acquired data.


### Elements of the Display

The following image shows the LTE ACLR display.



| Item | Display element | Description   |
|------|-----------------|---|
| A    | Pass/Fail       | Indicates Pass or Fail based on Absolute and Relative limits set by the user. The failure condition is set by Mask in the Offset and Limits Table tab of the Settings control panel. By default, the mask is set to Abs & Rel, meaning failure is reported only when both absolute and relative results fail against the respective limits. |
| B    | Ref Detection   | Set to Avg (VRMS) or +Peak based on the choice made in Processing tab of the Settings control panel.  |
| C    | Normal          | Indicates how the result is presented over multiple sweeps. This selection is made in the Processing tab of the control panel. Displays Average Count if Avg (VRMS) is checked in the Processing tab in the Settings control panel.   |
| D    | Clear           | Resets measurement. Clears all values.  |
| E    | Plot            | Shows the reference channel and adjacent channels and the regions in between them. The Absolute and Relative limit lines are also shown. The integrated power is shown as a band. Scalar results are also shown.  |
| F    | Span            | Adjust the span of the graph in symbols.  |
| G    | Results Table   | Tabulates the results in each adjacent band. Table shows the offset, bandwidth, Integrated Absolute and Relative power in the upper and lower adjacent channel regions.   |
| H    | CF              | Center Frequency at which the measurement is performed.   |
| I    | Ref Total Power | Gives the power in the Reference channel.   |
| J    | Autoscale       | Adjusts the vertical and horizontal settings so that the entire trace fits in the graph.  |
| K    | Bottom readout  | Displays the value indicated by the bottom of graph.  |
| L    | dB / div        | Shows the dB per each division in the Y axis of the plot.   |
| M    | Top of graph    | The vertical scale is normalized with appropriate power units.  |
| N    | Expand button   | Shows the results table in its own window with larger font size.  |

Clicking on the **Expand** button allows you to view the results table in a separate window, as shown in the following image.



| Channel | Freq Offset | Bandwidth | Lower Abs  | Lower Rel | Upper Abs  | Upper Rel |
|---------|-------------|-----------|------------|-----------|------------|-----------|
| AL&AU   | 7.50 MHz    | 3.84 MHz  | -61.53 dBm | -46.95 dB | -61.11 dBm | -46.53 dB |
| BL&BU   | 12.50 MHz   | 3.84 MHz  | -61.53 dBm | -46.95 dB | -61.32 dBm | -46.73 dB |

## LTE ACLR Settings

### Front Panel: Settings

### Application Toolbar:

The LTE ACLR Settings control panel provides access to settings that control parameters of the LTE ACLR display.

| Settings tab  | Description  |
|---|--|
| <a href="#">Channels (see page 402)</a>                   | Allows you to control how the measurement is performed. When in Real Time, the RBW and VBW settings apply for all channels (including offset regions) and the other parameters apply only for the reference channel. When Non-Real Time mode is selected, all information in this tab, including RBW and VBW, only applies to the reference channel. |
| <a href="#">Parameters (see page 404)</a>                 | Specifies several characteristics that control how the measurement is made. These parameters are used by the instrument to analyze the input signal.   |
| <a href="#">Processing (see page 406)</a>                 | Specifies settings for detection on the reference channel and the offsets. Specifies the function setting on how calculations are done across multiple sweeps.   |
| <a href="#">Offsets &amp; Limits Table (see page 407)</a> | Allows you to select the characteristics of offsets and mask limits.   |
| <a href="#">Scale (see page 408)</a>                      | Specifies the vertical and horizontal scale settings.  |
| <a href="#">Prefs (see page 410)</a>                      | Allows you to select to show or hide the graticule, power level, limits, and marker readouts.  |

## LTE Channel Spectrum display


The LTE Channel Spectrum display shows the spectrum of the input signal across one channel. Channel power measures the average power within the selected bandwidth and is expressed in dBm. Other units can be selected from the Units tab of the Analysis control panel. The measured occupied bandwidth is calculated as the bandwidth containing 99% of the total integrated power within the selected span around the selected center frequency.

To show the LTE Channel Spectrum display:

1. Press the **Displays** button or select **Setup > Displays**.
2. From the **Measurements** box, select **LTE Analysis**.
3. Double-click the **LTE Channel Spectrum** icon in the **Available Displays** box. This moves the LTE Channel Spectrum icon to the **Selected displays** box.
4. Click **OK** button to view the display.
5. Select **Setup > Settings** to display the control panel.
6. Select **Replay/Run** to take measurements on the acquired data.

### Elements of the Display

The following image shows the LTE Channel Spectrum display.



| Item | Display element | Description  |
|------|-----------------|--|
| A    | Normal          | Displays Average Count if Average is turned on (set to Time Domain or Frequency Domain) from the Measurement Params tab in the Settings control panel. |
| B    | Clear           | Resets measurement. Clears all values.   |
| C    | Plot            | Specifies the value shown at the center of the graph display.  |
| D    | Scale           | Adjust the span of the graph.  |
| E    | Position        | Displays the horizontal position of the trace on the graph display.  |
| F    | Scalar results  | Shows the Channel Power and Occupied Bandwidth measurement results.  |
| G    | Autoscale       | Adjusts the vertical and horizontal settings so that the entire trace fits in the graph.   |
| H    | Bottom readout  | Displays the value indicated by the bottom of graph.   |
| I    | dB / div        | Shows the dB per each division in the Y axis of the plot.  |
| J    | Top of graph    | The vertical scale is normalized with appropriate power units.   |

## LTE Channel Spectrum Settings

### Front Panel: Settings

#### Application Toolbar:

The settings for the LTE Channel Spectrum display are shown in the following table.

| Settings tab                                      | Description  |
|---|--|
| <a href="#">Freq &amp; RBW (see page 413)</a>     | Allows you to specify the frequency and resolution bandwidth used for the measurement.   |
| <a href="#">Measurement Params (see page 414)</a> | Allows you to set averaging (Time Domain, Off, or Frequency Domain) and to set the number value associated with the Average setting. |
| <a href="#">Channels (see page 402)</a>           | Allows you to set the Channel BW for the Channel Power measurement.  |
| <a href="#">Scale (see page 408)</a>              | Allows you to specify the horizontal and vertical scale settings.  |
| <a href="#">Prefs (see page 410)</a>              | Allows you to select to show or hide certain display elements.   |

## LTE Constellation display


The LTE Constellation display shows the constellation of Primary (PSS) and Secondary (SSS) Synchronization Signals. Scalar results include Cell ID, Group ID, Sector ID, and frequency error.

1. Press the **Displays** button or select **Setup > Displays**.
2. In the **Select Displays** dialog, select **LTE Analysis** in the **Measurements** box.
3. In the Available displays box, double-click the **LTE Constellation** icon or select the icon and click **Add**. The LTE Constellation icon will be added to the **Selected displays** box and will no longer appear under Available displays.

4. Click **OK** to show the display.
5. Select Replay/Run to take measurements on the acquired data.

### Elements of the Display

The following image shows the LTE Constellation display.



| Item | Display element | Description  |
|------|-----------------|--|
| A    | Scalar results  | Gives results of Cell ID, Group and Sector ID information along with Frequency error.          |
| B    | Plot            | Displays PSS and SSS (Primary and Second Synchronization Signal) constellation.                |
| C    | Status message  | Shows relevant status messages.<br>(You can read about status messages <a href="#">here</a> .) |
| D    | Marker readout  | Shows marker type, magnitude, phase and subcarrier number.                                     |

## LTE Constellation Settings

### Front Panel: Settings

#### Application Toolbar:

The settings for the LTE Constellation display are shown in the following table.

| Settings tab                                     | Description   |
|--|---|
| <a href="#">Modulation Params (see page 415)</a> | Specifies the frame structure and channel bandwidth.  |
| <a href="#">Analysis Params (see page 415)</a>   | Allows you to enable Equalization based on PSS (Primary Synchronization Signal).  |
| <a href="#">Analysis Time (see page 416)</a>     | Specifies parameters that define the portion of the acquisition record that is used for analysis. Also allows you to specify the Units (Symbols or Seconds) and Time Zero Reference (Trigger or Acquisition Start) for LTE Analysis displays. |
| <a href="#">Trace (see page 417)</a>             | Allows you to set the display characteristics of the traces.  |
| <a href="#">Scale (see page 408)</a>             | Specifies the vertical and horizontal scale settings.   |

## LTE Power vs Time display

The LTE Power vs Time display shows the filtered power of the data and marks where the  $T_{OFF}$  measurement is done. The  $T_{OFF}$  scalar results are also shown.

You can select **Presets > Standards > LTE** to view the display.

---

**NOTE.** Loading the LTE Power vs Time display from the **Presets > Standards > LTE** menu is recommended. This loads the control settings based on the selected options.

---


You can also load the LTE Power vs Time display as follows:

1. Press the **Displays** button or select **Setup > Displays**.
2. In the **Select Displays** dialog, select **LTE Analysis** in the **Measurements** box.

3. In the Available displays box, double-click the **LTE Power vs Time** icon or select the icon and click **Add**. The LTE Power vs Time icon will be added to the **Selected displays** box and will no longer appear under Available displays.
4. Click **OK** to show the display.
5. Select Replay/Run to take measurements on the acquired data.

### Elements of the Display

The following image shows the LTE Power vs Time display.



| Item | Display element | Description   |
|------|-----------------|---|
| A    | Pass / Fail     | Shows Pass or Fail based on whether the $T_{OFF}$ measurement (off slot power) is below the recommended limit set in the Limit tab of the Settings control panel or not. The Pass/Fail information is displayed only when the chosen frame structure is TDD and if a valid off slot is found. |
| B    | Plot            | Displays the filtered signal in case of TDD and the region in which the off slot power is measured is graphically shown. When the frame structure is chosen as FDD, the power in the input signal is displayed.   |
| C    | Scale           | Adjust the span of the graph in symbols.  |
| D    | Analysis Start  | Gives the start of analysis region.   |
| E    | Toff            | The $T_{OFF}$ (off slot power) is measured and displayed here when a valid off slot is found for a TDD signal.  |
| F    | Autoscale       | Adjusts the vertical and horizontal settings so that the entire trace fits in the graph.  |
| G    | Bottom readout  | Displays the value indicated by the bottom of graph.  |
| H    | dB / div        | Shows the dB per each division in the Y axis of the plot.   |
| I    | Top of graph    | The vertical scale is normalized with appropriate power units.  |

## LTE Power vs Time Settings

### Front Panel: Settings

#### Application Toolbar:

The settings for the LTE Power vs Time display are shown in the following table.

| Settings tab                                     | Description   |
|--|---|
| <a href="#">Modulation Params (see page 415)</a> | Specifies the frame structure and channel bandwidth.  |
| <a href="#">Analysis Time (see page 416)</a>     | Specifies parameters that define the portion of the acquisition record that is used for analysis. Also allows you to specify the Units (Symbols or Seconds) and Time Zero Reference (Trigger or Acquisition Start) for LTE Analysis displays. |
| <a href="#">Prefs (see page 410)</a>             | Allows you to select to show or hide the graticule and marker readouts.   |
| <a href="#">Scale (see page 408)</a>             | Specifies the vertical and horizontal scale settings.   |
| <a href="#">Limit (see page 418)</a>             | Specifies the mask limits.  |

## LTE Analysis Measurement Settings

### Front Panel: Settings

#### Application Toolbar:

The control panel tabs in this section are shared between the displays in LTE Analysis (Setup > Displays). Some tabs are shared by all the displays and some tabs are shared by only a subset of displays. The settings available on some tabs change depending on the selected display.

### Controls for LTE Analysis displays

| Settings tab  | Description  |
|---|--|
| <a href="#">Modulation Params (see page 415)</a>          | Allows you to set the frame structure and channel bandwidth.<br><i>Available for these displays: LTE Power vs Time, LTE Constellation.</i>   |
| <a href="#">Analysis Params (see page 415)</a>            | Allows you to enable Equalization based on PSS (Primary Synchronization Signal).   |
| <a href="#">Analysis Time (see page 416)</a>              | Specifies parameters that define the portion of the acquisition record that is used for analysis. Also allows you to specify the Units (Symbols or Seconds) for LTE Analysis displays.<br><i>Available for these displays: LTE Power vs Time, LTE Constellation.</i> |
| <a href="#">Trace (see page 417)</a>                      | Specifies the trace type (Vectors, Points, and Lines), the number of points per symbol when chosen to view as Vectors or Lines, and to control the view as Freq Dev or IQ.<br><i>Available for the LTE Constellation display only.</i>                               |
| <a href="#">Prefs (see page 410)</a>                      | Allows preferences with Radix display and marker readouts.<br><i>Available for these displays: LTE Power vs Time, LTE ACLR, LTE Channel Spectrum, SEM.</i>   |
| <a href="#">Scale (see page 408)</a>                      | Defines the vertical and horizontal axes.<br><i>Available for all LTE displays.</i>  |
| <a href="#">Offsets &amp; Limits Table (see page 407)</a> | <b>NOTE.</b> <i>Available for the LTE ACLR display only.</i>   |
| <a href="#">Freq &amp; RBW (see page 413)</a>             | Specifies the frequency and resolution bandwidth used for the measurement.<br><b>NOTE.</b> <i>Available for the LTE Channel Spectrum display only.</i>   |
| <a href="#">Parameters (see page 404)</a>                 | Specifies several characteristics that control how the measurement is made.  |
| <a href="#">Processing (see page 406)</a>                 | Specifies settings for detection on the reference channel and the offsets. Specifies the function setting on how calculations are done across multiple sweeps.   |
| <a href="#">Limit (see page 418)</a>                      | Allows you to define limits for pass/fail comparison with calculated values. The default values are as recommended in the test specification.  |
| <a href="#">Channels (see page 402)</a>                   | Specifies the Channel BW for the Channel Power measurement. Allows you to set RBW, VBW, power reference, as well.<br><i>Available for these displays: LTE ACLR, LTE Channel Spectrum.</i>  |
| <a href="#">Measurement Params (see page 414)</a>         | Allows you to set averaging (Time Domain, Off, or Frequency Domain) and to set the number value associated with the Average setting.   |

## Channels tab - LTE

The Channels tab allows you to control how the measurement is performed.

The following image shows the Channels tab for the LTE ACLR display. When in Real Time, the RBW and VBW settings apply for all channels (including offset regions) and the other parameters such as

Power Reference, Channel Width, and Integrated bandwidth apply only to the reference channel. When Non-Real Time mode is selected, all information in this tab, including RBW and VBW, only applies to the reference channel.

| Settings   | Description  |
|--|--|
| Reference  |  |
| Power Reference                                      | The value used to calculate relative measurements. When Auto is selected, the calculated reference channel power is displayed. When Auto is unchecked, you can enter a value for the reference power, and the measured reference power is not used or displayed.   |
| Channel Width  | Specifies the width of the reference channel.  |
| Integration BW                                       | Specifies the integration bandwidth used to compute the total power in the reference channel.  |
| All Channels (displayed for Real-Time mode)          |  |
| Reference Channel (displayed for Non-Real Time mode) |  |
| RBW  | Sets the RBW   |
| VBW  | Enables/disables the Video Bandwidth filter. VBW is used in traditional swept analyzers to reduce the effect of noise on the displayed signal. The VBW algorithm in the analyzer emulates the VBW filters of traditional swept analyzers. When the check box next to VBW is not checked, the VBW filter is not applied.<br><br>The maximum VBW value is the current RBW setting. The minimum VBW value is 1/10,000 of the RBW setting. |

The following image shows the Channels tab for the LTE Channel Spectrum display. This tab allows you to set the channel bandwidth, which in turn sets the span for LTE Channel Spectrum.

| Settings          | Description   |
|-------------------|---|
| Channel Bandwidth | Specifies the Channel Bandwidth, which in turn sets the span of the LTE Channel Spectrum. |

## Parameters tab - LTE


The Parameters tab enables you to specify several parameters that control the LTE ACLR measurements are made.

The screenshot shows the 'Parameters' tab of the 'LTE ACLR settings' dialog. The dialog has a title bar with tabs: 'Channels', 'Parameters', 'Processing', 'Offsets & Limits Table', 'Scale', and 'Prefs'. The 'Parameters' tab is active. Inside the tab, there are several controls: 'Meas Freq' is set to '1.00000 GHz'; 'Filter shape' is a dropdown menu set to 'Gaussian'; 'Step' is set to '2.0000 MHz' with an 'Auto' checkbox checked; a 'Real-Time' checkbox is unchecked; a 'Measure Noise Floor' button is present; and an 'Apply noise correction' checkbox is unchecked. On the left side of the dialog, there is a 'Standard Presets' button.

| Setting                | Description   |
|------------------------|---|
| Meas Freq              | Specify the frequency of the signal to be measured.   |
| Step                   | Sets the increment size when changing the Frequency using the knob or mouse wheel.<br>Auto: When Auto is enabled, the step size is adjusted automatically based on the span setting.  |
| Real-Time              | When Real-Time mode is enabled, the entire LTE ACLR span is measured using a real-time/contiguous acquisition. Not all described parameters are available in Real-Time mode.<br><br>When Real-Time is disabled (non-Real-Time mode), a separate acquisition for each region is taken and analyzed. Non-Real-Time mode is also useful when the bandwidth offered by the instrument does not support the span requirement for this measurement. The span requirement is guided by the choice of adjacent channel type and channel bandwidth selected in LTE Standards Presets.  |
| Measure Noise Floor    | Takes preliminary acquisitions to measure the instrument noise floor. This initiates a noise correction. A noise correction signal is created by switching off the RF input and performing acquisitions of the instrument's internal noise. Fifty acquisitions are averaged to create the noise reference signal. The noise reference signal is measured for the Reference channel and each Offset is defined by the measurement settings.  |
| Apply noise correction | This item is enabled and the check box automatically checked after the noise reference signal is taken when the Measure Noise floor button is clicked. This initiates noise reference subtraction from the incoming signal power for each region to create the corrected result. All calculations are performed in Watts and then converted to the desired units.<br><br>The amount of noise correction is limited to 20 dB to avoid the possibility of a negative power measurement. This is a rare condition that could occur if the subtraction of the reference power from the channel power results in a negative value (or "infinite" dBm). The noise reference for a region is subtracted from each trace point in the channel, rather than offsetting the entire region by a single amount. This produces a smooth trace with no discontinuities at the region edges.<br><br><b>NOTE.</b> If any relevant settings (such as reference level, frequency, span, RBW) are changed once the noise reference is measured, the following warning message will be displayed to notify you that Noise Correction was not applied: Noise correction not applied - select Measure Noise Floor for new noise correction. |
| Filter shape           | Specifies the shape of the filter determined by the window that is applied to the data record, in the spectrum analysis, to reduce spectral leakage. 3GPP specifies a Gaussian window shape be applied to the reference channel measurements.<br><br><b>Gaussian:</b> This filter shape provides optimal localization in the frequency domain.<br><b>Rectangular:</b> This filter shape provides the best frequency, worst magnitude resolution. This is essentially the same as no window.   |

## Processing tab - LTE

The Processing tab controls the detection settings for the Reference Channel and Offsets, as well as selecting the function for the LTE ACLR display.




| Setting                           | Description  |
|-----------------------------------|--|
| Detection (in each sweep)         |  |
| Reference Channel                 | Specifies the Detection method used for the Reference Channel. Detection is used to produce the desired measurement result (peak or average) and to reduce the results of a measurement to the desired number of trace points.               |
| Avg (VRMS)                        | For each sweep, each point of the trace is the result of determining the RMS voltage value for the last 'n counts' of the collected traces for the same point. When 'n count' has not been reached, partial averaging results are displayed. |
| +Peak                             | Selects the +Peak detection method. With this method, the highest value is selected from the results to be compressed into a trace point.  |
| Offsets                           | Specifies the detection method used for the offsets.   |
| Avg (VRMS)                        | Selects the Average Vrms detection method. With this method, each point on the trace in each offset is the result of determining the RMS Voltage value for all of the results values it includes.  |
| +Peak                             | Selects the +Peak detection method. With this method, the highest value is selected from the results to be compressed into a trace point.  |
| Function (across multiple sweeps) |  |
| Normal                            | When a new trace has been computed, it replaces the previous trace.  |
| Max Hold                          | With each sweep, each trace point in the new trace is compared to the point's value in the old trace and the greater value is retained for display and subsequent comparisons.   |
| Avg (VRMS)                        | For each sweep, each point on the trace is the result of determining the RMS Voltage value for all of the collected traces' values for the same point.   |
| Count                             | Enter the Avg (VRMS) value. Displayed only when Avg (VRMS) is selected.  |


## Offsets and Limits Table tab - LTE

The Offsets and Limits Table tab is used to specify parameters that define Offsets and masks for the LTE ACLR display. The following images show the tab and expanded view of the tab, respectively.

The following image shows the Offsets and Limits Table tab when Real-Time is selected in the Parameters tab.



The following image shows the Offsets and Limits Table tab when Real-Time is not selected in the Parameters tab (referred to as non-Real-Time mode).



The following image shows the expanded view of the Offsets and Limits Table (click the **Expand** button to view).



| Setting              | Description   |
|----------------------|---|
| <b>Buttons</b>       |   |
| Expand (button)      | Displays the Offsets and Limits Table in a new, resizable window.   |
| Reset Layout         | Allows you to reorder columns in the Offsets & Limits Table by dragging the columns to a new position. Clicking Reset Layout returns the column order to the factory default order.   |
| Load                 | Click to load a saved Offsets & Limits table from a file.   |
| Save As              | Click to save the current Offsets & Limits table to a file.   |
| Done                 | When the table is expanded, click Done when you have finished editing the table to save your changes and close the expanded table display.  |
| <b>Table columns</b> |   |
| On                   | Specifies whether or not measurements are taken in the specified offset.  |
| Freq Offset (Hz)     | Specifies the frequency offset for the offset region (adjacent channel region) from the Center frequency of the Reference Channel. This offset is always specified from the center frequency of the Reference Channel to the center frequency of the Offset region. |
| Start (Hz)           | Start Frequency of the selected offset.   |
| Stop (Hz)            | Stop Frequency of the selected offset.  |
| Bandwidth (Hz)       | Specifies the bandwidth of the Offset region.   |
| Side                 | Specifies whether the specified range appears on both side of the carrier frequency or just one side (left or right).   |
| RBW (Hz)             | Specifies the RBW for the selected range.   |
| n(RBWS)              | An integer value that specifies how many times to multiply the RBW to set the Filter bandwidth.   |
| Filter BW (n x RBW)  | Displays the Filter BW. Filter BW is the equivalent BW of each point in the offset. When $n > 1$ , an integration technique is used to achieve the Filter BW using narrower RBWs.   |
| VBW (Hz)             | Adjusts the VBW (Video Bandwidth) value. VBW Maximum: RBW current value; VBW Minimum: 1/10,000 RBW setting.   |
| VBW (On)             | Specifies whether the VBW filter is applied.  |
| Mask                 | Select the type of limits used for Pass/Fail testing. Signal excursions that exceed the mask settings are considered violations. The available choices are Abs, Rel and Abs & Rel.  |
| Abs Limit (dBm)      | The offset region integrated power is compared against the Abs Limit value mentioned here.  |
| Rel Limit (dB)       | The offset region integrated power relative to the reference channel integrated power is compared with this limit.  |

## Scale tab - LTE

The Scale tab allows you to change the scale settings that control how the trace appears on the display but does not change control settings. In effect, these controls operate like pan and zoom controls. The Scale tab values are unique to each display. Also, note that each display uses horizontal and vertical units that are appropriate for that display. There are three versions of the Scale tab for LTE displays.

The following image shows the Scale tab for the LTE Power vs Time display.

**LTE Power vs Time Settings**

Modulation Params Analysis Time Prefs **Scale** Limit

**Vertical**

Scale: 100.00 dB

Position: 0.00 dBm

Reset

**Horizontal Full Burst**

Start: 100.000 us

Stop: 0.000 s

Reset

Standard Presets

| Settings              | Description  |
|-----------------------|--|
| Vertical              | Controls the vertical position and scale of the trace display. |
| Scale                 | Changes the vertical scale units.                              |
| Position              | Adjusts the reference level away from top of the graph.        |
| Reset                 | Resets the scale of the vertical axis to default values.       |
| Horizontal Full Burst | Controls the horizontal span.                                  |
| Start                 | Specifies the horizontal axis start point.                     |
| Stop                  | Specifies the horizontal axis end point.                       |
| Reset                 | Resets the scale of the horizontal axis to default values.     |

The following image shows the Scale tab for the LTE Channel Spectrum and LTE ACLR displays.

**LTE Channel Spectrum Settings**

Freq & RBW Measurement Params Channels **Scale** Prefs

**Vertical**

Scale: 100.00 dB

Position: 0.00 dBm

Reset Scale

Autoscale

**Horizontal**

Scale: 6.27 MHz


Position: 1.00000 GHz

Autoscale

Standard Presets

| Settings    | Description   |
|-------------|---|
| Vertical    | Controls the vertical position and scale of the trace display.                |
| Scale       | Changes the vertical scale units.   |
| Position    | Adjusts the reference level away from top of the graph.                       |
| Autoscale   | Resets the scale of the vertical axis to contain the complete trace.          |
| Horizontal  | Controls the horizontal span of the trace display and position of the trace.  |
| Scale       | Specifies the frequency range displayed in the graph                          |
| Position    | Specifies the frequency shown at the center of the graph.                     |
| Autoscale   | Resets the scale of the horizontal axis to optimize the display of the trace. |
| Reset Scale | Resets all settings to their default values.                                  |

The following image shows the Scale tab for the LTE Constellation display.



| Settings            | Description   |
|---------------------|---|
| Zoom                |   |
| Scale               | Zooms in and out of the constellation. As the scale is increased, it will zoom out. |
| Vertical Position   | Adjusts the vertical position.  |
| Horizontal Position | Adjusts the horizontal position.  |

## Prefs tab - LTE

The Prefs tab enables you to change parameters of the measurement display. The parameters available on the Prefs tab vary depending on the selected display, but include such items as enabling/disabling Marker Readout, switching the graticule display on/off, and Marker Noise mode. Some parameters appear with most displays, while others appear with only one display.

The following image shows the Prefs tab for the LTE ACLR display.

The screenshot shows the 'LTE ACLR settings' window with the 'Prefs' tab selected. The window has a sidebar with 'Standard Presets' and a main area with tabs: 'Channels', 'Parameters', 'Processing', 'Offsets & Limits Table', 'Scale', and 'Prefs'. The 'Prefs' tab contains the following settings:

- Max Trace points: 801 (per range)
- ☒ Show graticule
- ☒ Show Marker readout in graph (selected marker)
- ☒ Show Power Levels In Graph
- Show limits: Line only

The following image shows the Prefs tab for the LTE Channel Spectrum display.

The screenshot shows the 'LTE Channel Spectrum Settings' window with the 'Prefs' tab selected. The window has a sidebar with 'Standard Presets' and a main area with tabs: 'Freq & RBW', 'Measurement Params', 'Channels', 'Scale', and 'Prefs'. The 'Prefs' tab contains the following settings:

- Trace points: 801
- Trace detection: +Peak
- ☒ Show graticule
- ☒ Show Marker readout in graph (selected marker)

The following image shows the Prefs tab for the LTE Power vs Time display.

The screenshot shows the 'LTE Power vs Time Settings' window with the 'Prefs' tab selected. The window has a sidebar with 'Standard Presets' and a main area with tabs: 'Modulation Params', 'Analysis Time', 'Prefs', 'Scale', and 'Limit'. The 'Prefs' tab contains the following settings:

- ☒ Average over offslot
- ☒ Show graticule
- ☒ Show marker readout

| Setting  | Description   |
|--|---|
| Show graticule                                 | Shows or hides the graticule.   |
| Trace / Max Trace points                       | When the spectrum analysis produces more than the selected maximum number of points, the method specified in Detection control is used to decimate the result. This setting applies to both the reference channel and offsets.    |
| Show Marker readout in graph (selected marker) | Shows or hides the readout for the selected marker in the graph area.   |
| Show Power Levels In Graph                     | This display the power level (calculated integrated power in the offset region) in the graph.   |
| Show limits                                    |   |
| Shaded   | Shows limits using a shaded area. Absolute limit and Relative limit shading colors can be set under Presets > Options.  |
| Line only                                      | Shows limits using only a line. Absolute limit and Relative limit line colors can be set under Presets > Options.   |
| None   | No lines or shading are used to indicate limits in the graph. Violations of the mask are still identified by red shading.   |
| Average Over Offslot                           | Enables for averaging of offslot power in non-overlapping 70 $\mu$ s windows in the entire offslot region. If this option is disabled, the off slot power is measured only in the center 70 $\mu$ s window of the offslot region. |

# Freq and RBW tab - LTE

The Freq and RBW tab specifies frequency parameters for the Channel Power and ACPR measurements and the MCPR measurement. It is available for the LTE Channel Spectrum display.

LTE Channel Spectrum Settings

Standard Presets

Freq & RBW

Measurement Params

Channels

Scale

Prefs

Meas Freq: 1.00000 GHz


RBW: 3.000 kHz ☐ Auto

Step: 2.0000 MHz ☒ Auto ☐ VBW:

| Settings  | Description   |
|-----------|---|
| Meas Freq | Specifies the center/measurement frequency  |
| RBW       | <p>Select Auto or Manual. Adjusts the RBW for the entire measurement. This setting is Independent of the Spectrum view's RBW setting.</p> <p>Auto: When Auto is enabled, the step size is adjusted automatically based on the span setting.</p>                                     |
| Step      | <p>The Step control sets the increment/decrement size for the adjustment of the center frequency. If Auto is enabled, the analyzer will adjust the Step size as required.</p> <p>Auto: When Auto is enabled, the step size is adjusted automatically based on the span setting.</p> |
| VBW       | <p>Adjusts the Video Bandwidth value.</p> <p>The maximum VBW value is the current RBW setting. The minimum VBW value is 1/10,000 of the RBW setting.</p>  |

# Measurement Params tab - LTE


The Measurement Params tab allows you to set parameters that control the ACPR and MCPR measurements. This is available for the LTE Channel Spectrum display.



| Settings         | Description   |
|------------------|---|
| Average          |   |
| Off              |   |
| Frequency-domain | This setting takes the average linear value of the traces (so that RMS values are preserved). The number of averages is user-defined. Frequency domain averaging is available in spans larger (or smaller) than the maximum real time bandwidth. This is the mode to use unless you need to extract maximum dynamic range from an ACPR measurement.   |
| Time-domain      | This setting takes the average linear value of the traces. It is useful if you need to extract maximum dynamic range from an ACPR measurement. The number of traces is user defined, however, the signals must be triggered and repeating (meaning the signal needs to be exactly the same for each acquisition). When this condition is met, each waveform contains the same signal, but the random noise changes from acquisition to acquisition and the average value of the random noise is lowered, while the signal value remains constant. Time domain averaging is not available in spans wider than the maximum real-time bandwidth. |

## Modulation Params tab - LTE


The Modulation Params tab specifies the type of modulation used by the input signal and other parameters that define the signal format. This tab is available for the LTE Constellation and Power vs Time displays.



| Settings                     | Description  |
|------------------------------|--|
| Frame Structure              | Select the frame structure to set how the demodulation is done.            |
| Downlink FDD                 |  |
| Downlink TDD                 |  |
| Bandwidth and Spacing        |  |
| Channel Bandwidth            | Select from 6 bandwidths: 1.4, 3, 5, 10, 15, 20 MHz                        |
| Subcarrier Spacing (readout) | This readout shows the subcarrier spacing. It is always at 15 kHz for LTE. |

## Analysis Params tab - LTE

The Analysis Params tab allows you to enable equalization on the demodulated LTE packet. The equalization is based on PSS (Primary Synchronization Signal). This tab is only available for the LTE Constellation display.



| Settings            | Description  |
|---------------------|--|
| Enable Equalization | Allows you to enable Equalization based on PSS (Primary Synchronization Signal). |

## Analysis Time tab - LTE

The Analysis Time tab contains parameters that define how the signal is analyzed in the LTE Analysis displays. This tab is available for the LTE Constellation and Power vs Time displays.

The screenshot shows the 'LTE Power vs Time Settings' dialog box with the 'Analysis Time' tab selected. The 'Analysis Offset' is set to '0.000 Symbol' with an 'Auto' checkbox. The 'Analysis Length' is set to '187.500 Symbol' with an 'Auto' checkbox. The 'Time Zero Reference' is set to 'Trigger' and 'Units' is set to 'Symbols'. An 'Actual' field shows '0.000 Symbol'. A 'Standard Presets' button is located on the left side of the dialog.

| Settings            | Description  |
|---------------------|--|
| Analysis Offset     | Specifies the location of the first time sample to use in measurements.  |
| Auto                | When enabled, causes the instrument to set the Analysis Offset value based on the requirements of the selected display.  |
| Analysis Length     | Specifies the length of the analysis period to use in measurements. Length is specified in either symbols or seconds, depending on the Units setting.                                      |
| Auto                | When enabled, causes the instrument to set the Analysis Length value based on the requirements of the selected display.  |
| Actual              | This is a displayed value, not a setting. It is the Analysis Length (time or symbols) being used by the analyzer. This value may not match the Analysis Length requested (in manual mode). |
| Time Zero Reference | Specifies the zero point for the analysis time.  |
| Units               | Specifies the units of the Analysis Length to either Symbols or Seconds.   |

### Analysis Offset

Use analysis offset to specify where measurements begin. Be aware that you cannot set the Analysis Offset outside the range of time covered by the current acquisition data. (All time values are relative to the Time Zero Reference.)

You can set the Analysis Length so that the requested analysis period falls partly or entirely outside the current range of acquisition data settings. When the next acquisition is taken, its Acquisition Length will be increased to cover the new Analysis Length, as long as the Sampling controls are set to Auto. If the Sampling parameters are set to manual, or if the instrument is analyzing saved data, the actual analysis length will be constrained by the available data length, but in most cases, measurements are able to be made anyway. The instrument will display a notification when measurement results are computed from less data than requested. Range: 0 to [(end of acquisition) - Analysis Length]. Resolution: 1 effective sample (or symbol).

## Analysis Length

Use the analysis length to specify how long a period of time is analyzed. As you adjust this value, the actual amount of time for Analysis Length, in Symbol or Seconds units, is shown below the control in the "Available" readout. This setting is not available when Auto is checked. Range: minimum value depends on the standard. Resolution: 1 symbol.


## Time Zero Reference

All time values are measured from this point (such as marker position or horizontal position (in Y vs Time displays). Choices are: Acquisition Start or Trigger.

| Parameter         | Description  |
|-------------------|--|
| Acquisition Start | Time zero starts from the point at which the acquisition begins. |
| Trigger           | Time zero starts from the trigger point.                         |

## Trace tab - LTE

The Trace tab allows you to set the trace display characteristics of the LTE Constellation display.



| Setting          | Description   |
|------------------|---|
| >                | The arrow indicates which measurement name is active so that you can check it (on) or uncheck it (off). |
| On               | Check the box next the measurement name you want to turn on.  |
| Measurement Name | Selections are PSS (Primary Synchronization Signal) or SSS (Secondary Synchronization Signal)           |

## Limit tab - LTE

The Limit tab is only available for the LTE Power vs Time display. It enables you to specify the limit used for comparison against measured values of  $T_{OFF}$  measurement.

LTE Power vs Time Settings

Modulation Params Analysis Time Prefs Scale Limit

Toff psd: -83 dBm

Standard Presets

| Setting  | Description   |
|----------|---|
| Toff psd | Specifies the limit that is used for comparison against measured values of $T_{OFF}$ measurement. Pass/ Fail is determined with this limit. |

# Bluetooth Analysis

## Overview

The Bluetooth Analysis option allows you to evaluate short range RF signals to ensure that they meet Bluetooth standard per Bluetooth Special Interests Group (SIG) Test Specifications for version 4.1. This analysis option enables measurements and supports detection, demodulation, and decoding of packet information for the following three standards: Basic Rate, Low Energy, and Enhanced Data Rate (EDR). You can also select from three RF test presets for the Low Energy standard and four test presets for the Basic Rate standard. This analysis option includes modulation, power, carrier drift, and spectral measurements as mentioned in the Radio Frequency test specification document by the Bluetooth SIG. These measurements are also compared with the limits provided by the standard to give pass/fail results. (There are no pass/fail results for EDR.)

With the Bluetooth Analysis test suite, test engineers can simplify the execution of a number of transmitter tests while still enabled to modify signal parameters for in-depth signal analysis. The analysis results give multiple views of Bluetooth signal characteristics to allow the diagnosis of signal imperfections and impairments quickly and easily. Display controls allow you to selectively display the analysis results to help locate trouble spots in the signal.

## Bluetooth topics in this Help

The following information about the Bluetooth Analysis option is available:

- [Bluetooth key features \(see page 419\)](#)
- [Reference table of supported Bluetooth measurements](#)
- [Bluetooth measurements and test setups \(see page 423\)](#)
- [Bluetooth status messages \(see page 428\)](#)
- [Bluetooth Standards presets \(see page 420\)](#)
- [Bluetooth displays \(see page 423\)](#)
- [Bluetooth Settings \(see page 454\)](#)

## Bluetooth key features

- Modulation/Carrier Drift/Output Power test setup in Bluetooth Standards Preset that will calculate results and compare against recommended or user defined limits or Basic Rate and Low Energy.
- In-band Emission measurement preset with capability to compare results against recommended or user defined limits for Basic Rate and Low Energy.
- 20 dB Bandwidth and Frequency Range measurement preset with capability to compare results against recommended or user defined limits for Basic Rate.
- Appropriate masks for Out-of-band Spurious measurements for ETSI and FCC specifications.

- Enhanced Data Rate (EDR) demodulation of GFSK and PSK data and relative power measurement preset.
- [Automatic detection of standards \(see page 420\).](#)
- Automatic detection of test pattern.
- Automatic determination of packet type.
- Summary display with decoded preamble, access code, packet header, payload header, and CRC packet information along with other scalar measurements.
- Symbol Table display with color coding to show different regions in the packet.
- Freq Dev vs Time display with zoomed Octet viewing to help you understand how the modulation characteristics measurement is done.
- CF Offset & Drift display with offset (calculated in preamble and 10 bit intervals in payload) and drift results in a tabular view.
- Summary display that allows you to edit pass/fail limits for comparison with actual results. (Default is recommended limits from the Bluetooth SIG Radio Frequency test specification document.)

**Automatic detection of standards.** This function can be enabled from the Standard Parameters control panel. This automatically detects whether the Bluetooth standard is Basic Rate, Low Energy, or Enhanced Data Rate (EDR). This function is not available as a test setup from the Standards Presets window.

### Supported automated Bluetooth measurements

| Bluetooth measurement              | Bluetooth standard(s)    |
|------------------------------------|--------------------------|
| Modulation characteristics         | Basic Rate               |
|                                    | Low Energy               |
| Carrier frequency offset and drift | Basic Rate               |
|                                    | Low Energy               |
| Output power                       | Basic Rate               |
|                                    | Low Energy               |
| In-band emission /ACPR             | Basic Rate               |
|                                    | Low Energy               |
| Out-of-band spurious emission      | Basic Rate               |
|                                    | Low Energy               |
| 20 dB bandwidth                    | Basic Rate               |
| Frequency range                    | Basic Rate               |
| Power density                      | Basic Rate               |
| Relative power                     | Enhanced Data Rate (EDR) |

## Bluetooth Standards presets

**Presets > Standards**

The Bluetooth Standards preset allows you to access displays preconfigured for the Bluetooth standard and test setup you select. The test setups load the displays and control setting options suggested by the standard to perform the measurements. You can read more about how Presets work ([see page 22](#)).

There are four test setups for the Low Energy standard and eight test setups for the Basic Rate standard. There is one preset (no test setups) for the EDR standard.

The following table shows the set of Preset displays that load when the specified standards and test setups are selected.

**Table 9: Bluetooth standards, test setups, and Preset displays**


| Standard                 | Test setup                                     | Displays loaded with preset  |
|--------------------------|--|--|
| Low Energy               | Modulation/Frequency Offset/Drift/Output Power | BT Eye Diagram, BT Freq Dev vs Time, BT Summary, BT CF Offset and Drift, Spectrum, Time Overview |
|                          | In-band Emissions                              | BT InBand Emission, Spectrum, Time Overview  |
|                          | Non-compliance                                 | BT Freq Dev vs Time, BT Constellation, BT Summary, Spectrum, Spectrogram, Time Overview          |
| Basic Rate               | Modulation/Frequency Offset/Drift/Output Power | BT Eye Diagram, BT Freq Dev vs Time, BT Summary, BT CF Offset and Drift, Spectrum, Time Overview |
|                          | In-band Emissions                              | BT InBand Emission, Spectrum, Time Overview  |
|                          | Tx Output Spectrum — 20dB Bandwidth            | Bluetooth 20 dB BW, Spectrum, Time Overview  |
|                          | Non-compliance                                 | BT Freq Dev vs Time, BT Constellation, BT Summary, Spectrum, Spectrogram, Time Overview          |
| Enhanced Data Rate (EDR) | Generic  | BT Eye Diagram, BT Constellation, BT Summary, BT Symbol Table, Spectrum, Time Overview           |

### Retain current center frequency and reference level settings

The **Retain current Center Frequency setting** appears when the Bluetooth Standards Preset is chosen. This setting allows you to retain the previously used center frequency.

The **Retain current Ref Level setting** appears when the Bluetooth Standards Preset is chosen. This setting allows you to retain the previously used reference level.

To activate these settings, check the box next to the desired setting.



By default, the Ref Level setting box is checked.

By default, the Center Frequency setting box is unchecked and the Bluetooth preset displays will load with one of the center frequency values shown in the following table.

**Table 10: Default center frequencies for Bluetooth test setups**

| Standard           | Test setup                                     | Center frequency |
|--------------------|--|------------------|
| Basic Rate         | Modulation/Frequency Offset/Drift/Output Power | 2.441 GHz        |
| Low Energy         |  |                  |
| Basic Rate         | In-band Emissions                              | 2.441 GHz        |
| Low Energy         |  |                  |
| Basic Rate         | Out-of-band Spurious Emissions                 | 2.441 GHz        |
| Low Energy         |  |                  |
| Basic Rate         | Tx Output Spectrum — 20dB Bandwidth            | 2.441 GHz        |
| Basic Rate         | Tx Output Spectrum — Power Density             | 2.441 GHz        |
| Basic Rate         | Non-compliance                                 | 2.441 GHz        |
| Low Energy         |  |                  |
| Basic Rate         | Tx Output Spectrum — Frequency Range Lower     | 2.402 GHz        |
| Basic Rate         | Tx Output Spectrum — Frequency Range Higher    | 2.480 GHz        |
| Enhanced Data Rate | Generic  | 2.441 GHz        |

---

**NOTE.** *Changing analysis and display parameters recalculates the measurement results, but does not affect acquisitions.*

---

## Bluetooth displays

The displays in Bluetooth Analysis (Setup > Displays > Measurements: Bluetooth Analysis) are:

- [Bluetooth 20dB Bandwidth](#)
- [Bluetooth Center Frequency Offset and Drift](#)
- [Bluetooth Constellation](#)
- [Bluetooth Eye Diagram](#)
- [Bluetooth Frequency Dev vs Time](#)
- [Bluetooth InBand Emission](#)
- [Bluetooth Summary](#)
- [Bluetooth Symbol Table](#)
- [Time Overview](#)

## Bluetooth measurements and test setups

A variety of measurements and test setups are provided by the Bluetooth Analysis option for use in performance testing of transmitters. These test setups allow the analyzer to compare the measurement result to the standards limit. Test engineers can select from the test setups and measurements described here. The following topics contain important information you should know about specific Bluetooth measurements and test setups.

---

**NOTE.** *Although the following information describes test setups for measurements recommended by the standard document, other measurement results may also be provided as additional information for a given measurement. For example, carrier frequency offset and drift results may be provided for modulation characteristics test setups.*

---

### Modulation characteristics (Basic Rate and Low Energy)

This measurement verifies that the modulation characteristics of the transmitted signal are correct. The Bluetooth test specification recommends this measurement be done using fixed frequency, also known as *hopping off*, and can be done in Direct Tx mode. This measurement can only be done if the payload has the bit pattern 10101010 or 11110000. This measurement and test setup are set to Free Run mode by default. However, you can set the analyzer to Triggered mode (Setup > Trigger) to do the measurements.

More information about this measurement:

- Select the Modulation/Frequency Offset/Drift/Output Power test setup from Presets > Standards to perform modulation characteristics measurements. This will load the displays found in the [Bluetooth](#)

[standards, test setups, and Preset displays table](#) . Additionally, you can choose to load the BT Symbol Table and BT Constellation displays, which can provide useful information.

- The Bluetooth Analysis splits the frequency deviation results in the payload region into octets (8 bit intervals). If the payload pattern is 10101010, the maximum frequency deviation is calculated in every bit interval within an octet and recorded as  $\Delta f_{max}$ . The  $\Delta f_{max}$  results are averaged to give  $\Delta f_{2avg}$ . The percentage of  $\Delta f_{max}$  values greater than recommended lower limit is also found.
- If the payload is 11110000, the average frequency deviation is found in 4 bit intervals (2, 3, 6 and 7) and the average of these values is calculated as  $\Delta f_{1avg}$ . The ratio of  $\Delta f_{2avg}$  and  $\Delta f_{1avg}$  is also found.
- BT Freq Dev vs Time display has an octet view (zoomed view of the 8 bit intervals in Payload) and also shows the bit intervals in which the  $\Delta f_{max}$  values are computed. The octet view option also allows you to see the regions in which the calculations are done.

---

**NOTE.** *The octet view option is only available when Preamble is detected.*

---

- Specify an octet to view by entering a value in the Octet # field in the BT Freq Dev vs Time display. This also shows the maximum number of octets used for analysis. The start and end point for every octet can be clearly seen using the octet view.
- In the Summary display, the scalar results ( $\Delta f_{1avg}$ ,  $\Delta f_{2avg}$ , Ratio, Percentage of  $\Delta f_{max}$  values greater than a limit) averaged over the last 10 packets analyzed are shown along with the packet count. These averaged results can be compared against limits set in the Limits tab of the Settings control panel.
- The Points/symbol option in the Trace tab can be used to control how many samples per symbol should be used for computation of these measurements. The standard recommends 32 samples per symbol for Low Energy and 4 samples per symbol for Basic Rate. These are the default values provided for the respective standards in the Bluetooth Standards presets.

### Carrier frequency offset and drift (Basic Rate and Low Energy)

This measurement verifies that the carrier frequency offset and carrier drift of the transmitted signal is within the specified limits for the specified standard. The Bluetooth test specification recommends this measurement be done using fixed frequency, also known as *hopping off*, and can be done in Direct Tx mode. This test can be done only if the payload contains 10101010 bit pattern. This measurement and test setup are set to Free Run mode by default. However, you can set the analyzer to Triggered mode (Setup > Trigger) to do the measurements.

More information about this measurement:

- Select the Modulation/Frequency Offset/Drift/Output Power test setup from Presets > Standards to perform carrier frequency offset and drift measurements. This will load the displays found in the

[Bluetooth standards, test setups, and Preset displays table](#) . Additionally, you can choose to load the BT Symbol Table and BT Constellation displays, which can provide useful information.

- This measurement calculates the following:
  - Frequency offset in Preamble (initial Carrier Frequency Offset).
  - Frequency offset in Payload (calculated in 10 bit intervals in Payload).
  - Drift (the difference in frequency offset calculated in Payload from that calculated in Preamble).
  - Drift/50  $\mu$ s (the difference in frequency offset calculated in two 10 bit intervals in Payload separated by 50  $\mu$ s (5 intervals away)).
- The BT CF Offset and Drift display shows all results in a table.
- The measurements shows the following scalar results:
  - Preamble frequency offset (initial carrier frequency offset).
  - Maximum frequency offset in Payload (along with the index number at which the offset is at maximum).
  - Drift of frequency offset from Preamble to first 10 bit interval in Payload ( $f_1-f_0$ ).
  - Maximum drift between Payload and Preamble ( $f_n-f_0$ ) (along with index number at which the drift is maximum).
  - Maximum drift between two 10 bit intervals spaced 50  $\mu$ s away ( $f_n-f_{n-5}$ ) (along with index number at which the drift is maximum).
- The Preamble region in which the Preamble offset is calculated is shown in the BT Freq Dev vs Time display along with markings to show the region in which maximum drift occurred. Maximum drift between Payload and Preamble is shown from the end of Preamble to the end of 10 bit interval in Payload in which maximum drift occurred. Maximum drift in payload (in 50  $\mu$ s duration) is also shown.
- In the Summary display, the scalar results (mentioned above) averaged over the last 10 packets are shown along with the packet count. The results are compared against recommended limits. You can set these limits in the Limits tab of the Settings control panel.

### In-band Emissions (Basic Rate and Low Energy)

This measurement verifies that the in-band spectral emissions are within limits. The standard document recommends that this measurement be done with Hopping off, finding the integrated power in 1 MHz band (with RBW 100 kHz) in 80 channels starting from 2401 MHz to 2481 MHz. The test can be done in Direct Tx mode. The integrated power values calculated in the adjacent channels are compared against recommended limits (except the three channels around transmitted frequency). This measurement is referred to as ACPR in the Radio Frequency Test Specification for Basic Rate. It is recommended to do this measurement in Free Run mode (set in Setup > Trigger) if the analyzer does not support the required BW (80 MHz). The measurement internally searches for burst and aligns the measurement to the on-slot region.

---

**NOTE.** For analyzers with limited bandwidth, this measurement will be done over multiple acquisitions to generate the data required for the entire span. Therefore, it is important for the user to set an analysis length that will ensure that on slot region is always included.

---

More information about this measurement:

- Select the In-band Emissions test setup from Presets > Standards to perform the in-band emissions measurements. This will load the displays found in the [Bluetooth standards, test setups, and Preset displays table](#).
- The In-band Emissions display will set the Center Frequency to 2.441 GHz by default. When the Retain CF option is not enabled from Standards Presets, 80 bands (each of bandwidth 1 MHz) are analyzed and the integrated power is found (indicated by Blue lines in the plot).
- The integrated power is compared against limits for comparison (shown by white masks) in all bands except the three channels around the transmitted frequency. If any band fails, that band is shown in red color. If more than three channels fail to meet the recommended limits, a FAIL is shown in the top left corner of the display.
- The table below the plot shows the integrated power and also the limit for comparison.
- The limits for comparison are set to the values given by the standard by default, but you can manually set them from the Limits tab of the Settings control panel.
- This measurement is done only on the on slot region if a ramp-up and ramp-down portions are available. However, the measurement is done on the complete analysis region if a ramp up or ramp down is not available. It is recommended that a complete packet is included in the analysis region.
- When the analyzer does not support the 80 MHz bandwidth requirement, multiple acquisitions are taken in different bands and stitched together. This measurement is therefore best done in Free Run mode (set in Setup > Trigger) as Triggered mode might not find a proper edge in an acquisition taken for stitching purposes.

### Output Power (Basic Rate and Low Energy)

This test verifies the maximum peak and average power emitted from the EUT. The standards document recommends this test be done for a PRBS payload pattern. Also, the measurement must be done over the duration of a burst starting at preamble position. The Bluetooth test specification recommends this measurement be done using fixed frequency, also known as *hopping off*, and can be done in Direct Tx mode.

More information about this measurement:

- Select the Modulation/Frequency Offset/Drift/Output Power test setup from Presets > Standards to perform carrier frequency offset and drift measurements. This will load the displays found in the [Bluetooth standards, test setups, and Preset displays table](#).
- Calculates the average power and peak power for any Bluetooth signal. The peak power is the peak power value in the chosen on slot region.
- The measurement is done only on preamble detection; otherwise, it is invalid.
- The BT Summary display shows the averaged result over 10 packets.

## 20dB Bandwidth (Basic Rate)

This measurement verifies if the emissions inside the operating frequency range are within limits. This measurement is done with Hopping off. The difference between frequency points at which the power level drops to 20 dB below the peak power of the emission is found at 20 dB Bandwidth.

More information about this measurement:

- Select the 20dB Bandwidth test setup from Presets > Standards to perform the 20 dB emissions measurements. This will load the displays found in the [Bluetooth standards, test setups, and Preset displays table](#).
- The BT 20 dB BW display allows you to set the relative power level (x dB BW) compared to the peak power at which the frequency points are to be found.
- When you select the 20dB Bandwidth test setup, the center frequency (CF), span, and relative bandwidth (RBW) are set as recommended by the standard document.
- It is recommended this measurement be done over three different center frequencies. You can do this by changing the center frequency as needed.
- From the Params tab of the Settings control panel, you can select that the search for the x dB relative power (20 dB in this case) be performed in one of two modes: from the edges towards the peak power (inwards) or vice versa (outwards). The default for this setting is Inwards (from edges to peak power) when this measurement is chosen from the Standards Preset menu, as recommended.
- The blue markings in the plot indicate the 20 dB bandwidth and the frequency points at which the power level drops down by 20 dB from the peak power.
- The resulting 20 dB bandwidth is compared with the limits recommended by the standard document and Pass/Fail is shown in the top left corner of the plot. This is not user defined.

If the highest power value measured in step d) is equal or higher than 0 dBm:  $f = |f_H - f_L| \leq 1.0 \text{ MHz}$ .

If the highest power value measured in step d) is lower than 0 dBm:  $f = |f_H - f_L| \leq 1.5 \text{ MHz}$ .

## Frequency Range (Basic Rate)

These measurements verify if the emissions inside the operating frequency range are within limits. These measurements are done with Hopping off in two steps. The first measurement is done at the lower frequency spectrum of the Bluetooth band. (Start – 2399 MHz and Stop 2405 MHz). The second is done at the higher frequency spectrum of the Bluetooth band (Start -2475 MHz and Stop at 2485 MHz). The difference between the frequency point at which the power level drops to -30 dBm from the center frequency is measured.  $f_L$  is the difference on the lower side from the center frequency when the lower end of the spectrum is being analyzed.  $f_H$  is the difference from the center frequency on the higher side when the higher end of the spectrum is analyzed.  $f_H - f_L$  is the Frequency Range.

More information about these measurements:

- The BT 20dB BW display allows you to set the absolute power level (x dBm level) from the Parameters tab of the Settings control panel. The Bluetooth SIG Radio Frequency test specification document suggests that this value be set to -30 dBm for doing the Frequency Range measurement. If you chose to set up this measurement using start and stop frequencies and RBW, those should be

set as recommended by the standard. Separate start and stop frequencies are given for the two steps of this measurement.

- You can select for the analysis to find the x dBm frequency difference from the center frequency either on the lower side, higher side, or both sides. This option (Meas Range) can be set from the Params tab of the Settings control panel.
- The blue marking indicates the frequency difference from the center frequency based on the Meas Range option choice. You can set this option to Low while analyzing in the lower range of the frequency spectrum, and set it to High when analyzing the higher range of the frequency spectrum.

### Power Density (Basic Rate)

This measurement verifies the maximum RF output power density.

More information about this measurement:

- This measurement can be done using the Spectrum display (Displays > Measurements > General Signal Viewing) and by setting appropriate parameters as recommended in the Bluetooth SIG Radio Frequency test specification document.
- The Bluetooth test specification recommends this measurement be done using fixed frequency, also known as *hopping off*, and can be done in Direct Tx mode.

### Out-of-band Spurious Emission (Basic Rate and Low Energy)

This measurement can be done for FCC or ETSI masks using the [Spurious](#) display.

### Relative Power (Enhanced Data Rate)

This measurement calculates the relative power in the GFSK and PSK part of the Enhanced Data Rate (EDR) signal. This measurement is supported only when an EDR signal is detected.

Select the Enhanced Data Rate standard from Presets > Standards to perform the measurement. This will load the displays found in the [Bluetooth standards, test setups, and Preset displays table](#).

## Bluetooth Status Messages

The following status messages may appear. Each message indicates the related condition (Description) shown in the following table.

| Status message   | Description  |
|--|--|
| Bluetooth Analysis: Preamble Not Detected                              | Shown in the display and status bar when no valid Sync and Preamble are found for the chosen standard. Also shown when no preamble is found for any of the standards in Automatic Standard Detection mode. |
| Bluetooth Analysis: Low Energy Sequence Detected                       | Shown in the status bar when the Low Energy standard is detected in Automatic Standard Detection mode.   |
| Bluetooth Analysis: Basic Rate Sequence Detected                       | Shown in the status bar when the Basic Rate standard is detected in Automatic Standard Detection mode.   |
| Bluetooth Analysis: EDR Signal Detected                                | Shown in the status bar when the EDR standard is detected in Automatic Standard Detection mode.  |
| Bluetooth Analysis: EDR Sync Word Not Detected                         | Shown in the display and status bar when EDR Sync Word is not found (when EDR standard is selected).   |
| Bluetooth Analysis: Not enough payload data for Bluetooth measurements | Shown in the display and status bar when Preamble and Access Address/Sync Word is detected but there is not enough payload data for Modulation Characteristics and CF Offset and Drift measurement.        |
| Bluetooth Analysis: Meas BW may be too low                             | Shown in the InBand Emission display and in status bar when Measurement Bandwidth is too low.  |
| Bluetooth Analysis: BT x dB BW detection failed                        | Shown in the BT 20 dB BW display when x dB measurement detection fails.  |
| Bluetooth Analysis: BT x dBm BW detection failed                       | Shown in the BT 20 dB BW display when x dBm measurement detection fails.   |

## Bluetooth Constellation display

The BT Constellation display shows a digitally modulated signal in constellation form. The constellation can be viewed either as IQ or Freq Dev. The default view is Freq Dev. The frequency deviation result that is shown in the Constellation display is compensated for the frequency error (shown as readout).

To show the BT Constellation display you can select **Presets > Standards > Bluetooth** or do the following:

1. Press the **Displays** button or select **Setup > Displays**.
2. In the **Select Displays** window, select **Bluetooth Analysis** in the **Measurements** box.
3. In the Available displays box, double-click the **BT Constellation** icon or select the icon and click **Add**. The BT Constellation icon will appear in the **Selected displays** box and will no longer appear under Available displays.
4. Click the **OK** button to show the display.
5. Select **Setup > Settings** to display the control panel.
6. Select the **Standard Params** tab. Set the Standard, Power class (if required), and Measurement and Reference filters as appropriate for the input signal. You can also check the **Auto Detect Standard** box.
7. Select the **Trace** tab and set the **Trace Type** to Freq Dev or IQ, as required.
8. Select **Replay/Run** to take measurements on the acquired data.


## Elements of the Display

The following image shows the BT Constellation display of QPSK data in an EDR signal. Trace Type is set to IQ.



| Item | Display element                            | Description  |
|------|--|--|
| A    | Trace                                      | Select and enable a trace.   |
| B    | Trace Type                                 | Specifies Trace type (choice between Freq Dev and IQ). This option appears only when EDR signal is detected.   |
| C    | Freq Error / Freq Offset / marker readouts | <p>The Freq Error, Freq Offset, and marker readout values appear here or below the plot, depending on the display view and size.</p> <p>The Freq Error value is the difference between the maximum and the minimum values of signal frequency during the measurement time. The Freq Error (Auto) value is displayed when the Auto check box is selected in the <b>Analysis Params</b> tab in the Settings control panel.</p> <p>The Freq Offset (Manual) value is displayed when the Auto check box is unselected in the <b>Analysis Params</b> tab in the Settings control panel.</p> <p>When markers are enabled, marker information (including time, magnitude, phase, symbol, and value) is displayed.</p> |
| D    | Avg Dev                                    | Shows the average deviation result at every symbol point.  |
| E    | Plot                                       | Bluetooth constellation graph. Shown as either IQ or as Frequency Deviation. The trace type is controlled from the <b>Settings &gt; Trace</b> tab.   |
| F    | Show Vectors                               | Specifies whether to show or hide the selected trace.  |

The following image shows the BT Constellation display of GFSK data in a Basic Rate signal. Trace Type is set to Freq Dev.



## BT Constellation Settings

### Front Panel: Settings

#### Application Toolbar:

The BT Constellation Settings control panel provides access to settings that control parameters of the Constellation Display.

| Settings tab    | Description   |
|-----------------|---|
| Standard Params | Select the standard, measurement filters, and power class (when applicable). You can also set the analyzer to auto detect the standard.   |
| Analysis Params | Specifies frequency error, frequency offset, measurement BW, and test patterns. You can also select to auto detect test patterns. These parameters are used by the instrument to analyze the input signal.  |
| Analysis Time   | Specifies parameters that define the portion of the acquisition record that is used for analysis. Also allows you to specify the Units (Symbols or Seconds) and Time Zero Reference (Trigger or Acquisition Start) for Bluetooth Analysis displays. |
| Trace           | Allows you to select the trace to view, the trace type (IQ or Freq Dev), the number of points per symbol (when trace content is set to Vectors or Lines), and to show or freeze the selected trace.   |
| Prefs           | Allows you to select to show or hide the graticule and marker readouts.   |

## Bluetooth Eye Diagram display


The BT Eye Diagram display shows a digitally modulated signal overlapped on itself to reveal variations in the signal.

To show the BT Eye Diagram display:

1. Press the **Displays** button or select **Setup > Displays**.
2. From the **Measurements** box, select **Bluetooth Analysis**.
3. Double-click the **BT Eye Diagram** icon in the **Available Displays** box. This adds the BT Eye Diagram icon to the **Selected displays** box.
4. Click **OK** button. This displays the BT Eye Diagram view.
5. Select **Setup > Settings** to display the control panel.
6. Select the **Standard Params** tab. Set the Standard, Power class (when appropriate), and Measurement and Reference filters as appropriate for the input signal. You can also check the **Auto Detect Standard** box.
7. Select the **Trace** tab and set the **Trace Type** to Freq Dev or IQ, as required..
8. Select Replay/Run to take measurements on the acquired data.


### Elements of the Display

The following image shows the BT Eye display of a basic rate signal with Trace Type set to IQ.



| Item | Display element          | Description  |
|------|--------------------------|--|
| A    | Freq Error / Freq Offset | The Freq Error value is the difference between the maximum and the minimum values of signal frequency during the measurement time. The Freq Error (Auto) value is displayed when the Auto check box is selected in the <b>Analysis Params</b> tab in the Settings control panel.<br>The Freq Offset (Manual) value is displayed when the Auto check box is unselected in the <b>Analysis Params</b> tab in the Settings control panel. |
| B    | Top of graph             | The vertical scale is normalized with no units (for IQ) and with Hz (for Freq Dev).  |
| C    | Position                 | Specifies the value shown at the center of the graph display.  |
| D    | Bottom Readout           | Displays the value indicated by the bottom of graph.   |
| E    | Autoscale                | Adjusts the vertical and horizontal settings to provide the best display.  |
| F    | Position                 | Displays the horizontal position of the trace on the graph display.  |
| G    | Scale                    | Adjusts the span of the graph in symbols.  |
| H    | Plot                     | Displays the input signal.   |

The following image shows the BT Eye display of a Basic Rate signal with Trace Type set to Freq Dev.



## BT Eye Diagram Settings

### Front Panel: Settings

#### Application Toolbar:

The settings for the BT Eye Diagram display are shown in the following table.

**NOTE.** You might save time configuring the BT Eye Diagram display by selecting the **Standard Presets** button in the Bluetooth Eye Diagram Settings control panel. This allows you to select a preset optimized for the selected standard.

| Settings tab    | Description   |
|-----------------|---|
| Standard Params | Select the standard, measurement filters, and power class (when applicable). You can also set the analyzer to auto detect the standard.   |
| Analysis Params | Specifies frequency error, frequency offset, measurement BW, and test patterns. You can also select to auto detect test patterns. These parameters are used by the instrument to analyze the input signal.  |
| Analysis Time   | Specifies parameters that define the portion of the acquisition record that is used for analysis. Also allows you to specify the Units (Symbols or Seconds) and Time Zero Reference (Trigger or Acquisition Start) for Bluetooth Analysis displays. |
| Trace           | Allows you to select the trace to view, the trace type (IQ or Freq Dev), the number of points per symbol (when trace content is set to Vectors or Lines), and to show or freeze the selected trace.   |
| Scale           | Allows you to specify the horizontal and vertical scale settings.   |
| Prefs           | Allows you to select to show or hide the graticule and marker readouts.   |

## Bluetooth Carrier Frequency Offset and Drift display

The BT CF (Carrier Frequency) Offset and Drift display shows the frequency offset preamble, payload regions, and the drift result values are shown in this display. You can also view the values of frequency offset calculated in every ten bit interval in the payload and look for excursions.

To show the BT CF (Carrier Frequency) Offset and Drift display:

1. Press the **Displays** button or select **Setup > Displays**.
2. In the **Select Displays** dialog, select **Bluetooth Analysis** in the **Measurements** box.
3. In the Available displays box, double-click the **BT CF Offset and Drift** icon or select the icon and click **Add**. The BT CF Offset and Drift icon will appear in the **Selected displays** box and will no longer appear under Available displays.
4. Click **OK** to show the display.
5. Select **Setup > Settings** to display the control panel.
6. Select the **Standard Params** tab. Set the Standard, Power class (when appropriate), and Measurement and Reference filters as appropriate for the input signal. You can also check the **Auto Detect Standard** box.
7. Select the **Limits** tab and set the limits , as required.
8. Select **Replay/Run** to take measurements on the acquired data.

## Elements of the Display

The following image shows the frequency offset and drift values of a basic rate signal in the BT Freq Offset and Drift display.

| Limits Interval# | $f_n$<br>(-150.0kHz to +150.0kHz) | $f_n - f_0$ (-40.00kHz to +40.00kHz)<br>$f_1 - f_0$ (-20.00kHz to +20.00kHz) | $f_n - f(n-5)$<br>(-20.00kHz to +20.00kHz) |
|------------------|-----------------------------------|--|--|
| 0 (Preamble)     | 139.2751                          |  |  |
| 1                | -96.15504                         | -235.4302  |  |
| 2                | 7.167774                          | -132.1074  |  |
| 3                | -0.01093826                       | -139.2861  |  |
| 4                | 0.01212997                        | -139.263   |  |
| 5                | -0.003742599                      | -139.2789  |  |
| 6                | 0.007567215                       | -139.2676  | 96.16261                                   |
| 7                | -0.002672958                      | -139.2778  | -7.170447                                  |
| 8                | -0.007925415                      | -139.2831  | 0.003012847                                |
| 9                | -0.01031799                       | -139.2855  | -0.02244797                                |
| 10               | 0.001925278                       | -139.2732  | 0.005667677                                |
| 11               | 0.001144409                       | -139.274   | -0.006422806                               |
| 12               | -0.004858398                      | -139.28  | -0.00218544                                |
| 13               | 0.001594162                       | -139.2736  | 0.009519577                                |
| 14               | -0.001327515                      | -139.2765  | 0.008990479                                |
| 15               | 0.01260147                        | -139.2625  | 0.01067619                                 |
| 16               | 0.008859253                       | -139.2663  | 0.007714843                                |
| 17               | -0.003403854                      | -139.2785  | 0.001454544                                |
| 18               | -0.0007839203                     | -139.2759  | -0.002378082                               |
| 19               | 0.001569366                       | -139.2736  | 0.002896881                                |
| 20               | 0.007740402                       | -139.2674  | -0.004861068                               |
| 21               | -0.001542282                      | -139.2767  | -0.01040153                                |
| 22               | 0.009552765                       | -139.2656  | 0.01295662                                 |
| 23               | -0.006316376                      | -139.2815  | -0.005532456                               |
| 24               | -0.001724243                      | -139.2769  | -0.00329361                                |
| 25               | -0.01467628                       | -139.2898  | -0.02241669                                |
| 26               | -0.004405975                      | -139.2796  | -0.002863693                               |
| 27               | 0.0005382538                      | -139.2746  | -0.009014511                               |
| 28               | -0.01394348                       | -139.2891  | -0.007627105                               |

|                            |                 |                         |                 |                               |
|----------------------------|-----------------|-------------------------|-----------------|-------------------------------|
| Marker: MR                 |                 | Time: 3.133 ms          | Interval# 0     | Value: 139.2                  |
| Freq Offset Preamble:      | 139.3 Hz        | Drift $f_1 - f_0$ :     | -235.4 Hz       | Max Drift $f_1 - f_0 - f_5$ : |
| Max Freq Offset (Payload): | -127.7 Hz / 271 | Max Drift $f_1 - f_0$ : | -267.0 Hz / 271 |                               |

| Display element | Description   |
|-----------------|---|
| Column 1        | Indicates ten bit interval numbers. Zero indicates preamble region and the non-zero number indicate the ten bit intervals in payload.   |
| Column 2        | Shows the calculated frequency offset in the preamble and the ten bit intervals in the payload.   |
| Column 3        | Shows the drift of offset from the preamble and over a 50 $\mu$ s duration.   |
| Column 4        | Shows the drift of offset over a 50 $\mu$ s duration.   |
| —               | Results that do not meet the limits are shown with a red colored background, indicating failure. (Limits can be set in the Limits tab from the Settings control panel.)                       |
| Scalar results  | The scalar results show below the table. They are as follows:<br>Freq Offset Preamble<br>Max Freq Offset (Payload)<br>Drift $f_1 - f_0$<br>Max Drift $f_n - f$<br>Max Drift $f_n - f_{(n-5)}$ |

## BT CF Offset and Drift Settings

### Front Panel: Settings

#### Application Toolbar:

The settings for the BT CF Offset and Drift display are shown in the following table.

| Settings tab    | Description  |
|-----------------|--|
| Standard Params | Select the standard, measurement filters, and power class (when applicable). You can also set the analyzer to auto detect the standard.  |
| Analysis Params | Specifies frequency error, frequency offset, measurement BW, and test patterns. You can also select to auto detect test patterns. These parameters are used by the instrument to analyze the input signal.   |
| Analysis Time   | Specifies parameters that define the portion of the acquisition record that is used for analysis. Also allows you to specify the Units (Symbols or Seconds) and Time Zero Reference (Trigger or Acquisition Start) for Bluetooth Analysis displays.                    |
| Limits          | Allows you to load and define Bluetooth measurement limits for Pass/Fail comparison. You can save defined limits as a .csv file and also load previously saved .csv files. These limits also apply to the Settings control panel Limits tab of the BT Summary display. |

## Bluetooth Summary display

The BT Summary display shows a summary of all the scalar measurements done on the acquired test pattern. The summary display and contents will vary according to the selected standard.

Pass/Fail information is also provided in this display for all enabled scalar measurements. You can set limits and choose which measurement to compare for Pass/Fail from the Limits tab of the BT Summary Settings control panel. The default limits come from the performance recommendation limits given by the Standards document. The default limits can be reloaded by selecting the Bluetooth Standards Preset option.

The results presented in the BT Summary display include scalar results of modulation characteristics, frequency offset and drift and, output power measurements (these are averaged over the last 10 acquired packets). The display also shows packet related information for the current acquired packet.

To show the BT Summary display you can select **Presets > Standards > Bluetooth** or do the following:

1. Press the **Displays** button or select **Setup > Displays**.
2. In the **Select Displays** dialog, select **Bluetooth Analysis** in the **Measurements** box.
3. In the Available displays box, double-click the **BT Summary** icon or select the icon and click **Add**. The BT Summary icon will appear in the **Selected displays** box and will no longer appear under Available displays.
4. Click **OK** to show the display.
5. Select **Setup > Settings** to display the control panel.
6. Select the **Standard Params** tab. Set the Standard, Power class (when appropriate), and Measurement and Reference filters as appropriate for the input signal. You can also check the **Auto Detect Standard** box.
7. Select the **Limits** tab and set the limits , as required.
8. Select **Replay/Run** to take measurements on the acquired data.

### Elements of the display

The following image shows all average scalar results and packet information for a Basic Rate signal in the BT Summary display. Modulation characteristics, carrier frequency offset and drift, and power measurements, along with packet information, are shown. The packet information will vary depending on the chosen or detected standard. For more information about specific measurement results, see the Bluetooth measurements and test setups section [here \(see page 385\)](#).

Bluetooth Summary Clear

Standard: Basic Rate Class3

| Modulation Characteristics [10 packet-average] |                |      |
|--|----------------|------|
| $\Delta F1_{avg}$ :                            | -- Hz -- of 10 |      |
| $\Delta F2_{avg}$ :                            | 135.4 kHz      |      |
| $\Delta F2_{Max\%} \geq 115$ kHz:              | 100.0 %        | PASS |
| $\Delta F2_{avg}/\Delta F1_{avg}$ :            | --             | N/A  |

| Frequency Offset and Drift [10 packet-average] |           |      |
|--|-----------|------|
| Freq Offset (Preamble):                        | 139.3 Hz  | PASS |
| Max FreqOffset:                                | -127.7 Hz | PASS |
| Drift $f_1-f_0$ :                              | -235.4 Hz | PASS |
| Max Drift $f_1-f_0$ :                          | -267.0 Hz | PASS |
| Max Drift $f_1-f_1-s_1$ :                      | -127.7 Hz | PASS |

| Output Power [10 packet-average] |                    |      |
|----------------------------------|--------------------|------|
| Peak Power Ppk:                  | -29.72 dBm         | PASS |
| Average Power Pavg:              | -30.00 dBm 2 of 10 | PASS |

| Packet Information      |                    |
|-------------------------|--------------------|
| Packet Type             | DH5                |
| Preamble (4 bits)       | 0101               |
| Sync Word (64 bits)     | 0x4F36F2CEE85390CB |
| Packet Header (18 bits) |                    |
| LT_ADDR (3 bits)        | 010                |
| Type (4 bits)           | 1111               |
| Flow (1 bit)            | 0                  |
| ARQN (1 bit)            | 1                  |
| SEQN (1 bit)            | 0                  |
| HEC (8 bits)            | 00110001           |
| PayLoad Length          | 0101010011         |
| CRC (16 bits)           | 0x9DB0             |

| Element  | Description   |
|--|---|
| Standard   | Display of the selected standard or the detected standard (if the Auto Detect Standard setting is chosen in the Standard Params tab of the Settings control panel).   |
| Power class  | Display of the Power class in the Standard Params tab of the Settings control panel. Only available when Basic Rate is the selected or detected standard.   |
| Clear  | Click button to reset measurement. Clears all values and all results in the queue used for average computation.   |
| Modulation characteristics<br>(ten packet average) | <p>Display of a group of scalar results related to Modulation Characteristics measurements. These scalar results are only populated in the BT Summary display if preamble is detected in the acquired data. The four scalar results are:</p> <p><math>\Delta f1_{avg}</math> is calculated when analyzed test pattern (payload) is 11110000.</p> <p><math>\Delta f2_{avg}</math> and the percentage of <math>\Delta f2_{max}</math> value (greater than a given limit) are calculated only when analyzed test pattern (payload) is 10101010.</p> <p>The ratio of <math>\Delta f2_{avg}</math> and <math>\Delta f1_{avg}</math> is calculated provided both the results are available (or have been done before).</p> <p>All the results given are averaged over the last ten packets. The scalar results can be compared with pass/fail limits set in the Limits tab of the Settings control panel.</p> <p>More information can be found <a href="#">here (see page 385)</a>.</p> |
| Frequency offset and drift<br>(ten packet average) | <p>Display of a group of five scalar results related to the carrier frequency offset and drift measurement. These scalar results are only populated in the BT Summary display if preamble is detected and if the test pattern payload is a 10101010 pattern. All the results given are averaged over the last ten packets. The scalar results can be compared with pass/fail limits set in the Limits tab of the Settings control panel.</p> <p>The five scalar results are:</p> <p>Preamble offset.</p> <p>Maximum offset in Payload (calculated over 10 bit intervals).</p> <p>Drift (<math>f_1 - f_0</math>)</p> <p>Max Drift (<math>f_n - f_0</math>).</p> <p>Max Drift (<math>f_n - f_n - 5</math>).</p> <p>For more information about this measurement, see the <a href="#">Carrier frequency offset and drift (Basic Rate and Low Energy) (see page 424)</a> topic in <i>Bluetooth Measurements and test setups</i>.</p>   |
| Output power                                       | <p>Display of Average Power and Peak Power scalar results for Output Power. All the results given are averaged over the last ten packets. The scalar results can be compared with pass/fail limits set in the Limits tab of the Settings control panel.</p> <p>For Enhanced Data Rate signals, the relative power in GFSK and PSK regions is measured.</p> <p>For more information about this measurement, see the <a href="#">Output Power (Basic Rate and Low Energy) (see page 391)</a> topic in <i>Bluetooth Measurements and test setups</i>.</p>  |
| Packet information                                 | The Packet information includes the Packet type, Preamble, Synchronization Word or Access Code, Packet Header, Payload length, and CRC details. The decoded bits are shown for the current analyzed packet.   |

The following image shows all average scalar results and packet information for a Low Energy signal in the BT Summary display.

|   |                               |                                 |            |
|---|-------------------------------|---------------------------------|------------|
| <b>Standard:</b> Low Energy                           |                               | <b>Clear</b>                    |            |
| <b>Modulation Characteristics [10 packet-average]</b> |                               | <b>Packet Information</b>       |            |
| $\Delta f_{1avg}$ :                                   | -- Hz -- of 10                | <b>Packet Type</b>              | BLE_TEST   |
| $\Delta f_{2avg}$ :                                   | 231.9 kHz                     | <b>Preamble( 8 bits)</b>        | 01010101   |
| $\Delta f_{2Max\%} \geq 185$ kHz:                     | 100.0 % 1 of 10               | <b>Access Address (32 bits)</b> | 0x71764129 |
| $\Delta f_{2avg}/\Delta f_{1avg}$ :                   | --                            | <b>PDU Header (16 bits)</b>     |            |
| <b>Frequency Offset and Drift [10 packet-average]</b> |                               | <b>PDU Type (4 bits)</b>        | 0010       |
| <b>Freq Offset (Preamble):</b>                        | -19.59 Hz <b>PASS</b>         | <b>Tx Address (1 bit)</b>       | 0          |
| <b>Max FreqOffset:</b>                                | -110.5 Hz                     | <b>Rx Address (1 bit)</b>       | 0          |
| <b>Drift <math>f_1-f_0</math>:</b>                    | -51.84 Hz                     | <b>Length (6 bits)</b>          | 100101     |
| <b>Max Drift <math>f_1-f_0</math>:</b>                | -90.95 Hz                     | <b>CRC (24 bits)</b>            | 0x8CB786   |
| <b>Max Drift <math>f_1-f_1-s_1</math>:</b>            | -148.5 Hz 1 of 10 <b>PASS</b> |                                 |            |
| <b>Output Power [10 packet-average]</b>               |                               |                                 |            |
| <b>Peak Power Ppk:</b>                                | 4.192 dBm <b>PASS</b>         |                                 |            |
| <b>Average Power Pavg:</b>                            | 3.695 dBm 1 of 10 <b>PASS</b> |                                 |            |

The following image shows all average scalar results and packet information for an EDR signal in the BT Summary display.

|   |                           |                                |                    |
|---|---------------------------|--------------------------------|--------------------|
| <b>Standard:</b> EDR                                  |                           | <b>Clear</b>                   |                    |
| <b>Modulation Characteristics [10 packet-average]</b> |                           | <b>Packet Information</b>      |                    |
| $\Delta f_{avg}$ :                                    | -- Hz 0 of 10             | <b>Packet Type</b>             | EDR_2DH1           |
| <b>Frequency Offset and Drift [10 packet-average]</b> |                           | <b>Preamble (4 bits)</b>       | 0101               |
| <b>Freq Offset (Preamble):</b>                        | -- Hz <b>N/A</b>          | <b>Sync Word (64 bits)</b>     | 0x4F36F2CEE85390CB |
| <b>Max FreqOffset:</b>                                | -- Hz                     | <b>Packet Header (18 bits)</b> |                    |
| <b>Drift <math>f_1-f_0</math>:</b>                    | -- Hz                     | <b>LT_ADDR (3 bits)</b>        | 001                |
| <b>Max Drift <math>f_1-f_0</math>:</b>                | -- Hz                     | <b>Type (4 bits)</b>           | 0100               |
| <b>Max Drift <math>f_1-f_1-s_1</math>:</b>            | -- Hz -- of 10 <b>N/A</b> | <b>Flow (1 bit)</b>            | 1                  |
| <b>Output Power [10 packet-average]</b>               |                           | <b>ARQN (1 bit)</b>            | 0                  |
| <b>POWER GFSK:</b>                                    | -28.42 dBm                | <b>SEQN (1 bit)</b>            | 1                  |
| <b>POWER DPSK:</b>                                    | -28.51 dBm 1 of 10        | <b>HEC (8 bits)</b>            | 11010111           |
|   |                           | <b>PayLoad Length</b>          | 110110             |
|   |                           | <b>CRC (16 bits)</b>           | --                 |

## BT Summary Settings

### Front Panel: Settings

### Application Toolbar:

The settings for the BT Summary display are shown in the following table.

| Settings tab    | Description  |
|-----------------|--|
| Standard Params | Select the standard, measurement filters, and power class (when applicable). You can also set the analyzer to auto detect the standard.  |
| Analysis Params | Specifies frequency error, frequency offset, measurement BW, and test patterns. You can also select to auto detect test patterns. These parameters are used by the instrument to analyze the input signal.   |
| Analysis Time   | Specifies parameters that define the portion of the acquisition record that is used for analysis. Also allows you to specify the Units (Symbols or Seconds) and Time Zero Reference (Trigger or Acquisition Start) for Bluetooth Analysis displays.                      |
| Limits          | Allows you to load and define Bluetooth measurement limits for Pass/Fail comparison. You can save defined limits as a .csv file and also load previously saved .csv files. The frequency offset and drift limits set here also apply to the CF Offset and Drift display. |

## Bluetooth Symbol Table display

The BT Symbol Table display shows decoded data values for each data symbol in the analyzed signal packet. It is like the BT Constellation display except that a text table is used to display data instead of a graph.

To show the BT Symbol Table display:

1. Press the **Displays** button or select **Setup > Displays**.
2. In the **Select Displays** dialog, select **Bluetooth Analysis** in the **Measurements** box.
3. In the Available displays box, double-click the **BT Symbol Table** icon or select the icon and click **Add**. The BT Symbol Table icon will appear in the **Selected displays** box and will no longer appear under Available displays.
4. Click **OK** to show the display.
5. Select **Setup > Settings** to display the control panel.
6. Select the **Standard Params** tab. Set the Standard, Power class (when appropriate), and Measurement and Reference filters as appropriate for the input signal. You can also check the **Auto Detect Standard** box.

7. Select the **Limits** tab and set the limits , as required.
8. Select Replay/Run to take measurements on the acquired data.


### Regions of the Display

The BT Symbol Table is color coded to indicate different regions of the packet. The legend at the bottom of the symbol table lists the packet contents in their associated colors. The following image shows values for a Basic Rate signal.


| Bluetooth Symbol Table         |   |   |   |                |   |   |   |   |   |         |   |   |   |   |   |   |        |
|--------------------------------|---|---|---|----------------|---|---|---|---|---|---------|---|---|---|---|---|---|--------|
| 0                              | 1 | 0 | 1 | 0              | 1 | 1 | 0 | 1 | 0 | 0       | 1 | 1 | 0 | 0 | 0 | 0 |        |
| 16                             | 1 | 0 | 0 | 1              | 1 | 1 | 0 | 0 | 1 | 0       | 1 | 0 | 0 | 0 | 0 | 1 |        |
| 32                             | 0 | 1 | 1 | 1              | 0 | 1 | 1 | 1 | 0 | 0       | 1 | 1 | 0 | 1 | 0 | 0 |        |
| 48                             | 1 | 1 | 1 | 1              | 0 | 1 | 1 | 0 | 1 | 1       | 0 | 0 | 1 | 1 | 1 | 1 |        |
| 64                             | 0 | 0 | 1 | 0              | 0 | 0 | 0 | 0 | 0 | 0       | 0 | 0 | 0 | 1 | 1 | 1 |        |
| 80                             | 1 | 1 | 1 | 1              | 1 | 1 | 1 | 1 | 1 | 1       | 1 | 1 | 1 | 1 | 1 | 0 |        |
| 96                             | 0 | 0 | 1 | 1              | 1 | 0 | 0 | 0 | 0 | 0       | 0 | 0 | 0 | 0 | 1 | 1 |        |
| 112                            | 1 | 1 | 1 | 1              | 0 | 0 | 0 | 0 | 0 | 0       | 1 | 1 | 1 | 1 | 1 | 0 |        |
| 128                            | 0 | 1 | 0 | 1              | 0 | 1 | 0 | 0 | 0 | 0       | 1 | 1 | 1 | 1 | 0 | 0 |        |
| 144                            | 0 | 0 | 1 | 1              | 1 | 1 | 0 | 0 | 0 | 0       | 1 | 1 | 1 | 1 | 0 | 0 |        |
| 160                            | 0 | 0 | 1 | 1              | 1 | 1 | 0 | 0 | 0 | 0       | 1 | 1 | 1 | 1 | 0 | 0 |        |
| 176                            | 0 | 0 | 1 | 1              | 1 | 1 | 0 | 0 | 0 | 0       | 1 | 1 | 1 | 1 | 0 | 0 |        |
| 192                            | 0 | 0 | 1 | 1              | 1 | 1 | 0 | 0 | 0 | 0       | 1 | 1 | 1 | 1 | 0 | 0 |        |
| 208                            | 0 | 0 | 1 | 1              | 1 | 1 | 0 | 0 | 0 | 0       | 1 | 1 | 1 | 1 | 0 | 0 |        |
|                                |   |   |   |                |   |   |   |   |   |         |   |   |   |   |   |   |        |
| Preamble                       |   |   |   | Packet Header  |   |   |   |   |   | Payload |   |   |   |   |   |   |        |
| Synchronization/Access Address |   |   |   | Payload Header |   |   |   |   |   | MIC/CRC |   |   |   |   |   |   |        |
|                                |   |   |   |                |   |   |   |   |   |         |   |   |   |   |   |   |        |
| Marker:                        |   |   |   | Time:          |   |   |   |   |   | Symbol: |   |   |   |   |   |   | Value: |

| Packet standard          | Associated regions  |
|--------------------------|---|
| Basic rate               | 4 bit preamble<br>64 bit access code<br>4 bit trailer<br>54 bit packet header<br>8 or 16 bit payload header based on packet type<br>Variable length payload data<br>16 bit CRC (Cyclic Redundancy Check)  |
| Low energy               | 8 bit preamble<br>32 bit access address / synchronization word<br>16 bit payload header<br>Variable length payload data (37 bytes)<br>24 bit CRC (Cyclic Redundancy Check)  |
| Enhanced Data Rate (EDR) | 4 bit preamble<br>64 bit access code<br>4 bit trailer (if present, shown in black)<br>54 bit packet header<br>Guard region (shown as X)<br>EDR sync sequence (QPSK or 8PSK symbols)<br>16 bit payload header (QPSK or 8PSK symbols)<br>Variable length payload data (QPSK or 8PSK symbols)<br>16 bit CRC (Cyclic Redundancy Check) (QPSK or 8PSK symbols) |

The following image shows values for a Low Energy signal.



The following image shows values for an EDR signal.



**Markers and the BT Symbol Table.** Markers are indicators in the display that you can position on a trace to measure values such as frequency, power, and time. A Marker always displays its position and, if the Delta Marker (M1-M4) readout is enabled, will display the difference between its position and that of the Marker Reference (MR). In the BT Symbol Table, colored cells indicate the location of markers. The selected Marker is highlighted with a light green background. All other markers are highlighted with a light gray background. In the BT Symbol Table, the marker readout below the table shows the marker location in time, symbol numbers and symbol value.

# BT Symbol Table Settings

## Front Panel: Settings

### Application Toolbar:

The settings for the BT Symbol Table display are shown in the following table.

| Settings tab    | Description   |
|-----------------|---|
| Standard Params | Select the standard, measurement filters, and power class (when applicable). You can also set the analyzer to auto detect the standard.   |
| Analysis Params | Specifies frequency error, frequency offset, measurement BW, and test patterns. You can also select to auto detect test patterns. These parameters are used by the instrument to analyze the input signal.  |
| Analysis Time   | Specifies parameters that define the portion of the acquisition record that is used for analysis. Also allows you to specify the Units (Symbols or Seconds) and Time Zero Reference (Trigger or Acquisition Start) for Bluetooth Analysis displays. |
| Prefs           | Allows you to select to show or hide the marker readouts and set the radix of shown symbols.  |

## Bluetooth Frequency Dev vs Time display


The BT Freq Dev vs. Time display shows how the signal frequency varies with time.

To display the BT Freq Dev vs. Time display:

1. Select the **Displays** button or **Setup > Displays**.
2. In the **Select Displays** dialog, select **Bluetooth Analysis** in the **Measurements** box.
3. In the **Available displays** box, double-click the **BT Freq Dev vs. Time** icon or select the icon and click **Add**. The BT Freq Dev vs. Time icon will appear in the **Selected displays** box and will no longer appear under Available displays.
4. Click **OK** to show the display.
5. Select the **Standard Params** tab. Set the Standard, Power class (when appropriate), and Measurement and Reference filters as appropriate for the input signal. You can also check the **Auto Detect Standard** box.
6. Select the **Trace** tab and set the **Content** to Vectors or Points, as required.
7. Select **Replay/Run** to take measurements on the acquired data.

### Elements of the Display


The following image shows the frequency deviation of the complete packet of a Basic Rate signal in the BT Freq Dev vs. Time display.




| Item | Display element       | Description  |
|------|-----------------------|--|
| A    | Top of graph          | The vertical scale is normalized with Hz.  |
| B    | View                  | (This option is visible only when Preamble is detected.)<br>Select one of the following:<br>Packet: view frequency deviation for an entire packet.<br>Octet: view frequency deviation for a specified octet duration (8 $\mu$ s). The Octet is specified in the Octet # field.   |
| C    | Octet # (xx of total) | (This option is visible only when Preamble is detected and Octet is selected as the View.)<br>Specifies the Octet number. (The total number of octets is also indicated.)<br>When View is set to Octet, you can enter a particular octet in the Octet # field for zoom view. The plot will show only 8 $\mu$ s of information corresponding to the octet number chosen in the payload. The regions that are used for doing the exact measurements are highlighted in the Octet View. For F0F0 (Low Deviation) pattern in payload, bit intervals 2,3,6,7 regions are highlighted and for 1010 (high deviation) pattern, all bit regions are highlighted with appropriate results. The offset calculated for every octet region (f1ccf or f2ccf) are also shown. |
| D    | Position (center)     | Adjust the frequency shown at the center of the display.   |
| F    | Autoscale button      | Adjusts the offset and range for both vertical and horizontal to provide the best display.   |
| G    | Position              | Displays the horizontal position of the trace on the graph display.  |
| H    | Scale                 | Adjusts the horizontal scale (time).   |
| I    | Plot                  | Displays the last analyzed complete packet or the selected octet (Octet zoom view) of the signal.  |
| J    | Scalar results        | The Scalar results for Modulation Characteristics and Frequency Offset and Drift measurements are captured here. These results are from the last analyzed packet and therefore could be different from the Averaged Scalar results shown in the BT Summary display.<br>For Drift results, the index is also shown where the maximum drift occurred. The Drift results are shown only when the preamble is detected and the test pattern payload detected is 10101010. $\Delta F_{1avg}$ or $\Delta F_{2avg}$ is shown only when the detected test pattern payload is (10101010) or (11110000).   |

**NOTE.** This display shows the frequency deviation results without compensating for the frequency error.

The following image shows the frequency deviation of the first octet in the payload of a Basic Rate signal (high deviation).



The following image shows the frequency deviation of the first octet in the payload of a Basic Rate signal (low deviation).



## BT Frequency Dev Vs Time Settings

**Menu Bar:** Setup > Settings

**Front Panel:** Settings

**Application Toolbar:** 

The Setup settings for BT Freq Dev vs. Time are shown in the following table.

| Settings tab    | Description   |
|-----------------|---|
| Standard Params | Select the standard, measurement filters, and power class (when applicable). You can also set the analyzer to auto detect the standard.   |
| Analysis Params | Specifies frequency error, frequency offset, measurement BW, and test patterns. You can also select to auto detect test patterns. These parameters are used by the instrument to analyze the input signal.  |
| Analysis Time   | Specifies parameters that define the portion of the acquisition record that is used for analysis. Also allows you to specify the Units (Symbols or Seconds) and Time Zero Reference (Trigger or Acquisition Start) for Bluetooth Analysis displays. |
| Trace           | Allows you to select the number of points per symbol, content, and to choose between entire packet view or zoom Octet view for a chosen octet number in a packet.   |
| Scale           | Allows you to specify the horizontal and vertical scale settings.   |
| Prefs           | Allows you to select to show or hide the graticule and marker readouts.   |

## Bluetooth 20dB Bandwidth display

The BT 20dB BW display shows the results of two Bluetooth measurements: 20dB Bandwidth and Frequency Range. When the xdB BW is chosen, the display shows the x dB bandwidth from the peak power. More detailed information about this measurement is available [here \(see page 385\)](#).


To display the BT 20dB BW display:

1. Select the **Displays** button or **Setup > Displays**.
2. In the **Select Displays** dialog, select **Bluetooth Analysis** in the **Measurements** box.
3. In the **Available displays** box, double-click the **BT 20dB BW** icon or select the icon and click **Add**. The BT 20dB BW icon will appear in the **Selected displays** box and will no longer appear under Available displays.
4. Click **OK** to show the display.
5. Select the **Standard Params** tab. Set the Standard, Power class (when appropriate), and Measurement and Reference filters as appropriate for the input signal. You can also check the **Auto Detect Standard** box.

6. Select the **Trace** tab and set the **Content** to Vectors or Points, as required.
7. Select Replay/Run to take measurements on the acquired data.

### Elements of the Display

The following image of the BT 20dB BW display shows a Basic Rate signal that would allow you to measure 20 dB bandwidth.



| Item | Display element     | Description   |
|------|---------------------|---|
| A    | PASS / FAIL         | Indicates Pass or Fail for the 20 dB BW measurement.  |
| B    | Autoscale           | Adjusts the vertical and horizontal settings so that the entire trace fits in the graph.  |
| C    | Normal / MaxHold    | Indicates whether the measurement is done with a MaxHold or a Normal condition.   |
| D    | Clear               | Resets count for Average and MaxHold functions. Enabled only when Averaging or MaxHold is enabled. Pressing Clear will clear the trace and, if acquisition is running, restart the averaging or hold process.   |
| E    | Main results area   | Shows the xdBm and xdB results. The requested dB or dBm value can be set in the controls below the result readout. Use the two radio buttons to select which of the two results are illustrated in the graph with the blue lines and arrows. RBW also can be set using the control. VBW enables the VBW (Video Bandwidth) filter. Displays current VBW filter setting. See Setup > Settings > Freq & RBW tab. |
| F    | Position and dB/div | Position sets the top of graph value. The dB/div setting is the vertical scale value.   |
| G    | Bottom readout      | Displays the value indicated by the bottom of graph.  |
| H    | Position            | Displays the horizontal position of the trace on the graph display.   |
| I    | Scale               | Adjust the span of the graph in symbols.  |
| J    | Detailed results    | Displays the following additional measurements results:<br><br><b>Freq Error:</b> The frequency difference between the measured carrier frequency of the signal and the user-selected center frequency of the analyzer.<br><br><b>x dB BW Ref Power:</b> The peak power measured within the measurement bandwidth.  |
| K    | Grid divider        | Determines the portion of the display allocated to the graph and detailed results area. You can move the grid divider all the way to the top or bottom and any position in between.   |
| L    | Blue lines          | Indicates the measurement positions.  |
| M    | Plot                | Displays the input signal. Shaded areas indicate the measurement bandwidth (Settings > Parameters tab > Measurement BW).  |

## BT 20dB BW settings

**Menu Bar:** Setup > Settings

**Front Panel:** Settings

**Application Toolbar:** 

The Setup settings for BT 20dB BW are shown in the following table.

| Settings tab | Description   |
|--------------|---|
| Freq & RBW   | Allows you to specify the frequency, resolution bandwidth (RBW), step, and VBW used for the MCPR measurements.  |
| Parameters   | Allows you to specify the x dB level, Measurement Direction, Measurement BW, xdBm level, xdBm Range, and to enable averaging and the Max Hold function. |
| Scale        | Allows you to define the vertical and horizontal axes.  |
| Prefs        | Allows you to specify the radix of the marker readout and whether elements of the graphs are displayed.   |

## Bluetooth InBand Emission display


The BT InBand Emission display verifies whether the emissions inside the operating frequency range are within limits. The power in adjacent channels of 1 MHz bandwidth around the frequency of transmission are calculated and compared against limits. The integrated power in the 1 MHz band for the adjacent channels is shown in a table below the plot.

To display the BT InBand Emission display:

1. Select the **Displays** button or **Setup > Displays**.
2. In the **Select Displays** dialog, select **BT Inband Emission** in the **Measurements** box.
3. In the **Available displays** box, double-click the **BT Inband Emission** icon or select the icon and click **Add**. The BT Inband Emission icon will appear in the **Selected displays** box and will no longer appear under Available displays.
4. Click **OK** to show the display.
5. Select the **Standard Params** tab and select the appropriate standard.
6. Select the **Measurement Params** tab and turn on averaging, if desired.
7. Select the **Limits** tab and set the desired limits.
8. Select **Replay/Run** to take measurements on the acquired data.

### Elements of the Display

The following image shows the BT Inband Emission display of a Basic Rate signal operating at 2.441 GHz (showing 80 adjacent bands).



| Item | Display element | Description   |
|------|-----------------|---|
| A    | PASS / FAIL     | Indicates Pass or Fail based on the number of adjacent channels that are below an upper limit. A maximum of 3 adjacent channels can have integrated power higher than an upper limit (Pass); otherwise, Fail is reported.   |
| B    | Normal          | Displays Average Count if Average is turned on (set to Time Domain or Frequency Domain) from the Measurement Params tab in the Settings control panel.  |
| C    | Clear           | Resets measurement. Clears all values.  |
| D    | Top of graph    | The vertical scale is normalized with appropriate power units.  |
| E    | dB/div          | The vertical scale value.   |
| F    | Bottom readout  | Displays the value indicated by the bottom of graph.  |
| G    | Autoscale       | Adjusts the vertical and horizontal settings so that the entire trace fits in the graph.  |
| H    | Position        | Displays the horizontal position of the trace on the graph display.   |
| I    | Scale           | Adjust the span of the graph in symbols.  |
| J    | Plot            | Divides the spectrum into different adjacent channels (each of 1 MHz bandwidth) as suggested in the standard document. You can select to show integrated power by checking the <i>Show power levels in graph</i> box in the Prefs tab in the Settings control panel). The integrated power level is shown in blue and the prescribed limits for comparison (set in the Limits tab of the Settings control panel) are shown in white. The region around the frequency of transmission is shown in a different color. |
| K    | Results table   | Reports the integrated power results (Channel frequency, Integrated power, and Limits) from the display in a table.   |

## Bluetooth Analysis Measurement Settings

### Front Panel: Settings

#### Application Toolbar:

The control panel tabs in this section are shared between the displays in Bluetooth Analysis (Setup > Displays). Some tabs are shared by all the displays, some tabs are shared by only a subset of displays. The settings available on some tabs change depending on the selected display.

### Common controls for Bluetooth Analysis displays

| Settings tab                                      | Description  |
|---|--|
| <a href="#">Standard Params (see page 455)</a>    | Specifies the input signal standard and additional user-settable signal parameters.  |
| <a href="#">Analysis Params (see page 457)</a>    | Specifies parameters used by the application to analyze the input signal.  |
| <a href="#">Analysis Time (see page 458)</a>      | Specifies parameters that define the portion of the acquisition record that is used for analysis. Also allows you to specify the Units (Symbols or Seconds) for Bluetooth Analysis displays. |
| <a href="#">Trace (see page 461)</a>              | Specifies the trace type (Vectors, Points, and Lines), the number of points per symbol when chosen to view as Vectors or Lines, and to control the view as Freq Dev or IQ.                   |
| <a href="#">Prefs (see page 463)</a>              | Allows preferences with Radix display and marker readouts.   |
| <a href="#">Scale (see page 460)</a>              | Defines the vertical and horizontal axes.  |
| <a href="#">Parameters (see page 464)</a>         | Specifies parameters used to analyze the signal.   |
| <a href="#">Freq RBW (see page 466)</a>           | Allows you to set Frequency and RBW settings for the BT 20 dB BW display.  |
| <a href="#">Limits (see page 459)</a>             | Allows you to define limits for pass/fail comparison with calculated values. The default values are as recommended in the test specification.  |
| <a href="#">Measurement Params (see page 467)</a> | Allows you to set average (Time Domain, Off, or Frequency Domain) and to correct for noise floor.  |

## Standard Params tab - BT

The Standard Params tab allows you to specify the standard, measurement and reference filters, power class (for some standards), and set standard auto detect.

The following image shows the tab for all of the Bluetooth displays except for the BT InBand Emission display. The Power Class menu only appears when Basic Rate is the selected standard.

The screenshot shows the 'Standard Params' tab selected in a software interface. The settings are as follows:

- Standard:** Basic Rate (dropdown menu)
- Measurement Filter:** None (dropdown menu)
- Channel Bandwidth:** 1MHz
- Reference Filter:** None (dropdown menu)
- Power Class:** 1 (dropdown menu)

The following image shows the tab for the BT InBand Emission display.

|                                   |                 |               |       |       |
|-----------------------------------|-----------------|---------------|-------|-------|
| Standard Params                   | Analysis Params | Analysis Time | Trace | Prefs |
| Standard: <span>Basic Rate</span> |                 |               |       |       |

| Settings                                       | Description   |
|--|---|
| Standard                                       | Select the appropriate standard: Basic Rate, Low Energy, or Enhanced Data Rate.   |
| Auto Detect Standard                           | Check box to enable automatic detection of the Bluetooth standard (Basic Rate, Low Energy, or Enhanced Data Rate). A status message will appear indicating which of the three standards is detected and the result will show as Standard selection. |
| Measurement Filter                             | Specifies the filter used for measurement. The default filters for the Low Energy and Basic Rate standards are LE-Recommended and BR-Recommended, respectively.   |
| Reference Filter                               | Specifies the filter used as a reference.   |
| Power Class<br>(only available for Basic Rate) | Select one out of three available power classes. This sets the default limits for comparison (set in the Limits tab of the Settings control panel) for Average and Peak power measurements.   |

### Power class for the Basic Rate standard

The power class for Basic Rate is the reference receive power range as specified by the standard. It sets the default limits for comparison in the Limits tab of the control panel for Average and Peak power measurements. The power classes are specified as follows:

**Class 1:** max power 20 dBm (100 mW) with mandatory power control.

**Class 2:** max power 6 dBm (4 mW) with optional power control.

**Class 3:** max power 0 dBm (1 mW) with optional power control.

### Recommended measurement and reference filters

The available measurement and reference filters depend on the selected standard. You can use the filters recommended by the standard (shown in following table) or load your own filters by selecting one of the User defined filters from a file.

The measurement filter is applied before the demodulation bit is detected and the ideal reference is calculated.

The reference filter is applied to the internally generated ideal reference signal.



**CAUTION.** Although there may be other filter types listed in the drop down menu, if you select any filter other than that which is recommended, the measurement results may not be accurate.

| Standard           | Measurement filters | Reference filters |
|--------------------|---------------------|-------------------|
| Basic Rate         | BR-Recommended      | None              |
| Low Energy         | LE-Recommended      | None              |
| Enhanced Data Rate | BR-Recommended      | None              |

### How to select filters

In a signal transmitter/receiver system, the baseband signal might be filtered for bandwidth limiting or for another kind of necessary shaping that needs to be applied. Normally, a filter in the transmitter (Ft) and a filter in the receiver (Fr) are applied.

The Measurement Filter setting in the analyzer corresponds to the baseband filter in the receiver (Fr): This setting tells the analyzer what filter your receiver uses. When the analyzer is set to the same filter used by the receiver, the analyzer sees the signal as your receiver would. The Measurement Filter setting should be the same as the filter used in the receiver under normal operation (as opposed to testing).

## Analysis Params tab - Bluetooth

The Analysis Params tab contains parameters that control the analysis of the input signal. The Test Patterns menu only appears when the Auto Detect Test Pattern is not checked. This tab is available for all displays except for the BT 20dB BW and the BT InBand Emission displays.

The screenshot shows the 'Analysis Params' tab with the following settings:

- Frequency Offset:** 552.5 Hz, with an 'Auto' checkbox.
- Measurement BW:** 6.401 MHz, with a dropdown menu set to 'Auto'.
- Auto detect Test Pattern:** An unchecked checkbox.
- Test Patterns:** A dropdown menu showing 'Other' as the selected option, with a list of other patterns: 'High Deviation (01010101)', 'Low Deviation (00001111)', and 'Other'.

| Settings                 | Description  |
|--------------------------|--|
| Frequency Error          | When the Auto box is checked, the analysis determines the Frequency Error and the measured Frequency Error is displayed. When the Auto box is unchecked, the entered value is used by the analysis as a fixed frequency offset. This is useful when the exact frequency offset of the signal is known. |
| Measurement BW           | Specifies the bandwidth about the center frequency at which measurements are made. Select Manual, Auto, or Link to Span.   |
| Auto Detect Test Pattern | When this box is checked, automatic detection of the test pattern is enabled. When this box is unchecked (Manual mode), automatic detection is disabled and you can select the test pattern and specify the $\Delta f_{avg}$ for the other pattern for ratio computation.                              |
| Test Pattern             | Allows you to select the test pattern to be analyzed. This choice is available only when Auto Detect Test Pattern is disabled.   |

## Analysis Time tab - Bluetooth

The Analysis Time tab contains parameters that define how the signal is analyzed in the Bluetooth Analysis displays. This tab is available for all displays except for the BT 20dB BW and the BT InBand Emission displays.

The screenshot shows the 'Analysis Time' tab with the following settings:

- Analysis Offset:** 0.000 s, ☒ Auto
- Time Zero Reference:** Trigger (dropdown)
- Analysis Length:** 3.500 ms, ☐ Auto
- Units:** Seconds (dropdown)
- Available:** 0.000 s

| Settings            | Description  |
|---------------------|--|
| Analysis Offset     | Specifies the location of the first time sample to use in measurements.  |
| Auto                | When enabled, causes the instrument to set the Analysis Offset value based on the requirements of the selected display.  |
| Analysis Length     | Specifies the length of the analysis period to use in measurements. Length is specified in either symbols or seconds, depending on the Units setting.                                      |
| Auto                | When enabled, causes the instrument to set the Analysis Length value based on the requirements of the selected display.  |
| Available           | This is a displayed value, not a setting. It is the Analysis Length (time or symbols) being used by the analyzer. This value may not match the Analysis Length requested (in manual mode). |
| Time Zero Reference | Specifies the zero point for the analysis time.  |
| Units               | Specifies the units of the Analysis Length to either Symbols or Seconds.   |

### Analysis Offset

Use analysis offset to specify where measurements begin. Be aware that you cannot set the Analysis Offset outside the range of time covered by the current acquisition data. (All time values are relative to the Time Zero Reference.)

You can set the Analysis Length so that the requested analysis period falls partly or entirely outside the current range of acquisition data settings. When the next acquisition is taken, its Acquisition Length will be increased to cover the new Analysis Length, as long as the Sampling controls are set to Auto. If the Sampling parameters are set to manual, or if the instrument is analyzing saved data, the actual analysis length will be constrained by the available data length, but in most cases, measurements are able to be made anyway. The instrument will display a notification when measurement results are computed from less data than requested. Range: 0 to [(end of acquisition) - Analysis Length]. Resolution: 1 effective sample (or symbol).

## Analysis Length

Use the analysis length to specify how long a period of time is analyzed. As you adjust this value, the actual amount of time for Analysis Length, in Symbol or Seconds units, is shown below the control in the "Available" readout. This setting is not available when Auto is checked. Range: minimum value depends on the standard. Resolution: 1 symbol.

## Time Zero Reference

All time values are measured from this point (such as marker position or horizontal position (in Y vs Time displays). Choices are: Acquisition Start or Trigger.


| Parameter         | Description  |
|-------------------|--|
| Acquisition Start | Time zero starts from the point at which the acquisition begins. |
| Trigger           | Time zero starts from the trigger point.                         |

## Limits tab - Bluetooth


This tab is only available for the BT CF Offset and Drift, BT InBand Emission, and BT Summary displays. It enables you to load an existing limits table, save a limits table, or edit limits values. The content under Measurement Name varies based on the chosen standard and power class.

Some measurements are done only when a specific test pattern is detected. If the specific pattern is not detected, then N/A appears in blue in the BT Summary display because the measurement is not done. If the measurement is done, Pass or Fail is shown in the BT Summary display.

The following image shows the Limits tab for the BT CF Offset and Drift and BT Summary displays.



The following image shows the Limits tab for the BT InBand Emission display.



| Settings | Description  |
|----------|--|
| Load     | Click to load a saved Limits table from a .csv file.   |
| Save     | Click to save the current Limits table to a .csv file. |

### Edit limits


To directly edit measurement limits in the table, click on the value in the Limit column that you want to change. The following table describes the parameters that are set in the Limits Table.

| Settings         | Description  |
|------------------|--|
| On               | Click on the cell in the On column next to the measurement to specify whether or not measurements are selected for limit comparison to indicate Pass or Fail. A check mark means the measurement will be taken. An empty box means it will not be taken. |
| Measurement Name | Specifies the name of the measurement related to the limit. (Not editable.) The content under Measurement Name varies based on the chosen standard and power class.  |
| Limit            | Specifies the value of the limit to the related measurement. When the cell is selected, the value is shown along with the units.   |


## Scale tab - Bluetooth

The Scale tab allows you to change the scale settings that control how the trace appears on the display but does not change control settings such as Measurement Frequency. In effect, these controls operate like pan and zoom controls. The Scale tab values are unique to each display. Also, note that each display uses horizontal and vertical units that are appropriate for that display. There are three versions of the Scale tab for Bluetooth displays.


The following image shows the Scale tab for the BT 20dB BW and BT InBand Emission displays.



The following image shows the Scale tab for the BT Eye Diagram display.



The following image shows the Scale tab for the BT Freq Dev vs Time display.



| Settings    | Description   |
|-------------|---|
| Vertical    | Controls the vertical position and scale of the trace display.  |
| Scale       | Changes the vertical scale of the graph.  |
| Position    | Adjusts the reference level away from top of the graph.   |
| Autoscale   | Resets the scale of the vertical axis to contain the complete trace.  |
| Horizontal  | Controls the horizontal span of the trace display and position of the trace.  |
| Scale       | Allows you to, in effect, change the horizontal span.   |
| Position    | Allows you to pan a zoomed trace without changing the frequency.  |
| Autoscale   | Resets the scale of the horizontal axis to contain the complete trace.  |
| Auto        | When Auto is checked, the scale and position values for the Symbols graph are automatically adjusted to maintain the optimal display. |
| Reset Scale | Resets all settings to their default values.  |

**NOTE.** The Units used for the horizontal scale for the BT Freq Dev vs Time display can be either Seconds or Symbols. To set the units for the horizontal scale, display the Analysis Time tab. On the tab, select the appropriate units from the Units drop-down list.

## Trace tab - Bluetooth

The Trace tab allows you to set the trace display characteristics of the BT Constellation, BT Eye Diagram, and BT Freq Dev vs Time displays. The selections vary depending on the selected display.

The following image shows the tab for the BT Constellation display.

Standard Params Analysis Params Analysis Time **Trace** Prefs

Trace: Trace1 ☒ Show ☐ Freeze

Points/Symbol: 4 Trace Type: Freq Dev

Content: Vectors

The following image shows the tab for the BT Eye Diagram display.

Standard Params Analysis Params Analysis Time Trace **Scale** Prefs

Points/Symbol: 4 Trace Type: Freq Dev

The following image shows the tab for the BT Freq Dev vs Time display.

Standard Params Analysis Params Analysis Time **Trace** Scale Prefs

Points/symbol: 4


Content: Vectors

| Setting         | Description   |
|-----------------|---|
| Trace           | Select the trace to display.  |
| Show            | Shows / hides the selected by trace.  |
| Freeze          | Halts updates to the selected trace.  |
| Points/Symbol   | Select how many points to use between symbols when connecting the dots. Values: 1, 2, 4, 8, 16, and 32.   |
| Content         | Select whether to display the trace as vectors (points connected by lines), points (symbols only without lines), or lines (lines drawn between symbols, but no symbols are displayed). The choices available depend on the display. |
| Trace Type      | Select to specify whether the plot is shown as IQ or as Frequency Deviation.  |
| View            | Allows you to see the full packet or only the chosen Octet.   |
| Octet # (of xx) | Allows you to view the specified octet. You can also select from the list of the available number of Payload octets in the packet.  |

## Comparing two traces in the BT Constellation display

You can use the Traces tab to enable the display of a second trace. The second trace is a version of the current acquisition. You can choose to freeze a trace in order to display the current live trace to an earlier version of itself, you can display the trace as a second trace, or you can choose to display both traces frozen in order to compare the trace to itself at different times.

To display a second trace in the BT Constellation display:


1. If more than one display is present, select the BT Constellation display to ensure it is the selected display.
2. Click .
3. Select the Trace tab.
4. Select Trace 2 from the Trace drop-down list.
5. Click the Show check box so that it is checked.
6. Specify the Content as desired.

The Trace 2 lines will appear in blue to help you distinguish Trace 2 from Trace 1.


## Prefs tab - Bluetooth

The Prefs tab enables you to change appearance characteristics of some of the Bluetooth Analysis displays.

The following image shows the Prefs tab for the BT 20dB BW, BT Freq Dev vs Time, BT Eye Diagram, and BT Constellation displays.



The following image shows the Prefs tab for the BT InBand Emission display.



The following image shows the Prefs tab for the BT Symbol Table display.

| Setting  | Description  |
|--|--|
| Show graticule                                 | Shows or hides the graticule.  |
| Show Marker readout in graph (selected marker) | Shows or hides the readout for the selected marker in the graph area.  |
| Radix  | Specifies how symbols are displayed in the Symbol Table display and in the Marker readout in the display. There are three choices for Radix: Binary (0,1) and Modulation Symbols (+1,-1). When EDR is detected, the symbol table will always show results in Hex for the PSK region (after guard). |
| Trace detection                                | <b>+Peak:</b> Shows the peak power in a bin (of chosen RBW) if there are multiple points to choose from within a bin.<br><b>Avg (VRMS):</b> Shows the average power in a bin (of chosen RBW) if there are multiple points to choose from within a bin.   |
| Show power levels graph                        | Displays the calculated power levels in graph in each band.  |

## Parameters tab - Bluetooth

The Parameters tab enables you to specify several parameters that control signal acquisition in the BT 20dB BW display.

| Setting                | Description   |
|------------------------|---|
| x dB level             | x dB level defines the x dB BW level search threshold.  |
| Meas Direction         | Specifies which way the search for the x dB level is done. Selecting <b>Inwards</b> directs the search for x dB from the edges towards maximum power. Selecting <b>Outwards</b> directs the search for x dB from maximum power (x dB ref power) to edges. |
| x dBm level            | x dBm level defines the x dBm BW level search threshold.  |
| x dBm range            | Specifies the search direction for the x dBm level. The choices are Lower, Higher, or Both.   |
| Measurement BW         | Specifies the frequency range used by the measurement.  |
| Max Hold spectral data | Enables the Max Hold function.  |
| Average results        | Enables/disables results average across acquisitions. (This is averaging of the results, not of the trace.)   |
| Count                  | Specifies the number of results averaged to calculate the Occupied BW. Range: 2 to 10,000.  |

### x dB level

The x dB level determines the x dB bandwidth. The analyzer analyses the spectrum trace to locate the frequencies at which the level is x dB down from the peak level, calculated over the measurement bandwidth. The frequency difference between the upper and lower crossing thresholds is the x dB BW.

**Range:** -80.0 to -1.0 dB; **Resolution:** 0.1%; Inc/dec small: 0.1%, large: 1%; **Default:** -20 dB

### Meas Direction

The search for the x dB level or x dBm level can be done by selecting Inwards or Outwards in Meas Direction. Selecting Inwards directs the search from the edges towards maximum power. Selecting Outwards directs the search from maximum power (x dB ref power or CF for x dBm) to edges.

### x dBm level

The x dBm level determines the x dBm bandwidth. The analyzer analyzes the spectrum trace to locate the frequencies at which the level is x dBm down from the Center frequency. The frequency range is calculated based on the choice of x dBm Range. The value of x dBm is set to -30 dBm when the BT 20dB BW display is launched from the Bluetooth Standards Presets.

### Meas Range (Higher, Lower, Both)

This determines the search range for the x dB level. The options are as follows:

- **Lower:** Indicates the frequency range from center to the lower frequency at which the power level drops to x dBm.
- **Higher:** Indicates the frequency range from center to the higher frequency at which power level drops to x dBm.
- **Both:** Indicates the frequency between the upper and the lower crossing thresholds at which the power level drops to x dBm.

Max Hold spectral data

Max Hold displays the maximum value in the acquisition record for each display point. Each new trace display point is compared to the previous maximum value and the greater value is retained for display and subsequent comparisons.

Freq & RBW tab - Bluetooth

The Freq & RBW tab allows you to specify the bandwidth parameters used for setting measurement bandwidth in the BT 20dB BW display. This determines what acquisition bandwidth the measurement will request.

Freq & RBW

Parameters

Scale

Prefs

Meas Freq: 2.44100 GHz

RBW: 30 kHz

☒ Auto

Step: 1.0000 MHz


☒ Auto

☐ VBW:

| Setting   | Description  |
|-----------|--|
| Meas Freq | Specifies the measurement frequency.   |
| RBW       | Select Auto or Manual. Adjusts the resolution bandwidth for the entire measurement. This setting is independent of the Spectrum display's RBW setting. |
| Step      | Sets the increment/decrement size for the adjustment of the center frequency. If Auto is enabled, the analyzer will adjust the step size as required.  |
| VBW       | Adjusts the video bandwidth.   |

## Measurement Params tab - Bluetooth

The Measurement Params tab is only available for the BT InBand Emission display. It allows you to select the average domain and if you want to correct for noise floor.



The screenshot shows a software interface with a tabbed menu at the top. The tabs are: Standard Params, Freq & RBW, Measurement Params (which is the active tab), Scale, Prefs, and Limits. Below the tabs, there are three settings: 'Average:' with a dropdown menu showing 'Frequency Domain', 'Number:' with a text input field containing '10', and a checkbox labeled 'Correct for Noise Floor' which is currently unchecked.

| Setting                 | Description  |
|-------------------------|--|
| Average                 | Specify the average domain (Time Domain, Frequency Domain) or set Average to Off. If a domain is selected, then you can also specify the number. |
| Number                  | Specify the number value associated with the Average setting.  |
| Correct for Noise Floor | Check the box next to this setting to correct for noise floor. This is disabled (unchecked) by default.  |



## Overview

Audio Analysis measures basic time- and frequency-domain parameters of analog audio signals modulated on a carrier (AM, FM and PM modulation) or unmodulated (non-carrier) audio signals (Direct).

For modulated signals, the measurement analysis first demodulates the signal to provide the *Audio signal* waveform. Direct input signals bypass the demodulation step. For FM and PM demodulation, the carrier frequency error is estimated during demodulation.

The Audio signal waveform excursions are then measured to determine the Peak and RMS waveform parameters. Next, the analysis detects the highest-amplitude frequency component within the audio bandwidth, and makes a high-accuracy frequency measurement of the frequency component. This value is called the *Audio Frequency*.

A spectral analysis of the Audio signal waveform is performed to determine the presence and level of harmonically- and/or non-harmonically-related narrowband spurs and wideband noise. The Audio signal, harmonic and non-harmonic spurs, and noise level data are combined to produce signal summary parameters including SINAD, Modulation Distortion, Signal-to-Noise, Total Harmonic Distortion, and Total Non-Harmonic Distortion.

Controls are provided to allow the user to select audio filters of Low Pass, High Pass, FM De-emphasis, or Standard-defined response, as well as completely user-definable filter response. Filtering can be applied as needed to modify the audio spectrum result before measurement to remove unwanted spurs or noise.

Flexible control parameters are provided to allow setting the Audio Bandwidth for analysis, the Resolution Bandwidth (RBW) and RBW filter type of the spectral analysis, and number and level qualifications for Harmonic and Non-harmonic spur detection. Multiple-spectrum averaging can be enabled to provide a smoothed spectrum for results with less variability than single-spectrum results.

The Audio Spectrum display shows the frequency spectrum waveform with detected harmonic and non-harmonic components identified by markers, and a corresponding table of frequency and level values for the spur components. The markers and table provide easy visualization of the significant spurs and their relation to the fundamental Audio frequency signal.

## Audio Spectrum Display


The Audio Spectrum display shows audio modulation characteristics. You can choose to show just the spectrum of the audio signal or show the audio spectrum of the signal and the results of distortion measurements. The Audio Spectrum display can show a table listing the frequency of a Harmonic Distortion (HD) and Non-Harmonic Distortion (NHD) and its level. The Spectrum graph indicates these harmonics and non-harmonics with special markers.

To display the Audio Spectrum display:

1. Press the **Displays** button or select **Setup > Displays**. This shows the **Select Displays** dialog box.
2. From the **Measurements** box, select **Audio Analysis**.

- 3. Double-click the **Audio Spectrum** icon in the **Available Displays** box. This adds the Audio Spectrum icon to the **Selected displays** box.
- 4. Click the **OK** button. This shows the Audio Spectrum display.

Elements of the Audio Spectrum Display



| Item | Display element                  | Description  |
|------|----------------------------------|--|
| 1    | Vertical position                | Sets the top of graph value.   |
| 2    | dB/div                           | Sets the vertical scale value. The maximum value is 20.00 dB/division.   |
| 3    | Audio BW                         | Specifies the measurement bandwidth of the Audio Spectrum display, which in turn can influence the acquisition bandwidth.                                      |
| 4    | Table                            | Displays a table that shows the distortion measurement results and displays indicators on the graph that highlight the location of the harmonics on the trace. |
| 5    | Bottom of graph readout          | Displays the bottom of graph value.  |
| 6    | Non-harmonic threshold indicator | Displays the threshold for detecting non-harmonic components.  |
| 7    | Analysis results                 | Display of the audio analysis results.   |

Audio Spectrum Settings

Menu Bar: Setup > Settings

Front Panel: Settings

Application Toolbar: 

The measurement settings for the Audio Spectrum display are shown in the following table.


| Settings tab                                     | Description  |
|--|--|
| <a href="#">Params1 Tab (see page 473)</a>       | Specifies signal type, Audio Bandwidth, RBW, RBW filter, and Ref Audio Frequency.      |
| <a href="#">Params2 Tab (see page 476)</a>       | Specifies Harmonics and Non-Harmonics measurement parameters.                          |
| <a href="#">Audio Filters Tab (see page 476)</a> | Specifies the audio filter characteristics.  |
| <a href="#">Scale Tab (see page 479)</a>         | Sets vertical and horizontal scale and position parameters.                            |
| <a href="#">Prefs Tab (see page 482)</a>         | Specifies vertical units, and whether on not some features are displayed in the graph. |

## Audio Summary Display

To display the Audio Summary display:

1. Press the **Displays** button or select **Setup > Displays**. This shows the **Select Displays** dialog box.
2. From the **Measurements** box, select **Audio Analysis**.
3. Double-click the **Audio Summary** icon in the **Available Displays** box. This adds the Audio Summary icon to the **Selected displays** box.
4. Click the **OK** button. This shows the Audio Summary display.

### The Audio Summary Display



| Item | Display element     | Description   |
|------|---------------------|---|
| 1    | Measurement results | Measurement results are displayed in this area. The displayed measurements vary with signal type. |
| 2    | Hum & Noise results | Hum & Noise analysis readout. Hum & Noise appears only when enabled on the Hum & Noise tab        |

## Audio Summary Displayed Measurements

Table 11: Audio Summary Measurements

| Signal type        | Item          | Description   |
|--------------------|---------------|---|
| AM, FM, PM         | Carrier Power | Average power of the carrier signal with modulation removed.  |
| Direct             | Signal Power  | Average power of the input signal   |
| FM, PM             | Carr Freq Err | Carrier frequency error   |
| AM, FM, PM. Direct | Audio Freq    | Fundamental audio frequency   |
|                    | +Peak         | +Peak modulation excursion (where the modulation excursion readout depends on the signal type) <sup>1</sup>                                   |
|                    | -Peak         | -Peak modulation excursion (where the modulation excursion readout depends on the signal type) <sup>1</sup>                                   |
|                    | Peak-Peak/2   | Half peak-peak modulation excursion (where the modulation excursion readout depends on the signal type) <sup>1</sup>                          |
|                    | RMS           | RMS modulation excursion (where the modulation excursion readout depends on the signal type) <sup>1</sup>                                     |
|                    | SINAD         | Signal to noise and distortion  |
|                    | Mod Distor    | Modulation distortion   |
|                    | S/N           | An estimate of the Signal level to Noise (only) level, with the HD and NHD components removed   |
|                    | THD           | Total harmonic distortion   |
|                    | TNHD          | Total non-harmonic distortion   |
|                    | Ref           | Ref is the RMS modulation value stored when the <b>Capture Reference</b> button is pressed. (Displayed only when Hum & Noise is enabled.)     |
|                    | Diff          | Diff is the difference between the current RMS mod value and the Ref value captured previously. (Displayed only when Hum & Noise is enabled.) |

<sup>1</sup> For AM signal types, modulation excursion is "% Modulation Depth". For FM signal types, modulation excursion is "Frequency Deviation". For PM signal types, modulation excursion is "Phase Deviation". For Direct, there is no modulation excursion, it is actually "signal excursion".

## Audio Summary Settings

### Front Panel: Settings

#### Application Toolbar:

The measurement settings for the Audio Summary display are shown in the following table.

| Settings tab                                     | Description  |
|--|--|
| <a href="#">Params1 Tab (see page 473)</a>       | Specifies signal type, Audio Bandwidth, RBW, RBW filter, and Ref Audio Frequency.  |
| <a href="#">Params2 Tab (see page 476)</a>       | Specifies Harmonics and Non-Harmonics measurement parameters.  |
| <a href="#">Audio Filters Tab (see page 476)</a> | Specifies the audio filter characteristics.  |
| <a href="#">Hum Noise Tab (see page 481)</a>     | Specifies whether or not Hum & Noise is measured and enables the capture of a signal to be used as a reference of the Hum & Noise measurement. |

## Audio Analysis Measurement Settings

The control panel tabs in this section are shared by the displays in the Audio Analysis folder (Setup > Displays).

### Common controls for Audio Analysis displays

| Settings tab  | Description  |
|---|--|
| <a href="#">Params1 Tab (see page 473)</a>                  | Specifies characteristics about the audio signal and how measurements are made.                |
| <a href="#">Params2 Tab (see page 476)</a>                  | Specifies parameters that control how measurements are made on harmonics.                      |
| <a href="#">Audio Filters Tab (see page 476)</a>            | Specifies characteristics of filters applied to the signal before measurements are taken.      |
| <a href="#">Scale Tab (see page 479)</a>                    | Sets vertical and horizontal scale and position parameters.                                    |
| <a href="#">Prefs Tab for Audio Analysis (see page 482)</a> | The Prefs tab enables you to change appearance characteristics of the Audio Analysis displays. |

## Params1 Tab

The Params1 tab is used to specify characteristics of the audio signal to be measured and how the signal will be measured.

Params1 Params2 Audio Filters Hum & Noise

Signal Type:

Audio BW:  RBW:  ☒ Auto

Ref Audio Freq:  ☒ Auto RBW Filter:

Params1 tab for AM signal types

| Setting                                  | Description   |
|--|---|
| Signal Type                              | Specifies the type of signal to be analyzed. The available choices are AM, FM, PM, and Direct.  |
| Audio BW                                 | Specifies the bandwidth used for audio analysis.  |
| Ref Audio Freq                           | A measured value when Auto is selected. If you want to specify the reference audio frequency, uncheck Auto and enter a value manually. If Ref Audio Freq is set manually, be aware that the automated detection is still performed, but it is limited to a frequency range of $\pm 1\%$ of the Audio BW centered around the manually specified value. |
| Carrier Freq Error / Carrier Freq Offset | (FM and PM only) A measured value, when Auto is selected. If Auto is unchecked, you can specify the Carrier Frequency Offset.   |
| RBW                                      | Displays the Resolution Bandwidth for Audio measurements. This value is automatically set by default to 1/500 of the measurement bandwidth. To manually specify the RBW, uncheck Auto. The minimum RBW value is limited to the larger of 1 Hz or AudioBW/10000. The maximum is limited to AudioBW/100.  |
| RBW Filter                               | Specifies the windowing method used for the transform.  |

### Setting Frequency for Direct Signal Types

Direct (unmodulated) signal analysis is only possible with the instrument Frequency control set to 0 Hz. You will receive a warning to set Frequency to 0 Hz when Direct signal type is selected, if you haven't already done so. Modulated signal types may be selected with Frequency set to 0 Hz, but results are not meaningful in that case. For modulated signals, Frequency should always be set to a value  $\geq$  Audio Bandwidth to avoid self-interference of the signal due to spectral folding.

### Setting Audio Bandwidth

For AM, FM and PM, the Audio Bandwidth control sets not only the demodulated signal bandwidth, but also determines the pre-demodulation bandwidth. Set it to a value at least half the pre-demodulation signal bandwidth, as in this equation:

$$\text{Audio BW} \geq \text{Signal Bandwidth} / 2$$

**NOTE.** When performing audio analysis, you should allow the Audio BW control to automatically set the acquisition bandwidth, rather than manually adjusting the Acq BW control on the Sampling Parameters tab of the Acquire control panel. The audio measurement will cause the Acq BW setting to be  $\geq$  Audio BW (Direct) or  $\geq 2 \times$  Audio BW (AM, FM, PM).

For Direct signals, set Audio Bandwidth large enough to include any significant harmonics/non-harmonics or other signal component of interest. For example, to measure up to the 10th harmonic of a signal with a 5 kHz fundamental component, set Audio Bandwidth to  $10 \times 5 \text{ kHz} = 50 \text{ kHz}$ .

For modulated signals, Audio Bandwidth must be set wide enough to include all significant signal modulation components in addition to the desired audio analysis bandwidth. For AM this is similar to Direct. For example, to measure up to the 10th harmonic of an AM signal with 3 kHz fundamental component, set Audio Bandwidth to  $10 \times 3 \text{ kHz} = 30 \text{ kHz}$ . This ensures that the bandwidth of the data provided by the system to the measurement will be at least of 60 kHz ( $2 \times 30 \text{ kHz}$ ) which is sufficient for this signal and analysis requirement.

FM and PM are more complex. For FM, the analysis bandwidth needs to be at least twice as wide as the sum of peak Frequency Deviation and the Fundamental Frequency (Carsons rule). This is a parallel condition along with setting Audio Bandwidth large enough for the maximum audio bandwidth to analyze. Therefore, for FM, Audio Bandwidth should be:

$$\text{AudioBW (FM)} = \text{MAX}(\text{MaxAudioAnalysisFreq}, \text{FreqDeviation} + \text{FundamentalFreq})$$

where MaxAudioAnalysisFreq is the highest audio frequency desired in the analysis. For example, for an FM signal with fundamental signal of 5 kHz and peak frequency deviation (one-sided) of 10 kHz, Acquisition Bandwidth should be at least  $(2 \times (10\text{k} + 5\text{k})) = 30\text{ kHz}$ , or an Audio Bandwidth of 15 kHz. Also if the analysis should extend to the 8th harmonic, then the Audio Bandwidth needs to be at least  $8 \times 5\text{ kHz} = 40\text{ kHz}$ . So Audio Bandwidth should be set to 40 kHz. Using the equation:

$$\text{AudioBW(FM)} = \text{MAX}(8 \times 5\text{kHz}, (10 + 5)\text{kHz}) = \text{MAX}(40\text{ kHz}, 15\text{kHz}) = 40\text{ kHz}$$

The formula for PM is:

$$\text{AudioBW (PM)} = \text{MAX}(\text{MaxAudioAnalysisFreq}, \text{PMFreqDeviation} + \text{FundamentalFreq})$$

where

$$\text{PMFreqDeviation} = \text{PMPhaseDeviationInRadians} \times \text{FundamentalFreq}$$

### RBW Filter Shape

Select Kaiser in most cases for best measurement performance. Select Flattop only if you want to use standard markers to measure signal amplitude with highest accuracy.

## Params2 Tab

The Params2 tab is used to specify how the signal harmonics are measured and to control spectrum averaging.

The screenshot shows the Params2 tab with the following settings:

- Params1 | **Params2** | Audio Filters | Scale | Prefs
- No. of Harmonics: 12
- No. of Non-Harmonics: 12
- ☐ Averaging: 10
- Non-Harmonics only:
  - Ignore region: 0.0 Hz
  - Threshold: -65.0 dBc
  - Excursion: 6.0 dB

| Setting                | Description  |
|------------------------|--|
| No. of Harmonics       | Specifies the number of harmonics to detect. The detected harmonics are tagged with a number on the spectrum trace. The available range is 1–20.                         |
| No. of Non-Harmonics   | Specifies the number of non-harmonics to detect. The detected non-harmonics are tagged with a number on the spectrum trace. The available range is 0–20.                 |
| Averaging              | Specifies the number of averages used to compute the results. Range: 2–100.  |
| Ignore region          | Specifies the region about the signal frequency where the instrument will ignore non-harmonics.  |
| Non-Harmonic Threshold | Specifies the level which a spectrum peak must exceed to be declared a non-harmonic signal component.  |
| Non-Harmonic Excursion | Specifies the difference in level between a spectrum peak and the average noise level that must be exceeded for the peak to be declared a non-harmonic signal component. |

### About Averaging In Audio Analysis Displays

Analysis averaging is implemented using a “block” method. This means that the entire record required for multiple spectrum computations is acquired and analyzed within one analysis cycle. The result of each analysis update is a complete, independent result from a set of N spectrums averaged together, where N is the Averaging control value.. Since each update is a fully averaged result, no partially averaged results are output before a final result is available, so each output is fully valid. However, with large Averaging values, acquisition record sizes and measurement times may become large, so care should be taken to select the minimum amount of averaging needed.

## Audio Filters Tab

The Audio Filters tab is used to specify filters to be applied to the acquired audio signal before measurements are taken. You can select from pre-defined filters or use a filter you define in a text file. You can also specify the de-emphasis time constant applied to the audio signal and the telecom weighting filters used to measure noise.

## Pre-defined Filters

You can specify low-pass filter (LPF) and high-pass filter (HPF) settings, a de-emphasis time constant and/or a telecom weighting filter to match the response of your receiver. Alternatively, you can create a text file to specify the frequency response points.

To use pre-defined filters:

1. Select the **Pre-defined Filters** option button.
2. Select the LPF, HPF, De-emphasis and Standard check boxes as appropriate.
3. Select the desired filter parameter from the drop-down list for each of the enabled filters or select User from the list if you wish to use a custom value. For LPF and HPF, the listed frequencies represent the 3dB cutoff point of the filter.
4. If you select **User** from the drop-down list, enter a value in the text entry box that appears.

To disable all filtering:

1. Select the **Pre-defined Filters** option button.
2. Deselect all four filter check boxes.

**Table 12: Predefined audio filters**

| Filter type  | Available settings |
|--|--------------------|
| LPF (Low Pass Filter)<br>(5th-order Butterworth response)  | 300 Hz             |
|  | 3 kHz              |
|  | 15 kHz             |
|  | 30 kHz             |
|  | 80 kHz             |
|  | 300 kHz            |
|  | User <sup>1</sup>  |
| HPF (High Pass Filter)<br>(5th-order Butterworth response) | 20 Hz              |
|  | 50 Hz              |
|  | 300 Hz             |
|  | 400 Hz             |
|  | User <sup>1</sup>  |

**Table 12: Predefined audio filters (cont.)**

| Filter type           | Available settings               |
|-----------------------|----------------------------------|
| De-emphasis (FM only) | 25 $\mu$ s                       |
|                       | 50 $\mu$ s                       |
|                       | 75 $\mu$ s                       |
|                       | 750 $\mu$ s                      |
|                       | User. Range: 25 $\mu$ s to 10 ms |
| Standard              | CCITT                            |
|                       | C-message                        |

<sup>1</sup> User-entered values are restricted to maximum of 0.9 \* Audio BW, and will automatically adjust downward to meet this limit (if required) when AudioBW is decreased.

**Standard.** Use this setting to specify the telecom weighting filter. The characteristics of these filters are described in ITU-T Recommendation O.41, Psophometer for Use on Telephone-type Circuits.

### Using Custom Audio Filters

If you want to use audio filters that have a different response shape than the pre-defined filters, you can create a custom audio filter using a text or CSV file to specify the desired filter frequency response.

To use a custom audio filter:

1. Verify that Audio Spectrum or Audio Summary is the selected display.
2. Click the Settings icon or select **Setup > Settings**.
3. Select the **Audio Filters** tab.
4. Click the **File** button.
5. Click the ... button and navigate to the location of the custom audio filter file you wish to use. Select the file you want to use and click Open.
6. Acquire a new trace or replay a saved file to see the effect of the custom filter.

**Creating a Custom Audio Filter.** A custom audio filter file is either a plain text file or a CSV format file. The file contains frequency (in Hz) and amplitude (in dB) value pairs which specify the filter frequency response shape. Each pair must be on a separate line. The filter response between the points is interpolated using a cubic spline fit.

The following table shows the first few lines of a custom audio filter text file.

|        |       |
|--------|-------|
| 0,     | -100  |
| 16.66, | -85   |
| 50,    | -63   |
| 100,   | -41   |
| 200,   | -21   |
| 300,   | -10.6 |

|      |      |
|------|------|
| 400, | -6.3 |
| 500, | -3.6 |
| 600, | -2   |
| 700, | -0.9 |
| 800, | 0    |
| ...  | ...  |

Format and rule-checking on custom audio filter files is performed as follows:

- The maximum number of frequency and amplitude pairs is 1000.
- Column 1 (frequency values in Hertz).
  - Non-negative values only (zero is allowed).
  - Strictly increasing order of frequencies (frequency value on each line > frequency value on previous line).
  - There is no upper limit on the frequency value.
- Column 2 (amplitude values in dB units, where gain is a positive value and attenuation is a negative value).
  - Values are restricted to the range -200 to +20 dB.

## Scale Tab


The Scale tab allows you to change the vertical and horizontal scale settings. Changing the scale settings changes how the trace appears on the display but does not change acquisition control settings. In effect, these controls operate like pan and zoom controls.

The screenshot shows the 'Scale' tab in a software interface. It has five tabs at the top: 'Params1', 'Params2', 'Audio Filters', 'Scale' (selected), and 'Prefs'. The 'Scale' tab is divided into two main sections: 'Vertical' and 'Horizontal'. In the 'Vertical' section, there is a 'Scale' input field set to '100 dB' and a 'Position' input field set to '101 dBHz²'. Below these is an 'Autoscale' button. In the 'Horizontal' section, there is a 'Left' input field set to '0.0000 Hz' and a 'Right' input field set to '200.0 kHz'. Below these is a 'Log' checkbox (unchecked) and an 'Autoscale' button. A 'Reset Scale' button is located between the 'Vertical' and 'Horizontal' sections.

| Setting     | Description   |
|-------------|---|
| Vertical    |   |
| Scale       | Changes the range shown between the top and bottom of the graph.  |
| Position    | Adjusts the level shown at the top of the graph for linear units or the top of the graph for log units (for example, dBm).                |
| Autoscale   | Resets the Position so that the highest trace points are in the graph. For linear units (Volts, Watts), the Autoscale also adjusts Scale. |
| Horizontal  |   |
| Left        | Changes the frequency shown at the left side of the graph.  |
| Right       | Changes the frequency shown at the right side of the graph.   |
| Autoscale   | Resets the Left and Right settings to show the entire trace.  |
| Log         | Sets the graph horizontal axis to a logarithmic scale.  |
| Reset Scale | Resets the Vertical and Horizontal settings to their default values.  |

## Hum & Noise Tab

Hum & Noise (available only when the Audio Summary display is the active display) is useful for comparing residual power or modulation if the Ref value is captured when the Signal is On (Direct) or modulated (AM/FM/PM). When the signal is turned off (Direct) or modulation turned off (AM/FM/PM), Diff indicates how much residual Hum and Noise are still present in the measurement value.



To display Hum & Noise measurement:

1. Select **Setup > Displays**.
2. In the Select Displays window, select the **Audio Analysis** folder.
3. Double-click the **Audio Summary** icon so that it appears in the **Selected displays** box.
4. Click **OK**
5. With the Audio Summary display selected, select **Setup > Settings**.
6. Select the **Hum & Noise** tab. Click on the **Measure Hum & Noise** checkbox so that it is checked.  
The Hum & Noise values appear at the bottom list of measurements in the Audio Summary display.
7. Acquire an appropriate signal.
8. While the analyzer is analyzing a signal you want to use as a reference, click the **Capture reference now** button to save a reference value.

The Hum & Noise measurement compares the value of a specific signal quantity captured by the **Capture reference now** button with the current measured value of that quantity. For Direct signal types (set on the Params1 tab), the Signal Level is captured and compared. The Diff measurement is:

$$\text{SignalLevel}(\text{current}) - \text{SignalLevel}(\text{Ref})$$

in dB.

For AM, FM, and PM signal types, the RMS modulation value (related to Modulation Depth, Frequency Deviation or Phase Deviation) is captured and compared. The Diff measurement is:

$$20 \times \log_{10} (\text{RMS}(\text{current})/\text{RMS}(\text{Ref}))$$

in dB.

# Prefs Tab

The Prefs tab enables you to change appearance characteristics of the Audio Spectrum display.

Params1

Params2

Audio Filters

Scale

Prefs

Units: 

dBHz<sup>2</sup>

Trace points: 

801

☒ Show graticule

☒ Show Marker readout in graph (selected marker)

Show Non-Harm Threshold: 

Line Only

| Setting  | Description   |
|--|---|
| Units:   | Specifies the vertical scale units. The units available depend on the signal type selected. |
| Show graticule                                 | Select to display or hide the graticule.  |
| Trace points                                   | Sets the number of trace points used for marker measurements and for results export.        |
| Show Marker readout in graph (selected marker) | Shows or hides the readout for the selected marker in the graph area.                       |
| Show Non-Harm Threshold                        | Shows or hides the non-harmonic threshold line.   |

## Overview

The displays in General Purpose (GP) Digital Modulation (Displays > Measurements > GP Digital Modulation) are:

- Constellation
- Demod I & Q vs Time
- EVM vs Time
- Eye Diagram
- Frequency Deviation vs Time
- Magnitude Error vs Time
- Phase Error vs Time
- Signal Quality
- Symbol Table
- Trellis Diagram

The General Purpose Digital Modulation Analysis (Option 21) provides vector signal analyzer functionality. A wide variety of modulation types are supported, allowing you to view your signals in Constellation, Eye and Trellis diagrams, measure the quality of the modulation, display time-domain waveforms for demodulated I & Q signals, EVM, Phase Error, Magnitude Error, and more.

### Modulation Measurements

---

**NOTE.** *A maximum of approximately 80,000 samples can be analyzed by the General Purpose Digital Modulation measurements (the actual value varies with modulation type).*

---

| Measurement      | Description   |
|------------------|---|
| EVM              | The normalized RMS value of the error vector between the measured signal and the ideal reference signal over the analysis length. The EVM is generally measured on symbol or chip instants and can be reported in units of percent or dB. EVM is usually measured after best-fit estimates of the frequency error and a fixed phase offset have been removed. |
| Phase Error      | The RMS phase difference between the measured signal and the ideal reference signal.  |
| Magnitude Error  | The RMS magnitude difference between the measured signal and the ideal reference signal.  |
| IQ Origin Offset | The magnitude of the DC offset of the signal measured at the symbol times. It indicates the magnitude of the carrier feed-through signal.   |
| Gain Imbalance   | The gain difference between the I and Q channels in the signal generation path. Constellations with gain imbalance show a pattern with a width that is different from height.   |
| Rho p            | The normalized correlated power of the measured signal and the ideal reference signal. Like EVM, Rho is a measure of modulation quality. The value of Rho is less than 1 in all practical cases and is equal to 1 for a perfect signal measured in a perfect receiver.  |
| Frequency Error  | The frequency difference between the measured carrier frequency of the signal and the user-selected center frequency of the analyzer.   |
| Quadrature Error | The orthogonal error between the I and Q channels. The error shows the phase difference between I and Q channels away from the ideal 90 degrees expected from the perfect I/Q modulation. Constellations with quadrature error will show some leakage of I into Q and vice versa.   |


## Constellation Display

The Constellation Display shows a digitally-modulated signal in constellation form.

To show the Constellation Display:

1. Select the **Displays** button or select **Setup > Displays**. This shows the **Select Displays** dialog box.
2. From the **Measurements** box, select **GP Digital Modulation**.
3. Double-click the **Constellation** icon in the **Available Displays** box. This adds the Constellation icon to the **Selected displays** box.
4. Click the **OK** button. This shows the Constellation display.

## Elements of the Constellation Display



| Item | Display element      | Description   |
|------|----------------------|---|
| 1    | Check mark indicator | The check mark indicator in the upper, left-hand corner of the display shows when the Constellation display is the optimized display.<br><b>NOTE.</b> When Best for multiple windows is selected in the Amplitude control panel's RF & IF Optimization control, none of the measurement displays shows a checkmark, as there is not a single optimized measurement. |
| 2    | Marker Readout       | Located to the left of the constellation plot or below it, depending on the size of the window. If markers are enabled, the marker readout shows the time, mag, phase, symbol marker and symbol value of the point with the selected marker.  |
| 3    | EVM Readouts         | The EVM readouts are located below the Constellation plot. The readout shows EVM Peak (%) and location, RMS (%).  |
| 4    | Plot                 | Constellation graph.  |

[Changing Constellation Settings \(see page 485\)](#)

## Constellation Settings

**Menu Bar:** Setup > Settings

**Front Panel:** Settings

**Application Toolbar:** 

The settings for the Constellation view are shown in the following table.

**NOTE.** You might be able to save time configuring the Constellation display by pressing the *Standard Settings* button from the *Settings* control panel. This allows you to select a preset optimized for a standard from the *Select Standard* dialog box. See [Standard Settings Button. \(see page 509\)](#)

| Settings tab                                     | Description  |
|--|--|
| <a href="#">Modulation Params (see page 510)</a> | Specifies the type of modulation, symbol rate, and filters to be used in demodulating the input signal.  |
| <a href="#">Freq &amp; BW (see page 514)</a>     | Sets values for frequency error/offset, measurement bandwidth, and frequency deviation (not every control is present for every modulation type).   |
| <a href="#">Equalizer Tab (see page 515)</a>     | Enable the Equalizer and adjust its parameters.  |
| <a href="#">Advanced Params (see page 517)</a>   | Specifies additional parameters that are less frequently used.   |
| <a href="#">Find (see page 521)</a>              | Used to set parameters for finding a burst within the data record and for entering a Synch word.   |
| <a href="#">Analysis Time (see page 521)</a>     | Contains parameters that define the portion of the acquisition record that is used for analysis. Also allows you to specify the Units (Seconds or Symbols) for the GP Digital Modulation displays. |
| <a href="#">Trace (see page 523)</a>             | Allows you to set the trace display characteristics.   |
| <a href="#">Prefs (see page 526)</a>             | Enables you to set characteristics of the measurement display.   |


## Demod I & Q vs Time Display

The Demod I & Q vs Time displays demodulated I and Q vs. Time. You can choose to display I only, Q only, or both.

### Elements of the Display

To show the Demod I & Q vs Time display:

1. Select the **Displays** button or select **Setup > Displays**. This shows the **Select Displays** dialog box.
2. From the **Measurements** box, select **GP Digital Modulation**.
3. Double-click the **Demod I&Q vs Time** icon in the **Available Displays** box. This adds the Demod I&Q vs Time icon to the **Selected displays** box.
4. Click the **OK** button. This shows the Demod I&Q vs Time display.



| Item | Element              | Description   |
|------|----------------------|---|
| 1    | Check mark indicator | The check mark indicator in the upper, left-hand corner of the display shows when the Demod I & Q vs Time display is the optimized display.<br><br><b>NOTE.</b> When Best for multiple windows is selected in the Amplitude control panel's RF & IF Optimization control, none of the measurement displays shows a checkmark, as there is not a single optimized measurement.   |
| 2    | Top of Graph         | Sets the I and Q amplitude value indicated at the top of the graph. Changing the top value affects the bottom of graph readout. Also, note that the top of graph setting interacts with the internal vertical scale setting (which is not user settable) such that the range between the top and bottom of the graph increases or decreases automatically.  |
| 3    | Position             | Specifies the I and Q amplitude value shown at the center of the graph display.   |
| 4    | Bottom Readout       | Displays the I and Q amplitude value shown at the bottom of graph.  |
| 5    | Autoscale            | Adjusts the vertical and horizontal settings to provide the best display.   |
| 6    | Position             | Specifies the horizontal position of the trace on the graph display.  |
| 7    | Scale                | Adjusts the span of the graph. By decreasing the scale (time per division), the graph essentially becomes a window that you can move over the acquisition record by adjusting the offset.   |
| 8    | Freq Error           | This readout can show Freq Error or Freq Offset. When it displays Freq Error, it shows the difference between the instrument Frequency setting and the measured value of the signal's carrier frequency. When it displays Freq Offset, it shows the frequency offset specified on the <a href="#">Settings &gt; Freq &amp; BW</a> (see page 514) tab. If Freq Error is displayed, it also indicates that the Carrier frequency detection setting is Auto. If Freq Offset is displayed, it indicates that the Carrier frequency detection setting is manual. |

[Changing Demod I&Q Settings. \(see page 488\)](#)

## Demod I & Q vs Time Settings

**Menu Bar:** Setup > Settings

**Front Panel:** Settings

**Application Toolbar:** 

The settings for the Demod I & Q vs Time display are shown in the following table.

---

**NOTE.** You might be able to save time configuring the Demod I & Q display by pressing the Standard Settings button from the Settings control panel. This allows you to select a preset optimized for a standard from the Select Standard dialog box. See [Standard Settings Button. \(see page 509\)](#)

---

| Settings tab                                     | Description  |
|--|--|
| <a href="#">Modulation Params (see page 510)</a> | Specifies the type of modulation used for the input signal and other parameters.   |
| <a href="#">Freq &amp; BW (see page 514)</a>     | Specifies settings for frequency error, measurement BW and Frequency Deviation. Each of these settings is set internally when set to Auto. Alternatively, you can specify values appropriate for specific measurement needs. |
| <a href="#">Equalizer Tab (see page 515)</a>     | Enable the Equalizer and adjust its parameters.  |
| <a href="#">Advanced Params (see page 517)</a>   | Specifies additional parameters.   |
| <a href="#">Find (see page 521)</a>              | Used to set parameters for finding a burst within the data record and for entering a Synch word.   |
| <a href="#">Analysis Time (see page 521)</a>     | Contains parameters that define the portion of the acquisition record that is used for analysis. Also allows you to specify the Units (Seconds or Symbols) for the GP Digital Modulation displays.                           |
| <a href="#">Trace (see page 523)</a>             | Allows you to set the trace display characteristics.   |
| <a href="#">Scale (see page 525)</a>             | Specifies the horizontal and vertical scale settings.  |
| <a href="#">Prefs (see page 526)</a>             | Enables you to set characteristics of the measurement display.   |

## EVM vs Time Display

The EVM vs. Time Display shows the Error Vector Magnitude plotted over Time.

---


**NOTE.** A maximum of approximately 80,000 samples can be analyzed by the General Purpose Digital Modulation measurements (the actual value varies with modulation type).

---

To show an EVM vs. Time display:

1. Press the **Displays** button or select **Setup > Displays**. This shows the **Select Displays** dialog box.
2. From the **Measurements** box, select **GP Digital Modulation**.
3. Double-click the **EVM vs. Time** icon in the **Available Displays** box. This adds the EVM vs. Time icon to the **Selected displays** box.
4. Click the **OK** button. This displays the EVM vs. Time view.

### Elements of the EVM vs Time Display



| Item | Display element            | Description  |
|------|----------------------------|--|
| 1    | Top of graph adjustment    | Use the knob to adjust the vertical scale.   |
| 2    | Position                   | Adjusts the vertical position.   |
| 3    | Autoscale                  | Adjusts the Horizontal and Vertical scale to show the entire trace.  |
| 4    | Peak and RMS value readout | Shows the maximum result, the time it occurred, and the RMS of the result over the entire analysis length. |
| 5    | Scale                      | Sets the length of time shown in the graph.  |

[Changing the EVM vs Time Display Settings \(see page 489\)](#)

## EVM vs Time Settings

**Menu Bar:** Setup > Settings

**Front Panel:** Settings

**Application Toolbar:** 

The settings for the EVM vs. Time display are shown in the following table.

**NOTE.** You might be able to save time configuring the EVM vs. Time display by pressing the *Standard Settings* button from the *Settings* control panel. This allows you to select a preset optimized for a standard from the *Select Standard* dialog box. See [Standard Settings Button.](#) (see page 509)

| Settings tab                                     | Description  |
|--|--|
| <a href="#">Modulation Params (see page 510)</a> | Specifies the type of modulation used in the input signal and other parameters that controls the demodulation of the input signal.   |
| <a href="#">Freq &amp; BW (see page 514)</a>     | Specifies settings for frequency error, measurement BW and Frequency Deviation. Each of these settings is set internally when set to Auto. Alternatively, you can specify values appropriate for specific measurement needs. |
| <a href="#">Equalizer Tab (see page 515)</a>     | Enable the Equalizer and adjust its parameters.  |
| <a href="#">Advanced Params (see page 517)</a>   | Specifies Freq Offset, Magnitude normalization parameters, and enables swapping I and Q.   |
| <a href="#">Find (see page 521)</a>              | The Find tab is used to set parameters for finding bursts within the data record.  |
| <a href="#">Analysis Time (see page 521)</a>     | The Analysis Time tab contains parameters that define how the signal is analyzed in the general purpose digital modulation displays.   |
| <a href="#">Trace (see page 523)</a>             | Specifies the display characteristics of the displayed trace.  |
| <a href="#">Scale (see page 525)</a>             | Specifies the horizontal and vertical scale settings.  |
| <a href="#">Prefs (see page 526)</a>             | Specifies whether certain display elements are visible.  |


## Eye Diagram Display

The Eye Diagram Display shows a digitally modulated signal overlapped on itself to reveal variations in the signal.

To show an Eye Diagram display:

1. Press the **Displays** button or select **Setup > Displays**. This shows the **Select Displays** dialog box.
2. From the **Measurements** box, select **GP Digital Modulation**.
3. Double-click the **Eye Diagram** icon in the **Available Displays** box. This adds the Eye Diagram icon to the **Selected displays** box.
4. Click the **OK** button. This displays the Eye Diagram view.

## Elements of the Display



| Item | Element              | Description   |
|------|----------------------|---|
| 1    | Check mark indicator | The check mark indicator in the upper, left-hand corner of the display shows when the Frequency Deviation vs Time display is the optimized display.<br><br><b>NOTE.</b> When Best for multiple windows is selected in the Amplitude control panel's RF & IF Optimization control, none of the measurement displays shows a checkmark, as there is not a single optimized measurement. |
| 2    | Top of Graph         | The vertical scale is normalized with no units (except for nFSK and C4FM modulation types where the vertical units are Hz).   |
| 3    | Position             | Specifies the value shown at the center of the graph display.   |
| 4    | Bottom Readout       | Displays the value indicated by the bottom of graph.  |
| 5    | Autoscale            | Adjusts the vertical and horizontal settings to provide the best display.   |
| 6    | Position             | Displays the horizontal position of the trace on the graph display.   |
| 7    | Scale                | Adjusts the span of the graph in symbols.   |
| 8    | Freq Error           | Displays the difference between the maximum and minimum measured values of the signal frequency during the Measurement Time. The displayed frequency error is followed by either <b>Auto</b> or <b>Manual</b> . This indicates the selected carrier frequency detection method (see Settings > Freq & BW).  |

[Changing Eye Diagram Settings \(see page 492\)](#)

## Eye Diagram Settings

**Menu Bar:** Setup > Settings

**Front Panel:** Settings

**Application Toolbar:** 

The settings for the Eye Diagram display are shown in the following table.

---

**NOTE.** You might be able to save time configuring the Eye Diagram display by pressing the *Standard Settings* button from the Settings control panel. This allows you to select a preset optimized for a standard from the *Select Standard* dialog box. See [Standard Settings Button. \(see page 509\)](#)

---


| Settings tab                                     | Description  |
|--|--|
| <a href="#">Modulation Params (see page 510)</a> | Specifies the type of modulation used for the input signal and other parameters.   |
| <a href="#">Freq &amp; BW (see page 514)</a>     | Specifies settings for frequency error, measurement BW and Frequency Deviation. Each of these settings is set internally when set to Auto. Alternatively, you can specify values appropriate for specific measurement needs. |
| <a href="#">Advanced Params (see page 517)</a>   | Specifies additional parameters.   |
| <a href="#">Find (see page 521)</a>              | Used to set parameters for finding a burst within the data record and for entering a Synch word.   |
| <a href="#">Analysis Time (see page 521)</a>     | Contains parameters that define the portion of the acquisition record that is used for analysis. Also allows you to specify the Units (Seconds or Symbols) for the GP Digital Modulation displays.                           |
| <a href="#">Trace (see page 523)</a>             | Allows you to set the trace display characteristics.   |
| <a href="#">Scale (see page 525)</a>             | Specifies the horizontal and vertical scale settings.  |
| <a href="#">Prefs (see page 526)</a>             | Enables you to set characteristics of the measurement display.   |

## Frequency Deviation vs Time Display

To show a Frequency Deviation vs Time display:

1. Press the **Displays** button or select **Setup > Displays**. This shows the **Select Displays** dialog box.
2. From the **Measurements** box, select **Frequency Deviation vs Time**.
3. Double-click the **Frequency Deviation vs Time** icon in the **Available Displays** box. This adds the Frequency Deviation vs Time icon to the **Selected displays** box.
4. Click the **OK** button. This displays the Frequency Deviation vs Time view.

## Elements of the Display



| Item | Element              | Description  |
|------|----------------------|--|
| 1    | Check mark indicator | The check mark indicator in the upper, left-hand corner of the display shows when the Frequency Deviation vs Time display is the optimized display.<br><br><b>NOTE.</b> When <i>Best for multiple windows</i> is selected in the <i>Amplitude control</i> panel's <i>RF &amp; IF Optimization</i> control, none of the measurement displays shows a checkmark, as there is not a single optimized measurement. |
| 2    | Top of Graph         | Sets the frequency deviation value indicated at the top of the graph. Changing the top value affects the bottom of graph readout. Also, note that the top of graph setting interacts with the internal vertical scale setting (which is not user settable) such that the range between the top and bottom of the graph increases or decreases automatically.   |
| 3    | Position             | Specifies the frequency deviation value shown at the center of the graph display.  |
| 4    | Bottom Readout       | Displays the value of the frequency deviation value shown at the bottom of graph.  |
| 5    | Autoscale            | Adjusts the vertical and horizontal settings to provide the best display.  |
| 6    | Position             | Specifies the horizontal position of the trace on the graph display.   |
| 7    | Scale                | Adjusts the span of the graph. By decreasing the scale (time per division), the graph essentially becomes a window that you can move over the acquisition record by adjusting the offset.  |
| 8    | Freq Error           | Displays the difference between the maximum and minimum measured values of the signal frequency during the Measurement Time. The displayed frequency error is followed by either <b>Auto</b> or <b>Manual</b> . This indicates the selected carrier frequency detection method (see Settings > Freq & BW).   |

[Changing Frequency Deviation vs Time Settings \(see page 494\)](#)

## Frequency Deviation vs Time Settings

**Menu Bar:** Setup > Settings

**Front Panel:** Settings

**Application Toolbar:** 

The Setup settings for Frequency Deviation vs. Time are shown in the following table.

**NOTE.** You might be able to save time configuring the Frequency vs. Time display by pressing the Standard Settings button from the Settings control panel. This allows you to select a preset optimized for a standard from the Select Standard dialog box. See [Standard Settings Button. \(see page 509\)](#)

| Settings tab                                     | Description  |
|--|--|
| <a href="#">Modulation Params (see page 510)</a> | Specifies the type of modulation used for the input signal and other parameters.   |
| <a href="#">Freq &amp; BW (see page 514)</a>     | Specifies settings for frequency error, measurement BW and Frequency Deviation. Each of these settings is set internally when set to Auto. Alternatively, you can specify values appropriate for specific measurement needs. |
| <a href="#">Equalizer Tab (see page 515)</a>     | Enable the Equalizer and adjust its parameters.  |
| <a href="#">Advanced Params (see page 517)</a>   | Specifies additional parameters.   |
| <a href="#">Find (see page 521)</a>              | Used to set parameters for finding a burst within the data record and for entering a Synch word.   |
| <a href="#">Analysis Time (see page 521)</a>     | Contains parameters that define the portion of the acquisition record that is used for analysis. Also allows you to specify the Units (Seconds or Symbols) for the GP Digital Modulation displays.                           |
| <a href="#">Trace (see page 523)</a>             | Allows you to set the trace display characteristics.   |
| <a href="#">Scale (see page 525)</a>             | Specifies the horizontal and vertical scale settings.  |
| <a href="#">Prefs (see page 526)</a>             | Enables you to set characteristics of the measurement display.   |

## Magnitude Error vs Time Display


The Magnitude Error displays the magnitude of the symbol error. The amplitude appears on the vertical axis while time is plotted along the horizontal axis.

**NOTE.** A maximum of approximately 80,000 samples can be analyzed by the General Purpose Digital Modulation measurements (the actual value varies with modulation type).

To display Magnitude Error vs. Time:

1. Select the **Displays** button or **Setup > Displays**. This displays the Select Displays dialog box.
2. Select **GP Digital Modulation** in the **Measurements** box.
3. Double-click the **Mag Error vs. Time** icon or select the icon and click **Add**. The icon will appear in the **Selected displays** box and will no longer appear under Available displays.
4. Click **OK**.

### Elements of the Display



| Item | Display element            | Description   |
|------|----------------------------|---|
| 1    | Top of graph adjustment    | Use the knob to adjust the value of the vertical scale.   |
| 2    | Position                   | Adjusts the level shown at the bottom of the display.   |
| 3    | Autoscale button           | Adjusts the vertical and horizontal settings to provide the best display.   |
| 4    | Horizontal Position        | Adjusts the horizontal position of the signal. Units can be either Symbols or Seconds (Settings > Analysis Time tab > Units).   |
| 5    | Peak and RMS value readout | Displays the Peak value of the magnitude error, the RMS value of the magnitude error, and the time at which it occurs within the acquisition. Units can be either Symbols or Seconds (Settings > Analysis Time tab > Units).  |
| 6    | Horizontal Scale           | Sets the time spanned by the graph. Units can be either Symbols or Seconds (Settings > Analysis Time tab > Units).  |
| 7    | Freq Error                 | Freq Error is the difference between the Center Frequency and the measured frequency of the signal being tested. This readout will be Freq Offset if the Freq Offset parameter on the Settings > <a href="#">Advanced Params</a> (see page 517) tab is set to Manual. |

[Changing Magnitude Error vs Time Display Settings \(see page 496\)](#)

## Magnitude Error vs Time Settings

**Menu Bar:** Setup > Settings

**Front Panel:** Settings

**Application Toolbar:** 

The Setup settings for Magnitude Errors vs. Time are shown in the following table.

**NOTE.** You might be able to save time configuring the Magnitude vs. Time display by pressing the Standard Settings button from the Settings control panel. This allows you to select a preset optimized for a standard from the Select Standard dialog box. See [Standard Settings Button. \(see page 509\)](#)

| Settings tab                                     | Description  |
|--|--|
| <a href="#">Modulation Params (see page 510)</a> | Specifies the type of modulation used for the input signal and other parameters.   |
| <a href="#">Freq &amp; BW (see page 514)</a>     | Specifies settings for frequency error, measurement BW and Frequency Deviation. Each of these settings is set internally when set to Auto. Alternatively, you can specify values appropriate for specific measurement needs. |
| <a href="#">Equalizer Tab (see page 515)</a>     | Enable the Equalizer and adjust its parameters.  |
| <a href="#">Advanced Params (see page 517)</a>   | Specifies additional parameters.   |
| <a href="#">Find (see page 521)</a>              | Used to set parameters for finding a burst within the data record and for entering a Synch word.   |
| <a href="#">Analysis Time (see page 521)</a>     | Contains parameters that define the portion of the acquisition record that is used for analysis. Also allows you to specify the Units (Seconds or Symbols) for the GP Digital Modulation displays.                           |
| <a href="#">Trace (see page 523)</a>             | Allows you to set the trace display characteristics.   |
| <a href="#">Scale (see page 525)</a>             | Specifies the horizontal and vertical scale settings.  |
| <a href="#">Prefs (see page 526)</a>             | Enables you to set characteristics of the measurement display.   |

## Phase Error vs Time Display


The Phase Error vs. Time display shows the phase angle of the symbol error over time. The phase is plotted along the vertical axis while time is plotted along the horizontal axis.

**NOTE.** A maximum of approximately 80,000 samples can be analyzed by the General Purpose Digital Modulation measurements (the actual value varies with modulation type).

To show the Phase Error display:

1. Press the **Displays** button or select **Setup > Displays**.
2. In the **Select Displays** dialog, select **GP Digital Modulation** in the **Measurements** box.
3. In the **Available displays** box, double-click the **Phase Error** icon or select the icon and click **Add**.  
The Phase Error icon will appear in the **Selected displays** box and will no longer appear under Available displays.
4. Click **OK** to display the Phase Error.

### Elements of the Phase Error vs Time Display



| Item | Display element         | Description   |
|------|-------------------------|---|
| 1    | Freq Error              | Freq Error is the difference between the Center Frequency and the measured frequency of the signal being tested. This readout will be Freq Offset if the Freq Offset parameter on the Settings > <a href="#">Advanced Params</a> (see page 517) tab is set to Manual. |
| 2    | Top of graph adjustment | Adjusts the phase angle shown at the top of the graph.  |
| 3    | Position                | Adjusts the vertical offset.  |
| 4    | Autoscale               | Adjusts the vertical and horizontal settings so that the entire trace fits in the graph.  |
| 5    | Offset                  | Adjusts the horizontal offset.  |
| 6    | Peak and RMS readouts   | Displays the Peak value of the phase error and the time at which it occurred. Also displays the RMS value over the analysis length.   |
| 7    | Scale                   | Sets the time spanned by the graph.   |

[Changing the Phase Error vs Time Display Settings \(see page 498\)](#)

## Phase Error vs. Time Settings

**Menu Bar:** Setup > Settings

**Front Panel:** Settings

**Application Toolbar:** 

The settings for the Phase Error vs. Time display are shown in the following table.

---


**NOTE.** You might be able to save time configuring the Phase Error vs. Time display by pressing the Standard Settings button from the Settings control panel. This allows you to select a preset optimized for a standard from the Select Standard dialog box. See [Standard Settings Button. \(see page 509\)](#)

---


| Settings tab                                     | Description  |
|--|--|
| <a href="#">Modulation Params (see page 510)</a> | Specifies the type of modulation used for the input signal and other parameters.   |
| <a href="#">Freq &amp; BW (see page 514)</a>     | Specifies settings for frequency error, measurement BW and Frequency Deviation. Each of these settings is set internally when set to Auto. Alternatively, you can specify values appropriate for specific measurement needs. |
| <a href="#">Equalizer Tab (see page 515)</a>     | Enable the Equalizer and adjust its parameters.  |
| <a href="#">Advanced Params (see page 517)</a>   | Specifies additional parameters.   |
| <a href="#">Find (see page 521)</a>              | Used to set parameters for finding a burst within the data record and for entering a Synch word.   |
| <a href="#">Analysis Time (see page 521)</a>     | Contains parameters that define the portion of the acquisition record that is used for analysis. Also allows you to specify the Units (Seconds or Symbols) for the GP Digital Modulation displays.                           |
| <a href="#">Trace (see page 523)</a>             | Allows you to set the trace display characteristics.   |
| <a href="#">Scale (see page 525)</a>             | Specifies the horizontal and vertical scale settings.  |
| <a href="#">Prefs (see page 526)</a>             | Enables you to set characteristics of the measurement display.   |

## Signal Quality Display


The Signal Quality display shows several measurements of signal quality. The measurements displayed depend on the modulation type. There is a set of measurements displayed for all modulation types except nFSK and C4FM. There is a second set of measurements displayed for nFSK and C4FM modulation types.




Signal Quality display for all modulation types except nFSK, C4FM, OQPSK, and SOQPSK



Signal Quality display for OQPSK and SOQPSK modulation types



Signal Quality display for nFSK modulation type



Signal Quality display for C4FM modulation type

## Elements of the Display

### Measurements for all modulation types except nFSK, C4FM, OQPSK and SOQPSK

| Measurement      | Description   |
|------------------|---|
| EVM              | The normalized RMS value of the error vector between the measured signal and the ideal reference signal over the analysis length. The EVM is generally measured on symbol or chip instants and is reported in units of percent and dB. EVM is usually measured after best-fit estimates of the frequency error and a fixed phase offset have been removed. These estimates are made over the analysis length. Displays RMS and Peak values with location of Peak value. |
| Phase Error      | The RMS phase difference between the measured signal and the ideal reference signal. Displays RMS and Peak values with location of Peak value.  |
| Mag Error        | The RMS magnitude difference between the measured signal and the reference signal magnitude. Displays RMS and Peak values with location of Peak value.  |
| MER (RMS)        | The MER is defined as the ratio of I/Q signal power to I/Q noise power; the result is indicated in dB.  |
| IQ Origin Offset | The magnitude of the DC offset of the signal measured at the symbol times. It indicates the magnitude of the carrier feed-through signal.   |
| Frequency Error  | The frequency difference between the measured carrier frequency of the signal and the user-selected center frequency of the instrument.   |
| Gain Imbalance   | The gain difference between the I and Q channels in the signal generation path. Constellations with gain imbalance show a pattern with a width that is different from height.   |
| Quadrature Error | The orthogonal error between the I and Q channels. The error shows the phase difference between I and Q channels away from the ideal 90 degrees expected from the perfect I/Q modulation. Not valid for BPSK modulation type.   |
| Rho              | The normalized correlated power of the measured signal and the ideal reference signal. Like EVM, Rho is a measure of modulation quality. The value of Rho is less than 1 in all practical cases and is equal to 1 for a perfect signal measured in a perfect receiver.  |

**Measurements for QPSK and SOQPSK modulation types**

| Measurement      | Description   |
|------------------|---|
| EVM              | The normalized RMS value of the error vector between the measured signal and the ideal reference signal over the analysis length. The EVM is generally measured on symbol or chip instants and is reported in units of percent and dB. EVM is usually measured after best-fit estimates of the frequency error and a fixed phase offset have been removed. These estimates are made over the analysis length. Displays RMS and Peak values with location of Peak value.                       |
| Offset EVM       | Offset EVM is like EVM except for a difference in the time alignment of the I and Q samples. For EVM, I and Q samples are collected at the same time, for every symbol decision point (twice the symbol rate for offset modulations). For Offset EVM, the I and Q symbol decision points are time-aligned before collecting the I and Q samples. In this case, one I and one Q sample is collected for each symbol (half as many samples as the same number of symbols for (non-offset) EVM). |
| Phase Error      | The RMS phase difference between the measured signal and the ideal reference signal. Displays RMS and Peak values with location of Peak value.  |
| Mag Error        | The RMS magnitude difference between the measured signal and the reference signal magnitude. Displays RMS and Peak values with location of Peak value.  |
| MER (RMS)        | The MER is defined as the ratio of I/Q signal power to I/Q noise power; the result is indicated in dB.  |
| IQ Origin Offset | The magnitude of the DC offset of the signal measured at the symbol times. It indicates the magnitude of the carrier feed-through signal.   |
| Frequency Error  | The frequency difference between the measured carrier frequency of the signal and the user-selected center frequency of the instrument.   |
| Gain Imbalance   | The gain difference between the I and Q channels in the signal generation path. Constellations with gain imbalance show a pattern with a width that is different from height.   |
| Quadrature Error | The orthogonal error between the I and Q channels. The error shows the phase difference between I and Q channels away from the ideal 90 degrees expected from the perfect I/Q modulation. Not valid for BPSK modulation type.   |
| Rho              | The normalized correlated power of the measured signal and the ideal reference signal. Like EVM, Rho is a measure of modulation quality. The value of Rho is less than 1 in all practical cases and is equal to 1 for a perfect signal measured in a perfect receiver.  |

**Measurements for nFSK modulation types**

| Measurement       | Description   |
|-------------------|---|
| Peak FSK err      | Peak value of the frequency deviation error at the symbol point.  |
| RMS FSK Err       | RMS value of the frequency deviation error at the symbol point.   |
| Peak Mag Err      | The Peak magnitude difference between the measured signal and the reference signal magnitude.   |
| RMS Mag Err       | The RMS magnitude difference between the measured signal and the reference signal magnitude.  |
| Freq Error        | The frequency difference between the measured carrier frequency of the signal and the user-selected center frequency of the instrument.   |
| Freq Deviation    | Frequency distance from the center frequency at the symbol point.   |
| Symbol Rate Error | This compares the user-entered symbol rate to the instrument calculated symbol rate of the analyzed signal.   |
| Symbol Rate       | When in Auto-symbol rate, the instrument calculates the symbol rate of the signal and the instrument calculates the error between the user entered value and the instrument calculated value. |

**Measurements for C4FM modulation type**

| Measurement             | Description  |
|-------------------------|--|
| RMS Error Magnitude     | RMS value of the frequency deviation error at the symbol point.                  |
| Carrier Frequency Error | Frequency difference between averaged signal frequency and the center frequency. |
| Deviation               | Frequency distance from the center frequency at the symbol point.                |
| Length                  | Number of symbols in the analysis area.  |

[Changing the Signal Quality Display Settings \(see page 503\)](#)

## Signal Quality Settings

**Menu Bar:** Setup > Settings

**Front Panel:** Settings

**Application Toolbar:** 

The Setup settings for Signal Quality are accessible only when the Signal Quality display is selected.

---

**NOTE.** You might be able to save time configuring the Signal Quality display by pressing the Standard Settings button from the Settings control panel. This allows you to select a preset optimized for a standard from the Select Standard dialog box. See [Standard Settings Button. \(see page 509\)](#)

---


| Settings tab                                     | Description  |
|--|--|
| <a href="#">Modulation Params (see page 510)</a> | The Modulation tab specifies the type of modulation used for the input signal and other parameters.  |
| <a href="#">Freq &amp; BW (see page 514)</a>     | Specifies settings for frequency error, measurement BW and Frequency Deviation. Each of these settings is set internally when set to Auto. Alternatively, you can specify values appropriate for specific measurement needs. |
| <a href="#">Equalizer Tab (see page 515)</a>     | Enable the Equalizer and adjust its parameters.  |
| <a href="#">Advanced Params (see page 517)</a>   | The Advanced Params tab specifies frequency offset, magnitude normalization method and allows you to swap the I and Q signals.   |
| <a href="#">Find (see page 521)</a>              | Find tab is used to set parameters for finding bursts within the data record.  |
| <a href="#">Analysis Time (see page 521)</a>     | The Analysis Time tab contains parameters that define the portion of the acquisition record that is used for analysis.   |
| <a href="#">Prefs (see page 526)</a>             | The Prefs tab enables you to set characteristics of the measurement display.   |

## Symbol Table Display

The Symbol Table Display is like the Constellation Display except that a text table is used to display data bits at a symbol rather than a graph. The Synch Word characters, if used, are in bold font.

To display the Symbol Table:

1. Select the **Displays** button or select **Setup > Displays**. This displays the **Select Displays** dialog box.
2. From the **Measurements** box, select **GP Digital Modulation**.
3. Double-click the **Symbol Table** icon in the **Available Displays** box. This adds the Symbol Table icon to the **Selected displays** box.
4. Click the **OK** button. This displays the Symbol Table view.



|     |    |    |    |    |    |    |    |    |    |    |    |    |    |    |    |    |
|-----|----|----|----|----|----|----|----|----|----|----|----|----|----|----|----|----|
| 0   | 10 | 01 | 11 | 10 | 00 | 10 | 01 | 10 | 11 | 00 | 00 | 11 | 01 | 11 | 01 | 11 |
| 16  | 01 | 10 | 00 | 01 | 10 | 11 | 10 | 00 | 01 | 10 | 10 | 00 | 00 | 01 | 10 | 11 |
| 32  | 10 | 00 | 01 | 11 | 01 | 10 | 10 | 00 | 11 | 00 | 00 | 10 | 00 | 11 | 10 | 01 |
| 48  | 11 | 10 | 01 | 01 | 11 | 11 | 11 | 01 | 01 | 00 | 11 | 00 | 00 | 11 | 01 | 11 |
| 64  | 01 | 01 | 11 | 11 | 10 | 11 | 10 | 10 | 01 | 10 | 10 | 10 | 00 | 10 | 01 | 11 |
| 80  | 10 | 10 | 00 | 11 | 11 | 01 | 01 | 01 | 00 | 00 | 00 | 10 | 10 | 10 | 10 | 01 |
| 96  | 10 | 11 | 11 | 01 | 11 | 01 | 11 | 10 | 01 | 00 | 00 | 00 | 00 | 11 | 00 | 01 |
| 112 | 01 | 01 | 10 | 00 | 11 | 01 | 01 | 01 | 10 | 11 | 01 | 01 | 10 | 11 | 11 | 01 |
| 128 | 01 | 00 | 11 | 00 | 11 | 10 | 00 | 01 | 01 | 01 | 01 | 10 | 01 | 11 | 01 | 10 |
| 144 | 01 | 10 | 00 | 10 | 11 | 10 | 00 | 01 | 10 | 11 | 10 | 10 | 01 | 01 | 00 | 10 |
| 160 | 11 | 10 | 10 | 11 | 00 | 00 | 11 | 10 | 00 | 00 | 11 | 00 | 10 | 10 | 01 | 11 |
| 176 | 10 | 01 | 10 | 11 | 11 | 00 | 10 | 10 | 11 | 10 | 11 | 01 | 01 | 00 | 00 | 10 |
| 192 | 00 | 11 | 11 | 00 | 00 | 00 | 11 | 00 | 00 | 10 | 10 | 00 | 11 | 00 | 10 | 00 |
| 208 | 11 | 10 | 01 | 11 | 10 | 10 | 01 | 11 | 11 | 00 | 01 | 10 | 11 | 11 | 10 | 11 |
| 224 | 10 | 00 | 10 | 10 | 00 | 10 | 10 | 10 | 11 | 01 | 10 | 10 | 10 | 11 | 00 | 01 |
| 240 | 11 | 00 | 11 | 00 | 01 | 01 | 01 | 11 | 11 | 10 | 10 | 10 | 10 | 01 | 00 | 10 |

Marker: M1      Symbol: 0  
Time: 0.00 Sym      Value: 10

## Using Markers

Markers are indicators in the display that you can position on a trace to measure values such as frequency, power, and time. A Marker always displays its position and, if the Delta readout is enabled, will display the difference between its position and that of the Marker Reference. Within the Symbol Table, colored cells indicate the location of markers. The selected Marker is highlighted with a light green background. All other markers are highlighted with a light gray background. In the Symbol Table, the marker readout below the table shows the marker location in time, symbol numbers and symbol value.

[Changing the Symbol Table Display Settings \(see page 505\)](#)

## Symbol Table Settings

**Menu Bar:** Setup > Settings

**Front Panel:** Settings

**Application Toolbar:** 

The Setup settings for the Symbol Table view are shown in the following table.

**NOTE.** You might be able to save time configuring the Symbol Table display by pressing the Standard Settings button from the Settings control panel. This allows you to select a preset optimized for a standard from the Select Standard dialog box. See [Standard Settings Button. \(see page 509\)](#)


| Settings tab                                     | Description  |
|--|--|
| <a href="#">Modulation Params (see page 510)</a> | The Modulation tab specifies the type of modulation used for the input signal and other parameters.  |
| <a href="#">Freq &amp; BW (see page 514)</a>     | Specifies settings for frequency error, measurement BW and Frequency Deviation. Each of these settings is set internally when set to Auto. Alternatively, you can specify values appropriate for specific measurement needs. |
| <a href="#">Equalizer Tab (see page 515)</a>     | Enable the Equalizer and adjust its parameters.  |
| <a href="#">Advanced Params (see page 517)</a>   | The Advanced Params tab specifies additional parameters.   |
| <a href="#">Find (see page 521)</a>              | Find tab is used to set parameters for finding bursts within the data record.  |
| <a href="#">Analysis Time (see page 521)</a>     | The Analysis Time tab contains parameters that define the portion of the acquisition record that is used for analysis.   |
| <a href="#">Prefs (see page 526)</a>             | The Prefs tab enables you to set characteristics of the measurement display.   |

## Trellis Diagram Display

To show an Trellis Diagram display:

1. Press the **Displays** button or select **Setup > Displays**. This shows the **Select Displays** dialog box.
2. From the **Measurements** box, select **GP Digital Modulation**.
3. Double-click the **Trellis Diagram** icon in the **Available Displays** box. This adds the Trellis Diagram icon to the **Selected displays** box.
4. Click the **OK** button. This displays the Trellis Diagram view.

### Elements of the Display



| Item | Element              | Description   |
|------|----------------------|---|
| 1    | Check mark indicator | The check mark indicator in the upper, left-hand corner of the display shows when the Frequency Deviation vs Time display is the optimized display.<br><br><b>NOTE.</b> When Best for multiple windows is selected in the Amplitude control panel's RF & IF Optimization control, none of the measurement displays shows a checkmark, as there is not a single optimized measurement. |
| 2    | Top of Graph         | Sets the phase value indicated at the top of the graph. Changing the top value affects the bottom of graph readout. Also, note that the top of graph setting interacts with the internal vertical scale setting (which is not user settable) such that the range between the top and bottom of the graph increases or decreases automatically.  |
| 3    | Position             | Specifies the phase value shown at the center of the graph display.   |
| 4    | Bottom Readout       | Displays the value of the phase value shown at the bottom of graph.   |
| 5    | Autoscale            | Adjusts the vertical and horizontal settings to provide the best display.   |
| 6    | Position             | Displays the horizontal position of the trace on the graph display.   |
| 7    | Scale                | Adjusts the span of the graph in symbols.   |
| 8    | Freq Error           | Displays the difference between the maximum and minimum measured values of the signal frequency during the Measurement Time. The displayed frequency error is followed by either <b>Auto</b> or <b>Manual</b> . This indicates the selected carrier frequency detection method (see Settings > Freq & BW).  |

[Changing the Trellis Diagram Settings \(see page 507\)](#)

## Trellis Diagram Settings

**Menu Bar:** Setup > Settings

**Front Panel:** Settings

**Application Toolbar:** 

The settings for the Trellis Diagram display are shown in the following table.

**NOTE.** You might be able to save time configuring the Trellis Diagram display by pressing the Standard Settings button from the Settings control panel. This allows you to select a preset optimized for a standard from the Select Standard dialog box. See [Standard Settings Button. \(see page 509\)](#)

| Settings tab                                     | Description  |
|--|--|
| <a href="#">Modulation Params (see page 510)</a> | Specifies the type of modulation used for the input signal and other parameters.   |
| <a href="#">Freq &amp; BW (see page 514)</a>     | Specifies settings for frequency error, measurement BW and Frequency Deviation. Each of these settings is set internally when set to Auto. Alternatively, you can specify values appropriate for specific measurement needs. |
| <a href="#">Equalizer Tab (see page 515)</a>     | Enable the Equalizer and adjust its parameters.  |
| <a href="#">Advanced Params (see page 517)</a>   | Specifies additional parameters.   |
| <a href="#">Find (see page 521)</a>              | Used to set parameters for finding a burst within the data record and for entering a Synch word.   |
| <a href="#">Analysis Time (see page 521)</a>     | Contains parameters that define the portion of the acquisition record that is used for analysis. Also allows you to specify the Units (Seconds or Symbols) for the GP Digital Modulation displays.                           |
| <a href="#">Trace (see page 523)</a>             | Allows you to set the trace display characteristics.   |
| <a href="#">Scale (see page 525)</a>             | Specifies the horizontal and vertical scale settings.  |
| <a href="#">Prefs (see page 526)</a>             | Enables you to set characteristics of the measurement display.   |

## GP Digital Modulation Shared Measurement Settings

The displays in the GP Digital Modulation folder (Setup > Displays) are each a different format for presenting the results of a single underlying analysis. For this reason, all controls that affect the analysis parameters are shared by all the displays in the GP Digital Modulation folder.

Changing a setting on one tab changes that setting for all the GP Digital Modulation displays. For example, if you change the Modulation Type for the Constellation Display, it also changes the Modulation type setting for the Signal Quality display. There are some controls that affect only the way an individual display presents its results, such as graph scaling.

### Common controls for GP digital modulation displays

| Settings tab                                     | Description  |
|--|--|
| <a href="#">Modulation Params (see page 510)</a> | Specifies the type of modulation, symbol rate, and filters to be used in demodulating the input signal.  |
| <a href="#">Freq &amp; BW (see page 514)</a>     | Sets values for frequency error/offset, measurement bandwidth, and frequency deviation (not every control is present for every modulation type).   |
| <a href="#">Equalizer Tab (see page 515)</a>     | Enable the Equalizer and adjust its parameters.  |
| <a href="#">Advanced Params (see page 517)</a>   | Specifies additional parameters that are less frequently used.   |
| <a href="#">Find (see page 521)</a>              | Used to set parameters for finding a burst within the data record and for entering a Synchronizing word.   |
| <a href="#">Analysis Time (see page 521)</a>     | Contains parameters that define the portion of the acquisition record that is used for analysis. Also allows you to specify the Units (Seconds or Symbols) for the GP Digital Modulation displays. |
| <a href="#">Trace (see page 523)</a>             | Allows you to set the trace display characteristics.   |
| <a href="#">Prefs (see page 526)</a>             | Enables you to set characteristics of the measurement display.   |

## Standard Settings Button

On every GP Digital Modulation control panel there is a button labeled **Standard Settings**. This button is used to recall settings optimized for analyzing the selected standard. See the following table for a list of the standards for which standard settings are available. Choosing a standard from the dialog box changes only settings for GP Digital Modulation displays.

All of the presets in the Standard Settings Dialog make the following settings:

- Analysis Length: Auto
- Points per Symbol: 4
- Data Differential: No
- Burst Mode: Off
- Burst Detection Threshold: -10 dBc

## Parameter values set by presets in the standard settings dialog

| Standard           | Modulation | Symbol Rate | Meas. Filter    | Reference Filter | Filter Parameter | Other        |
|--------------------|------------|-------------|-----------------|------------------|------------------|--------------|
| 802.15.4           | OQPSK      | 1e6         | None            | Half sine        | NA               |              |
| SBPSK-MIL          | SBPSK      | 2.4e3       | None            | SBPSK-MIL        | NA               |              |
| SOQPSK-MIL         | SOQPSK     | 2.4e3       | None            | SOQPSK-MIL       | NA               | <sup>1</sup> |
| CPM-MIL            | CPM        | 19.2e3      | None            | None             | NA               |              |
| SOQPSK-ARTM Tier 1 | SOQPSK     | 2.5e6       | None            | SOQPSK-ARTM      | NA               | <sup>1</sup> |
| Project25 Phase 1  | C4FM       | 4.8e3       | C4FM-P25        | RC               | 0.2              |              |
| CDMA2000–Base      | QPSK       | 1.2288e6    | IS-95 TXE-Q_MEA | IS-95 REF        | NA               |              |
| W-CDMA             | QPSK       | 3.84e6      | RRC             | RC               | 0.22             |              |

<sup>1</sup> Center Symbol Position, Half Shift Removed

## Modulation Params Tab

### Menu bar: Setup > Settings > Modulation Params

The Modulation Params tab specifies the type of modulation on the input signal and other parameters that control the demodulation of the input signal.

The screenshot shows the 'Modulation Params' tab with the following settings:

- Modulation type: QPSK (dropdown)
- Measurement Filter: Root raised cosine (dropdown)
- Symbol Rate: 40 kHz (text input)
- Reference Filter: Raised cosine (dropdown)
- Filter Parameter: 0.300 (text input)

| Parameter          | Description   |
|--------------------|---|
| Modulation type    | Specifies the type of modulation on the input signal.                                     |
| Symbol Rate        | Specifies the symbol rate in Hertz.   |
| Measurement Filter | Specifies the filter used for measurements.   |
| Reference Filter   | Specifies the filter used as a reference.   |
| Filter Parameter   | Enter a value used for defining the Reference Filter. (Not present for some filter types) |
| Modulation index   | (Present only for CPM modulation type)  |

### Modulation Type

The modulation types that can be demodulated and analyzed are:

| Modulation type | Description                                     |
|-----------------|---|
| QPSK            | Quadrature Phase Shift Keying                   |
| 8PSK            | 8-Phase Shift Keying                            |
| D8PSK           | Differential Eight Phase Shift Keying           |
| D16PSK          | Differential Sixteen Phase Shift Keying         |
| PI/2DBPSK       | Pi/2 Differential Binary Phase Shift Keying     |
| DQPSK           | Differential Quadrature Phase Shift Keying      |
| PI/4DQPSK       | Pi/4 Differential Quadrature Phase Shift Keying |
| BPSK            | Binary Phase Shift Keying                       |
| OQPSK           | Offset Quadrature Phase Shift Keying            |
| 16QAM           | 16-state Quadrature Amplitude Modulation        |
| 32QAM           | 32-state Quadrature Amplitude Modulation        |
| 64QAM           | 64-state Quadrature Amplitude Modulation        |
| 128QAM          | 128-state Quadrature Amplitude Modulation       |
| 256QAM          | 256-state Quadrature Amplitude Modulation       |
| MSK             | Minimum Shift Keying                            |
| 2FSK            | 2-Frequency Shift Keying                        |
| 4FSK            | 4-Frequency Shift Keying                        |
| 8FSK            | 8-Frequency Shift Keying                        |
| 16FSK           | 16-Frequency Shift Keying                       |
| CPM             | Continuous Phase Modulation                     |
| SOQPSK          | Shaped Offset Quadrature Phase Shift Keying     |
| SBPSK           | Shaped Binary Phase Shift Keying                |
| C4FM            | Constant Envelope 4-Level Frequency Modulation  |
| 16APSK          | 16-symbol Amplitude Phase Shift Keying          |
| 32APSK          | 32-symbol Amplitude Phase Shift Keying          |

## Symbol Rate

Specifies the symbol rate for demodulating digitally modulated signals. The symbol rate and the bit rate are related as follows:

$$(\text{Symbol rate}) = (\text{Bit rate}) / (\text{Number of bits per symbol})$$

For example, the number of bits per symbol is 3 for 8PSK.

## Measurement and Reference Filters

The available measurement and reference filters depend on the selected modulation type. If a particular filter is not practical for a selected modulation type, it is not presented as an available filter. To determine which filters are available, make certain that your desired modulation type is selected. See the following table.

| Modulation type | Measurement filters | Reference filters  |
|-----------------|---------------------|--------------------|
| BPSK            | None                | None               |
| PI/2DBPSK       | RootRaisedCosine    | RaisedCosine       |
| 8PSK            | RaisedCosine        | Gaussian           |
| D8PSK           | Gaussian            | User               |
| DQPSK           | User                | Rectangular (freq) |
| PI/4DQPSK       | Rectangular (freq)  | IS-95REF           |
| 16QAM           | IS-95TX_MEA         |                    |
| 32QAM           | IS-95TXEQ_MEA       |                    |
| 64QAM           |                     |                    |
| 128QAM          |                     |                    |
| 256QAM          |                     |                    |
| QPSK            |                     |                    |
| 16APSK          |                     |                    |
| 32APSK          |                     |                    |
| MSK             | None                | None               |
|                 | Root Raised Cosine  | Gaussian           |
|                 | RaisedCosine        | User               |
|                 | Gaussian            |                    |
|                 | User                |                    |
|                 | Rectangular (freq)  |                    |
| OQPSK           | None                | None               |
|                 | RootRaisedCosine    | Half sine          |
|                 | User                | RaisedCosine       |
|                 | IS-95TX_MEA         | User               |
|                 | IS-95TXEQ_MEA       | IS-95REF           |
| HDQPSK          | HDQPSK-P25          | None               |
|                 |                     | RaisedCosine       |
|                 |                     | Gaussian           |
|                 |                     | User               |
|                 |                     | Rectangular (freq) |
| SOQPSK          | None                | SOQPSK-MIL         |
|                 | User                | SOQPSK-ARTM        |
|                 |                     | User               |
| CPM             | None                | None               |
|                 | User                | User               |

| Modulation type | Measurement filters | Reference filters |
|-----------------|---------------------|-------------------|
| 2FSK            | None                | None              |
| 4FSK            | Gaussian            | Gaussian          |
| 8FSK            | RootRaisedCosine    | RaisedCosine      |
| 16FSK           | RaisedCosine        | User              |
|                 | Rectangular (freq)  |                   |
|                 | User                |                   |
| C4FM            | C4FM-P25            | RaisedCosine      |
|                 |                     | User              |
| SBPSK           | None                | SBPSK-MIL         |
|                 | User                | User              |

The measurement filter is applied before the demodulation bit is detected and the ideal reference is calculated.

The reference filter is applied to the internally generated ideal reference signal before the EVM is calculated.

### How to Select Filters

In a signal transmitter/receiver system, the baseband signal might be filtered for bandwidth limiting or for another kind of necessary shaping that needs to be applied. Normally, a filter in the transmitter ( $F_t$ ) and a filter in the receiver ( $F_r$ ) are applied.

The Measurement Filter setting in the analyzer corresponds to the baseband filter in the receiver ( $F_r$ ): This setting tells the analyzer what filter your receiver uses. When the analyzer is set to the same filter used by the receiver, the analyzer sees the signal as your receiver would. The Measurement Filter setting should be the same as the filter used in the receiver under normal operation (as opposed to testing).

The Reference Filter setting in the analyzer corresponds to the baseband filter in the transmitter-receiver combination ( $F_r * F_t$ ). The baseband filter for the transmitter-receiver combination is often referred to as the *System Filter*. This filter is called the reference filter because it is used to recreate a reference signal that is compared to the received signal. This recreated reference signal is the *ideal signal* with  $F_r * F_t$  applied; differences between this *ideal signal* and the received signal enables the determination of signal quality, such as EVM measurements.

The following is an example of a hypothetical signal that is transmitted into a vector signal analyzer for analysis:

Assume that a signal is transmitted using a baseband filter ( $F_t$ ). It then travels through a transmission medium (air/cable/etc) where it may be affected by the communication channel ( $F_c$ ). The signal is received and filtered by the receiver's filter ( $F_r$ ). At this point, the signal has passed through  $F_t$  and  $F_r$ , and in addition, the communication channel might have affected it (so:  $F_t * F_r * F_c$ ). This double-filtered signal is demodulated as it was received to determine the symbols/bits in it. The obtained bits are used to regenerate a baseband ideal signal that can be compared against the received signal to determine signal quality. However, to determine the effect of the environment on the signal quality, the ideal signal must be filtered by the REFERENCE FILTER ( $F_t * F_r$ ), so that the ideal signal and the filtered signal differ only by the effect of the environment. So, the received signal is the ideal signal filtered by

$F_t * F_r * F_c$  and the reference signal is the ideal signal filtered by  $F_t * F_r$ , since they only differ by the effect of  $F_c$ , the comparison will show the effect of the communication channel on the signal. The communication channel can also include the hardware path the signal follows after (Tx) or before (Rx) digitizing; this would account for Tx/Rx hardware linear and non-linear distortion.

Common examples of how these filters are used are shown below:

- For Transmit Filter = Root Raised Cosine (RRC), Measurement Filter = RRC, the Reference Filter =  $RRC^2$  = Raised Cosine
- For Transmit Filter = Raised Cosine (RC), Measurement Filter = None, the Reference Filter = Raised Cosine (When the Measurement Filter = None, the Reference Filter = Transmit Filter)
- For Transmit Filter = Gaussian, Measurement Filter = None, the Reference Filter = Gaussian

### Filter Parameter

The filter parameter specifies the alpha for the Root Raised Cosine or Raised Cosine filter, or the bandwidth-time product (BT) for the Gaussian filter, when selected as the Reference filter. Some filter types have a fixed parameter value that is specified by industry standard, while other filter types by definition have no filter parameter. For filter types with no filter parameter, there is no filter parameter control present in the control panel.

## Freq & BW Tab

The Freq & BW tab specifies a group of settings that affect how measurements are made.

| Modulation Params  | Freq & BW | Advanced Params | Find | Analysis Time | Trace | Scale | Prefs |
|--|-----------|-----------------|------|---------------|-------|-------|-------|
| <p>Frequency Error: <input type="text" value="133.3 Hz"/> <input checked="" type="checkbox"/> Auto</p> <p>Measurement BW: <input type="text" value="30.72 MHz"/> <input type="button" value="Auto"/></p> <p>Frequency Deviation: <input type="text" value="465.7 kHz"/> <input checked="" type="checkbox"/> Auto</p> |           |                 |      |               |       |       |       |

Freq & BW tab with nFSK or C4FM modulation type selected and Frequency Error readout enabled (Auto selected)

Modulation Params Freq & BW Advanced Params Find Analysis Time Trace Scale Prefs

Frequency Error:  ☒ Auto

Measurement BW:

Freq & BW tab with SOQPSK modulation type selected and Frequency Offset enabled (Auto deselected)

| Setting                            | Description  |
|------------------------------------|--|
| Frequency Error / Frequency Offset | When Auto is enabled, this readout displays frequency error and the measurement is made at the calculated frequency. When Auto is disabled, this setting changes to Frequency Offset. When set to Frequency Offset, this setting is used to demodulate a signal that is not at the center frequency. The measurement is made at the user-entered offset. The Measurement Filter (if any, specified on the Modulation Params tab) is applied about the offset frequency whether set automatically or manually.                                |
| Measurement BW                     | This setting allows you to override the automatic bandwidth calculation and directly enter a bandwidth value. If you enter a value for the measurement bandwidth, be aware that the actual bandwidth of data provided to the measurement will be at least as wide as the value you request and may be as much as two times wider than requested. This override of the selected measurement bandwidth is done so that the instrument uses sufficient bandwidth relative to the chosen symbol rate to ensure good signal quality measurements. |
| Frequency Deviation                | For nFSK modulation types, this setting specifies the frequency deviation. Select Auto to make the instrument do this automatically. Deselect Auto to enter a value manually.<br>This setting is present for only nFSK and C4FM modulation types.  |

## Equalizer Tab

The Equalizer tab enables you to apply an adaptive equalizer to a digitally modulated signal to compensate for linear distortions in the signal. The Equalizer is available only for displays in the GP Digital Modulation folder (Select Displays window). The analyzer implements a decision directed, feed-forward FIR filter to correct linear distortion in the input signal.

Modulation Params Freq & BW Equalizer Advanced Params Find Analysis Time Trace Scale Prefs

☐ Enable Equalization Mode:  Taps/Symbol:

Convergence:  Taps:

☒ Enable EQ Export Length:

| Parameter           | Description   |
|---------------------|---|
| Enable Equalization | This setting turns the Equalizer on and off.  |
| Reset Equalization  | Initializes the equalizer filter for training.  |
| Enable EQ Export    | Exports a text file with equalizer taps in I/Q pairs.   |
| Mode                | Specifies whether the equalizer is in learning (Train) mode or analysis (Hold) mode.                              |
| Convergence         | Specifies the update rate. Maximum value: 0.002. Default value: 0.0005  |
| Taps/Symbol         | The number of filter coefficients per symbol used by the filter. Available choices are 1, 2, 4, and 8.            |
| Taps                | The number of filter coefficients. Range: 3 to 100 (you can set a higher number, but 100 is the practical limit). |
| Length              | Specifies the number of symbols analyzed (or filter length).  |

### Selecting the Mode

When enabled, the Equalizer is in either Train mode or Hold mode. When the equalizer is in Train mode, it will update internal filter parameters whenever you adjust the Convergence, Taps/Symbol, Taps, or Length values. When it is in Hold mode, the Equalizer uses the parameter values (both internal and the values accessible on the Equalizer tab) in effect when it was placed into Hold mode.

---

**NOTE.** *The Equalizer does not need to be retrained if the modulation type is changed. You can train the Equalizer by using a simpler modulation type (such as QPSK), place the Equalizer into Hold mode and can then measure more complex modulation types such as QAM.*

---

### Training the Equalizer

To obtain the desired results using the Equalizer, you must first train the Equalizer. This is an iterative process where you adjust some filter parameters (and the analyzer adjusts internal parameters) to achieve the lowest error possible on the acquired signal.

To configure the Equalizer:

1. Press the **Displays** button or select **Setup > Displays**.
2. In the **Select Displays** window, select **GP Digital Modulation** from **Measurements**.
3. Add **Signal Quality** to the **Selected Displays** and select **OK**.
4. With the Signal Quality display selected, select **Setup > Settings**.
5. Select the **Modulation Params** tab set the parameters as necessary for the signal.
6. Select the **Equalizer** tab.
7. Set the **Convergence** value to 0.0005.
8. Set **Taps/Symbol** to 2.
9. Click the **Reset Equalization** button to reset the equalizer.
10. Set the **Mode** to **Train**.

11. Click **Enable Equalization** so it is checked.
12. On the Signal Quality display, examine the value for EVM.
13. Change the Convergence, Taps/Symbol, and Taps values iteratively to achieve a minimum EVM value.

---

**NOTE.** *Changing the Taps/symbol or Taps values resets the equalizer.*

---

### Using the Equalizer

To use the equalizer:

1. Select and configure a **GP Digital Modulation** display.
2. Select **Setup > Settings**.
3. Select the **Equalizer** tab.
4. Select **Enable Equalization** so it is checked.
5. Verify that Mode is set to **Hold** if you have previously trained the Equalizer. If you have not previously trained the Equalizer, [train the Equalizer \(see page 516\)](#) and then set the Mode to Hold.

### Exporting EQ Files


If the **Enable EQ Export** box is checked, the following outputs files are automatically generated to these locations after each measurement update cycle if the EQ coefficients change: *c:\temp\EqTaps.txt* and *c:\temp\EqTapsHdr.txt*. These files have EQ taps in I/Q pairs (one pair per row).

The *Hdr* version file has a five line header before the EQ tap data. Each line in the five line header of the *Hdr* version file has two data values, with the second one always being "0":

```
<line1>Number_of_Header_Lines 0
<line2>EQ_Sample_Rate_Hz 0
<line3>Demod_Symbol_Rate_Sps 0
<line4>EQ_Samples_per_Symbol 0
<line5>Number_of_EQ_Taps_Following 0
```

## Advanced Params Tab

The Advanced Params tab specifies additional parameters that control the demodulation of the signal.




Modulation Params Freq & BW Advanced Params Find Analysis Time Trace Scale Prefs

Mag Normalize: RMS Symbol Magnitude ☐ Swap I and Q

☐ User Symbol Map (per Modulation Type):  
 ...

Advanced Params tab for all modulation types except nFSK, C4FM, and APSK




Modulation Params Freq & BW Advanced Params Find Analysis Time Trace Scale Prefs

☐ Swap I and Q

☐ Symbol rate search

☐ User Symbol Map (per Modulation Type):  
 ...

Advanced Params tab for modulation type nFSK




Modulation Params Freq & BW Advanced Params Find Analysis Time Trace Scale Prefs

☐ Swap I and Q

☐ User Symbol Map (per Modulation Type):  
 ...

Advanced Params tab for modulation type C4FM



Modulation Params Freq & BW Equalizer Advanced Params Find Analysis Time Traces Prefs

Mag Normalize: RMS Symbol Magnitude ☐ Swap I and Q

APSK Parameters

| Ring | Radius | Rotation |
|------|--------|----------|
| 0    | 1      | 45       |

☐ User Symbol Map (per Modulation Type):  
 ...

Advanced Params tab for modulation type APSK

| Parameter  | Description   |
|--|---|
| Mag Normalize<br>(not present for nFSK or C4FM modulation types) | Select RMS Symbol Magnitude or Max Symbol Magnitude. This setting applies to Mag Error and EVM.   |
| Swap I and Q   | When enabled, the I and Q data are exchanged before demodulating.   |
| APSK Parameters<br>(Present only for APSK modulation types)      | Sets Constellation display parameters for APSK modulation types.  |
| Symbol rate search<br>(Present only for nFSK modulation types)   | Determines whether to automatically detect or manually set the symbol rate. When selected, automatically detects the symbol rate to perform analysis. The calculated symbol rate is displayed in the Signal Quality display. The Symbol Rate Error is also calculated and displayed when Symbol rate search is enabled. |
| User Symbol Map (per Modulation Type)                            | Enables the use of custom symbol maps. This enables you to specify the location of symbols in the display. This control can be set independently for each of the modulation types.  |

### Mag Normalize

Specifies whether Magnitude Normalization uses the RMS Symbol Magnitude or the Maximum Symbol Magnitude as the basis for normalization. Use RMS Symbol Magnitude on QPSK modulations (equal magnitude symbol locations), and use Maximum Symbol Magnitude for signals that have a large difference in magnitude among the symbol locations (such as 128QAM). It prevents the instrument from using the very low magnitude center symbols when normalizing the constellation. The outer symbols are a better normalization reference than the center in this case.

### Swap I and Q

Use the Swap I and Q control to correct a signal sourced by a downconverter that inverts the frequency of the signal under test.

### APSK Parameters

Three parameters can be set for Constellation display of APSK modulation types.

| Parameter | Description   |
|-----------|---|
| Ring      | Use to select which ring the radius and rotation values are being specified for. Values are 0 and 1 for 16APSK. 0, 1, and 2 for 32APSK. |
| Radius    | Sets the Ring ratio relative to Ring 0 (inner most ring). Set after selecting the appropriate Ring setting.                             |
| Rotation  | Sets the rotation value from the I axis. Set after selecting the appropriate Ring setting.  |

### User Symbol Map

A User Symbol Map is a text file that specifies the location of symbols in the display. The symbol map is unique for each modulation type. The easiest way to create a custom symbol map is to start with the default symbol map and modify it. The default symbol map file is located at C:\RSA6100B Files.

The default symbol map file is named **DefaultSymbolMaps.txt**. See [Symbol Maps \(see page 526\)](#) for illustrations of the default symbol mapping.

To specify a user symbol map:

1. Click the ... button.
2. Navigate to the directory containing the user symbol text file you want to use.
3. Select the desired file in the Open window and click **Open**.
4. Select **User Symbol Map** to enable the user symbol map.

**Editing the User Symbol Map.** The symbol map is a plain text file and can be edited with any plain text editor.




---

**CAUTION.** *Whenever you reinstall the program software, the existing DefaultSymbolMaps.txt file will be overwritten. To create a custom symbol map, you should make a copy of the default symbol map file, edit the copy to suit your needs, and save it with a new name. Guidance on how to edit the symbol map file is contained within the default symbol map file.*

---

The following excerpt from the default symbol map file explains the structure of the file and how to edit it.

```
## Symbol Mapping Definitions
## Version 1.2

## This file defines the mapping of modulation states to symbol values.
##
## File Format :
## -----
## 1. Comments begin with '##' and may appear after the last field in a line
## 2. A symbol map begins and ends with a line containing the name of the
## modulation type. These names must exactly match the name of one of the
## modulation types in the RSA software
## 3. Empty cells may be included to preserve the constellation shape.
## 4. Blank lines are ignored.
## 5. A modulation type which does not match the name of an existing
## type will be ignored.
##
## Usage :
## -----
## 1. The file is intended to be edited with Notepad or similar text editor
```

The following text is an example of a symbol map.

```
## Symbol Map for 32 QAM
## (Resembles the shape of the constellation)
      32QAM
      00011 00010 00001 00000
01001 01000 00111 00110 00101 00100
01111 01110 01101 01100 01011 01010
10101 10100 10011 10010 10001 10000
```

11011 11010 11001 11000 10111 10110  
 11111 11110 11101 11100  
 32QAM

## Find Tab

The Find tab is used to set parameters for finding bursts within the data record. This is a post-acquisition operation. Synch Word search controls are also on this tab.

| Setting                    | Description   |
|----------------------------|---|
| Burst Detection: Mode      | <p>Select whether to analyze bursts</p> <ul style="list-style-type: none"> <li>- Auto: If a burst is found, analyze just that burst period. If a burst is not found, analyze the whole analysis length.</li> <li>- On: If a burst is found, analyze just that burst period. If a burst is not found, display an error message.</li> <li>- Off: Analyze the whole analysis length.</li> </ul> <p>If the signal isn't adequate for demodulation, an error message is shown.</p> |
| Burst Detection: Threshold | Sets the level required for the signal to qualify as a burst. Enter a value in dBc down from top of the signal.   |
| Use Synch Word             | When enabled, specifies the string of symbols to look for. Enter the search string with external keyboard or the on-screen keyboard.  |
| Clear                      | Blanks the search string field.   |

## Analysis Time Tab

The Analysis Time tab contains parameters that define how the signal is analyzed in the general purpose digital modulation displays.

|  |           |           |                 |      |               |       |       |       |
|--|-----------|-----------|-----------------|------|---------------|-------|-------|-------|
| Modulation Params  | Freq & BW | Equalizer | Advanced Params | Find | Analysis Time | Trace | Scale | Prefs |
| Analysis Offset: <input type="text" value="-66.667 ns"/> <input checked="" type="checkbox"/> Auto      Time Zero Reference: <input type="text" value="Trigger"/> |           |           |                 |      |               |       |       |       |
| Analysis Length: <input type="text" value="128 Sym"/> <input checked="" type="checkbox"/> Auto      Units: <input type="text" value="Symbols"/>                  |           |           |                 |      |               |       |       |       |
| Actual: 128 Sym  |           |           |                 |      |               |       |       |       |

The settings values on this tab are the same as those on the main Analysis control panel for the instrument with the only difference being that Analysis Length can be set in either Seconds or Symbols in this location.

| Setting             | Description   |
|---------------------|---|
| Analysis Offset     | Specifies the location of the first time sample to use in measurements.   |
| Auto                | When enabled, causes the instrument to set the Analysis Offset value based on the requirements of the selected display.   |
| Analysis Length     | Specifies the length of the analysis period to use in measurements. Length is specified in either seconds or symbols, depending on the Units setting. For most modulation types, the Analysis Length set when Auto is enabled is 128 symbols. For some modulation types, a longer length is used. |
| Auto                | When enabled, causes the instrument to set the Analysis Length value based on the requirements of the selected display.   |
| Time Zero Reference | Specifies the zero point for the analysis time.   |
| Actual              | This is a displayed value, not a setting. It is the Analysis Length (time or symbols) being used by the analyzer; this value may not match the Analysis Length requested (in manual mode).  |
| Units               | Sets the units of the Analysis Length to either Symbols or Seconds.   |

## Analysis Offset

Use analysis offset to specify where measurements begin. Be aware that you cannot set the Analysis Offset outside the range of time covered by the current acquisition data. (all time values are relative to the Time Zero Reference).

You can set the Analysis Length so that the requested analysis period falls partly or entirely outside the current range of acquisition data settings. When the next acquisition is taken, its Acquisition Length will be increased to cover the new Analysis Length, as long as the Sampling controls are set to Auto. If the Sampling parameters are set to manual, or if the instrument is analyzing saved data, the actual analysis length will be constrained by the available data length, but in most cases, measurements are able to be made anyway. The instrument will display a notification when measurement results are computed from less data than requested. Range: 0 to [(end of acquisition) - Analysis Length]. Resolution: 1 effective sample (or symbol).

## Analysis Length

Use the analysis length to specify how long a period of time is analyzed by a measurement. As you adjust this value, the actual amount of time for Analysis Length, in Symbol or Seconds units, is shown below the control in the "Actual" readout. This setting is not available when Auto is checked. Range: minimum

value depends on modulation type. Resolution: 1 symbol. A maximum of approximately 80,000 samples can be analyzed (the actual value varies with modulation type).

### Time Zero Reference

All time values are measured from this point (such as marker position or horizontal position (in Y vs Time displays). Choices are: Acquisition Start or Trigger.

| Parameter         | Description  |
|-------------------|--|
| Acquisition Start | Offset is measured from the point at which acquisition begins. |
| Trigger           | Offset is measured from the trigger point.                     |

## Trace Tab

### Menu Bar: Setup > Settings > Trace

The Trace tab allows you to set the trace display characteristics of the selected display.

The screenshot shows the 'Traces' tab in the software interface. The top menu bar includes 'Modulation Params', 'Freq & BW', 'Equalizer', 'Advanced Params', 'Find', 'Analysis Time', 'Traces', and 'Prefs'. The 'Traces' tab is active. The settings are as follows:

- Trace: Trace 2 (dropdown)
- Show: ☒ (checkbox)
- Freeze: ☒ (checkbox)
- Content: Lines (dropdown)
- Q Offset: Use shared Pref (dropdown)
- All GP Digital Modulation displays:
  - Points/symbol: 4 (dropdown)
  - Symbol point location: Center (dropdown)

Example traces tab for constellation display set to SOQPSK modulation type

Note that some settings are not present for all modulation types and some settings are not present for all displays.

The screenshot shows the 'Traces' tab in the software interface. The top menu bar includes 'Modulation Params', 'Freq & BW', 'Equalizer', 'Advanced Params', 'Find', 'Analysis Time', 'Traces', and 'Prefs'. The 'Traces' tab is active. The settings are as follows:

- Trace: Trace 1 (dropdown)
- Show: ☒ (checkbox)
- Freeze: ☐ (checkbox)
- Content: Vectors (dropdown)
- Phase multiplier: 1 (dropdown)
- All GP Digital Modulation displays:
  - Points/symbol: 4 (dropdown)

Example trace tab for constellation display set to CPM modulation type

|  |           |           |                 |      |               |       |       |       |
|--|-----------|-----------|-----------------|------|---------------|-------|-------|-------|
| Modulation Params  | Freq & BW | Equalizer | Advanced Params | Find | Analysis Time | Trace | Scale | Prefs |
| Trace: <span>I</span> <input checked="" type="checkbox"/> Show<br>Points/symbol: <span>8</span><br>Content: <span>Points</span> Symbol point location: <span>Center</span> |           |           |                 |      |               |       |       |       |

Example trace tab for Demod I&amp;Q display

| Setting               | Description  |
|-----------------------|--|
| Trace                 | Selects the trace that is hidden or displayed based on whether or not Show is selected.  |
| Show                  | Specifies whether the trace selected by Trace is displayed or hidden.  |
| Freeze                | Halts updates to the trace selected by the Trace setting. Present for the Constellation display only.  |
| Content               | Selects whether to display the trace as vectors (points connected by lines), points (symbols only without lines), or lines (lines drawn between symbols, but no symbols are displayed). The choices available depend on the display.   |
| Q Offset              | For traces with offset modulation (OQPSK and SOQPSK), this setting enables the trace to be displayed with Q offset or without Q offset. Choices available are <b>Remove Q offset</b> , <b>Include Q Offset</b> , and <b>Use Shared Pref</b> . Selecting Use Shared Pref causes the analyzer to add or remove Q offset according to the <b>Remove Q Offset</b> setting located on the Prefs tab. Setting the Q Offset on the Traces tab changes the Q offset only for the Constellation display. If other GP Digital Modulation displays are shown, they will use the use the Remove Q Offset setting on the Prefs tab regardless of the Q Offset setting on the traces tab for the Constellation display. The Q Offset setting is available only when the modulation type is set to OQPSK or SOQPSK. |
| Points/symbol         | Select how many points to use between symbols when connecting the dots. Values: 1, 2, 4, 8.  |
| Symbol point location | Selects whether to evaluate the symbol value at the center or the end of the eye opening. This control is only present for some of the supported modulation types.   |
| Phase Multiplier      | Sets the multiplication constant for the phase multiplication display: ×1 (default), ×2, ×4, ×8, ×16, or ×32. The phase multiplication display facilitates observation of noisy CPM signals by multiplying measurement signal phase by the constant to reduce the number of phase states and expand the phase difference between adjacent symbols.   |

## Comparing Two Traces in the Constellation Display

When the Constellation display is the selected display, you can use the Traces tab to enable the display of a second trace. The second trace is a version of the current acquisition. You can choose to freeze a trace in order to display the current live trace to an earlier version of itself, you can display the trace as a second trace with or without Q Offset, or you can choose to display both traces frozen in order to compare the trace to itself at different times.

To display a second trace in the Constellation display:

1. If more than one display is present, select the Constellation display to ensure it is the selected display.
2. Click the settings icon or select **Setup > Settings** from the menu bar.

3. Select the **Traces** tab.
4. Select **Trace 2** from the **Trace** drop-down list.
5. Click the **Show** checkbox so that it is checked.
6. Specify the **Content** as desired. Trace 2 lines appear in blue to aid in distinguishing Trace 2 from Trace 1.

## Scale Tab

The Scale tab allows you to change the vertical and horizontal scale settings. Changing the scale settings changes how the trace appears on the display but does not change control settings such as Measurement Frequency.

The screenshot shows the 'Scale' tab in a software interface. It has two main sections: 'Vertical' and 'Horizontal'.  
 In the 'Vertical' section, there is a 'Scale' input field set to '200 mV', a 'Position' input field set to '0.000 V', and an 'Autoscale' button.  
 In the 'Horizontal' section, there is a 'Scale' input field set to '128.00 Sym', a 'Position' input field set to '-1.00 Sym', a checked 'Auto' checkbox, and an 'Autoscale' button.  
 At the top of the window, there are several tabs: 'Modulation Params', 'Freq & BW', 'Equalizer', 'Advanced Params', 'Find', 'Analysis Time', 'Trace', 'Scale' (which is active), and 'Prefs'.


| Setting    | Description  |
|------------|--|
| Vertical   | Controls the vertical position and scale of the trace display.   |
| Scale      | Changes the vertical scale units.  |
| Position   | Position adjusts the reference level away from top of the graph.   |
| Autoscale  | Resets the scale of the vertical axis to contain the complete trace.   |
| Horizontal | Controls the span of the trace display and position of the trace.  |
| Scale      | Allows you to, in effect, change the span.   |
| Position   | Allows you to pan a zoomed trace without changing the Measurement Frequency. Position is only enabled when the span, as specified by Freq/div, is less than the acquisition bandwidth. |
| Autoscale  | Resets the scale of the horizontal axis to contain the complete trace.   |
| Auto       |  |

## A Note About Units

The Units used for the horizontal scale can be either Seconds or Symbols. To set the units for the horizontal scale, display the **Analysis Time** tab. On the tab, select the appropriate units from the **Units** drop-down list.

Prefs Tab

The Prefs tab enables you to change appearance characteristics of the GP Digital Modulation displays.



| Setting  | Description  |
|--|--|
| Show graticule                                 | Shows or hides the graticule.  |
| Show Marker readout in graph (selected marker) | Shows or hides the readout for the selected marker in the graph area.  |
| Radix  | Specifies how symbols are displayed in the Symbol Table display and in the Marker readout in the Constellation display.  |
| Remove Q offset                                | The I and Q data traces are displayed with an offset of half a symbol when the modulation type is set to SOQPSK or OQPSK. You can remove this offset by selecting <b>Remove Q offset</b> . (Which is only present when the modulation type is set to OQPSK or SOQPSK.) |
| Show EVM and Offset EVM results                | Adds EVM results in the display when enabled (Offset EVM is always displayed). Present only in Signal Quality display and with SOQPSK and OQPSK modulation types only  |

Symbol Maps

This topic shows the symbol mapping for each digital modulation technique.

| QPSK |   | OQPSK |    | SOQPSK |    |
|------|---|-------|----|--------|----|
| 1    | 3 | 01    | 11 | 01     | 11 |
| 0    | 2 | 00    | 10 | 00     | 10 |

0109-010

**8PSK**

|      |  |   |  |   |   |  |       |   |
|------|--|---|--|---|---|--|-------|---|
| Left |  |   |  | 3 |   |  | Right |   |
|      |  |   |  |   |   |  |       |   |
|      |  | 2 |  |   | 1 |  |       |   |
| 6    |  |   |  |   |   |  |       | 0 |
|      |  | 7 |  |   | 4 |  |       |   |
|      |  |   |  | 5 |   |  |       |   |

0100-000

**BPSK**

|   |   |
|---|---|
| 1 | 0 |
|---|---|

**SBPSK**

|   |   |
|---|---|
| 1 | 0 |
|---|---|

0100-010

**D8PSK**

| Phase shift<br>(radians) | Symbol value<br>(binary) |
|--------------------------|--------------------------|
| 0                        | 000                      |
| $\pi/4$                  | 001                      |
| $\pi/2$                  | 011                      |
| $3\pi/4$                 | 010                      |
| $\pi$                    | 110                      |
| $5\pi/4$                 | 111                      |
| $3\pi/2$                 | 101                      |
| $7\pi/4$                 | 100                      |

**Pi/2 DBPSK**

| Phase shift<br>(radians) | Symbol value<br>(binary) |
|--------------------------|--------------------------|
| $+\pi/2$                 | 0                        |
| $-\pi/2$                 | 1                        |

DQPSK

| Phase shift<br>(radians) | Symbol value<br>(binary) |
|--------------------------|--------------------------|
| 0                        | 00                       |
| $\pi/2$                  | 01                       |
| $\pi$                    | 11                       |
| $3\pi/2$                 | 10                       |

Pi/4 DQPSK

| Phase shift<br>(radians) | Symbol value<br>(binary) |
|--------------------------|--------------------------|
| $+\pi/4$                 | 00                       |
| $+3\pi/4$                | 01                       |
| $-\pi/4$                 | 10                       |
| $-3\pi/4$                | 11                       |

MSK

| Phase shift<br>direction | Symbol value<br>(binary) |
|--------------------------|--------------------------|
| -                        | 0                        |
| +                        | 1                        |

16QAM

|      |   |       |   |
|------|---|-------|---|
| Left |   | Right |   |
| 3    | 2 | 1     | 0 |
| 7    | 6 | 5     | 4 |
| B    | A | 9     | 8 |
| F    | E | D     | C |

0169-010

**32QAM**

| Left |    |    |    | Right |    |
|------|----|----|----|-------|----|
|      | 3  | 2  | 1  | 0     |    |
| 9    | 8  | 7  | 6  | 5     | 4  |
| F    | E  | D  | C  | B     | A  |
| 15   | 14 | 13 | 12 | 11    | 10 |
| 1B   | 1A | 19 | 18 | 17    | 16 |
|      | 1F | 1E | 1D | 1C    |    |

0109-011

**64QAM**

| Left |    |    |    |    |    | Right |    |
|------|----|----|----|----|----|-------|----|
| 7    | 6  | 5  | 4  | 3  | 2  | 1     | 0  |
| F    | E  | D  | C  | B  | A  | 9     | 8  |
| 17   | 16 | 15 | 14 | 13 | 12 | 11    | 10 |
| 1F   | 1E | 1D | 1C | 1B | 1A | 19    | 18 |
| 27   | 26 | 25 | 24 | 23 | 22 | 21    | 20 |
| 2F   | 2E | 2D | 2C | 2B | 2A | 29    | 28 |
| 37   | 36 | 35 | 34 | 33 | 32 | 31    | 30 |
| 3F   | 3E | 3D | 3C | 3B | 3A | 39    | 38 |

0109-012

**128QAM**

|    |    |    |    |    |    |    |    |    |    |    |    |
|----|----|----|----|----|----|----|----|----|----|----|----|
|    |    | 5D | 5F | 4F | 4D | 1A | 1B | 0B | 0A |    |    |
|    |    | 5C | 5E | 4E | 4C | 18 | 19 | 09 | 08 |    |    |
| 4A | 48 | 54 | 56 | 46 | 44 | 10 | 11 | 15 | 14 | 1C | 1D |
| 4B | 49 | 55 | 57 | 47 | 45 | 12 | 13 | 17 | 16 | 1E | 1F |
| 5B | 59 | 51 | 53 | 43 | 41 | 02 | 03 | 07 | 06 | 0E | 0F |
| 5A | 58 | 50 | 52 | 42 | 40 | 00 | 01 | 05 | 04 | 0C | 0D |
| 6D | 6C | 64 | 65 | 61 | 60 | 20 | 22 | 32 | 30 | 38 | 3A |
| 6F | 6E | 66 | 67 | 63 | 62 | 21 | 23 | 33 | 31 | 39 | 3B |
| 7F | 7E | 76 | 77 | 73 | 72 | 25 | 27 | 37 | 35 | 29 | 2B |
| 7D | 7C | 74 | 75 | 71 | 70 | 24 | 26 | 36 | 34 | 28 | 2A |
|    |    | 68 | 69 | 79 | 78 | 2C | 2E | 3E | 3C |    |    |
|    |    | 6A | 6B | 7B | 7A | 2D | 2F | 3F | 3D |    |    |

0169-017

**256QAM**

Left

Right

|    |    |    |    |    |    |    |    |    |    |    |    |    |    |    |    |
|----|----|----|----|----|----|----|----|----|----|----|----|----|----|----|----|
| EF | FD | EB | F9 | E7 | F5 | E3 | F1 | 0F | 3F | 4F | 7F | 8F | BF | CF | FF |
| CE | DC | CA | D8 | C6 | D4 | C2 | D0 | 0C | 3C | 4C | 7C | 8C | BC | CC | FC |
| AF | BD | AB | B9 | A7 | B5 | A3 | B1 | 0B | 3B | 4B | 7B | 8B | BB | CB | FB |
| 8E | 9C | 8A | 98 | 86 | 94 | 82 | 90 | 08 | 38 | 48 | 78 | 88 | B8 | C8 | F8 |
| 6F | 7D | 6B | 79 | 67 | 75 | 63 | 71 | 07 | 37 | 47 | 77 | 87 | B7 | C7 | F7 |
| 4E | 5C | 4A | 58 | 46 | 54 | 42 | 50 | 04 | 34 | 44 | 74 | 84 | B4 | C4 | F4 |
| 2F | 3D | 2B | 39 | 27 | 35 | 23 | 31 | 03 | 33 | 43 | 73 | 83 | B3 | C3 | F3 |
| 0E | 1C | 0A | 18 | 06 | 14 | 02 | 10 | 00 | 30 | 40 | 70 | 80 | B0 | C0 | F0 |
| E1 | D1 | A1 | 91 | 61 | 51 | 21 | 11 | 01 | 13 | 05 | 17 | 09 | 1B | 0D | 1F |
| E2 | D2 | A2 | 92 | 62 | 52 | 22 | 12 | 20 | 32 | 24 | 36 | 28 | 3A | 2C | 3E |
| E5 | D5 | A5 | 95 | 65 | 55 | 25 | 15 | 41 | 53 | 45 | 57 | 49 | 5B | 4D | 5F |
| E6 | D6 | A6 | 96 | 66 | 56 | 26 | 16 | 60 | 72 | 64 | 76 | 68 | 7A | 6C | 7E |
| E9 | D9 | A9 | 99 | 69 | 59 | 29 | 19 | 81 | 93 | 85 | 97 | 89 | 9B | 8D | 9F |
| EA | DA | AA | 9A | 6A | 5A | 2A | 1A | A0 | B2 | A4 | B6 | A8 | BA | AC | BE |
| ED | DD | AD | 9D | 6D | 5D | 2D | 1D | C1 | D3 | C5 | D7 | C9 | DB | CD | DF |
| EE | DE | AE | 9E | 6E | 5E | 2E | 1E | E0 | F2 | E4 | F6 | E8 | FA | EC | FE |

0169-013

**2FSK**

Left

Right

|   |   |
|---|---|
| 0 | 1 |
|---|---|

**4FSK**

Left

Right

|   |   |   |   |
|---|---|---|---|
| 0 | 1 | 3 | 2 |
|---|---|---|---|

0169-014

**8FSK**

Left

Right

|   |   |   |   |   |   |   |   |
|---|---|---|---|---|---|---|---|
| 0 | 1 | 2 | 3 | 7 | 6 | 5 | 4 |
|---|---|---|---|---|---|---|---|

0169-015

**16FSK**

Left

Right

|   |   |   |   |   |   |   |   |    |    |    |    |    |    |   |   |
|---|---|---|---|---|---|---|---|----|----|----|----|----|----|---|---|
| 0 | 1 | 2 | 3 | 4 | 5 | 6 | 7 | 15 | 14 | 13 | 12 | 11 | 10 | 9 | 8 |
|---|---|---|---|---|---|---|---|----|----|----|----|----|----|---|---|

0169-016

**C4FM**

Left

Right

|    |    |    |    |
|----|----|----|----|
| 11 | 10 | 00 | 01 |
|----|----|----|----|

0169-020

**CPM**

**Phase shift**  
(h = modulation index)

**Symbol value**  
(binary)

-3h

11

-h

10

+h

01

+3h

00

## Overview: User Defined Measurement and Reference Filters

The Modulation Parameters control tab for GP Digital Modulation displays enables you to load custom measurement and reference filters. If the existing filters do not meet your requirements, you can create your own filters for use in the measurement and reference settings. This section describes the structure of user filters and provides two examples of customized filters. See [User Filter File Format \(see page 533\)](#).

### Loading a User Measurement Filter

To load a your own measurement filter:

1. From the Modulation Params control tab (Settings > Modulation Params), click on the drop-down list for **Measurement Filter**.
2. Select one of the filter names that starts with **User**. This displays the **Manage user filters** window.
3. Enter a name for the filter in one of the **Name (editable)** boxes. This name will appear in the drop-down list on the Modulation Params tab, prefaced with **User**. The maximum number of characters for the filter name is 20.

4. Click the **Browse** button and navigate to the directory containing the filter you want to load. Select the filter and click **Open**. If you wish to use a filter that is not in the list, select **User other** and locate and open the file you wish to use.
5. Click **OK** to load the filter and return to the Modulation Params page.

## User Filter File Format

The filter file is selected on the Modulation Params control panel tab used by the GP Digital Modulation displays (Option 21 only). It stores the user-defined measurement or reference filter coefficient data in CSV format. The following figure shows the file structure.

```
# Rate
10
# FilterI,FilterQ
0.97321,0.01947
0.89559,0.04051
0.77497,0.05942
0.62333,0.07202
0.45524,0.07438
0.28614,0.06354
...
0.13045,0.01947
```

} Oversampling rate of the filter (samples/symbol)

} IQ pairs (1 to 1024) of the filter coefficient in time domain

0169-023

### User filter file structure

A filter file is a plain text file, in comma-separated-variable format. The file extension must be CSV.

The filter file contains the following variables:

**Rate.** Specifies the oversampling rate (the number of samples per symbol). The filter coefficient data will be interpolated by the specified rate.

**FilterI,FilterQ.** Specifies IQ pairs (1 to 1024) of the filter coefficient in time domain.

### Rules for Creating a Filter File

- A line beginning with “#” is a comment line.
- Enter a positive value for the oversampling rate.
- A decimal number can be expressed by fixed point or floating point. For example, 0.01 and 1.0E-2 are both valid.
- “0” (zero) and “,0” (comma zero) can be omitted. For example, “1.5,0”, “1.5,”, and “1.5” are equivalent.
- Lines with only a comma and blank lines are skipped.

**Example filters.** For your reference, two example filters, Raised Cosine and Gaussian, are shown here. Both filters contain 65 data points with an oversampling rate of 8.

**Raised Cosine ( $\alpha = 0.3$ )**

| (Row 1 to 18)     | (Row 19 to 36) | (Row 37 to 54) | (Row 55 to 68) |
|-------------------|----------------|----------------|----------------|
| # Rate            | 0.0383599,0    | 0.973215,0     | 0.0743803,0    |
| 8                 | 0,0            | 0.895591,0     | 0.0720253,0    |
| # FilterI,FilterQ | -0.047715,0    | 0.774975,0     | 0.0594205,0    |
| 0,0               | -0.0984502,0   | 0.623332,0     | 0.0405144,0    |
| -0.0062255,0      | -0.143898,0    | 0.455249,0     | 0.0194761,0    |
| -0.0136498,0      | -0.174718,0    | 0.286147,0     | 0,0            |
| -0.0209294,0      | -0.181776,0    | 0.130455,0     | -0.0151973,0   |
| -0.0263419,0      | -0.157502,0    | 0,0            | -0.0246357,0   |
| -0.0280807,0      | -0.0971877,0   | -0.0971877,0   | -0.0280807,0   |
| -0.0246357,0      | 0,0            | -0.157502,0    | -0.0263419,0   |
| -0.0151973,0      | 0.130455,0     | -0.181776,0    | -0.0209294,0   |
| 0,0               | 0.286147,0     | -0.174718,0    | -0.0136498,0   |
| 0.0194761,0       | 0.455249,0     | -0.143898,0    | -0.0062255,0   |
| 0.0405144,0       | 0.623332,0     | -0.0984502,0   | 0,0            |
| 0.0594205,0       | 0.774975,0     | -0.047715,0    |                |
| 0.0720253,0       | 0.895591,0     | 0,0            |                |
| 0.0743803,0       | 0.973215,0     | 0.0383599,0    |                |
| 0.063548,0        | 1,0            | 0.063548,0     |                |

**Gaussian (BT = 0.5)**

| (Row 1 to 18)     | (Row 19 to 36) | (Row 37 to 54) | (Row 55 to 68) |
|-------------------|----------------|----------------|----------------|
| # Rate            | 0.00191127,0   | 0.978572,0     | 0.000401796,0  |
| 8                 | 0.00390625,0   | 0.917004,0     | 0.000172633,0  |
| # FilterI,FilterQ | 0.00764509,0   | 0.822878,0     | 7.10E--05,0    |
| 2.33E-10,0        | 0.0143282,0    | 0.707107,0     | 2.80E-05,0     |
| 9.11E-10,0        | 0.0257149,0    | 0.581862,0     | 1.06E-05,0     |
| 3.42E-09,0        | 0.0441942,0    | 0.458502,0     | 3.81E-06,0     |
| 1.23E- 08,0       | 0.0727328,0    | 0.345977,0     | 1.32E-06,0     |
| 4.21E-08,0        | 0.114626,0     | 0.25,0         | 4.37E-07,0     |
| 1.39E-07,0        | 0.172989,0     | 0.172989,0     | 1.39E-07,0     |
| 4.37E-07,0        | 0.25,0         | 0.114626,0     | 4.21E-08,0     |
| 1.32E-06,0        | 0.345977,0     | 0.0727328,0    | 1.23E-08,0     |
| 3.81E-06,0        | 0.458502,0     | 0.0441942,0    | 3.42E-09,0     |
| 1.06E-05,0        | 0.581862,0     | 0.0257149,0    | 9.11E-10,0     |
| 2.80E-05,0        | 0.707107,0     | 0.0143282,0    | 2.33E-10,0     |
| 7.10E-05,0        | 0.822878,0     | 0.00764509,0   |                |
| 0.000172633,0     | 0.917004,0     | 0.00390625,0   |                |
| 0.000401796,0     | 0.978572,0     | 0.00191127,0   |                |
| 0.000895512,0     | 1,0            | 0.000895512,0  |                |




## Using Markers

Markers are indicators in the display that you can position on a trace to measure values for the X and Y axes, such as frequency, power, and time. A Marker always displays its position and, if enabled, will display the difference between its position and that of the Marker Reference (MR).


You can display up to five markers including the reference marker. Markers can all be placed on the same trace or they can be placed on different traces. There are three types of Markers: Reference, Delta, and Power Markers. The Marker Reference (labeled MR in the graph) makes absolute measurements and is also used for calculating differences when Delta or Power readouts are enabled. The Delta Markers (labeled M1 to M4 in the graph) are used to measure other points on the trace or the difference between the Marker Reference and the Delta marker. The Power Markers (labeled M1 to M4 in the graph) function the same way as the Delta Markers, except they show power density and integrated power density (dBm/Hz) instead of power level (dBm).

The following two tables show the appearance of the five types of marker readouts.


### Marker Reference Readout




### Absolute Marker Readout




### Delta Marker Readout



### Power Marker Readout



### DPX Bitmap Marker Readout



The following table shows the appearance of the marker indicators as they appear on the trace. Whichever marker is active will appear as a solid diamond.

## Marker Reference




## Absolute, Delta, and Power Marker



## Controlling Markers with the Touchscreen Actions Menu

In addition to controlling the marker actions from the front panel or screen menu items, you can use the touch screen actions menu to move markers or add and delete markers.




To use the Touchscreen Actions menu, touch the display and hold for one second, then remove your finger. You can also use a mouse to display the Touchscreen Action menu by clicking the right mouse button.

| Menu item       | Description   |
|-----------------|---|
| Marker to peak  | Moves the selected marker to the highest peak. If no marker is turned on, this control automatically adds a marker.               |
| Next Peak       | Moves the selected marker to the next peak. Choices are Next left, Next right, Next lower (absolute), and Next higher (absolute). |
| Add marker      | Defines a new marker located at the horizontal center of the graph.   |
| Delete marker   | Removes the last added marker.  |
| All markers off | Removes all markers.  |

## Measuring Signal Density, Frequency and Power on a DPX Bitmap Trace

To measure the signal density, frequency and power at a point on a DPX Bitmap trace:

1. Select **Markers > Define Markers**.
2. Select the **Marker Properties** tab.

3. Click **Add**. This displays a drop-down list under **MR** (Marker Reference), found inside the **Reference** box. The first marker defined is always designated the Marker Reference. Subsequently defined markers are Delta Markers for which readouts can be selected for Delta, Power, or Absolute. You can also select and adjust markers by clicking on an existing marker on a trace.
4. From the drop-down list, select **Bitmap**.
5. Click  to remove the Define Markers control panel.
6. Touch or click on the MR marker on the trace to activate the marker.
7. Drag the marker to the desired location on the bitmap.

---

**NOTE.** Delta markers on the DPX Bitmap trace can be placed on any point in the display, unlike markers in other displays that can be placed only on a trace. This enables you to measure the signal density, frequency, and power level at any point on the graph.


---



---

**NOTE.** Power markers can only be placed on a trace. To use Power markers in the DPX display, click the down arrow in the top left portion of the display labeled Bitmap and select a Trace from the drop down menu.

---




**Searching for peaks on the DPX bitmap trace.** To locate peaks on the DPX Bitmap trace, use the arrow buttons on the Marker toolbar. The left / right arrows move the selected marker to the next peak lower or higher in frequency. The up and down arrows search for peaks at the same frequency. The peaks on the DPX Bitmap trace are defined by the settings for the DPX Signal Density on the Markers > Define Peaks (see page 542) tab.

## Measuring Frequency and Power in the Spectrum Display

To measure the frequency and power at a point on a Spectrum trace:

1. Select **Markers > Define Markers**.
2. Select the **Marker Properties** tab.
3. Click **Add**. This displays a drop-down list under **MR** (Marker Reference), found inside the **Reference** box. The first marker defined is always designated the Marker Reference. Subsequently defined markers are Delta Markers for which readouts can be selected for Delta, Power, or Absolute. You can also select and adjust markers by clicking on an existing marker on a trace.
4. From the drop-down list, select the trace to which you want to assign the marker.



5. Click the close control panel button to remove the Define Markers control panel.
6. Click on the marker on the trace to activate that marker.
7. Use the knob or the arrow keys on the front panel to move the marker to the desired location on the trace. You can also click on the marker location text box in the Markers toolbar at the bottom of the screen and adjust it with your mouse wheel. The Peak button in the Markers toolbar and the arrow buttons to the right of it control marker peak searching on the trace.
8. Read the frequency and power level of the marker position on the display.
9. Read the signal density, frequency, and power level of the marker position on the display.
10. If Power markers are selected, read the point power density and the integrated power density.

## Marker Action Controls

Controls for enabling and moving markers and for initiating marker peak searches are found in several locations. There are buttons for a few of the most common marker activities on the front panel of the instrument, used along with the knob for adjusting marker positions. The Markers menu contains selections for peak searches and the Marker toolbar also has buttons for peak searches. All graphs that support the markers offer a pop-up menu for marker actions.

### Peak

Selecting **Peak** from the Markers menu moves the selected marker to the highest level peak within the acquisition record.

### Next Peak

Selecting Next Peak displays a submenu that enables you to move the selected marker to the next peak.

| Setting                | Description   |
|------------------------|---|
| Next Left              | Moves the selected marker to the next peak to the left of the current marker position.  |
| Next Right             | Moves the selected marker to the next peak to the right of the current marker position. |
| Next Lower (absolute)  | Moves the selected marker to the lower level peak (in absolute terms) on the trace.     |
| Next Higher (absolute) | Moves the selected marker to the higher level peak (in absolute terms) on the trace.    |

## Marker to Center Frequency

Changes the center frequency to match the frequency of the selected marker.


## Enabling Markers and Setting Marker Properties

The Define Markers Control Panel is used to enable markers and set their properties. You can set up to five markers including the marker reference. Markers are shown in most displays.

Markers have three types of on-screen readouts: Absolute, Delta, and Power. When **Readouts** is set to **Absolute**, each readout displays only the marker's position on the trace. In Frequency displays, this means the marker readout shows the frequency and power of the trace at the marker position. When **Readouts** is set to **Delta**, each delta marker (M1-M4) readout displays both the marker's position on the trace and the difference between its position and the position of the Reference Marker (MR). When **Readouts**

is set to **Power**, each delta marker (M1-M4) readout displays the marker's position on the trace, the difference between its position and the position of the Reference Marker (MR), the point power density, and the integrated power density.

## Defining Markers



1. Select **Markers > Define Markers** to display the Define Markers control panel.
2. Select **Add** to turn on the next marker. A drop-down list under the marker label allows you to assign the marker to a trace.

---


**NOTE.** The first marker defined will always be **MR**. The MR marker is the reference for delta marker readouts.

---

3. Select the trace to which the marker should be attached from the drop-down list.
4. Click **Add** to add additional markers.
5. Click on the **Readouts** drop down menu to select the readout to view. If you select Delta from the drop down menu, you can check the **dBc/Hz** box to measure noise in dBc/Hz and show that for the delta value. Checking this box if other readouts are selected will force the readout to Delta. You can read more about dBc/Hz mode [here \(see page 544\)](#).
6. Click the close button to remove the Define Markers control panel.

## Defining Peaks

You can specify two amplitude values that define peaks. For the DPX display, you can also define peaks based on signal density characteristics.



**Amplitude**

|                   |  |
|-------------------|--|
| Peak Threshold    | Peak Threshold specifies the level that the signal must exceed to be considered a peak.                              |
| Minimum Excursion | Minimum Excursion specifies how much the signal must decrease and then increase before another peak can be declared. |


**DPX Signal Density**

|                            |  |
|----------------------------|--|
| Density Threshold          | Density Threshold specifies the signal density (number of hits per displayed pixel) that the DPX bitmap must exceed to be considered a trigger event.  |
| Minimum Excursion          | Minimum Excursion specifies how much the signal density must decrease and increase again before another peak can be declared.  |
| Smoothing (pixels squared) | Smoothing specifies the number of pixels around the marker that are averaged together to reduce "noise" in the readout of signal density. The value of this control is the number of pixels on each side of the square area used for averaging. With Smoothing = 1, no averaging is done and the marker z-axis readout is the hit count (density) of a single pixel. Use this control to characterize how wide or narrow a range of pixels should be averaged to determine the signal density. |

1. Select **Markers > Define Markers** to display the Define Markers control panel.
2. Select the **Define Peaks** tab.
3. To define the level for Peak Threshold, enter a value in the Peak Threshold number entry box.
4. To define the amount the trace must dip, enter a value in the Peak Excursion number entry box.
5. Click the close button to remove the Define Markers control panel.


## Using the Markers Toolbar

**Front Panel: Markers**












**Application toolbar:** 

**Menu bar: View > Marker Toolbar**

Select **Marker Toolbar** to display or hide the Marker Toolbar in the application window. The Marker Toolbar enables you to operate existing markers or define new markers.



Enabling a marker or adjusting the position of a marker automatically opens the Markers toolbar.

| Icon / Readout  | Description   |
|---|---|
|    | Opens the Define Markers control panel.   |
|    | Selected marker readout. This readout shows which marker is selected. The pop-up menu allows you to choose the selected marker, add markers, and turn all markers off.  |
|    | Marker position controls. For frequency displays, this readout shows the marker position in Hertz. For time displays, this readout shows the marker position in seconds. The position of the selected marker can be changed by selecting the numeric readout and using the knob to adjust the value.                    |
|    | Changes the analyzer's Center Frequency to the frequency of the selected marker. Not selectable for time markers.   |
|    | Moves the marker to the highest peak on the signal. On displays that scale about zero on the vertical axis (for example, Magnitude Error, EVM, and Frequency vs. Time), the highest peak selected by the Peak button is an "absolute value", therefore, negative peaks are included in the search for the highest peak. |
|    | Moves the selected marker to the next peak to the <b>left</b> of the current position.  |
|    | Moves the selected marker to the next peak to the <b>right</b> of the current position.   |
|    | Moves the selected marker to the next <b>lower</b> peak value. The peak value here refers to the numeric value of the peak amplitude. Thus, when repeatedly moving the marker, it can move to the right or left depending on the location of the next lower value.  |
|    | Moves the selected marker to the next <b>higher</b> peak value. The peak value here refers to the numeric value of the peak amplitude. Thus, when repeatedly moving the marker, it can move to the right or left depending on the location of the next higher value.  |
|   | Displays/hides the marker table from the display.   |
|  | Removes the Marker Toolbar from the display.  |

## Measuring Noise Using Delta Markers in the Spectrum Display


In the Spectrum display, you can set Markers to dBc/Hz to measure noise on the trace. Markers in this mode operate just as they do in normal mode, but the readouts for the markers are in dBm/Hz and dBc/Hz.

dBm/Hz is the power in milliwatts referenced to a 1 Hz bandwidth. To make this measurement, the analyzer assumes that the measured signal is random noise. It then converts the measured power (made at any RBW) to the power that would be measured had a 1 Hz filter been applied for the measurement.


dBc/Hz represents dBm/Hz referenced to a carrier. Here, it is assumed that the carrier is a CW signal, and its signal level does not change when the RBW is changed, so the Reference Marker measurement on the carrier is unchanged from any other marker measurement. However, the delta marker values are converted to dBm/Hz, and then a difference value, in dBc/Hz, is calculated between each delta measurement and the reference.

### Measuring Noise

To measure noise on a Spectrum trace:

1. Select **Markers > Define** to display the Markers control panel.
2. Click **Add**. The first marker defined is always designated the Marker Reference. Subsequently defined markers are Delta Markers for which readouts can be selected for Delta, Power, or Absolute. You can also select and adjust markers by clicking on an existing marker on a trace.
3. Click **Add** again so that there are at least two markers defined.
4. Check that **Readouts** is set to Delta and check the **dBc/Hz** box.
5. If you have more than one trace defined, use the drop-down list for each marker to set it to the trace on which you want to measure noise.
6. Notice that Detection is set to **Average** for the trace you are using for this measurement.
7. Check that the Amplitude units are set to dBm (that is the default). If they are not, click the  (Analysis) icon, select the **Units** tab, and select **dBm**.
8. Click the Close button to remove the control panel.
9. Move the markers to the desired locations on the trace.
10. Read the frequency and power level for the selected marker in the upper corners of the display. To display the delta measurement in dBc/Hz, select the delta marker (M1, M2, M3 or M4) by clicking on it or pressing the front-panel Select button until it is selected.

The following image shows the Spectrum display with a Delta Marker.



The following image shows the Spectrum display with Power Markers.




# The Mask Test Tool

You can specify pass/fail and mask parameter conditions for the Spectrum, Spurious, DPX Spectrum, Noise Figure, Gain, Frequency Settling Time and Phase Settling Time displays. When these conditions are met, the instrument can perform actions such as stopping acquisitions or saving data.

## Mask Test Settings

**Menu Bar: Tools > Mask Test**

Selecting **Mask Test** displays the Mask Test control panel. These settings define test parameters and specify actions to be performed when the test conditions are met. Checking the **Enable Test** box sets the application to perform the test once the conditions are defined.




| Setting                                | Description  |
|--|--|
| Enable Test                            | Select to perform a test, then set the test conditions.      |
| <a href="#">Define (see page 547)</a>  | Specifies which result to test and what to test for.         |
| <a href="#">Actions (see page 553)</a> | Specifies the action to take when the test condition is met. |


## Define Tab (Mask Test)

The Define tab sets the parameters for tests. From this tab, you specify which result to test and what kind of violation to test for.

The following image shows the Define tab with *is great than* selected. After you select that test, you will enter the desired signal level. The procedure is the same if you select *is less than*.



The following image shows the Define tab with *is outside mask* selected. After you select that test, you need to click the Edit limits button and set the desired test limits. The procedure is the same if you select *is inside mask*.




| Setting   | Description   |
|-----------|---|
| Search in | Specifies which result to test.   |
| Test for  | Specifies what to test for. You can specify a test based on a signal level ("is greater than", "is less than") or a mask ("is outside mask", "is inside mask"). Options for Spurious and Settling Time are Pass and Fail. |

Search In

The possible choices for **Search in** include traces from Spectrum, DPX, Noise Figure, Spurious, and Setting Time displays. The available choices are Trace 1, Trace 2, Trace 3, Math , Spectrogram Trace, Spurious, and Settling Time. The available choices include only results from displays that are currently open.

**NOTE.** If you select a result that is not the selected trace or result in the target display, you will not see the results of the test. To see the results, select the trace from the drop-down menu in the target measurement display and check the Show box:



Test For

The **Test for** setting has selections that vary based on which display's results you are testing.


Pass/Fail Tests for Spurious and Settling Time

The Frequency Settling Time, Phase Settling Time measurements, and Spurious measurements test functions provide pass/fail results.

Greater Than/Less Than Tests for Spectrum


If you select a greater than/less than test, you also specify the level that defines a violation. When you select either **is greater than** or **is less than**, a text entry box appears to the right of the drop-down list. Use the text entry box to specify the signal level you wish to test for. While not as flexible as mask testing, this type of test is quicker to set up.

The following figure shows the results of an **is greater than** test. The vertical red bar highlights results that match the test definition.




### Outside/Inside Mask Tests

If you specify a mask-based test, then you need to edit the mask to specify the levels that define a violation. When you select **is outside mask** or **is inside mask**, an **Edit limits** button is displayed. Click the **Edit limits** button to display the Mask Test Editor window.




**To set up and save mask test limits for a test.** Perform the following procedure to set up a test using the mask limits.


1. After choosing the desired **Search in** and **Test for** items, click **Edit limits** to display the Mask Test Editor window.
2. Click the **New** button to create a table. This clears the existing points and loads the default table. You can also click **Open** to open an existing table.
3. To edit values, add points, or delete points in a table:
  - a. To edit an existing value, double-click on the cell you want to edit and enter the desired value. The active point shows as a red point on the plot.



- b. To add a new point, check the box next to the target mask (located below the **Auto draw** button).
- c. Select **Add points** located below the **Open** button.
- d. Select the target mask (Upper Mask Table or Lower Mask Table) from the drop-down menu.
- e. Click in the desired location on the plot to add the point.



- f. To use the auto draw feature to automatically place points on the chosen mask, enable the desired mask from the drop-down menu and then click the **Auto draw** button.




---

**NOTE.** Units may be changed for measurements other than Noise Figure in Setup > Analysis > Units.

---

4. To delete a point from the table, select the point to be removed, and click the **X** button.
5. To save the mask to a file for later recall, click **Save As**.
6. From the Save As dialog, name the file and save it in the desired location.

---


**NOTE.** Masks are saved in XML format with a .msk extension.

---


7. To dismiss the Mask Test Editor window, click **Done**.

### Mask Test Result Examples


The following figure shows the results of a Spectrum **is outside mask** test with Upper Mask enabled. The vertical red bar highlights results that have exceeded the test limits. The shaded region shows the mask area.



The following figure shows the results of a passing Noise Figure **is outside mask** test with Upper Mask enabled. Markers are also in use.




The following figure shows the results of a failing Noise Figure **is outside mask** test with Lower Mask and Upper Mask enabled. The red area highlights the area of failure, which occurs at the lower mask. Markers are also in use.



## Actions Tab

The Actions tab allows you to trigger the application when a signal in the frequency domain violates the mask.



| Setting               | Description  |
|-----------------------|--|
| During Run or Replay  | Actions specified here will occur in either Run mode or Replay mode.   |
| Beep                  | The analyzer beeps when the test condition is met.   |
| Stop                  | The analyzer stops when the test condition is met.   |
| During Run Only       | Actions specified here are taken only during Run mode (while acquiring live data).   |
| Save acquisition data | Saves acquisition data to a file when the test condition is met. Use the drop-down list to specify the format of the saved data. The available file formats are: <a href="#">TIQ, CSV, and MAT (see page 607)</a> .  |
| Save trace            | Saves Trace data to a file when the test condition is met.   |
| Save picture          | Saves a screen capture to a file when the test condition is met. Use the drop-down list to specify the format of the saved picture. The available file formats are: <a href="#">PNG, JPG, and BMP (see page 607)</a> . Note that no trace will be saved if the tested trace isn't a saveable trace type. For example, a Spurious trace is not saveable.  |
| Max files             | Specifies the number of times a test action stores a file. After this limit is reached, no more files are saved. The instrument will continue to run, but no additional files are saved when test conditions are met.<br><br>Keep in mind when setting this number that picture files are counted as part of the total number of files. For example, if you set Max files to 100, the instrument will save 100 acquisitions if only acquisitions are saved or only pictures are saved. But, if both acquisitions and pictures are saved, then 50 acquisitions and 50 pictures will be saved. |

### AutoSave File Naming

When one of the AutoSave actions is enabled, the name of the saved file is automatically incremented even if the **Automatically generate filenames** option (Tools > Options > Save and Export) is not enabled. When the file is saved, it will be saved to *C:\<instrument name>Files\*.




## Analysis Settings

**Menu Bar:** Setup > Analysis

**Application Toolbar:** 


The Analysis control panel provides access to settings that are used by all displays. These settings affect only post processing and they do not control hardware settings.



| Setting                                      | Description   |
|--|---|
| <a href="#">Analysis Time (see page 555)</a> | Specifies the length of time to use in measurements.  |
| <a href="#">Spectrum Time (see page 557)</a> | Specifies whether the Spectrum Analysis display uses the same Analysis Time parameters as all the other displays or if it uses a different Offset and Length. |
| <a href="#">Frequency (see page 557)</a>     | Specifies the measurement frequency (center frequency).   |
| <a href="#">Units (see page 561)</a>         | Specifies the Power units for all displays.   |

## Analysis Time Tab

The Analysis Time tab contains parameters that define the portion of the acquisition record that is used for analysis.



| Setting             | Description   |
|---------------------|---|
| Analysis Offset     | Specifies the location of the first time sample to use in measurements.   |
| Auto                | When enabled, causes the instrument to set the Analysis Offset value based on the requirements of the selected display.   |
| Analysis Length     | Specifies the length of time to use in measurements.  |
| Auto                | When enabled, causes the instrument to set the Analysis Length value based on the requirements of the selected display.   |
| Time Zero Reference | Specifies the zero point for the analysis time.   |
| Actual              | This is a displayed value, not a setting. It is the Analysis Length (time) being used by the analyzer; this value may not match the Analysis Length requested (in manual mode). |

### Analysis Offset

Use analysis offset to specify where measurements begin. Range: 0 to [(end of acquisition) - Analysis Length)]. Resolution: 1 effective sample (or symbol).

### Analysis Length

Use the analysis length to specify how long a period of time is analyzed by a measurement. After you enter a value, this box changes to show the actual value in use, which is constrained by Acquisition Time. This setting is not available when Auto is checked. Range: minimum value depends on modulation type to Acquisition Length. Resolution: 1 effective sample (or symbol).

### Time Zero Reference

The analysis offset is measured from this point. Choices are: Acquisition Start or Trigger Point.

| Parameter         | Description  |
|-------------------|--|
| Acquisition Start | Offset is measured from the point at which acquisition begins. |
| Trigger Point     | Offset is measured from the trigger point.                     |

## Spectrum Time Tab

This tab controls whether the Spectrum display uses the same Analysis Time parameters as all the other views or uses a different Offset and Length. These settings do not apply to the DPX display because it has no acquisition record or analysis time.


| Settings  | Description   |
|---|---|
| <b>Spectrum Time Mode</b>   |   |
| Independent   | Causes the spectrum analysis views to use the settings unique to those displays.  |
| Use Analysis Time settings  | Causes the spectrum analysis views to use the settings on the Analysis Time tab.  |
| <b>Spectrum Time (only available when <i>Independent</i> is selected)</b> |   |
| Spectrum Offset   | Sets the beginning of Spectrum Time with respect to the selected time reference point (selectable in the Analysis Time tab as either Acquisition Start or Trigger).   |
| Spectrum Length   | The amount of data, in terms of time, from which spectrum traces are computed.  |
| Auto  | When enabled, causes the instrument to set the Spectrum Length value based on the RBW setting.  |
| Actual  | This is a displayed value, not a setting. It is the Spectrum Length (time) being used by the analyzer; this value may not match the Spectrum Length requested (in manual mode). The actual spectrum length is always an integer multiple of the time needed to support the RBW value. |

## Frequency Tab


The Frequency tab specifies two frequency values: the Measurement Frequency and the Spectrum Center Frequency. The Measurement Frequency is the frequency at which most displays take measurements. The Spectrum Center Frequency is the center frequency used by the Spectrum, DPX Spectrum, Spectrogram and Time Overview displays.

| Analysis Time   | Spectrum Time | Frequency | Units |
|---|---------------|-----------|-------|
| Measurement Frequency: 1.500000 GHz   |               |           |       |
| <input checked="" type="checkbox"/> Lock Center Frequency of Spectrum displays to Measurement Frequency |               |           |       |
| Spectrum Center Frequency: 1.500000 GHz   |               |           |       |


The following screen capture shows a display with both the Measurement Frequency and the Spectrum Center Frequency locked together.



Normally, the Measurement Frequency and the Spectrum Center Frequency are locked together so that both have the same setting. But in some situations, for example, where a signal contains a set of channels, it is useful to unlock the Measurement Frequency from the Spectrum Center Frequency. When the Spectrum Center Frequency is unlocked from the Measurement Frequency, you can adjust the Measurement Frequency so that measurements can be taken at different frequencies without resetting the center frequency. The following screen capture shows the magenta-colored measurement frequency indicator still located at the center frequency.



The following screen capture shows the measurement frequency indicator located at 2.0402 GHz while the Spectrum Center Frequency is still located at 2.0000 GHz.



You can drag the Measurement Frequency indicator on the screen to set the measurement frequency. Note the base of the Measurement Frequency indicator. The width of the box indicates the widest measurement bandwidth in use among the open displays. You can see how the width of this box changes with the measurement bandwidth by, in this example, adjusting the Measurement BW setting for the Constellation

display (Settings > Freq & BW tab). As you adjust the setting, you will see how the width of the box at the base of the Measurement Frequency Indicator changes.

The Measurement Frequency indicator is useful for interpreting system behavior when MeasFreq is unlocked. If a measurement has a wide bandwidth relative to the spectrum span, and the Measurement Frequency is far from spectrum center, the measurement is likely to fail because its required frequency range exceeds the frequency range of the available data. In such a case, the navigation control will show that the measurement bandwidth extends outside the Spectrum's span.

There are interactions between frequency unlocking and RF & IF Optimization (see the [Amplitude control panel \(see page 569\)](#)). When **Best for multiple displays** is the selected optimization, the instrument is allowed to use its full bandwidth to meet the needs of all open displays. This is the most user-friendly optimization because it decreases the number of **Acq BW too small** errors, but it can increase noise and slightly decrease measurement accuracies. For all other optimization types, the instrument optimizes the acquisition bandwidth for the selected display, improving measurement quality somewhat, but reducing concurrent measurement capability.

There are also interactions with trigger settings. When the Spectrum Center Frequency is unlocked from Measurement Frequency, the RF triggers (Trigger Source = RF Input) can tune to either Spectrum Center Frequency or Measurement Frequency. When a spectrum display is selected, the trigger frequency is the same as the Spectrum Center Frequency. When one of the other displays that uses Measurement Frequency is selected, the trigger is tuned to the Measurement Frequency also. This allows you to trigger on the signal you are measuring. But, there is an exception: when the selected RF & IF Optimization is **Best for multiple displays**, the trigger is always tuned to the Spectrum Center Frequency.

The reason the trigger frequency is affected by Measurement Frequency and RF & IF Optimization, is that these functions control how the acquisition is tuned. The optimization **Best for multiple displays** keeps the acquisition centered about the Spectrum Center Frequency at all times. To accommodate off-center Measurement Frequencies, it just widens the acquisition bandwidth. Other optimizations tune the acquisition frequency to match that of whichever display is currently selected. The RF trigger module receives the same acquisition data as all the measurements, tuned to the center of the current acquisition bandwidth.

## Units Tab

The Units tab specifies the global Amplitude units for all the views in the analysis window.





---

**NOTE.** If you are using an External Loss Table for antenna corrections ( $Ampl > External\ Gain/Loss\ Correction$ ), set the Amplitude units to  $dBuA/m$  or  $dBuV/m$ .

---

## Replay Overview

The Replay function enables you to reanalyze data with different settings and even different measurements. You can replay all the acquisitions in memory, a single acquisition, a single frame within an acquisition (if Fast Frame is enabled), or any contiguous set of data records from acquisition history. You can replay the DPX Spectra that make up a DPXogram display.



### Displaying the Replay Control Panel

To display the Replay control panel:

- From the menu bar, select **Replay > Select data records from history**.

### Selecting the Data Type to Replay

You can choose to replay either acquisition data or DPX spectra. Choose **Acq Data** to replay any acquisitions except DPXogram acquisitions. Choose **DPX Spectra** to replay a DPXogram acquisition.

## Select Data Records Tab

The Select data records tab is used to select which data records to replay. A data record is the smallest unit that can be replayed. Note that what constitutes a data record can vary. If FastFrame is not enabled, a data record consists of a single acquisition. If FastFrame is enabled, each acquisition can contain multiple frames, and a data record consists of a single frame.

| Select data records  | Acquisition Info         | Replay Speed |
|--|--------------------------|--------------|
| <b>Select acquisitions</b> <input type="button" value="Select All"/> |                          |              |
| 1 : 7/25/2011 17:39:29.72  |                          |              |
| Start: 1   | 7/25/2011<br>17:39:29.72 |              |
| Stop: 304  | 7/25/2011<br>17:39:33.79 |              |
| 304 : 7/25/2011 17:39:33.79  |                          |              |

Without FastFrame enabled

| Select data records  | Acquisition Info        | Replay Speed |
|--|-------------------------|--------------|
| <b>Select acquisitions</b> <input type="button" value="Select All"/> |                         |              |
| 1 : 3/10/2010 4:47:40.22   |                         |              |
| Start: 1   | 3/10/2010<br>4:47:40.22 |              |
| Stop: 1  | 3/10/2010<br>4:47:40.22 |              |
| 1 : 3/10/2010 4:47:40.22   |                         |              |
| <b>Select frames</b>   |                         |              |
| 1 : 3/10/2010 4:47:40.22   |                         |              |
| Start: 1   | 3/10/2010<br>4:47:40.22 |              |
| Stop: 10   | 3/10/2010<br>4:47:51.74 |              |
| 10 : 3/10/2010 4:47:51.74  |                         |              |

With FastFrame enabled

**Date and Time Stamps.** The line that appears above each Start box (one for acquisitions and one for frames when FastFrame is enabled) is a date stamp that displays the date and time the first record in the current acquisition was acquired. The line that appears below each Stop box is a date stamp that displays the date and time the last record in the current acquisition was acquired.

The line that appears to the right of each Start box displays the date and time the selected acquisition or frame was acquired. The line that appears to the right of each Stop box displays the date and time the selected record or frame in the current acquisition was acquired.

**Select All button.** Selecting Select All resets the Start and Stop values. The Start value is reset to the first acquisition or frame of the current acquisition. The Stop value is reset to the last acquisition or frame in the current acquisition.

## Select DPX Spectra Tab

The Select DPX Spectra tab is used to select which DPXogram spectra to replay. A spectra is the smallest DPXogram unit that can be replayed. It is equivalent to one line in the DPXogram display.

|  |                       |                  |              |
|--|-----------------------|------------------|--------------|
| Select data records                    | Select DPX Spectra    | Acquisition Info | Replay Speed |
| Select spectra <span>Select All</span> |                       |                  |              |
| 1 : 9/29/2011 17:36:53.05              |                       |                  |              |
| Start: 1                               | 9/29/2011 17:36:53.05 |                  |              |
| Stop: 602                              | 9/29/2011 17:37:7.82  |                  |              |
| 602 : 9/29/2011 17:37:7.82             |                       |                  |              |

**Date and Time Stamps.** The line that appears above the Start box is a date stamp that displays the date and time the first DPX spectra in the current acquisition was acquired. The line that appears below the Stop box is a date stamp that displays the date and time the last DPX Spectra in the current acquisition was acquired.

The line that appears to the right of each Start box displays the date and time the selected DPX Spectra was acquired. The line that appears to the right of each Stop box displays the date and time the selected DPX Spectra in the current acquisition was acquired.

**Select All button.** Selecting Select All resets the Start and Stop values. The Start value is reset to the first DPX Spectra of the current acquisition. The Stop value is reset to the last DPX Spectra in the current acquisition.

### Acquisition Info Tab

Displays information about acquisition settings and sets the number of decimal places used for displaying the time stamp.

|                             |                    |                     |                        |
|-----------------------------|--------------------|---------------------|------------------------|
| Select data records         | Select DPX Spectra | Acquisition Info    | Replay Speed           |
| Timestamp decimal places: 2 |                    |                     |                        |
|                             |                    | Delete all Acq data | Delete all DPX spectra |
| Acq Center Freq:            | 1.90 GHz           | Preamp:             | Off                    |
| Acq Bandwidth:              | 85.0 MHz           | RF Atten:           | 15.00 dB               |
| Sampling Rate:              | 150 MHz            | Ref Level:          | 0.00 dBm               |
| Dither:                     | On                 | Acq Length:         | 5.19 us                |

### Replay Speed Tab

Sets the speed at which data records are replayed.

|                              |                    |                  |              |
|------------------------------|--------------------|------------------|--------------|
| Select data records          | Select DPX Spectra | Acquisition Info | Replay Speed |
| Slow <span>—————</span> Fast |                    |                  |              |

## Replay Menu

The Replay menu provides controls that let you choose how to replay acquisitions. The replay function enables you to, in effect, “rerun” an analysis while applying different measurements to the same set of acquisition data.

| Menu item                     | Description   |
|-------------------------------|---|
| Acq Data                      | Select Acq Data to replay acquisitions.   |
| DPX Spectra                   | Select DPX Spectra to replay a DPXogram display.  |
| Replay all selected records   | Replays the sequence of records specified by Start and Stop on the Select data records tab.   |
| Loop overall selected records | Replays the sequence of records specified by Start and Stop on the Select data records tab continuously.  |
| Replay current record         | Replays the currently selected acquisition (or frame).  |
| Replay from selected          | Displays a submenu that you use to specify which records are to be replayed.  |
| First record                  | Replays the first record within the selected set.   |
| Previous                      | Replays the previous record within the selected set.  |
| Next                          | Replays the next record within the selected set.  |
| Last record                   | Replays the last record within the selected set.  |
| Pause                         | Suspends replay of the data records as soon as the current record's replay action is completed. Press Pause again to begin replay with the next record in the sequence. |
| Stop                          | Halts the replay of acquisitions. If replay is started after Stop has been selected, replay starts from the first acquisition.  |
| Select all                    | Selects all acquisitions for replay.  |
| Select records from history   | Displays the Select data records tab of the Replay control panel. Use the Select data records tab to specify which acquisitions and frames you would like to replay.    |
| Replay toolbar                | Displays or hides the Replay toolbar.   |

## Acq Data

Selecting **Acq Data** selects acquisition data as the source for replay. Selecting Acq Data does not start replay, it only selects the type of acquisition data that will be replayed.

## DPX Spectra

Selecting **DPX Spectra** selects DPX spectra as the source for replay. Selecting DPX Spectra does not start replay, it only selects the type of acquisition data that will be replayed.

## Replay All Selected Records

Selecting **Replay all selected records** replays all the selected data records. The set of selected records may comprise a single record, all records in acquisition history, or a subset of the records in history.

## Replay Current Record

Selecting **Replay current record** replays the current data record. You can identify the current acquisition record by looking at the Replay toolbar. The first number to the right of the Select button identifies the current data record. For example, if the number is 2:10, it means the current record is the tenth frame of the second acquisition in history.

## Replay from Selected

Select **Replay from selected** to replay records as selected from the submenu. The records replayed can be from the acquisition memory (history) or from a saved acquisition data file that has already been recalled as the current acquisition data.

## Pause

Select Pause to suspend playback. Selecting Pause again restarts the replay at the point it was paused.

## Stop

Select **Stop** to halt the replay of data. Selecting any Replay action restarts replay of records from the beginning.

## Select All

Select **Select all** to select all data records for replay.

## Select Records from History










Selecting **Select records from history** displays the **Select data records** tab of the **Replay** control panel. The Select data records tab allows you to specify which records in the acquisition history will be used when the Replay button is selected.



## Replay Toolbar

Displays or hides the Replay toolbar that appears below the main tool bar.



Replay toolbar

| Item  | Description   |
|---|---|
| Replay  | Selects data type to be replayed. Select DPX Spectra to replay a DPXogram. Select Acq Data for all other types of replay.   |
|  | Replays the first record in the selected set.   |
|  | Replays the previous record in the selected set.  |
|  | Replays the current record in the selected set.   |
|  | Replays the next record in the selected set.  |
|  | Replays the last record in the selected set.  |
|  | Replays all records in the selected set.  |
|  | Replays all records in the selected set continuously until stopped.   |
|  | Pauses replay. Pressing pause suspends replay with the current record. Selecting pause again starts Replay with the next record.  |
|  | Stops replay. Starting any replay action after pressing stop starts a new Replay action rather than continuing from the record at which the previous replay action was stopped.                       |
| Select All  | Pressing <b>Select all</b> selects all records in history for replay. Selecting Select all resets the <b>Start</b> and <b>Stop</b> values on the Select data records tab of the Replay control panel. |

| Item  | Description   |
|---|---|
|  | Pressing <b>Select</b> displays the Select data records tab of the Replay control panel. The Select data records tab allows you to select records from acquisition history for replay.  |
| 1:10      3/10/2010 4:47:51.74  | This readout shows information about the data record being replayed. The information displayed shows the acquisition and frame number and time stamp for the current data record.   |
|  | The Info button displays the Acquisition Info tab of the Replay control panel. The Acquisition Info tab displays information about the acquisition data such as acquisition bandwidth, sampling rate and acquisition length. All data records in the acquisition history were acquired with identical parameters. When any of these parameters are changed, all records in history are deleted as soon as the first record acquired under the new parameter values is received. |




## Amplitude Settings

**Menu Bar:** Setup > Amplitude

**Application Toolbar:** 


The Amplitude control panel provides access to power-related settings that are used by all displays within the analyzer application.



| Setting  | Description   |
|--|---|
| <a href="#">Internal Settings (see page 569)</a>             | Sets the parameters that control the acquisition of the signal.   |
| <a href="#">External Gain/Loss Correction (see page 572)</a> | Specifies whether a correction is applied to the signal to compensate for the use of external equipment.  |
| <a href="#">External Probe Correction (see page 576)</a>     | Specifies whether a correction is applied to the signal if the analyzer detects a probe is connected to the instrument. (Probes are connected to the analyzer through the RTPA2A Real-Time Spectrum Analyzer TekConnect Probe Adapter.) |

## Internal Settings Tab

On the Internal Settings tab, you can specify the Reference Level for the signal analyzer's RF front end, the settings for the internal attenuator, enable/disable the optional internal preamp, and specify the RF & IF Optimization.



## Internal Attenuator

The Internal Attenuator setting can be set to Auto or manual. When Auto mode is enabled, the attenuation setting is displayed but is not adjustable. When Auto is disabled, the attenuator operates in manual mode, allowing you to set the internal attenuation. The internal attenuation setting range is 0 - 75 dB.

## Internal Preamp

If your instrument has the optional internal preamp installed, use this setting to switch the preamp on or off.

---

**NOTE.** *To ensure the best accuracy, turn on the Internal Preamp and allow it to run for 20 minutes to thermally stabilize before taking critical measurements.*

---

## Monitor Attenuator Use

The input attenuator relays are rated for several million cycles. Some measurement settings result in high rates of attenuator switching and could prematurely age the relays. Enabling this function causes the instrument to monitor the attenuator switching rate. When a high rate of switching is maintained for more than an hour, the instrument will halt acquisitions and display a message about the condition. To restart acquisitions, press Run. If the check box is not selected, the monitoring function is disabled.

---

**NOTE.** *This feature is provided to prevent accidental long-term high-rate attenuator switching (thus extending attenuator life). If you need to run the instrument for extended period, such as overnight or over a weekend, disable the setting to prevent the halting of acquisitions.*

---

## RF & IF Optimization

Specifies how the gain and bandwidth should be optimized (in the RF and IF stages of the front end).

You can select RF & IF optimizations to achieve the best dynamic range, fastest sweep speed, or minimum displayed noise. These choices affect the RF and IF attenuation levels, the selection of the acquisition bandwidth, digitizer and the resolution of the mathematical operation performed on the acquisitions used in varying conditions of span, reference level and resolution bandwidth.

The following paragraphs describe the effects of choosing each of the available choices.

**Best for multiple displays (Default).** This setting is the instrument default and is set as part of the Preset function. When this is selected, the instrument considers all the open displays and chooses an Acquisition Bandwidth wide enough to satisfy as many measurements as possible. Instead of optimizing the measurement performance of the selected display, this setting decreases the incidence of non-selected displays being unable to run because the acquisition bandwidth is set too small by the selected display.

**Auto.** The Auto setting is an optimized trade-off of the competing performance choices.

**Minimize sweep time.** Selecting Minimize Sweep Time causes the instrument to use the widest available acquisition bandwidth when sweeping across spans greater than the maximum available real-time span. For instruments without Option 110 (110 MHz acquisition BW), this control has no effect. For instruments

with Option 110, the swept update rate is improved by approximately 3X compared to all other RF/IF optimization choices.

**Maximize dynamic range.** Selecting Maximize Dynamic Range has several effects:

- Sets RF attenuation to allow higher drive levels at the RF mixer.
- Sets IF gain levels to optimize the IF for the higher signal levels that result from allowing higher drive levels at the mixer.
- Selects the acquisition path used in the instrument taking into account the user-specified Resolution Bandwidth (RBW). When an RBW of less than 30 kHz is selected, the instrument uses a narrow-band (312.5 kHz) acquisition path, regardless of the selected span or measurement bandwidth. This results in the best rejection of out-of-band signals in the IF stage of the analyzer, but comes at the cost of slower update rates for all signal analyzer spans and measurement bandwidths greater than 312.5 kHz.
- Increases the resolution of the mathematics used in calculations of all signal results from 16 bits to 28 bits whenever span or measurement bandwidth is set to  $\leq 312.5$  kHz. This is made possible by the high levels of decimation applied to signals processed in these narrow bandwidths.

**Minimize noise.** This mode increases the system gain at points past the RF mixer. This has the effect of reducing the displayed average noise level.

The table below summarizes the conditions.

| Settings                   | RF auto attenuation<br>(min 10 dB, max 55 dB) | Selected Acquisition Bandwidth   | IF Gain/Loss settings  | Notes and Trade-offs  |
|----------------------------|---|--|--|---|
| Best for Multiple Displays | Ref Level +15 dB                              | Acquisition Bandwidth is set to the value required to service as many of the open displays as possible.  | Neutral  | Swept Speed, Dynamic Range, Noise Floor   |
| Auto                       | Ref Level +15 dB                              | Driven by the measurement bandwidth of the selected display. If span is greater than the available acquisition BW, unit sweeps using 40 MHz acquisitions   | Neutral  | Swept Speed, Dynamic Range, Noise Floor may not be optimal  |
| Minimize Sweep Time        | Ref Level +15 dB                              | Unit uses widest available acquisition BW for spans greater than maximum acquisition BW.   | Neutral  | Instruments equipped with Option 110 (110 MHz), swept update rate is approximately 3 times faster than Auto. No effect in units without Option 110. |
| Maximize Dynamic Range     | Ref Level +20 dB                              | Driven by span and user-selected RBW when a spectrum-type display is selected, or by measurement bandwidth and filter bandwidth when other display types are selected. For RBW < 30 kHz, the high dynamic range 312.5 kHz path is used regardless of selected span or measurement bandwidth. Sweeps are made in 40 MHz acquisitions. | IF attenuation increased to improve dynamic range. This causes higher noise level. | Might result in very slow sweep speeds in wide spans when RBW < 30 kHz.   |
| Minimize Noise             | Ref Level +15 dB                              | Same as Auto   | Set to minimize noise level.   | Might result in earlier IF overload under some signal conditions.   |

## External Gain/Loss Correction Tab

The External Gain/Loss Correction tab allows you to apply a correction to a signal to compensate for the use of external equipment, such as an amplifier or attenuator. You can specify whether the corrections on this tab apply to all acquired data or only to a set group of displays.

## External Gain Value

Use the External Gain Value setting to apply a flat gain/loss correction to the signal. Positive values represent a gain and negative values represent a loss. The range is  $-50$  to  $+30$  dB. Resolution is 0.1 dB.

---

**NOTE.** Selecting Preset clears the check box, but it does not change the value.

---

## Apply External Corrections To

Use the two selections to specify whether the enabled external loss tables are applied to all acquired *data* or only to *traces* in the Spectrum, Spectrogram, Spurious, and Amplitude vs Time displays.

When choosing whether to correct sample data or traces, keep the following in mind:

- If sample data is acquired while data correction is enabled, that data record is permanently corrected. If the acquisition is saved and recalled, the recalled acquisition includes corrections. Corrections are not saved separately from the raw data; they are used during the acquisition process to adjust the data values.
- Trace corrections can be applied at any time, as they are part of the measurement's computations. Using Replay will apply trace corrections to a trace if this control is enabled (and an external loss table is specified and enabled).
- Be aware that if you recall acquisition data that was saved with data corrections and you use Replay with trace corrections selected and an external loss table enabled, the trace will be calculated using corrected data and also corrected with trace corrections. This probably creates incorrect traces.
- When only Traces are being corrected, triggering operates on uncorrected acquisition data. Frequency masks and trigger level lines in the displays are also not corrected, so there is a visual discrepancy between data traces and trigger indicators in the graphs under this condition.
- When only Traces are being corrected, triggering operates on uncorrected acquisition data. Frequency masks and trigger level lines in the displays are also not corrected, so there is a visual discrepancy between data traces and trigger indicators in the graphs under this condition.

## Trace Corrections Versus Data Corrections

Because of the real-time nature of the instrument and desire to frequently reanalyze stored acquisitions, it is possible to apply amplitude corrections in real time. Applying real-time corrections to the incoming data means that, the correction factors present during the acquisition are already part of the data when

the data is stored and then reanalyzed. This also allows the DPX display to maintain its update rate while presenting an amplitude-corrected display.

Real-time gain/loss factors are applied by creating a correction filter that is applied by the same digital signal processing hardware used to correct for calibration and alignment factors in the instrument. This process does have limitations. In creating a digital filter to represent the gain/loss table, a limited number of filter taps are used in order for the filter to be practically realizable. If your amplitude corrections contain discontinuities or extremely sharp transitions, the filter used will contain errors and ringing.

If the above is the case, we recommend that you use trace corrections. These are very similar to a conventional spectrum analyzer, and may be applied to the spectrum, spectrogram, spurious search and amplitude vs. time displays. These traces may contain transitions of any type within the maximum and minimum correction values allowed in the tables.

## External Loss Tables

Use an External Loss Table to apply a frequency dependent gain/loss correction to the signal. When an External Loss Table is selected, the analyzer adjusts the signal according to the values in the table. An external loss table allows you to compensate the signal level for variations in cable loss, antenna frequency response or preamp frequency response.

You can create external loss tables from the analyzer application and save them as files. External loss tables are saved as plain text files in CSV (Comma Separated Value) format. The tables have a CSV file extension.

You can change the title for each table. However, note that the title is only a label. It is not tied to the file name of any table you may have loaded or saved.

---

**NOTE.** *If you are using an External Loss Table for antenna corrections, be sure to set the Amplitude units (Setup > Analysis > Units) to dBuA/m or dBuV/m.*


---

### Creating an External Loss Table

You can create an external loss table with the analyzer application. No external application is required.

To create an external loss table:

1. Click **Edit.** to display the Table Editor.



2. Click **New Table** to create a table. This clears all entries in the existing table and loads the default table.
3. To edit values in the table:
  - a. To edit an existing value, double-click on the cell you want to edit.
  - b. To add a new point to the table, click on **Add Point**. Edit the values as required.
  - c. Click **Apply** if you want to test your table without closing the editor (the table must be enabled with the check box in the control panel and measurements must already be running).

To specify a frequency multiplier, you can type K, M, or G.

4. To delete a point from the table, select the point to be removed, and click **Delete**.
5. Select the Frequency Interpolation method:
  - Select Linear when the frequency scale of the spectrum or spurious measurements are linear.
  - Select Log when the frequency scale of the spectrum is logarithmic.
6. When you have entered all the necessary values, click **Save As** (to save the table in a file) or click **OK**.
7. From the Save As dialog, name the file and save it in the desired location.

External loss tables are saved in CSV (Comma Separated Value) format.

### Loading an External Loss Table


To load an external loss table file:

1. Click **Edit** to display the Table Editor.
2. Click **Load** to display the Open dialog box.

3. Navigate to the location of the desired file, select it and click Open.
4. Click **OK**.

## External Probe Correction Tab

The External Probe Correction tab is used to specify whether a correction factor is applied to the signal when a Tektronix P7000 Series probe is connected to the analyzer through the Tektronix RTPA2A Real-Time Spectrum Analyzer TekConnect Probe Adapter.




Internal Settings | External Gain/Loss Correction | External Probe Correction

Probe connected

☒ Correct data for probe attenuation: 20.2 dB

The analyzer can detect when an RTPA2A TekConnect probe adapter is connected to the instrument. The only setting available on the tab specifies whether a correction is applied to the signal or not. The amount of correction is specific to the probe and is communicated to the signal analyzer through the RTPA2A TekConnect probe adapter.

If the signal analyzer does not detect a probe adapter, the setting to correct for a probe is dimmed and **No probe detected** is displayed on the tab.



Internal Settings | External Gain/Loss Correction | External Probe Correction

No probe detected

☐ Correct data for probe attenuation: 0.0 dB

## Continuous Versus Single Sequence

### Menu Bar: **Run > Single Sequence / Continuous**

Selecting Single Sequence sets the acquisition mode so that when you press Run, a single acquisition sequence is performed and the instrument stops once the acquisition sequence is completed. Selecting **Continuous** sets the acquisition mode so that an acquisition sequence is started as soon as you select Continuous and as one acquisition sequence completes, another begins.

Note that an acquisition sequence can require more than one acquisition. For example, in a spectrum view, the trace function might be set to Average 100 acquisitions. Thus, a complete acquisition sequence would consist of 100 acquisitions that are averaged together to create the trace that is displayed.

## Run

### Menu Bar: **Run > Run**

### Application Bar / Front Panel: **Run**

Selecting **Run** begins a new acquisition/measurement cycle.

## Resume

### Menu Bar: **Run > Resume**

Restarts data acquisition, but does not reset accumulated results, such as Average or MaxHold. This allows you to stop acquisitions temporarily, then continue. If the accumulation is already complete, for example, 10 acquisitions or 10 averages have already been completed, each subsequent Resume command will cause one more acquisition to be taken, and its results added to the accumulation. Not available if instrument settings have been changed.

## Abort

### Menu Bar: **Run > Abort**


Selecting **Abort** immediately halts the current acquisition/measurement cycle. In-process measurements and acquisitions are not allowed to complete. Visibility and accuracy of results is unspecified after an abort.

# The Acquire Control Panel

**Menu Bar:** Setup > Acquire

**Application Toolbar:** 

Selecting **Acquire** displays the Acquire control panel. These settings control the hardware acquisition parameters for the signal analyzer. Normally, these parameters are automatically adjusted for selected measurements. You can use Acquire to change these parameters manually if necessary.



| Tab  | Description   |
|--|---|
| <a href="#">Sampling Parameters (see page 579)</a> | Sets the parameters that control the acquisition of the signal. |
| <a href="#">Advanced (see page 580)</a>            | Sets the Dither parameters.                                     |
| <a href="#">FastSave (see page 584)</a>            | Sets the FastSave parameters. Initiates FastSave acquisitions.  |

## Run

Run mode specifies whether the analyzer will stop acquiring data after it completes a measurement sequence.

- **Continuous** - In Continuous mode, once the analyzer completes a measurement sequence, it begins another.
- **Single** - In Single mode, once the analyzer completes a measurement sequence, it stops.

**NOTE.** A measurement sequence can require more than one acquisition. If the analyzer is configured to average 100 traces together, the measurement sequence will not be completed until 100 traces have been acquired and averaged.

## Sampling Parameters Tab

The Sampling Parameters tab enables you to set the controls for data acquisition. There are two settings on the Sampling Parameters Tab, the Adjust setting and FastFrame. Depending on the setting chosen for Adjust, two additional parameters can be set. Normally, the best results are achieved by leaving the Adjust control set to All Auto. Manually overriding the automatic settings affects both real-time and swept acquisitions.

### Sampling Control

There are three sampling parameters that interact with each other: Acquisition Bandwidth, Acquisition Length, and Acquisition Samples. Acquisition Bandwidth is the frequency range of the acquisition. Acquisition Length refers to the time over which the acquisition occurs. Acquisition Samples is the number of samples acquired over the acquisition time. The parameters available for Adjust and how the parameters are determined are shown in the following table.

| Adjust               | User sets |                    | Analyzer calculates  |
|----------------------|-----------|--------------------|--|
| All Auto             | N/A       | N/A                | All values based on the open displays, with the highest priority given to the selected display |
| Acq BW / Acq Samples | Acq BW    | Acq Samples        | Acq Length   |
| Acq BW / Acq Length  | Acq BW    | Acquisition Length | Acq Samples  |

### Acquisition Memory Usage

The center portion of the tab shows how the acquisition memory is used.

| Readout               | Description   |
|-----------------------|---|
| Samples/s, sec/Sample | Readout of the acquisition sample rate and sample period.   |
| Capacity              | The maximum period of time and number of samples that could be acquired with the current sampling parameters. |
| Using                 | The total amount of acquisition memory that will be used based on the current settings.                       |

## FastFrame

FastFrame is a feature that allows you to segment the acquisition record into a series of frames, or data records. A typical use of FastFrame is to record data samples around signal events of interest, while not wasting memory on irrelevant data between these events. FastFrame is only valid for real-time acquisitions. If any measurement is configured for swept acquisitions, an error message is displayed: Disabled: *FastFrame doesn't support swept settings*.

| Setting    | Description   |
|------------|---|
| FastFrame  | Enables FastFrame acquisition mode.   |
| Max Frames | Specifies the maximum number of frames to record in a single acquisition process. |

The Actual readout shows the actual number of frames that will be acquired into the acquisition memory. This number changes based on the Acq BW, Acq Samples, and Acq Length values. This number will never be greater than the value set by Max Frames.

## Advanced Tab (Acquire)

Provides access to Advanced Acquire settings.

| Setting | Description   |
|---------|---|
| Dither  | Can be set to Auto or Manual. In Auto mode, each measurement sets Dither as appropriate for the measurement. In Manual mode, the user setting overrides the measurement preference. |

## Dither

Dither is the process of adding noise to the analog signal prior to the analog-to-digital conversion. The purpose of adding noise is to randomize the quantizing process of the ADC (analog-to-digital converter) so it does not appear as discrete components in locations where the distortion components of the ADC would

fall. All ADCs produce quantizing errors that appear as noise (quantizing noise) or as discrete components which are less desirable. Without dither and in the presence of tones that are close to sub-multiples of the sample rate or correlated to the sample rate, all the quantizing energy can fall in the same locations as discrete distortion components. This can increase what appears as spurious tones but they are not related to the normal spurious mechanisms or true distortion. Adding noise can randomize the quantization process so those tones aren't produced.

In general, you should leave the Dither setting on Auto. In the Auto mode, individual measurements will select the best guess dither setting (on or off). In Manual mode, the user setting overrides individual measurements' dither settings enabling the user to determine the best use for the best measurement.

---

**NOTE.** *If the [RF & IF Optimization \(see page 570\)](#) is set to Best for multiple displays, you cannot change the Dither setting.*


---

Dither is not always necessary. The signal itself can be used to randomize quantizing errors. Since the input signal is not under control of the instrument, the use of dither is signal dependent and its use must be determined by the user. Typically, pure tones create the most problems with quantizing errors. Dither should be used for those input types. Many digital modulation types are noise-like, so dither is usually not needed for modulation types such as CDMA, GSM and many others. Dither also uses up a bit of headroom in the ADC, typically 2-3 dB. If the largest dynamic range to the noise floor is needed, that is Signal to Noise Ratio, the dither can probably be turned off. The measurements for Channel Power/ACLR and MCPDR automatically turn dither off for just that reason. Signals in the range of -6 dB to -40 dB below the reference level usually show the greatest benefit from having dither present. The user can find the best use by observing the results for the particular input signals while setting dither on or off. There are cases where dither can make things worse. If dither makes an improvement in the measurement, it should be left on. It is doing its job to improve dynamic range. If it makes it worse, dither should be left off.

## FastSave

FastSave (also known as Fast Save) is a way to automatically save acquisitions to a file. It is a way to collect data, not to analyze the data. Analyzing the data require recalling the saved acquisition data file.

When making acquisitions and performing analysis, the instrument creates a cache file with a copy of the acquisition that it uses as the source for analysis and displays. To Save a waveform as a .TIQ or .MAT file, the instrument pulls data from the cache, formats the data for the chosen file type, and writes again to the disk with a .TIQ or .MAT extension. Normal operation requires creation of the cache file and reading of a cache file prior to analysis or save functions. These hard-drive read and write cycles can be time consuming for large files. If you need only to save data for later analysis, a mode named FastSave is available.



FastSave mode bypasses cache file creation and writes directly to a file. This eliminates two of three disk read or write operations. Depending on the file size, FastSave reduces the time required for data acquisition up to 65%. Using FastSave means that *only* real-time processes will be used. The instrument triggers and DPX display are not affected. All other measurements are suspended, and an information message (Disabled - FastSave is selected) is displayed in the graph area of each display.

### Preparing for a FastSave Acquisition

1. Configure the instrument as necessary to capture the signal event you are interested in.
2. On the **Setup > Acquire > FastSave** tab, specify the **Base file name**.
3. Click the ... button to specify the location where the acquisition files will be saved.
4. Set the number of files to save by entering a number in the **Max files saved per Run** box.
5. Specify the file format (TIQ or Matlab).

### Starting a FastSave Acquisition

To start a FastSave acquisition:

1. Select the **Acq** button to display the Acquire control panel.
2. Select the **FastSave** tab.
3. Select the **FastSave acquisition data** checkbox to enable FastSave acquisition.
4. Start a FastSave acquisition:
  - a. If the instrument is running when you enable FastSave, FastSave acquisitions begin immediately.
  - b. If the instrument is paused when you select FastSave, select **Run** to begin FastSave acquisitions.

---

**NOTE.** *When using FastSave in Single Sequence, it is possible for more than one acquisition to be saved under the following conditions.*

- *The DPX display is open*
- *Acquisition length is less than about 50 ms*

*FastSave can store more than one short acquisition in the time it takes DPX to acquire, process and display one set of DPX spectrums (about 50 milliseconds). Thus, you may find that FastSave creates more than one file under these conditions. To limit FastSave to a single acquisition per process, set **Max files saved per run** to 1 in the FastSave tab (Setup > Acquire > FastSave).*

---

### FastSave Status Messages

Once the FastSave acquisition begins, a status message appears in the graph area. The message that appears depends on whether the instrument was running or paused when FastSave acquisition data was clicked.

- **Disabled - FastSave is selected:** This appears when acquisition is paused. It means that the display update is disabled until FastSave is disabled. FastSave acquisition will not begin until you select **Run**.
- **Saving X of XX** - FastSave is active and saving files. This appears only while FastSave is active.
- **Done saving - restart with Stop, then Run .. More** - Displayed when the specified number of files have been saved. FastSave can be restarted by selecting Run to begin a new set of acquisitions.

## FastSave Tab

Use FastSave tab is used to configure and initiate [FastSave \(see page 581\)](#) acquisitions.

| Setting                   | Description   |
|---------------------------|---|
| FastSave acquisition data | Enables/disables FastSave acquisitions. If acquisitions are running, selecting the FastSave acquisition data checkbox immediately begins saving acquisition data to files. If the instrument is paused, FastSave acquisitions will not begin until you select <b>Run</b> .                        |
| Format                    | Specifies the file format of the FastSave acquisition data files. Files can be saved in <a href="#">TIQ</a> or <a href="#">MAT</a> format.  |
| Base file name            | Specifies the base name for acquisition data filenames. When acquisition data files are saved, a 24-character timestamp is appended to the base file name. The maximum base file name length cannot exceed 232 characters.  |
| Location                  | Specifies the directory location where files will be saved. Files can be saved on the local hard drive, a network location or even a USB- connected drive. However, due to the time required to transfer files across a network or by USB, the local hard drive is the only recommended location. |
| Max files saved per Run   | Specifies the number of files to be saved each time a FastSave acquisition is started. The number of files that can be saved is limited only by the storage space available. Each saved acquisition file uses about 3.3 MB of space.  |

## Acquire

**Menu Bar:** Setup > Acquire

**Application Toolbar:** 

Selecting **Acquire** displays the Acquire control panel.

The Acquire control panel changes based on whether you are connected to an RSA306 or not.

### Acquire (RSA306 not connected)

The Acquire control panel contains the Acquisition Data tab and Playback tab.

Acquire

ACQUISITION Data Playback

Signal Input:  
RF

Sample Rate: 56.000 MS/s  
Record Length: 119 Samples

**Signal Input.** The Signal Input control is not used by SignalVu-PC.

| Setting  | Description   |
|--|---|
| <a href="#">Acquisition Data</a>                                       | Displays the sample rate and record length reported by the oscilloscope.                                |
| <a href="#">Playback</a><br>(requires option SV56)                     | Plays saved files that were recorded with the Record feature.<br>Playback only plays .r3f file formats. |
| <b>NOTE.</b> The Record tab only displays when connected to an RSA306. |   |

### Acquire (RSA306 connected)

Selecting **Acquire** displays the Acquire control panel. These settings control the hardware acquisition parameters for the RSA306 signal analyzer.

Acquire

Run:  
Continuous

IQ Sampling Parameters Frequency Reference Record

Adjust: All Auto (recommended)

Acq BW: 40.00 MHz

Acq Length: 1.906 us

Samples: 107 Samples

56.0 MS/s  
17.9 nsec/Sample  
Capacity: 1 s  
56.6 MSamples  
Using: 1.91 us, 107 Samples, 0.0 % of capacity

| Tab  | Description   |
|--|---|
| <a href="#">IQ Sampling Parameters</a>             | Sets the controls for real-time acquisition.  |
| <a href="#">Frequency Reference</a>                | Specifies the source of the reference frequency.  |
| <a href="#">Record</a>                             | Records the acquisition data directly to a file.  |
| <a href="#">Playback</a><br>(requires option SV56) | Plays the saved files recorded with the Record feature.<br>Playback only plays .r3f file formats. |

### Run

Run mode specifies whether the analyzer will stop acquiring data after it completes a measurement sequence.

- **Continuous** - In Continuous mode, once the analyzer completes a measurement sequence, it begins another.
- **Single** - In Single mode, once the analyzer completes a measurement sequence, it stops.

---

**NOTE.** *A measurement sequence can require more than one acquisition. If the analyzer is configured to average 100 traces together, the measurement sequence will not be completed until 100 traces have been acquired and averaged.*

---

## Triggering

The analyzer has two triggering modes: Free Run and Triggered.

### Free Run

In Free Run mode, the instrument initiates acquisitions without considering any trigger conditions. It is a fast and easy way to see the signals. Free Run is usually adequate for the Spectrum display unless you need to specify a particular time at which to collect the data record.

### Triggered

In the Triggered mode, the instrument initiates an acquisition when a trigger event is recognized. The conditions that define a trigger event depend on the selected trigger source. There are several source selections available for choosing the signal to monitor for a trigger event. If the RF input is the selected trigger source, you can choose to use power, frequency mask triggering, DPX Density, or Runt as the trigger types. If other inputs are selected as the trigger source, you can select the slope and level of the trigger point. With the Gated source, you also specify the gating signal level.

**Power triggering.** Power triggering triggers the instrument on time-domain signal transitions. The incoming data is compared to a user-selected level in dBm. You can select the time-domain bandwidth and trigger on the rising or falling edge.

**Frequency Edge triggering.** Frequency Edge triggering triggers the instrument when the frequency of the input signal changes relative to the instrument center frequency.

**Frequency mask triggering.** Frequency Mask Triggering allows you to trigger the instrument when a signal in the frequency domain violates the mask. You can draw a mask to define the conditions within the real-time bandwidth that will generate the trigger event. It allows you to trigger on weak signals in the presence of strong signals. This triggering is also useful for capturing intermittent signals.

**DPX Density triggering.** DPX Density triggering uses characteristics of the DPX display to specify trigger parameters. When using DPX Density triggering, the percentage of time the signal falls within the measurement box is used to define a trigger event. The measurement box is defined by frequency and amplitude values.

**Runt triggering.** Runt triggering allows you to define a trigger event based on a pulse amplitude that crosses one threshold but fails to cross a second threshold before recrossing the first.

You can define the following event parameters by selecting the respective tab in the Trigger control panel:

- Event source, event type, and event definition.
- Time parameters that define what portion of the data record is used for pretrigger samples and whether to delay the trigger for a set amount of time after the recognized event.
- Advanced parameters define where a trigger occurs within the acquired memory or trigger each segment in the swept acquisition mode.

## Save on Trigger

An additional feature of the triggering capability is to save acquisition data when a trigger event occurs. When the analyzer is run in Triggered mode, you can configure the instrument to save acquisition data to a file. You can also set the instrument to save a picture of the screen when the trigger event occurs. See [Action Tab \(Triggering\) \(see page 605\)](#) for details on configuring Save on trigger.

## Trigger Front Panel Status Lights


The analyzer has three lights on the front panel that indicate the Trigger status:

- Arm – The instrument is collecting pre-trigger data.
- Ready – Enough pretrigger data has been collected and the instrument is waiting for the next trigger event.
- Trig'd – A trigger even has been received, and the instrument is collecting post-trigger data and transferring the complete data record out of acquisition memory.

## Setting Up Triggering

To set up triggering, use the Trigger control panel. To display the Trigger control panel:

- Press **Trigger** on the front panel or click **Trig** in the Application toolbar or select **Setup > Trigger**.



## Using Free Run

To set Triggering to Free Run:

- Click the **Free Run** option button.

The signal analyzer will acquire a signal without regard to any triggering conditions.

### Using Power Trigger

To trigger acquisitions based on the power of the signal:

1. Click the **Triggered** option button.
2. Select **RF Input** from the Source drop-down list.
3. Select **Power** from the **Type** drop-down list.
4. For further information on setting the power level parameters, see [Event Tab \(Triggering\)](#). (see [page 594](#))

### Using Frequency Edge Triggering

To trigger acquisitions based on input signal frequency changes:

1. Click the **Triggered** option button.
2. Select **RF Input** from the **Source** drop-down list.
3. Select **Frequency Edge** from the **Type** drop-down list.
4. Set the **Freq Slope** as appropriate. Select **Rise** to trigger when the signal frequency is increasing. Select **Fall** to trigger when the signal frequency is decreasing.
5. Set the **Freq Level**. The Freq Level is the frequency that you want to generate a trigger.
6. If appropriate, enable **Power Threshold** and set the value. Use this setting to ensure that only sufficiently large signals, not noise, can generate a trigger.
7. If appropriate, enable **Time-domain BW** and set the level. Use this setting to process the trigger input signal before the trigger system analyzes the signal.

For further information on setting the Frequency Edge parameters, see [Event Tab \(see page 594\)](#).

### Using Frequency Mask Triggering

To trigger acquisitions based on violations of a user-defined mask:

1. Click the **Triggered** option button.
2. Select **RF Input** from the **Source** drop-down list.
3. Select **Frequency Mask** from the **Type** drop-down list.

For further information on setting the frequency mask parameters, see [Event Tab \(Triggering\)](#) (see [page 594](#)) and [Mask Editor](#) (see [page 590](#)). If Frequency Mask triggering is installed, you can change the resolution bandwidth of the frequency transform used to compare to the trigger mask. The standard FMT uses a fixed length FFT for frequency transforms.

### Using DPX Density Triggering

To trigger acquisitions based on the density of the DPX Bitmap trace:

1. Click the **Triggered** option button.
2. Select **RF Input** from the Source drop-down list.
3. Select **DPX Density** from the **Type** drop-down list.
4. Set the DPX Density parameters as appropriate. You can do this two ways:
  - Entering values for Threshold, Frequency, and Amplitude using the front-panel controls.
  - Select **Trigger On This** from the Touchscreen Actions menu (right-click in the graph or press and hold on the display for two seconds). Move and size the measurement box that appears to set the values for Frequency and Amplitude.

---

**NOTE.** *If the Amplitude units (Tools > Analysis > Units) is set to Watts or Volts, you cannot move the measurement box using a mouse or the touchscreen. Set the amplitude units to any unit other than Watts or Volts to move or resize the measurement box.*

---

For further information on setting the power level parameters, see [Event Tab \(Triggering\) \(see page 594\)](#).

### Using Runt Trigger

To trigger acquisition based on a runt signal:

1. Click the **Triggered** option button.
2. Select **RF Input** from the Source drop-down list.
3. Select **Runt** from the **Type** drop-down list.
4. For further information on setting the power level parameters, see [Event Tab \(Triggering\) \(see page 594\)](#).

## Frequency Mask Trigger

The frequency mask trigger monitors all changes in frequency occupancy within the capture bandwidth (40 MHz or 110 MHz).

The following topics describe the frequency mask trigger controls:

[Event tab \(see page 594\)](#): explains how to configure frequency mask triggering parameters.

[Mask Editor \(see page 590\)](#): explains how to create and edit frequency masks.

[Triggering \(see page 586\)](#): explains how to initiate frequency mask triggering.

## Mask Editor (Frequency Mask Trigger)

The Mask Editor is used to create masks that are used by the Frequency Mask Trigger function. The Frequency Mask Trigger ( FMT ) function is used to define trigger events to capture signal anomalies based on their frequency and amplitude.

### Displaying the Mask Editor

To display the Mask Editor:

1. Press **Trigger** on the front panel or select **Trig** from the Application toolbar.
2. Select the **Event** tab.
3. Set **Source** to **RF Input**.
4. Set **Type** to **Frequency Mask**.
5. Select **Mask Editor**.


### Creating a Mask

The simplest way to create a mask is to use the **Auto draw** function.

To create a new mask using **Auto draw**:

1. Select the trace on which to base the mask from the **Reference trace** drop-down list. (Only traces that are shown can be selected from the drop-down list. To show a trace that isn't currently in the list, open its display and enable the trace of interest.)
2. Specify the margins to be applied between the mask and the reference trace. Enter values for the X Margin and the Y Margin.
3. Click **Auto draw**.

The following screen shows a mask created with Auto draw.



To create a new mask without using Auto draw:

- Click the **New** button.

## Changing a Mask

A mask is a collection of points specified by their frequency (defined by their offset from the center frequency) and their amplitude. You can change a mask by adding (or removing) points and editing points.

---

**NOTE.** When placing points in the frequency mask, keep in mind that in most cases, the Frequency Mask Trigger has a range of 80 dB below Reference Level. However, for values of Reference Level lower than -50 dBm, this range decreases as the RefLevel decreases. The FMT range is also reduced when the Acquisition Bandwidth contains frequencies near 0 Hz. In both these situations, the underlying reason the trigger range is reduced is because the instrument is using various amounts of "digital gain" in addition to hardware gain. You are not prevented from placing mask points outside the usable range of the instrument, but at the time the mask is applied, the instrument adjusts its range. When the mask is adjusted, you can see this in the Spectrum display and the Mask Editor. The mask's shaded fill shows the usable vertical range. Points outside this range are shown, but will not cause a trigger.

---

## Adding Points

To add a point to a mask:

1. Click the **Add points** option button. (Alternatively, you can right-click on the desired point location (or left-click and hold for about a second) and release.)
2. Click in the graph area where you want the new point added to the mask.

You can also add a point to the mask while it is in Edit Points mode:

- Right-click in the graph area where you wish to add a point to the mask. You can also left-click and hold the button down for one second. You can duplicate this method on the touchscreen by touching the screen for one second.

If you add a point by accident, click **Undo**. You can click Undo until the Undo button is grayed out.

## Editing Points

When you edit a point, you change its offset from the center frequency and/or its amplitude. There are two ways to edit the offset and amplitude of a point on a mask. First, you can drag a point to the desired position. Second, you can double-click on the offset or amplitude value in the table and enter a new value.

To edit a point on a mask:

1. Click the **Edit Points** option button.
2. Click on the point you wish to change.
3. Change the position of the point using the following methods:
  - a. Dragging it with the mouse to the desired position.
  - b. Clicking the arrow keys in the Mask Editor window to move the point to the desired position. Use the left/right keys to change the frequency offset. Use the up/down arrow keys to change the amplitude. Once you click an arrow key, you can use the front panel knob for fine adjustments.
4. To edit another point, click **Next**.

You can also edit a point by double-clicking on the value in the table that you want to change and typing in a new value.

## Deleting Points

To delete points from the mask:

1. Click the **Edit Points** option button.
2. Click on the point to be removed and click the **X** button to delete the point.

You can also delete a point by clicking on the point in the table that you want to delete and clicking the **X** button.

If you delete a point by accident, click **Undo**. You can click Undo until the Undo button is grayed out.

## Saving a Mask

To save a mask:

- Click the **Save** button.

To save an edited mask under a new name:

- Click the **Save As** button.

### Opening an Existing Mask

To open an existing mask:

1. Click on **Open**. This displays the Open dialog box.
2. Navigate to the location of the mask file.
3. Select the desired file and click **Open**.

## Trigger Settings

**Menu Bar:** Setup > Trigger

**Application Toolbar:** Trig

**Front Panel:** Trigger

The Trigger control panel allows you to set the parameters that define trigger events and how the instrument responds to them.

### Trigger Modes

There are two trigger modes: Free Run and Triggered.

| Mode      | Description   |
|-----------|---|
| Free Run  | Acquires and displays the signal without triggering.                  |
| Triggered | Acquires a data record after the defined trigger event is recognized. |


### Trigger Tabs

Use the Trigger tabs to access controls for setting the different trigger parameters.

| Tab   | Description   |
|---|---|
| <a href="#">Event (see page 594)</a>          | Sets parameters that control triggering on the selected trigger source.               |
| <a href="#">Time Qualified (see page 603)</a> | Sets values used to adjust time of acquisition with respect to time of trigger event. |
| <a href="#">Advanced (see page 604)</a>       | Sets controls for triggering in swept acquisitions.                                   |
| <a href="#">Actions (see page 605)</a>        | Sets parameters for Save on trigger function.   |

## Event Tab

The Event tab is used to set the parameters that define trigger events. The parameters that appear on the Event tab change depending on the selected Trigger source and type.



### Source


Source specifies the source of the signal to be monitored for a trigger event. The choices available are: RF Input, Trig In (front), Trig 2 In (rear), Gated, and Line. The Source drop-down list is always visible on the Event tab.

### Type

Type defines the trigger type as Power, Frequency Edge, Frequency Mask, DPX Density, or Runt. The Type selection is only present when Source is set to RF Input.

- When Type is set to Power, a transition through an adjustable level of the selected source determines a trigger event.
- When Type is set to Frequency Edge, the input signal changing in frequency relative to the center frequency defines a trigger event.
- When Type is set to Frequency Mask, where and how the input signal crosses or fails to cross the edge of the mask defines a trigger event.
- When Type is set to DPX Density, the amount of time a signal is present at a particular point in the DPX Spectrum graph defines a trigger event.
- When Type is set to Runt, a signal that exceeds a specified level and then fails to exceed a second specified level defines a trigger event.

### Source = RF input.




The following table describes the settings for each triggering type available when Source is set to RF Input. Type specifies whether a Power level, Frequency Edge violation, Frequency Mask violation, DPX Density threshold violation, or Runt signal will define a trigger event.

**Table 13: Trigger types**

| Event Type     | Setting         | Description   |
|----------------|-----------------|---|
| Power          |                 | For the Power setting, if the RF input signal crosses the specified level (in the direction set by Slope), a trigger event occurs. A trigger event does not occur if the power level of the signal simply exceeds the level setting, there must be a power level transient that crosses the specified level.  |
|                | Slope           | The slope control determines whether the signal analyzer finds the trigger point on the rising or falling edge of the RF input signal.  |
|                | Level           | The Level setting determines the power level that will be recognized as a trigger event. The Level is indicated in the Amplitude vs Time and Time Overview displays with a dark cyan colored-line.  |
|                | Time-domain BW  | Time-domain BW is a filter used to process the trigger input signal before the trigger system analyzes the signal. The frequency edge trigger point must lie within the range of time domain bandwidth. This makes the range of the frequency edge trigger = Center Frequency $\pm(0.5 \times \text{Time Domain Bandwidth})$                            |
| Frequency Edge |                 | For the Frequency Edge setting, if the RF input signal crosses the specified frequency (in the direction set by Freq Slope), a trigger event occurs. For Frequency Edge trigger, the frequency that initiates a trigger event is identified by the Measurement Frequency + Frequency Level. Frequency Level can be either a positive or negative value. |
|                | Freq Slope      | The slope control determines whether the analyzer finds the trigger point on the rising (increasing) or the falling (decreasing) edge. This setting is analogous to the Slope control on the Power trigger type.  |
|                | Freq Level      | The Freq Level setting specifies the difference in frequency between the input signal and the analyzer center frequency required to generate a trigger event. The Freq Level setting is limited to half the real-time bandwidth.  |
|                | Power Threshold | Specifies the power level the signal must exceed for a frequency variation to be considered valid.  |
|                | Time-domain BW  | Time-domain BW is a filter used to process the trigger input signal before the trigger system analyzes the signal.  |
| Frequency Mask |                 | For the Frequency Mask setting, events where the trace crosses the edge of the mask when it should not or when the trace does not cross the edge of the mask when it should define the trigger event. Use <b>Mask Editor</b> (on the Event tab when Type is set to Frequency Mask) to create and modify Frequency Masks.                                |
|                | Violation       | Specifies the type of transition that defines a violation.  |
|                | RBW             | Specifies the resolution bandwidth.   |
|                | Auto            | Specifies whether the RBW is adjusted automatically or manually.  |

| Event Type  | Setting        | Description  |
|-------------|----------------|--|
| DPX Density |                | For DPX Density, the percentage of time the signal falls within the measurement box is used to define a trigger event. The measurement box is defined by frequency and amplitude values.   |
|             | Density        | Specifies whether a trigger event is defined by the DPX Density value exceeding (Higher) the specified Threshold value or is defined by the DPX Density value falling below (Lower) the specified Threshold value.               |
|             | Threshold      | Threshold specifies the density value used for determining a trigger event. Density is a measure of how much time the signal is present in a particular region in the DPX Spectrum graph. The range for this value is 0 to 100%. |
|             | Frequency      | Specifies the frequency at the center of the measurement box. The +/- value specifies the edges of the measurement box above and below the Frequency setting.  |
|             | Amplitude      | Specifies the amplitude of the center of the measurement box. The +/- value specifies the edges of the measurement box above and below the Amplitude setting.  |
| Runt        |                | Defines a trigger event based on a pulse amplitude that crosses one threshold but fails to cross a second threshold before recrossing the first.   |
|             | Pulse          | Specifies whether the runt trigger is defined by a positive-going pulse or a negative-going pulse.   |
|             | High level     | The High level defines the upper voltage limit (first threshold) of a runt pulse.  |
|             | Low level      | The Low level defines the lower voltage limit (second threshold) of a runt pulse.  |
|             | Time-domain BW | Time-domain BW is a filter used to process the trigger input signal before the trigger system analyzes the signal.   |

**Source = Trig In (front).**



| Setting | Description  |
|---------|--|
| Slope   | The slope control determines whether the analyzer finds the trigger point on the rising or the falling edge of the trigger input signal. |
| Level   | The Level setting determines the signal amplitude that will be recognized as a trigger event.  |

**Source = Trig 2 In (rear).**

The screenshot shows the 'Trigger' configuration window. On the left, 'Triggered' is selected under the 'Trigger' section. The 'Source' dropdown is set to 'Trig 2 In (rear)'. Under the 'Slope' section, 'Rise' is selected. To the right, there are two waveforms: a rising edge and a falling edge, with a red dashed line indicating the trigger level. The text 'Triggers at TTL threshold' is displayed.

**Setting****Description**

|       |   |
|-------|---|
| Slope | The slope control determines whether the analyzer finds the trigger point on the rising or the falling edge of the trigger input signal. The trigger level is automatically set at a TTL level. |
|-------|---|

**Source = Gated.** When the Source is set to gated, the front panel **Trig In** source is monitored for the trigger event and a signal on the rear panel **Gate / Trig 2 In** is used to determine when a trigger event is valid. A transition through the selected Level on the TrigIn signal input (selectable as rising or falling edge) will only be recognized as a trigger event if it occurs while the Trig2 signal is True (selectable as High or Low). Note that only edges are trigger events, not stable signal levels.

The screenshot shows the 'Trigger' configuration window with 'Source' set to 'Gated'. Under 'Slope', 'Rise' is selected. The 'Level' is set to '1.60 V'. The 'Input Impedance' is '50 Ω'. The 'Trigger when Trig In (front):' section has 'while Trig 2 In (rear):' set to 'High'. A link 'use menu Setup > Configure In/Out' is visible.

**Setting****Description**

|                        |  |
|------------------------|--|
| Slope                  | The slope control determines whether the analyzer finds the trigger point on the rising or the falling edge of a signal.                                     |
| Level                  | The Level setting determines the signal amplitude that will be recognized as a trigger event.  |
| while Trig 2 In (rear) | Specifies whether a High or Low signal at the <b>Trig 2 In</b> input (rear panel) allows a trigger event on the <b>Trig In</b> (front) input to be accepted. |

**Source = Line.**

The screenshot shows the 'Trigger' configuration window with 'Source' set to 'Line'. The 'Triggered' option is selected under the 'Trigger' section.

The analyzer triggers off of the AC line supply.


**Source = RF Input and Type = Frequency Edge.**


**Table 14:**

| Setting         | Description  |
|-----------------|--|
| Freq Slope      | The slope control sets whether the analyzer triggers on the rising (increasing) or the falling (decreasing) edge. This setting is analogous to the Slope control on the Power trigger type.  |
| Freq Level      | The Freq Level setting specifies the difference in frequency between the input signal and the analyzer center frequency required to generate a trigger event. The Freq Level setting is limited to half the real-time bandwidth. For example, if your analyzer has an 85 MHz real-time bandwidth, the Freq Level is limited to $\pm 42.5$ MHz. If your analyzer has a 25 MHz real-time bandwidth, the Freq Level is limited to $\pm 12.5$ MHz. |
| Actual          | Readout of the Freq Level being used. This number will be different from the Freq Level setting when Time-domain BW is enabled.  |
| Hysteresis      | Readout of the current hysteresis value.   |
| Power Threshold | Specifies the power level the signal must exceed for a frequency variation to be considered valid.   |
| Time-domain BW  | Specifies the Time-domain BW. Time-domain BW is a filter used to process the trigger input signal before the trigger system analyzes the signal.<br>The frequency edge trigger point must lie within the range of time domain bandwidth. This makes the range of the frequency edge trigger = Center Frequency $\pm (0.5 \times \text{Time Domain Bandwidth})$   |
| Actual          | Readout of the current acquisition bandwidth (see Setup > Acquire > Sampling Parameters). If Time-domain BW is enabled, this readout mirrors the setting for the Time-domain BW.   |

### Freq Slope


The following figures show how the Freq Slope setting affects the trigger.






### Hysteresis

The following figures show how the hysteresis value affects when a trigger is generated. In the first figure, the signal frequency crosses through the User Trigger Level (Freq Level setting) several times but never generates a trigger because it never crosses the lower hysteresis level followed by crossing the upper hysteresis level.



The following figure shows that a trigger is generated when the signal frequency changes sufficiently to exceed the hysteresis value (the difference between the upper and lower thresholds) and cross the lower threshold followed by crossing the upper threshold.



The hysteresis value is calculated as:

$$\text{Hysteresis} = (0.2 \times \text{Full Trigger Level range}) / 10$$


This amounts to 0.2 divisions if the full range is set to 10 divisions.

- If Time-domain BW is not enabled and the ACQ BW is 110 MHz, the full range is 110 MHz, so the hysteresis is 2.2 MHz
- If Time-domain BW = 1 MHz, then full range is 1 MHz, resulting in a hysteresis value of 20 kHz

Frequency edge hysteresis only changes when the acquisition bandwidth or time-domain bandwidth changes. Hysteresis does not change with the vertical scale factor.


Using a wider acquisition bandwidth allows you to monitor a wider frequency range but it also allows for more noise which requires more hysteresis. To achieve the best results, use the smallest acquisition bandwidth (or time-domain bandwidth) that contains the frequency variations of interest. Using a wide acquisition bandwidth with a small frequency variation can result in the frequency variation not exceeding the hysteresis. This will result in no trigger being generated.

### Source = RF Input and Type = Frequency Mask.



| Setting        | Description   |
|----------------|---|
| Violation      | These selections define the state or transitions of the signal used to define a trigger event.  |
| Mask Editor    | Displays the Mask Editor so that you can create and edit Frequency Masks.   |
| Load from file | Invokes the Open dialog for selecting a previously saved frequency mask file.   |
| RBW            | Sets the Resolution Bandwidth value to be used in the spectrum view. The value is italicized when Auto is selected.                                     |
| Actual         | This is a displayed value, not a setting. It is the RBW being used by the signal analyzer; this value may not match the RBW requested (in manual mode). |

**Violation.** Violations are state changes (or a sequence of state changes) defined by whether a point on the signal falls within the frequency mask or outside the frequency mask. A True violation occurs if even a single point on the trace falls inside the mask (above the mask's lower edge). Conversely, a False violation occurs if no points on a signal fall inside the mask (all trace points are below the mask's lower edge).



A trigger event is initiated based on the type of violation that occurs and the sequence of violations that occur.

| Violation            | Description  |
|----------------------|--|
| T                    | Only one state is required to initiate a trigger event. The signal has at least one data point inside (above) the mask. The trigger event occurs when the instrument detects the first signal with a point violating the mask. A trigger event could occur on the first acquisition.                               |
| F                    | Only one state change is required to initiate a trigger event. The signal has all data points outside (below) the mask. The trigger event occurs when the instrument has detected the first spectral waveform that lies completely below (outside) the mask. A trigger event could occur on the first acquisition. |
| T - F (Out)          | Two states are required to initiate a trigger event. The signal must be inside the mask and then passes out of the mask. The trigger event occurs when the first non-violating waveform is detected.   |
| F - T (In)           | Two states are required to initiate a trigger event. The signal must be outside the mask and then passes into the mask. The trigger event occurs when the first non-violating waveform is detected.  |
| T - F - T (Out & In) | Three states are required to initiate a trigger event. The signal starts inside the mask and then passes out of the mask. Next, the signal must pass into the mask. The trigger event occurs at the second transition where the signal passes back into the mask.  |
| F - T - F (In & Out) | Three states are required to initiate a trigger event. The signal starts outside the mask and then passes into the mask. Next, the signal must pass back outside the mask. The trigger event occurs at the second transition where the signal passes back out of the mask.   |

**Mask Editor.** Launches the [Mask Editor \(see page 590\)](#) that enables you to create and edit Frequency Masks.

**Load from File.** Enables you to load an existing frequency mask for use with Frequency Mask Trigger.

**Source = RF Input and Type = DPX Density.** DPX Density triggering analyzes the density of the DPX Bitmap trace to define trigger events. You specify the portion of the DPX Bitmap trace to analyze by

defining a measurement box. The measurement box is the only area of the DPX Spectrum graph that is analyzed for trigger events.


| Setting   | Description   |
|-----------|---|
| Density   | Specifies whether a density value higher or lower than the threshold value defines a trigger event.   |
| Threshold | Specifies the DPX Density threshold that defines a trigger event. The DPX Density must be higher or lower (specified by the Density selection) than this value to define a trigger event. |
| Frequency | Specifies the center of the measurement box on the frequency axis (horizontal).   |
| +/-       | Specifies the width of the measurement box.   |
| Amplitude | Specifies the center of the measurement box on the power level (vertical) axis.   |
| +/-       | Specifies the height of the measurement box.  |

**Source = RF Input and Type = Runt.** Runt triggering defines a trigger event based on whether or not a signal transition fails to cross a user-defined level. Two levels are defined for Runt triggering. A Runt signal is one that passes through both levels in one cycle but only passes through one level on the next cycle.

| Setting    | Description   |
|------------|---|
| Positive   | Specifies that a second signal transition cycle that fails to exceed the High level will define a trigger event.    |
| Negative   | Specifies that a second signal transition cycle that fails to fall below the Low level will define a trigger event. |
| High level | Specifies the High level used to define the trigger event.  |
| Low level  | Specifies the Low level used to define the trigger event.   |



## Time Qualified Tab

The Time Qualified tab allows you to specify whether triggers are accepted based on timing characteristics. Trigger events that pass the timing qualification result in an acquisition. Trigger events that do not pass the timing qualification are ignored. Time-qualification can be applied to Power, Frequency Mask, Runt and External triggers.



| Setting | Description   |
|---------|---|
| None    | No timing qualification is applied to trigger events. Acquisitions occur whenever a trigger event occurs (as specified on the Event tab).   |
| Shorter | If the length of the trigger event is shorter than the time specified by the Time 1 setting, then the trigger event is accepted and an acquisition is completed.  |
| Longer  | If the length of the trigger event is longer than specified by the Time 1 setting, then the trigger event is accepted and an acquisition is completed.  |
| Inside  | If the length of the trigger event is longer than the time specified by the Time 1 setting and shorter than the time specified by the Time 2 setting, then the trigger event is accepted and an acquisition is completed. |
| Outside | If the length of the trigger event is shorter than the time specified by the Time 1 setting or longer than the time specified by the Time 2 setting, then the trigger event is accepted and an acquisition is completed.  |
| Time 1  | Enter the Time 1 reference value here. Range: 20 ns-28 s.   |
| Time 2  | Enter the Time 2 reference value here. Range: 20 ns-28 s.   |

The following table shows how the appearance of the Time Qualified tab changes depending on whether time qualified triggering is enabled or not enabled.

| Tab appearance   | Description  |
|--|--|
|  Time Qualified | When None is selected on the Time Qualified tab, the watch icon is dimmed.                                 |
|  Time Qualified | When any selection other than None is selected on the Time Qualified tab, the watch icon appears in color. |

**Minimum off time between trigger events.** This readout shows how long the trigger conditions must be absent between the end of one trigger event and the start of the next trigger event to be ensure that the end of the first event is recognized. For example, if Freq Mask is selected and its RBW causes a minimum off time of 500  $\mu$ s, and your signal violates the mask for 600  $\mu$ s of every 1 ms, (so that the OFF time is


400  $\mu$ s), it will appear to the analyzer that the signal is continuously in violation. Shorter and Inside settings will not produce a trigger in this situation.

## Advanced Tab (Triggering)

**Menu Bar: Setup > Trigger > Advanced**

**Application Toolbar: Trig > Advanced**

Allows you to set parameters for trigger position and holdoff settings.



| Setting  | Description  |
|--|--|
| Position   | Sets the point along the acquisition record the trigger event should occur.  |
| Auto   | Enables the instrument to set the position of the trigger event. To manually specify the position of the trigger event, unselect Auto.               |
| Delay  | Sets how long after the trigger event specified by the Position setting, the analyzer should wait before actually initiating the trigger event.      |
| Holdoff  | Prevents triggers until there have been no trigger events for at least the specified time. Range: 20 ns-10 s.  |
| Trigger each frequency segment in Swept acquisition mode | Sets the analyzer to wait for a trigger for each acquisition in a swept spectrum or other measurement requiring multiple acquisitions worth of span. |

### Holdoff

Holdoff on the analyzer does not work exactly like the traditional Holdoff setting on an oscilloscope. On an oscilloscope, once the instrument triggers, the holdoff setting causes the oscilloscope to refuse to accept another trigger until the holdoff time has expired. On the analyzer, once a trigger event is identified, the instrument determines how long it has been since the previous trigger event occurred and if the time since the previous trigger event is greater than the holdoff time, the trigger event is accepted and an acquisition is completed. If the time since the previous trigger event is less than the holdoff time, the latest trigger event is ignored (no acquisition occurs) and the holdoff time count restarts.

The most common reason for using holdoff is ensuring that the instrument triggers on the first pulse in a burst. For this application, the difference between the signal analyzer holdoff and the traditional oscilloscope holdoff has no effect on acquisitions.

The case where the difference can be noticed is when the time between trigger events is always less than the specified holdoff time. An oscilloscope will trigger on the first event after the holdoff time expires,

no matter how short the time has been since the most recent rejected trigger event. In such a situation, the analyzer will not ever trigger because it requires that there are no trigger events during the holdoff time before an accepted trigger event.

## Actions Tab (Triggering)

The Actions tab sets parameters for the Save on trigger function. The Save on trigger function allows you to save an acquisition data file and/or a screen capture when a trigger event occurs.

Trigger events are defined on the [Event \(see page 594\)](#) and [Time Qualified \(see page 603\)](#) tabs.

### Action tab settings

| Setting                       | Description  |
|-------------------------------|--|
| Save acq data on Trigger      | <p>Enables the Save on Trigger function. The list box to the right of the <b>Save acq data on Trigger</b> check box specifies the format of the saved file. Acquisition data files can be saved in TIQ (native) format, CSV format, and MAT (MATLAB) format. For descriptions of these file formats, see <a href="#">Data, Settings and Picture File Formats (see page 609)</a>.</p> <p>When TIQ is the selected file type, a file is saved after a complete acquisition, which can include multiple frames. When CSV or MAT is the selected file type, a separate file is saved for each frame.</p> <p>Files are saved in the last location a file was saved. If <b>Automatically increments filename/number</b> is enabled (see <a href="#">Save and Export (see page 30)</a>), files will be saved without user interaction. If Automatically increments filename/number is not enabled, you will be prompted for a filename when a trigger event occurs.</p> |
| Save picture on Trigger       | <p>Enables saving a picture of the display when a trigger event occurs. Files can be saved in PNG, JPG, and BMP format. For descriptions of these file formats, see <a href="#">Data, Settings, and Picture File Formats (see page 609)</a>.</p>   |
| Max total files saved per Run | <p>Specifies the number of times a Save on Trigger action stores a file. After this limit is reached, no more files are saved. The instrument will continue to run, but no additional files are saved when a trigger event occurs.</p> <p>Keep in mind when setting this number that picture files are counted as part of the total number of files. For example, if you set Max total files to 100, the instrument will save 100 acquisitions if only acquisitions are saved or only pictures are saved. But, if both acquisitions and pictures are saved, then 50 acquisitions and 50 pictures will be saved.</p>  |



## Saving and Recalling Data, Settings, and Pictures

You can save different types of data for later recall and analysis.


| Data type                                 | Description  | Save as type   |
|---|--|--|
| Setup files                               | Saves all of the setup information for all displays, except those settings that are not part of Preset.  | Setup (.Setup)   |
| Picture of Selected Display (PNG/JPG/BMP) | Saves a capture of the screen in the specified format. This option is useful for including the graphic in reports or other applications. Marker readouts and other information are included.   | Picture (.BMP, .PNG, .JPG)   |
| Results Export files                      | Saves the trace and numeric data for the selected display. The trace and numeric data are saved as CSV files.  | Results export (various)   |
| Measurement Settings                      | Saves a list of settings relevant to the selected measurement to a text file. This option is useful for including the measurement settings in reports.   | Measurement settings export (TXT)  |
| Trace                                     | Saves a trace for later recall into the display from which it was saved.   | Selected trace (various)   |
| Data                                      | Saves data for reanalysis by RSA6100B Series Real-Time Spectrum Analyzers, RSA5100B Series Real-Time Signal Analyzers, SPECMONB Series Real-Time Spectrum Analyzers, or as data to use with external software (either CSV (comma-separated value) or MAT (MATLAB format)). For fastest capture of data files in TIQ or MAT format, use FastSave. This function disables caching and analysis of incoming data, greatly decreasing the time required to save files to local or network storage. | Acquisition data with setup (TIQ); or Acquisition data export (CSV or MAT) |

### Saving Files

Saving files follows the same procedure regardless of the type of data being saved. To save setups (including application presets), pictures, results, or acquisition data:

1. If you are going to save DPX Spectra (DPXogram) acquisition data, stop acquisitions.  
Saving while running is allowed by the analyzer, but data files saved while running will not contain any DPX spectral data.
2. Select **Save As** from the **File** menu to open the **Save As** dialog box.

3. Navigate to the folder where you want to save the setups, or use the default location.
  - To save setups so that they appear in the User Presets list, save your setup in the folder **C:\RSA6100B Files\User Presets**. The saved setup will appear in the User Presets window with the name you give the file.
4. Enter a file name.
5. Select the type of file to save from the **Save as type** drop-down list.
6. Click **Save**.
7. If you save acquisition data (Acq data with setup (TIQ), Acq data export (CSV), and Acq data export (MAT)), the File Contents window appears. In the File Contents window, you need to specify the data type, data records, and data length to save the acquisition data.



- a. Select the data type: Choose IQ data, DPX Spectra, or both. DPX spectral data only exists while either the DPXogram or Split DPX display is open, and acquisitions are stopped. IQ data is not available when FastSave is turned on. In cases where the selected data type is not available, the file will be saved without that data.
- b. Select the data records you want saved. Frames can be saved only when FastFrame was enabled during acquisition.
- c. Select the desired data length. The data length controls are enabled only when IQ data is being saved.
- d. Click Save to save the acquisition data file.

### Saving Data with FastSave

To save data using FastSave, use the procedure [Preparing for a FastSave Acquisition \(see page 582\)](#).

### Recalling Files

You can recall three types of files: Setup files, Selected Trace, and Acquisition data with setup. Recalling data follows the same procedure regardless of the type of data being recalled.

1. Select **Recall** from the **File** menu to open the **Open** dialog box.
2. Navigate to the folder containing the file you want to recall.
3. Select the type of file to recall from the **Files of type** drop-down list. This selection determines the files that appear in the Open dialog box.
4. Select the file to recall.
5. Click **Open**.

### Setup Files

You can set up the instrument as desired and then save the settings in a setup file. This enables you to quickly setup the instrument by recalling previously saved setups.

### Exporting Results

Save for further analysis of results in other programs, such as MATLAB or Excel.

### Pictures of the Selected Display

You can save pictures of the instrument display for documentation purposes. When saving pictures of the display, you can select from three file types: BMP, JPG, or PNG.

### Measurement Settings

You can save a list of settings relevant to the selected display to a file for documentation purposes. The exported file uses tab characters to separate values. The settings included in the file depend on the selected display. The contents of the file are the same as a Results export except it does not include the results.

### Saving Acquisition Data with Setup

Save for later analysis with the analyzer. The setup is saved as part of the acquisition data file so you can choose when recalling data whether or not to also recall the associated acquisition and analysis parameters.

### Saving Acquisition Data Export in CSV Format

Save for examining results in other programs, such as MATLAB or Excel.

## Data, Settings, and Picture File Formats

You can save different types of data for later recall and analysis.

## Saved File Types

| File type                   | File extension      | Description  |
|-----------------------------|---------------------|--|
| Setup files                 | .setup              | Setup files enable you to save instrument settings for later recall. Using these files saves setup time and ensures consistent measurements.   |
| Picture                     | .png/.jpg/.bmp      | Screen capture of analyzer application window. They are useful for documenting results in test reports.  |
| Results Export              | .csv                | Results files contain the trace points and numeric values that were produced by the selected measurement. The header of a Results file includes key acquisition and analysis settings to aid in the interpretation of the data. The file is saved as a plain text file, but the contents are formatted as CSV (comma-separated values). Results files can be opened from applications such as Microsoft Excel and MATLAB.  |
| Measurement Settings        | .txt                | Measurement settings files contain a list of settings that describe how the instrument is set up for the selected measurement. The list contains measurement settings (for example, Span), trace settings (for example, whether or not a trace is selected) and global settings (for example, Frequency and Reference Level). The list of settings contained in the file varies depending on which display is selected.  |
| Selected Trace              | varies with display | Trace files contain the trace results data in binary format. These files are only readable by the analyzer application. Several of the instrument's measurements allow you to recall a Trace file for visually comparing a saved trace to a live trace.  |
| Acquisition Data with Setup | .tiq                | These files contain the acquisition data record and complete instrument setup (in binary format). Use these files to save data for later recall and analysis. When you save an Acquisition Data file, the current Setup is always saved with the file. At the time of recall, you will be asked whether you want to restore only the acquisition data or both data and setup. If the instrument is already set up for a specific measurement, you will probably want to recall only the acquisition data. Recalling both data and setup returns the instrument to the same state it was in at the time you saved the file. |
| Acquisition Data            | .csv                | These files contain IQ sample data before it has been processed by a measurement. The acquisition data points are saved as IQ pairs. Use this format to import the acquisition data into Microsoft Excel for further analysis.   |
| Acquisition Data            | .mat                | These files contain IQ sample data before it has been processed by a measurement. The acquisition data points are saved as IQ pairs. Use this format to import the acquisition data into MATLAB for further analysis.  |
| Range file                  | .csv                | These files contain a list of settings that describe how the instrument is set up for the Spurious measurement. The file contains the number of ranges enabled, the start and stop frequencies for each enabled range, the limits for each enabled range, the mask type and more.  |

## Results Export File Format

The Results Export format contains trace points and/or scalar results produced by the measurement. The file contains general information about the measurement settings at the top of the file and the results data in the second part of the file. Groups of settings or results are headed with [text], as described below.

At the top of the file is the measurement name and the date and time data was acquired.

The first group of settings is [Global Parameters]. These settings include the Measurement Frequency, Acquisition Data, Reference Level and others.

The second group of settings is [Parameters]. These settings are specific parameters which vary depending on the measurement.

The next group is [Trace Parameters], which may not be present, depending on the measurement. Within this group, there are parameters specific to a trace. There will be a Trace Parameters group for each trace shown on the display.

The next group is [Results]. These are scalar results for the measurement.

Next is [Traces]. The Traces group consists of one or more [Trace] groups. There is one Trace group for each trace. Each [Trace] group contains background information about the trace (units, number of points, x scaling, and others depending on the measurement) at the top of the group, followed by the trace points.

## Acquisition Data with Setup File (.tiq) Format

The file consists of two parts - the header that is in XML and the data that is in binary format.

**Header.** The header consists entirely of the DataFile element. This element is broken down into sub-elements.

1. DataSetsCollection
2. Setup

**DataSetsCollection.** The DataSetsCollection contains a DataDescription element, a DataSet element, and a ProductSpecific element.

**Binary data.** Binary data is a sequence of Int32 values, which are IQ pairs (I,Q,I,Q, ...) in binary Little Endian format.

To convert these values to volts, multiply the individual values by the Scaling value in the DataDescription.

The binary data starts at an offset into the file == the "offset" attribute in the DataFile element. This offset number will always start at the 19th character into the file (starting with 1), and will always be 9 characters long.

---

**NOTE.** You should not casually modify the XML header, because this will change the offset to the start of the binary data.

---

### Acquisition Data Files (.csv)

The acquisition data files have two sections. At the top of the file is the following information:

- SamplingFrequency - The sampling frequency in Hertz.
- NumberSamples - The number of IQ samples in the file.
- DateTime - When the data was acquired.
- Frequency - The center frequency in Hertz.
- AcquisitionBandwidth - The acquisition bandwidth Hz.

Following the AcquisitionBandwidth are the data. The data are IQ pairs, in volts.

Groups of settings or results are headed with [some text].

The first thing in the file is the measurement name and the date/time when the acquisition was taken.

The first group [Parameters] are global parameters. Measurement Bandwidth in this group is the measurement bandwidth used by the General Signal Viewing measurements (Spectrum, Amplitude vs. Time, etc). It also includes some source settings, like Acq BW, Dither, Preamp, and RF Attenuation.

The second group [Parameters] are measurement-specific parameters.

Another group which can occur is [Trace Parameters]. Within this group, there will be a set of parameters, one for each trace that is currently shown.

Another group is [Results]. These are scalar results for the measurement.

Another group is [Traces]. It has [Trace] groups under it, one for each trace. Each [Trace] group has some background information about the trace (units, number of points, x scaling, etc), and the trace points themselves.

### Acquisition Data Files (.mat)

The acquisition data files saved in MATLAB format contain the following MATLAB variables:

- InputCenter - The center frequency in Hertz.
- XDelta - The sample period in seconds.
- Y - A complex array containing IQ pairs.
- InputZoom - The acquisition bandwidth in Hertz.

The MATLAB format used to save acquisition data has the following properties and limitations:

- Files are stored in MATLAB Level 5 format.
- MATLAB Level 5 file size is limited to 2 GB.

- The instrument acquisition memory limit is 1 GB (1.024 Gigabytes). Because of the way acquisition data is processed, 1 GB of acquisition memory requires 2 GB of memory to store. When header information (instrument settings) is added to the acquisition data, the resulting file size is >2 GB.
- If acquisition memory is filled (1 GB), exporting the data to MATLAB format will result in a file that exceeds 2 GB in MATLAB level 5 format. A PC with sufficient memory is required to open the file. If memory is not sufficient, MATLAB will warn the user. To check how much acquisition memory is filled, select Setup > Acquire > Sampling Parameters. When the instrument is stopped, or you are analyzing a recalled data file, you can find the acquisition length under File > Acquisition Data Info.

## Printing Screen Shots

You can print screen shots (screen captures) two ways: use File > Print or save a picture file and print the file using a separate graphics program. Printing a screen capture is the same as printing with any windows program. For details on the available file formats for saving a screen capture, refer to [Data, Settings, and Picture File Formats \(see page 609\)](#). For details on saving a picture to a file, see [Saving and Recalling Data, Settings, and Pictures \(see page 607\)](#).

To print a screen shot:

1. Select **File > Print**.
2. Select **File > Print Preview** if you wish to review the screen shot before sending it to the printer.
3. Select **File > Print** to print the file to a printer.




## Online Help

### Menu Bar: Help > User Manual

This menu item displays this help. The help is a standard Windows help system. The Online Help menu item is the only method available to display the help; there are no other links to the help within the software.

## About the Tektronix RTSA

This window displays information about the software and hardware versions for the analyzer.



### Version

At the top of the window is a line that displays the version of the analyzer application software.

### Installed Options

This text box lists the instrument options installed on the analyzer. This window does not list every option installed on the instrument, just those that can be detected by the software.

### Hardware Information

This text box lists information about the instrument hardware versions.


### Copy Info

The Copy Info button copies information about the instrument to the Windows clipboard. This information may be useful if you need to have your instrument serviced.

## Configure In/Out Settings

**Menu bar:** Setup > Configure In/Out.

The Configure In/Out control panel enables you to configure the analyzer inputs and outputs. The settings in this control panel are shared by all application windows.



| Settings tab                                 | Description   |
|--|---|
| Trigger I/O                                  | Specifies the impedance for the Trig 1 (front panel) input. You can select between 50 Ω and 5 kΩ.   |
| Frequency Reference                          | Specifies the source of the Frequency Reference. You can select between an Internal (10 MHz) source and an external source.                 |
| <a href="#">Other Outputs (see page 621)</a> | Specifies whether the 28 VDC, Analog IF and Digital I and Q outputs are on or off. Additional settings are available for IF and IQ outputs. |

## Connecting an RF Signal

The analyzer has a single RF signal input located on the front panel. The input impedance is 50 Ω. The input frequency range for each analyzer model is listed in the following table. The maximum measurable continuous input power for all models is 1 W.

| Model    | Input frequency range |
|----------|-----------------------|
| RSA6106B | 9 kHz to 6.2 GHz      |
| RSA6114B | 9 kHz to 14 GHz       |
| RSA6120B | 9 kHz to 20 GHz       |

## About RF Input Protection

The following paragraphs describe how the analyzer behaves to protect the RF input during power-off, power-on, power-on self test (POST), and attenuator switching:

- When switched off (or when power is removed), the RF input is electronically disconnected from the input connector. The analyzer always has a minimum 10 dB attenuation applied during power-up and power-on-self-test. This is accomplished by always switching the attenuation control to Auto during these states. The minimum Auto attenuation value is 10 dB and is applied prior to the pre-amplifier (Option 50/51) in the signal path.
- When switching attenuator values, the action is performed in a 'make before break' sequence, such that the attenuation never switches to zero during an attenuator switching sequence. For example, consider switching from 20 dB to 15 dB attenuation. This change requires the instrument to switch from using one 20 dB attenuator to using one 10 dB and one 5 dB attenuator. This attenuator switching is done such that the 10 dB and 5 dB attenuators are switched in prior to switching out the 20 dB attenuator.

To connect an RF signal:

1. If the source signal is greater than 1 Watt (+30 dBm), connect the source signal to an attenuator. Connect the attenuator output to the RF Input connector.



**CAUTION.** To prevent damage to the analyzer, do not connect signals greater than +30 dBm (1 Watt).



**CAUTION.** To prevent damage to the analyzer, do not turn the equipment under test on or off with a LISN (Line Impedance Stabilization Network) RF cable connected to the analyzer. The power cycling of the equipment under test causes surge currents that can damage the sensitive input of the analyzer.

2. Connect the appropriate Planar Crown connector if applicable for your instrument model.
3. Connect the signal to the RF INPUT connector.

The maximum limits for the input signal are as follows:

| RSA6100B Series                |   |
|--------------------------------|---|
| Characteristic                 | Description   |
| Maximum DC Voltage             | ± 40 V (RF input)   |
| Maximum Measurable Input Power | + 30 dBm (RF input, RF ATT Auto)<br>75 Watts peak (RF Band, RF ATT Auto (<10 ms pulse width, 1% Duty Cycle repetitive pulses) |

## Changing the RF Input Connector

RSA6100B Series instruments use a Planar Crown® input connector. This connector allows you to easily change the input connector if you need a different connector type or the connector is damaged. The Planar Crown input connector consists of two parts. The first part is the Planar Bulkhead, which is mounted in the instrument front panel. The second part of the input connector is the Planar Crown, which mates to

the Planar Bulkhead. To change the RF input connector, you remove the Planar Crown part of the input connector and replace it with type needed.

To change the RF input Planar Crown input connector:

1. Disconnect the input signal cable, if one is attached.
2. Grasp the Planar Crown connector and turn counterclockwise to remove it. You do not need to use tools to remove the connector.

---

**NOTE.** *Be careful not to touch the inner surfaces of the connector halves.*

---

3. Attach the new Planar Crown connector to the bulkhead connector and tighten finger tight. Do not use tools to tighten the connector.

If you need to clean the connector, use only a 75% isopropyl alcohol solution to clean the Planar Crown connector, if needed. Do not use tap water to clean the connector. Do not submerge the connector in a cleaning solution. Do not use abrasive compounds to clean the connector.

## Connecting a Signal Using a TekConnect Probe

You can connect a signal to the analyzer using a TekConnect Probe by using an RTPA2A Real-Time Spectrum Analyzer TekConnect Probe Adapter. The RTPA2A enables you to use Tektronix P7000 Series active and differential probes to connect to signals. See the RTPA2A Instruction Manual for instructions on how to connect the probe adapter and a probe to the analyzer.

To use the RTPA2A probe adapter with the RSA6100B Series instruments, you may need install a device driver for the RTPA2A probe adapter.

### Installing the RTPA2A Probe Adapter Device Driver

When the RTPA2A probe adapter is connected to the instrument's USB port for the first time, the Windows Add Hardware wizard may automatically start. If the wizard does start, use the procedure below to install the device driver for the new hardware. (If the wizard does not start, the probe adapter device driver is already installed.)

To properly install the probe adapter device driver:

1. Select the option for Windows to install from a specific location (then select **Next**).
2. Only search for the driver in the following location depending on your instrument:
  - **C: \Program Files\Tektronix\RSA6100B\RSA\ProbeAdapter**
3. If Windows warns about Windows Logo testing, select **Continue Anyway**.
4. Select **Finish** after the Found New Hardware wizard completes.

## Connecting External Trigger Signals

External trigger signals are connected to the analyzer using the **Trig In** connector on the front panel. There is also an additional trigger input on the rear panel, the Trig 2 In.

To connect a trigger signal to the analyzer, connect the signal to the **Trig In** connector on the front panel, or the **Trig 2 In** connector on the rear panel.

**Table 15: External trigger signal input requirements**

| Input     | Level            | Impedance   |
|-----------|------------------|---|
| Trig In   | -2.5 V to +2.5 V | 50 $\Omega$ or 5,000 $\Omega$<br>Set in Setup > Configure In/Out > Trigger tab. |
| Trig 2 In | +5 V (TTL)       | 10,000 $\Omega$   |

### Enabling External Trigger

To use an external trigger signal, select either **Trig In (front)** or **Trig 2 In (rear)** in the **Source** control on the Trigger control panel's [Event tab \(see page 594\)](#).

## Digital I and Q Outputs

The Digital I and Q outputs are a means for you to obtain the analyzer digital baseband I/Q signals in a real-time data stream. This output enables you to use the analyzer as a down-converter and then pass the down-converted baseband I/Q signals to external equipment for either additional processing or storage.

The rear panel I and Q output connectors are LVDS connectors. For full details on the connectors, see the Specifications and Performance Verification manual. A PDF file on the manual is located on the Documentation CD that shipped with your instrument. The manual can be downloaded from the Tektronix Web site, [www.tektronix.com/manuals](http://www.tektronix.com/manuals).

---

**NOTE.** *The Digital I and Q outputs are disabled when an alignment is running.*

---

## IF Output

The Analog IF Output is a means for you to observe the analyzer analog IF signal with an oscilloscope.

The output level is amplified and filtered especially for use with oscilloscopes. The output frequency varies with changes in center frequency of the analyzer. Additionally, the output spectrum of the IF is inverted from the input spectrum to the analyzer. This means that a signal above the center frequency of the input is converted to a frequency below the center of the IF. The output-frequency variations and spectral inversion makes the IF output unsuitable for use with receivers.

The IF bandwidth can be switched between 150 MHz (wide open) and 60 MHz (Gaussian-shaped to -12 dB). The 'wide-open' selection will result in the fastest transient response for measurement of rapidly changing levels such as pulse rise times. However, the 150 MHz filter may create significant overshoot and ringing in the output signal. To control transient responses, a 60 MHz filter with an approximately Gaussian shape is available. Selection of the 60 MHz filter reduces IF-induced overshoot to less than approximately 5% for a rectangular-shaped input envelope.

---

**NOTE.** *Spectral inversion is performed in the IF conversion process. Signals above the center frequency of the instrument will appear below the center frequency of the IF output.*

---

---

**NOTE.** *The IF output is disabled when an alignment is running.*

---

## Other Outputs

The Other Outputs tab is accessed from the Config In/Out control panel (Setup > Configure In/Out > Other Outputs tab). The IF Output and IQ Output settings appear only when Option 05 is installed.

The screenshot shows the 'Other Outputs' configuration panel with three main sections:

- 28V DC power:** Two radio buttons, 'Off' (selected) and 'On'.
- IF Output:** Two radio buttons, 'Off' (selected) and 'On (500 MHz)'. Below them is a dropdown menu for 'BW:' set to '120MHz (Flat-Top)'.
- IQ Output:** Three radio buttons: 'Off' (selected), 'On', and 'Test (Counting Pattern)'.

### 28V DC power

The 28V DC power provides a voltage source to power an external noise source. The output connector is located on the rear panel and provides 28 VDC at 140 mA.

### IF Output

The IF Output enables you to use the analyzer as a high-quality down converter. The instrument shifts the RF input signal down to a frequency that an oscilloscope can display.

The BW drop-down list allows you to apply filters to the signal. The 120 MHz flat-top filter selection provides the widest possible bandwidth out, which means faster rise times but results in more overshoot and other aberrations. The Gaussian filter reduces the bandwidth out and reduces aberrations. The BW setting is enabled only when the IF Output is On. The selections are **120 MHz (Flat-Top)** and **60 MHz (Gaussian)**. Default: 120 MHz.

### IQ Output

The IQ Output transmits the digitized analyzer data for capture and storage by external equipment. The I/Q data is provided in 16-bit, two's complement format, along with clock and data-valid signals.

There are three selections available for the IQ Output: Off, On, and Test. When in the “Off” position, the IQ Output is disabled, and all lines are static. When in the “On” position, the data is output as described in the Specifications and Performance Verification manual.

Use the “Test” position to verify operation when setting up your equipment. When in the “Test (Counting Pattern)” position, the I channel data will be counting up and the Q channel data will be counting down. The EXT\_IQ\_CLK and EXT\_IQ\_DAV signals will operate at a rate consistent with the instrument SPAN setting.

---

**NOTE.** The analog IF output and IQ outputs are disabled when an alignment is running.


---

## Mapping Measurements

The RSA Map software lets you record measurements on a geo-referenced map or a bitmap map for documentation purposes. If an external GPS receiver is connected to the analyzer, measurements can be mapped at the GPS coordinates where the measurement was taken. Using an external GPS receiver, RSA Map can perform repeat measurements using either time or distance to initiate measurements. RSA Map software can also save results for future reference or analysis.

To map measurements:

- Select Tools > RSA Map. This launches the RSAMap software in a separate window.



Full documentation on how to use RSA Map is provided in the RSA Map Help. In RSA Map, select **Help > User Manual**.

## Menu Overview

The main menus are:

| Menu                                   | Description  |
|--|--|
| <a href="#">File (see page 623)</a>    | Select measurements, open and save files, print documents, and preset.                             |
| <a href="#">View (see page 627)</a>    | Change display size, display the Marker toolbar and Status bar.                                    |
| <a href="#">Run (see page 629)</a>     | Start, stop and abort acquisitions, select single or continuous acquisition mode.                  |
| <a href="#">Replay (see page 564)</a>  | Replay measurements; select which record(s) to play, Replay/Stop/Pause, or enable continuous loop. |
| <a href="#">Markers (see page 631)</a> | Define markers and search for signal peaks.  |

| Menu                                  | Description   |
|---------------------------------------|---|
| <a href="#">Setup (see page 631)</a>  | Change settings for acquisition, trigger, analysis, measurements, amplitude, alignments, and configure in/out settings. |
| <a href="#">Presets (see page 22)</a> | Load and configure presets.   |
| <a href="#">Tools (see page 632)</a>  | Perform searches, access diagnostics, install upgrades and configure user preferences.                                  |
| <a href="#">Window (see page 634)</a> | Controls the size and layout of displays within the analyzer application.   |
| <a href="#">Help (see page 634)</a>   | Access the help, display information about the instrument hardware and software.  |

## File Menu

| Command  | Description   |
|--|---|
| <a href="#">Recall (see page 623)</a>                  | The Recall dialog enables you to recall saved data, setups and traces.  |
| <a href="#">Save (see page 624)</a>                    | Saves a file without asking for a file parameters (based on most recent settings).  |
| <a href="#">Save As (see page 624)</a>                 | Displays the Save dialog enabling you to specify the parameters of the save operation.  |
| <a href="#">Save on trigger (see page 605)</a>         | Displays the Actions tab of the Trigger control panel which allows you to configure the Save on Trigger function.   |
| <a href="#">Acquisition save options (see page 30)</a> | Displays the Save and Export tab of the Options control panel which allows you to specify how much data is saved in acquisition files.  |
| <a href="#">FastSave Acquisitions (see page 584)</a>   | Displays the FastSave tab of the Acquire control panel which allows you to configure FastSave acquisitions.   |
| <a href="#">Acquisition data info (see page 563)</a>   | Displays the <a href="#">Acquisition Info tab (see page 563)</a> of the Replay control panel. The info on this tab describes such acquisition parameters as acquisition bandwidth, sampling rate, RF attenuation, and acquisition length. |
| <a href="#">Measurement Data Info (see page 626)</a>   | Displays the characteristics of the most recently analyzed record in the display.   |
| <a href="#">Print (see page 626)</a>                   | Prints the selected display.  |
| <a href="#">Print Preview (see page 626)</a>           | Displays a preview of the print output.   |
| Exit   | Closes the analyzer application. Selecting Exit does not shut down the instrument.  |

## Recall

**Menu Bar:** File > Recall


**Front Panel:** Recall

**Application Toolbar:** 

Use the Recall command to load previously saved acquisition data, setups or trace data.

### To Recall Data or Setups

1. Select **File > Recall**. This displays the **Open** dialog box.



2. Select the file type to be recalled and click **Open**.

## Save / Save As

### Menu Bar: File > Save / Save As

Use Save / Save As to store acquisition data, setups, and traces. Save is also used to export traces, results and pictures of the display for use in other programs.

### Difference Between Save and Save As

Use Save As to specify what kind of data you want to save and where the data should be saved. Use Save to quickly save the same data as you saved the last time you executed a Save, without having to specify the data type and location.

For example, suppose you want to save a picture of a spectrum trace each time you adjust a circuit to document how the adjustments affect the output of the circuit. The first time you want to save a picture of the display, you will need to select Save As. From the Save As dialog box, you specify the type of data you want to save (Picture of Selected Display) and specify the location of the saved file. As long as the [Save and Export option](#) (see page 29) is set to automatically name saved files, the next time you want to save a picture of the display, you can just press Save on the front panel and a picture of the selected display will be saved without requiring you to type a file name or the location of the file to be saved.

## What Data Types Can Be Saved

| Data type                           | Description  |
|-------------------------------------|--|
| Acquisition Data                    | Data collected during acquisition that can be recalled for later analysis. Data is saved in a format readable only by RSA6100/RSA5100/SPECMON Series instruments or oscilloscopes running SignalVu software. You can save acquisition data as IQ records only, DPX spectra (DPXogram data) only, or both IQ records and DPX Spectra. |
| Setup                               | Configuration information detailing instrument settings. Data can be saved in a format readable only by SignalVu-PC, oscilloscopes running SignalVu software, or RSA6100/RSA5100/SPECMON Series instruments.   |
| Selected Trace                      | Saves the selected trace for later analysis by the analyzer. Data is saved in a format readable only by SignalVu-PC, oscilloscopes running SignalVu software, or RSA6100/RSA5100/SPECMON Series instruments.   |
| Exported Traces and Numeric Results | Save traces and results in a file format that can be used by other programs.   |
| Pictures of the Display             | Save screen images in graphic image file formats that can be used in other programs.   |
| Exported Acquisition Data           | Save acquisition data records in a file format that can be used by other programs. Acquisition data can be saved in either comma-separated-variable format or MATLAB format.   |


[Data, Settings, and Picture File Formats \(see page 609\).](#)

## Options for Saving Pictures of the Display

| Option       | Setting | Description   |
|--------------|---------|---|
| Image Format | PNG     | Saves exported screen captures in Portable Network Graphics format.               |
|              | JPG     | Saves exported screen captures in Joint Photographic Experts Group (JPEG) format. |
|              | BMP     | Saves exported screen captures in Windows bitmap format.                          |

## Measurement Data Info

The Measurement Data Info command in the File menu displays a listing of acquisition-related information about the last data analyzed by the selected measurement. The last data can be from the current acquisition or it could be from a recalled data file.



## Print

**Menu Bar:** File > Print

**Front Panel:** Print

Print displays the Windows Print dialog box for printing a screen capture of the display. To save ink when printing, use the Colors tab to set the color scheme. See [Options Settings \(see page 29\)](#).



## Print Preview

**Menu Bar:** File > Print Preview

Print Preview shows how a print out will appear when it is printed.

## View Menu

The View menu enables you to control the display of items in the application window.

| Command  | Description   |
|--|---|
| <a href="#">Full Screen (see page 628)</a>     | Toggles all views between full-screen size and user-selected size.  |
| Navigator View                                 | <p>Selecting Navigator View adds the Time Overview display to the existing measurement displays to provide a better perspective of the signal.</p> <p>When the Navigator View is enabled, the Time Overview display is always located above any other measurement displays and uses the maximum horizontal resolution of the display area, regardless of the Window Tile setting.</p>   |
| Numeric Keypad                                 | <p>When selected, any time you place the cursor within a control or setting that takes an numeric value, you are presented with a dialog box to easily enter a value.</p> <p>For instance, placing the cursor in the Frequency setting, displays the following dialog screen.</p>  <p>Use the arrow buttons to change the value.</p> <p>Press the calculator icon to display a keypad that allows you to enter values and suffixes.</p> |
| <a href="#">Replay Toolbar (see page 566)</a>  | Shows or hides the Replay toolbar.  |
| Traces Toolbar                                 | <p>Shows or hides the Traces toolbar.</p> <p>The Traces toolbar allows quick access to the trace controls of the selected measurement display. This example is the DPX Spectrum trace toolbar.</p>  <p>The Traces toolbar can also be controlled from the Basic Controls Toolbar.</p>   |
| <a href="#">Markers Toolbar (see page 543)</a> | <p>Shows or hides the Marker toolbar. Enables you to define Markers and perform Peak searches.</p> <p>The Markers toolbar can also be controlled from the Basic Controls Toolbar.</p>   |

| Command                                   | Description  |
|---|--|
| Basic Controls Toolbar                    | <div>Shows or hides the toolbar containing the basic acquisition settings and controls.</div> <div><div><div>■</div> Name of selected measurement window</div><div><div>■</div> Various controls based on the selected measurement display</div><div><div>■</div> Show or hide the Markers toolbar</div><div><div>■</div> Show or hide the Traces toolbar</div><div><div>■</div> Show or hide the Settings Control Panel</div></div> <div>The Basic Controls toolbar allows quick access to controls and settings of the selected measurement display. This example is the DPX Spectrum Basic Controls toolbar.</div> <div><div><div>DPX Spectrum</div><div>Frequency 1.85000 GHz</div><div>Ref Lev 1.00 dBm</div><div>Span 10.00 MHz</div><div>Res BW 100 kHz</div><div>Markers</div><div>Traces</div><div></div></div></div> |
| <a href="#">Status Bar (see page 628)</a> | Shows or hides the Status bar.   |

Full Screen


Menu Bar: View > Full Screen

When unchecked, clicking **Full Screen** resizes the application window to fill the screen. Full Screen mode maximizes the application window, and turns off the application title bar.

When checked, clicking **Full Screen** restores the application window to its previous size. The application title bar is restored.

Status Bar

The Status Bar displays information on specific instrument settings. It contains only status information; it does not display any error information. The Status bar has no controls. It can be hidden.



## Elements of the Status Bar

| Area within Bar             | Description   |
|-----------------------------|---|
| Run Status                  | Displays the running state. For example, some run states are Arm, Ready, Triggered, Analyzing, or Stopped.  |
| Status message area         | Displays status messages (for example, ADC Overrange). When there no status messages to report, displays current acquisition parameters (for example, Acq Length: 22.320 us).   |
| Acquisition mode            | Displays either Real Time or Swept. Indicates how the next acquisition will be acquired.  |
| Trigger Settings            | Displays Trigger settings. For example, you will see Free Run, Power, Trig In (front), Frequency Mask and others.   |
| Frequency Reference Setting | Indicates whether the internal or external reference source is selected. Displays Ref: Int or Ref: Ext.   |
| Internal Attenuator setting | Displays the setting of the internal attenuator.  |
| Preamplifier Setting        | If the optional preamp is installed, this indicates whether the preamp is on or off. If the preamp is on, displays Preamp: ON, 3 GHz max. as a reminder that the input frequency range is restricted when using the preamp. |

## Run Status Indicators

| Indicator | Description  |
|-----------|--|
| Acquiring | The analyzer is capturing the signal.  |
| Analyzing | The analyzer has captured the signal and is processing the signal record.  |
| Arm       | The trigger circuitry is filling the pretrigger portion of the signal record.  |
| File Data | The analyzer is analyzing recalled acquisition records.  |
| Replaying | The analyzer is analyzing recalled waveform or acquisition records.  |
| Stopped   | If Stopped is displayed, Signal acquisition has been halted. This can occur because the Run button has been pressed or because a trigger event has occurred, signal acquisition has occurred and Run mode was set to Single. |
| Triggered | If Triggered is displayed, the analyzer has recognized a valid trigger and is filling the posttrigger portion of the signal record.  |

## Showing or Hiding the Status Bar

- Select **View > Status Bar** to toggle the display of the Status bar.

## Run Menu

The Run menu provides access to commands that control the signal acquisition.

| Command  | Description  |
|--|--|
| <a href="#">Run (see page 577)</a>             | If acquisition mode is Stopped, selecting Run begins a new measurement/acquisition cycle. If acquisition mode is Run, pressing Run halts the current measurement/acquisition cycle after it completes.   |
| Resume   | Restarts data acquisition, but does not reset accumulated results, such as Average or MaxHold. This allows you to stop acquisitions temporarily, then continue. If the accumulation is already complete, for example, 10 acquisitions or 10 averages have already been completed, each subsequent Resume command will cause one more acquisition to be taken, and its results added to the accumulation. Resume is not available if instrument settings have been changed. |
| <a href="#">Abort (see page 577)</a>           | Immediately halts the current measurement/acquisition cycle.   |
| <a href="#">Single Sequence (see page 630)</a> | Selects the single-sequence acquisition mode. This is only a mode selector; it does not initiate an acquisition.   |
| <a href="#">Continuous (see page 630)</a>      | Selects the continuous acquisition mode. Selecting Continuous does initiate acquisitions.  |

## Single Sequence Acquisition Mode

### Menu Bar: Run > Single

Selecting **Single** sets the Run mode to Single. In Single mode, as soon as one acquisition sequence completes, acquisition stops.

Note that a single acquisition sequence can require more than one acquisition. For example, in a spectrum view, the trace function might be set to Average 100 acquisitions. Thus, a complete acquisition sequence would consist of 100 acquisitions to produce 100 intermediate traces that are averaged together to create the final trace that is displayed. Once the 100 acquisitions have been completed, acquisition stops.

This is only a mode selector, it does not initiate an acquisition.

## Continuous Acquisition Mode

### Menu bar: Run > Continuous

Selecting **Continuous** places the analyzer in the Continuous acquisition mode. In Continuous mode, the analyzer acquires and displays acquisitions repeatedly. The Continuous and Single Sequence acquisition modes are mutually exclusive.

Selecting Continuous restarts acquisitions.

## Replay Menu

The [Replay Menu \(see page 564\)](#) allows you to rerun measurements using the data in acquisition memory or a recalled TIQ file. You can use this to compute new results for old data after you change settings or measurements.

## Markers Menu

The Markers menu provides to settings that define and control the location of markers.

| Setting                                       | Description   |
|---|---|
| <a href="#">Peak (see page 541)</a>           | Moves the selected marker to the highest peak on the trace.             |
| Next Peak >                                   | Moves the selected marker to next peak depending on the setting chosen. |
| Marker to Center Frequency                    | Sets center frequency to the frequency of the selected marker.          |
| <a href="#">Define Markers (see page 541)</a> | Displays the Define Marker control panel.                               |

## Setup Menu

The Setup menu provides access to control panels that specify parameters for numerous signal analyzer functions.

| Command   | Description   |
|---|---|
| <a href="#">Displays (see page 33)</a>          | Displays the Displays control panel.                          |
| <a href="#">Settings (see page 646)</a>         | Displays the Settings control panel for the selected display. |
| <a href="#">Trigger (see page 593)</a>          | Displays the Trigger control panel.                           |
| <a href="#">Acquire (see page 578)</a>          | Displays the Acquire control panel.                           |
| <a href="#">Analysis (see page 555)</a>         | Displays the Analysis control panel.                          |
| <a href="#">Amplitude (see page 569)</a>        | Displays the Amplitude control panel.                         |
| <a href="#">Configure In/Out (see page 616)</a> | Displays the Configure In/Out panel.                          |

## Presets Menu

The Presets menu provides you access to instrument presets and preset options.

| Command                                      | Description   |
|--|---|
| <a href="#">Main (see page 22)</a>           | Resets the instrument to factory defaults. Acquisition data and settings that have not been saved will be lost. |
| DPX  | Recalls the DPX presets.  |
| <a href="#">Standards (see page 25)</a>      | Recalls the Standards presets.  |
| <a href="#">Application (see page 22)</a>    | Recalls the Application presets.  |
| <a href="#">User (see page 25)</a>           | Recalls the User presets.   |
| <a href="#">Preset Options (see page 22)</a> | Displays a control panel where you can configure options for Presets.   |

## Tools Menu


Provides access to several utilities for controlling instrument functions.

| Command   | Description  |
|---|--|
| <a href="#">Mask Test (see page 547)</a>        | Enables you to locate and highlight specified signal levels in specified displays. |
| <a href="#">Alignments (see page 21)</a>        | Displays the Alignments dialog box.  |
| <a href="#">Diagnostics (see page 633)</a>      | Displays the Diagnostics dialog box.   |
| <a href="#">Install Upgrades (see page 634)</a> | Used to activate upgrades you have purchased.                                      |
| <a href="#">Options (see page 29)</a>           | Displays the Options control panel.  |

## Alignments Dialog

### Menu Bar: Tools > Alignments

The analyzer can adjust itself, optimizing the internal signal path used to acquire the signal you measure. Alignment optimizes the analyzer to take more accurate measurements based on the ambient temperature. The Alignments dialog enables you to specify when alignments occur, either automatically or only when manually initiated.




## Elements of the Alignments Dialog

| Setting  | Description   |
|--|---|
| Automatically align as needed                          | Specifies that alignments run as ambient conditions require.        |
| Run alignments only when "Align Now button is pressed. | Specifies that alignments will never run unless manually initiated. |
| Align Now  | Use to manually initiate an alignment.                              |
| Abort  | Stops the alignment process.  |
| Close  | Closes the Alignments dialog box.                                   |

## Diagnostics

### Menu Bar: Tools > Diagnostics

The Diagnostics menu item displays the Diagnostics Tests window, which enables you to run variety of tests to verify the operation of the signal analyzer. For detailed information on running diagnostics, refer to the Service Manual. The manual can be downloaded from the Tektronix Web site, [www.tektronix.com/manuals](http://www.tektronix.com/manuals).



## Install Upgrades

You can install upgrades on your analyzer as your needs change. To install upgrades, you will need an Option Installation Key. For information about available options or to make a purchase in North America, call 1-800-833-9200. Worldwide, visit [www.tektronix.com](http://www.tektronix.com) to find contacts in your area.

Full instructions on how to install your upgrade and activate the option with the Option Installation Key are provided with the upgrade kit.

## Window Menu

Use the Window menu to arrange how windows are displayed. Displays can be set to appear full screen (one display at a time) or with all (selected) displays visible at once. When all displays are visible at once, you can rearrange the displays by dragging the title bar of a window (deselect Window > Lock Windows to move displays around).

| Command           | Description  |
|-------------------|--|
| Close View        | Closes the selected view.  |
| Cascade           | Positions windows in a cascade view (not available when Lock Windows is selected).   |
| Tile Horizontally | Positions windows in a horizontal orientation (top to bottom).   |
| Tile Vertically   | Positions windows in a vertical orientation (side by side).  |
| Lock Windows      | Locks the windows into their current position, preventing them from being moved. If the windows are locked, the Cascade arrangement is not selectable. |
| (List of windows) | A numbered list of open windows.   |

## Help Menu

The Help menu provides access to the help and version information about the analyzer hardware and software.

| Command                             | Description  |
|-------------------------------------|--|
| User Manual On-line                 | Displays the help.   |
| User Manual (PDF)                   | Displays a PDF version of the help.                                  |
| Application Reference (PDF)         | Displays a PDF version of the analyzer Applications Manual.          |
| Quick Start Manual (PDF)            | Displays a PDF version of the analyzer Quick Start User Manual.      |
| About Tektronix Real-Time Analyzers | Displays information about the analyzer model and installed options. |

## Error and Information Messages

The following list describes some of the common error and information messages that might appear during instrument operation. Messages that apply specifically to one or more measurements appear in the displays. Messages that pertain globally, such as those about hardware status, are shown in the Status Bar at the bottom of the analyzer application window.

### Acq BW too small for current setup

The display needs a wider acquisition bandwidth than what the current data record contains. This can be due to any of the following reasons:

- The sampling parameters are being manually controlled.  
In the Acquire control panel > Sampling Parameters tab, set the Adjust control to All Auto to allow the software to pick the sample rate and record length that it needs.
- A display other than the one you intended has been selected. The selected display has requested a smaller acquisition bandwidth to achieve a better accuracy or dynamic range for its particular measurement.  
Select the display that contains the message. Click Run if the instrument is not already acquiring data.
- Acquisitions are not running and the measurement now requests a wider bandwidth than the last acquisition.  
Click Run to perform a new acquisition with a wider bandwidth.
- The data is from a recalled TIQ file.  
There is no way to increase the acquisition bandwidth for saved IQ data. You must adjust the measurement settings so that less bandwidth is required.
- The current Frequency setting is different than that of the data record, causing the measurement bandwidth to fall at least partly outside the bandwidth of the data.

### Analysis failure: <description of error>

The instrument is unable to complete a measurement due to difficulty in characterizing the signal. For example, due to either the signal or settings, the instrument may not be able to recognize a pulse so it can compute the pulse measurements.

- Try changing settings to improve analysis. For example, when Pulse Trace is displayed, try changing the settings on the Settings > Params.

**Analysis failure: Carrier detection failed**

The instrument was unable to locate a carrier signal. Try adjusting the Carrier Threshold or Integration BW values, if the measurement has these controls.

**Analysis Length was limited**

This message appears if the "Results Length", the time over which the measurement computed its results, is less than the Actual Analysis Length reported in the Analysis Time control panel tab and the Time Overview display. The Results Length is indicated by the magenta line below the Time Overview graph. This can occur because the acquisition contained more data samples than a measurement can process (for example, digital demodulation is limited to 80,000 samples) or the measurement had to use some of the first and last samples for pre-measurement processes.

**Analysis length too small for current setup**

Increase the Analysis Length or decrease the RBW (Spectrum, ACPR, MCPR).

**Avg Tx not available in Volts units**

The Average Transmitted Power measurement is not defined for Volts.

**Cache file missing. Unable to create new cache file. Not enough free space (need XXX GB). Remove unnecessary files, defragment the hard drive using a standard disk management utility, then relaunch application.**

The instrument is unable to create a required cache file. To resolve this issue, use a standard disk management utility to defragment the hard disk. If this error persists, backup your data (including any setup files you have saved), any applications and associated files you have installed and perform an OS restore on the hard disk. After restoring the OS, reinstall the analyzer product software.

---

**NOTE.** The specific OS restore procedure required for your RSA6100B Series instrument depends on when your instrument was manufactured. Go to [www.tektronix.com/manuals](http://www.tektronix.com/manuals) and download the RSA6100B Series Service Manual, Tektronix part number 077-0648-00 or later. Review the OS restore procedure in the manual to determine the procedure appropriate to your instrument.

---

**Can't get acquisition data record**

No acquisition record currently exists in memory (run an acquisition or recall a file), or an error has occurred. Repeat the acquisition.

**Can't replay data from swept acq**

The measurement could not produce results because it is a real-time only measurement but was asked to reanalyze (Replay) an acquisition taken in swept mode.

**Can't replay. Live data needed for swept settings**

The measurement could not produce results because it was asked to reanalyze (Replay) the acquisition but it is in Swept acquisition mode. Swept-mode measurements cannot analyze recalled data.

**CISPR accuracy limited by acq memory. Adjust RBW or freq range**

The CISPR function was applied, but the available data did not represent a long enough time to satisfy CISPR requirements. Increasing RBW reduces the amount of time needed for analysis. Reducing frequency range (for example, by reducing Span), decreases sample rate, allowing the available memory to cover a longer time period.

**Data acquired during hardware failure**

An acquisition was completed even though a hardware failure was detected. This message refers to the acquisition data currently being analyzed, but not necessarily to the current status of the instrument. See the explanation for "Hardware Failure".

If this data is from a saved file, the error cannot be cleared.

**Data acquired during RF ADC overrange**

An acquisition was completed but the signal was outside the range of the analog-to-digital converter at the time this data was acquired. This message refers to the acquisition data currently being analyzed, but not necessarily to the current status of the instrument. See the explanation for "RF ADC overrange".

If this data is from a saved file, this error cannot be cleared.

**Data acquired during RF digital gain overflow**

This data was acquired when the input signal contained peaks greater than 6 dB above the Reference Level setting.

See the definition for "RF digital gain overflow".

If the data is from a file, this error cannot be cleared.

**Data acquired during warm-up**

This data was acquired while the instrument was warming up. The warm-up period is 20 minutes. Until this warm-up period is completed, the instrument is considered uncalibrated.

**Data from unaligned instrument**

The acquisition data was captured when the instrument was not aligned. This message refers to the acquisition data currently being analyzed, but not necessarily to the current status of the instrument. Measurements made on this data might not be accurate.

**Data from uncalibrated instrument**

The acquisition data was captured when the instrument was not calibrated. This message refers to the acquisition data currently being analyzed, but not necessarily to the current status of the instrument. See the definition for "Not calibrated".

If this data is from a file, the error cannot be cleared.

**Disabled: data is from swept acquisition**

The display needs to run in real-time mode. The display associated with this message cannot run now because it is not the selected display, and the selected display is performing multiple acquisitions (it is in swept mode).

- Change the settings of the selected display so it is performing real-time acquisitions.

- Select the associated display to make it the selected display. When it is selected, it will force the acquisition parameters to change to meet its own needs.

**Disabled: Freq Mask Trigger in use**

The DPX display cannot run while the instrument is set to use Frequency Mask Trigger.

- To run the DPX display, you must turn off Frequency Mask Trigger.

**DPX disabled: RefLev too low for Volts/Watts units**

The Reference Level value is set too low. It is possible to set a RefLev value so low that DPX can't run if volts or watts are the selected amplitude units. Reset the Reference Level to a higher value.

**Frequency exceeds preamp range**

This is a warning that signals below the minimum preamp operating frequency are likely to be severely attenuated (this is 100 kHz for option 51, 1 MHz for option 50).

**Needs swept acq or larger AcqBW - Acquire data while display is selected**

The display is not running because it needs to perform multiple acquisitions (it must be in swept mode) but it is not the selected display. Only the selected display can perform multiple acquisitions.

- Select the display showing this message to give it control of acquisitions parameters.

Sometimes, only one display can work and the others will be blank and show errors. This happens when different displays have conflicting demands on the acquisition data record.

**Dither: manual control**

This is a reminder that the [Dither control \(see page 580\)](#) has been set to manual, which might affect the quality of measurement results.

**Export failure: file not saved**

An error occurred while exporting Results. The file was not created.

- Export the results again.

**Export failure: unable to open results file for export. File not saved.**

The Export Results file could not be opened for writing, so the export of results was not completed.

- Verify that there is sufficient free space on hard disk.

**External Frequency Reference signal not valid.**

The instrument is no longer locked to the external reference signal because it can no longer detect a signal with proper amplitude and frequency.

- Verify that the external reference signal is still connected to the instrument.

**Hardware failure detected by Diagnostics**

A hardware failure was detected while running Diagnostics. To see a description of the event initiated by the failure, [display the Windows Event Viewer \(see page 643\)](#). There is a shortcut to the Windows Event Viewer on the Windows desktop.

**Hardware failure - see Windows Event Viewer**

A hardware failure has occurred. To see a description of the event initiated by the failure, [display the Windows Event Viewer \(see page 643\)](#). There is an shortcut to the Windows Event Viewer on the Windows desktop.

**Insufficient data for CISPR. Acquire while display is selected**

This message appears when a measurement is not the selected measurement and CISPR filters are selected. Set the measurement to be the selected measurement and reacquire the signal.

**IQ Processing Error: 8012**

This message occurs in GP Digital Modulation displays. The most likely cause for this message is that there are not enough symbols to analyze. This can happen if:

- The Analysis Length is set too short. Increase the Analysis Length on the Analysis Time tab.
- The Analysis Offset has pushed the Analysis Time so far out that the actual Analysis Length is too short, even though the user-requested Analysis Length would have been long enough without the excessive offset. Decrease the Analysis Offset on the Analysis Time tab.
- The input signal is bursted, and the burst does not contain enough symbols.

**Locking to ext freq ref signal**

This a status message displayed while the instrument is attempting to lock to an external reference signal. The message disappears when the instrument locks to the reference signal.

**Needs swept acq or larger Acq BW - Acquire data while display is selected**

The display has one of two problems: It is not the selected display, which prevents it from controlling the hardware acquisition parameters, and setting the acquisition mode to Swept; or its settings require a wider data bandwidth.

- Select the display that you are interested in and it will change the acquisition to meet its own needs.
- Increase the acquisition bandwidth manually or by changing the selected display's settings to cause the wider bandwidth.

**No Math trace: unmatched trace lengths**

A math trace could not produced because the traces selected to generate the math trace do not have the same number of points. This can easily happen if both traces are recalled, but were saved under different "Points" settings. This can also occur if one of the selected traces is a live trace and the other trace is a recalled trace.

- In a Spectrum display, as long as one trace is live, you can change the "Points" setting (Setup > Settings > Freq & Span tab) to match the recalled trace.
- If you are using two recalled traces to generate the Math trace. You must recreate at least one of the traces.

**Not aligned**

The instrument needs to be aligned. When the instrument is not aligned, measurements might not be accurate.

- Select Tools > Alignments > Align Now to alignment the instrument.

**No burst detected**

The [Burst Detection Mode \(see page 521\)](#) is On, but no burst was detected in the signal.

- Check that the Threshold setting is properly set.

**Not calibrated**

The instrument requires calibration. Calibration can be performed only at a Tektronix Service center. [Contact Tektronix](#) ) to schedule a calibration.

**Not enough samples for current setup**

The measurement was not able to run because the combination of analysis length, offset, and measurement bandwidth relative to acquisition bandwidth, were such that not enough samples were available for the measurement to analyze. This can occur when two or more displays are shown and one display requires a wide acquisition bandwidth and another display requires a much narrower bandwidth. The display requiring the much narrower bandwidth must decimate and filter the acquisition record which can result in too few samples left for the measurement.

- Increase the Analysis Length to provide more samples.
- Close any displays you don't currently need.

**Not enough samples – try increasing MeasBW**

The measurement was not able to run because there are not enough samples available for the measurement to analyze. The Settling Time measurement requires at least 256 samples. This can occur when two or more displays are shown and one display requires a wide acquisition bandwidth and another display requires a much narrower bandwidth. The display requiring the much narrower bandwidth must decimate and filter the acquisition record which can result in too few samples left for the measurement.

- Increase the Analysis Length to provide more samples.
- Close any displays you don't currently need.

**No FFT (not all pulses have results)**

If a pulse cannot be measured (because its shape is too indistinct or it does not meet the [parameters that define a pulse \(see page 326\)](#)), its results will be “-” for every measurement on that pulse. The instrument cannot compute an FFT.

**No pulses found**

The instrument was unable to find any complete pulses in the signal.

- Make sure the [analysis length \(see page 555\)](#) includes at least one complete pulse cycle, from before one rising edge until after the next rising edge.

**Pulse Detection Error**

The instrument was unable to detect a pulse.

- The pulse [Measurement Filter \(see page 326\)](#) needs to be smaller. Try reducing the bandwidth and/or selecting the Gaussian filter.

- Detection threshold is not set to the proper level for the signal. Adjust the [Power threshold to detect pulses \(see page 326\)](#).
- The pulse interval is too long for the current settings. Try decreasing the [filter bandwidth \(see page 326\)](#), as this may reduce the number of data points to a manageable quantity.

**RBW conflict. Increase Span or Analysis Length**

The measurement is not running because the actual RBW used by the measurement is too large for the current acquisition span. Typically, the analysis length is too short as well.

- Either increase the span or increase the Analysis Length.

**RBW decreased**

The current span or acquisition bandwidth is too small to allow a wider RBW filter.

- Increase the span or acquisition bandwidth if the decreased RBW is not acceptable.

**RBW increased**

The current Spectrum Length (or Analysis Length if Spectrum Length is not Independent) is too small to allow the requested RBW.

- Increase the Spectrum Length (or Analysis Length) if the increased RBW is not acceptable.

**RBW limited by AcqBW to: XX Hz**

The requested RBW is too close to the acquisition BW. Increase the frequency range of the measurement (for example, Span).

**RBW too small/large for current Acq BW**

If the RBW is set manually, it is possible for the acquisition bandwidth to be incompatible with the RBW setting.

- Change the RBW setting.
- Adjust the Acq BW setting, either directly (Setup > Acquire > Sampling Params: select on of the manual modes) or by adjusting the measurement bandwidth of the selected display (Setup > Acquire > Sampling Params: All Auto).

**Recall error: Setup not completely restored**

An error occurred while recalling a Setup file. Thus, the current setup may be a combination of settings from the Setup file and the previous Setup.

- Recall the setup again.

**Recall failure: problem with file or file contents**

An error occurred while recalling a Setup, Trace or Data file. This can occur because of a problem opening the file (operating system error) or because of a problem with the contents of the file.

- Recall the file again.

**Restoring acquisition data**

This is a status message displayed while data is being restored from a file.

**RF ADC overrange**

This input signal power is outside the range of the analog-to-digital converter circuitry.

- Decrease the input signal power or increase the internal attenuation setting.
- Increase the Reference Level setting.

**RF Attenuator: manual control**

This message is just a reminder that the [internal attenuator \(see page 569\)](#) has been set to manual mode.

**RF digital gain overflow**

The input signal contains peaks greater than 6 dB above the Reference Level setting.

- Increase Reference Level
- Decrease input signal power

**Save failure: file not saved**

An error occurred while saving a Setup or Data file.

- Save the file again.

**Saving acquisition data**

This is a status message displayed while data is being saved to a file.

**Selected VBW does not use full Spectrum Length**

This message can occur when the Spectrum Length is greater than required for the VBW filter. If you look at the Time Overview display, the Magenta line for Results Length indicates the part of the Spectrum Length that was actually used. The measurement results are correct, but don't include some of the data in the selected Spectrum Length. To clear this message, you can set the Spectrum Length to Auto.

**Setup error: <description of error>**

When this message appears, it includes text that explains the problem. For example, the ACPR display might show: "Setup error: channels can't overlap". Setup errors are the result of conflicts in instrument settings. These types of errors occur when a user makes manual changes to settings. In the example above, the channel settings in the Channel Power and ACPR display have been set so that the channels overlap in frequency.

- Adjust the instrument settings, or change a setting back to Auto, to eliminate the error.

**Setup error: Integration BW exceeds Measurement BW**

When this message appears, it includes text that explains the problem. Setup errors are the result of conflicts in instrument settings. These types of errors occur when a user makes manual changes to settings.

- Adjust the instrument settings, or change a setting back to Auto, to eliminate the error.

**Setup error: Measurement time for Freq & Phase results**

The [Measurement time for Freq & Phase results \(see page 328\)](#) specifies how far across the pulse top the instrument should wait before measuring the Phase Difference and Frequency Difference for each pulse. If this value is set too large for any of the pulses in the signal, the measurement point ends up on the falling edge or during the pulse off time.

- Decrease the Measurement time for Freq & Phase results setting (Settings > Define tab).

#### **Swept: RF Trig invalid for most signals**

This is an status message that appears when the instrument is in swept acquisition mode and Power Trigger mode at the same time. When the acquisition is swept, it captures the requested bandwidth using multiple acquisitions, each for a different portion of the bandwidth. The trigger system evaluates each of these separate acquisitions for trigger events, so unless a valid event exists in each of these smaller frequency ranges, the acquisition will not complete.

**Unexpected software error. Please cycle power and try again. If the problem persists, contact your Tektronix Service Center.**

An unrecoverable error has occurred, and the instrument application software will shut down.

- Switch the instrument off and restart it or relaunch the analyzer application.

#### **Unable to lock to External Frequency Reference.**

The instrument is no longer locked to the external reference signal.

- Verify that the external reference signal is still a valid signal.
- Select External on the Frequency Reference tab in the Config In/Out control panel.

#### **VBW not applied - Acq BW too small**

Increase VBW or measurement bandwidth. Make sure Sampling Parameters are set to Auto.

#### **VBW not applied - Spectrum Length too short**

This message occurs when the requested VBW can't be produced because the Spectrum Length is too short.

To clear this problem, set the Spectrum Length to Auto or manually increase the Spectrum Length (see Setup > Analysis > Spectrum Time tab). If Spectrum Length is coupled to Analysis Length, set Analysis Length to Auto or manually increase it (see Setup > Analysis > Analysis Time tab).


## Displaying the Windows Event Viewer

When the analyzer generates an error message, information about the error is logged to the Windows Event Viewer.

To display the Windows Event Viewer:


1. Select Start > Control Panel.
2. Select Administrative Tools. (If your instrument displays control panels in Category View, select System and Security, and then select Administrative Tools.)

3. Double-click **Event Viewer**.
4. From the Event Viewer window, select **Windows Logs**, then select **Application**. This displays a list of all errors that have been reported to the operating system from applications.



Errors reported to the Event Viewer from the analyzer application appear under Source as TekRSA.

5. Double-click the last error reported for TekRSA to see details on the most recently reported error. Please note that many items reported as “errors” are simply informational and do not mean that your instrument is impaired. Contact the Tektronix Customer Support Center or Service Center if you are concerned about an error shown in the Event Viewer. Do not send an instrument out for repair based solely on these event reports.



## Dealing with Sluggish Instrument Operation

Over time you might notice that acquisitions seem to be occurring sluggishly. This can occur if the disk is too full or if someone has accidentally deleted the instrument cache file. While the instrument automatically recreates a missing 4 GB cache file, if the disk is too full when the replacement cache file is created, the cache file can be fragmented and result in sluggish performance.

If your instrument appears to be performing acquisitions slowly, do the following:

- Delete unnecessary files to create free space. After deleting unnecessary files, run a commercial disk defragmentation program.
- If removing unnecessary files and defragging the hard drive does not improve instrument performance, reinstall the operating system and application software. To do this, back up all your data and files, and then reinstall the instrument operating system. After reinstalling the operating system, reinstall the application software.

## On/Standby Switch

Use this button to power the instrument on or to shut down (set the instrument to standby).

Before shutting down, you should first exit the instrument application and any other active Windows applications to avoid error messages on shutdown or the next power up.

## How to Find Out if Software Upgrades are Available

Software upgrades might be available can be downloaded from the Tektronix Web site. To see if a software upgrade is available for your product, use your browser to go to [www.tektronix.com/software](http://www.tektronix.com/software). Search by the product model number.


To add additional software options or features, you will need an option key from Tektronix. In some instances, you may need to upgrade the product software before adding the new option key. Follow the installation instructions provided with your software option.

## Settings


### Menu Bar: Setup > Settings

### Front Panel / Application Toolbar: Settings

The Settings menu item enables access to control panels that allow you to change settings for the selected display. The control panel that appears when you select **Settings** depends on the selected display. The Settings control panel for the Spectrum display is shown in the following figure.



The control panel for the CCDF display is shown in the following figure.



### Settings Control Panels

[Acquire \(see page 578\)](#)

[AM Settings \(see page 130\)](#)

[Amplitude Settings \(see page 569\)](#)

[Amplitude vs Time Settings \(see page 109\)](#)

[Analysis Settings \(see page 555\)](#)

[Audio Spectrum Settings \(see page 470\)](#)

[Audio Summary Settings \(see page 472\)](#)

[CCDF Settings \(see page 186\)](#)

[Channel Power and ACPR Settings \(see page 156\)](#)

[OFDM Channel Response Settings \(\[OFDM \\(see page 271\\)\]\(#\)\) \(\[WLAN \\(see page 163\\)\]\(#\)\)](#)

[Configure In/Out Settings \(see page 616\)](#)

[Constellation Settings \(see page 485\)](#)

[DPX Spectrum Settings \(see page 76\)](#)

[EVM vs Time Settings \(see page 489\)](#)

[Eye Diagram Settings \(see page 492\)](#)

[FM Settings \(see page 137\)](#)

[Frequency Deviation vs Time Settings \(see page 494\)](#)

[Frequency vs Time Settings \(see page 111\)](#)

[Magnitude Error Settings \(see page 496\)](#)

[MCPR Settings \(see page 163\)](#)

[Occupied Bandwidth Settings \(see page 173\)](#)

[OFDM Constellation Settings \(see page 273\)](#)

[OFDM EVM Settings \(see page 274\)](#)

[OFDM Magnitude Error Settings \(see page 276\)](#)

[OFDM Phase Error Settings \(see page 280\)](#)

[OFDM Power Settings \(see page 282\)](#)

[OFDM Summary Settings \(see page 284\)](#)

[OFDM Symbol Table Settings \(see page 286\)](#)

[Phase Error vs Time Settings \(see page 498\)](#)

[Phase Noise Settings \(see page 192\)](#)

[Phase vs Time Settings \(see page 113\)](#)

[PM Settings \(see page 145\)](#)

[Pulse Statistics Settings \(see page 325\)](#)

[Pulse Table Settings \(see page 320\)](#)


[Pulse Trace Settings \(see page 323\)](#)  
[RF I Q vs Time Settings \(see page 115\)](#)  
[Mask Test Limits Settings \(see page 547\)](#)  
[Spectral Emissions Mask Settings \(see page 218\)](#)  
[Settling Time Settings \(see page 206\)](#)  
[Signal Quality Settings \(see page 503\)](#)  
[Spectrogram Settings \(see page 104\)](#)  
[Spectrum Settings \(see page 100\)](#)  
[Spurious Settings \(see page 178\)](#)  
[Symbol Table Settings \(see page 505\)](#)  
[Time Overview Settings \(see page 93\)](#)  
[Trellis Diagram Settings \(see page 507\)](#)

## Remote Login

You can access the instrument across a network using Windows Remote Desktop Connection.

To access the instrument using Windows Remote Desktop Connection for the first time:

1. On the PC from which you want to remotely access the instrument, select **Start > All Programs > Accessories > Remote Desktop Connection**.
2. In the Remote Desktop Connection window, click the **Options** button to display all the available settings if they are not already visible.



3. Enter the name of the instrument in the **Computer** field. (To find out the name of the instrument, go to the Windows desktop, right-click the Computer icon, and select **Properties**. The computer name will appear on the control panel window.)
4. Enter **OEM** in the **User name** field.
5. Click **Connect** to initiate the connection.
6. Enter **Gladius** as the password when prompted.



## Accuracy

The closeness of the indicated value to the true value.

## ACLR

Adjacent Channel Leakage power Ratio is the ratio of the RRC (Root Raised Cosine) filtered mean power centered on the assigned channel frequency to the RRC filtered mean power centered on an adjacent frequency (defined in 3GPP).

## ACPR Measurement

Adjacent Channel Power Ratio (ACPR) is the ratio of the mean power centered on the assigned channel frequency to the mean power centered on an adjacent channel frequency. In the 3GPP specification, it is called ACLR (Adjacent Channel Level Ratio) and both the main channel and adjacent channels are required to be filtered with RRC (Root Raised Cosine) filters.

## Acquisition

A series of time-contiguous frames. This is also called a Block.

## Analysis Length

The length of time in the Analysis Time.

## Analysis Time

The portion of the acquisition record over which one or more measurements are calculated.

## ASK

Acronym for Amplitude Shift Keying. The process, or result of a process, in which the amplitude of the carrier is varied in accordance with the state of a digital input signal.

## Block

An integer number of time-contiguous frames. See also: Acquisition.

## Calibrator

A signal generator producing a specified output used for calibration purposes.

## Carrier

The RF signal upon which modulation resides.

## Carrier Frequency

The frequency of the CW component of the carrier signal.

## Carrier Signal

The electrical signal, typically a sine wave, upon which modulation is impressed.

## Carrier-to-Noise Ratio (C/N)

The ratio of carrier signal power to average noise power in a given bandwidth surrounding the carrier; usually expressed in decibels.

## CCDF - Complimentary Cumulative Distribution Function

The Complementary Cumulative Distribution Function (CCDF) represents the probability that the peak power above average power of a measured signal exceeds a threshold.

CCDF is a plot of the percent of time that a signal's power value exceeds its average value versus the amount by which it exceeds the average. The CCDF plot has a log of probability on the Y-axis (100% at the top) and dB above average amplitude on the X-axis (0 at the left).

## CDMA

Acronym for Code Division Multiple Access.

## Center Frequency

The frequency corresponding to the center of a frequency span of the analyzer display.

## Check Mark Indicator

The check mark indicator in the upper-left corner of the display indicates the display for which the acquisition hardware is optimized. When you have more than one display open, the display with the check mark indicator has control over the acquisition hardware. To give a display priority over any others, click its title bar.

When *Best for multiple windows* is selected in the Amplitude control panel's RF & IF Optimization control, none of the measurement displays shows a checkmark, as there is not a single optimized measurement.

## CISPR

International special committee on radio interference. (Comité international spécial des perturbations radioélectriques)

## CW

Acronym for Continuous Wave.

## CW Signal

Continuous wave signal - a sine wave.

## DANL

Acronym for Displayed Average Noise Level. See [Sensitivity \(see page 664\)](#).

## dBfs

A unit to express power level in decibels referenced to full scale. Depending on the context, this is either the full scale of the display screen or the full scale of the analog-to-digital converter (ADC).

## dBm

A unit of expressed power level in decibels referenced to 1 milliwatt.

## dBmV

A unit to express voltage levels in decibels referenced to 1 millivolt.

## dBuV

A unit to express voltage levels in decibels referenced to 1 microvolt.

## Decibel

Ten times the logarithm of the ratio of one electrical power to another.

## Detection

The process by which a long waveform is decimated (reduced) down to the desired number of trace points, by dividing the waveform into intervals and choosing a single value to represent each interval in the trace.

## Display Reference Level

A designated vertical position representing a specified input level. The level may be expressed in dBm, volts, or any other units.

## Distortion

Degradation of a signal, often a result of nonlinear operations, resulting in unwanted signal components. Harmonic and intermodulation distortion are common types.

## Dynamic Range

The maximum ratio of the levels of two signals simultaneously present at the input which can be measured to a specified accuracy.

## EVM

Acronym for Error Vector Magnitude.

## Export

Save data to a file in a format other than application-native.

## FastFrame

FastFrame segments the acquisition record into a series of frames and then captures acquisitions as single frames. You can then view and measure each frame individually.

## Field Strength

Field Strength measures the absolute electromagnetic power at the antenna for a specified frequency and bandwidth.

## FFT

Fast Fourier Transform - a mathematical process to calculate the frequency spectrum of a discrete number of time domain sample points.

## Filter

A circuit that separates electrical signals or signal components based on their frequencies.

## FM

Acronym for Frequency Modulation.

## Frame

A series of time-contiguous samples, long enough in duration and at a sufficient sample rate to produce a spectrum view of a specified span and RBW.

## Frequency

A series of time-contiguous samples, long enough in duration and at a sufficient sample rate to produce a spectrum view of a specified span and RBW.

## Frequency Band

The continuous range of frequencies extending between two limiting frequencies, expressed in hertz.

## Frequency Domain View

The representation of the power of the spectral components of a signal as a function frequency; the spectrum of the signal.

## Frequency Drift

Gradual shift or change in displayed frequency over the specified time due to internal changes in the analyzer, where other conditions remain constant. Expressed in hertz per second.

## Frequency Mask Trigger

A flexible real-time trigger based on specific events that occur in the frequency domain.

## Frequency Range

The range of frequencies over which the performance of the instrument is specified.

## Frequency Span

A continuous range of frequencies extending between two frequency limits.

## Frequency Settling Time

The time measured from a reference point to when the signal of interest settles to within user-defined tolerance.

## GPB

Acronym for General Purpose Interface Bus, the common name for the communications interface system defined in IEEE Std. 488.

## Graticule

The calibrated grid overlaying the display screen of analyzers, oscilloscopes, and other test instruments.

## Grayed Out

An on-screen control is “Grayed Out” if it is not adjustable.

## I/Q

Acronym for In-phase / Quadrature.

## IF

Acronym for Intermediate Frequency.

## Import

Bring data into the application from a file of some format other than application-native.

## Impulse Response

The Impulse Response trace display shows normalized power on the vertical axis and time on the horizontal axis.

## Input Impedance

The impedance at the desired input terminal. Usually expressed in terms of VSWR, return loss, or other related terms for low impedance devices and resistance-capacitance parameters for high impedance devices.

## LISN

Acronym for Line Impedance Stabilization Network.

## Local Oscillator (LO)

An oscillator which produces the internal signal that is mixed with an incoming signal to produce the IF signal.

## Marker

A visually identifiable point on a waveform trace, used to extract a readout of domain and range values represented by that point.

## Max Hold

Digitally stored display mode which, at each frequency address, compares the incoming signal level to the stored level and retains the greater level. In this mode, the display indicates the peak level at each frequency after several successive acquisitions.

## MCPR (Multiple Carrier Power Ratio)

The ratio of the signal power in the reference channel or group of channel to the power in adjacent channels.

## Min Hold

Digitally stored display mode which, at each frequency address, compares the incoming signal level to the stored level and retains the lower level. In this mode, the display indicates the minimum level at each frequency after several successive sweeps.

## Modulate

To regulate or vary a characteristic of a signal, typically in order to transmit information.

## Modulating Signal

The signal which modulates a carrier. The signal which varies or regulates some characteristic of another signal.

## Modulation

The process of varying some characteristic of a signal with a second signal.

## Noise

Unwanted random disturbances superimposed on a signal which tend to obscure it.

## Noise Bandwidth (NBW)

The exact bandwidth of a filter that is used to calculate the absolute power in dBm/Hz.

## Noise Floor

The noise intrinsic to a system that represents the minimum limit at which input signals can be observed; ultimately limited by thermal noise (kTB). The analyzer noise floor appears as a “grassy” baseline in the display, even when no signal is present.

## Open (Recall)

Bring data into the application from a file of application-native format.

## OQPSK

Acronym for Offset QPSK (Quadrature Phase Shift Keying).

## Phase Settling Time

The time measured from a reference point to when the signal of interest settles to within user-defined tolerance.

## PM

Acronym for Phase Modulation.

## Primary Marker

The marker displayed in the Single Marker mode whose frequency and/or position is changed when tuning with the general purpose knob.

## PSK

Acronym for Phase Shift Keying. The process, or result of a process, in which the carrier phase is varied discretely in accordance with a digital code.

## QAM

Acronym for Quadrature Amplitude Modulation. The process, or result of a process, in which the amplitude and phase of the carrier are varied concurrently by synthesizing two orthogonal ASK waves (see ASK).

## Real-Time Analysis

Measurement technique based on triggering on an RF signal, seamlessly capturing it into memory, and analyzing it in the frequency, time, and modulation domains.

## Real-Time Bandwidth

The frequency span over which real-time seamless capture can be performed, which is a function of the digitizer and the IF bandwidth of a Real-Time Signal Analyzer.

## Real-Time Seamless Capture

The ability to acquire and store an uninterrupted series of time domain samples that represent the behavior of an RF signal over a long period of time.

## Reference Level

The signal level represented by the uppermost graticule line of the analyzer display.

## Residual FM (Incidental FM)

Short term displayed frequency instability or jitter due to instability in the analyzer local oscillators. Given in terms of peak-to-peak frequency deviation and expressed in hertz or percent of the displayed frequency.

## Residual Response


A spurious response in the absence of an input signal. (Noise and zero pip are excluded.)

## RBW

The RBW determines how well the analyzer can resolve or separate two or more closely spaced signal components.

## Ripple

The Ripple measurement result is displayed in either Watts or Volts. The amplitude units selected on the Setup > Analysis > Units tab determine whether the measurement is presented in Watts or Volts. Volts are shown for linear units (for example, volts or amps); Watts are shown for non-linear units (for example, watts or dBm).



The Ripple measurement, in Watts, is calculated as follows:

$$\%Ripple (Watts) = 100 \times (RatioPos + RatioNeg)$$

Where:

$$RatioNeg = \left( \left( \left( \frac{DelNeg}{RefNeg} \right) + 1 \right)^2 \right) - 1$$

$$RatioPos = \left( \left( \left( \frac{DelPos}{RefPos} \right) + 1 \right)^2 \right) - 1$$

- DelPos = Delta Positive in Volts
- RefPos = Reference Positive in Volts
- DelNeg = Delta Negative in Volts (this is a positive value)
- RefNeg = Reference Negative in Volts

The Ripple measurement, in Volts, is calculated as follows:

$$\%Ripple (Volts) = 100 \times (RatioPosV + RatioNegV)$$

Where:

- RatioPosV = DelPos/RefPos
- RatioNegV = DelNeg/RefNeg

## RTSA

Acronym for Real-Time Signal Analyzer or Real-Time Spectrum Analyzer.

## Secondary Marker

The “second” marker displayed only in the Delta Marker mode.

## Sensitivity

Measure of a analyzer’s ability to display minimum level signals, usually expressed as displayed average noise level ([DANL \(see page 654\)](#)).

## Shape Factor (Skirt Selectivity)

The ratio of the frequency separation of the two (60 dB/3 dB) down points on the response curve to the static resolution bandwidth.

## Signal

As used in this help, the signal refers to the input signal before it is processed. The signal is an input.

## Signal Strength

Signal Strength locates the direction of a signal using a directional antenna and an audio tone and a visual display.

## Span

Span is the range of frequencies displayed in a spectrum window. Span, start frequency and stop frequency are related by the following equation:  $\text{Span} = (\text{stop frequency}) - (\text{start frequency})$ . The settings for center, start and stop frequencies are related to the setting for span; when one parameter is changed, the others are changed automatically.

## Span Per Division (Span/Div)

Frequency difference represented by each major horizontal division of the graticule.

## Spectrogram

Frequency vs. time vs. amplitude display where the frequency is represented on the x-axis and time on the y-axis. The power level is indicated by variations in color.

## Spectrum

The frequency domain representation of a signal showing the power distribution of its spectral component versus frequency.

## Spectrum Analysis

The technique or process of determining the frequency content of an RF signal.

## Spectrum Analyzer

A device for determining the frequency components of a signal.

## Spectrum Time

Analysis Time for spectrum analysis views. Spectrum time can be the same as Analysis Time, but it can be different.

## Spur

A spur is a signal peak that exceeds a user-definable threshold (See Spurious > Settings > Ranges) and excursion setting. A spur that also exceeds a limit (either Absolute or Relative) specified on the Settings > Limits tab is considered a violation.

## Spurious Response

A response to a analyzer wherein the displayed frequency is not related to the input frequency.

## Time Measurement

This is the time in seconds relative to the time reference point in the first acquisition record in the data set.

## Time Reference

The point in time during the acquisition record used as the zero point for counting time. The time reference can be set to either the start of the acquisition record or the trigger point.

## Trace

As used in this help, trace refers to the displayed signal. The displayed signal can be a processed version of the input signal (for example, it may be averaged.) The trace is a result or output.

## Vector Signal Analyzer

Like a spectrum analyzer, a device for determining the frequency components of a signal. However, unlike a standard spectrum analyzer, the vector signal analyzer is optimized for analyzing digitally modulated RF signals.

## Vertical Scale Factor, Vertical Display Factor

The number of dB, volts, etc., represented by one vertical division of a spectrum analyzer display screen.

## Violation

A violation is a spur that exceeds either an Absolute or Relative limit (depending on the selected mask) specified on the Settings > Limits tab. A spur is a signal peak that exceeds a user-definable threshold (See Spurious > Settings > Ranges) and excursion setting.



# Index

## Symbols and Numbers

- 6dB RBW (MIL) Filter shape
  - Spectrum display BW tab, 122
- 20 dB bandwidth
  - Bluetooth, 42
- 20dB Bandwidth display
  - Bluetooth, 449
- 28 VDC output, 621
- 3-D Spectrogram display, 101
- 3-D Waterfall
  - Spectrogram/DPXogram, 106
  - Spectrogram/DPXogram display, 86

## A

- Abort acquisition, 577
- About the RTSA, 615
- Abs
  - Ranges and Limits table, 183
- Abs & Rel
  - Ranges and Limits table, 183
- Abs OR Rel
  - Ranges and Limits table, 183
- Abs Same
  - Ranges and Limits table, 183
- Abs Start (dBm)
  - Ranges and Limits table, 183
- Abs Stop (dBm)
  - Ranges and Limits table, 183
- Accessories, 3
- accuracy, 651
- ACLR, 651
  - LTE, 42
- ACLR measurement, 156
- ACPR measurement, 651
- Acq Data
  - Replay menu, 564
- Acq Data/DPX Spectra, 561
- Acquire control panel, 578
- acquisition, 651

- acquisition data
  - saving, 607
- Acquisition data files, 30
- Acquisition Info tab, 563
- acquisition memory usage, 579
- Actions tab
  - Search, 553
  - Triggering, 605
- Actual BW
  - Freq & BW tab, 78
- adaptive equalizer, 515
- Adjacent Channel Power
  - measurement, 156
- Adjacent Channel Power Ratio, 35
- Advanced Params tab
  - OFDM Analysis displays, 288
- Advanced Params Tab
  - GP Digital Modulation displays, 517
- Advanced tab
  - Triggering, 604
- Advanced tab (Acquire), 580
- Alignments, 632
  - Align Now, 21
  - Auto button, 21
  - Off button, 21
- +AM, 40
  - AM display, 40
- AM
  - AM display, 40
- AM display, 129
- AM measurements, 40
- Ampl Scale tab
  - DPXOgram display, 86
- Amplitude control panel, 569
- Amplitude vs. Time display, 108
- Amplitude vs. Time display
  - Settings, 109
- Analog IF output, 620
- Analysis control panel, 555
- analysis length, 651

- Analysis Params tab
  - Bluetooth, 457
  - LTE, 415
  - P25 Analysis displays, 371
  - WLAN Analysis displays, 260
- analysis time, 651
- Analysis Time tab
  - Bluetooth, 458
  - GP Digital Modulation displays, 521
  - LTE, 416
  - OFDM Analysis displays, 289
  - Options control panel, 29
  - P25 Analysis displays, 371
  - WLAN Analysis displays, 263
- Analysis Time tab (Analysis control panel), 555
- APCO P25 measurements, 41
- application presets
  - saving, 607
- Application presets, 22
- Apply ampl corrections, 329
- APSK parameters, 519
- Arm trigger status light, 587
- arranging windows, 634
- ASK, 652
- Attenuator settings tab, 569
- Audio BW, 474
  - setting, 474
- Audio Demod tab, 89
- audio filters
  - custom, 478
- Audio Filters tab
  - Audio Analysis displays, 476
- Audio Freq
  - Audio Summary display, 472
- Audio Spectrum display, 469
- Audio Spectrum settings, 470
- Audio Summary settings, 472
- Auto
  - Ranges and Limits table, 183

- Auto button (Alignments dialog), 21
- automatically generate filenames, 29
- Automatically increment filename/number, 30
- Autosave acquisition on Trigger, 604
- Average Channel Power measurement, 156
- Average ON Power, 43
- Average Transmitted Power, 43
- Averaging
  - Audio Analysis displays, 476
  - Params2 tab, 476

## B

- Base file name
  - FastSave tab, 584
- Basic Controls Toolbar, 628
- black text
  - settling time measurements, 201
- Blackman-Harris 4B Filter Shape
  - Spectrum display BW tab, 122
- Blizzard color scheme, 30
- block, 652
- Blue background (cells), 174
- Bluetooth
  - Analysis Params tab, 457
  - Analysis Time tab, 458
  - BT 20dB Bandwidth display, 449
  - BT InBand Emission, 452
  - Freq & RBW tab, 466
  - Frequency Dev vs Time (BT) display, 445
  - Limits tab, 459
  - Measurement Params tab, 467
  - Parameters tab, 464
  - Power class, 456
  - Prefs tab, 463
  - Scale tab, 460
  - Trace tab, 461
- Bluetooth measurements, 42

- BT 20dB BW Settings, 451
- BT CF Offset and Drift
  - display, 434
  - settings, 401
- BT Constellation display, 429
- BT settings, 431
- BT Eye Diagram display, 431
- BT settings, 433
- BT limits, 437
- BT Summary display, 436
- BT settings, 441
- BT Symbol Table display, 441
- BT settings, 444
- Burst detect threshold, 139
- BW (Hz)
  - Ranges and Limits table, 183
- BW Tab
  - Amplitude vs. Time display, 110
  - DPXOgram display, 122
  - Spectrum display, 122

## C

- C4FM modulation type
  - measurements, 40
- C4FM signal quality
  - measurements, 503
- calibration, 22
  - See also* Alignments
- calibrator, 652
- Carr Freq Err
  - Audio Summary display, 472
- carrier, 652
- carrier frequency, 652
- Carrier frequency detection, 139
- Carrier Frequency Error, 40
- Carrier frequency offset and drift
  - Bluetooth, 42
- Carrier Power
  - Audio Summary display, 472
- carrier signal, 652
- carrier-to-noise ratio (C/N), 652
- CCDF, 35
- CCDF - Complimentary Cumulative Distribution Function, 653
- CCDF display, 185
- CDMA, 653
- Cell ID detection
  - LTE, 42
- center frequency, 653
- Center Frequency
  - Freq & BW tab, 78
- changing the input connector, 617
- Channel Power, 35
  - LTE, 43
- Channel Power and ACPR (Adjacent Channel Power Ratio) display, 153
- Channel Power and ACPR Settings control panel, 156
- Channel Power measurement, 156
- Channel Response
  - OFDM, 36
  - WLAN, 37
- Channels tab
  - Channel Power & ACPR display, 157
  - LTE Analysis displays, 402
- Channels tab (MCPR), 168
- check mark indicator, 653
- Check mark indicator, 15
- CISPR, 653
- CISPR Filter shape
  - Spectrum display BW tab, 122
- Classic color scheme, 30
- Color (DPXogram)
  - DPXogram display, 86
- Configure In/Out control panel, 616
- configuring Presets, 25
- connecting signals, 616
- connecting trigger signals, 619
- Constellation
  - OFDM, 36
  - WLAN, 37
- Constellation display, 484
  - displaying a second trace, 524

- Constellation display
  - settings, 485
- Continuous Acquisition
  - Mode, 630
- Convergence
  - Equalizer tab, 516
- Correct for Noise Floor
  - Channel Power & ACPR
    - display, MCPR
    - display, 166
- correction tables, 572
- creating a mask, 590
- creating presets, 26
- Creating presets, 22
- CW, 653
- CW signal, 654
- D**
- DANL, 654
- Data file formats, 609
- Data length
  - File Contents window, 608
- Data properties
  - File menu, 626
- Data Range tab
  - OFDM Analysis
    - displays, 289
  - WLAN Analysis
    - displays, 262
- Data records
  - File Contents window, 608
- Data types
  - File Contents window, 608
- dBfs, 654
- dBm, 654
- dBmV, 654
- dBuV, 654
- De-emphasis
  - Audio Filters tab, 477
- decibel, 654
- Default
  - BT limits table, 437
  - P25 limits table, 358
- default settings, restore, 21
- Define Markers, 541
- Define Peaks tab, 542
- Define tab, 547
  - Pulsed RF displays, 328
- Settling Time displays, 207
- Delay
  - Triggering, 604
- Delta Frequency, 44
- Delta Markers, 541
- Demod I&Q vs Time display
  - settings, 488
- Density tab, 88
- detect, 228
  - AM display, 133
  - FM display, 140
  - General Signal Viewing
    - displays, 119
  - Time Overview, 96
- Detect
  - PM display, 148
- detection, 228
  - AM display, 133
  - FM display, 140
  - General Signal Viewing
    - displays, 119
  - Time Overview, 96
- Detection
  - DPX display, 82
  - PM display, 148
- detector, 228
  - AM display, 133
  - FM display, 140
  - General Signal Viewing
    - displays, 119
  - Time Overview, 96
- Detector
  - PM display, 148
  - Ranges and Limits table, 183
- Deviation, 40
- Diagnostics, 633
- Diff
  - Audio Summary display, 472
- Digital I and Q outputs, 619
- Digital Phosphor Spectrum, 67
- Disabled - FastSave is
  - selected, 583
- display reference level, 655
- Displays
  - arranging, 33
  - BT CF Offset and Drift, 434
  - BT Constellation, 429
  - BT Eye Diagram, 431
- BT Summary, 436
- BT Symbol Table, 441
- Frequency Settling Time, 198
- interactions between, 33
- LTE ACLR, 392
- LTE Channel Spectrum, 395
- LTE Constellation, 399
- OFDM Channel
  - Response, 269
- OFDM Constellation, 272
- OFDM EVM, 273
- OFDM Mag Error, 277
- OFDM Phase Error, 279
- OFDM Power, 281
- OFDM Spectral Flatness, 275
- OFDM Summary, 283
- OFDM Symbol Table, 285
- P25 Constellation, 351
- P25 Eye Diagram, 354
- P25 Power, 356
- P25 Summary, 358
- P25 Symbol Table, 361
- Phase Settling Time, 198
- selecting, 33
- SEM (Spectrum Emission
  - Mask), 214
- WLAN Channel
  - Response, 235
- WLAN Constellation, 238
- WLAN EVM, 240
- WLAN Mag Error, 242
- WLAN Phase Error, 245
- WLAN Power, 246
- WLAN Spectral Flatness, 249
- WLAN Summary, 251
- WLAN Symbol Table, 256
- distortion, 655
- Dither Control, 580
- documentation, 3
- Done
  - Ranges and Limits table, 183
- Done saving - restart with Stop,
  - then Run, 583
- Dot Persistence, 78
- DPX
  - Bitmap settings, 78
  - DPXogram Settings, 78
  - presets, 22

- DPX Density, 88
- DPX Density Trigger, 588
- DPX display
  - DPXogram, 67
  - Frequency, 67
  - Phase, 67
  - Spectrum, 67
  - Zero Span, 67
- DPX Spectra
  - Replay menu, 564
- DPX Spectra/Acq Data, 561
- DPX Spectrum settings, 76
- DPXogram
  - DPX display, 67
- Droop, 43
- Droop dB, 43
- Duty Factor (%) measurement, 43
- Duty Factor (Ratio)
  - measurement, 43
- Dwell Time, 118
- dynamic range, 655

**E**

- Elements of the display, 15
- Enable EQ Export
  - Equalizer tab, 516
- Enable Equalization
  - Equalizer tab, 516
- Equalizer tab
  - GP Digital Modulation displays, 515
- error message
  - Windows Event Viewer, 643
- Error messages, 635
- Event tab (Triggering), 594
- Event Viewer, 643
- EVM, 38
  - OFDM, 36
  - WLAN, 37
- EVM tab
  - WLAN Analysis displays, 267
- EVM vs. Time display, 488
- EVM vs. Time Settings, 489
- Excursn (dB)
  - Ranges and Limits table, 183
- Expand
  - Ranges and Limits Tab, 182

- export, 655
- Ext Gain/Loss Table tab, 311
- External Gain value, 572
- External Gain/Loss Correction tab, 572
- External Loss Tables, 572
- External Probe Correction tab, 576
- external trigger
  - enabling, 619
- external trigger signals
  - connecting, 619
- Eye Diagram display settings, 492

## F

- Fall Time, 43
- Fast Save, 581
  - See also* FastSave
- FastFrame, 655
- FastSave
  - configuring, 582
  - overview, 581
  - starting, 582
  - status messages, 583
- FastSave acquisition data
  - FastSave tab, 584
- FastSave tab, 584
- FFT, 656
- field strength, 656
- File Contents window, 608
- File formats, 609
- File Menu, 623
- File types, 609
- filter, 656
- Filter Parameter
  - GP Digital Modulation displays, 510
- Filter Shape
  - Ranges and Limits table, 183
  - Spectrum display BW tab, 122
- Find tab
  - GP Digital Modulation displays, 521
- finding bursts in data
  - GP Digital Modulation displays, 521

Flatness

- OFDM, 36
- WLAN, 37

FlatTop (CW ampl) Filter Shape

- Spectrum display BW tab, 122

FM, 656

FM display, 135

FM measurements, 40

FMT (Frequency Mask Trigger), 586

Format

- FastSave tab, 584

frame, 656

Freq & BW tab

- DPX display, 78
- RF I & Q vs Time display, 116

Freq & BW Tab

- GP Digital Modulation displays, 514

Freq & RBW tab

- Bluetooth, 466

Freq & Span tab

- General Signal Viewing displays, 118

Freq and RBW tab

- LTE, 413

Freq Deviation, 44

Freq Diff, Phase Diff setting, 330

Freq Error, 40

Freq Estimation tab

- Pulsed RF displays, 333

Freq Setup tab, 310

Freq, Span & RBW tab

- Channel Power & ACPR display, MCPR display, 165
- RF Measurements displays, 226

Freq-domain Linearity

- setting, 328

frequency, 656

Frequency

- DPX display, 67
- of local oscillator, 297

frequency band, 656

Frequency Dev vs Time (BT)  
   display  
     Bluetooth, 445  
 Frequency Dev vs Time (P25)  
   display, 364  
 Frequency Dev vs. Time (BT)  
   Settings, 449  
 Frequency Dev vs. Time (P25)  
   Settings, 366  
 Frequency deviation  
   P25, 41  
 Frequency Deviation vs. Time  
   settings, 494  
 frequency domain view, 657  
 frequency drift, 657  
 Frequency Edge triggering, 598  
 Frequency Error, 501  
   GP Digital Modulation  
     displays, 514  
 Frequency Mask Trigger  
   (FMT), 657  
 Frequency Offset  
   GP Digital Modulation  
     displays, 514  
 frequency range, 657  
 Frequency range  
   Bluetooth, 42  
 frequency settling time, 657  
   measuring, 203  
 Frequency Settling Time  
   display, 198  
 Frequency Settling Time  
   settings, 206  
 frequency span, 657  
 Frequency tab, 557  
 Frequency tab (Phase Noise), 193  
 Frequency vs. Time display, 110  
 Frequency vs. Time Settings, 111  
 Front panel connectors, 9  
 Front panel controls, 9  
 Full Screen, 628

## G

Gain and Noise Figure  
   measurements, 296  
 Gain display, 298  
 Gain display settings, 299  
 Gain Imbalance, 501

GPIO, 657  
 GPIO tab, 29  
 graticule, 658  
 Graticule On/Off  
   RF Measurements  
     displays, 231  
 grayed out, 658

## H

Hanning Filter Shape  
   Spectrum display BW  
     tab, 122  
 HCPM transmitter logical channel  
   time alignment  
   P25, 42  
 Help Menu, 634  
 Hide sensitive readouts, 29  
 Hide Sensitive readouts, 29  
 Histogram readout  
   Pulsed RF displays, 336  
 Holdoff  
   Triggering, 604  
 HPF  
   Audio Filters tab, 477  
 Hum & Noise  
   Audio Summary display, 481  
 hysteresis, 599

## I

I/Q, 658  
 IF, 658  
 IF output, 620  
 Ignore Region  
   Params2 tab, 476  
 import, 658  
 Impulse Response Amplitude, 44  
 Impulse Response setting, 329  
 Impulse Response Time, 44  
 In-band emission / ACPR  
   Bluetooth, 42  
 InBand Emission  
   Bluetooth, 452  
 Information messages, 635  
 input impedance, 658  
 input protection, 617  
 Inside  
   Time Qualified tab, 603

Installed options, 5  
 Intermediate Values tab  
   Uncertainty Calculator  
     Display, 308  
 Internal Attenuator  
   Internal Settings tab, 569  
 Internal attenuator (Internal  
   Settings tab), 569  
 Internal Preamp  
   Internal Settings tab, 570  
 Internal Settings tab, 569  
 IQ Origin Offset, 38  
 IQ output, 621  
 IQ outputs, 619  
 IQ vs. time view, 114  
 is greater than (Test), 548  
 is greater than mask, 547  
 is inside mask, 547  
 is inside mask (Test), 549  
 is less than (Test), 548  
 is less than mask, 547  
 is outside mask, 547  
 is outside mask (Test), 549

## K

Kaiser (RBW) Filter Shape  
   Spectrum display BW  
     tab, 122  
 Keep-out time, 329  
 Keypad, 627

## L

Length, 40  
   Equalizer tab, 516  
 Levels tab  
   Pulsed RF displays, 333  
 Limit lines, 547  
 Limits  
   BT default, 437  
   P25 default, 358  
   Spurious display, 184  
 Limits tab  
   Bluetooth, 459  
   P25 Analysis displays, 378  
 Link to Span, 116  
 LISN, 659

- Load
    - Ranges and Limits Tab, 182
    - Ranges and Limits table, 183
  - Local oscillator, 297
  - local oscillator (LO), 659
  - Location
    - FastSave tab, 584
  - Lock Center Frequency of
    - Spectrum displays, 557
  - Lock Center Frequency of
    - Spectrum displays to
    - Measurement Frequency
    - Frequency tab, 558
  - Log
    - Audio Spectrum display, 480
  - Log (Horizontal)
    - Spectrum display, 101
  - Longer
    - Time Qualified tab, 603
  - Loop over all selected records
    - Replay menu, 564
  - LPF
    - Audio Filters tab, 477
  - LTE
    - Analysis Params tab, 415
    - Analysis Time tab, 416
    - Channels tab, 402
    - Freq and RBW tab, 413
    - Measurement Params
      - tab, 414
    - Modulation Params tab, 415
    - Offsets and Limits Table
      - tab, 407
    - Parameters tab, 404
    - Prefs tab, 410
    - Processing tab, 406
    - Scale tab, 408
    - Trace tab, 417
    - VBW (Channels tab), 403
    - VBW (Freq & RBW tab), 413
  - LTE ACLR
    - display, 392
    - settings, 395
  - LTE Channel Spectrum
    - display, 395
    - settings, 397
  - LTE Constellation
    - display, 399
  - LTE measurements, 42
- M**
- Mag Error, 38
    - OFDM, 36
    - WLAN, 37
  - Magnitude Error vs. Time
    - display, 494
  - Magnitude Error vs. Time
    - settings, 496
  - Main
    - presets, 22
  - manuals, 5
  - marker, 659
  - Marker Noise mode, 544
  - Marker Properties tab, 541
  - Marker to Center Frequency, 541
  - Markers, 543
    - in the BT Symbol Table, 444
  - Marker Toolbar, 543
  - Using Markers, 537
  - Using Markers in Noise
    - Mode, 544
  - Markers menu, 631
  - Markers snap to peaks when
    - dragged, 31
  - Markers Toolbar, 627
  - Mask
    - Ranges and Limits table, 183
  - Mask Editor, 590
  - Mask tab
    - Settling Time displays, 210
  - Mask Test, 547
    - is greater than mask, 547
    - is inside mask, 547
    - is less than mask, 547
    - is outside mask, 547
  - mask testing
    - settling time
      - measurement, 194
  - MATLAB file format, 609
  - Max files saved per Run
    - FastSave tab, 584
  - Max Freq Error, 44
  - max hold, 659
  - Max number of pulses
    - Pulse Table display, 326
  - Max Number of Pulses
    - Params tab, 327
  - Max Phase Error, 44
  - Max total files saved per Run, 605
  - Max trace points, 112
    - Settling Time displays, 214
  - MCPR (multiple carrier power
    - ratio), 659
  - MCPR display, 159
  - MCPR settings, 163
  - Meas BW
    - Frequency Settling Time, 207
  - Meas Frequency
    - Frequency Settling Time, 207
  - Measurement BW, no filter, 116
    - Freq & BW tab, 78
  - Measurement Filter
    - GP Digital Modulation
      - displays, 510
    - Params tab, 327
  - Measurement Frequency, 557
    - Frequency tab, 558
  - Measurement Length
    - Time Params tab, 209
  - Measurement Params tab
    - Bluetooth, 467
    - Channel Power & ACPR
      - display, MCPR
      - display, 166
    - LTE, 414
  - Measurement point, 330
  - Measurement settings export, 607
  - Measurements, 35
    - APCO P25, 41
    - Bluetooth, 42
    - LTE, 42
  - Measurements tab, 321
  - measuring settling time, 203
  - Menus, 622
  - MER (RMS), 502
  - min hold, 659
  - Min Settled Duration
    - Time Params tab, 209
  - Mod Distor
    - Audio Summary display, 472
  - Mode
    - Equalizer tab, 516
  - modulate, 660

- modulating signal, 660
- modulation, 660
- Modulation characteristics
  - Bluetooth, 42
- Modulation fidelity
  - P25, 41
- modulation measurements, 483
- Modulation measurements, 38
- Modulation Params tab
  - GP Digital Modulation
    - displays, 510
  - LTE, 415
  - OFDM Analysis
    - displays, 287
  - P25 Analysis displays, 455
  - WLAN Analysis
    - displays, 259
- Modulation types supported, 510
- Monitor Attenuator use, 569
- Monitor Attenuator Use
  - Internal Settings tab, 570
- moving windows, 634
- Multi-Carrier Power Ratio, 35
- Multiple Carrier Power Ratio, 163

## N

- Navigator View, 627
- Network Connections, 19
- nFSK modulation type
  - measurements, 40
- nFSK signal quality
  - measurements, 502
- No. of Harmonics
  - Params2 tab, 476
- No. of Non-Harmonics
  - Params2 tab, 476
- noise, 660
- Noise
  - Params tab, 313
- noise bandwidth (NBW), 660
- Noise Figure and Gain
  - measurements, 296
- Noise Figure display, 300
- Noise Figure display settings, 301
- noise floor, 660
- Noise markers, 544
- Noise measurements, 41
- Noise Source tab, 311

- Noise Table display, 296
- Noise Table settings, 297
- Noise Temperature display, 302
- Noise Temperature display
  - settings, 303
- Non-Harmonic Excursion
  - Params2 tab, 476
- Non-Harmonic Threshold
  - Params2 tab, 476
- None
  - trace motion, 77
- Normal
  - trace motion, 77
- Numeric keypad, 627

## O

- OBW Level Left/Right, 170
- OBW Ref Power, 170
- Occupied bandwidth
  - LTE, 43
- Occupied Bandwidth, 173
- Occupied Bandwidth
  - settings, 173
- Occupied BW display, 170
- OFDM Channel Response
  - display, 269
  - settings, 271
- OFDM Constellation
  - display, 272
  - settings, 273
- OFDM EVM
  - display, 273
  - settings, 274
- OFDM Mag Error
  - display, 277
  - settings, 278
- OFDM Phase Error
  - display, 279
- OFDM Power
  - display, 281
  - settings, 282
- OFDM Spectral Flatness
  - display, 275
  - settings, 276
- OFDM Summary
  - display, 283
  - settings, 284
- OFDM Symbol Table
  - display, 285
  - settings, 286
- Off
  - Ranges and Limits table, 183
- Off button (Alignments
  - dialog), 21
- Offset & Limits Table tab
  - SEM display, 221
- Offset EVM, 502
- Offsets and Limits Table tab
  - LTE, 407
- On
  - Ranges and Limits table, 183
- On/Standby switch, 645
- online help, 615
- open (recall), 660
- Operating band unwanted
  - emission
    - LTE, 43
- Operating frequency accuracy
  - P25, 41
- operating system restore, 31
- Optimization
  - Channel Power & ACPR
    - display, MCPR
    - display, 166
- Options
  - installed, 5
- Options control panel, 29
- OQPSK, 661
- OQPSK and SOQPSK modulation
  - type measurements, 39
- Other Outputs tab, 621
- Out-of-band spurious emission
  - Bluetooth, 42
- Output power
  - Bluetooth, 42
- Outputs, 621
  - 28 VDC, 621
  - IF, 621
  - IQ, 621
- Outside
  - Time Qualified tab, 603
- Overlap, 104
- Overshoot, 44
- Overshoot dB, 44
- Overshoot settings, 331

Overview tab, 309

## P

P25

presets, 22

P25 Constellation

display, 351

settings, 353

P25 Eye Diagram

display, 354

settings, 355

P25 limits, 358

P25 measurements, 41

P25 Power vs Time

display, 356

settings, 357

P25 Summary

display, 358

settings, 361

P25 Symbol Table

display, 361

settings, 363

Parameters tab

AM display, 131

Bluetooth, 464

CCDF display, 187

FM display, 139

LTE, 404

Occupied BW display, 174

Phase Noise display, 194

PM display, 146

SEM display, 219

Spurious display, 179

Params tab

DPX display, 77

Noise, 313

Pulsed RF displays, 326

Params1 tab

Audio Analysis displays, 473

Params2 tab

Audio Analysis displays, 476

Pass /Fail readout

SEM display, 216

Pass/Fail Limit Testing

Spurious display, 184

Pass/Fail tests, 548

Pause

Replay menu, 564

+Peak

Audio Summary display, 472

-Peak

Audio Summary display, 472

Peak Excursion, 541

Peak FSK err, 40

Peak Mag Err, 40

Peak Power, 43

Peak Threshold, 541

Peak/Avg Ratio, 35

Peak-Peak/2

Audio Summary display, 472

Phase

DPX display, 67

Phase Deviation, 44

Phase Error, 502

OFDM, 36

WLAN, 37

Phase Error vs. Time display, 496

Phase Error vs. Time settings, 498

Phase Noise display, 187

Phase Noise settings, 192

phase settling time, 661

measuring, 203

Phase Settling Time display, 198

Phase Settling Time settings, 202

Phase vs. Time display, 112

Phase vs. Time settings, 113

pictures

saving, 607

+Pk

FM display, 40

PM display, 41

-Pk

FM display, 40

PM display, 41

Pk-Pk

FM display, 40

PM display, 41

Pk-Pk/2

FM display, 40

Planar Crown input

connector, 617

PM, 661

PM measurements, 41

Position

Triggering, 604

Power

OFDM, 36

Power class

Bluetooth, 456

Power density

Bluetooth, 42

Power Markers, 541

power measurements, 153

Power measurements, 35

Power Threshold to Detect Pulses

Params tab, 327

Power vs Time

WLAN, 37

Pre-defined Filters

Audio Filters tab, 477

Prefs tab, 317

AM display, 135

Bluetooth, 463

FM display, 143

General Signal Viewing

Displays, 125

GP Digital Modulation

displays, 526

LTE, 410

OFDM Analysis

displays, 293

Options control panel, 31

P25 Analysis displays, 377

PM display, 151

Pulsed RF displays, 336

RF Measurements

displays, 231

Settling Time displays, 214

WLAN Analysis

displays, 268

presets, 26

Presets, 607

Application, 22

configuring User, 25

DPX, 22

Main, 22

P25, 25

recalling, 25

Standards, 25

User, 22

WLAN, 22

Presets tab

Options control panel, 29

primary marker, 661  
 Print, 626  
 Print Preview, 626  
 Probes, 618  
 Processing tab  
   LTE, 406  
   SEM display, 220  
 PSK, 661  
 Pulse measurements, 319  
 Pulse Statistics display, 323  
 Pulse Statistics display  
   settings, 325  
 Pulse Table display, 319  
 Pulse Table settings, 320  
 Pulse Trace display, 321  
 Pulse Trace display settings, 323  
 Pulse Width, 43  
 Pulse-Pulse Freq Difference, 44  
 Pulse-Pulse Phase Difference, 44  
 pulses  
   how many can be  
     analyzed, 327

## Q

QAM, 661  
 Quadrature Error, 38

## R

Range file, 609  
 Ranges and Limits Tab  
   Spurious display, 181  
 RBW, 662  
   Channel Power & ACPR  
     display, MCPR  
     display, 165  
 RBW (Time-domain BW) filter  
   Freq & BW tab, 78  
 Ready trigger status light, 587  
 real-time analysis, 661  
 real-time bandwidth, 662  
 real-time seamless capture, 662  
 rear panel connectors, 19  
 recalling Presets, 25  
 Red background (cells), 174  
 red text  
   settling time  
     measurements, 201

Ref  
   Audio Summary display, 472  
 Ref Channel tab  
   SEM display, 221  
 Ref level (Internal Settings  
   tab), 569  
 Reference Filter  
   GP Digital Modulation  
     displays, 510  
 reference level, 662  
 Reference tab  
   Spurious display, 179  
 Rel  
   Ranges and Limits table, 183  
 Rel Same  
   Ranges and Limits table, 183  
 Rel Start (dB)  
   Ranges and Limits table, 183  
 Rel Stop (dB)  
   Ranges and Limits table, 183  
 Relative power  
   Bluetooth, 42  
 Repeat, 577  
 Repetition Interval, 43  
 Repetition Rate, 43  
 Replay all selected records  
   Replay menu, 564  
 Replay control panel, 561  
 Replay current record  
   Replay menu, 564  
 Replay from selected  
   Replay menu, 564  
 Replay menu, 564  
 Replay Speed tab, 563  
 Replay toolbar, 566  
   Replay menu, 564  
 Reset Equalization  
   Equalizer tab, 516  
 Reset layout  
   Ranges and Limits Tab, 182  
 residual FM (incidental FM), 662  
 residual response, 662  
 results  
   saving, 607  
 Resume, 577  
 RF & IF Optimization  
   Internal Settings tab, 570

RF Channel Power  
   measurement, 155  
 RF I & Q vs. Time display, 114  
 RF I & Q vs. Time settings, 115  
 RF input connector, 617  
 RF input protection, 617  
 RF output power  
   P25, 41  
 Rho, 501  
 Ripple, 43  
 Ripple dB, 43  
 ripple measurement  
   calculation, 663  
 Ripple settings, 331  
 Rise Time, 43  
 RMS  
   Audio Summary display, 472  
   Detection, 228  
   FM display, 40  
   PM display, 41  
 RMS Error Magnitude, 40  
 RMS FSK Err, 40  
 RMS Mag Err, 40  
 RMS Phase Error, 44  
 Roll  
   trace motion, 77  
 RTPA2A, 618  
 RTSA, 664  
 Run, 577  
 Run menu, 629

## S

S/N  
   Audio Summary display, 472  
 Sampling Parameters Tab, 579  
 Save  
   Ranges and Limits Tab, 182  
 Save acq data on Trigger, 605  
 Save and Export tab, 29  
   Options control panel, 30  
 Save As  
   Ranges and Limits table, 183  
 Save on trigger, 605  
 Save picture on Trigger, 605  
 saved file types, 607  
 saving  
   acquisition data, 607  
   pictures, 607

- presets, 26
- results, 607
- setups, 607
- Saving
  - presets, 22
- Saving and Recalling Data, 607
- Saving X of XX, 583
- Scale tab, 316
  - AM display, 134
  - Audio Analysis displays, 479
  - Bluetooth, 460
  - DPX Spectrum display, 84
  - FM display, 142
  - General Signal Viewing displays, 124
  - GP Digital Modulation displays, 525
  - LTE, 408
  - OFDM Analysis displays, 291
  - P25 Analysis displays, 374
  - PM display, 150
  - Pulsed RF displays, 335
  - RF Measurements displays, 230
  - SEM display, 224
  - Settling Time displays, 213
  - Spectrogram display, 106
  - Spectrum display, 101
  - WLAN Analysis displays, 265
- Search control panel, 553
- Search In, 548
- secondary marker, 664
- Security tab, 29
  - Options control panel, 30
- Select all
  - Replay menu, 564
- Select data records tab, 561
- Select DPX Spectra tab, 562
- Select records from history
  - Replay menu, 564
- SEM display, 214
- sensitivity, 664
- Set to max BW, 116
- Settings, 616
  - AM display, 130
  - Audio Spectrum display, 470
  - Audio Summary display, 472
  - BT 20dB BW, 451
  - BT CF Offset and Drift, 401
  - BT Constellation, 431
  - BT Eye Diagram, 433
  - BT Frequency Dev vs. Time, 449
  - BT Summary, 441
  - BT Symbol Table, 444
  - CCDF display, 186
  - Constellation display, 485
  - Demod I&Q vs Time display, 488
  - DPX Spectrum, 76
  - EVM vs. Time display, 489
  - Eye Diagram display, 492
  - FM display, 137
  - Frequency Deviation vs. Time, 494
  - Frequency Settling Time, 206
  - Frequency vs. Time, 111
  - Gain display, 299
  - LTE ACLR, 395
  - LTE Channel Spectrum, 397
  - Magnitude Error vs. Time, 496
  - MCPR, 163
  - Noise Figure display, 301
  - Noise Table display, 297
  - Noise Temperature display, 303
  - Occupied Bandwidth, 173
  - OFDM Channel Response, 269
  - OFDM Constellation, 273
  - OFDM EVM, 274
  - OFDM Flatness, 276
  - OFDM Mag Error, 278
  - OFDM Phase Error, 280
  - OFDM Power, 282
  - OFDM Summary, 284
  - OFDM Symbol Table, 286
  - Options control panel, 29
  - P25 Constellation, 353
  - P25 Eye Diagram, 355
  - P25 Frequency Dev vs. Time, 366
  - P25 Power vs Time, 357
  - P25 Summary, 361
  - P25 Symbol Table, 363
  - Phase Error vs. Time, 498
  - Phase Noise, 192
  - Phase Settling Time, 202
  - Phase vs. Time, 113
  - Pulse Statistics display, 325
  - Pulse Table display, 320
  - Pulse Trace display, 323
  - restoring defaults, 21
  - RF I & Q vs. Time, 115
  - Signal Quality, 503
  - Spectrogram display, 104
  - Spectrum display, 100
  - Spectrum Emission Mask, 218
  - Spurious display, 178
  - Symbol Table display, 505
  - Time Overview display, 93
  - Trellis Diagram display, 507
  - Trigger, 593
  - Uncertainty Calculator display, 307
  - WLAN Channel Response, 235
  - WLAN Constellation, 239
  - WLAN EVM, 241
  - WLAN Mag Error, 243
  - WLAN Phase Error, 245
  - WLAN Power vs Time, 248
  - WLAN Spectral Flatness, 250
  - WLAN Summary, 255
  - WLAN Symbol Table, 258
  - Y Factor display, 304
- Settings tab, 314
- Settled Frequency
  - definition, 194
- Settled Phase
  - definition, 194
- settling time
  - measuring, 203
- Settling Time, 194
  - definition, 194
- Settling Time decimal places
  - Settling Time displays, 214
- settling time measurement
  - text color, 201
- Setup Menu, 631

- setups
    - saving, 607
  - shape factor (skirt selectivity), 664
  - Shorter
    - Time Qualified tab, 603
  - Show graticule
    - General Signal Viewing Displays, 125
    - Pulsed RF displays, 336
    - Settling Time displays, 214
  - Show Marker Readout in Graph
    - General Signal Viewing Displays, 125
    - Pulsed RF displays, 336
    - RF Measurements displays, 231
    - Settling Time displays, 214
  - Show Parameter Readouts, 126
  - Show recalled trace, 119
    - AM display, 132
    - FM display, 140
    - PM display, 147
    - Time Overview display, 96
  - signal, 664
  - Signal Power
    - Audio Summary display, 472
  - Signal Quality display, 498
  - Signal quality measurements, 501
  - Signal Quality settings, 503
  - signal strength, 664
  - SINAD
    - Audio Summary display, 472
  - Single, 630
  - sluggish operation, 645
  - span, 665
  - span per division, 665
  - Span/RBW Ratio
    - Spectrum display BW tab, 122
  - spectrogram, 665
  - Spectrogram display, 101
  - Spectrogram display settings, 104
  - spectrum, 665
  - Spectrum
    - DPX display, 67
  - spectrum analysis, 665
  - spectrum analyzer, 665
  - Spectrum Center Frequency, 557
    - Frequency tab, 558
  - Spectrum display, 98
  - Spectrum display settings, 100
  - Spectrum Emission Mask
    - display, 214
    - settings, 218
  - Spectrum Monitor
    - Spectrogram, 107
  - spectrum time, 666
  - Spectrum Time tab, 557
  - Spectrums/line, 104
  - Split
    - DPX display, 67
  - spur, 666
  - Spur
    - defining, 184
  - Spurious, 174
  - Spurious display settings, 178
  - spurious response, 666
  - Standard
    - Audio Filters tab, 477
  - Standard Settings button, 509
  - Standards
    - presets, 22
  - Standards presets
    - P25, 25
    - WLAN, 25
  - Start (Hz)
    - Ranges and Limits table, 183
  - Status Bar, 628
  - Step size
    - Freq & BW tab, 78
  - Stop
    - Replay menu, 564
  - Stop (Hz)
    - Ranges and Limits table, 183
  - Summary
    - WLAN, 37
  - Sweep time, 77
  - symbol maps, 526
  - Symbol Rate, 40
  - Symbol rate accuracy
    - P25, 41
  - Symbol Rate Error, 40
  - Symbol Table
    - WLAN, 37
  - Symbol Table display, 504
  - Symbol Table display
    - settings, 505
- ## T
- Taps
    - Equalizer tab, 516
  - Taps/Symbol
    - Equalizer tab, 516
  - Target Reference
    - Frequency Settling Time, 208
  - Test For, 548
  - Test Patterns tab
    - P25 Analysis displays, 373
  - Test using a mask, 549
  - THD
    - Audio Summary display, 472
  - Thrshld (dBm)
    - Ranges and Limits table, 183
  - Thunderstorm color scheme, 30
  - Time, 44
  - Time & Freq Scale tab
    - Spectrogram display, 106
  - Time 1 / Time 2
    - Time Qualified tab, 603
  - time measurement, 666
  - Time Overview, 91
  - Time Overview display
    - settings, 93
  - Time Params tab
    - Settling Time displays, 209
  - Time Qualified tab, 603
  - time reference, 666
  - time scale status readout, 104
  - TIQ acquisition data files, 30
  - TIQ acquisition data files
    - parameters, 29
  - TNHD
    - Audio Summary display, 472
  - TOFF
    - LTE, 43
  - Tolerance (+/-)
    - Frequency Settling Time, 208
  - Tools Menu, 632
  - Total AM
    - AM display, 40
  - Touch Screen, 13
  - Touch Screen Off, 9
  - Touch-screen Actions, 13

- trace, 666
  - Trace 2
    - Constellation display, 524
  - Trace motion for Sweep
    - ≥ 1 sec, 77
  - Trace tab
    - AM display, 132
    - Bluetooth, 461
    - FM display, 140
    - GP Digital Modulation displays, 523
    - LTE, 417
    - OFDM Analysis displays, 291
    - P25 Analysis, 374
    - PM display, 147
    - Settling Time displays, 210
    - Time Overview display, 96
    - WLAN Analysis displays, 264
  - Traces tab, 315
    - DPX Bitmap, 78
    - DPX display, 81
    - DPXogram view, 78
    - General Signal Viewing displays, 119
    - RF Measurements display, 227
    - Spectrogram display, 105
    - WLAN Channel Response display, 264
  - Traces tab - Math Trace
    - General Signal Viewing displays, 122
  - Traces Toolbar, 627
  - Transform Window
    - Spectrum display BW tab, 122
  - Transmitter frequency behavior
    - P25, 42
  - Transmitter power and encoder attack time
    - P25, 41
  - Transmitter power and encoder attack time with busy/idle
    - P25, 41
  - Transmitter throughput delay
    - P25, 42
  - Trellis Diagram display
    - settings, 507
  - Trig Meas tab
    - P25 Analysis displays, 376
  - Trig'd trigger status light, 587
  - Trigger On This, 588
  - Trigger settings, 593
    - Advanced tab, 604
    - Event tab, 594
  - trigger signals
    - connecting, 619
  - trigger status lights, 587
  - triggering
    - frequency mask trigger, 586
  - troubleshooting, 633
  - Tutorials
    - YouTube videos, 6
- U**
- Uncertainty Calculator
    - display, 305
  - Uncertainty Calculator
    - settings, 307
  - Uncertainty tab
    - Uncertainty Calculator
      - Display, 308
  - Uniform (None) Filter Shape
    - Spectrum display BW tab, 122
  - Units tab, 561
  - Unwanted emissions (ACPR)
    - P25, 41
  - Upgrades, 634
  - User Filters
    - file format, 533
    - loading, 532
  - User presets, 26
  - User Symbol Map, 519
- V**
- VBW, 170
    - LTE, 413
    - RF Measurements displays, 226
    - Spectrum display BW tab, 122
  - VBW (Hz)
    - Ranges and Limits table, 183
  - VBW On
    - Ranges and Limits table, 183
  - vector signal analyzer, 667
  - vertical scale factor, vertical display factor, 667
  - Video tutorials, 6
  - View
    - Basic Controls Toolbar, 628
    - Full Screen, 627
    - Markers Toolbar, 627
    - Navigator View, 627
    - Numeric keypad, 627
    - Replay Toolbar, 627
    - Status Bar, 628
    - Traces Toolbar, 627
  - View Menu, 627
  - violation, 667
  - violation (frequency mask trigger), 600
- W**
- watch icon, 603
  - Waterfall
    - Spectrogram/DPXogram, 106
    - Spectrogram/DPXogram display, 86
  - Window menu, 634
  - Windows Event Viewer, 643
  - WLAN
    - presets, 25
    - video tutorials, 6
  - WLAN Channel Response
    - display, 235
    - settings, 235
  - WLAN Constellation
    - display, 238
    - settings, 239
  - WLAN EVM
    - display, 240
    - settings, 241
  - WLAN Mag Error
    - display, 242
    - settings, 243
  - WLAN Phase Error
    - display, 245

WLAN Power vs Time  
  display, 246  
  settings, 248  
WLAN Spectral Flatness  
  display, 249  
  settings, 250  
WLAN Summary  
  display, 251  
  settings, 255  
WLAN Symbol Table  
  display, 256

settings, 258

## X

x dB BW Ref Power, 170  
x dB Level, 174

## Y

Y Factor display, 304  
Y Factor display settings, 304

YouTube  
  Tektronix channel, 6

## Z

Zero Span  
  DPX display, 67