
Read This First

Electrostatic discharge notice

The RSA7100A Real-Time Spectrum Analyzers may reboot if subject to electrostatic discharge. If this occurs, it will typically take 6 minutes to return to the previous operating state.

